

Certificate of Calibration

Customer: TMI - TAMPA

12530 TELECOM DRIVE

TEMPLE TERRACE, FL 33637

813-978-3054

P.O. Number: N/R

ID Number: L2720021



Description: TEMP/HUMIDITY METER

Manufacturer: VAISALA

Model Number: MI70/HMP76

Serial Number: L2720021 / L3020096

Technician: STEVE TORRES

On-Site Calibration: ..

Comments:

Calibration Date: 11/06/2018

Calibration Due: 03/06/2020

Procedure: TMI-M-HYGROTHERMOGRAPHS

Rev: 2/22/2011

Temperature: 71 F

Humidity: 42 % RH

As Found Condition: IN TOLERANCE

Calibration Results: IN TOLERANCE

Limiting Attribute:

This instrument has been calibrated using standards traceable to the SI units through the National Institute of Standards and Technology (NIST) or other National Metrological Institute (NMI). The method of calibration is direct comparison to a known standard, derived from natural physical constants, ratio measurements or compared to consensus standards.

Reported uncertainties are expressed as expanded uncertainty values at an approximately 95% confidence level using a coverage factor of k=2. Statements of compliance are based on test results falling within specified limits with no reduction by the uncertainty of the measurement.

TMI's Quality System is accredited to ISO/IEC 17025:2017 and ANSI/NCSL Z540-1-1994. ISO/IEC 17025:2017 is written in a language relevant to laboratory operations, meeting the principles of ISO 9001 and aligned with its pertinent requirements. This calibration is within the current Scope of Accreditation and complies with the requirements of ISO/IEC 17025:2017 and TMI's Quality Manual, QM-1.

Results contained in this document relate only to the item calibrated. Calibration due dates appearing on the certificate or label are determined by the client for administrative purposes and do not imply continued conformance to specifications.

This certificate shall not be reproduced, except in full, without the written permission of Technical Maintenance, Inc.

Measurements not currently on TMI's Scope of Accreditation are identified with an asterisk.

FRANK BAHMANN, BRANCH MANAGER

Scott Chamberlain, QUALITY MANAGER

Calibration Standards

<u>Asset Number</u>	<u>Manufacturer</u>	<u>Model Number</u>	<u>Date Calibrated</u>	<u>Cal Due</u>
05535	FLUKE	5609-12-D	7/3/2018	7/3/2019
0710649	THUNDER SCIENTIFIC	2500ST	11/2/2018	6/23/2019
A88072	FLUKE/HART	1502A	9/5/2018	12/17/2018



Technical Maintenance, Inc.

12530 TELECOM DRIVE, TEMPLE TERRACE, FL 33637

Phone: 813-978-3054 Fax 813-978-3758

www.tmicalibration.com



ISO/IEC 17025
ANSI/NCSL Z540-1
ANSI/NCSL Z540.3

AC-2080

Certificate of Calibration

Data Sheet

<u>Parameter</u>	<u>Nominal</u>	<u>Minimum</u>	<u>Maximum</u>	<u>As Found</u>	<u>As Left</u>	<u>Uncertainty</u>	<u>Unit</u> <u>ADJ/FAIL</u>
Temperature Accuracy, °C	10.00	9.80	10.20	10.05	10.05	0.026 C	°C
Temperature Accuracy, °C	25.00	24.80	25.20	24.86	24.86	0.026 C	°C
Temperature Accuracy, °C	40.00	39.80	40.20	39.84	39.84	0.026 C	°C
Humidity Accuracy, %RH, low range	33.0	32.0	34.0	32.1	32.1	0.51%	%RH
Humidity Accuracy, %RH, low range	50.0	49.0	51.0	50.9	50.9	0.51%	%RH
Humidity Accuracy, %RH, high range	75.0	74.0	76.0	75.9	75.9	0.51%	%RH



Technical Maintenance, Inc.

12530 TELECOM DRIVE, TEMPLE TERRACE, FL 33637

Phone: 813-978-3054 Fax 813-978-3758

www.tmicalibration.com



ISO/IEC 17025
ANSI/NCSL Z540-1
ANSI/NCSL Z540.3

AC-2080