



# CERTIFICATE OF ACCREDITATION

**The ANSI National Accreditation Board**

Hereby attests that

**Technical Maintenance, Inc.**  
12530 Telecom Drive  
Temple Terrace, FL 33637  
(and satellite sites as shown on the scope)

Fulfills the requirements of

**ISO/IEC 17025:2017**

and national standards

**ANSI/NCSL Z540-1-1994 (R2002) and  
ANSI/NCSL Z540.3-2006 (R2013)**

In the field (s) of

**CALIBRATION and DIMENSIONAL MEASUREMENT**

This certificate is valid only when accompanied by a current scope of accreditation document.  
The current scope of accreditation can be verified at [www.anab.org](http://www.anab.org).

Jason Stine, Vice President

Expiry Date: 20 September 2025

Certificate Number: AC-2080



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.  
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory  
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



**SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017**

**AND**

**ANSI/NCSL Z540-1-1994 (R2002)**

**ANSI/NCSL Z540.3-2006 (R2013)**

**Technical Maintenance, Inc.**

12530 Telecom Drive

Temple Terrace, FL 33637

Wally Gynn (Branch Manager) Phone: 813-978-3054

Scott Chamberlain (Quality Manager) Phone: 321-242-0890

**CALIBRATION AND DIMENSIONAL MEASUREMENT**

Valid to: **September 20, 2025**

Certificate Number: **AC-2080**

**Satellite locations in:**

[Peachtree Corners, GA](#)

[Madison, AL](#)

[Rockford, IL](#)

[Windsor, CT](#)

[Hilliard, OH](#)

[Fort Lauderdale, FL](#)

[Melbourne, FL](#)

[White, GA](#)

[Colorado Springs, CO](#)

[Morrisville, NC](#)



ANSI National Accreditation Board

**Technical Maintenance, Inc.**

12530 Telecom Drive

Temple Terrace, FL 33637

Wally Gynn (Branch Manager) Phone: 813-978-3054

Scott Chamberlain (Quality Manager) Phone: 321-242-0890

**CALIBRATION AND DIMENSIONAL MEASUREMENT**

**CALIBRATION**

**Acoustics and Vibration**

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Accelerometers – Acceleration (7 < 10) Hz (10 < 30) Hz (30 < 2000) Hz (2 to 10) kHz	(0.01 to 10) g	2 % of reading 1.5 % of reading 1 % of reading 2.5 % of reading	Accelerometer Calibrator

**Chemical Quantities**

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
pH meters <sup>1,5</sup>	4 pH 7 pH 10 pH	0.018 pH 0.036 pH 0.069 pH	pH buffer solutions
Conductivity Meters <sup>1,5</sup>	2 µS/cm 100 µS/cm 1 000 µS/cm 10 000 µS/cm 100 000 µS/cm 150 000 µS/cm	0.2 µS/cm 0.74 µS/cm 3.6 µS/cm 35 µS/cm 350 µS/cm 580 µS/cm	Conductivity solutions

**Electrical – DC/Low Frequency**

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Phase Angle – Generate 5 V Equal Amplitude	(0 to 360)° 1 Hz to 1 kHz (1 to 6.25) kHz (6.25 to 50) kHz (50 to 200) kHz	0.005° 0.01° 0.015° 0.04°	Clarke-Hess 5500-2 Phase Standard
Phase Angle – Generate 50 mV to 120 V– Unequal Amplitude	(0 to 360)° 1 Hz to 1 kHz (1 to 6.25) kHz (6.25 to 50) kHz (50 to 200) kHz	0.005° 0.01° 0.015° 0.04°	Clarke-Hess 5500-2 Phase Standard
Phase Angle – Measure 10 mV to 630 V	(0 to 360)° 5 Hz to 2 kHz (2 to 5) kHz (5 to 10) kHz (10 to 50) kHz (50 to 100) kHz (100 to 200) kHz	0.02° 0.03° 0.04° 0.05° 0.1° 0.2°	Clarke-Hess 6000A Phase Meter
Phase Angle – Measure 50 mV to 120 V	(0 to 360) ° 1 Hz to 1 kHz (1 to 50) kHz	0.001° 0.002°	Clarke-Hess 5002A Phase Meter
	(1 to 50) kHz (50 to 200) kHz	0.002° 0.008°	Clarke-Hess 5002B Phase Meter
	(5 to 50) kHz (50 to 200) kHz	0.002° 0.008°	Clarke-Hess 5002C Phase Meter
	(5 to 50) kHz (50 to 200) kHz	0.002° 0.008°	Clarke-Hess 5002D Phase Meter
DC Current – Generate <sup>1</sup>	(1 to 2.2) nA (2.2 to 22) nA (22 to 220) nA (0.22 to 2.2) μA (2.2 to 10) μA	93 μA/A + 0.007 nA 92 μA/A + 0.007 nA 92 μA/A + 0.01 nA 36 μA/A + 0.1 nA 15 μA/A + 1 nA	Fluke 5730A Multifunction Calibrator & Fluke 5522A Multiproduct Calibrator
DC Current – Generate <sup>1</sup>	(10 to 220) μA 220 μA to 2.2 mA (2.2 to 22) mA (22 to 100) mA (100 to 220) mA (0.22 to 1) A (1 to 2.2) A	38 μA/A + 5 nA 30 μA/A + 7 nA 30 μA/A + 44 nA 38 μA/A + 0.7 μA 45 μA/A + 0.7 μA 68 μA/A + 12 μA 105 μA/A + 12 μA	Fluke 5730A Multifunction Calibrator

**Electrical – DC/Low Frequency**

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Current – Generate <sup>1</sup>	(2.2 to 11) A	274 $\mu$ A/A + 365 $\mu$ A	Fluke 5730A Multifunction Calibrator /5725A Amplifier
DC Current – Generate <sup>1</sup>	(11 to 20.5) A (20 to 100) A	77 $\mu$ A/A + 761 $\mu$ A 77 $\mu$ A/A + 3.8 mA	Fluke 5522A Multiproduct Calibrator /52120A Transconductance Amplifier
DC Current Clamp Meters Toroidal-Wound	(0 to 1 025) A	0.21 % of output + 0.05A	Fluke 5522A Multiproduct Calibrator /Coil5500A x50
DC Current Clamp Meters Other	(0 to 1 025) A	0.39 % of output + 0.38A	Fluke 5522A Multiproduct Calibrator /Coil5500A x50
DC Current – Measure <sup>1</sup>	(1 to 10) nA (10 to 100) nA (0.1 to 1) $\mu$ A (1 to 10) $\mu$ A (10 to 100) $\mu$ A (0.1 to 1) mA (1 to 10) mA (10 to 100) mA	19 $\mu$ A/A + 0.16 pA 7.1 $\mu$ A/A + 1.6 pA 6.1 $\mu$ A/A + 0.016 nA 5.3 $\mu$ A/A + 0.016 nA 5.3 $\mu$ A/A + 0.16 nA 5.3 $\mu$ A/A + 0.001 6 $\mu$ A 5.3 $\mu$ A/A + 0.016 $\mu$ A 10 $\mu$ A/A + 0.16 $\mu$ A	Agilent 3458A Multimeter Option 002, 5450A Resistance Calibrator
DC Current – Measure <sup>1</sup>	(10 to 100) $\mu$ A (0.1 to 1) mA (1 to 10) mA (10 to 100) mA (0.1 to 1) A	20 $\mu$ A/A + 0.8 nA 20 $\mu$ A/A + 5 nA 20 $\mu$ A/A + 0.05 $\mu$ A 35 $\mu$ A/A + 0.5 $\mu$ A 110 $\mu$ A/A + 10 $\mu$ A	Agilent 3458A Multimeter
DC Current – Measure <sup>1</sup>	(1 to 2.0) A (2 to 20) A (20 to 25) A	36.5 $\mu$ A/A + 21 $\mu$ A 33.7 $\mu$ A/A + 21 $\mu$ A 28.6 $\mu$ A/A + 35 $\mu$ A	Agilent 3458A Multimeter, Guildline 9230A-50 Current Shunts
DC Current – Measure <sup>1</sup>	(25 to 100) A	0.25 % of reading + 1 mA	Agilent 3458A Multimeter Option 002, Empro 250A Current Shunt
DC Current – Measure <sup>1</sup>	(100 to 1 000) A	0.25 % of reading + 8 mA	Agilent 3458A Multimeter Option 002, Empro 1000A Current Shunt
DC Voltage – Generate <sup>1</sup>	(0 to 220) mV 220 mV to 2.2 V (2.2 to 11) V (11 to 22) V (22 to 220) V (220 to 1 100) V	6.8 $\mu$ V/V + 0.8 $\mu$ V 4.6 $\mu$ V/V + 0.9 $\mu$ V 3 $\mu$ V/V + 2.5 $\mu$ V 3 $\mu$ V/V + 3.9 $\mu$ V 4.6 $\mu$ V/V + 38 $\mu$ V 6.1 $\mu$ V/V + 385 $\mu$ V	Fluke 5730A Multifunction Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Voltage – Measure <sup>1</sup>	(0 to 100) mV 100 mV to 1 V (1 to 10) V (10 to 100) V (100 to 1 000) V	5.0 $\mu\text{V/V} + 1.2 \mu\text{V}$ 4.0 $\mu\text{V/V} + 1.2 \mu\text{V}$ 4.0 $\mu\text{V/V} + 2.1 \mu\text{V}$ 6.0 $\mu\text{V/V} + 30 \mu\text{V}$ 6 $\mu\text{V/V} + 100 \mu\text{V} + 12 \mu\text{V/V} \times (\text{Vin}/1\ 000)^2$	Agilent 3458A Option 002 Multimeter
DC Voltage – Measure <sup>1</sup>	(1 to 60) kV	0.035 % of reading	Ross VD60 High Voltage Divider, HP 3458A Multimeter
DC Voltage – Measure <sup>1</sup>	(12 to 120) kV	0.018 % of reading	Ross VD120 High Voltage Divider, HP 3458A Multimeter
DC Voltage – Measure <sup>1</sup> Fixed Points	100 mV 1 V 10 V 100 V 1 000 V	3.1 $\mu\text{V/V}$ 0.46 $\mu\text{V/V}$ 0.3 $\mu\text{V/V}$ 0.34 $\mu\text{V/V}$ 0.48 $\mu\text{V/V}$	Fluke 732B Reference Standard, 752A Divider
AC Voltage – Generate <sup>1</sup>	(0.22 to 2.2) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz (2.2 to 22) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	228 $\mu\text{V/V} + 3.9 \mu\text{V}$ 88 $\mu\text{V/V} + 3.9 \mu\text{V}$ 76 $\mu\text{V/V} + 3.9 \mu\text{V}$ 190 $\mu\text{V/V} + 3.9 \mu\text{V}$ 457 $\mu\text{V/V} + 4.6 \mu\text{V}$ 989 $\mu\text{V/V} + 9.2 \mu\text{V}$ 1.3 mV/V + 19 $\mu\text{V}$ 2.6 mV/V + 19 $\mu\text{V}$ 228 $\mu\text{V/V} + 3.9 \mu\text{V}$ 88 $\mu\text{V/V} + 3.9 \mu\text{V}$ 76 $\mu\text{V/V} + 3.9 \mu\text{V}$ 190 $\mu\text{V/V} + 3.9 \mu\text{V}$ 457 $\mu\text{V/V} + 4.6 \mu\text{V}$ 989 $\mu\text{V/V} + 9.2 \mu\text{V}$ 1.3 mV/V + 19 $\mu\text{V}$ 2.6 mV/V + 19 $\mu\text{V}$	Fluke 5730A Multifunction Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Generate <sup>1</sup>	(22 to 220) mV		Fluke 5730A Multifunction Calibrator
	(10 to 20) Hz	228 $\mu\text{V/V} + 11.4 \mu\text{V}$	
	(20 to 40) Hz	88 $\mu\text{V/V} + 6.1 \mu\text{V}$	
	40 Hz to 20 kHz	53 $\mu\text{V/V} + 6.1 \mu\text{V}$	
	(20 to 50) kHz	114 $\mu\text{V/V} + 6.1 \mu\text{V}$	
	(50 to 100) kHz	304 $\mu\text{V/V} + 15.2 \mu\text{V}$	
	(100 to 300) kHz	609 $\mu\text{V/V} + 19 \mu\text{V}$	
	(300 to 500) kHz	1.3 mV/V + 23 $\mu\text{V}$	
	500 kHz to 1 MHz	2.5 mV/V + 46 $\mu\text{V}$	
	220 mV to 2.2 V		
	(10 to 20) Hz	228 $\mu\text{V/V} + 38 \mu\text{V}$	
	(20 to 40) Hz	84 $\mu\text{V/V} + 15 \mu\text{V}$	
	40 Hz to 20 kHz	37 $\mu\text{V/V} + 8 \mu\text{V}$	
	(20 to 50) kHz	61 $\mu\text{V/V} + 9 \mu\text{V}$	
	(50 to 100) kHz	76 $\mu\text{V/V} + 30 \mu\text{V}$	
	(100 to 300) kHz	304 $\mu\text{V/V} + 76 \mu\text{V}$	
	(300 to 500) kHz	913 $\mu\text{V/V} + 190 \mu\text{V}$	
	500 kHz to 1 MHz	1.5 mV/V + 304 $\mu\text{V}$	
	(2.2 to 22) V		
	(10 to 20) Hz	228 $\mu\text{V/V} + 380 \mu\text{V}$	
	(20 to 40) Hz	84 $\mu\text{V/V} + 152 \mu\text{V}$	
	40 Hz to 20 kHz	37 $\mu\text{V/V} + 54 \mu\text{V}$	
	(20 to 50) kHz	61 $\mu\text{V/V} + 91 \mu\text{V}$	
	(50 to 100) kHz	76 $\mu\text{V/V} + 190 \mu\text{V}$	
	(100 to 300) kHz	228 $\mu\text{V/V} + 609 \mu\text{V}$	
	(300 to 500) kHz	913 $\mu\text{V/V} + 1.9 \text{ mV}$	
	500 kHz to 1 MHz	1.4 mV/V + 3 mV	
(22 to 220) V			
(10 to 20) Hz	228 $\mu\text{V/V} + 3.8 \text{ mV}$		
(20 to 40) Hz	84 $\mu\text{V/V} + 1.5 \text{ mV}$		
40 Hz to 20 kHz	49 $\mu\text{V/V} + 0.6 \text{ mV}$		
(20 to 50) kHz	76 $\mu\text{V/V} + 0.9 \text{ mV}$		
(50 to 100) kHz	137 $\mu\text{V/V} + 2.3 \text{ mV}$		
(220 to 750) V			
40 Hz to 1 kHz	68 $\mu\text{V/V} + 3 \text{ mV}$		
(1 to 20) kHz	126 $\mu\text{V/V} + 5 \text{ mV}$		
(20 to 50) kHz	457 $\mu\text{V/V} + 8 \text{ mV}$		
(50 to 100) kHz	1.8 mV/V + 34 mV		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Generate <sup>1</sup>	(750 to 1 000) V 40 Hz to 1 kHz (1 to 20) kHz (20 to 30) kHz	68 $\mu$ V/V + 3 mV 126 $\mu$ V/V + 5 mV 457 $\mu$ V/V + 8 mV	Fluke 5730A Multifunction Calibrator /5725A Amplifier
AC Voltage – Generate <sup>1</sup> Wideband Absolute	(0.3 to 1.1) mV (10 to 30 Hz) 30 Hz to 500 kHz (0.5 to 1.2) MHz (1.2 to 2) MHz (2 to 12) MHz (12 to 20) MHz (20 to 30) MHz (1.1 to 3.3) mV (10 to 30 Hz) 30 Hz to 500 kHz (0.5 to 1.2) MHz (1.2 to 2) MHz (2 to 12) MHz (12 to 20) MHz (20 to 30) MHz (3.3 to 11) mV (10 to 30 Hz) 30 Hz to 500 kHz (0.5 to 1.2) MHz (1.2 to 2) MHz (2 to 12) MHz (12 to 20) MHz (20 to 30) MHz (11 to 33) mV (10 to 30 Hz) 30 Hz to 500 kHz (0.5 to 1.2) MHz (1.2 to 2) MHz (2 to 12) MHz (12 to 20) MHz (20 to 30) MHz	0.65 % of output + 1.5 $\mu$ V 0.61 % of output + 1.5 $\mu$ V 0.63 % of output + 3.8 $\mu$ V 0.63 % of output + 3.8 $\mu$ V 0.68 % of output + 3.8 $\mu$ V 0.76 % of output + 3.8 $\mu$ V 1.3 % of output + 13 $\mu$ V 0.58 % of output + 2.3 $\mu$ V 0.53 % of output + 2.3 $\mu$ V 0.54 % of output + 4.6 $\mu$ V 0.54 % of output + 4.6 $\mu$ V 0.58 % of output + 4.6 $\mu$ V 0.65 % of output + 4.6 $\mu$ V 1.3 % of output + 4.6 $\mu$ V 0.58 % of output + 6.1 $\mu$ V 0.53 % of output + 6.1 $\mu$ V 0.54 % of output + 8.4 $\mu$ V 0.54 % of output + 8.4 $\mu$ V 0.55 % of output + 8.4 $\mu$ V 0.61 % of output + 8.4 $\mu$ V 0.93 % of output + 8.4 $\mu$ V 0.52 % of output + 12 $\mu$ V 0.46 % of output + 12 $\mu$ V 0.47 % of output + 14 $\mu$ V 0.47 % of output + 14 $\mu$ V 0.49 % of output + 14 $\mu$ V 0.55 % of output + 14 $\mu$ V 0.89 % of output + 14 $\mu$ V	Fluke 5730A Multifunction Calibrator /5725A Amplifier



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment		
AC Voltage – Generate <sup>1</sup> Wideband Absolute	(33 to 110) mV (10 to 30 Hz)	0.52 % of output + 30 μV	Fluke 5730A Multifunction Calibrator		
	30 Hz to 500 kHz (0.5 to 1.2) MHz	0.46 % of output + 30 μV 0.4 % of output + 33 μV			
	(1.2 to 2) MHz (2 to 12) MHz	0.4 % of output + 33 μV 0.42 % of output + 33 μV			
	(12 to 20) MHz (20 to 30) MHz	0.55 % of output + 33 μV 0.89 % of output + 33 μV			
	(110 to 330) mV (10 to 30 Hz)	0.45 % of output + 0.1 mV			
	30 Hz to 500 kHz (0.5 to 1.2) MHz	0.38 % of output + 0.1 mV 0.4 % of output + 0.1 mV			
	(1.2 to 2) MHz (2 to 12) MHz	0.4 % of output + 0.1 mV 0.42 % of output + 0.1 mV			
	(12 to 20) MHz (20 to 30) MHz	0.49 % of output + 0.1 mV 0.85 % of output + 0.1 mV			
	(0.33 to 1.1) V (10 to 30 Hz)	0.45 % of output + 0.3 mV			
	30 Hz to 500 kHz (0.5 to 1.2) MHz	0.38 % of output + 0.3 mV 0.4 % of output + 0.3 mV			
	(1.2 to 2) MHz (2 to 12) MHz	0.4 % of output + 0.3 mV 0.42 % of output + 0.3 mV			
	(12 to 20) MHz (20 to 30) MHz	0.49 % of output + 0.3 mV 0.85 % of output + 0.3 mV			
	(1.1 to 3.5) V (10 to 30 Hz)	0.39 % of output + 0.4 mV			
	30 Hz to 500 kHz (0.5 to 1.2) MHz	0.3 % of output + 0.4 mV 0.32 % of output + 0.4 mV			
	(1.2 to 2) MHz (2 to 12) MHz	0.32 % of output + 0.4 mV 0.35 % of output + 0.4 mV			
	(12 to 20) MHz (20 to 30) MHz	0.44 % of output + 0.4 mV 0.82 % of output + 0.4 mV			
	AC Voltage – Measure <sup>1</sup>	Up to 2.2 mV (10 to 20) Hz		1.3 mV/V + 1.2 μV	Fluke 5790B AC/DC Transfer Standard
		(20 to 40) Hz		563 μV/V + 1.2 μV	
		40 Hz to 20 kHz (20 to 50) kHz		320 μV/V + 1.2 μV 616 μV/V + 1.7 μV	
		(50 to 100) kHz (100 to 300) kHz		913 μV/V + 2 μV 1.8 mV/V + 3.1 μV	
(300 to 500) kHz 500 kHz to 1 MHz		1.8 mV/V + 6.1 μV 2.7 mV/V + 6.1 μV			



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure <sup>1</sup>	(2.2 to 7) mV		Fluke 5790B AC/DC Transfer Standard
	(10 to 20) Hz	647 $\mu\text{V/V} + 1.2 \mu\text{V}$	
	(20 to 40) Hz	282 $\mu\text{V/V} + 1.2 \mu\text{V}$	
	40 Hz to 20 kHz	160 $\mu\text{V/V} + 1.2 \mu\text{V}$	
	(20 to 50) kHz	304 $\mu\text{V/V} + 1.7 \mu\text{V}$	
	(50 to 100) kHz	457 $\mu\text{V/V} + 2 \mu\text{V}$	
	(100 to 300) kHz	913 $\mu\text{V/V} + 3.1 \mu\text{V}$	
	(300 to 500) kHz	989 $\mu\text{V/V} + 6.1 \mu\text{V}$	
	500 kHz to 1 MHz	1.8 mV/V + 6.1 $\mu\text{V}$	
	(7 to 22) mV		
	(10 to 20) Hz	221 $\mu\text{V/V} + 1.2 \mu\text{V}$	
	(20 to 40) Hz	145 $\mu\text{V/V} + 1.2 \mu\text{V}$	
	40 Hz to 20 kHz	84 $\mu\text{V/V} + 1.2 \mu\text{V}$	
	(20 to 50) kHz	160 $\mu\text{V/V} + 1.7 \mu\text{V}$	
	(50 to 100) kHz	236 $\mu\text{V/V} + 2 \mu\text{V}$	
	(100 to 300) kHz	616 $\mu\text{V/V} + 3.1 \mu\text{V}$	
	(300 to 500) kHz	677 $\mu\text{V/V} + 6.1 \mu\text{V}$	
	500 kHz to 1 MHz	1.3 mV/V + 6.1 $\mu\text{V}$	
	(22 to 70) mV		
	(10 to 20) Hz	183 $\mu\text{V/V} + 1.3 \mu\text{V}$	
	(20 to 40) Hz	91 $\mu\text{V/V} + 1.3 \mu\text{V}$	
	40 Hz to 20 kHz	49 $\mu\text{V/V} + 1.3 \mu\text{V}$	
	(20 to 50) kHz	99 $\mu\text{V/V} + 1.7 \mu\text{V}$	
	(50 to 100) kHz	198 $\mu\text{V/V} + 2 \mu\text{V}$	
	(100 to 300) kHz	388 $\mu\text{V/V} + 3.1 \mu\text{V}$	
	(300 to 500) kHz	510 $\mu\text{V/V} + 6.1 \mu\text{V}$	
	500 kHz to 1 MHz	837 $\mu\text{V/V} + 6.1 \mu\text{V}$	
	(70 to 220) mV		
	(10 to 20) Hz	160 $\mu\text{V/V} + 1.3 \mu\text{V}$	
	(20 to 40) Hz	65 $\mu\text{V/V} + 1.3 \mu\text{V}$	
40 Hz to 20 kHz	29 $\mu\text{V/V} + 1.3 \mu\text{V}$		
(20 to 50) kHz	53 $\mu\text{V/V} + 1.7 \mu\text{V}$		
(50 to 100) kHz	122 $\mu\text{V/V} + 2 \mu\text{V}$		
(100 to 300) kHz	190 $\mu\text{V/V} + 3.1 \mu\text{V}$		
(300 to 500) kHz	289 $\mu\text{V/V} + 6.1 \mu\text{V}$		
500 kHz to 1 MHz	761 $\mu\text{V/V} + 6.1 \mu\text{V}$		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure <sup>1</sup>	(220 to 700) mV		Fluke 5790B AC/DC Transfer Standard
	(10 to 20) Hz	160 $\mu\text{V/V} + 1.3 \mu\text{V}$	
	(20 to 40) Hz	58 $\mu\text{V/V} + 1.3 \mu\text{V}$	
	40 Hz to 20 kHz	25 $\mu\text{V/V} + 1.3 \mu\text{V}$	
	(20 to 50) kHz	39 $\mu\text{V/V} + 1.7 \mu\text{V}$	
	(50 to 100) kHz	60 $\mu\text{V/V} + 2 \mu\text{V}$	
	(100 to 300) kHz	137 $\mu\text{V/V} + 3.1 \mu\text{V}$	
	(300 to 500) kHz	228 $\mu\text{V/V} + 6.1 \mu\text{V}$	
	500 kHz to 1 MHz	730 $\mu\text{V/V} + 6.1 \mu\text{V}$	
	700 mV to 2.2 V		
	(10 to 20) Hz	152 $\mu\text{V/V} + 0.7 \mu\text{V}$	
	(20 to 40) Hz	50 $\mu\text{V/V} + 0.7 \mu\text{V}$	
	40 Hz to 20 kHz	18 $\mu\text{V/V} + 0.7 \mu\text{V}$	
	(20 to 50) kHz	35 $\mu\text{V/V} + 0.7 \mu\text{V}$	
	(50 to 100) kHz	54 $\mu\text{V/V} + 0.7 \mu\text{V}$	
	(100 to 300) kHz	122 $\mu\text{V/V} + 0.7 \mu\text{V}$	
	(300 to 500) kHz	198 $\mu\text{V/V} + 0.7 \mu\text{V}$	
	500 kHz to 1 MHz	685 $\mu\text{V/V} + 0.7 \mu\text{V}$	
	(2.2 to 7) V		
	(10 to 20) Hz	152 $\mu\text{V/V} + 0.9 \mu\text{V}$	
	(20 to 40) Hz	51 $\mu\text{V/V} + 0.9 \mu\text{V}$	
	40 Hz to 20 kHz	18 $\mu\text{V/V} + 0.9 \mu\text{V}$	
	(20 to 50) kHz	37 $\mu\text{V/V} + 0.9 \mu\text{V}$	
	(50 to 100) kHz	62 $\mu\text{V/V} + 0.9 \mu\text{V}$	
	(100 to 300) kHz	145 $\mu\text{V/V} + 0.9 \mu\text{V}$	
	(300 to 500) kHz	304 $\mu\text{V/V} + 0.9 \mu\text{V}$	
	500 kHz to 1 MHz	913 $\mu\text{V/V} + 0.9 \mu\text{V}$	
	(7 to 22) V		
(10 to 20) Hz	152 $\mu\text{V/V} + 0.9 \mu\text{V}$		
(20 to 40) Hz	51 $\mu\text{V/V} + 0.9 \mu\text{V}$		
40 Hz to 20 kHz	21 $\mu\text{V/V} + 0.9 \mu\text{V}$		
(20 to 50) kHz	37 $\mu\text{V/V} + 0.9 \mu\text{V}$		
(50 to 100) kHz	62 $\mu\text{V/V} + 0.9 \mu\text{V}$		
(100 to 300) kHz	145 $\mu\text{V/V} + 0.9 \mu\text{V}$		
(300 to 500) kHz	304 $\mu\text{V/V} + 0.9 \mu\text{V}$		
500 kHz to 1 MHz	913 $\mu\text{V/V} + 0.9 \mu\text{V}$		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure <sup>1</sup>	(22 to 70) V		Fluke 5790B AC/DC Transfer Standard
	(10 to 40) Hz	152 $\mu\text{V/V} + 5.7 \mu\text{V}$	
	40 Hz to 20 kHz	52 $\mu\text{V/V} + 5.7 \mu\text{V}$	
	(20 to 50) kHz	24 $\mu\text{V/V} + 5.7 \mu\text{V}$	
	(50 to 100) kHz	43 $\mu\text{V/V} + 5.7 \mu\text{V}$	
	(100 to 300) kHz	72 $\mu\text{V/V} + 5.7 \mu\text{V}$	
	(300 to 500) kHz	152 $\mu\text{V/V} + 5.7 \mu\text{V}$	
	(70 to 220) V		
	(10 to 20) Hz	152 $\mu\text{V/V} + 5.7 \mu\text{V}$	
	(20 to 40) Hz	52 $\mu\text{V/V} + 5.7 \mu\text{V}$	
	40 Hz to 20 kHz	24 $\mu\text{V/V} + 5.7 \mu\text{V}$	
	(20 to 50) kHz	53 $\mu\text{V/V} + 5.7 \mu\text{V}$	
	(50 to 100) kHz	75 $\mu\text{V/V} + 5.7 \mu\text{V}$	
	(220 to 700) V		
40 Hz to 20 kHz	31 $\mu\text{V/V} + 57 \mu\text{V}$		
(20 to 50) kHz	99 $\mu\text{V/V} + 57 \mu\text{V}$		
(50 to 100) kHz	380 $\mu\text{V/V} + 57 \mu\text{V}$		
(700 to 1 000) V			
40 Hz to 20 kHz	29 $\mu\text{V/V} + 57 \mu\text{V}$		
(20 to 30) kHz	99 $\mu\text{V/V} + 57 \mu\text{V}$		
AC Voltage – Measure <sup>1</sup>	60 Hz (1 to 10) kV (10 to 42) kV	0.5 % of reading + 0.002 kV 0.5 % of reading + 0.02 kV	Ross VD60 High Voltage Divider, HP 34401A Multimeter
AC Voltage – Measure <sup>1</sup>	60 Hz (1 to 10) kV (10 to 85) kV	0.5 % of reading + 0.002 kV 0.5 % of reading + 0.02 kV	Ross VD120 High Voltage Divider, HP 34401A Multimeter
AC Current – Generate <sup>1</sup>	(9 to 220) $\mu\text{A}$		Fluke 5730A/03 Multifunction Calibrator
	(10 to 20) Hz	228 $\mu\text{A/A} + 15 \text{ nA}$	
	(20 to 40) Hz	152 $\mu\text{A/A} + 10 \text{ nA}$	
	40 Hz to 1 kHz	91 $\mu\text{A/A} + 8 \text{ nA}$	
	(1 to 5) kHz	266 $\mu\text{A/A} + 12 \text{ nA}$	
	(5 to 10) kHz	989 $\mu\text{A/A} + 61 \text{ nA}$	
	(0.22 to 2.2) mA		
	(10 to 20) Hz	228 $\mu\text{A/A} + 39 \text{ nA}$	
	(20 to 40) Hz	152 $\mu\text{A/A} + 31 \text{ nA}$	
	40 Hz to 1 kHz	91 $\mu\text{A/A} + 31 \text{ nA}$	
	(1 to 5) kHz	183 $\mu\text{A/A} + 99 \text{ nA}$	
	(5 to 10) kHz	989 $\mu\text{A/A} + 609 \text{ nA}$	

**Electrical – DC/Low Frequency**

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Generate <sup>1</sup>	(2.2 to 22) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	228 $\mu$ A/A + 385 nA 152 $\mu$ A/A + 310 nA 91 $\mu$ A/A + 310 nA 183 $\mu$ A/A + 536 nA 989 $\mu$ A/A + 4.6 $\mu$ A	Fluke 5730A/03 Multifunction Calibrator
	(22 to 220) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	228 $\mu$ A/A + 4 $\mu$ A 152 $\mu$ A/A + 3 $\mu$ A 91 $\mu$ A/A + 2 $\mu$ A 183 $\mu$ A/A + 3 $\mu$ A 989 $\mu$ A/A + 9 $\mu$ A	
AC Current – Generate <sup>1</sup>	(0.22 to 2.2) A 20 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	228 $\mu$ A/A + 31 $\mu$ A 380 $\mu$ A/A + 76 $\mu$ A 6.1 mA/A + 152 $\mu$ A	Fluke 5730A Multifunction Calibrator
AC Current – Generate <sup>1</sup>	(2.2 to 11) A 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	350 $\mu$ A/A + 141 $\mu$ A 723 $\mu$ A/A + 295 $\mu$ A 2.7 mA/A + 573 $\mu$ A	Fluke 5730A Multifunction Calibrator /5725A Amplifier
AC Current – Generate <sup>1</sup>	(11 to 20.5) A (45 to 100) Hz 100 Hz to 1 kHz (1 to 5) kHz	0.09 % of output + 5 mA 0.11 % of output + 5 mA 2.3 % of output + 5 mA	Fluke 5522A Multiproduct Calibrator
AC Current – Generate <sup>1</sup>	(11 to 20) A (10 to 300) Hz 300 Hz to 1 kHz (1 to 3) kHz (3 to 6) kHz (6 to 10) kHz	255 $\mu$ A/A + 10 mA 769 $\mu$ A/A + 11 mA 0.23 % of output + 31 mA 0.76 % of output + 62 mA 2.3 % of output + 92 mA	Fluke 5522A Multiproduct Calibrator / 52120A Transconductance Amplifier
AC Current – Generate <sup>1</sup>	(20 to 120) A (10 to 300) Hz 300 Hz to 1 kHz (1 to 3) kHz (3 to 6) kHz (6 to 10) kHz	255 $\mu$ A/A + 28 mA 769 $\mu$ A/A + 92 mA 0.23 % of output + 228 mA 0.76 % of output + 411 mA 3 % of output + 685 mA	Fluke 5522A Multiproduct Calibrator / 52120A Transconductance Amplifier
AC Current Clamps – Toroidal-Wound	(16.5 to 1 025) A (45 to 65) Hz (16.5 to 150) A (65 to 440) Hz	0.23 % of output + 0.069 A 0.64 % of output + 0.076 A	Fluke 5522A Multiproduct Calibrator /x50 Coil

**Electrical – DC/Low Frequency**

Temple Terrace, FL

<b>Parameter/Equipment</b>	<b>Range</b>	<b>Expanded Uncertainty of Measurement (+/-)</b>	<b>Reference Standard, Method, and/or Equipment</b>
AC Current Clamps – Other	(16.5 to 1 025) A (45 to 65) Hz (16.5 to 150) A (65 to 440) Hz	0.44 % of output +0.53 A 0.79 % of output +0.69 A	Fluke 5522A Multiproduct Calibrator /x50 Coil
AC Current Clamps – Rogowski	(10 to 100) A (10 to 1 000) Hz (1 to 3) kHz (100 to 1000) A (10 to 1 000) Hz (1 to 3) kHz	0.53 % of output +0.011 A 0.61 % of output +0.015 A 0.53 % of output +0.11 A 0.61 % of output +0.15 A	Fluke 52120 Transconductance Amplifier, Coil 6KA Current Coil
AC Current Clamps – Rogowski	(1 000 to 6 000) A (10 to 1 000) Hz (1 000 to 3 500) A (1 to 3) kHz	0.53 % of output +0.69 A 0.61 % of output +0.95 A	Fluke 52120 Transconductance Amplifier, Coil 6KA Current Coil
AC Current – Measure <sup>1</sup>	(2 to 200) $\mu$ A 10 Hz to 10 kHz 200 $\mu$ A to 2 mA 10 Hz to 10 kHz	0.047 % of reading + 0.018 $\mu$ A 0.028 % of reading + 0.18 $\mu$ A	Fluke 8508A Multimeter
AC Current – Measure <sup>1</sup>	(0.19 to 1) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz	306 $\mu$ A/A 306 $\mu$ A/A 259 $\mu$ A/A 419 $\mu$ A/A	Agilent 3458A Multimeter / Fluke A40 Current Shunts



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Measure <sup>1</sup>	(1 to 10) mA		Fluke 5790B AC/DC Transfer Standard / Holt HCS-1 Current Shunts
	(10 to 20) Hz	163 $\mu$ A/A	
	(20 to 40) Hz	70 $\mu$ A/A	
	40 Hz to 20 kHz	40 $\mu$ A/A	
	(20 to 50) kHz	65 $\mu$ A/A	
	(10 to 50) mA		
	(10 to 20) Hz	165 $\mu$ A/A	
	(20 to 40) Hz	69 $\mu$ A/A	
	40 Hz to 20 kHz	39 $\mu$ A/A	
	(20 to 50) kHz	66 $\mu$ A/A	
	(50 to 500) mA		
	(10 to 20) Hz	167 $\mu$ A/A	
	(20 to 40) Hz	70 $\mu$ A/A	
	40 Hz to 20 kHz	40 $\mu$ A/A	
(20 to 50) kHz	72 $\mu$ A/A		
AC Current – Measure <sup>1</sup>	(0.5 to 2) A		Fluke 5790B AC/DC Transfer Standard / Holt HCS-1 Current Shunts
	(10 to 20) Hz	170 $\mu$ A/A	
	(20 to 40) Hz	77 $\mu$ A/A	
	40 Hz to 20 kHz	62 $\mu$ A/A	
AC Current – Measure <sup>1</sup>	(2 to 10) A		Fluke 5790B AC/DC Transfer Standard / Holt HCS-1 Current Shunts
	(10 to 20) Hz	181 $\mu$ A/A	
	(20 to 40) Hz	94 $\mu$ A/A	
	40 Hz to 20 kHz	79 $\mu$ A/A	
AC Current – Measure <sup>1</sup>	(20 to 50) kHz	130 $\mu$ A/A	Agilent 3458A Multimeter / Fluke Y5020 Current Shunt
	(10 to 20) A		
	(10 to 20) Hz	368 $\mu$ A/A	
	(20 to 40) Hz	368 $\mu$ A/A	
AC Current – Measure <sup>1</sup>	40 Hz to 1 kHz	302 $\mu$ A/A	Fluke 376 Clamp Meter
	(1 to 5) kHz	424 $\mu$ A/A	
	(100 to 1 000) A		
AC Current – Measure <sup>1</sup>	(10 to 100) Hz	1.6 % of reading + 0.5A	Fluke 376 Clamp Meter
	(100 to 500) Hz	1.9 % of reading+ 0.5A	
AC Current – Measure <sup>1</sup>	(200 to 600) A		Fluke 376 Clamp Meter W/i2500 flex probe
	(10 to 500) Hz	2.3 % of reading + 0.5A	
	(600 to 2 500) A		
Resistance – Generate <sup>1</sup> Fixed Points	(10 to 500) Hz	2.3 % of reading + 5A	LN 422x Resistance Standards
	0.001 $\Omega$	10 $\mu\Omega/\Omega$	
	0.01 $\Omega$	6.6 $\mu\Omega/\Omega$	
	0.1 $\Omega$	25.3 $\mu\Omega/\Omega$	



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Resistance – Generate <sup>1</sup> Fixed Points	1 Ω	1 μΩ/Ω	Fluke 742A, IET SRL Resistance Standards
	10 Ω	2 μΩ/Ω	
	100 Ω	0.44 μΩ/Ω	
	1 kΩ	0.67 μΩ/Ω	
	10 kΩ	0.86 μΩ/Ω	
	100 kΩ	1.2 μΩ/Ω	
	1 MΩ	2 μΩ/Ω	
	10 MΩ	2.5 μΩ/Ω	
	100 MΩ	19 μΩ/Ω	
	1 GΩ	84 μΩ/Ω	
Resistance – Generate <sup>1</sup> Fixed Points	(1, 1.9) Ω	84 μΩ/Ω + 0.1 μΩ	Fluke 5730A Multifunction Calibrator
	(10, 19) Ω	21 μΩ/Ω + 1 μΩ	
	(100, 190) Ω	9.1 μΩ/Ω + 6 μΩ	
	(1, 1.9) kΩ	6.1 μΩ/Ω + 60 μΩ	
	(10, 19) kΩ	6.1 μΩ/Ω + 0.6 mΩ	
	100 kΩ	7.6 μΩ/Ω + 6 mΩ	
	190 kΩ	9.1 μΩ/Ω + 6 mΩ	
	1 MΩ	11 μΩ/Ω + 60 mΩ	
	1.9 MΩ	16 μΩ/Ω + 60 mΩ	
	10 MΩ	35 μΩ/Ω + 0.6 Ω	
	19 MΩ	42 μΩ/Ω + 0.6 Ω	
	100 MΩ	91 μΩ/Ω + 6 Ω	
Resistance – Generate <sup>1</sup>	Up to 11 Ω	30 μΩ/Ω + 0.001 Ω	Fluke 5522A Multiproduct Calibrator
	(11 to 33) Ω	23 μΩ/Ω + 0.001 Ω	
	(33 to 110) Ω	21 μΩ/Ω + 0.001 Ω	
	(110 to 330) Ω	21 μΩ/Ω + 0.002 Ω	
	330 Ω to 1.1 kΩ	21 μΩ/Ω + 0.002 Ω	
	(1.1 to 3.3) kΩ	21 μΩ/Ω + 0.02 Ω	
	(3.3 to 11) kΩ	21 μΩ/Ω + 0.02 Ω	
	(11 to 33) kΩ	21 μΩ/Ω + 0.2 Ω	
	(33 to 110) kΩ	21 μΩ/Ω + 0.2 Ω	
	(110 to 330) kΩ	24 μΩ/Ω + 2 Ω	
	(0.33 to 1.1) MΩ	24 μΩ/Ω + 2 Ω	
	(1.1 to 3.3) MΩ	46 μΩ/Ω + 23 Ω	
	(3.3 to 11) MΩ	99 μΩ/Ω + 38 Ω	
	(11 to 33) MΩ	190 μΩ/Ω + 1.9 kΩ	
	(33 to 110) MΩ	380 μΩ/Ω + 2.3 kΩ	
	(110 to 330) MΩ	0.23 % of reading + 76 kΩ	
(0.33 to 1.1) GΩ	1.1 % of reading + 380 kΩ		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Resistance – Generate <sup>1</sup> Fixed Points	100 V		TMI RB Resistance Standard
	100 kΩ	1 % of output	
	(100 to 1 000) V		
	1 MΩ	1 % of output	
	10 MΩ	1 % of output	
	100 MΩ	1 % of output	
	1 GΩ	1 % of output	
Resistance – Measure <sup>1</sup> Fixed Points	10 GΩ	1.2 % of output	Fluke 8508A Multimeter, Fluke 742A, IET SRL Resistance Standards
	100 GΩ	3.2 % of output	
	(1, 1.9) Ω	4.8 μΩ/Ω	
	(10, 19) Ω	2.6 μΩ/Ω	
	(100, 190) Ω	0.7 μΩ/Ω	
	(1, 1.9) kΩ	0.9 μΩ/Ω	
	(10, 19) kΩ	1 μΩ/Ω	
	(100, 190) kΩ	1.3 μΩ/Ω	
	(1, 1.9) MΩ	2.3 μΩ/Ω	
Resistance – Measure <sup>1</sup>	(10, 19) MΩ	2.9 μΩ/Ω	Fluke 8508A Multimeter
	100 MΩ	22 μΩ/Ω	
	1 GΩ	123 μΩ/Ω	
	Up to 2 Ω	15 μΩ/Ω + 3.8 μΩ	
	(2 to 20) Ω	8.8 μΩ/Ω + 14 μΩ	
	(20 to 200) Ω	7.2 μΩ/Ω + 46 μΩ	
	200 Ω to 2 kΩ	7.2 μΩ/Ω + 457 μΩ	
	(2 to 20) kΩ	7.2 μΩ/Ω + 4.6 mΩ	
	(20 to 200) kΩ	7.2 μΩ/Ω + 46 mΩ	
	200 kΩ to 2 MΩ	8 μΩ/Ω + 0.9 Ω	
	(2 to 20) MΩ	15 μΩ/Ω + 9 Ω	
(20 to 200) MΩ	57 μΩ/Ω + 0.9 kΩ		
200 MΩ to 2 GΩ	148 μΩ/Ω + 91 kΩ		
(2 to 20) GΩ	514 μΩ/Ω + 9.1 MΩ		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Resistance – Measure <sup>1</sup>	100 V		Quadtech 1865 Megohmmeter
	100 kΩ	0.53 % of reading	
	(100, 1 000) V		
	1 MΩ	0.5 % of reading	
	10 MΩ	0.5 % of reading	
	100 MΩ	0.5 % of reading	
	1 GΩ	0.5 % of reading	
	10 GΩ	0.52 % of reading	
	(500, 1 000) V		
	100 GΩ	0.59 % of reading	
Resistance – Measure <sup>1</sup>	1 000 V		IET 1693 RLC Meter
	1 TΩ	0.7 % of reading	
	10 TΩ	2.5 % of reading	
	50 Hz		
	(1 to 10) Ω	0.17 % of reading + 0.003 Ω	
	(10 to 100) Ω	0.062 % of reading + 0.006 Ω	
	(100 to 1000) Ω	0.051 % of reading + 0.06 Ω	
	(1 to 10) kΩ	0.05 % of reading + 0.000 6 kΩ	
	(10 to 100) kΩ	0.054 % of reading + 0.006 kΩ	
	(0.1 to 1) MΩ	0.09 % of reading + 0.000 06 MΩ	
	(1 to 10) MΩ	0.45% of reading + 0.000 6 MΩ	
	100/120 Hz		
	(1 to 10) Ω	0.13 % of reading + 0.003 Ω	
	(10 to 100) Ω	0.049 % of reading + 0.006 Ω	
	(100 to 1000) Ω	0.041 % of reading + 0.06 Ω	
	(1 to 10) kΩ	0.04 % of reading + 0.000 6 kΩ	
	(10 to 100) kΩ	0.043 % of reading + 0.006 kΩ	
	(0.1 to 1) MΩ	0.07 % of reading + 0.000 06 MΩ	
	(1 to 10) MΩ	0.34 % of reading + 0.000 6 MΩ	
	1 000 Hz		
(1 to 10) Ω	0.05 % of reading + 0.003 Ω		
(10 to 100) Ω	0.023 % of reading + 0.006 Ω		
(100 to 1000) Ω	0.02 % of reading + 0.06 Ω		
(1 to 10) kΩ	0.02 % of reading + 0.000 6 kΩ		
(10 to 100) kΩ	0.021 % of reading + 0.006 kΩ		
(0.1 to 1) MΩ	0.03 % of reading + 0.000 06 MΩ		
(1 to 10) MΩ	0.12 % of reading + 0.000 6 MΩ		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Resistance – Measure <sup>1</sup>	10 kHz		IET 1693 RLC Meter
	(1 to 10) Ω	0.17 % of reading + 0.003 Ω	
	(10 to 100) Ω	0.062 % of reading + 0.006 Ω	
	(100 to 1000) Ω	0.051 % of reading + 0.06 Ω	
	(1 to 10) kΩ	0.05 % of reading + 0.000 6 KΩ	
	(10 to 25.6) kΩ	0.051 % of reading + 0.006 KΩ	
	(25.6 to 100) kΩ	0.186 % of reading + 0.006 KΩ	
	(0.1 to 1) MΩ	0.33 % of reading + 0.000 06 MΩ	
	(1 to 10) MΩ	1.8 % of reading + 0.0006 MΩ	
	100 kHz		
	(1 to 10) Ω	0.77 % of reading + 0.003 Ω	
	(10 to 100) Ω	0.257 % of reading + 0.006 Ω	
	(100 to 1000) Ω	0.21 % of reading + 0.06 Ω	
	(1 to 10) kΩ	0.23 % of reading + 0.000 6 kΩ	
(10 to 100) kΩ	0.52 % of reading + 0.006 kΩ		
(0.1 to 1) kΩ	3.4 % of reading + 0.000 06 MΩ		
Capacitance – Generate <sup>1</sup>	(220 to 400) pF	0.38 % of output + 7.6 pF	Fluke 5522A Multiproduct Calibrator
	(0.4 to 3.299 9) nF	0.38 % of output + 0.01 nF	
	(3.3 to 10.999 9) nF	0.19 % of output + 0.01 nF	
	(11 to 32.999 9) nF	0.19 % of output + 0.08 nF	
	(33 to 109.999) nF	0.19 % of output + 0.08 nF	
	(110 to 329.999) nF	0.19 % of output + 0.23 nF	
	(0.33 to 1.099 99) μF	0.19 % of output + 0.76 nF	
	(1.1 to 3.299 99) μF	0.19 % of output + 2.3 nF	
	(3.3 to 10.999 9) μF	0.19 % of output + 7.6 nF	
	(11 to 32.999 9) μF	0.3 % of output + 23 nF	
	(33 to 109.999) μF	0.34 % of output + 76 nF	
	(110 to 329.999) μF	0.34 % of output + 228 nF	
	(0.33 to 1.099 99) mF	0.34 % of output + 0.76 μF	
	(1.1 to 3.299 99) mF	0.34 % of output + 2.3 μF	
(3.3 to 10.999 9) mF	0.34 % of output + 7.6 μF		
(11 to 32.999 9) mF	0.57 % of output + 23 μF		
(33 to 110) mF	0.84 % of output + 76 μF		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Capacitance – Measure 12 Hz to 200 kHz	50 Hz		IET 1693 LCR Meter
	(50 to 100) pF	2.6 % of reading + 0.015 pF	
	(100 to 400) pF	1.3 % of reading + 0.06 pF	
	(400 to 1 000) pF	0.37 % of reading + 0.06 pF	
	(1 to 10) nF	0.18 % of reading + 0.000 6 nF	
	(10 to 100) nF	0.063 % of reading + 0.006 nF	
	(100 to 1000) nF	0.051 % of reading + 0.06 nF	
	(0.1 to 1) μF	0.051 % of reading + 0.000 06 μF	
	(1 to 10) μF	0.05% of reading + 0.000 6 μF	
	(10 to 100) μF	0.054 % of reading + 0.006 μF	
	(100 to 1 000) μF	0.09 % of reading + 0.06 μF	
	100 Hz		
	(20 to 100) pF	2.4 % of reading + 0.015 pF	
	(100 to 400) pF	0.52 % of reading + 0.06 pF	
	(400 to 1 000) pF	0.16 % of reading + 0.06 pF	
	(1 to 10) nF	0.088 % of reading + 0.000 6 nF	
	(10 to 100) nF	0.045 % of reading + 0.006 nF	
	(100 to 1000) nF	0.041 % of reading + 0.06 nF	
	(0.1 to 1) μF	0.041 % of reading + 0.000 06 μF	
	(10 to 100) μF	0.046 % of reading + 0.006 μF	
	(100 to 1 000) μF	0.1 % of reading + 0.06 μF	
	120 Hz		
	(20 to 100) pF	2 % of reading + 0.01 5pF	
	(100 to 400) pF	0.44 % of reading + 0.06 pF	
	(400 to 1 000) pF	0.14 % of reading + 0.06 pF	
	(1 to 10) nF	0.08 % of reading + 0.000 6 nF	
	(10 to 100) nF	0.044 % of reading + 0.006 nF	
(100 to 1000) nF	0.04 % of reading + 0.06 nF		
(0.1 to 1) μF	0.04 % of reading + 0.000 06 μF		
(1 to 10) μF	0.041 % of reading + 0.000 6 μF		
(10 to 100) μF	0.047 % of reading + 0.006 μF		
(100 to 1 000) μF	0.11 % of reading + 0.06 μF		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Capacitance – Measure 12 Hz to 200 kHz	1 000 Hz		IET 1693 LCR Meter
	(1 to 10) pF	1.6 % of reading + 0.014 pF	
	(10 to 20) pF	0.18 % of reading + 0.015 pF	
	(20 to 100) pF	0.1 % of reading + 0.015 pF	
	(100 to 400) pF	0.036 % of reading + 0.06 pF	
	(400 to 1 000) pF	0.024 % of reading + 0.06 pF	
	(1 to 10) nF	0.022 % of reading + 0.000 6 nF	
	(10 to 100) nF	0.02 % of reading + 0.006 nF	
	(100 to 1 000) nF	0.02 % of reading + 0.06 nF	
	(0.1 to 1) μF	0.02 % of reading + 0.000 06 μF	
	(1 to 10) μF	0.020 % of reading + 0.000 6 μF	
	(10 to 100) μF	0.04 % of reading + 0.006 μF	
	(100 to 1 000) μF	0.22 % of reading + 0.06 μF	
	10 kHz		
	(20 to 100) pF	0.3 % of reading + 0.015 pF	
	(100 to 622) pF	0.2 % of reading + 0.06 pF	
	(622 to 1 000) pF	0.051 % of reading + 0.06 pF	
	(1 to 10) nF	0.051 % of reading + 0.000 6 nF	
	(10 to 100) nF	0.051 % of reading + 0.006 nF	
	(100 to 1 000) nF	0.058 % of reading + 0.06 nF	
	(0.1 to 1) μF	0.058 % of reading + 0.000 06 μF	
	(1 to 10) μF	0.13 % of reading + 0.000 6 μF	
	(10 to 100) μF	0.85 % of reading + 0.006 μF	
	(100 to 500) μF	4.1 % of reading + 0.06 μF	
	100 kHz		
	(20 to 100) pF	0.49 % of reading + 0.015 pF	
	(100 to 400) pF	0.26 % of reading + 0.06 pF	
	(400 to 1 000) pF	0.21 % of reading + 0.06 pF	
	(1 to 10) nF	0.21 % of reading + 0.000 6 nF	
	(10 to 100) nF	0.24 % of reading + 0.006 nF	
(100 to 1 000) nF	0.58 % of reading + 0.06 nF		
(0.1 to 1) μF	0.58 % of reading + 0.000 06 μF		
(1 to 10) μF	4 % of reading + 0.000 6 μF		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Capacitance – Generate <sup>1</sup> Fixed Points	1 pF		Agilent 16381A Capacitor Set
	1 kHz	0.028 % of output	
	1 MHz	0.028 % of output	
	2 MHz	0.034 % of output	
	3 MHz	0.047 % of output	
	4 MHz	0.069 % of output	
	5 MHz	0.096 % of output	
	10 MHz	0.31 % of output	
Capacitance – Generate <sup>1</sup> Fixed Points	10 pF		Agilent 16382A Standard Air Capacitor
	1 kHz	0.019 % of output	
	1 MHz	0.019 % of output	
	2 MHz	0.019 % of output	
	3 MHz	0.019 % of output	
	4 MHz	0.019 % of output	
	5 MHz	0.019 % of output	
	10 MHz	0.022 % of output	
Capacitance – Generate <sup>1</sup> Fixed Points	100 pF		Agilent 16383A Standard Air Capacitor
	1 kHz	0.017 % of output	
	1 MHz	0.017 % of output	
	2 MHz	0.018 % of output	
	3 MHz	0.018 % of output	
	4 MHz	0.02 % of output	
	5 MHz	0.024 % of output	
	10 MHz	0.053 % of output	
Capacitance – Generate <sup>1</sup> Fixed Points	1 000 pF		Agilent 16384A Standard Air Capacitor
	1 kHz	0.018 % of output	
	1 MHz	0.019 % of output	
	2 MHz	0.024 % of output	
	3 MHz	0.035 % of output	
	4 MHz	0.051 % of output	
	5 MHz	0.069 % of output	
	10 MHz	0.22 % of output	
13 MHz	0.32 % of output		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Inductance – Measure <sup>1</sup> 12 Hz to 200 kHz	100 Hz		IET 1693 LCR Meter
	(40 to 100) $\mu$ H	3.8 % of reading + 0.006 $\mu$ H	
	(0.1 to 0.4) mH	1.6 % of reading + 0.000 06 mH	
	(0.4 to 1) mH	0.42 % of reading + 0.000 06 mH	
	(1 to 10) mH	0.19 % of reading + 0.000 6 mH	
	(10 to 100) mH	0.056 % of reading + 0.006 mH	
	(0.1 to 1) H	0.042 % of reading + 0.000 06 H	
	(1 to 10) H	0.041 % of reading + 0.000 6 H	
	(10 to 100) H	0.042 % of reading + 0.006 H	
	1 000 Hz		
	(4 to 10) $\mu$ H	1.3 % of reading + 0.003 $\mu$ H	
	(10 to 40) $\mu$ H	0.53 % of reading + 0.006 $\mu$ H	
	(40 to 100) $\mu$ H	0.15 % of reading + 0.006 $\mu$ H	
	(0.1 to 0.4) mH	0.071 % of reading + 0.000 06 mH	
	(0.4 to 1) mH	0.033 % of reading + 0.000 06 mH	
	(1 to 10) mH	0.025 % of reading + 0.000 6 mH	
	(10 to 100) mH	0.021 % of reading + 0.006 mH	
	(0.1 to 1) H	0.02 % of reading + 0.000 06 H	
	(1 to 10) H	0.021 % of reading + 0.000 6 H	
	(10 to 100) H	0.026 % of reading + 0.006 H	
	10 kHz		
	(1 to 4) $\mu$ H	2.1 % of reading + 0.003 $\mu$ H	
	(4 to 10) $\mu$ H	0.56 % of reading + 0.003 $\mu$ H	
	(10 to 40) $\mu$ H	0.25 % of reading + 0.006 $\mu$ H	
	(40 to 100) $\mu$ H	0.1 % of reading + 0.006 $\mu$ H	
	(0.1 to 0.4) mH	0.071 % of reading + 0.000 06 mH	
	(0.4 to 1) mH	0.056 % of reading + 0.000 06 mH	
(1 to 10) mH	0.052 % of reading + 0.000 6 mH		
(10 to 100) mH	0.051 % of reading + 0.006 mH		
(0.1 to 0.407) H	0.051 % of reading + 0.000 06 H		
(0.407 to 1) H	0.18 % of reading + 0.000 6 H		
(1 to 10) H	0.27 % of reading + 0.000 6 H		
(10 to 100) H	1.2 % of reading + 0.006 H		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Inductance – Measure <sup>1</sup> 12 Hz to 200 kHz	100 kHz (1 to 4) $\mu$ H (4 to 10) $\mu$ H (10 to 40) $\mu$ H (40 to 100) $\mu$ H (0.1 to 0.4) mH (0.4 to 1) mH (1 to 10) mH (10 to 100) mH (0.1 to 1) H	1.2 % of reading + 0.003 $\mu$ H 0.44 % of reading + 0.003 $\mu$ H 0.3 % of reading + 0.006 $\mu$ H 0.23 % of reading + 0.006 $\mu$ H 0.21 % of reading + 0.000 06 mH 0.21 % of reading + 0.000 06 mH 0.22 % of reading + 0.000 6 mH 0.39 % of reading + 0.006 mH 2.1 % of reading + 0.000 06 H	IET 1693 LCR Meter
Inductance – Generate <sup>1</sup> Fixed Point	50 mH 100 Hz 1 kHz	0.028 mH 0.021 mH	GR 1482-K Standard Inductor
	100 mH 100 Hz 1 kHz	0.069 mH 0.063 mH	GR 1481-G Standard Inductor
	200 mH 100 Hz 1 kHz	0.09 mH 0.06 mH	GR 1482-M Standard Inductor
	500 mH 100 Hz 1 kHz	0.23 mH 0.15 mH	GR 1482-N Standard Inductor
	5 H 100 Hz 1 kHz	0.021 H 0.021 H	GR 1482-R Standard Inductor
Oscilloscopes Calibration <sup>1</sup> Generate Voltage DC - 50 $\Omega$	(1 to 24.999) mV (25 to 109.99) mV (110mV to 2.1999) V (2.2 to 6.6) V	0.19 % of Output + 31 $\mu$ V 0.19 % of Output + 36 $\mu$ V 0.19 % of Output + 87 $\mu$ V 0.19 % of Output + 0.6 mV	Fluke 5820A Oscilloscope Calibrator w/ GHz Option
DC - 1M $\Omega$	(1 to 24.999) mV (25 to 109.99) mV (110mV to 2.1999) V (2.2 to 10.999) V (11 to 130) V	0.019 % of reading + 20 $\mu$ V 0.019 % of reading + 25 $\mu$ V 0.019 % of reading + 76 $\mu$ V 0.019 % of reading + 0.6 mV 0.019 % of reading + 6 mV	
Square Wave 10Hz to 10kHz - 50 $\Omega$	(1 to 24.999) mVpp (25 to 109.99) mVpp (110mV to 2.1999) Vpp (2.2 to 6.6) Vpp	0.19 % of Output + 31 $\mu$ V 0.19 % of Output + 36 $\mu$ V 0.19 % of Output + 87 $\mu$ V 0.19 % of Output + 0.6 mV	



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Oscilloscopes Calibration <sup>1</sup> Generate  Square Wave 10 Hz to 1 kHz – 1 MΩ  Square Wave 1 to 10 kHz - 1MΩ	(1 to 24.999) mV (25 to 109.99) mV (110mV to 2.199 9) V (2.2 to 10.999) V (11 to 130) V  (1 to 24.999) mV (25 to 109.99) mV (110mV to 2.199 9) V (2.2 to 10.999) V (11 to 130) V	0.038 % of reading + 4 μV 0.038 % of reading + 9 μV 0.038 % of reading + 60 μV 0.038 % of reading + 0.6 mV 0.038 % of reading + 6 mV  0.19 % of reading + 31 μV 0.19 % of reading + 36 μV 0.19 % of reading + 87 μV 0.19 % of reading + 0.6 mV 0.19 % of reading + 6 mV	Fluke 5820A Oscilloscope Calibrator w/ GHz Option
Oscilloscopes Calibration– Generate <sup>1</sup>  Leveled Sine Flatness 50 kHz to 10 MHz Reference	3 dB Bandwidth (5 to 50 mV) Vpp 50 kHz to 100 MHz (100 to 300) MHz (300 to 500) MHz (500 to 600) MHz (600 to 16 00) MHz (1 600 to 2 100) MHz	0.34 dB 0.36 dB 0.42 dB 0.46 dB 0.5 dB 0.56 dB	Fluke 5820A Oscilloscope Calibrator w/ GHz Option
Oscilloscopes Calibration Generate <sup>1</sup>  Leveled Sine Flatness 50 kHz to 10 MHz Reference	3 dB Bandwidth 50 mVpp to 3.5 Vpp 50 kHz to 100) MHz (100 to 300) MHz (300 to 500) MHz (500 to 600) MHz (600 to 16 00) MHz (1 600 to 2 100) MHz (3.5 to 5) Vpp 50 kHz to 100 MHz (100 to 300) MHz (300 to 500) MHz (500 to 600) MHz	0.24 dB 0.24 dB 0.32 dB 0.34 dB 0.4 dB 0.44 dB  0.24 dB 0.24 dB 0.32 dB 0.34 dB	Fluke 5820A Oscilloscope Calibrator w/ GHz Option
Oscilloscopes Calibration Generate <sup>1</sup> Leveled Sine Flatness	3 dB Bandwidth 50 mV to 3.5Vpp (2 100 to 4 000) MHz (4 000 to 8 000) MHz (8 000 to 18 000) MHz	0.25 dB 0.35 dB 0.46 dB	EPM Power Meter w/ E Series Power Sensors

**Electrical – DC/Low Frequency**

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Oscilloscopes Calibration Generate <sup>1</sup> Time Marker	500 ps to 20 ms 50 ms to 5 s	0.25 $\mu$ s/s 1.9 $\mu$ s/s + 3.8 $\mu$ Hz	Fluke 5820A Oscilloscope Calibrator w/ GHz Option
Rise Time	< 150 ps	+0 / -50 ps	
Input Impedance	(40 to 60) $\Omega$		
Resistance	500 k $\Omega$ to 1.5M $\Omega$	0.08 % of reading 0.08 % of reading	
Capacitance	(5 to 50) pF	3.8 % of reading + 0.4pF	
Leakage	(0 to 5.99) V	0.038 % of reading + 0.8 mV	
Oscilloscopes Calibration <sup>1</sup> Generate Voltage DC - 50 $\Omega$ DC - 1M $\Omega$	(1mV to 5V) (1mV to 200V)	0.025 % of output + 25 $\mu$ V 0.025 % of output + 25 $\mu$ V	Fluke 9500B with 9510 Active Head
Square Wave 10Hz to 10kHz – 1M $\Omega$	(1mV to 200Vpp)	0.1 % of output + 10 $\mu$ V	
Oscilloscopes Calibration <sup>1</sup> Generate	5 mV to 5 Vpp (0.1 Hz to 300) MHz (300 to 550) MHz	0.18 dB 0.22 dB	
Leveled Sine Flatness 50 kHz to 10 MHz Reference	(5 mV to 3Vpp) (550 to 1 100) MHz	0.3 dB	Fluke 9500B with 9510 Active Head
Oscilloscopes Calibration <sup>1</sup> Generate Time Marker	9 ns to 55 s	0.25 $\mu$ s/s	Fluke 9500B with 9510 Active Head
Oscilloscopes Calibration <sup>1</sup> Measure	(10 to 40) $\Omega$ (40 to 90) $\Omega$ (90 to 150) $\Omega$ (50 to 800) K $\Omega$ (0.8 to 1.2) M $\Omega$ (1.2 to 12) M $\Omega$	0.5 % of reading 0.1 % of reading 0.5 % of reading 0.5 % of reading 0.1 % of reading 0.5 % of reading	Fluke 9500B with 9510 Active Head



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Calibration of Thermocouple Indicating Devices <sup>1</sup>	Type B		Fluke 7526A Process Calibrator
	(600 to 800) °C	0.27 °C	
	(800 to 1 550) °C	0.22 °C	
	(1 550 to 1 820) °C	0.17 °C	
	Type C		
	(0 to 1000) °C	0.13 °C	
	(1 000 to 1 800) °C	0.18 °C	
	(1 800 to 2 000) °C	0.2 °C	
	(2 000 to 2 316) °C	0.27 °C	
	Type E		
	(-250 to -200) °C	0.19 °C	
	(-200 to -100) °C	0.1 °C	
	(-100 to 0) °C	0.07 °C	
	(0 to 600) °C	0.07 °C	
	(600 to 1 000) °C	0.08 °C	
	Type J		
	(-210 to -100) °C	0.11 °C	
	(-100 to 800) °C	0.07 °C	
	(800 to 1 200) °C	0.08 °C	
	Type K		
	(-250 to -200) °C	0.35 °C	
(-200 to -100) °C	0.13 °C		
(-100 to 800) °C	0.08 °C		
(800 to 1 372) °C	0.1 °C		
Type L			
(-200 to -100) °C	0.08 °C		
(-100 to 900) °C	0.07 °C		
Type N			
(-250 to -200) °C	0.56 °C		
(-200 to -100) °C	0.18 °C		
(-100 to 0) °C	0.1 °C		
(0 to 100) °C	0.09 °C		
(100 to 800) °C	0.08 °C		
(800 to 1 300) °C	0.1 °C		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Calibration of Thermocouple Indicating Devices <sup>1</sup>	Type R		Fluke 7526A Process Calibrator
	(-50 to -25) °C	0.42 °C	
	(-25 to 0) °C	0.34 °C	
	(0 to 100) °C	0.3 °C	
	(100 to 400) °C	0.22 °C	
	(400 to 600) °C	0.17 °C	
	(600 to 1 000) °C	0.16 °C	
	(1 000 to 1 600) °C	0.15 °C	
	(1 600 to 1 767) °C	0.18 °C	
	Type S		
	(50 to -25) °C	0.39 °C	
	(-25 to 0) °C	0.33 °C	
	(0 to 100) °C	0.29 °C	
	(100 to 400) °C	0.22 °C	
	(400 to 600) °C	0.18 °C	
	(600 to 1 600) °C	0.17 °C	
	(1 600 to 1 767) °C	0.2 °C	
	Type T		
	(-250 to -200) °C	0.26 °C	
	(-200 to -100) °C	0.13 °C	
(-100 to 0) °C	0.09 °C		
(0 to 400) °C	0.07 °C		
Type U			
(-200 to 0) °C	0.13 °C		
(0 to 600) °C	0.08 °C		
Electrical Calibration of RTD Indicating Devices <sup>1</sup>	Pt 385, 100 Ω		Fluke 7526A Process Calibrator
	(-200 to 800) °C	0.05 °C	
	Pt 3926, 100 Ω		
	(-200 to 630) °C	0.05 °C	
	Pt 3916, 100 Ω		
	(-200 to 630) °C	0.05 °C	
	Pt 385, 200 Ω		
	(-200 to 400) °C	0.4 °C	
(400 to 630) °C	0.5 °C		
Pt 385, 500 Ω			
(-200 to 630) °C	0.17 °C		
Pt 385, 1 000 Ω			
(-200 to 630) °C	0.09 °C		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure Wideband Flatness Relative to 1 kHz	(0.6 to 2.2) mV		Fluke 5790B AC/DC Measurement Standard
	(10 to 30) Hz	0.08 % of reading + 0.1 $\mu$ V	
	30 Hz to 120 kHz	0.04 % of reading + 0.1 $\mu$ V	
	120 kHz 2 MHz	0.05 % of reading + 0.8 $\mu$ V	
	(2 to 10) MHz	0.13 % of reading + 0.8 $\mu$ V	
	(10 to 20) MHz	0.23 % of reading + 0.8 $\mu$ V	
	(20 to 30) MHz	0.53 % of reading + 1.6 $\mu$ V	
	(2.2 to 7) mV		
	(10 to 30) Hz	0.08 % of reading + 0.1 $\mu$ V	
	30 Hz to 120 kHz	0.04 % of reading + 0.1 $\mu$ V	
	120 kHz 2 MHz	0.05 % of reading + 0.8 $\mu$ V	
	(2 to 10) MHz	0.08 % of reading + 0.8 $\mu$ V	
	(10 to 20) MHz	0.13 % of reading + 0.8 $\mu$ V	
	(20 to 30) MHz	0.28 % of reading + 0.8 $\mu$ V	
	(7 to 22) mV		
	(10 to 30) Hz	0.08 % of reading + 0.1 $\mu$ V	
	30 Hz to 120 kHz	0.04 % of reading + 0.1 $\mu$ V	
	120 kHz 2 MHz	0.05 % of reading + 0.1 $\mu$ V	
	(2 to 10) MHz	0.08 % of reading + 0.1 $\mu$ V	
	(10 to 20) MHz	0.13 % of reading + 0.1 $\mu$ V	
	(20 to 30) MHz	0.28 % of reading + 0.1 $\mu$ V	
	(22 to 70) mV		
	(10 to 30) Hz	0.08 % of reading + 0.6 $\mu$ V	
	30 Hz to 2 MHz	0.04 % of reading + 0.6 $\mu$ V	
	(2 to 10) MHz	0.08 % of reading + 0.6 $\mu$ V	
	(10 to 20) MHz	0.11 % of reading + 0.6 $\mu$ V	
	(20 to 30) MHz	0.27 % of reading + 0.6 $\mu$ V	
	(70 to 220) mV		
(10 to 30) Hz	0.08 % of reading		
30 Hz to 500 kHz	0.03 % of reading		
500 kHz to 2 MHz	0.04 % of reading		
(2 to 10) MHz	0.08 % of reading		
(10 to 20) MHz	0.11 % of reading		
(20 to 30) MHz	0.27 % of reading		
(220 to 700) mV			
(10 to 30) Hz	0.08 % of reading		
30 Hz to 500 kHz	0.02 % of reading		
500 kHz to 2 MHz	0.04 % of reading		
(2 to 10) MHz	0.08 % of reading		
(10 to 20) MHz	0.11 % of reading		
(20 to 30 MHz)	0.27 % of reading		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure Wideband Flatness Relative to 1 kHz	(0.7 to 2) V		Fluke 5790B AC/DC Measurement Standard
	(10 to 30) Hz	0.08 % of reading	
	30 Hz to 500 kHz	0.02 % of reading	
	500 kHz to 2 MHz	0.04 % of reading	
	(2 to 10) MHz	0.08 % of reading	
	(10 to 20) MHz	0.11 % of reading	
	(20 to 30) MHz	0.27 % of reading	
	(2.2 to 7) V		
	(10 to 30) Hz	0.08 % of reading	
	30 Hz to 500 kHz	0.02 % of reading	
	500 kHz to 2 MHz	0.04 % of reading	
	(2 to 10) MHz	0.08 % of reading	
	(10 to 20) MHz	0.11 % of reading	
	(20 to 30) MHz	0.27 % of reading	
Power Meter Range Calibration <sup>1</sup>	3 $\mu$ W to 100 mW	0.25 % of reading	HP 11683A Power Meter Calibrator

Electrical – RF/Microwave

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Flatness – Measure	9 kHz to 2 000 MHz		Agilent EPM Series Power Meter w/E9304A H18 Power Sensor
	(20 to -10) dBm	0.1 dB	
	(-10 to -30) dBm	0.1 dB	
	(-30 to -40) dBm	0.11 dB	
	(-40 to -42) dBm	0.12 dB	
	(2 to 14) GHz		
	(20 to -10) dBm	0.1 dB	
	(-10 to -30) dBm	0.09 dB	
	(-30 to -40) dBm	0.1 dB	
	(-40 to -42) dBm	0.11 dB	
	(14 to 18) GHz		
	(20 to -10) dBm	0.11 dB	
	(-10 to -30) dBm	0.12 dB	
	(-30 to -40) dBm	0.12 dB	
	(-40 to -42) dBm	0.13 dB	



ANSI National Accreditation Board

Electrical – RF/Microwave

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Thermal Noise – Measure ENR	5 dB, 15 dB, or 21 dB		Agilent 346C Noise Source Option H13
	0.1 GHz	0.21 dB	
	1 GHz	0.21 dB	
	2 GHz	0.22 dB	
	3 GHz	0.23 dB	
	4 GHz	0.24 dB	
	5 GHz	0.25 dB	
	6 GHz	0.26 dB	
	7 GHz	0.27 dB	
	8 GHz	0.28 dB	
	9 GHz	0.29 dB	
	10 GHz	0.3 dB	
	11 GHz	0.31 dB	
	12 GHz	0.32 dB	
	13 GHz	0.33 dB	
	14 GHz	0.34 dB	
15 GHz	0.35 dB		
16 GHz	0.36 dB		
Thermal Noise – Measure ENR	5 dB, 15 dB, or 21 dB		Agilent 346C Noise Source Option H13
	17 GHz	0.37 dB	
	18 GHz	0.38 dB	
	19 GHz	0.39 dB	
	20 GHz	0.4 dB	
	21 GHz	0.41 dB	
	22 GHz	0.42 dB	
	23 GHz	0.43 dB	
	24 GHz	0.44 dB	
	25 GHz	0.45 dB	
26 GHz	0.46 dB		
26.5 GHz	0.47 dB		

**Electrical – RF/Microwave**

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Attenuation – Measure <sup>1</sup>	(30 to 3 050) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.12 dB	
	(90 to 100) dB	0.13 dB	
	(100 to 110) dB	0.13 dB	
	(3 050 to 6 600) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.12 dB	
	(90 to 100) dB	0.13 dB	
	(100 to 110) dB	0.19 dB	
	(6 600 to 13 200) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
(50 to 60) dB	0.076 dB		
(60 to 70) dB	0.081 dB		
(70 to 80) dB	0.12 dB		
(80 to 90) dB	0.12 dB		
(90 to 100) dB	0.13 dB		
(100 to 110) dB	0.25 dB		

**Electrical – RF/Microwave**

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Attenuation – Measure <sup>1</sup>	(13 200 to 19 200) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.12 dB	
	(90 to 100) dB	0.13 dB	
	(100 to 110) dB	0.31 dB	
	(19 200 to 26 500) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.14 dB	
	(90 to 100) dB	0.36 dB	
	(100 to 110) dB	0.82 dB	
	(26 500 to 31 150) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
(50 to 60) dB	0.076 dB		
(60 to 70) dB	0.081 dB		
(70 to 80) dB	0.12 dB		
(80 to 90) dB	0.13 dB		
(90 to 100) dB	0.33 dB		
(100 to 110) dB	0.77 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Attenuation – Measure <sup>1</sup>	(31 150 to 41 000) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.14 dB	
	(80 to 90) dB	0.36 dB	
	(41 000 to 45 000) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.11 dB	
	(70 to 80) dB	0.24 dB	
	(80 to 90) dB	0.6 dB	
	(45 000 to 50 000) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.11 dB	
(60 to 70) dB	0.29 dB		
(70 to 80) dB	0.7 dB		
(80 to 90) dB	1.4 dB		
RF Power – Measure <sup>1</sup>	9 kHz to 14 000 MHz		Agilent EPM Series Power Meter w/E9304A H18 Power Sensor
	(20 to 0) dB	0.13 dB	
	(0 to -40) dB	0.15 dB	
	(-40 to -50) dB	0.35 dB	
	(-50 to -55) dB	0.93 dB	
	(14 000 to 18 000) MHz		
	(20 to 0) dB	0.12 dB	
	(0 to -40) dB	0.16 dB	
	(-40 to -50) dB	0.35 dB	
	(-50 to -55) dB	0.93 dB	



ANSI National Accreditation Board

Electrical – RF/Microwave

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Measure <sup>1</sup>	(50 to 100) MHz		Agilent EPM Series Power Meter RF Power / Keysight N8487A Power Sensor
	(20 to 10) dB	0.07 dB	
	(10 to 0) dB	0.07 dB	
	(0 to -10) dB	0.06 dB	
	(-10 to -20) dB	0.07 dB	
	(-20 to -25) dB	0.11 dB	
	(100 to 6 000) MHz		
	(20 to 10) dB	0.07 dB	
	(10 to 0) dB	0.07 dB	
	(0 to -10) dB	0.07 dB	
	(-10 to -20) dB	0.07 dB	
	(-20 to -25) dB	0.11 dB	
	(6 000 to 12 400) MHz		
	(20 to 10) dB	0.08 dB	
	(10 to 0) dB	0.08 dB	
	(0 to -10) dB	0.07 dB	
	(-10 to -20) dB	0.08 dB	
	(-20 to -25) dB	0.12 dB	
	(12 400 to 18 000) MHz		
	(20 to 10) dB	0.09 dB	
	(10 to 0) dB	0.08 dB	
	(0 to -10) dB	0.08 dB	
	(-10 to -20) dB	0.08 dB	
	(-20 to -25) dB	0.12 dB	
(18 000 to 26 500) MHz			
(20 to 10) dB	0.11 dB		
(10 to 0) dB	0.1 dB		
(0 to -10) dB	0.1 dB		
(-10 to -20) dB	0.1 dB		
(-20 to -25) dB	0.13 dB		
(26 500 to 33 000) MHz			
(20 to 10) dB	0.12 dB		
(10 to 0) dB	0.12 dB		
(0 to -10) dB	0.11 dB		
(-10 to -20) dB	0.12 dB		
(-20 to -25) dB	0.14 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Measure <sup>1</sup>	(33 000 to 40 000) MHz		Agilent EPM Series Power Meter RF Power / Keysight N8487A Power Sensor
	(20 to 10) dB	0.13 dB	
	(10 to 0) dB	0.12 dB	
	(0 to -10) dB	0.12 dB	
	(-10 to -20) dB	0.13 dB	
	(-20 to -25) dB	0.15 dB	
	(40 000 to 50 000) MHz		
	(20 to 10) dB	0.18 dB	
	(10 to 0) dB	0.18 dB	
	(0 to -10) dB	0.18 dB	
RF Power – Measure <sup>1</sup>	100 kHz to 2 000 MHz		Agilent EPM Series Power Meter RF Power / HP 8482B Power Sensor
	(44 to 35) dB	0.18 dB	
	(35 to 30) dB	0.07 dB	
	(30 to 20) dB	0.07 dB	
	(20 to 10) dB	0.08 dB	
	(10 to 5) dB	0.15 dB	
	(2 000 to 4 200) MHz		
	(44 to 35) dB	0.18 dB	
	(35 to 30) dB	0.07 dB	
	(30 to 20) dB	0.07 dB	
RF Power – Measure <sup>1</sup>	(30 to 2 000) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor
	(30 to 20) dB	0.36 dB	
	(20 to 0) dB	0.2 dB	
	(0 to -58) dB	0.22 dB	
	(-58 to -78) dB	0.23 dB	
	(-78 to -110) dB	0.25 dB	
	(-110 to -115) dB	0.27 dB	
	(-115 to -120) dB	0.33 dB	
(-120 to -125) dB	0.46 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Measure <sup>1</sup>	(2 000 to 3 050) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor
	(30 to 20) dB	0.37 dB	
	(20 to 0) dB	0.21 dB	
	(0 to -58) dB	0.23 dB	
	(-58 to -78) dB	0.24 dB	
	(-78 to -110) dB	0.26 dB	
	(-110 to -115) dB	0.28 dB	
	(-115 to -120) dB	0.34 dB	
	(-120 to -125) dB	0.7 dB	
	(3 050 to 6 600) MHz		
	(30 to 20) dB	0.37 dB	
	(20 to 0) dB	0.21 dB	
	(0 to -58) dB	0.23 dB	
	(-58 to -78) dB	0.24 dB	
	(-78 to -110) dB	0.29 dB	
	(-110 to -115) dB	0.38 dB	
	(-115 to -120) dB	0.53 dB	
	(6 600 to 13 200) MHz		
	(30 to 20) dB	0.37 dB	
	(20 to 0) dB	0.21 dB	
	(0 to -58) dB	0.23 dB	
	(-58 to -78) dB	0.24 dB	
	(-78 to -110) dB	0.34 dB	
	(-110 to -115) dB	0.46 dB	
	(-115 to -120) dB	0.65 dB	
	(13 200 to 18 000) MHz		
	(30 to 20) dB	0.37 dB	
	(20 to 0) dB	0.21 dB	
	(0 to -58) dB	0.23 dB	
	(-58 to -78) dB	0.24 dB	
	(-78 to -90) dB	0.26 dB	
	(-90 to -95) dB	0.26 dB	
	(-95 to -100) dB	0.26 dB	
	(-100 to -105) dB	0.29 dB	
	(-105 to -110) dB	0.38 dB	
	(-110 to -115) dB	0.53 dB	
(-115 to -120) dB	0.75 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Measure <sup>1</sup>	(18 000 to 19 200) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor
	(30 to 20) dB	0.39 dB	
	(20 to 0) dB	0.25 dB	
	(0 to -58) dB	0.24 dB	
	(-58 to -78) dB	0.25 dB	
	(-78 to -90) dB	0.27 dB	
	(-90 to -95) dB	0.27 dB	
	(-95 to -100) dB	0.27 dB	
	(-100 to -105) dB	0.3 dB	
	(-105 to -110) dB	0.38 dB	
	(-110 to -115) dB	0.53 dB	
	(-115 to -120) dB	0.75 dB	
	(19 200 to 26 500) MHz		
	(30 to 20) dB	0.39 dB	
	(20 to 0) dB	0.25 dB	
	(0 to -58) dB	0.24 dB	
	(-58 to -78) dB	0.25 dB	
	(-78 to -90) dB	0.28 dB	
	(-90 to -95) dB	0.33 dB	
	(-95 to -100) dB	0.43 dB	
	(-100 to -105) dB	0.61 dB	
	(-105 to -110) dB	0.85 dB	
	(-110 to -115) dB	1.2 dB	
	(-115 to -120) dB	1.5 dB	
	(26 500 to 31 150) MHz		
	(30 to 20) dB	0.42 dB	
	(20 to 0) dB	0.3 dB	
(0 to -58) dB	0.34 dB		
(-58 to -78) dB	0.34 dB		
(-78 to -90) dB	0.36 dB		
(-90 to -95) dB	0.39 dB		
(-95 to -100) dB	0.46 dB		
(-100 to -105) dB	0.61 dB		
(-105 to -110) dB	0.82 dB		
(-110 to -115) dB	1.1 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Measure <sup>1</sup>	(31 150 to 41 000) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor
	(30 to 20) dB	0.42 dB	
	(20 to 0) dB	0.3 dB	
	(0 to -58) dB	0.34 dB	
	(-58 to -78) dB	0.35 dB	
	(-78 to -90) dB	0.48 dB	
	(-90 to -95) dB	0.64 dB	
	(-95 to -100) dB	0.87 dB	
	(-100 to -105) dB	1.2 dB	
	(-105 to -110) dB	1.5 dB	
	(41 000 to 45 000) MHz		
	(30 to 20) dB	0.42 dB	
	(20 to 0) dB	0.3 dB	
	(0 to -58) dB	0.34 dB	
	(-58 to -78) dB	0.38 dB	
	(-78 to -90) dB	0.68 dB	
	(-90 to -95) dB	0.93 dB	
	(-95 to -100) dB	1.2 dB	
(-100 to -105) dB	1.6 dB		
Amplitude Modulation – Measure <sup>1</sup>	(45 000 to 50 000) MHz		
	(30 to 20) dB	0.42 dB	
	(20 to 0) dB	0.3 dB	
	(0 to -58) dB	0.34 dB	
	(-58 to -78) dB	0.67 dB	
	(-78 to -90) dB	1.4 dB	
	100 kHz to 10 MHz		
	Rate 50 Hz to 10 kHz		
	(5 to 99) % Depth	0.75 % of reading + 0.3 digits	
	10 MHz to 3 GHz		
	Rate 50 Hz to 100 kHz		
	(5 to 20) % Depth	2.5 % of reading + 0.4 digits	
(20 to 99) % Depth	1.5 % of reading + 0.4 digits		
(3 to 26.5) GHz			
Rate 50 Hz to 100 kHz			
(5 to 20) % Depth	4.5 % of reading + 0.4 digits		
(20 to 99) % Depth	1.5 % of reading + 0.4 digits		
(26.5 to 31.15) GHz			
Rate 50 Hz to 100 kHz			
(5 to 20) % Depth	6.8 % of reading + 0.4 digits		
(20 to 99) % Depth	1.9 % of reading + 0.4 digits		



ANSI National Accreditation Board

Electrical – RF/Microwave

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Amplitude Modulation – Measure <sup>1</sup>	(31.15 to 50) GHz Rate 50 Hz to 100 kHz (5 to 20) % Depth (20 to 99) % Depth	26 % of reading + 0.4 digits 6 % of reading + 0.4 digits	Agilent N5531S Measuring Receiver
Frequency Modulation – Measure <sup>1</sup> $\beta$ = deviation / rate	250 kHz to 10 MHz Rates 20 Hz to 10 kHz Peak Dev 200 to 40 kHz 10 MHz to 6.6 GHz Rates 50 Hz to 200 kHz Peak Dev 250 to 400 kHz (6.6 to 13.2) GHz Rates 50 Hz to 200 kHz Peak Dev 250 to 400 kHz (13.2 to 26.5) GHz Rates 50 Hz to 200 kHz Peak Dev 250 to 400 kHz (13.2 to 31.15) GHz Rates 50 Hz to 200 kHz Peak Dev 250 to 400 kHz (31.15 to 50) GHz Rates 50 Hz to 200 kHz Peak Dev 250 to 400 kHz	$\beta > 0.2$ - 1.5 % of reading + 2 Hz $\beta > 1.2$ - 1 % of reading + 2 Hz $\beta > 0.20$ - 1.5 % of reading + 2 Hz $\beta > 0.45$ - 1 % of reading + 2 Hz $\beta > 0.2$ - 2.5 % of reading + 4 Hz $\beta > 8.0$ - 1 % of reading + 4 Hz $\beta > 0.2$ - 3.8 % of reading + 9 Hz $\beta > 16$ - 1 % of reading + 9 Hz $\beta > 0.2$ - 3.8 % of reading + 9 Hz $\beta > 16$ - 1 % of reading + 9 Hz $\beta > 0.2$ - 8.5 % of reading + 17 Hz $\beta > 16$ - 1 % of reading + 17 Hz	Agilent N5531S Measuring Receiver
Phase Modulation – Measure <sup>1</sup>	100 kHz to 6.6 GHz Deviations > 0.3 rad Deviations > 0.7 rad (6.6 to 13.2) GHz Deviations > 0.6 rad Deviations > 2.0 rad (13.2 to 26.5) GHz Deviations: > 1.2 rad Deviations > 4.0 rad (26.5 to 31.15) GHz Deviations: > 1.3 rad Deviations > 4.0 rad (31.15 to 50) GHz Deviations: > 2.4 rad Deviations > 8.0 rad	3 % of reading + 0.002 rad 1 % of reading + 0.002 rad 3 % of reading + 0.005 rad 1 % of reading + 0.005 rad 3 % of reading + 0.009 rad 1 % of reading + 0.009 rad 3 % of reading + 0.009 rad 1 % of reading + 0.009 rad 3 % of reading + 0.018 rad 1 % of reading + 0.018 rad	Agilent N5531S Measuring Receiver



ANSI National Accreditation Board

Electrical – RF/Microwave

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AM Distortion Measure <sup>1</sup> Rate 20 Hz to 1 kHz	(0.1 to 10) MHz		Agilent N5531S Measuring Receiver
	AM Depth > 1 %		
	(0 to -20) dB	1.2 dB	
	(-20 to -30) dB	2.2 dB	
	AM Depth > 3 %		
	(0 to -20) dB	1 dB	
	(-20 to -30) dB	1.3 dB	
	(-30 to -40) dB	2.4 dB	
	10 MHz to 26.5 GHz		
	AM Depth > 1 %		
	(0 to -20) dB	1.3 dB	
	(-20 to -30) dB	2.5 dB	
	AM Depth > 3 %		
(0 to -20) dB	1.1 dB		
(-20 to -30) dB	1.4 dB		
(-30 to -40) dB	3 dB		
(26.5 MHz to 50.0 GHz)			
AM Depth > 3 %			
(0 to -20) dB	1.8 dB		
AM Depth > 5%			
(0 to -20) dB	1.5 dB		
(-20 to -30) dB	3.5 dB		
FM Distortion Measure <sup>1</sup> Rate 20 Hz to 1 kHz	(1 to 6 600) MHz		Agilent N5531S Measuring Receiver
	Dev 500 Hz to 2 kHz		
	(0 to -20) dB	0.26 dB	
	(-20 to -30) dB	0.79 dB	
	(-30 to -40) dB	2.3 dB	
	Dev > 2 kHz		
	(0 to -20) dB	0.09 dB	
	(-20 to -30) dB	0.27 dB	
	(-30 to -40) dB	0.83 dB	
	(-40 to -50) dB	2.4 dB	
	(6.6 to 13.2) GHz		
	Dev > 2.3 kHz		
	(0 to -20) dB	0.26 dB	
(-20 to -30) dB	0.79 dB		
(-30 to -40) dB	2.3 dB		

Electrical – RF/Microwave

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
FM Distortion Measure <sup>1</sup> Rate 20 Hz to 1 kHz	(6.6 to 13.2) GHz Dev > 4.5 K kHz Hz		Agilent N5531S Measuring Receiver
	(0 to -20) dB	0.09 dB	
	(-20 to -30) dB	0.27 dB	
	(-30 to -40) dB	0.83 dB	
	(-40 to -50) dB	2.4 dB	
	(13.2 to 31.15) GHz Dev > 2.7 kHz		
	(0 to -20) dB	0.26 dB	
	(-20 to -30) dB	0.79 dB	
	(-30 to -40) dB	2.3 dB	
	Dev > 6.0 kHz		
	(0 to -20) dB	0.09 dB	
	(-20 to -30) dB	0.27 dB	
	(-30 to -40) dB	0.83 dB	
	(-40 to -50) dB	2.4 dB	
	(31.15 to 50.0) GHz Dev > 4 kHz		
	(0 to -20) dB	0.26 dB	
(-20 to -30) dB	0.79 dB		
(-30 to -40) dB	2.3 dB		
(31.15 to 50.0) GHz Dev > 12.0 kHz			
(0 to -20) dB	0.09 dB		
(-20 to -30) dB	0.27 dB		
(-30 to -40) dB	0.83 dB		
(-40 to -50) dB	2.4 dB		
PM Distortion Measure <sup>1</sup>	(1 to 6 600) MHz Rate (20 to 500) Hz Dev > 0.8 rad		Agilent N5531S Measuring Receiver
	(0 to -20) dB	0.26 dB	
	(-20 to -30) dB	0.79 dB	
	(-30 to -40) dB	2.3 dB	
	Dev > 2.5 rad		
	(0 to -20) dB	0.09 dB	
	(-20 to -30) dB	0.27 dB	
	(-30 to -40) dB	0.83 dB	
(-40 to -50) dB	2.3 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment	
PM Distortion Measure <sup>1</sup>	(1 to 6 600) MHz			
	Rate (500 to 1 000) Hz			
	Dev > 0.4 rad		0.26 dB	Agilent N5531S Measuring Receiver
	(0 to -20) dB		0.79 dB	
	(-20 to -30) dB		2.3 dB	
	(-30 to -40) dB			
	Dev > 1.0 rad			
	(0 to -20) dB		0.09 dB	
	(-20 to -30) dB		0.27 dB	
	(-30 to -40) dB		0.83 dB	
	(-40 to -50) dB		2.3 dB	
	(6.6 to 13.2) GHz			
	Rate (20 to 500) Hz			
	Dev > 1.8 rad		0.26 dB	
	(0 to -20) dB		0.79 dB	
	(-20 to -30) dB		2.3 dB	
	(-30 to -40) dB			
	Dev > 5.5 rad			
	(0 to -20) dB		0.09 dB	
	(-20 to -30) dB		0.27 dB	
	(-30 to -40) dB		0.83 dB	
	(-40 to -50) dB		2.3 dB	
	(6.6 to 13.2) GHz			
	Rate (500 to 1 000) Hz			
Dev > 0.8 rad		0.26 dB		
(0 to -20) dB		0.79 dB		
(-20 to -30) dB		2.3 dB		
(-30 to -40) dB				
Dev > 2.5 rad				
(0 to -20) dB		0.09 dB		
(-20 to -30) dB		0.27 dB		
(-30 to -40) dB		0.83 dB		
(-40 to -50) dB		2.3 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
PM Distortion Measure <sup>1</sup>	(13.2 to 31.15) GHz		Agilent N5531S Measuring Receiver
	Rate (20 to 500) Hz		
	Dev > 3.5 rad		
	(0 to -20) dB	0.26 dB	
	(-20 to -30) dB	0.79 dB	
	(-30 to -40) dB	2.3 dB	
	Dev > 10.0 rad		
	(0 to -20) dB	0.09 dB	
	(-20 to -30) dB	0.27 dB	
	(-30 to -40) dB	0.83 dB	
	(-40 to -50) dB	2.3 dB	
	Rate (500 to 1 000) Hz		
	Dev > 1.2 rad		
	(0 to -20) dB	0.26 dB	
(-20 to -30) dB	0.79 dB		
(-30 to -40) dB	2.3 dB		
Dev > 4.0 rad			
(0 to -20) dB	0.09 dB		
(-20 to -30) dB	0.27 dB		
(-30 to -40) dB	0.83 dB		
(-40 to -50) dB	2.3 dB		
Harmonics Measure <sup>1</sup>	(-10 to -80) dB		Agilent E4448A Measuring Receiver
	2 <sup>nd</sup> through 5 <sup>th</sup> Harmonic		
	1kHz to 600MHz	0.37 dB	
	(600 to 1 320) MHz	1.1 dB	
	(1 320 to 2 200) MHz	1.4 dB	
	(2 200 to 3 000) MHz	1.4 dB	
	(3 000 to 4 400) MHz	1.7 dB	
	(4 400 to 5 300) MHz	1.9 dB	
	(5 300 to 10 000) MHz	2.1 dB	
	2 <sup>nd</sup> through 4 <sup>th</sup> Harmonic		
(10 000 to 12 500) MHz	2.1 dB		
2 <sup>nd</sup> through 3 <sup>rd</sup> Harmonic			
(12 500 to 16 667) MHz	2.1 dB		
2 <sup>nd</sup> Harmonic			
(16 667 to 25 000) MHz	2.3 dB		
Attenuation Generate @ 30 MHz	10 dB	5.6 mdB	HP 11812A Calibration Kit
	20 dB	7.6 mdB	
	30 dB	6.4 mdB	
	40 dB	7.4 mdB	
	50 dB	8.6 mdB	

**Electrical – RF/Microwave**

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Measure <sup>1</sup> 50 MHz	1.0 mW	0.37 % of reading	HP 478A/Opt H76 Power Meter
RF Power – Generate	10 Hz to 100 kHz (24 to -48) dBm 100 kHz to 9.99 MHz (-48 to 24) dBm (-48 to -74) dBm (-74 to -94) dBm (10 MHz to 128) MHz (24 to -48) dBm (-48 to -84) dBm (-84 to -94) dBm (-94 to -124) dBm (128 to 300) MHz (20 to -48) dBm (-48 to -74) dBm (-74 to -84) dBm (-84 to -94) dBm (-94 to -124) dBm (300 to 1 400) MHz (20 to -48) dBm (-48 to -74) dBm (-74 to -84) dBm (-84 to -94) dBm (-94 to -124) dBm (1.4 to 3.0) GHz (14 to -48) dBm (-48 to -74) dBm (-74 to -94) dBm (-94 to -124) dBm	0.06 dB 0.07 dB 0.16 dB 0.39 dB 0.07 dB 0.09 dB 0.24 dB 0.54 dB 0.08 dB 0.09 dB 0.24 dB 0.39 dB 1.2 dB 0.16 dB 0.31 dB 0.39 dB 0.77 dB 1.2 dB 0.24 dB 0.39 dB 0.77 dB 1.2 dB	Fluke 9640A-LPNX RF Reference Source
RF Power – Generate <sup>1</sup>	(30 to 2 000) MHz (20 to 0) dB (0 to -58) dB (-58 to -78) dB (-78 to -110) dB	0.29 dB 0.3 dB 0.31 dB 0.32 dB	Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor, E8257D Signal Generator

**Electrical – RF/Microwave**

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Generate <sup>1</sup>	(2 000 to 3 050) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor, E8257D Signal Generator
	(20 to 0) dB	0.34 dB	
	(0 to -58) dB	0.35 dB	
	(-58 to -78) dB	0.36 dB	
	(-78 to -110) dB	0.39 dB	
	(3 050 to 6 600) MHz		
	(20 to 0) dB	0.34 dB	
	(0 to -58) dB	0.35 dB	
	(-58 to -78) dB	0.36 dB	
	(-78 to -110) dB	0.39 dB	
	(6 600 to 13 200) MHz		
	(20 to 0) dB	0.34 dB	
	(0 to -58) dB	0.35 dB	
	(-58 to -78) dB	0.36 dB	
	(-78 to -110) dB	0.42 dB	
	(13 200 to 18 000) MHz		
	(15 to 0) dB	0.34 dB	
	(0 to -58) dB	0.35 dB	
	(-58 to -78) dB	0.36 dB	
	(-78 to -110) dB	0.46 dB	
	(18 000 to 19 200) MHz		
	(15 to 0) dB	0.41 dB	
	(0 to -58) dB	0.41 dB	
	(-58 to -78) dB	0.42 dB	
	(-78 to -110) dB	0.5 dB	
	(19 200 to 26 500) MHz		
	(15 to 0) dB	0.41 dB	
	(0 to -58) dB	0.41 dB	
(-58 to -78) dB	0.42 dB		
(-78 to -110) dB	0.9 dB		
(26 500 to 31 150) MHz			
(15 to 0) dB	0.62 dB		
(0 to -58) dB	0.63 dB		
(-58 to -78) dB	0.64 dB		
(-78 to -110) dB	0.96 dB		
(31 150 to 41 000) MHz			
(10 to 0) dB	0.82 dB		
(0 to -58) dB	0.83 dB		
(-58 to -78) dB	0.84 dB		
(-78 to -100) dB	1.1 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Generate <sup>1</sup>	(41 000 to 45 000) MHz (10 to 0) dB (0 to -58) dB (-58 to -78) dB (-78 to -100) dB (45 000 to 50 000) MHz (10 to 0) dB (0 to -58) dB (-58 to -78) dB (-78 to -90) dB	0.82 dB 0.83 dB 0.85 dB 1.4 dB 0.82 dB 0.83 dB 1 dB 1.5 dB	Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor, E8257D Signal Generator
RF Power – Power Sensor (Cal Factor)	Type-N (50 Ω) 0.1 MHz 0.3 MHz 0.5 MHz 1 MHz 3 MHz 5 MHz 10 MHz 30 MHz 50 MHz 100 MHz 300 MHz 500 MHz 1 000 MHz 1 500 MHz 2 000 MHz 2 500 MHz 3 000 MHz 3 500 MHz 3 700 MHz 4 000 MHz 4 200 MHz	1.1 % of reading + M 0.94 % of reading + M 0.93 % of reading + M 0.93 % of reading + M 0.9 % of reading + M 0.9 % of reading + M 0.89 % of reading + M 0.89 % of reading + M 0.91 % of reading + M 0.89 % of reading + M 0.94 % of reading + M 0.94 % of reading + M 0.96 % of reading + M 0.96 % of reading + M 0.96 % of reading + M 1.1 % of reading + M 1.2 % of reading + M 1.2 % of reading + M 1.2 % of reading + M 1.4 % of reading + M 1.4 % of reading + M	M is the mismatch Uncertainty Agilent EPM Series Power Meter RF Power / Keysight 8482A H84 Power Sensor



ANSI National Accreditation Board

Electrical – RF/Microwave

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Power Sensor (Cal Factor)	Type-N (50 Ω)		M is the mismatch Uncertainty  Agilent EPM Series Power Meter RF Power / Keysight 8481A H84 Power Sensor
	100 MHz	1.2 % of reading + M	
	500 MHz	1.2 % of reading + M	
	1 000 MHz	1.2 % of reading + M	
	2 000 MHz	1.2 % of reading + M	
	3 000 MHz	1.3 % of reading + M	
	4 000 MHz	1.3 % of reading + M	
	5 000 MHz	1.4 % of reading + M	
	6 000 MHz	1.4 % of reading + M	
	7 000 MHz	1.4 % of reading + M	
	8 000 MHz	1.4 % of reading + M	
	9 000 MHz	1.5 % of reading + M	
	10 000 MHz	1.5 % of reading + M	
	11 000 MHz	1.5 % of reading + M	
	12 000 MHz	1.5 % of reading + M	
	13 000 MHz	1.7 % of reading + M	
	14 000 MHz	1.7 % of reading + M	
15 000 MHz	1.7 % of reading + M		
16 000 MHz	1.8 % of reading + M		
17 000 MHz	1.8 % of reading + M		
18 000 MHz	1.9 % of reading + M		
RF Power – Power Sensor (Cal Factor)	3.5 mm (50Ω)		M is the mismatch Uncertainty  Agilent EPM Series Power Meter RF Power / Keysight 8485A H84 Power Sensor
	50 MHz	1.3 % of reading + M	
	100 MHz	1.2 % of reading + M	
	300 MHz	1.2 % of reading + M	
	500 MHz	1.3 % of reading + M	
	1 000 MHz	1.3 % of reading + M	
	1 500 MHz	1.3 % of reading + M	
	2 000 MHz	1.2 % of reading + M	
	3 000 MHz	1.3 % of reading + M	
	4 000 MHz	1.3 % of reading + M	
	5 000 MHz	1.3 % of reading + M	
6 000 MHz	1.4 % of reading + M		
7 000 MHz	1.4 % of reading + M		



ANSI National Accreditation Board

Electrical – RF/Microwave

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Power Sensor (Cal Factor)	3.5 mm (50Ω)		<p>M is the mismatch Uncertainty</p> <p>Agilent EPM Series Power Meter RF Power / Keysight 8485A H84 Power Sensor</p>
	8 000 MHz	1.5 % of reading + M	
	9 000 MHz	1.6 % of reading + M	
	10 000 MHz	1.6 % of reading + M	
	11 000 MHz	1.5 % of reading + M	
	12 000 MHz	1.5 % of reading + M	
	12 400 MHz	1.6 % of reading + M	
	13 000 MHz	1.6 % of reading + M	
	14 000 MHz	1.9 % of reading + M	
	15 000 MHz	1.7 % of reading + M	
	16 000 MHz	1.6 % of reading + M	
	17 000 MHz	1.8 % of reading + M	
	18 000 MHz	1.7 % of reading + M	
	19 000 MHz	2 % of reading + M	
	20 000 MHz	2.2 % of reading + M	
	21 000 MHz	2.1 % of reading + M	
	22 000 MHz	2.5 % of reading + M	
23 000 MHz	2.7 % of reading + M		
24 000 MHz	2.3 % of reading + M		
25 000 MHz	2 % of reading + M		
26 000 MHz	2 % of reading + M		
26 500 MHz	2.3 % of reading + M		
RF Power – Power Sensor (Cal Factor)	2.4 mm (50 Ω)		<p>M is the mismatch Uncertainty</p> <p>Agilent EPM Series Power Meter RF Power / Keysight N8487A H84 Power Sensor</p>
	50 MHz	1.1 % of reading + M	
	100 MHz	1.1 % of reading + M	
	300 MHz	1.1 % of reading + M	
	500 MHz	1.2 % of reading + M	
	1 000 MHz	1.2 % of reading + M	
	2 000 MHz	1.2 % of reading + M	
	3 000 MHz	1.2 % of reading + M	
	4 000 MHz	1.2 % of reading + M	
	5 000 MHz	1.3 % of reading + M	
	6 000 MHz	1.3 % of reading + M	
	7 000 MHz	1.3 % of reading + M	
	8 000 MHz	1.4 % of reading + M	
	9 000 MHz	1.5 % of reading + M	
	10 000 MHz	1.5 % of reading + M	
11 000 MHz	1.5 % of reading + M		
12 000 MHz	1.5 % of reading + M		
13 000 MHz	1.5 % of reading + M		
14 000 MHz	1.6 % of reading + M		



ANSI National Accreditation Board

Electrical – RF/Microwave

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Power Sensor (Cal Factor)	2.4 mm (50 Ω)		<p>M is the mismatch Uncertainty</p> <p>Agilent EPM Series Power Meter RF Power / Keysight N8487A H84 Power Sensor</p>
	15 000 MHz	1.6 % of reading + M	
	16 000 MHz	1.7 % of reading + M	
	17 000 MHz	1.7 % of reading + M	
	18 000 MHz	1.7 % of reading + M	
	19 000 MHz	1.9 % of reading + M	
	20 000 MHz	1.9 % of reading + M	
	21 000 MHz	1.9 % of reading + M	
	22 000 MHz	1.9 % of reading + M	
	23 000 MHz	1.9 % of reading + M	
	24 000 MHz	1.9 % of reading + M	
	25 000 MHz	1.9 % of reading + M	
	26 000 MHz	1.9 % of reading + M	
	27 000 MHz	2.4 % of reading + M	
	28 000 MHz	2.4 % of reading + M	
	29 000 MHz	2.4 % of reading + M	
	30 000 MHz	2.4 % of reading + M	
	31 000 MHz	2.4 % of reading + M	
	32 000 MHz	2.4 % of reading + M	
	33 000 MHz	2.4 % of reading + M	
	34 000 MHz	2.5 % of reading + M	
	34 500 MHz	2.5 % of reading + M	
	35 000 MHz	2.5 % of reading + M	
	36 000 MHz	2.6 % of reading + M	
	37 000 MHz	2.6 % of reading + M	
	38 000 MHz	2.6 % of reading + M	
	39 000 MHz	2.9 % of reading + M	
	40 000 MHz	3 % of reading + M	
	41 000 MHz	3.2 % of reading + M	
	42 000 MHz	3.2 % of reading + M	
	43 000 MHz	3.2 % of reading + M	
44 000 MHz	3.2 % of reading + M		
45 000 MHz	3.2 % of reading + M		
46 000 MHz	3.2 % of reading + M		
47 000 MHz	3.2 % of reading + M		
48 000 MHz	3.3 % of reading + M		
49 000 MHz	3.4 % of reading + M		
50 000 MHz	3.4 % of reading + M		
Phase Noise <sup>1</sup> Offset Frequency 5 MHz < f ≤ 18 GHz	≤ 100 kHz 100 kHz to 40 MHz	2.3 dB 4.6 dB	HP 3048A Phase Noise System with 866XA Source

**Electrical – RF/Microwave**

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Total Harmonic Distortion (THD)	(0 to -60) dB		HP 8903B Audio Analyzer
	20 Hz to 20 kHz		
	(0 to -40) dB	1 dB	
	(-40 to -50) dB	1 dB	
	(-50 to -60) dB	1.3 dB	
	(-60 to -65) dB	1.7 dB	
	(20 to 50) kHz		
	(0 to -40) dB	2 dB	
	(-40 to -50) dB	2.1 dB	
	(-50 to -60) dB	3 dB	
	(50 to 100) kHz		
	(0 to -40) dB	2 dB	
	(-40 to -50) dB	2.4 dB	
Return Loss (VSWR)			HP 8751A, HP 85107B Network Analyzers
	5 Hz to 45 MHz	1.1 dB	
	(0 to 120) dB	0.41 dB	
	(0 to 120) dB		

**Length – Dimensional Metrology**

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Gage Blocks <sup>2</sup>	Up to 20 in	(3.2 + 1.2L) μin	Master gage blocks, P&W universal measuring machine
Micrometers <sup>1,2</sup>	Up to 40 in	(30 + 4.8L) μin	Gage blocks (grade 2)
Micrometer Standards Length Rods	Up to 40 in	(4.3 + 3.8L) μin	Gage blocks (grade 2), P&W universal measuring machine
Calipers <sup>1,2</sup>	Up to 40 in	(280 + 2L) μin	Gage blocks (grade 2)
Dial Indicators <sup>1,2</sup> Resolution: ≥ 50μin < 50μin	Up to 10 in	(26 + 3.3L) μin	Gage blocks (grade 2)
	Up to 0.1 in	7.7 μin	
Height Gages <sup>1,2</sup>	Up to 40 in	(114 + 3.3L) μin	Gage blocks (grade 2)
Protractors <sup>1</sup>	(0 to 360) °	0.008 6°	Angle blocks
Rulers <sup>1</sup>	Up to 46 in	0.009 in	Gage blocks (grade 2)

**Length – Dimensional Metrology**

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Radius Gages	(0.016 to 1) in	220 μin	Optical Comparator
Tape Measures <sup>2</sup>	Up to 100 ft	(0.000 26 <i>F</i> + 0.025) in	Standard rule
Bore Micrometers <sup>2</sup> 2 point 3 point	Up to 12 in	(22 + 3.6 <i>L</i> ) μin (8.8 + 42 <i>L</i> ) μin	Master gage blocks, P&W universal measuring machine, Master Ring
Feeler Gage <sup>1</sup>	Up to 1 in	30 μin	Pratt & Whitney Supermicrometer C
Cylindrical Gages <sup>1,2</sup> – Plain Pin, Plugs, Rings	(0.04 to 13) in	(5.3 + 2.3 <i>D</i> ) μin	Master gage blocks, P&W universal measuring machine
Cylindrical Gages <sup>1,2</sup> – Plain Rings	(0 to 14) in	(8.1 + 2.9 <i>D</i> ) μin	
Thread Plugs <sup>1</sup> – Major Diameter Pitch Diameter	Up to 12 in Up to 12 in	36 μin 92 μin	B & S 599-246-00, Van Keuren thread wire set Gage blocks, Pratt & Whitney Labmaster
Solid Thread Rings Pitch Diameter	Up to 12 in	107 μin	Pratt & Whitney Labmaster Measuring Machine
Adjustable Thread Rings <sup>2,3</sup> Pitch Diameter (Tactile Fit)	Up to 12 in	(350 + 47 <i>D</i> ) μin	Thread setting plug gages
Thread Wires <sup>2</sup>	Up to 0.5 <i>D</i>	10 μin	Master gage blocks, P&W universal measuring machine
Surface Plates <sup>1</sup> – Overall Flatness Local Area Flatness	Up to 6 ft × 6 ft (-0.001 to 0.001) in	125 μin 68 μin	Planekator Repeat-O-Meter
Optical Comparators <sup>1</sup> – Angle Linearity	Up to 360 ° Up to 20 in (20 to 40) in	0.008 7 ° 210 μin 450 μin	Gage blocks, Angle Blocks, SI Industries glass scales
Magnification	(10 to 100) x	340 μin	
Coating Thickness Gages <sup>1,2</sup> Eddy Current & Magnetic Induction	(0.737 to 100) mils (100 to 243) mils	26 μin 240 μin	Coating thickness standards
Ultrasonic Thickness Gauges <sup>1</sup>	Up to 10 in	110 μin	Gage blocks



ANSI National Accreditation Board

**Length – Dimensional Metrology**

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Coating Thickness Shims <sup>2</sup>	(0 to 243) mils	56 μin	Pratt & Whitney Supermicrometer C

**Mass and Mass Related**

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Scales & Balances <sup>1,2,4</sup>	1 mg to 5 000 g (0.001 to 100) lb	(0.049 + 0.003 <i>X</i> ) mg (3.2 E <sup>-6</sup> + 3.1 E <sup>-6</sup> <i>W</i> ) lb	Class 1 weights
	Up to 1 000 lb	(0.000 2 + 0.000 12 <i>W</i> ) lb	Class F weights
Force <sup>1</sup> Tension and Compression	(0.5 to 750) lbf	0.09 % of reading	Class F weights
	(20 to 1 000) lbf (200 to 10 000) lbf (10 000 to 25 000) lbf (25 000 to 50 000) lbf	The greater of: 0.012 % of reading or 0.023 lbf 0.012 % of reading or 0.24 lbf 0.012 % of reading or 0.5 lbf 0.012 % of reading or 0.77 lbf	Morehouse Press with Load Cells
	(0 to 24) inH <sub>2</sub> O	0.001 5 inH <sub>2</sub> O	Dwyer 1425-25 Hook Gage
Pressure <sup>1</sup>	(-15 to 200) psi (0 to 1 000) psi	The greater of: 0.013 % of reading or 0.003 9% of reading of PPC4 Span	Fluke PPC4 Pressure Controller
	(10 to 16 000) psi	0.019 % of reading	Fluke P3125-PSI Dead Weight Tester
Pressure <sup>1</sup>	(16 000 to 40 000) psi	25 psi	Additel ADT672-10-GP40K Pressure Calibrator
Gas Flow	(1 to 10 000) sccm (10 to 1 000) slpm	0.2 % of reading 0.26 % of reading	Molbloc flow standards
Pipette <sup>2</sup>	(10 to 100) μL (100 to 1 000) μL (1 to 10) mL	(0.58 + 0.004 <i>V</i> ) μL (0.6 + 0.001 <i>V</i> ) μL (2.6 + 0.001 2 <i>V</i> ) μL	Mass Balance
Torque Tools <sup>1</sup>	(10 to 100) ozf-in	0.59 % of reading	CDI 1001 Torque Tester
	64 ozf-in to 1 000 lbf-ft	0.32 % of reading	Analyzer / CDI 5000 ST Torque Analyzer
Torque Analyzers <sup>1</sup>	1 lbf-in to 1 000 lbf-ft	0.075 % of reading	Torque Arm, Class F Weights



ANSI National Accreditation Board

Mass and Mass Related

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Rockwell Hardness Testers <sup>1</sup>	(< 60) HRBW (≥ 60 to < 80) HRBW (≥ 80) HRBW  (< 35) HRC (≥ 35 to < 60) HRC (≥ 60) HRC  (< 84) HREW (≥ 84 to < 93) HREW (≥ 93) HREW  (< 81) HR15TW (≥ 81 to < 87) HR15TW (> 87) HR15TW	3 HRBW 3 HRBW 1.3 HRBW  1.3 HRC 1.2 HRC 0.7 HRC  1.3 HREW 1.3 HREW 1.3 HREW  1.8 HR15TW 1.3 HR15TW 1.3 HR15TW	Indirect verification per ASTM E18
Durometers Scale (Force) Accuracy Types A, B, C, D, DO, O Type M Types O, OO  Indenter Geometry Length Diameter Angle	(0 to 100) duros  0.1 in 0.05 in (30 to 35)°	0.06 duros 0.07 duros 0.08 duros  130 μin 130 μin 0.12°	Direct Verification  Master balance  Optical comparator
Mass - Fixed Points Metric	(1, 2, 5, 10) mg (20, 50, 100, 500) mg (1, 2, 5) g 10 g 20 g 50 g 100 g 200 g 500 g 1 kg 5 kg	0.027 mg 0.027 mg 0.08 mg 0.13 mg 0.11 mg 0.22 mg 0.36 mg 0.64 mg 29 mg 30 mg 31 mg	Comparison to ASTM E617 Class 1 weights



ANSI National Accreditation Board

**Mass and Mass Related**

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Mass - Fixed Points Avoirdupois	(0.001, 0.002) lb	0.18 mg	Comparison to ASTM E617 Class 1 weights
	(0.005, 0.01, 0.02) lb	0.19 mg	
	0.05 lb	0.24 mg	
	0.1 lb	0.29 mg	
	0.2 lb	0.39 mg	
	(0.5, 1, 2) lb	27 mg	
	5 lb	28 mg	
	10 lb	28 mg	
	25 lb	0.23 g	
	50 lb	0.23 g	

**Thermodynamic**

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Relative Humidity Generate	(10 to 95) %RH	0.5 %RH	Thunder Scientific 2500 Humidity Chamber
Relative Humidity Measure	(0 to 90) %RH (90 to 100) %RH	1.2 %RH 2 %RH	Vaisala MI70/HMP76 Humidity Indicator and Probe
Temperature – Measure <sup>1</sup>	(-196 to 230) °C (-321 to 446) °F (230 to 420) °C (446 to 788) °F (420 to 660) °C (788 to 1 220) °F	0.023 °C 0.042 °F 0.04 °C 0.072 °F 0.056 °C 0.1 °F	Hart 1502A Indicator with 5616 & 5609 PRT
	(800 to 1 550) °C	2 °C	Type B Thermocouple
Temperature Uniformity Surveys <sup>1</sup>	Type J (100 to 900) °F Type K (100 to 1 000) °F (1 000 to 2 000) °F (2 000 to 2 200) °F	1.8 °F 2 °F 2.1 °F 2.7 °F	Datalogger and Thermocouples
Infrared (IR) Thermometry <sup>1</sup>	(20 to 100) °C (100 to 300) °C (300 to 420) °C (420 to 500) °C	1.5 °C 4.3 °C 6 °C 7.7 °C	Fluke 9132 Infrared Calibrator $\lambda = (8 \text{ to } 14) \mu\text{m}$ , $\epsilon = 0.95$



ANSI National Accreditation Board

Time and Frequency

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency – Measure	10 MHz	$1 \times 10^{-12}$ Hz/Hz	58503A/B GPS Receiver
Frequency – Generate <sup>1</sup>	1 to 10 Hz 10 to 100 Hz 100 to 1 000 Hz 1 to 10 kHz 10 to 100 kHz 0.1 to 1 MHz 1 to 10 MHz	$1 \times 10^{-12}$ Hz/Hz + 0.57 $\mu$ Hz $1 \times 10^{-12}$ Hz/Hz + 5.7 $\mu$ Hz $1 \times 10^{-12}$ Hz/Hz + 57 $\mu$ Hz $1 \times 10^{-12}$ Hz/Hz + 0.57 mHz $1 \times 10^{-12}$ Hz/Hz + 5.7 mHz $1 \times 10^{-12}$ Hz/Hz + 57 mHz $1 \times 10^{-12}$ Hz/Hz + 0.57 Hz	Agilent 33250A Function Generator / HP 58503A GPS Receiver
Frequency – Generate <sup>1</sup>	10 to 50 000 MHz	$1 \times 10^{-12}$ Hz/Hz + 0.57mHz	Agilent E8257D Opt 550 Signal Generator / HP 58503A GPS Receiver
Frequency – Measure <sup>1</sup>	1 to 10 Hz 10 to 100 Hz 100 to 1 000 Hz 1 to 10 kHz 10 to 100 kHz 100 to 200 kHz 0.2 to 3 000 MHz	$1.82 \times 10^{-9}$ Hz/Hz $0.59 \times 10^{-9}$ Hz/Hz $0.20 \times 10^{-9}$ Hz/Hz $74 \times 10^{-10}$ Hz/Hz $35 \times 10^{-10}$ Hz/Hz $23 \times 10^{-10}$ Hz/Hz $20 \times 10^{-10}$ Hz/Hz	Agilent 53132A Opt 030 Frequency Counter / HP 58503A GPS Receiver
	50 to 46 000 MHz	$1 \times 10^{-12}$ Hz/Hz + 1.6 Hz	Agilent 53152A Frequency Counter / HP 58503A GPS Receiver
	10 to 50 000 MHz	$1 \times 10^{-12}$ Hz/Hz + 0.1 Hz	Agilent E4448A Spectrum Analyzer / HP 58503A GPS Receiver
Time – Generate	1 pps	$1 \times 10^{-12}$ s/s + 750 ps	HP 58503A GPS Receiver
Type I (digital) Timers	(0 to 19.99) sec/day (0 to 599) sec/month	0.031 sec/day 1.1 sec/month	Helmut Klein Timometer 4500
Type II (mechanical) Timers	(0 to 320) sec/day	0.6 sec/day	
Tachometers – RPM <sup>1,2</sup>	Up to 100 000 RPM	0.001 % of reading + 0.6R	HP 33250A Signal Generator & LED

## DIMENSIONAL MEASUREMENT

### 1 Dimensional

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Length	X Axis (0.01 to 10) in Y Axis (0.01 to 6) in	162 $\mu$ in 123 $\mu$ in	Optical comparator

### 2 Dimensional

Temple Terrace, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Angle	Up to 360 °	0.004 °	Optical comparator

[Return to Site listing](#) (top)    [Go to Notes](#) (bottom)





ANSI National Accreditation Board

### Services performed at satellite laboratory

#### Technical Maintenance, Inc.

3000 Northwoods Parkway, Suite 270

Peachtree Corners, GA 30071

Milt Mosher (Branch Manager) Phone: 770-409-8348

Scott Chamberlain (Quality Manager) Phone: 321-242-0890

## CALIBRATION AND DIMENSIONAL MEASUREMENT

### CALIBRATION

#### Acoustics and Vibration

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Accelerometers – Acceleration <sup>1</sup>	(0.01 to 10) g (7 < 10) Hz (10 < 30) Hz (30 < 2 000) Hz (2 to 10) kHz	4 % of reading 3 % of reading 1.5 % of reading 4 % of reading	The Modal Shop 9100D Portable Vibration Calibrator
Sound Pressure Level – Generate <sup>1</sup>	114dB @ 251.2 Hz	0.26 dB	Larson Davis CAL250 Sound Level Calibrator

#### Chemical Quantities

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
pH Meters <sup>1,5</sup>	(4, 7, 10) pH	0.02 pH	Standard pH buffers

#### Electrical – DC/Low Frequency

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Voltage – Generate <sup>1</sup>	(0 to 220) mV (0.22 to 2.2) V (2.2 to 11) V (11 to 22) V (22 to 220) V (220 to 1 100) V	6.8 μV/V + 0.8 μV 4.6 μV/V + 0.9 μV 3 μV/V + 2.5 μV 3 μV/V + 3.9 μV 4.6 μV/V + 38 μV 6.1 μV/V + 385 μV	Fluke 5720A Multifunction Calibrator

**Electrical – DC/Low Frequency**

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Voltage – Measure <sup>1</sup>	Up to 100 mV (0.1 to 1) V (1 to 10) V (10 to 100) V (100 to 1 000) V	3.3 $\mu\text{V/V} + 1 \mu\text{V}$ 2.6 $\mu\text{V/V} + 1 \mu\text{V}$ 2.6 $\mu\text{V/V} + 1.5 \mu\text{V}$ 3.9 $\mu\text{V/V} + 20 \mu\text{V}$ 3.9 $\mu\text{V/V} + 66 \mu\text{V} + 12 \mu\text{V/V} \times (\text{Vin}/1\ 000)^2$	HP 3458A Opt 002 Multimeter
DC Voltage – Measure <sup>1</sup>	(1 to 60) kV	0.1 % of reading	Ross VD60 High Voltage Divider, HP 34401A Multimeter
DC Current – Generate <sup>1</sup>	(1 to 2.2) nA (2.2 to 22) nA (22 to 220) nA (0.22 to 2.2) $\mu\text{A}$ (2.2 to 10) $\mu\text{A}$	93 $\mu\text{A/A} + 0.007 \text{ nA}$ 92 $\mu\text{A/A} + 0.007 \text{ nA}$ 92 $\mu\text{A/A} + 0.01 \text{ nA}$ 36 $\mu\text{A/A} + 0.1 \text{ nA}$ 20 $\mu\text{A/A} + 1 \text{ nA}$	Fluke 5720A Multifunction Calibrator & Fluke 5522A Multifunction Calibrator
DC Current – Generate <sup>1</sup>	(10 to 220) $\mu\text{A}$ (0.22 to 2.2) mA (2.2 to 22) mA (22 to 100) mA (100 to 220) mA (0.22 to 1) A (1 to 2.2) A	38 $\mu\text{A/A} + 5 \text{ nA}$ 30 $\mu\text{A/A} + 7 \text{ nA}$ 30 $\mu\text{A/A} + 44 \text{ nA}$ 38 $\mu\text{A/A} + 0.7 \mu\text{A}$ 45 $\mu\text{A/A} + 0.7 \mu\text{A}$ 68 $\mu\text{A/A} + 12 \mu\text{A}$ 105 $\mu\text{A/A} + 12 \mu\text{A}$	Fluke 5720A Multifunction Calibrator
DC Current – Generate <sup>1</sup>	(2.2 to 11) A	274 $\mu\text{A/A} + 365 \mu\text{A}$	Fluke 5720A Multifunction Calibrator, 5725A Amplifier
DC Current – Generate <sup>1</sup>	(11 to 20.5) A	761 $\mu\text{A/A} + 578 \mu\text{A}$	Fluke 5522A Multifunction Calibrator
DC Current Clamp Meters <sup>1</sup> Toroidal-Wound	(0 to 1 025) A	0.21 % of output + 0.05 A	
DC Current Clamp Meters <sup>1</sup> Other	(0 to 1 025) A	0.39 % of output + 0.38 A	Fluke 5522A Multiproduct Calibrator / Coil x50
DC Current – Measure <sup>1</sup>	(1 to 10) nA (10 to 100) nA (0.1 to 1) $\mu\text{A}$ (1 to 10) $\mu\text{A}$	35 $\mu\text{A/A} + 0.12 \text{ pA}$ 18 $\mu\text{A/A} + 1.2 \text{ pA}$ 11 $\mu\text{A/A} + 0.01 \text{ nA}$ 9.1 $\mu\text{A/A} + 0.12 \text{ nA}$	Fluke 5720A Multifunction Calibrator, Agilent 3458A Multimeter Option 002
DC Current – Measure <sup>1</sup>	(10 to 100) $\mu\text{A}$ (0.1 to 1) mA 1 to 10 mA (10 to 100) mA (0.1 to 1) A	20 $\mu\text{A/A} + 0.8 \text{ nA}$ 20 $\mu\text{A/A} + 5 \text{ nA}$ 20 $\mu\text{A/A} + 0.05 \mu\text{A}$ 35 $\mu\text{A/A} + 0.5 \mu\text{A}$ 110 $\mu\text{A/A} + 10.1 \mu\text{A}$	Agilent 3458A Multimeter



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Current – Measure <sup>1</sup>	(1 to 10) A (10 to 100) A	41 $\mu$ A/A + 103 $\mu$ A 53 $\mu$ A/A + 103 $\mu$ A	Agilent 3458A Option 002 Precision Multimeter Standard Resistor L&N 4361 Current Shunt
DC Current – Measure <sup>1</sup>	(1 to 500) A	0.27 % of reading	Current shunts
Resistance – Fixed Points <sup>1</sup>	(0.01, 0.1, 1.0) $\Omega$	34 $\mu\Omega/\Omega$	Leeds & Northrup Resistor Set
	(1, 1.9) $\Omega$	84 $\mu\Omega/\Omega$ + 0.1 $\mu\Omega$	Fluke 5720A Multifunction Calibrator
	(10, 19) $\Omega$	21 $\mu\Omega/\Omega$ + 1 $\mu\Omega$	
	(100, 190) $\Omega$	9.1 $\mu\Omega/\Omega$ + 6 $\mu\Omega$	
	(1, 1.9) k $\Omega$	7.6 $\mu\Omega/\Omega$ + 60 $\mu\Omega$	
	(10, 19) k $\Omega$	7.6 $\mu\Omega/\Omega$ + 0.6 m $\Omega$	
	(100, 190) k $\Omega$	9.9 $\mu\Omega/\Omega$ + 6 m $\Omega$	
	1 M $\Omega$	17.5 $\mu\Omega/\Omega$ + 60 m $\Omega$	
	1.9 M $\Omega$	18.3 $\mu\Omega/\Omega$ + 60 m $\Omega$	
	10 M $\Omega$	35 $\mu\Omega/\Omega$ + 0.6 $\Omega$	
19 M $\Omega$	42 $\mu\Omega/\Omega$ + 0.6 $\Omega$		
100 M $\Omega$	91 $\mu\Omega/\Omega$ + 6 $\Omega$		
Resistance – Generate <sup>1</sup>	Up to 11 $\Omega$	30 $\mu\Omega/\Omega$ + 0.001 $\Omega$	Fluke 5522A Multiproduct Calibrator
	(11 to 33) $\Omega$	23 $\mu\Omega/\Omega$ + 0.001 $\Omega$	
	(33 to 110) $\Omega$	21 $\mu\Omega/\Omega$ + 0.001 $\Omega$	
	(110 to 330) $\Omega$	21 $\mu\Omega/\Omega$ + 0.002 $\Omega$	
	(0.33 to 1.1) k $\Omega$	21 $\mu\Omega/\Omega$ + 0.002 $\Omega$	
	(1.1 to 3.3) k $\Omega$	21 $\mu\Omega/\Omega$ + 0.02 $\Omega$	
	(3.3 to 11) k $\Omega$	21 $\mu\Omega/\Omega$ + 0.02 $\Omega$	
	(11 to 33) k $\Omega$	21 $\mu\Omega/\Omega$ + 0.2 $\Omega$	
	(33 to 110) k $\Omega$	21 $\mu\Omega/\Omega$ + 0.2 $\Omega$	
	(110 to 330) k $\Omega$	24 $\mu\Omega/\Omega$ + 2 $\Omega$	
	(0.33 to 1.1) M $\Omega$	24 $\mu\Omega/\Omega$ + 2 $\Omega$	
	(1.1 to 3.3) M $\Omega$	46 $\mu\Omega/\Omega$ + 23 $\Omega$	
	(3.3 to 11) M $\Omega$	99 $\mu\Omega/\Omega$ + 38 $\Omega$	
	(11 to 33) M $\Omega$	190 $\mu\Omega/\Omega$ + 1.9 k $\Omega$	
	(33 to 110) M $\Omega$	380 $\mu\Omega/\Omega$ + 2.3 k $\Omega$	
	(110 to 330) M $\Omega$	0.23 % of reading + 76 k $\Omega$	
(0.33 to 1.1) G $\Omega$	1.1 % of reading + 380 k $\Omega$		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Resistance – Measure <sup>1</sup>	Up to 12 Ω (10 to 120) Ω (0.1 to 1.2 kΩ 1 to 12 kΩ (10 to 120) kΩ (0.1 to 1.2) MΩ (1 to 12) MΩ (10 to 120) MΩ (0.1 to 1.2) GΩ	15 μΩ/Ω + 56 μΩ 12 μΩ/Ω + 0.5 mΩ 10 μΩ/Ω + 0.6 mΩ 10 μΩ/Ω + 5.6 mΩ 10 μΩ/Ω + 56 mΩ 15 μΩ/Ω + 2.2 Ω 50 μΩ/Ω + 120 Ω 500 μΩ/Ω + 1.2 kΩ 0.5 % of reading + 70 kΩ	HP 3458A Multimeter
Capacitance – Generate <sup>1</sup>	(220 to 400) pF (0.4 to 3.299 9) nF (3.3 to 10.999 9) nF (11 to 32.999 9) nF (33 to 109.999) nF (110 to 329.999) nF (0.33 to 1.099 99) μF (1.1 to 3.299 99) μF (3.3 to 10.999 9) μF (11 to 32.999 9) μF (33 to 109.999) μF (110 to 329.999) μF (0.33 to 1.099 99) mF (1.1 to 3.299 99) mF (3.3 to 10.999 9) mF (11 to 32.999 9) mF (33 to 110) mF	0.38% of output + 7.6 pF 0.38 % of output + 0.01 nF 0.19 % of output + 0.01 nF 0.19 % of output + 0.08 nF 0.19 % of output + 0.08 nF 0.19 % of output + 0.23 nF 0.19 % of output + 0.76 nF 0.19 % of output + 2.3 nF 0.19 % of output + 7.6 nF 0.3 % of output + 23 nF 0.34 % of output + 76 nF 0.34 % of output + 228 nF 0.34 % of output + 0.76 μF 0.34 % of output + 2.3 μF 0.34 % of output + 7.6 μF 0.57 % of output + 23 μF 0.84 % of output + 76 μF	Fluke 5522A Multiproduct Calibrator
Capacitance – Generate <sup>1</sup> Fixed Points	1 pF 1 kHz 1 MHz 2 MHz 3 MHz 4 MHz 5 MHz 10 MHz 13 MHz	0.028 % of output 0.028 % of output 0.034 % of output 0.047 % of output 0.069 % of output 0.096 % of output 0.31 % of output 0.5 % of output	Agilent 16381A Standard Air Capacitor



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Capacitance – Generate <sup>1</sup> Fixed Points	10 pF		Agilent 16382A Standard Air Capacitor
	1 kHz	0.019 % of output	
	1 MHz	0.019 % of output	
	2 MHz	0.019 % of output	
	3 MHz	0.019 % of output	
	4 MHz	0.019 % of output	
	5 MHz	0.019 % of output	
	10 MHz	0.022 % of output	
13 MHz	0.024 % of output		
Capacitance – Generate <sup>1</sup> Fixed Points	100 pF		Agilent 16383A Standard Air Capacitor
	1 kHz	0.017 % of output	
	1 MHz	0.017 % of output	
	2 MHz	0.018 % of output	
	3 MHz	0.018 % of output	
	4 MHz	0.02 % of output	
	5 MHz	0.024 % of output	
	10 MHz	0.053 % of output	
13 MHz	0.083 % of output		
Capacitance – Generate <sup>1</sup> Fixed Points	1 000 pF		Agilent 16384A Standard Air Capacitor
	1 kHz	0.018 % of output	
	1 MHz	0.019 % of output	
	2 MHz	0.024 % of output	
	3 MHz	0.035 % of output	
	4 MHz	0.051 % of output	
	5 MHz	0.069 % of output	
	10 MHz	0.22 % of output	
13 MHz	0.32 % of output		
Inductance – Generate <sup>1</sup> Fixed Point	500 μH		GR 1482-D Standard Inductor
	1 kHz	0.35 μH	
	20 mH		GR 1482-J Standard Inductor
	1 kHz	0.008 mH	
200 mH		GR 1482-M Standard Inductor	
100 Hz	0.11 mH		
2 H		GR 1482-Q Standard Inductor	
100 Hz	0.8 mH		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Generate <sup>1</sup>	Up to 22 mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	228 $\mu\text{V/V} + 3.9 \mu\text{V}$ 88 $\mu\text{V/V} + 3.9 \mu\text{V}$ 76 $\mu\text{V/V} + 3.9 \mu\text{V}$ 190 $\mu\text{V/V} + 3.9 \mu\text{V}$ 457 $\mu\text{V/V} + 4.6 \mu\text{V}$ 989 $\mu\text{V/V} + 9.2 \mu\text{V}$ 1.3 mV/V + 19 $\mu\text{V}$ 2.6 mV/V + 19 $\mu\text{V}$	Fluke 5720A Multifunction Calibrator
AC Voltage – Generate <sup>1</sup>	(22 to 220) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz (0.22 to 2.2) V (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz (2.2 to 22) V (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz (22 to 220) V (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz	228 $\mu\text{V/V} + 11.4 \mu\text{V}$ 88 $\mu\text{V/V} + 6.1 \mu\text{V}$ 76 $\mu\text{V/V} + 6.1 \mu\text{V}$ 190 $\mu\text{V/V} + 6.1 \mu\text{V}$ 457 $\mu\text{V/V} + 15.2 \mu\text{V}$ 837 $\mu\text{V/V} + 19 \mu\text{V}$ 1.3 mV/V + 23 $\mu\text{V}$ 2.5 mV/V + 46 $\mu\text{V}$ 228 $\mu\text{V/V} + 38 \mu\text{V}$ 84 $\mu\text{V/V} + 15 \mu\text{V}$ 40 $\mu\text{V/V} + 8 \mu\text{V}$ 68 $\mu\text{V/V} + 9 \mu\text{V}$ 99 $\mu\text{V/V} + 30 \mu\text{V}$ 380 $\mu\text{V/V} + 76 \mu\text{V}$ 913 $\mu\text{V/V} + 190 \mu\text{V}$ 1.5 mV/V + 304 $\mu\text{V}$ 228 $\mu\text{V/V} + 380 \mu\text{V}$ 84 $\mu\text{V/V} + 152 \mu\text{V}$ 37 $\mu\text{V/V} + 54 \mu\text{V}$ 61 $\mu\text{V/V} + 91 \mu\text{V}$ 76 $\mu\text{V/V} + 190 \mu\text{V}$ 228 $\mu\text{V/V} + 609 \mu\text{V}$ 913 $\mu\text{V/V} + 1.9 \text{ mV}$ 1.4 mV/V + 3 mV 228 $\mu\text{V/V} + 3.8 \text{ mV}$ 84 $\mu\text{V/V} + 1.5 \text{ mV}$ 49 $\mu\text{V/V} + 0.6 \text{ mV}$ 76 $\mu\text{V/V} + 0.9 \text{ mV}$ 137 $\mu\text{V/V} + 2.3 \text{ mV}$	Fluke 5720A Multifunction Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Generate <sup>1</sup>	(220 to 750) V 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (750 to 1 000) V 40 Hz to 1 kHz (1 to 20) kHz (20 to 30) kHz	68 $\mu$ V/V + 3 mV 126 $\mu$ V/V + 5 mV 457 $\mu$ V/V + 8 mV 1.8 mV/V + 34 mV 68 $\mu$ V/V + 3 mV 126 $\mu$ V/V + 5 mV 457 $\mu$ V/V + 8 mV	Fluke 5720A Multifunction Calibrator /5725A Amplifier
AC Voltage – Generate <sup>1</sup> Wideband Absolute	(0.3 to 1.1) mV (10 to 30) Hz 30 Hz to 500 kHz (0.5 to 1.2) MHz (1.2 to 2) MHz (2 to 12) MHz (12 to 20) MHz (20 to 30) MHz (1.1 to 3.3) mV (10 to 30) Hz 30 Hz to 500 kHz (0.5 to 1.2) MHz (1.2 to 2) MHz (2 to 12) MHz (12 to 20) MHz (20 to 30) MHz (3.3 to 11) mV (10 to 30) Hz 30 Hz to 500 kHz (0.5 to 1.2) MHz (1.2 to 2) MHz (2 to 12) MHz (12 to 20) MHz (20 to 30) MHz (11 to 33) mV (10 to 30) Hz 30 Hz to 500 kHz (0.5 to 1.2) MHz (1.2 to 2) MHz (2 to 12) MHz (12 to 20) MHz (20 to 30) MHz	0.65 % of output + 1.5 $\mu$ V 0.61 % of output + 1.5 $\mu$ V 0.63 % of output + 3.8 $\mu$ V 0.63 % of output + 3.8 $\mu$ V 0.68 % of output + 3.8 $\mu$ V 0.76 % of output + 3.8 $\mu$ V 1.3 % of output + 12.9 $\mu$ V 0.58 % of output + 2.3 $\mu$ V 0.53 % of output + 2.3 $\mu$ V 0.54 % of output + 4.6 $\mu$ V 0.54 % of output + 4.6 $\mu$ V 0.58 % of output + 4.6 $\mu$ V 0.65 % of output + 4.6 $\mu$ V 1.3 % of output + 4.6 $\mu$ V 0.58 % of output + 6.1 $\mu$ V 0.53 % of output + 6.1 $\mu$ V 0.54 % of output + 8.4 $\mu$ V 0.54 % of output + 8.4 $\mu$ V 0.55 % of output + 8.4 $\mu$ V 0.61 % of output + 8.4 $\mu$ V 0.93 % of output + 8.4 $\mu$ V 0.52 % of output + 12 $\mu$ V 0.46 % of output + 12 $\mu$ V 0.47 % of output + 14 $\mu$ V 0.47 % of output + 14 $\mu$ V 0.49 % of output + 14 $\mu$ V 0.55 % of output + 14 $\mu$ V 0.89 % of output + 14 $\mu$ V	Fluke 5720A Multifunction Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Generate <sup>1</sup> Wideband Absolute	(33 to 110) mV		Fluke 5720A Multifunction Calibrator
	(10 to 30) Hz	0.52 % of output + 30 μV	
	30 Hz to 500 kHz	0.46 % of output + 30 μV	
	(0.5 to 1.2) MHz	0.47 % of output + 33 μV	
	(1.2 to 2) MHz	0.47 % of output + 33 μV	
	(2 to 12) MHz	0.49 % of output + 33 μV	
	(12 to 20) MHz	0.55 % of output + 33 μV	
	(20 to 30) MHz	0.89 % of output + 33 μV	
	(110 to 330) mV		
	(10 to 30) Hz	0.45 % of output + 0.1 mV	
	30 Hz to 500 kHz	0.38 % of output + 0.1 mV	
	(0.5 to 1.2) MHz	0.4 % of output + 0.1 mV	
	(1.2 to 2) MHz	0.4 % of output + 0.1 mV	
	(2 to 12) MHz	0.42 % of output + 0.1 mV	
	(12 to 20) MHz	0.49 % of output + 0.1 mV	
	(20 to 30) MHz	0.85 % of output + 0.1 mV	
	0.33 to 1.1 V		
	(10 to 30) Hz	0.45 % of output + 0.3 mV	
	30 Hz to 500 kHz	0.38 % of output + 0.3 mV	
	(0.5 to 1.2) MHz	0.4 % of output + 0.3 mV	
(1.2 to 2) MHz	0.4 % of output + 0.3 mV		
(2 to 12) MHz	0.42 % of output + 0.3 mV		
(12 to 20) MHz	0.49 % of output + 0.3 mV		
(20 to 30) MHz	0.85 % of output + 0.3 mV		
(1.1 to 3.5) V			
(10 to 30) Hz	0.39 % of output + 0.4 mV		
30 Hz to 500 kHz	0.3 % of output + 0.4 mV		
(0.5 to 1.2) MHz	0.32 % of output + 0.4 mV		
(1.2 to 2) MHz	0.32 % of output + 0.4 mV		
(2 to 12) MHz	0.35 % of output + 0.4 mV		
(12 to 20) MHz	0.44 % of output + 0.4 mV		
(20 to 30) MHz	0.82 % of output + 0.4 mV		
AC Voltage – Measure <sup>1</sup>	(1 to 10) mV		Agilent 3458A Multimeter
	(1 to 40) Hz	0.03 % of reading + 3 μV	
	40 Hz to 1 kHz	0.02 % of reading + 1.1 μV	
	(1 to 20) kHz	0.03 % of reading + 1.1 μV	
	(20 to 50) kHz	0.1 % of reading + 1.1 μV	
	(50 to 100) kHz	0.5 % of reading + 1.1 μV	
	100 kHz to 1 MHz	1.2 % of reading + 5 μV	
	(1 to 4) MHz	7 % of reading + 7 μV	
(4 to 8) MHz	20 % of reading + 8 μV		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure <sup>1</sup>	10 mV to 100 mV		Agilent 3458A Multimeter
	(1 to 40) Hz	0.007 % of reading + 4 μV	
	40 Hz to 1 kHz	0.007 % of reading + 2 μV	
	(1 to 20) kHz	0.014 % of reading + 2 μV	
	(20 to 50) kHz	0.03 % of reading + 2 μV	
	(50 to 100) kHz	0.08 % of reading + 2 μV	
	(100 to 300) kHz	0.3 % of reading + 10 μV	
	300 kHz to 1 MHz	1 % of reading + 10 μV	
	(1 to 2) MHz	1.5 % of reading + 70 μV	
	(2 to 4) MHz	4 % of reading + 70 μV	
	(4 to 8) MHz	4 % of reading + 80 μV	
	(8 to 10) MHz	15 % of reading + 100 μV	
	(0.1 to 1) V		
	(1 to 40) Hz	0.007 % of reading + 40 μV	
	40 Hz to 1 kHz	0.07 % of reading + 20 μV	
	(1 to 20) kHz	0.014 % of reading + 20 μV	
	(20 to 50) kHz	0.03 % of reading + 20 μV	
	(50 to 100) kHz	0.08 % of reading + 20 μV	
	(100 to 300) kHz	0.3 % of reading + 100 μV	
	300 kHz to 1 MHz	1 % of reading + 100 μV	
	(1 to 2) MHz	1.5 % of reading + 0.7 mV	
	(2 to 4) MHz	4 % of reading + 0.7 mV	
	(4 to 8) MHz	4 % of reading + 0.8 mV	
	(8 to 10) MHz	15 % of reading + 1.0 mV	
	(1 to 10) V		
	(1 to 40) Hz	0.007 % of reading + 0.4 mV	
	40 Hz to 1 kHz	0.007 % of reading + 0.2 mV	
(1 to 20) kHz	0.014 % of reading + 0.2 mV		
(20 to 50) kHz	0.03 % of reading + 0.2 mV		
(50 to 100) kHz	0.08 % of reading + 0.2 mV		
(100 to 300) kHz	0.3 % of reading + 1.0 mV		
300 kHz to 1 MHz	1 % of reading + 1 mV		
(1 to 2) MHz	1.5 % of reading + 7 mV		
(2 to 4) MHz	4 % of reading + 7 mV		
(4 to 8) MHz	4 % of reading + 8 mV		
(8 to 10) MHz	15 % of reading + 10 mV		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure <sup>1</sup>	(10 to 100) V (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (100 to 700) V (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz	0.02 % of reading + 4 mV 0.02 % of reading + 2 mV 0.02 % of reading + 2 mV 0.035 % of reading + 2 mV 0.12 % of reading + 2 mV 0.4 % of reading + 10 mV 1.5 % of reading + 10 mV 0.04 % of reading + 40 mV 0.04 % of reading + 20 mV 0.06 % of reading + 20 mV 0.12 % of reading + 20 mV 0.3 % of reading + 20 mV	HP 3458A Multimeter
AC Voltage – Measure <sup>1</sup>	(1 to 10) kV 60 Hz (10 to 42) kV 60 Hz	0.5 % of reading + 0.002 kV 0.5 % of reading + 0.02 kV	Ross VD60 High Voltage Divider, HP 34401A Multimeter
AC Current – Generate <sup>1</sup>	(9 to 220) $\mu$ A (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (0.22 to 2.2) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (2.2 to 22) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	228 $\mu$ A/A + 15 nA 152 $\mu$ A/A + 10 nA 91 $\mu$ A/A + 8 nA 266 $\mu$ A/A + 12 nA 989 $\mu$ A/A + 61 nA 228 $\mu$ A/A + 39 nA 152 $\mu$ A/A + 31 nA 107 $\mu$ A/A + 31 nA 183 $\mu$ A/A + 99 nA 989 $\mu$ A/A + 609 nA 228 $\mu$ A/A + 385 nA 152 $\mu$ A/A + 310 nA 107 $\mu$ A/A + 310 nA 183 $\mu$ A/A + 536 nA 989 $\mu$ A/A + 4.6 $\mu$ A	Fluke 5720A Multifunction Calibrator

**Electrical – DC/Low Frequency**

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Generate <sup>1</sup>	(22 to 200) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	228 $\mu$ A/A + 4 $\mu$ A 152 $\mu$ A/A + 3 $\mu$ A 107 $\mu$ A/A + 2 $\mu$ A 183 $\mu$ A/A + 3 $\mu$ A 989 $\mu$ A/A + 9 $\mu$ A	Fluke 5720A Multifunction Calibrator
	(0.2 to 2.2) A 20 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	243 $\mu$ A/A + 31 $\mu$ A 380 $\mu$ A/A + 76 $\mu$ A 6.1 mA/A + 152 $\mu$ A	
AC Current – Generate <sup>1</sup>	(2.2 to 11) A 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	350 $\mu$ A/A + 141 $\mu$ A 723 $\mu$ A/A + 295 $\mu$ A 2.7 mA/A + 573 $\mu$ A	Fluke 5720A Multifunction Calibrator /5725A Amplifier
AC Current – Generate <sup>1</sup>	(11 to 20.5) A (45 to 100) Hz 100 Hz to 1 kHz (1 to 5) kHz	0.09 % of output + 5 mA 0.11 % of output + 5 mA 2.28 % of output + 5 mA	Fluke 5720A Multifunction Calibrator /5725A Amplifier
AC Current Clamps – Toroidal-Wound	(16.5 to 1 025) A (45 to 65) Hz (16.5 to 150) A (65 to 440) Hz	0.23 % of output + 0.103 A 0.64 % of output + 0.085 A	Fluke 5522A Multiproduct Calibrator, 5500A/Coil x50
AC Current Clamps – Other	(16.5 to 1 025) A (45 to 65) Hz (16.5 to 150) A (65 to 440) Hz	0.44 % of output + 0.53 A 0.79 % of output + 0.69 A	Fluke 5522A Multiproduct Calibrator, 5500A/Coil x50
AC Current – Measure <sup>1</sup>	(5 to 100) $\mu$ A (10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz (0.1 to 1) mA (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz (1 to 10) mA (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz	0.4 % of reading + 0.03 $\mu$ A 0.15 % of reading + 0.03 $\mu$ A 0.06 % of reading + 0.03 $\mu$ A 0.4 % of reading + 0.2 $\mu$ A 0.15 % of reading + 0.2 $\mu$ A 0.06 % of reading + 0.2 $\mu$ A 0.03 % of reading + 0.2 $\mu$ A 0.4 % of reading + 2 $\mu$ A 0.15 % of reading + 2 $\mu$ A 0.06 % of reading + 2 $\mu$ A 0.03 % of reading + 2 $\mu$ A	Agilent 3458A Multimeter

**Electrical – DC/Low Frequency**

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Measure <sup>1</sup>	(10 to 100) mA (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz	0.4 % of reading + 20 μA 0.15 % of reading + 20 μA 0.06 % of reading + 20 μA 0.03 % of reading + 20 μA	Agilent 3458A Multimeter
	(0.1 to 1) A (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz	0.4 % of reading + 0.2 mA 0.15 % of reading + 0.2 mA 0.06 % of reading + 0.2 mA 0.03 % of reading + 0.2 mA	
AC Current – Measure <sup>1</sup>	(1 to 3) A (3 to 5) Hz (5 to 10) Hz 10 Hz to 5 kHz	0.72 % of reading + 1.2 mA 0.23 % of reading + 1.2 mA 0.1 % of reading + 1.2 mA	Agilent 34401A Multimeter
AC Current – Measure <sup>1</sup>	(1 to 20) A (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz	368 μA/A 368 μA/A 302 μA/A 424 μA/A	Agilent 3458A Multimeter / Fluke Y5020 Current Shunt
AC Current – Measure <sup>1</sup>	(3 to 30) A 40 Hz to 1 kHz (1 to 5) kHz	0.3 % of reading + 0.07 A 5 % of reading + 0.14 A	Agilent 3458A Multimeter Keysight 34330A Current Shunt
AC Current – Measure <sup>1</sup>	(10 to 100) A (50 to 1000) Hz (100 to 600) A (50 to 1000) Hz	2 % of reading + 0.02 A 2 % of reading + 0.2 A	Fluke 80i-600 Clamp Meter / Agilent 3458A Multimeter
AC Current – Measure <sup>1</sup>	(100 to 2 000) A 55 to 65 Hz (2 000 to 6 000) A 55 to 65 Hz	1 % of reading + 0.5A 1 % of reading + 3.9 A	PEM CWT600B Rogowski Coil / HP 34401A Multimeter
Low Frequency Power – Generate <sup>1</sup>	Up to 20 kW (45 to 65) Hz, 1 PF	0.44 % of reading	Fluke 5522A Multiproduct Calibrator
	Up to 20 kW DC	0.13 % of reading	

Electrical – DC/Low Frequency

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Calibration of Thermocouple Indicators <sup>1</sup>	Type B		Fluke 5522A Multiproduct Calibrator
	(600 to 800) °C	0.34 °C	
	(800 to 1550) °C	0.26 °C	
	(800 to 1550) °C	0.23 °C	
	(1550 to 1820) °C	0.25 °C	
	Type C		
	(0 to 150) °C	0.23 °C	
	(150 to 650) °C	0.2 °C	
	(650 to 1 000) °C	0.24 °C	
	(1 000 to 1 800) °C	0.39 °C	
	(1 800 to 2 316) °C	0.64 °C	
	Type E		
	(-250 to -100) °C	0.38 °C	
	(-100 to -25) °C	0.13 °C	
	(-25 to 350) °C	0.11 °C	
	(350 to 650) °C	0.13 °C	
	(650 to 1 000) °C	0.16 °C	
	Type J		
	(-210 to -100) °C	0.21 °C	
	(-100 to -30) °C	0.13 °C	
	(-30 to 150) °C	0.11 °C	
	(150 to 760) °C	0.13 °C	
	(760 to 1 200) °C	0.18 °C	
	Type K		
	(-200 to -100) °C	0.25 °C	
	(-100 to -25) °C	0.14 °C	
	(-25 to 120) °C	0.13 °C	
	(120 to 1 000) °C	0.2 °C	
(1 000 to 1 372) °C	0.31 °C		
Type L			
(-200 to -100) °C	0.28 °C		
(-100 to 800) °C	0.2 °C		
(800 to 900) °C	0.13 °C		
Type N			
(-200 to -100) °C	0.31 °C		
(-100 to -25) °C	0.17 °C		
(-25 to 120) °C	0.15 °C		
(120 to 410) °C	0.14 °C		
(410 to 1 300) °C	0.21 °C		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Calibration of Thermocouple Indicators <sup>1</sup>	Type R (0 to 250) °C	0.43 °C	Fluke 5522A Multiproduct Calibrator
	(250 to 1 000) °C	0.27 °C	
	(1 000 to 1 400) °C	0.25 °C	
	(1 400 to 1 767) °C	0.31 °C	
	Type S (0 to 250) °C	0.36 °C	
	(250 to 1 000) °C	0.28 °C	
	(1 000 to 1 400) °C	0.28 °C	
	(1 400 to 1 767) °C	0.35 °C	
	Type T (-250 to -150) °C	0.48 °C	
	(-150 to 0) °C	0.18 °C	
	(0 to 120) °C	0.13 °C	
	(120 to 400) °C	0.11 °C	
	Type U (-200 to 0) °C	0.43 °C	
	(0 to 600) °C	0.21 °C	
Electrical Calibration of RTD Indicators <sup>1</sup>	Pt 385, 100 Ω (-200 to 0) °C	0.058 °C	Fluke 5522A Multiproduct Calibrator
	(0 to 100) °C	0.081 °C	
	(100 to 300) °C	0.1 °C	
	(300 to 400) °C	0.12 °C	
	(400 to 630) °C	0.14 °C	
	(630 to 800) °C	0.27 °C	
	Pt 3926, 100 Ω (-200 to 0) °C	0.07 °C	
	(0 to 100) °C	0.081 °C	
	(100 to 300) °C	0.1 °C	
	(300 to 400) °C	0.12 °C	
	(400 to 630) °C	0.14 °C	
	Pt 3916, 100 Ω (-200 to -190) °C	0.29 °C	
	(-190 to -80) °C	0.047 °C	
	(-80 to 0) °C	0.058 °C	
	(0 to 100) °C	0.07 °C	
	(100 to 260) °C	0.081 °C	
	(260 to 300) °C	0.093 °C	
	(300 to 400) °C	0.1 °C	
	(400 to 600) °C	0.12 °C	
(600 to 630) °C	0.27 °C		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Calibration of RTD Indicators <sup>1</sup>	Pt 385, 200 Ω		Fluke 5522A Multiproduct Calibrator
	(-200 to 100) °C	0.047 °C	
	(100 to 260) °C	0.058 °C	
	(260 to 300) °C	0.14 °C	
	(300 to 400) °C	0.15 °C	
	(400 to 600) °C	0.16 °C	
	(600 to 630) °C	0.19 °C	
	Pt 385, 500 Ω		
	(-200 to -80) °C	0.048 °C	
	(-80 to 100) °C	0.059 °C	
	(100 to 260) °C	0.07 °C	
	(260 to 400) °C	0.093 °C	
	(400 to 600) °C	0.11 °C	
	(600 to 630) °C	0.13 °C	
	Pt 385, 1000 Ω		
	(-200 to 0) °C	0.037 °C	
	(0 to 100) °C	0.048 °C	
	(100 to 260) °C	0.059 °C	
(260 to 300) °C	0.07 °C		
(300 to 600) °C	0.082 °C		
(600 to 630) °C	0.27 °C		
PtNi 385, 120 Ω (Ni120)			
(-80 to 100) °C	0.093 °C		
(100 to 260) °C	0.16 °C		
Cu 427, 10 Ω			
(100 to 260) °C	0.35 °C		
Oscilloscopes Calibration – Generate <sup>1</sup> Voltage DC - 50Ω	(1 to 24.999) mV	0.19 % of Output + 31 μV	Fluke 5820A Oscilloscope Calibrator w/ GHz Option
	(25 to 109.99) mV	0.19 % of Output + 36 μV	
	(110mV to 2.199 9) V	0.19 % of Output + 87 μV	
	(2.2 to 6.6) V	0.19 % of Output + 0.6 mV	
	(1 to 24.999) mV	0.019 % of reading + 20 μV	
	(25 to 109.99) mV	0.019 % of reading + 25 μV	
	(110mV to 2.199 9) V	0.019 % of reading + 76 μV	
	(2.2 to 10.999) V	0.019 % of reading + 0.6 mV	
DC - 1MΩ	(11 to 130) V	0.019 % of reading + 6 mV	



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Oscilloscopes Calibration – Generate <sup>1</sup> Square Wave 10 Hz to 10 kHz - 50Ω  Square Wave 10 Hz to 1 kHz - 1MΩ  Square Wave (1 to 10) kHz - 1MΩ	(1 to 24.999) mVpp (25 to 109.99) mVpp (110mV to 2.199 9) Vpp (2.2 to 6.6) Vpp  (1 to 24.999) mV (25 to 109.99) mV (110mV to 2.199 9) V (2.2 to 10.999) V (11 to 130) V  (1 to 24.999) mV (25 to 109.99) mV (110mV to 2.1999) V (2.2 to 10.999) V (11 to 130) V	0.19 % of Output + 31 μV 0.19 % of Output + 36 μV 0.19 % of Output + 87 μV 0.19 % of Output + 0.6 mV  0.038 % of reading + 4 μV 0.038 % of reading + 9 μV 0.038 % of reading + 60 μV 0.038 % of reading + 0.6 mV 0.038 % of reading + 6 mV  0.19 % of reading + 31 μV 0.19 % of reading + 36 μV 0.19 % of reading + 87 μV 0.19 % of reading + 0.6 mV 0.19 % of reading + 6 mV	Fluke 5820A Oscilloscope Calibrator w/ GHz Option
Oscilloscopes Calibration – Generate <sup>1</sup> Leveled Sine Flatness 50 kHz to 10 MHz Reference	3 dB Bandwidth (5 to 50) mVpp 50 kHz to 100) MHz (100 to 300) MHz (300 to 500) MHz (500 to 600) MHz (600 to 16 00) MHz (1 600 to 2 100) MHz 50 mV to 3.5 Vpp 50 kHz to 100 MHz (100 to 300) MHz (300 to 500) MHz (500 to 600) MHz (600 to 1 600) MHz (1 600 to 2 100) MHz	0.34 dB 0.36 dB 0.42 dB 0.46 dB 0.5 dB 0.56 dB  0.24 dB 0.24 dB 0.32 dB 0.34 dB 0.4 dB 0.44 dB	Fluke 5820A Oscilloscope Calibrator w/ GHz Option
Oscilloscopes Calibration – Generate <sup>1</sup> Leveled Sine Flatness	3 dB Bandwidth 50 mV to 3.5Vpp (2 100 to 4 000) MHz (4 000 to 8 000) MHz (8 000 to 18 000) MHz	0.25 dB 0.35 dB 0.46 dB	EPM Power Meter w/ E Series Power Sensors
Oscilloscopes Calibration – Generate <sup>1</sup> Time Marker	500 ps to 20 ms 50 ms to 5 s	0.25 μs/s 1.9 μs/s + 3.8 μHz	Fluke 5820A Oscilloscope Calibrator w/ GHz Option



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Oscilloscopes Calibration Measure <sup>1</sup> Input Impedance Resistance Leakage	(40 to 60) Ω 500 kΩ to 1.5MΩ (0 to 5.99) V	0.08 % of reading 0.08 % of reading 0.038 % of reading + 0.8 mV	Fluke 5820A Oscilloscope Calibrator w/ GHz Option
Oscilloscopes Calibration – Generate <sup>1</sup>  Voltage DC – 50 Ω DC – 1 MΩ  Square Wave 10 Hz to 10 kHz – 1 MΩ	1 mV to 5 V 1 mV to 200 V  1 mV to 200 Vpp	0.025 % of output + 25 μV 0.025 % of output + 25 μV  0.1 % of output + 10 μV	Fluke 9500B with 9510 Active Head
Oscilloscopes Calibration – Generate <sup>1</sup>  Leveled Sine Flatness 50 kHz to 10 MHz Reference	5 mV to 5 Vpp (0.1 Hz to 300) MHz (300 to 550) MHz  5 mV to 3 Vpp (550 to 1 100) MHz	0.18 dB 0.22 dB  0.3 dB	Fluke 9500B with 9510 Active Head
Oscilloscopes Calibration – Generate <sup>1</sup> Time Marker	9 ns to 55 s	0.25 μs/s	Fluke 9500B with 9510 Active Head
Oscilloscopes Calibration – Measure <sup>1</sup>  Input Impedance Resistance	(10 to 40) Ω (40 to 90) Ω (90 to 150) Ω (50 to 800) KΩ (0.8 to 1.2) MΩ (1.2 to 12) MΩ	0.5 % of reading 0.1 % of reading 0.5 % of reading 0.5 % of reading 0.1 % of reading 0.5 % of reading	Fluke 9500B with 9510 Active Head
Power Meters <sup>1</sup>	3 μW to 100 mW	0.3 % of reading	HP 11683A Range Calibrator

**Electrical – RF/Microwave**

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Attenuation – Measure <sup>1</sup>	(30 to 3050) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.12 dB	
	(90 to 100) dB	0.13 dB	
	(100 to 110) dB	0.13 dB	
	(3 050 to 6 600) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.12 dB	
	(90 to 100) dB	0.13 dB	
	(100 to 110) dB	0.19 dB	
	(6 600 to 13 200) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
(50 to 60) dB	0.076 dB		
(60 to 70) dB	0.081 dB		
(70 to 80) dB	0.12 dB		
(80 to 90) dB	0.12 dB		
(90 to 100) dB	0.13 dB		
(100 to 110) dB	0.25 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Attenuation – Measure <sup>1</sup>	(13 200 to 19 200) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.12 dB	
	(90 to 100) dB	0.13 dB	
	(100 to 110) dB	0.31 dB	
	(19 200 to 26 500) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.14 dB	
	(90 to 100) dB	0.36 dB	
	(100 to 110) dB	0.82 dB	
	(26 500 to 31 150) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
(50 to 60) dB	0.076 dB		
(60 to 70) dB	0.081 dB		
(70 to 80) dB	0.12 dB		
(80 to 90) dB	0.13 dB		
(90 to 100) dB	0.33 dB		
(100 to 110) dB	0.77 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Attenuation – Measure <sup>1</sup>	(31 150 to 41 000) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.14 dB	
	(80 to 90) dB	0.36 dB	
	(41 000 to 45 000) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.11 dB	
	(70 to 80) dB	0.24 dB	
	(80 to 90) dB	0.6 dB	
	(45 000 to 50 000) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
(30 to 40) dB	0.035 dB		
(40 to 50) dB	0.04 dB		
(50 to 60) dB	0.11 dB		
(60 to 70) dB	0.29 dB		
(70 to 80) dB	0.7 dB		
(80 to 90) dB	1.4 dB		
RF Flatness – Measure <sup>1</sup>	9 kHz to 2 000 MHz		Agilent EPM Series Power Meter w/E9304A Power Sensor
	(20 to -10) dBm	0.1 dB	
	(-10 to -30) dBm	0.1 dB	
	(-30 to -40) dBm	0.11 dB	
	(-40 to -42) dBm	0.12 dB	
	(2 000 to 6 000) MHz		
	(20 to -10) dBm	0.1 dB	
	(-10 to -30) dBm	0.09 dB	
	(-30 to -40) dBm	0.1 dB	
	(-40 to -42) dBm	0.11 dB	

**Electrical – RF/Microwave**

Peachtree Corners, GA

<b>Parameter/Equipment</b>	<b>Range</b>	<b>Expanded Uncertainty of Measurement (+/-)</b>	<b>Reference Standard, Method, and/or Equipment</b>
RF Power – Measure <sup>1</sup>	9 kHz to 14 000 MHz		Agilent EPM Series Power Meter w/E9304A H18 Power Sensor
	(20 to 0) dB	0.13 dB	
	(0 to -40) dB	0.15 dB	
	(-40 to -50) dB	0.35 dB	
	(-50 to -55) dB	0.93 dB	
	(14 000 to 18 000) MHz		
RF Power – Measure <sup>1</sup>	(20 to 0) dB	0.12 dB	Agilent EPM Series Power Meter 8487A Power Sensor
	(0 to -40) dB	0.16 dB	
	(-40 to -50) dB	0.35 dB	
	(-50 to -55) dB	0.93 dB	
	(100 to 2 000) MHz		
	(20 to 10) dB	0.14 dB	
RF Power – Measure <sup>1</sup>	(10 to 0) dB	0.06 dB	Agilent EPM Series Power Meter 8487A Power Sensor
	(0 to -10) dB	0.07 dB	
	(-10 to -20) dB	0.08 dB	
	(-20 to -25) dB	0.15 dB	
	(2 000 to 12 400) MHz		
	(20 to 10) dB	0.15 dB	
(10 to 0) dB	0.07 dB		
(0 to -10) dB	0.07 dB		
(-10 to -20) dB	0.08 dB		
(-20 to -25) dB	0.15 dB		
(12 400 to 18 000) MHz			
(20 to 10) dB	0.15 dB		
(10 to 0) dB	0.08 dB		
(0 to -10) dB	0.08 dB		
(-10 to -20) dB	0.09 dB		
(-20 to -25) dB	0.15 dB		
(18 000 to 26 500) MHz			
(20 to 10) dB	0.16 dB		
(10 to 0) dB	0.1 dB		
(0 to -10) dB	0.1 dB		
(-10 to -20) dB	0.11 dB		
(-20 to -25) dB	0.16 dB		
(26 500 to 40 000) MHz			
(20 to 10) dB	0.17 dB		
(10 to 0) dB	0.12 dB		
(0 to -10) dB	0.12 dB		
(-10 to -20) dB	0.13 dB		
(-20 to -25) dB	0.18 dB		

**Electrical – RF/Microwave**

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Measure <sup>1</sup>	(40 000 to 50 000) MHz (20 to 10) dB (10 to 0) dB (0 to -10) dB (-10 to -20) dB (-20 to -25) dB	0.22 dB 0.19 dB 0.19 dB 0.19 dB 0.23 dB	Agilent EPM Series Power Meter 8487A Power Sensor
RF Power – Measure <sup>1</sup>	(30 to 2 000) MHz (30 to 20) dB (20 to 0) dB (0 to -58) dB (-58 to -78) dB (-78 to -110) dB (-110 to -115) dB (-115 to -120) dB (-120 to -125) dB (2 000 to 3 050) MHz (30 to 20) dB (20 to 0) dB (0 to -58) dB (-58 to -78) dB (-78 to -110) dB (-110 to -115) dB (-115 to -120) dB (-120 to -125) dB (3 050 to 6 600) MHz (30 to 20) dB (20 to 0) dB (0 to -58) dB (-58 to -78) dB (-78 to -110) dB (-110 to -115) dB (-115 to -120) dB (6 600 to 13 200) MHz (30 to 20) dB (20 to 0) dB (0 to -58) dB (-58 to -78) dB (-78 to -110) dB (-110 to -115) dB (-115 to -120) dB	0.36 dB 0.2 dB 0.22 dB 0.23 dB 0.25 dB 0.27 dB 0.33 dB 0.46 dB 0.37 dB 0.21 dB 0.23 dB 0.24 dB 0.26 dB 0.28 dB 0.34 dB 0.7 dB 0.37 dB 0.21 dB 0.23 dB 0.24 dB 0.29 dB 0.38 dB 0.53 dB 0.37 dB 0.21 dB 0.23 dB 0.24 dB 0.34 dB 0.46 dB 0.65 dB	Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor



ANSI National Accreditation Board

Electrical – RF/Microwave

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Measure <sup>1</sup>	(13 200 to 18 000) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor
	(30 to 20) dB	0.37 dB	
	(20 to 0) dB	0.21 dB	
	(0 to -58) dB	0.23 dB	
	(-58 to -78) dB	0.24 dB	
	(-78 to -90) dB	0.26 dB	
	(-90 to -95) dB	0.26 dB	
	(-95 to -100) dB	0.26 dB	
	(-100 to -105) dB	0.29 dB	
	(-105 to -110) dB	0.38 dB	
	(-110 to -115) dB	0.53 dB	
	(-115 to -120) dB	0.75 dB	
	(18 000 to 19 200) MHz		
	(30 to 20) dB	0.39 dB	
	(20 to 0) dB	0.25 dB	
	(0 to -58) dB	0.24 dB	
	(-58 to -78) dB	0.25 dB	
	(-78 to -90) dB	0.27 dB	
	(-90 to -95) dB	0.27 dB	
	(-95 to -100) dB	0.27 dB	
	(-100 to -105) dB	0.3 dB	
	(-105 to -110) dB	0.38 dB	
	(-110 to -115) dB	0.53 dB	
	(-115 to -120) dB	0.75 dB	
	(19 200 to 26 500) MHz		
	(30 to 20) dB	0.39 dB	
	(20 to 0) dB	0.25 dB	
	(0 to -58) dB	0.24 dB	
	(-58 to -78) dB	0.25 dB	
	(-78 to -90) dB	0.28 dB	
(-90 to -95) dB	0.33 dB		
(-95 to -100) dB	0.43 dB		
(-100 to -105) dB	0.61 dB		
(-105 to -110) dB	0.85 dB		
(-110 to -115) dB	1.2 dB		
(-115 to -120) dB	1.5 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Measure <sup>1</sup>	(26 500 to 31 150) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor
	(30 to 20) dB	0.42 dB	
	(20 to 0) dB	0.3 dB	
	(0 to -58) dB	0.34 dB	
	(-58 to -78) dB	0.34 dB	
	(-78 to -90) dB	0.36 dB	
	(-90 to -95) dB	0.39 dB	
	(-95 to -100) dB	0.46 dB	
	(-100 to -105) dB	0.61 dB	
	(-105 to -110) dB	0.82 dB	
	(-110 to -115) dB	1.1 dB	
	(31 150 to 41 000) MHz		
	(30 to 20) dB	0.42 dB	
	(20 to 0) dB	0.3 dB	
	(0 to -58) dB	0.34 dB	
	(-58 to -78) dB	0.35 dB	
	(-78 to -90) dB	0.48 dB	
	(-90 to -95) dB	0.64 dB	
	(-95 to -100) dB	0.87 dB	
	(-100 to -105) dB	1.2 dB	
	(-105 to -110) dB	1.5 dB	
	(41 000 to 45 000) MHz		
	(30 to 20) dB	0.42 dB	
	(20 to 0) dB	0.3 dB	
	(0 to -58) dB	0.34 dB	
	(-58 to -78) dB	0.38 dB	
	(-78 to -90) dB	0.68 dB	
	(-90 to -95) dB	0.93 dB	
	(-95 to -100) dB	1.2 dB	
	(-100 to -105) dB	1.6 dB	
(45 000 to 50 000) MHz			
(30 to 20) dB	0.42 dB		
(20 to 0) dB	0.3 dB		
(0 to -58) dB	0.34 dB		
(-58 to -78) dB	0.67 dB		
(-78 to -90) dB	1.4 dB		

Electrical – RF/Microwave

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Generate <sup>1</sup>	(30 to 2 000) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor, E8257D Signal Generator
	(20 to 0) dB	0.29 dB	
	(0 to -58) dB	0.3 dB	
	(-58 to -78) dB	0.31 dB	
	(-78 to -110) dB	0.32 dB	
	(2 000 to 3 050) MHz		
	(20 to 0) dB	0.34 dB	
	(0 to -58) dB	0.35 dB	
	(-58 to -78) dB	0.36 dB	
	(-78 to -110) dB	0.39 dB	
	(3 050 to 6 600) MHz		
	(20 to 0) dB	0.34 dB	
	(0 to -58) dB	0.35 dB	
	(-58 to -78) dB	0.36 dB	
	(-78 to -110) dB	0.39 dB	
	(6 600 to 13 200) MHz		
	(20 to 0) dB	0.34 dB	
	(0 to -58) dB	0.35 dB	
	(-58 to -78) dB	0.36 dB	
	(-78 to -110) dB	0.42 dB	
	(13 200 to 18 000) MHz		
	(15 to 0) dB	0.34 dB	
	(0 to -58) dB	0.35 dB	
	(-58 to -78) dB	0.36 dB	
	(-78 to -110) dB	0.46 dB	
	(18 000 to 19 200) MHz		
	(15 to 0) dB	0.41 dB	
	(0 to -58) dB	0.41 dB	
(-58 to -78) dB	0.42 dB		
(-78 to -110) dB	0.5 dB		
(19 200 to 26 500) MHz			
(15 to 0) dB	0.41 dB		
(0 to -58) dB	0.41 dB		
(-58 to -78) dB	0.42 dB		
(-78 to -110) dB	0.9 dB		
(26 500 to 31 150) MHz			
(15 to 0) dB	0.62 dB		
(0 to -58) dB	0.63 dB		
(-58 to -78) dB	0.64 dB		
(-78 to -110) dB	0.96 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Generate <sup>1</sup>	(31 150 to 41 000) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor, E8257D Signal Generator
	(10 to 0) dB	0.82 dB	
	(0 to -58) dB	0.83 dB	
	(-58 to -78) dB	0.84 dB	
	(-78 to -100) dB	1.1 dB	
	(41 000 to 45 000) MHz		
	(10 to 0) dB	0.82 dB	
	(0 to -58) dB	0.83 dB	
	(-58 to -78) dB	0.85 dB	
	(-78 to -100) dB	1.4 dB	
	(45 000 to 50 000) MHz		
	(10 to 0) dB	0.82 dB	
	(0 to -58) dB	0.83 dB	
	(-58 to -78) dB	1 dB	
(-78 to -90) dB	1.5 dB		
Amplitude Modulation – Measure <sup>1</sup>	100 kHz to 10 MHz		Agilent N5531S Measuring Receiver
	Rate 50 Hz to 10 kHz		
	(5 to 99) % Depth	0.75 % of reading + 0.3 digits	
	10 MHz to 3 GHz		
	Rate 50 Hz to 100 kHz		
	(5 to 20) % Depth	2.5 % of reading + 0.4 digits	
	(20 to 99) % Depth	0.5 % of reading + 0.4 digits	
	(3 to 26.5) GHz		
	Rate 50 Hz to 100 kHz		
	(5 to 20) % Depth	4.5 % of reading + 0.4 digits	
	(20 to 99) % Depth	1.5 % of reading + 0.4 digits	
	(26.5 to 31.15) GHz		
	Rate 50 Hz to 100 kHz		
	(5 to 20) % Depth	6.8 % of reading + 0.4 digits	
(20 to 99) % Depth	1.9 % of reading + 0.4 digits		
(31.15 to 50) GHz			
Rate 50 Hz to 100 kHz			
(5 to 20) % Depth	26 % of reading + 0.4 digits		
(20 to 99) % Depth	6 % of reading + 0.4 digits		



ANSI National Accreditation Board

Electrical – RF/Microwave

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency Modulation – Measure <sup>1</sup> $\beta$ = deviation / rate	250 kHz to 10 MHz Rates 20 Hz to 10 kHz Peak Dev 200 to 40 kHz 10 MHz to 6.6 GHz Rates 50 Hz to 200 kHz Peak Dev 250 to 400 kHz (6.6 to 13.2) GHz Rates 50 Hz to 200 kHz Peak Dev 250 to 400 kHz (13.2 to 31.15) GHz Rates 50 Hz to 200 kHz Peak Dev 250 to 400 kHz (31.15 to 50.0) GHz Rates 50 Hz to 200 kHz Peak Dev 250 to 400 kHz	$\beta > 0.2 - 1.5$ % of reading + 2 Hz $\beta > 1.2 - 1$ % of reading + 2 Hz $\beta > 0.20 - 1.5$ % of reading + 2 Hz $\beta > 0.45 - 1$ % of reading + 2 Hz $\beta > 0.2 - 2.5$ % of reading + 4 Hz $\beta > 8.0 - 1$ % of reading + 4 Hz $\beta > 0.2 - 3.8$ % of reading + 9 Hz $\beta > 16 - 1$ % of reading + 9 Hz $\beta > 0.2 - 8.5$ % of reading + 17 Hz $\beta > 32 - 1$ % of reading + 17 Hz	Agilent N5531S Measuring Receiver
Phase Modulation – Measure <sup>1</sup>	100 kHz to 6.6 GHz Deviations > 0.3 rad Deviations > 0.7 rad (6.6 to 13.2) GHz Deviations > 0.6 rad Deviations > 2.0 rad (13.2 to 26.5) GHz Deviations: > 1.2 rad Deviations > 4.0 rad (26.5 to 31.15) GHz Deviations: > 1.3 rad Deviations > 4.0 rad (31.15 to 50) GHz Deviations: > 2.4 rad Deviations > 8.0 rad	3 % of reading + 0.002 rad 1 % of reading + 0.002 rad 3 % of reading + 0.005 rad 1 % of reading + 0.005 rad 3 % of reading + 0.009 rad 1 % of reading + 0.009 rad 3 % of reading + 0.009 rad 1 % of reading + 0.009 rad 3 % of reading + 0.018 rad 1 % of reading + 0.018 rad	Agilent N5531S Measuring Receiver
Audio Distortion	(-80 to 0) dB 400 Hz and 1 kHz	5.8 % of reading	HP 8902A Measuring Receiver



ANSI National Accreditation Board

Electrical – RF/Microwave

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AM Distortion Measure <sup>1</sup> Rate 20 Hz to 1 kHz	(0.1 to 10) MHz		Agilent N5531S Measuring Receiver
	AM Depth > 1%		
	(0 to -20) dB	1.2 dB	
	(-20 to -30) dB	2.2 dB	
	AM Depth > 3%		
	(0 to -20) dB	1 dB	
	(-20 to -30) dB	1.3 dB	
	(-30 to -40) dB	2.4 dB	
	10 MHz to 26.5 GHz		
	AM Depth > 1%		
	(0 to -20) dB	1.3 dB	
	(-20 to -30) dB	2.5 dB	
	AM Depth > 3%		
	(0 to -20) dB	1.1 dB	
(-20 to -30) dB	1.4 dB		
(-30 to -40) dB	3 dB		
(26.5 MHz to 50.0 GHz)			
AM Depth > 3%			
(0 to -20) dB	1.8 dB		
AM Depth > 5%			
(0 to -20) dB	1.5 dB		
(-20 to -30) dB	3.5 dB		
FM Distortion Measure <sup>1</sup> Rate 20 Hz to 1 kHz	(1 to 6 600) MHz		Agilent N5531S Measuring Receiver
	Dev 500 Hz to 2KHz		
	(0 to -20) dB	0.26 dB	
	(-20 to -30) dB	0.79 dB	
	(-30 to -40) dB	2.3 dB	
	Dev > 2 kHz		
	(0 to -20) dB	0.09 dB	
	(-20 to -30) dB	0.27 dB	
	(-30 to -40) dB	0.83 dB	
	(-40 to -50) dB	2.4 dB	



ANSI National Accreditation Board

Electrical – RF/Microwave

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
FM Distortion Measure <sup>1</sup> Rate 20 Hz to 1 kHz	(6.6 to 13.2) GHz		Agilent N5531S Measuring Receiver
	Dev > 2.3 kHz		
	(0 to -20) dB		
	(-20 to -30) dB	0.26 dB	
	(-30 to -40) dB	0.79 dB	
	Dev > 4.5 kHz	2.3 dB	
	(0 to -20) dB		
	(-20 to -30) dB	0.09 dB	
	(-30 to -40) dB	0.27 dB	
	(-40 to -50) dB	0.83 dB	
	(13.2 to 31.15) GHz	2.4 dB	
	Dev > 2.7 kHz		
	(0 to -20) dB		
	(-20 to -30) dB	0.26 dB	
	(-30 to -40) dB	0.79 dB	
	Dev > 6.0 kHz	2.3 dB	
	(0 to -20) dB		
	(-20 to -30) dB	0.09 dB	
	(-30 to -40) dB	0.27 dB	
	(-40 to -50) dB	0.83 dB	
(31.15 to 50.0) GHz	2.4 dB		
Dev > 4 kHz			
(0 to -20) dB			
(-20 to -30) dB	0.26 dB		
(-30 to -40) dB	0.79 dB		
(31.15 to 50.0) GHz	2.3 dB		
Dev > 12.0 kHz			
(0 to -20) dB	0.09 dB		
(-20 to -30) dB	0.27 dB		
(-30 to -40) dB	0.83 dB		
(-40 to -50) dB	2.3 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
PM Distortion Measure <sup>1</sup>	(1 to 6 600) MHz		
	Rate 20 to 500 Hz		
	Dev > 0.8 rad		
	(0 to -20) dB	0.26 dB	
	(-20 to -30) dB	0.79 dB	
	(-30 to -40) dB	2.3 dB	
	Dev > 2.5 rad		
	(0 to -20) dB	0.09 dB	
	(-20 to -30) dB	0.27 dB	
	(-30 to -40) dB	0.83 dB	
	(-40 to -50) dB	2.3 dB	
	Rate (500 to 1 000) Hz		
	Dev > 0.4 rad	0.26 dB	
	(0 to -20) dB	0.79 dB	
	(-20 to -30) dB	2.3 dB	
	(-30 to -40) dB		
	Dev > 1.0 rad		
	(0 to -20) dB	0.09 dB	
	(-20 to -30) dB	0.27 dB	
	(-30 to -40) dB	0.83 dB	
	(-40 to -50) dB	2.3 dB	
	(6.6 to 13.2) GHz		
	Rate (20 to 500) Hz		
	Dev > 1.8 rad		
(0 to -20) dB	0.26 dB		
(-20 to -30) dB	0.79 dB		
(-30 to -40) dB	2.3 dB		
Dev > 5.5 rad			
(0 to -20) dB	0.09 dB		
(-20 to -30) dB	0.27 dB		
(-30 to -40) dB	0.83 dB		
(-40 to -50) dB	2.3 dB		
Rate (500 to 1 000) Hz			
Dev > 0.8 rad			
(0 to -20) dB	0.26 dB		
(-20 to -30) dB	0.79 dB		
(-30 to -40) dB	2.3 dB		
Agilent N5531S Measuring Receiver			



ANSI National Accreditation Board

Electrical – RF/Microwave

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
PM Distortion Measure <sup>1</sup>	(6.6 to 13.2) GHz		
	Rate (500 to 1 000) Hz		
	Dev > 2.5 rad		
	(0 to -20) dB	0.09 dB	
	(-20 to -30) dB	0.27 dB	
	(-30 to -40) dB	0.83 dB	
	(-40 to -50) dB	2.3 dB	
	(13.2 to 31.15) GHz		
	Rate (20 to 500) Hz		
	Dev > 3.5 rad		
	(0 to -20) dB	0.26 dB	
	(-20 to -30) dB	0.79 dB	
	(-30 to -40) dB	2.3 dB	
	Dev > 10.0 rad		
	(0 to -20) dB	0.09 dB	
	(-20 to -30) dB	0.27 dB	
	(-30 to -40) dB	0.83 dB	
	(-40 to -50) dB	2.3 dB	
	Rate (500 to 1 000) Hz		
	Dev > 1.2 rad		
	(0 to -20) dB	0.26 dB	
	(-20 to -30) dB	0.79 dB	
	(-30 to -40) dB	2.3 dB	
	Dev > 4.0 rad		
(0 to -20) dB	0.09 dB		
(-20 to -30) dB	0.27 dB		
(-30 to -40) dB	0.83 dB		
(-40 to -50) dB	2.3 dB		
(31.15 to 50.0) GHz			
Rate 20 to 500 Hz			
Dev > 7.5 rad			
(0 to -20) dB	0.26 dB		
(-20 to -30) dB	0.79 dB		
(-30 to -40) dB	2.3 dB		
Dev > 19.0 rad			
(0 to -20) dB	0.09 dB		
(-20 to -30) dB	0.27 dB		
(-30 to -40) dB	0.83 dB		
(-40 to -50) dB	2.3 dB		

Agilent N5531S Measuring Receiver



ANSI National Accreditation Board

Electrical – RF/Microwave

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
PM Distortion Measure <sup>1</sup>	(31.15 to 50.0) GHz		Agilent N5531S Measuring Receiver
	Rate (500 to 1 000) Hz		
	Dev > 3.0 rad	0.26 dB	
	(0 to -20) dB	0.79 dB	
	(-20 to -30) dB	2.3 dB	
	(-30 to -40) dB		
	Dev > 8.0 rad		
Total Harmonic Distortion (THD)	(0 to -60) dB		HP 8903B Audio Analyzer
	20 Hz to 20 kHz		
	(0 to -40) dB	1 dB	
	(-40 to -50) dB	1 dB	
	(-50 to -60) dB	1.3 dB	
	(-60 to -65) dB	1.7 dB	
	(20 to 50) kHz		
	(0 to -40) dB	2 dB	
	(-40 to -50) dB	2.1 dB	
	(-50 to -60) dB	3 dB	
	(50 to 100) kHz		
	(0 to -40) dB	2 dB	
	(-40 to -50) dB	2.4 dB	
	Harmonics Measure <sup>1</sup>	(-10 to -80) dB	
2 <sup>nd</sup> through 5 <sup>th</sup> Harmonic			
1kHz to 600MHz		0.37 dB	
(600 to 1 320) MHz		1.1 dB	
(1 320 to 2 200) MHz		1.4 dB	
(2 200 to 3 000) MHz		1.4 dB	
(3 000 to 4 400) MHz		1.7 dB	
(4 400 to 5 300) MHz		1.9 dB	
(5 300 to 10 000) MHz		2.1 dB	
2 <sup>nd</sup> through 4 <sup>th</sup> Harmonic			
(10 000 to 12 500) MHz		2.1 dB	
2 <sup>nd</sup> through 3 <sup>rd</sup> Harmonic			
(12 500 to 16 667) MHz	2.1 dB		
2 <sup>nd</sup> Harmonic			
(16 667 to 25 000) MHz	2.3 dB		

**Length – Dimensional Metrology**

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Calipers <sup>1,2</sup>	Up to 46 in	$(280 + 1.8L) \mu\text{in}$	Grade 2 gage blocks
Micrometers <sup>1,2</sup>	Up to 46 in	$(31 + 4.3L) \mu\text{in}$	Grade 2 gage blocks
Micrometers Standards	(1 to 12) in	$(1.7 + 6.8L) \mu\text{in}$	Universal measuring machine, master gage block set
Bore Micrometers	(0.125 to 3.5) in	92 $\mu\text{in}$	Master ring gages
Dial Indicators <sup>1,2</sup>	Up to 0.2 in Up to 6 in Up to 1 in	15 $\mu\text{in}$ $(81 + 1.2L) \mu\text{in}$ 150 $\mu\text{in}$	Grade 2 gage blocks Starrett 716 Indicator Tester
Height Gages <sup>1,2</sup>	Up to 46 in	$(202 + 1.7L) \mu\text{in}$	Grade 2 gage blocks
Scales – Rulers <sup>1</sup>	Up to 46 in	0.009 in	Grade 2 gage blocks
Feeler Gages <sup>1</sup>	Up to 1 in	44 $\mu\text{in}$	Mitutoyo 293-369 Micrometer
Surface Plates <sup>1</sup> – Overall Flatness Local Area Flatness	(18 x 18) in to (6x6) ft (-0.001 to 0.001) in	79 $\mu\text{in}$ 68 $\mu\text{in}$	Mahr Leveling system Repeat-o-meter
Gage Blocks <sup>2</sup>	Up to 12 in	$(3.2 + 2.7L) \mu\text{in}$	Universal measuring machine, master gage block set
Protractors <sup>1</sup>	(0 to 360) <sup>o</sup>	0.014 <sup>o</sup>	Angle blocks
Radius Gages	(0.01 to 1) in	290 $\mu\text{in}$	Optical comparator
Cylindrical Gages <sup>1,2</sup> – Plain Pin, Plugs Rings	(0 to 12) in (0.04 to 14) in	$(5.4 + 2.2D) \mu\text{in}$ $(17 + 2.2D) \mu\text{in}$	Master gage blocks, P&W universal measuring machine
Thread Plugs – Major Pitch Diameter (6 to 80) TPI	Up to 12 in Up to 12 in	36 $\mu\text{in}$ 91 $\mu\text{in}$	Gage blocks, P & W universal measuring machine, Van Keuren thread wire set
Thread Rings <sup>2,3</sup>	Up to 12 in	$(350 + 47D) \mu\text{in}$	Thread setting plug gages



ANSI National Accreditation Board

Mass and Mass Related

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Scales & Balances <sup>1,2,4</sup>	1 g to 40 kg	(0.042 + 0.0045X) mg	Class 1 weights
	Up to 1 100 lb Up to 500 kg	(0.0003 + 0.00012W) lb (0.13 + 0.00012X) g	Class F weights
Pressure <sup>1</sup> – Measure	(0 to 30) inH2O	0.01 % of reading or 0.000 75 inH2O whichever is greater	Low Pressure Controller/Calibrator Ruska 7250LP
	(-150 to 150) inH2O (0 to 30) psig	0.073 inH2O 0.008 psi	Pressure Gage Additel ADT686-DP150 Additel ADT686-GP30
	(0 to 23) psia (0 to 30) inHg	0.02 psi 0.04 inHg	High Accuracy Pressure Standard Paroscientific 760-23A Paroscientific 760-23A
	(0 to 5 000) psig (0 to 10 000) psig	1.3 psi 2.6 psi	Pressure Gage Additel ADT686-GP5K Additel ADT686-GP10K
Pressure <sup>1</sup> – Generate	(10 to 16 000) psi	0.019 % of reading	Fluke P3125-PSI Deadweight Tester
Vacuum Gages	(0 to 30) inHg	0.075 inHg	Paroscientific 760-23A Pressure Standard
Torque Tools <sup>1</sup>	(4 to 1 000) lbf·in (25 to 1 000) lbf·ft	0.4 % of reading 0.49 % of reading	CDI torque system
Torque Analyzers	1 ozf·in to 250 lbf·ft	0.11 % of reading	Weights and arm
Force Gages Tension & Compression <sup>1</sup>	10 mgf to 540 lbf	0.06 % of reading	Class 1 and Class F weights
Mass – Class F	(5 to 225) g (2 to 10) lb (25 to 50) lb	0.49 mg 0.028 g 0.3 g	Master balances Class 1 Weights
Rockwell Hardness Testers <sup>1</sup>	(20 to 69) HRA (70 to 79) HRA (80 to 86) HRA (0 to 59) HRBW (60 to 79) HRBW (80 to 100) HRBW (20 to 39) HRC (40 to 59) HRC (60 to 70) HRC	1.3 HRA 1.3 HRA 0.73 HRA 3 HRBW 2.9 HRBW 1.2 HRBW 1.2 HRC 1.2 HRC 0.69 HRC	Indirect Verification using Hardness blocks



ANSI National Accreditation Board

**Mass and Mass Related**

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Durometers – Scale Accuracy (force) Types A, B, C, D, DO, O Type M, Types OO, OOO Indenter Geometry Length Diameter Angle	(0 to 100) duros  0.1 in 0.05 in (30 to 35)°	0.06 duros 0.09 duros 0.07 duros  180 μin 180 μin 0.006°	Direct Verification  Master balance  Optical comparator
Mass - Fixed Points Metric	(1,2) g 10 g 20 g 50 g 100 g 200 g 500 g 2 kg 5 kg 10 kg 20 kg	0.07 mg 0.15 mg 0.19 mg 0.3 mg 0.56 mg 1.2 mg 6.5 mg 14 mg 31 mg 82 mg 114 mg	Comparison to ASTM E617 Class 1 weights

**Thermodynamic**

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Temperature – Measure <sup>1</sup>	(-196 to 200) °C (200 to 420) °C	0.027 °C 0.038 °C	Hart 1502 Indicator with Fluke 5615 PRT
Temperature – Measuring Equipment	(-40 to 160) °C (160 to 420) °C	0.056 °C 0.067 °C	Hart 1502 Indicator with Fluke 5615 PRT and Additel Dry Well
	(420 to 660) °C	0.31 °C	Additel Dry Well
Humidity – Measure <sup>1</sup>	(0 to 90) % RH	1.7 % RH	Vaisala HM141/HMP46 Humidity Indicator and Probe
Humidity - Generate	(10 to 95) %RH	0.5 % RH	Thunder Scientific 2500



ANSI National Accreditation Board

**Thermodynamic**

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Infrared (IR) Thermometry <sup>1</sup>	(20 to 100) °C (100 to 300) °C (300 to 420) °C (420 to 500) °C	1.5 °C 4.3 °C 6 °C 7.7 °C	Fluke 9132 Infrared Calibrator ε= 0.95, λ = (8 to 14) μm

**Time and Frequency**

Peachtree Corners, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency – Generate <sup>1</sup>	10 MHz	1 x 1 <sup>e-12</sup> Hz/Hz	HP 58503A GPS Receiver
	(1 to 10) Hz (10 to 100) Hz (100 to 1 000) Hz (1 to 10) kHz (10 to 100) kHz (0.1 to 1) MHz (1 to 10) MHz	1 x 1 <sup>e-12</sup> Hz/Hz + 0.57 μHz 1 x 1 <sup>e-12</sup> Hz/Hz + 5.7 μHz 1 x 1 <sup>e-12</sup> Hz/Hz + 57 μHz 1 x 1 <sup>e-12</sup> Hz/Hz + 0.57 mHz 1 x 1 <sup>e-12</sup> Hz/Hz + 5.7 mHz 1 x 1 <sup>e-12</sup> Hz/Hz + 57 mHz 1 x 1 <sup>e-12</sup> Hz/Hz + 0.57 Hz	Agilent 33250A Function Generator / HP 58503A GPS Receiver
	(10 to 50 000) MHz	1 x 1 <sup>e-12</sup> Hz/Hz + 0.57 mHz	Agilent E8257D Opt 550 Signal Generator / HP 58503A GPS Receiver
	Time – Generate	1 pps	1 x 1 <sup>e-12</sup> s/s + 750 ps
Frequency – Measure	1 to 10 Hz 10 to 100 Hz 100 to 1 000 Hz 1 to 10 kHz 10 to 100 kHz 100 to 200 kHz 0.2 to 3 000 MHz	4.20 x 1 <sup>e-9</sup> Hz/Hz 1.47 x 1 <sup>e-9</sup> Hz/Hz 0.60 x 1 <sup>e-9</sup> Hz/Hz 0.33 x 1 <sup>e-9</sup> Hz/Hz 0.24 x 1 <sup>e-9</sup> Hz/Hz 0.21 x 1 <sup>e-9</sup> Hz/Hz 0.21 x 1 <sup>e-9</sup> Hz/Hz	Agilent 53131A Opt 030 Frequency Counter / HP 58503A GPS Receiver
Frequency – Measure <sup>1</sup>	50 to 46 000 MHz	1 x 1 <sup>e-12</sup> Hz/Hz + 1.6 Hz	Agilent 53152A Frequency Counter / HP 58503A GPS Receiver
Frequency – Measure <sup>1</sup>	10 to 50 000 MHz	1 x 1 <sup>e-12</sup> Hz/Hz + 0.1 Hz	Agilent E4448A Spectrum Analyzer / HP 58503A GPS Receiver
Tachometers – RPM <sup>1,2</sup>	Up to 100 000 RPM	0.01 of reading + 0.6R	HP 33250A Signal Generator & LED

## DIMENSIONAL MEASUREMENT

### 1 Dimensional

Peachtree Corners, GA

Specific Tests and / or Properties Measured	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Length (distance)	X Axis (0.01 to 8) in Y Axis (0.01 to 4) in	180 $\mu$ in 180 $\mu$ in	Optical comparator

### 2 Dimensional

Peachtree Corners, GA

Specific Tests and / or Properties Measured	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Angle	Up to 360°	0.006°	Optical comparator

[Return to Site listing](#) (top)    [Go to Notes](#) (bottom)





ANSI National Accreditation Board

### Services performed at satellite laboratory

#### Technical Maintenance, Inc.

117 Jetplex Circle, Suite C4  
Madison, AL 35758

Donny Prax (Branch Manager) Phone: 256-772-4115  
Scott Chamberlain (Quality Manager) Phone: 321-242-0890

## CALIBRATION AND DIMENSIONAL MEASUREMENT

### CALIBRATION

#### Acoustics and Vibration

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Accelerometers – Acceleration <sup>1</sup>	(0.01 to 10) g (7 to 100) Hz (100 to 2 500) Hz (2.5 to 10) kHz	1.5 % of reading 1.2 % of reading 2.5 % of reading	Accelerometer Calibrator
Sound Pressure Level – Generate <sup>1</sup>	94 dB @ 1 kHz 114 dB @ 1 kHz	0.65 dB 0.97 dB	Extech 407766 Sound Level Calibrator

#### Chemical Quantities

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
pH Meters <sup>1,5</sup>	4 pH 7 pH 10 pH	0.016 pH 0.025 pH 0.061 pH	pH buffer solutions
Conductivity Meters <sup>1,5</sup>	100 µS/cm 1 410 µS/cm 10 000 µS/cm	0.83 µS/cm 5.3 µS/cm 44 µS/cm	Conductivity solutions

**Electrical – DC/Low Frequency**

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Voltage – Generate <sup>1</sup>	(0 to 220) mV (0.22 to 2.2) V (2.2 to 11) V (11 to 22) V (22 to 220) V (220 to 1 100) V	6.8 $\mu\text{V/V} + 0.8 \mu\text{V}$ 4.6 $\mu\text{V/V} + 0.9 \mu\text{V}$ 3 $\mu\text{V/V} + 2.5 \mu\text{V}$ 3 $\mu\text{V/V} + 3.9 \mu\text{V}$ 4.6 $\mu\text{V/V} + 38 \mu\text{V}$ 6.1 $\mu\text{V/V} + 385 \mu\text{V}$	Fluke 5730A Multifunction Calibrator
DC Voltage – Measure <sup>1</sup>	(0 to 100) mV (0.1 to 1) V (1 to 10) V (10 to 100) V (100 to 1 000) V	5.0 $\mu\text{V/V} + 1.2 \mu\text{V}$ 4.0 $\mu\text{V/V} + 1.2 \mu\text{V}$ 4.0 $\mu\text{V/V} + 2.1 \mu\text{V}$ 6.0 $\mu\text{V/V} + 30 \mu\text{V}$ 6.0 $\mu\text{V/V} + 100 \mu\text{V} + 12 \mu\text{V/V} \times (\text{Vin}/1000) ^2$	Agilent 3458A Option 002 Multimeter
DC Voltage – Measure <sup>1</sup>	(1 to 30) kV	0.1 % of reading	Ross VD30 High Voltage Divider, Fluke 89 IV Multimeter
DC Voltage – Measure <sup>1</sup>	(1 to 150) kV	0.1 % of reading	Ross VD150 High Voltage Divider, HP 34401A Multimeter
DC Current – Generate <sup>1</sup>	(1 to 2.2) nA (2.2 to 22) nA (22 to 220) nA (0.22 to 2.2) $\mu\text{A}$ (2.2 to 10) $\mu\text{A}$	93 $\mu\text{A/A} + 0.007 \text{ nA}$ 92 $\mu\text{A/A} + 0.007 \text{ nA}$ 92 $\mu\text{A/A} + 0.01 \text{ nA}$ 36 $\mu\text{A/A} + 0.1 \text{ nA}$ 15 $\mu\text{A/A} + 1 \text{ nA}$	Fluke 5730A & 5522A Multifunction Calibrator
	(10 to 220) $\mu\text{A}$ 220 $\mu\text{A}$ to 2.2 mA (2.2 to 22) mA (22 to 100) mA (100 to 220) mA (0.22 to 1) A (1 to 2.2) A	38 $\mu\text{A/A} + 5 \text{ nA}$ 30 $\mu\text{A/A} + 7 \text{ nA}$ 30 $\mu\text{A/A} + 44 \text{ nA}$ 38 $\mu\text{A/A} + 0.7 \mu\text{A}$ 45 $\mu\text{A/A} + 0.7 \mu\text{A}$ 68 $\mu\text{A/A} + 12 \mu\text{A}$ 105 $\mu\text{A/A} + 12 \mu\text{A}$	Fluke 5730A Multifunction Calibrator
	(2.2 to 11) A	274 $\mu\text{A/A} + 365 \mu\text{A}$	Fluke 5730A/03 Multifunction Calibrator 5725A Amplifier
	(11 to 20.5) A	761 $\mu\text{A/A} + 578 \mu\text{A}$	Fluke 5522A Multiproduct Calibrator
	DC Current Clamp Meters <sup>1</sup> Toroidal-Wound	(0 to 1 025) A	0.21 % of output + 0.05A
DC Current Clamp Meters <sup>1</sup> Other	(0 to 1 025) A	0.39 % of output + 0.38A	Fluke 5522A Multiproduct Calibrator/5500A/Coil x50



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Current – Measure <sup>1</sup>	(1 to 10) nA (10 to 100) nA (0.1 to 1) $\mu$ A (1 to 10) $\mu$ A	35 $\mu$ A/A + 0.12 pA 13 $\mu$ A/A + 1.2 pA 9.1 $\mu$ A/A + 0.01 nA 7.9 $\mu$ A/A + 0.12 nA	Fluke 5730A Multifunction Calibrator Agilent 3458A Option 002 Multimeter
DC Current – Measure <sup>1</sup>	(10 to 100) $\mu$ A (0.1 to 1) mA (1 to 10) mA (10 to 100) mA (0.1 to 1) A	20 $\mu$ A/A + 0.8 nA 20 $\mu$ A/A + 5 nA 20 $\mu$ A/A + 0.05 $\mu$ A 35 $\mu$ A/A + 0.5 $\mu$ A 110 $\mu$ A/A + 10.1 $\mu$ A	Agilent 3458A Multimeter
DC Current – Measure <sup>1</sup>	(1 to 3) A (3 to 10) A	761 $\mu$ A/A + 462 $\mu$ A 1.1 mA/A + 614 $\mu$ A	Fluke 8845A Multimeter
	(1 to 600) A	0.27 % of reading	Current Shunts
AC Voltage – Generate <sup>1</sup>	(0.22 to 2.2) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	228 $\mu$ V/V + 3.9 $\mu$ V 88 $\mu$ V/V + 3.9 $\mu$ V 76 $\mu$ V/V + 3.9 $\mu$ V 190 $\mu$ V/V + 3.9 $\mu$ V 457 $\mu$ V/V + 4.6 $\mu$ V 989 $\mu$ V/V + 9.2 $\mu$ V 1.3 mV/V + 19 $\mu$ V 2.6 mV/V + 19 $\mu$ V	Fluke 5730A Multifunction Calibrator
	(2.2 to 22) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	228 $\mu$ V/V + 3.9 $\mu$ V 88 $\mu$ V/V + 3.9 $\mu$ V 76 $\mu$ V/V + 3.9 $\mu$ V 190 $\mu$ V/V + 3.9 $\mu$ V 457 $\mu$ V/V + 4.6 $\mu$ V 989 $\mu$ V/V + 9.2 $\mu$ V 1.3 mV/V + 19 $\mu$ V 2.6 mV/V + 19 $\mu$ V	
	(22 to 220) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	228 $\mu$ V/V + 11.4 $\mu$ V 88 $\mu$ V/V + 6.1 $\mu$ V 53 $\mu$ V/V + 6.1 $\mu$ V 114 $\mu$ V/V + 6.1 $\mu$ V 304 $\mu$ V/V + 15.2 $\mu$ V 609 $\mu$ V/V + 19 $\mu$ V 1.3 mV /V + 22.8 $\mu$ V 2.5 mV /V + 45.7 $\mu$ V	



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Generate <sup>1</sup>	(0.22 to 2.2) V		Fluke 5730A Multifunction Calibrator
	(10 to 20) Hz	228 $\mu\text{V/V}$ + 38 $\mu\text{V}$	
	(20 to 40) Hz	84 $\mu\text{V/V}$ + 15 $\mu\text{V}$	
	40 Hz to 20 kHz	37 $\mu\text{V/V}$ + 8 $\mu\text{V}$	
	(20 to 50) kHz	61 $\mu\text{V/V}$ + 9 $\mu\text{V}$	
	(50 to 100) kHz	76 $\mu\text{V/V}$ + 30 $\mu\text{V}$	
	(100 to 300) kHz	304 $\mu\text{V/V}$ + 76 $\mu\text{V}$	
	(300 to 500) kHz	913 $\mu\text{V/V}$ + 190 $\mu\text{V}$	
	500 kHz to 1 MHz	1.5 mV/V + 304 $\mu\text{V}$	
	(2.2 to 22) V		
	(10 to 20) Hz	228 $\mu\text{V/V}$ + 380 $\mu\text{V}$	
	(20 to 40) Hz	84 $\mu\text{V/V}$ + 152 $\mu\text{V}$	
	40 Hz to 20 kHz	37 $\mu\text{V/V}$ + 54 $\mu\text{V}$	
	(20 to 50) kHz	61 $\mu\text{V/V}$ + 91 $\mu\text{V}$	
	(50 to 100) kHz	76 $\mu\text{V/V}$ + 190 $\mu\text{V}$	
(100 to 300) kHz	228 $\mu\text{V/V}$ + 609 $\mu\text{V}$		
(300 to 500) kHz	913 $\mu\text{V/V}$ + 1.9 mV		
500 kHz to 1 MHz	1.4 mV/V + 3 mV		
(22 to 220) V			
(10 to 20) Hz	228 $\mu\text{V/V}$ + 3.8 mV		
(20 to 40) Hz	84 $\mu\text{V/V}$ + 1.5 mV		
40 Hz to 20 kHz	49 $\mu\text{V/V}$ + 0.6 mV		
(20 to 50) kHz	76 $\mu\text{V/V}$ + 0.9 mV		
(50 to 100) kHz	137 $\mu\text{V/V}$ + 2.3 mV		
AC Voltage – Generate <sup>1</sup>	(220 to 750) V		Fluke 5730A Multifunction Calibrator /5725A Amplifier
	40 Hz to 1 kHz	68 $\mu\text{V/V}$ + 3 mV	
	(1 to 20) kHz	126 $\mu\text{V/V}$ + 5 mV	
	(30 to 50) kHz	457 $\mu\text{V/V}$ + 8 mV	
	(50 to 100) kHz	1.8 mV/V + 34 mV	
	(750 to 1 000) V		
	40 Hz to 1 kHz	68 $\mu\text{V/V}$ + 3 mV	
(1 to 20) kHz	126 $\mu\text{V/V}$ + 5 mV		
(20 to 30) kHz	457 $\mu\text{V/V}$ + 8 mV		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Generate <sup>1</sup> Wideband Absolute	(0.3 to 1.1) mV		Fluke 5730A Multifunction Calibrator
	(10 to 30 Hz)	0.65 % of output + 1.5 μV	
	30 Hz to 500 kHz	0.61 % of output + 1.5 μV	
	(0.5 to 1.2) MHz	0.63 % of output + 3.8 μV	
	(1.2 to 2) MHz	0.63 % of output + 3.8 μV	
	(2 to 12) MHz	0.68 % of output + 3.8 μV	
	(12 to 20) MHz	0.76 % of output + 3.8 μV	
	(20 to 30) MHz	1.3 % of output + 12.9 μV	
	(1.1 to 3.3) mV		
	(10 to 30 Hz)	0.58 % of output + 2.3 μV	
	30 Hz to 500 kHz	0.53 % of output + 2.3 μV	
	(0.5 to 1.2) MHz	0.54% of output + 4.6 μV	
	(1.2 to 2) MHz	0.54 % of output + 4.6 μV	
	(2 to 12) MHz	0.58 % of output + 4.6 μV	
	(12 to 20) MHz	0.65 % of output + 4.6 μV	
	(20 to 30) MHz	1.3 % of output + 4.6 μV	
	(3.3 to 11) mV		
	(10 to 30 Hz)	0.58 % of output + 6.1 μV	
	30 Hz to 500 kHz	0.53 % of output + 6.1 μV	
	(0.5 to 1.2) MHz	0.54 % of output + 8.4 μV	
	(1.2 to 2) MHz	0.54 % of output + 8.4 μV	
	(2 to 12) MHz	0.55 % of output + 8.4 μV	
	(12 to 20) MHz	0.61 % of output + 8.4 μV	
	(20 to 30) MHz	0.93 % of output + 8.4 μV	
	(11 to 33) mV		
	(10 to 30 Hz)	0.52 % of output + 12 μV	
	30 Hz to 500 kHz	0.46 % of output + 12 μV	
	(0.5 to 1.2) MHz	0.47 % of output + 14 μV	
	(1.2 to 2) MHz	0.47 % of output + 14 μV	
	(2 to 12) MHz	0.49 % of output + 14 μV	
(12 to 20) MHz	0.55 % of output + 14 μV		
(20 to 30) MHz	0.89 % of output + 14 μV		
(33 to 110) mV			
(10 to 30 Hz)	0.52 % of output + 30 μV		
30 Hz to 500 kHz	0.46 % of output + 30 μV		
(0.5 to 1.2) MHz	0.47 % of output + 33 μV		
(1.2 to 2) MHz	0.47 % of output + 33 μV		
(2 to 12) MHz	0.49 % of output + 33 μV		
(12 to 20) MHz	0.55 % of output + 33 μV		
(20 to 30) MHz	0.89 % of output + 33 μV		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Generate <sup>1</sup> Wideband Absolute	(110 to 330) mV		Fluke 5730A Multifunction Calibrator
	(10 to 30 Hz)	0.45 % of output + 0.1 mV	
	30 Hz to 500 kHz	0.38 % of output + 0.1 mV	
	(0.5 to 1.2) MHz	0.4 % of output + 0.1 mV	
	(1.2 to 2) MHz	0.4 % of output + 0.1 mV	
	(2 to 12) MHz	0.42 % of output + 0.1 mV	
	(12 to 20) MHz	0.49 % of output + 0.1 mV	
	(20 to 30) MHz	0.85 % of output + 0.1 mV	
	(0.33 to 1.1) V		
	(10 to 30 Hz)	0.45 % of output + 0.3 mV	
	30 Hz to 500 kHz	0.38 % of output + 0.3 mV	
	(0.5 to 1.2) MHz	0.4 % of output + 0.3 mV	
	(1.2 to 2) MHz	0.4 % of output + 0.3 mV	
	(2 to 12) MHz	0.42 % of output + 0.3 mV	
	(12 to 20) MHz	0.49 % of output + 0.3 mV	
(20 to 30) MHz	0.85 % of output + 0.3 mV		
(1.1 to 3.5) V			
(10 to 30 Hz)	0.39 % of output + 0.4 mV		
30 Hz to 500 kHz	0.3 % of output + 0.4 mV		
(0.5 to 1.2) MHz	0.32 % of output + 0.4 mV		
(1.2 to 2) MHz	0.32 % of output + 0.4 mV		
(2 to 12) MHz	0.35 % of output + 0.4 mV		
(12 to 20) MHz	0.44 % of output + 0.4 mV		
(20 to 30) MHz	0.82 % of output + 0.4 mV		
AC Voltage – Measure <sup>1</sup>	(1 to 10) mV		Agilent 3458A Multimeter
	(1 to 40) Hz	0.03 % of reading + 3 μV	
	40 Hz to 1 kHz	0.02 % of reading + 1.1 μV	
	(1 to 20) kHz	0.03 % of reading + 1.1 μV	
	(20 to 50) kHz	0.1 % of reading + 1.1 μV	
	(50 to 100) kHz	0.5 % of reading + 1.1 μV	
	100 kHz to 1 MHz	1.2% of reading + 5 μV	
(1 to 4) MHz	7% of reading + 7 μV		
(4 to 8) MHz	20% of reading + 8 μV		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure <sup>1</sup>	(10 to 100) mV		Agilent 3458A Multimeter
	(1 to 40) Hz	0.007 % of reading + 4 $\mu$ V	
	40 Hz to 1 kHz	0.007 % of reading + 2 $\mu$ V	
	(1 to 20) kHz	0.014 % of reading + 2 $\mu$ V	
	(20 to 50) kHz	0.03 % of reading + 2 $\mu$ V	
	(50 to 100) kHz	0.08 % of reading + 2 $\mu$ V	
	(100 to 300) kHz	0.3 % of reading + 10 $\mu$ V	
	300 kHz to 1 MHz	1 % of reading + 10 $\mu$ V	
	(1 to 2) MHz	1.5 % of reading + 70 $\mu$ V	
	(2 to 4) MHz	4 % of reading + 70 $\mu$ V	
	(4 to 8) MHz	4 % of reading + 80 $\mu$ V	
	(8 to 10) MHz	15 % of reading + 100 $\mu$ V	
	(0.1 to 1) V		
	(1 to 40) Hz	0.007 % of reading + 40 $\mu$ V	
	40 Hz to 1 kHz	0.07 % of reading + 20 $\mu$ V	
	(1 to 20) kHz	0.014 % of reading + 20 $\mu$ V	
	(20 to 50) kHz	0.03 % of reading + 20 $\mu$ V	
	(50 to 100) kHz	0.08 % of reading + 20 $\mu$ V	
	(100 to 300) kHz	0.3 % of reading + 100 $\mu$ V	
	300 kHz to 1 MHz	10 % of reading + 100 $\mu$ V	
	(1 to 2) MHz	1.5 % of reading + 0.7 mV	
	(2 to 4) MHz	4 % of reading + 0.7 mV	
	(4 to 8) MHz	4 % of reading + 0.8 mV	
	(8 to 10) MHz	15 % of reading + 1 mV	
	(1 to 10) V		
	(1 to 40) Hz	0.007 % of reading + 0.4 mV	
	40 Hz to 1 kHz	0.007 % of reading + 0.2 mV	
	(1 to 20) kHz	0.014 % of reading + 0.2 mV	
	(20 to 50) kHz	0.03 % of reading + 0.2 mV	
	(50 to 100) kHz	0.08 % of reading + 0.2 mV	
(100 to 300) kHz	0.3 % of reading + 1 mV		
300 kHz to 1 MHz	10 % of reading + 1 mV		
(1 to 2) MHz	1.5 % of reading + 7 mV		
(2 to 4) MHz	40 % of reading + 7 mV		
(4 to 8) MHz	40 % of reading + 8 mV		
(8 to 10) MHz	15 % of reading + 10 mV		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure <sup>1</sup>	(10 to 100) V (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz	0.02 % of reading + 4 mV 0.02 % of reading + 2 mV 0.02 % of reading + 2 mV 0.035 % of reading + 2 mV 0.12 % of reading + 2 mV 0.4 % of reading + 10 mV 1.5 % of reading + 10 mV	Agilent 3458A/002 Precision Multimeter
	(100 to 700) V (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz	0.04 % of reading + 40 mV 0.04 % of reading + 20 mV 0.06 % of reading + 20 mV 0.12 % of reading + 20 mV 0.3 % of reading + 20 mV	
AC Voltage – Measure <sup>1</sup>	(1 to 21) kV 60 Hz	0.57 % of reading + 0.002 kV	Ross VD30 High Voltage Divider, Fluke 89IV Multimeter
AC Voltage – Measure <sup>1</sup>	(1 to 100) kV 60 Hz	0.5 % of reading + 0.02 kV	Ross VD150 High Voltage Divider, HP 34401A Multimeter
AC Current – Generate <sup>1</sup>	(9 to 220) $\mu$ A (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	228 $\mu$ A/A + 15 nA 152 $\mu$ A/A + 10 nA 91 $\mu$ A/A + 8 nA 266 $\mu$ A/A + 12 nA 989 $\mu$ A/A + 61 nA	Fluke 5730A Multifunction Calibrator
	(0.22 to 2.2) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	228 $\mu$ A/A + 39 nA 152 $\mu$ A/A + 31 nA 91 $\mu$ A/A + 31 nA 183 $\mu$ A/A + 99 nA 989 $\mu$ A/A + 609 nA	
AC Current – Generate <sup>1</sup>	(2.2 to 22) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	228 $\mu$ A/A + 385 nA 152 $\mu$ A/A + 310 nA 91 $\mu$ A/A + 310 nA 183 $\mu$ A/A + 536 nA 989 $\mu$ A/A + 4.6 nA	

**Electrical – DC/Low Frequency**

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Generate <sup>1</sup>	(22 to 220) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	228 $\mu$ A/A + 4 $\mu$ A 152 $\mu$ A/A + 3 $\mu$ A 91 $\mu$ A/A + 2 $\mu$ A 183 $\mu$ A/A + 3 $\mu$ A 989 $\mu$ A/A + 9 $\mu$ A	Fluke 5730A Multifunction Calibrator
	(0.22 to 2.2) A 20 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	228 $\mu$ A/A + 31 $\mu$ A 380 $\mu$ A/A + 76 $\mu$ A 6.1 mA/A + 152 $\mu$ A	
AC Current – Generate <sup>1</sup>	(2.2 to 11) A 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	350 $\mu$ A /A + 141 $\mu$ A 723 $\mu$ A /A + 295 $\mu$ A 2.7 mA /A + 573 $\mu$ A	Fluke 5730A Multifunction Calibrator /5725A Amplifier
AC Current – Generate <sup>1</sup>	(11 to 20.5) A 45 Hz to 100 Hz 100 Hz to 1 kHz (1 to 5) kHz	0.09 % of output + 5 mA 0.11 % of output + 5 mA 2.3 % of output + 5 mA	Fluke 5522A Multiproduct Calibrator
AC Current – Generate <sup>1</sup> Clamps – Toroidal-Wound	(16.5 to 1 025) A (45 to 65) Hz (16.5 to 150) A (65 to 440) Hz	0.23 % of output + 0.1 A 0.64 % of output + 0.085 A	Fluke 5522A Multiproduct Calibrator, 5500A/Coil x50
AC Current – Generate <sup>1</sup> Clamps – Other	(16.5 to 1 025) A (45 to 65) Hz (16.5 to 150) A (65 to 440) Hz	0.44 % of output + 0.53 A 0.79 % of output + 0.68 A	Fluke 5522A Multiproduct Calibrator, 5500A/Coil x50
AC Current – Measure <sup>1</sup>	(5 to 100) $\mu$ A (10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz 100 $\mu$ A to 1 mA (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz (1 to 10) mA (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz	0.4 % of reading + 0.03 $\mu$ A 0.15 % of reading + 0.03 $\mu$ A 0.06 % of reading + 0.03 $\mu$ A  0.4 % of reading + 0.2 $\mu$ A 0.15 % of reading + 0.2 $\mu$ A 0.06 % of reading + 0.2 $\mu$ A 0.03 % of reading + 0.2 $\mu$ A  0.4 % of reading + 2 $\mu$ A 0.15 % of reading + 2 $\mu$ A 0.06 % of reading + 2 $\mu$ A 0.03 % of reading + 2 $\mu$ A	Agilent 3458A/002 Precision Multimeter

**Electrical – DC/Low Frequency**

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Measure <sup>1</sup>	(10 to 100) mA (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz	0.4 % of reading + 20 $\mu$ A 0.15 % of reading + 20 $\mu$ A 0.06 % of reading + 20 $\mu$ A 0.03 % of reading + 20 $\mu$ A	Agilent 3458A/002 Precision Multimeter
	100 mA to 1 A (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz	0.4 % of reading + 0.2 mA 0.15 % of reading + 0.2 mA 0.06 % of reading + 0.2 mA 0.03 % of reading + 0.2 mA	
AC Current – Measure <sup>1</sup>	(1 to 3) A (3 to 5) Hz (5 to 10) Hz 10 Hz to 5 kHz	0.72 % of reading + 1.2 mA 0.23 % of reading + 1.2 mA 0.1 % of reading + 1.2 mA	Agilent 34401A Multimeter
AC Current – Measure <sup>1</sup>	(3 to 30) A 40 Hz to 1 kHz (1 to 5) kHz	0.3 % of reading + 0.07 A 5 % of reading + 0.14 A	Agilent 3458A/002 Precision Multimeter Keysight 34330A Current Shunt
AC Current – Measure <sup>1</sup>	(10 to 1 000) A (10 to 100) Hz (100 to 500) Hz	1.5 % of reading + 1 A 1.9 % of reading + 1 A	Fluke 376 Clamp Meter
AC Current – Measure <sup>1</sup>	(100 to 2 500) A (10 to 500) Hz	2.3 % of reading + 5 A	Fluke 376 Clamp Meter w/i2500 flex probe
Resistance – Generate <sup>1</sup> Fixed Points	(1, 1.9) $\Omega$ (10, 19) $\Omega$ (100, 190) $\Omega$ (1, 1.9) k $\Omega$ (10, 19) k $\Omega$ 100 k $\Omega$ 190 k $\Omega$ 1 M $\Omega$ 1.9 M $\Omega$ 10 M $\Omega$ 19 M $\Omega$ 100 M $\Omega$	84 $\mu\Omega/\Omega$ + 0.1 $\mu\Omega$ 21 $\mu\Omega/\Omega$ + 1 $\mu\Omega$ 9.1 $\mu\Omega/\Omega$ + 6 $\mu\Omega$ 6.1 $\mu\Omega/\Omega$ + 60 $\mu\Omega$ 6.1 $\mu\Omega/\Omega$ + 0.6 m $\Omega$ 7.6 $\mu\Omega/\Omega$ + 6 m $\Omega$ 9.1 $\mu\Omega/\Omega$ + 6 m $\Omega$ 11.4 $\mu\Omega/\Omega$ + 60 m $\Omega$ 16 $\mu\Omega/\Omega$ + 60 m $\Omega$ 35 $\mu\Omega/\Omega$ + 0.6 $\Omega$ 42 m $\Omega/\Omega$ + 0.6 $\Omega$ 91 m $\Omega/\Omega$ + 6 $\Omega$	Fluke 5730A Multifunction Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Resistance – Generate <sup>1</sup>	Up to 11 Ω (11 to 33) Ω (33 to 110) Ω (110 to 330) Ω 330 Ω to 1.1 kΩ (1.1 to 3.3) kΩ (3.3 to 11) kΩ (11 to 33) kΩ (33 to 110) kΩ (110 to 330) kΩ (0.33 to 1.1) MΩ (1.1 to 3.3) MΩ (3.3 to 11) MΩ (11 to 33) MΩ (33 to 110) MΩ (110 to 330) MΩ 330 MΩ to 1.1 GΩ	30 μΩ/Ω + 0.001 Ω 23 μΩ/Ω + 0.001 Ω 21 μΩ/Ω + 0.001 Ω 21 μΩ/Ω + 0.002 Ω 21 μΩ/Ω + 0.002 Ω 21 μΩ/Ω + 0.02 Ω 21 μΩ/Ω + 0.02 Ω 21 μΩ/Ω + 0.2 Ω 21 μΩ/Ω + 0.2 Ω 24 μΩ/Ω + 2 Ω 24 μΩ/Ω + 2 Ω 46 μΩ/Ω + 23 Ω 99 μΩ/Ω + 38 Ω 190 μΩ/Ω + 1.9 kΩ 380 μΩ/Ω + 2.3 kΩ 0.23 % of setting + 76 kΩ 1.1 % of reading + 380 kΩ	Fluke 5520A Multiproduct Calibrator
Resistance – Measure <sup>1</sup>	Up to 12 Ω (10 to 120) Ω 100 Ω to 1.2 kΩ (1 to 12) kΩ (10 to 120) kΩ 100 kΩ to 1.2 MΩ (1 to 12) MΩ (10 to 120) MΩ 100 MΩ to 1.2 GΩ	15 μΩ/Ω + 56 μΩ 12 μΩ/Ω + 0.5 mΩ 10 μΩ/Ω + 0.6 mΩ 10 μΩ/Ω + 5.6 mΩ 10 μΩ/Ω + 56 mΩ 15 μΩ/Ω + 2.2 Ω 50 μΩ/Ω + 120 Ω 500 μΩ/Ω + 1.2 kΩ 0.5 % of reading + 70 kΩ	Agilent 3458A Multimeter
Resistance – Generate <sup>1</sup> Fixed Points	100 V 100 kΩ (100 to 1 000) V 1 MΩ 10 MΩ 100 MΩ 1 GΩ 10 GΩ 100 GΩ	1 % of output 1 % of output 1 % of output 1 % of output 1 % of output 1.2 % of output 3.2 % of output	TMI RB Resistance Standard



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Capacitance – Generate <sup>1</sup>	(220 to 400) pF (0.4 to 3.299 9) nF (3.3 to 10.999 9) nF (11 to 32.999 9) nF (33 to 109.999) nF (110 to 329.999) nF (0.33 to 1.099 99) μF (1.1 to 3.299 99) μF (3.3 to 10.999 9) μF (11 to 32.999 9) μF (33 to 109.999) μF (110 to 329.999) μF (0.33 to 1.099 99) mF (1.1 to 3.299 99) mF (3.3 to 10.999 9) mF (11 to 32.999 9) mF (33 to 110) mF	0.38% of output + 7.6 pF 0.38 % of output + 0.01 nF 0.19 % of output + 0.01 nF 0.19 % of output + 0.08 nF 0.19 % of output + 0.08 nF 0.19 % of output + 0.23 nF 0.19 % of output + 0.76 nF 0.19 % of output + 2.3 nF 0.19 % of output + 7.6 nF 0.3 % of output + 23 nF 0.34 % of output + 76 nF 0.34 % of output + 228 nF 0.34 % of output + 0.76 μF 0.34 % of output + 2.3 μF 0.34 % of output + 7.6 μF 0.57 % of output + 23 μF 0.84 % of output + 76 μF	Fluke 5522A Multiproduct Calibrator
Capacitance – Measure <sup>1</sup>	100 Hz /120 Hz (16 to 400) pF (0.4 to 1) nF (1 to 100) nF (0.1 to 1) μF (1 to 100) μF (0.1 to 1) mF 1 000 Hz (16 to 400) pF (0.4 to 1) nF (1 to 100) nF (0.1 to 1) μF (1 to 100) μF (0.1 to 1) mF 10 kHz (16 to 400) pF (0.4 to 1) nF (1 to 100) nF (0.1 to 1) μF (1 to 100) μF (0.1 to 1) mF	2 % of reading + 0.3 pF 0.17 % of reading 0.13 % of reading 0.12 % of reading 0.18 % of reading 0.2 % of reading 0.43 % of reading + 0.3pF 0.1 % of reading 0.08 % of reading 0.07 % of reading 0.13 % of reading 0.45 % of reading 0.56 % of reading + 0.3pF 0.15 % of reading 0.12 % of reading 0.17 % of reading 0.69 % of reading 3.5 % of reading	Agilent 4263B LCR Meter



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Capacitance – Measure <sup>1</sup>	20 kHz (16 to 400) pF (0.4 to 1) nF (1 to 100) nF (0.1 to 1) μF (1 to 100) μF	2.1 % of reading + 0.3pF 0.62 % of reading 0.62 % of reading 0.43 % of reading 1.7 % of reading	Agilent 4263B LCR Meter
	100 kHz (16 to 400) pF (0.4 to 1) nF (1 to 100) nF (0.1 to 1) μF (1 to 10) μF	1.6 % of reading + 0.3 pF 1 % of reading 0.98 % of reading 1.4 % of reading 4.1 % of reading	
Inductance – Measure <sup>1</sup>	100 Hz /120 Hz (4 to 10) μH (10 to 40) μH (40 to 100) μH (0.1 to 0.4) mH (0.4 to 1) mH (1 to 10) mH (10 to 100) mH (0.1 to 1) H (1 to 10) H	5.2 % of reading + 0.03 μH 2.3 % of reading + 0.03 μH 0.57 % of reading + 0.03 μH 0.4 % of reading 0.32 % of reading 0.28 % of reading 0.23 % of reading 0.18 % of reading 0.22 % of reading	Agilent 4263B LCR Meter
	1 000 Hz (1 to 4) μH (4 to 10) μH (10 to 40) μH (40 to 100) μH (0.01to 0.4) mH (0.4 to 1) mH (1 to 10) mH (10 to 100) mH (0.1 to 1) H (1 to 10) H	1 % of reading + 0.03 μH 0.46 % of reading + 0.03 μH 0.74 % of reading + 0.03 μH 0.35 % of reading + 0.03 μH 0.3 % of reading 0.11 % of reading 0.12 % of reading 0.1 % of reading 0.1 % of reading 0.12 % of reading	



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Inductance – Measure <sup>1</sup>	10 kHz		Agilent 4263B LCR Meter
	(1 to 4) $\mu$ H	0.62 % of reading + 0.03 $\mu$ H	
	(4 to 10) $\mu$ H	0.4 % of reading + 0.03 $\mu$ H	
	(10 to 40) $\mu$ H	0.32 % of reading + 0.03 $\mu$ H	
	(40 to 100) $\mu$ H	0.18 % of reading + 0.03 $\mu$ H	
	(0.1 to 0.4) mH	0.2 % of reading	
	(0.4 to 1) mH	0.14 % of reading	
	(1 to 10) mH	0.17 % of reading	
	(10 to 100) mH	0.2 % of reading	
	(0.1 to 1) H	0.2 % of reading	
	(1 to 10) H	0.44 % of reading	
	20 kHz		
	(1 to 4) $\mu$ H	1.3 % of reading + 0.03 $\mu$ H	
	(4 to 10) $\mu$ H	0.68 % of reading + 0.03 $\mu$ H	
	(10 to 40) $\mu$ H	0.63 % of reading + 0.03 $\mu$ H	
	(40 to 100) $\mu$ H	0.53 % of reading + 0.03 $\mu$ H	
	(0.1 to 0.4) mH	0.49 % of reading	
	(0.4 to 1) mH	0.48 % of reading	
	(1 to 10) mH	0.65 % of reading	
	(10 to 100) mH	0.66 % of reading	
	(0.1 to 1) H	1.2 % of reading	
	(1 to 10) H	3.3 % of reading	
	100 kHz		
	(1 to 4) $\mu$ H	2.7 % of reading + 0.03 $\mu$ H	
	(4 to 10) $\mu$ H	0.86 % of reading + 0.03 $\mu$ H	
	(10 to 40) $\mu$ H	1.2 % of reading + 0.03 $\mu$ H	
	(40 to 100) $\mu$ H	0.68 % of reading + 0.03 $\mu$ H	
(0.1 to 0.4) mH	0.9 % of reading		
(0.4 to 1) mH	1.1 % of reading		
(1 to 10) mH	1.4 % of reading		
(10 to 100) mH	1.4 % of reading		
(0.1 to 1) H	6.3 % of reading		
Electrical Calibration of Thermocouple Indicating Devices <sup>1</sup>	Type B		Fluke 7526A Process Calibrator
	(600 to 800) $^{\circ}$ C	0.27 $^{\circ}$ C	
	(800 to 1 550) $^{\circ}$ C	0.22 $^{\circ}$ C	
	(1 550 $^{\circ}$ C to 1 820) $^{\circ}$ C	0.17 $^{\circ}$ C	
	Type C		
	(0 to 1000) $^{\circ}$ C	0.13 $^{\circ}$ C	
	(1 000 to 1 800) $^{\circ}$ C	0.18 $^{\circ}$ C	
(1 800 to 2 000) $^{\circ}$ C	0.2 $^{\circ}$ C		
(2 000 to 2 316) $^{\circ}$ C	0.27 $^{\circ}$ C		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Calibration of Thermocouple Indicating Devices <sup>1</sup>	Type E		Fluke 7526A Process Calibrator
	(-250 to -200) °C	0.19 °C	
	(-200 to -100) °C	0.1 °C	
	(-100 to 0) °C	0.07 °C	
	(0 to 600) °C	0.07 °C	
	(600 to 1 000) °C	0.08 °C	
	Type J		
	(-210 to -100) °C	0.11 °C	
	(-100 to 800) °C	0.07 °C	
	(800 to 1 200) °C	0.08 °C	
	Type K		
	(-250 to -200) °C	0.35 °C	
	(-200 to -100) °C	0.13 °C	
	(-100 to 800) °C	0.08 °C	
	(800 to 1 372) °C	0.1 °C	
	Type L		
	(-200 to -100) °C	0.08 °C	
	(-100 to 900) °C	0.07 °C	
	Type N		
	(-250 to -200) °C	0.56 °C	
	(-200 to -100) °C	0.18 °C	
	(-100 to 0) °C	0.1 °C	
	(0 to 100) °C	0.09 °C	
	(100 to 800) °C	0.08 °C	
	(800 to 1 300) °C	0.1 °C	
	Type R		
	(-50 to -25) °C	0.42 °C	
	(-25 to 0) °C	0.34 °C	
(0 to 100) °C	0.3 °C		
(100 to 400) °C	0.22 °C		
(400 to 600) °C	0.17 °C		
(600 to 1 000) °C	0.16 °C		
(1 000 to 1 600) °C	0.15 °C		
(1 600 to 1 767) °C	0.18 °C		
Type S			
(50 to -25) °C	0.39 °C		
(-25 to 0) °C	0.33 °C		
(0 to 100) °C	0.29 °C		
(100 to 400) °C	0.22 °C		
(400 to 600) °C	0.18 °C		
(600 to 1 600) °C	0.17 °C		
(1 600 to 1 767) °C	0.2 °C		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Calibration of Thermocouple Indicating Devices <sup>1</sup>	Type T (-250 to -200) °C (-200 to -100) °C (-100 to 0) °C (0 to 400) °C	0.26 °C 0.13 °C 0.09 °C 0.07 °C	Fluke 7526A Process Calibrator
	Type U (-200 to 0) °C (0 to 600) °C	0.13 °C 0.08 °C	
Electrical Calibration of RTD Indicating Devices <sup>1</sup>	Pt 385, 100 Ω (-200 to 800) °C	0.05 °C	Fluke 7526A Process Calibrator
	Pt 3926, 100 Ω (-200 to 630) °C	0.05 °C	
	Pt 3916, 100 Ω (-200 to 630) °C	0.05 °C	
	Pt 385, 200 Ω (-200 to 400) °C (400 to 630) °C	0.4 °C 0.5 °C	
	Pt 385, 500 Ω (-200 to 630) °C	0.17 °C	
	Pt 385, 1000 Ω (-200 to 630) °C	0.09 °C	
Oscilloscopes Calibration <sup>1</sup> – Generate Voltage DC - 50Ω	(1 to 24.999) mV	0.19 % of Output + 31 μV	Fluke 5820A Oscilloscope Calibrator w/ GHz Option
	(25 to 109.99) mV	0.19 % of Output + 36 μV	
	(110mV to 2.1999) V	0.19 % of Output + 87 μV	
	(2.2 to 6.6) V	0.19 % of Output + 0.6 mV	
	(1 to 24.999) mV	0.019 % of reading + 20 μV	
	(25 to 109.99) mV	0.019 % of reading + 25 μV	
	(110mV to 2.1999) V	0.019 % of reading + 76 μV	
	(2.2 to 10.999) V	0.019 % of reading + 0.6 mV	
	(11 to 130) V	0.019 % of reading + 6 mV	
	(1 to 24.999) mVpp	0.19 % of Output + 31 μV	
	(25 to 109.99) mVpp	0.19 % of Output + 36 μV	
	(110mV to 2.1999) Vpp	0.19 % of Output + 87 μV	
(2.2 to 6.6) Vpp	0.19 % of Output + 0.6 mV		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Oscilloscopes Calibration – Generate <sup>1</sup> Square Wave 10 Hz to 1 kHz – 1 MΩ  Square Wave 1 to 10 kHz – 1 MΩ	(1 to 24.999) mV (25 to 109.99) mV (110mV to 2.1999) V (2.2 to 10.999) V (11 to 130) V  (1 to 24.999) mV (25 to 109.99) mV (110mV to 2.1999) V (2.2 to 10.999) V (11 to 130) V	0.038 % of reading + 4 μV 0.038 % of reading + 9 μV 0.038 % of reading + 60 μV 0.038 % of reading + 0.6 mV 0.038 % of reading + 6 mV  0.19 % of reading + 31 μV 0.19 % of reading + 36 μV 0.19 % of reading + 87 μV 0.19 % of reading + 0.6 mV 0.19 % of reading + 6 mV	Fluke 5820A Oscilloscope Calibrator w/ GHz Option
Oscilloscopes Calibration – Generate <sup>1</sup> Leveled Sine Flatness 50 kHz to 10 MHz Reference	3 dB Bandwidth (5 to 50) mVpp 50 kHz to 100 MHz (100 to 300) MHz (300 to 500) MHz (500 to 600) MHz (600 to 1 600) MHz (1 600 to 2 100) MHz 50 mV to 3.5 Vpp 50 kHz to 100 MHz (100 to 300) MHz (300 to 500) MHz (500 to 600) MHz (600 to 1 600) MHz (1 600 to 2 100) MHz (3.5 to 5) Vpp 50 kHz to 100 MHz (100 to 300) MHz (300 to 500) MHz (500 to 600) MHz	0.34 dB 0.36 dB 0.42 dB 0.46 dB 0.5 dB 0.56 dB  0.24 dB 0.24 dB 0.32 dB 0.34 dB 0.4 dB 0.44 dB  0.24 dB 0.24 dB 0.32 dB 0.34 dB	Fluke 5820A Oscilloscope Calibrator w/ GHz Option
Oscilloscopes Calibration – Generate <sup>1</sup> Leveled Sine Flatness	3 dB Bandwidth 50 mV to 3.5Vpp (2 100 to 4 000) MHz (4 000 to 8 000) MHz (8 000 to 18 000) MHz	0.25 dB 0.35 dB 0.46 dB	EPM Power Meter w/ E Series Power Sensors
Oscilloscopes Calibration – Generate <sup>1</sup> Time Marker	500 ps to 20 ms 50 ms to 5 s	0.25 μs/s 1.9 μs/s + 3.8 μHz	Fluke 5820A Oscilloscope Calibrator w/ GHz Option

**Electrical – DC/Low Frequency**

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Oscilloscopes Calibration – Measure <sup>1</sup> Input Impedance Resistance Leakage	(40 to 60) Ω 500 kΩ to 1.5MΩ (0 to 5.99) V	0.08 % of reading 0.08 % of reading 0.038 % of reading + 0.8 mV	Fluke 5820A Oscilloscope Calibrator w/ GHz Option
Low Frequency Power – Generate <sup>1</sup> (45 to 65) Hz 1 PF  DC	Up to 20 kW	0.25 % of reading  0.21 % of reading	Fluke 5522A Multiproduct Calibrator
Power Meters <sup>1</sup>	3 μW to 100 mW	0.25 % of reading	HP 11683A Range Calibrator

**Electrical – RF/Microwave**

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Flatness – Measure <sup>1</sup>	9 kHz to 2 000 MHz (20 to -10) dBm (-10 to -30) dBm (-30 to -40) dBm (-40 to -42) dBm (2 000 to 14 000) MHz (20 to -10) dBm (-10 to -30) dBm (-30 to -40) dBm (-40 to -42) dBm (14 000 to 18 000) MHz (20 to -10) dBm (-10 to -30) dBm (-30 to -40) dBm (-40 to -42) dBm	0.1 dB 0.1 dB 0.11 dB 0.12 dB 0.1 dB 0.09 dB 0.1 dB 0.11 dB 0.11 dB 0.12 dB 0.12 dB 0.12 dB 0.13 dB	Agilent EPM Series Power Meter w/E9304A H18 Power Sensor



ANSI National Accreditation Board

Electrical – RF/Microwave

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Attenuation – Measure <sup>1</sup>	(30 to 3 050) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.12 dB	
	(90 to 100) dB	0.13 dB	
	(100 to 110) dB	0.13 dB	
	(3 050 to 6 600) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.12 dB	
	(90 to 100) dB	0.13 dB	
	(100 to 110) dB	0.19 dB	
	(6 600 to 13 200) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
(50 to 60) dB	0.076 dB		
(60 to 70) dB	0.081 dB		
(70 to 80) dB	0.12 dB		
(80 to 90) dB	0.12 dB		
(90 to 100) dB	0.13 dB		
(100 to 110) dB	0.25 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Attenuation – Measure <sup>1</sup>	(13 200 to 19 200) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.12 dB	
	(90 to 100) dB	0.13 dB	
	(100 to 110) dB	0.31 dB	
	(19 200 to 26 500) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.14 dB	
	(90 to 100) dB	0.36 dB	
	(100 to 110) dB	0.82 dB	
	(26 500 to 31 150) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
(50 to 60) dB	0.076 dB		
(60 to 70) dB	0.081 dB		
(70 to 80) dB	0.12 dB		
(80 to 90) dB	0.13 dB		
(90 to 100) dB	0.33 dB		
(100 to 110) dB	0.77 dB		

**Electrical – RF/Microwave**

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Attenuation – Measure <sup>1</sup>	(31 150 to 41 000) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.14 dB	
	(80 to 90) dB	0.36 dB	
	(41 000 to 45 000) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.11 dB	
	(70 to 80) dB	0.24 dB	
	(80 to 90) dB	0.6 dB	
	(45 000 to 50 000) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.11 dB	
(60 to 70) dB	0.29 dB		
(70 to 80) dB	0.7 dB		
(80 to 90) dB	1.4 dB		
RF Power – Measure <sup>1</sup>	9 kHz to 14 000 MHz		Agilent EPM Series Power Meter w/E9304A H18 Power Sensor
	(20 to 0) dB	0.13 dB	
	(0 to -40) dB	0.15 dB	
	(-40 to -50) dB	0.35 dB	
	(-50 to -55) dB	0.93 dB	
	(14 000 to 18 000) MHz		
	(20 to 0) dB	0.12 dB	
	(0 to -40) dB	0.16 dB	
	(-40 to -50) dB	0.35 dB	
	(-50 to -55) dB	0.93 dB	



ANSI National Accreditation Board

Electrical – RF/Microwave

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Measure <sup>1</sup>	(100 to 2 000) MHz (20 to 10) dB (10 to 0) dB (0 to -10) dB (-10 to -20) dB (-20 to -25) dB	0.14 dB 0.06 dB 0.07 dB 0.08 dB 0.15 dB	Agilent EPM Series Power Meter 8487A Power Sensor
RF Power – Measure <sup>1</sup>	(2 000 to 12 400) MHz (20 to 10) dB (10 to 0) dB (0 to -10) dB (-10 to -20) dB (-20 to -25) dB (12 400 to 18 000) MHz (20 to 10) dB (10 to 0) dB (0 to -10) dB (-10 to -20) dB (-20 to -25) dB (18 000 to 26 500) MHz (20 to 10) dB (10 to 0) dB (0 to -10) dB (-10 to -20) dB (-20 to -25) dB (26 500 to 40 000) MHz (20 to 10) dB (10 to 0) dB (0 to -10) dB (-10 to -20) dB (-20 to -25) dB (40 000 to 50 000) MHz (20 to 10) dB (10 to 0) dB (0 to -10) dB (-10 to -20) dB (-20 to -25) dB	0.15 dB 0.07 dB 0.07 dB 0.08 dB 0.15 dB 0.15 dB 0.08 dB 0.08 dB 0.09 dB 0.15 dB 0.16 dB 0.1 dB 0.1 dB 0.11 dB 0.16 dB 0.17 dB 0.12 dB 0.12 dB 0.13 dB 0.18 dB 0.22 dB 0.19 dB 0.19 dB 0.19 dB 0.23 dB	Agilent EPM Series Power Meter 8487A Power Sensor

**Electrical – RF/Microwave**

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Measure <sup>1</sup>	(30 to 2 000) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor
	(30 to 20) dB	0.36 dB	
	(20 to 0) dB	0.2 dB	
	(0 to -58) dB	0.22 dB	
	(-58 to -78) dB	0.23 dB	
	(-78 to -110) dB	0.25 dB	
	(-110 to -115) dB	0.27 dB	
	(-115 to -120) dB	0.33 dB	
	(-120 to -125) dB	0.46 dB	
	(2 000 to 3 050) MHz		
	(30 to 20) dB	0.37 dB	
	(20 to 0) dB	0.21 dB	
	(0 to -58) dB	0.23 dB	
	(-58 to -78) dB	0.24 dB	
	(-78 to -110) dB	0.26 dB	
	(-110 to -115) dB	0.28 dB	
	(-115 to -120) dB	0.34 dB	
	(-120 to -125) dB	0.7 dB	
	(3 050 to 6 600) MHz		
	(30 to 20) dB	0.37 dB	
	(20 to 0) dB	0.21 dB	
	(0 to -58) dB	0.23 dB	
	(-58 to -78) dB	0.24 dB	
	(-78 to -110) dB	0.29 dB	
	(-110 to -115) dB	0.38 dB	
	(-115 to -120) dB	0.53 dB	
	(6 600 to 13 200) MHz		
	(30 to 20) dB	0.37 dB	
	(20 to 0) dB	0.21 dB	
	(0 to -58) dB	0.23 dB	
(-58 to -78) dB	0.24 dB		
(-78 to -110) dB	0.34 dB		
(-110 to -115) dB	0.46 dB		
(-115 to -120) dB	0.65 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Measure <sup>1</sup>	(13 200 to 18 000) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor
	(30 to 20) dB	0.37 dB	
	(20 to 0) dB	0.21 dB	
	(0 to -58) dB	0.23 dB	
	(-58 to -78) dB	0.24 dB	
	(-78 to -90) dB	0.26 dB	
	(-90 to -95) dB	0.26 dB	
	(-95 to -100) dB	0.26 dB	
	(-100 to -105) dB	0.29 dB	
	(-105 to -110) dB	0.38 dB	
	(-110 to -115) dB	0.53 dB	
	(-115 to -120) dB	0.75 dB	
	(18 000 to 19 200) MHz		
	(30 to 20) dB	0.39 dB	
	(20 to 0) dB	0.25 dB	
	(0 to -58) dB	0.24 dB	
	(-58 to -78) dB	0.25 dB	
	(-78 to -90) dB	0.27 dB	
	(-90 to -95) dB	0.27 dB	
	(-95 to -100) dB	0.27 dB	
	(-100 to -105) dB	0.3 dB	
	(-105 to -110) dB	0.38 dB	
	(-110 to -115) dB	0.53 dB	
	(-115 to -120) dB	0.75 dB	
	(19 200 to 26 500) MHz		
	(30 to 20) dB	0.39 dB	
	(20 to 0) dB	0.25 dB	
	(0 to -58) dB	0.24 dB	
	(-58 to -78) dB	0.25 dB	
	(-78 to -90) dB	0.28 dB	
(-90 to -95) dB	0.33 dB		
(-95 to -100) dB	0.43 dB		
(-100 to -105) dB	0.61 dB		
(-105 to -110) dB	0.85 dB		
(-110 to -115) dB	1.2 dB		
(-115 to -120) dB	1.5 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Measure <sup>1</sup>	(26 500 to 31 150) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor
	(30 to 20) dB	0.42 dB	
	(20 to 0) dB	0.3 dB	
	(0 to -58) dB	0.34 dB	
	(-58 to -78) dB	0.34 dB	
	(-78 to -90) dB	0.36 dB	
	(-90 to -95) dB	0.39 dB	
	(-95 to -100) dB	0.46 dB	
	(-100 to -105) dB	0.61 dB	
	(-105 to -110) dB	0.82 dB	
	(-110 to -115) dB	1.1 dB	
	(31 150 to 41 000) MHz		
	(30 to 20) dB	0.42 dB	
	(20 to 0) dB	0.3 dB	
	(0 to -58) dB	0.34 dB	
	(-58 to -78) dB	0.35 dB	
	(-78 to -90) dB	0.48 dB	
	(-90 to -95) dB	0.64 dB	
	(-95 to -100) dB	0.87 dB	
	(-100 to -105) dB	1.2 dB	
	(-105 to -110) dB	1.5 dB	
	(41 000 to 45 000) MHz		
	(30 to 20) dB	0.42 dB	
	(20 to 0) dB	0.3 dB	
(0 to -58) dB	0.34 dB		
(-58 to -78) dB	0.38 dB		
(-78 to -90) dB	0.68 dB		
(-90 to -95) dB	0.93 dB		
(-95 to -100) dB	1.2 dB		
(-100 to -105) dB	1.6 dB		
(45 000 to 50 000) MHz			
(30 to 20) dB	0.42 dB		
(20 to 0) dB	0.3 dB		
(0 to -58) dB	0.34 dB		
(-58 to -78) dB	0.67 dB		
(-78 to -90) dB	1.4 dB		
RF Power – Measure <sup>1</sup> 50 MHz	1 mW	0.003 2 mW	HP 432A Power Meter & 8478B Power Sensor

**Electrical – RF/Microwave**

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Generate <sup>1</sup>	(30 to 2 000) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor, E8257D Signal Generator
	(20 to 0) dB	0.29 dB	
	(0 to -58) dB	0.3 dB	
	(-58 to -78) dB	0.31 dB	
	(-78 to -110) dB	0.32 dB	
	(2 000 to 3 050) MHz		
	(20 to 0) dB	0.34 dB	
	(0 to -58) dB	0.35 dB	
	(-58 to -78) dB	0.36 dB	
	(-78 to -110) dB	0.39 dB	
	(3 050 to 6 600) MHz		
	(20 to 0) dB	0.34 dB	
	(0 to -58) dB	0.35 dB	
	(-58 to -78) dB	0.36 dB	
	(-78 to -110) dB	0.39 dB	
	(6 600 to 13 200) MHz		
	(20 to 0) dB	0.34 dB	
	(0 to -58) dB	0.35 dB	
	(-58 to -78) dB	0.36 dB	
	(-78 to -110) dB	0.42 dB	
	(13 200 to 18 000) MHz		
	(15 to 0) dB	0.34 dB	
	(0 to -58) dB	0.35 dB	
	(-58 to -78) dB	0.36 dB	
	(-78 to -110) dB	0.46 dB	
	(18 000 to 19 200) MHz		
	(15 to 0) dB	0.41 dB	
	(0 to -58) dB	0.41 dB	
(-58 to -78) dB	0.42 dB		
(-78 to -110) dB	0.5 dB		
(19 200 to 26 500) MHz			
(15 to 0) dB	0.41 dB		
(0 to -58) dB	0.41 dB		
(-58 to -78) dB	0.42 dB		
(-78 to -110) dB	0.9 dB		
(26 500 to 31 150) MHz			
(15 to 0) dB	0.62 dB		
(0 to -58) dB	0.63 dB		
(-58 to -78) dB	0.64 dB		
(-78 to -110) dB	0.96 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Generate <sup>1</sup>	(31 150 to 41 000) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor, E8257D Signal Generator
	(10 to 0) dB	0.82 dB	
	(0 to -58) dB	0.83 dB	
	(-58 to -78) dB	0.84 dB	
	(-78 to -100) dB	1.1 dB	
	(41 000 to 45 000) MHz		
	(10 to 0) dB	0.82 dB	
	(0 to -58) dB	0.83 dB	
	(-58 to -78) dB	0.85 dB	
	(-78 to -100) dB	1.4 dB	
	(45 000 to 50 000) MHz		
	(0 to 10) dB	0.82 dB	
(-58 to 0) dB	0.83 dB		
(-78 to -58) dB	1 dB		
(-90 to -78) dB	1.5 dB		
Amplitude Modulation – Measure <sup>1</sup>	100 kHz to 10 MHz		Agilent N5531S Measuring Receiver
	Rate 50 Hz to 10 kHz		
	(5 to 99) % Depth	0.75 % of reading + 0.3 digits	
	10 MHz to 3 GHz		
	Rate 50 Hz to 100 kHz		
	(5 to 20) % Depth	2.5 % of reading + 0.4 digits	
	(20 to 99) % Depth	1.5 % of reading + 0.4 digits	
	(3 to 26.5) GHz		
	Rate 50 Hz to 100 kHz		
	(5 to 20) % Depth	4.5 % of reading + 0.4 digits	
	(20 to 99) % Depth	1.5 % of reading + 0.4 digits	
	(26.5 to 31.15) GHz		
Rate 50 Hz to 100 kHz			
(5 to 20) % Depth	6.8 % of reading + 0.4 digits		
(20 to 99) % Depth	1.9 % of reading + 0.4 digits		
(31.15 to 50) GHz			
Rate 50 Hz to 100 kHz			
(5 to 20) % Depth	26 % of reading + 0.4 digits		
(20 to 99) % Depth	6 % of reading + 0.4 digits		



ANSI National Accreditation Board

Electrical – RF/Microwave

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency Modulation – Measure <sup>1</sup> $\beta$ = deviation / rate	250 kHz to 10 MHz Rates 20 Hz to 10 kHz Peak Dev 200 to 40 kHz 10 MHz to 6.6 GHz Rates 50 Hz to 200 kHz Peak Dev 250 to 400 kHz (6.6 to 13.2) GHz Rates 50 Hz to 200 kHz Peak Dev 250 to 400 kHz (13.2 to 31.15) GHz Rates 50 Hz to 200 kHz Peak Dev 250 to 400 kHz (31.15 to 50.0) GHz Rates 50 Hz to 200 kHz Peak Dev 250 to 400 kHz	$\beta > 0.2 - 1.5$ % of reading + 2 Hz $\beta > 1.2 - 1$ % of reading + 2 Hz $\beta > 0.20 - 1.5$ % of reading + 2 Hz $\beta > 0.45 - 1$ % of reading + 2 Hz $\beta > 0.2 - 2.5$ % of reading + 4 Hz $\beta > 8.0 - 1$ % of reading + 4 Hz $\beta > 0.2 - 3.8$ % of reading + 9 Hz $\beta > 16 - 1$ % of reading + 9 Hz $\beta > 0.2 - 8.5$ % of reading + 17 Hz $\beta > 32 - 1$ % of reading + 17 Hz	Agilent N5531S Measuring Receiver
Phase Modulation <sup>1</sup> – Measure	100 kHz to 6.6 GHz Deviations > 0.3 rad Deviations > 0.7 rad (6.6 to 13.2) GHz Deviations > 0.6 rad Deviations > 2.0 rad (13.2 to 26.5) GHz Deviations: > 1.2 rad Deviations > 4.0 rad (26.5 to 31.15) GHz Deviations: > 1.3 rad Deviations > 4.0 rad (31.15 to 50) GHz Deviations: > 2.4 rad Deviations > 8.0 rad	3 % of reading + 0.002 rad 1 % of reading + 0.002 rad 3 % of reading + 0.005 rad 1 % of reading + 0.005 rad 3 % of reading + 0.009 rad 1 % of reading + 0.009 rad 3 % of reading + 0.009 rad 1 % of reading + 0.009 rad 3 % of reading + 0.018 rad 1 % of reading + 0.018 rad	Agilent N5531S Measuring Receiver
AM Distortion Measure <sup>1</sup> Rate 20 Hz to 1 kHz	(0.1 to 10) MHz AM Depth > 1 % (0 to -20) dB (-20 to -30) dB AM Depth > 3 % (0 to -20) dB (-20 to -30) dB (-30 to -40) dB	1.2 dB 2.2 dB 1 dB 1.3 dB 2.4 dB	Agilent N5531S Measuring Receiver



ANSI National Accreditation Board

Electrical – RF/Microwave

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AM Distortion Measure <sup>1</sup> Rate 20 Hz to 1 kHz	10 MHz to 26.5 GHz AM Depth > 1 % (0 to -20) dB	1.3 dB	Agilent N5531S Measuring Receiver
	(-20 to -30) dB	2.5 dB	
	AM Depth > 3 % (0 to -20) dB	1.1 dB	
	(-20 to -30) dB	1.4 dB	
	(-30 to -40) dB	3 dB	
	(26.5 MHz to 50.0 GHz) AM Depth > 3 % (0 to -20) dB	1.8 dB	
FM Distortion Measure <sup>1</sup> Rate 20 Hz to 1 kHz	(1 to 6 600) MHz Dev 500 Hz to 2 kHz (0 to -20) dB	0.26 dB	Agilent N5531S Measuring Receiver
	(-20 to -30) dB	0.79 dB	
	(-30 to -40) dB	2.3 dB	
	Dev > 2 kHz (0 to -20) dB	0.09 dB	
	(-20 to -30) dB	0.27 dB	
	(-30 to -40) dB	0.83 dB	
FM Distortion Measure <sup>1</sup> Rate 20 Hz to 1 kHz	(6.6 to 13.2) GHz Dev > 2.3 kHz (0 to -20) dB	0.26 dB	Agilent N5531S Measuring Receiver
	(-20 to -30) dB	0.79 dB	
	(-30 to -40) dB	2.3 dB	
	Dev > 4.5 kHz (0 to -20) dB	0.09 dB	
	(-20 to -30) dB	0.27 dB	
	(-30 to -40) dB	0.83 dB	
	(-40 to -50) dB	2.4 dB	



ANSI National Accreditation Board

Electrical – RF/Microwave

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
FM Distortion Measure <sup>1</sup> Rate 20 Hz to 1 kHz	(13.2 to 31.15) GHz		Agilent N5531S Measuring Receiver
	Dev > 2.7 kHz		
	(0 to -20) dB	0.26 dB	
	(-20 to -30) dB	0.79 dB	
	(-30 to -40) dB	2.3 dB	
	Dev > 6.0 kHz		
	(0 to -20) dB	0.09 dB	
	(-20 to -30) dB	0.27 dB	
	(-30 to -40) dB	0.83 dB	
	(-40 to -50) dB	2.4 dB	
	(31.15 to 50.0) GHz		
	Dev > 4 kHz		
	(0 to -20) dB	0.26 dB	
	(-20 to -30) dB	0.79 dB	
	(-30 to -40) dB	2.3 dB	
	Dev > 12.0 kHz		
	(0 to -20) dB	0.09 dB	
	(-20 to -30) dB	0.27 dB	
	(-30 to -40) dB	0.83 dB	
	(-40 to -50) dB	0.83 dB	
	(1 to 6 600) MHz	2.3 dB	
	Rate (20 to 500) Hz		
	Dev > 0.8 rad		
	(0 to -20) dB	0.26 dB	
(-20 to -30) dB	0.79 dB		
(-30 to -40) dB	2.3 dB		
Dev > 2.5 rad			
(0 to -20) dB	0.09 dB		
(-20 to -30) dB	0.27 dB		
(-30 to -40) dB	0.83 dB		
(-40 to -50) dB	2.3 dB		
Rate (500 to 1 000) Hz			
Dev > 0.4 rad			
(0 to -20) dB	0.26 dB		
(-20 to -30) dB	0.79 dB		
(-30 to -40) dB	2.3 dB		
Dev > 1.0 rad			
(0 to -20) dB	0.09 dB		
(-20 to -30) dB	0.27 dB		
(-30 to -40) dB	0.83 dB		
(-40 to -50) dB	2.3 dB		

**Electrical – RF/Microwave**

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
PM Distortion Measure <sup>1</sup>	(6.6 to 13.2) GHz		Agilent N5531S Measuring Receiver
	Rate (20 to 500) Hz		
	Dev > 1.8 rad		
	(0 to -20) dB	0.26 dB	
	(-20 to -30) dB	0.79 dB	
	(-30 to -40) dB	2.3 dB	
	Dev > 5.5 rad		
	(0 to -20) dB	0.09 dB	
	(-20 to -30) dB	0.27 dB	
	(-30 to -40) dB	0.83 dB	
	(-40 to -50) dB	2.3 dB	
	(6.6 to 13.2) GHz		
	Rate (500 to 1 000) Hz		
	Dev > 0.8 rad		
	(0 to -20) dB	0.26 dB	
	(-20 to -30) dB	0.79 dB	
	(-30 to -40) dB	2.3 dB	
	Dev > 2.5 rad		
	(0 to -20) dB	0.09 dB	
	(-20 to -30) dB	0.27 dB	
	(-30 to -40) dB	0.83 dB	
(-40 to -50) dB	2.3 dB		
(13.2 to 31.15) GHz			
Rate (20 to 500) Hz			
Dev > 3.5 rad			
(0 to -20) dB	0.26 dB		
(-20 to -30) dB	0.79 dB		
(-30 to -40) dB	2.3 dB		
Dev > 10.0 rad			
(0 to -20) dB	0.09 dB		
(-20 to -30) dB	0.27 dB		
(-30 to -40) dB	0.83 dB		
(-40 to -50) dB	2.3 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
PM Distortion Measure <sup>1</sup>	(13.2 to 31.15) GHz		
	Rate (500 to 1 000) Hz		
	Dev > 1.2 rad		
	(0 to -20) dB	0.26 dB	
	(-20 to -30) dB	0.79 dB	
	(-30 to -40) dB	2.3 dB	
	Dev > 4.0 rad		
	(0 to -20) dB	0.09 dB	
	(-20 to -30) dB	0.27 dB	
	(-30 to -40) dB	0.83 dB	
	(-40 to -50) dB	2.3 dB	
	(31.15 to 50.0) GHz		
	Rate 20 to 500 Hz		
	Dev > 7.5 rad		
	(0 to -20) dB	0.26 dB	
	(-20 to -30) dB	0.79 dB	
	(-30 to -40) dB	2.3 dB	
	Dev > 19.0 rad		
	(0 to -20) dB	0.09 dB	
	(-20 to -30) dB	0.27 dB	
(-30 to -40) dB	0.83 dB		
(-40 to -50) dB	2.3 dB		
Rate (500 to 1 000) Hz			
Dev > 3.0 rad			
(0 to -20) dB	0.26 dB		
(-20 to -30) dB	0.79 dB		
(-30 to -40) dB	2.3 dB		
Dev > 8.0 rad			
(0 to -20) dB	0.09 dB		
(-20 to -30) dB	0.27 dB		
(-30 to -40) dB	0.83 dB		
(-40 to -50) dB	2.3 dB		

Agilent N5531S  
Measuring Receiver

**Electrical – RF/Microwave**

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Total Harmonic Distortion (THD)	(0 to 60) dB		HP 8903B Audio Analyzer
	20 Hz to 20 kHz	1 dB	
	(-40 to 0) dB	1 dB	
	(-50 to -40) dB	1.3 dB	
	(-60 to -50) dB	1.7 dB	
	(-65 to -60) dB		
	(20 to 50) kHz	2 dB	
	(-40 to 0) dB	2.1 dB	
	(-50 to -40) dB	3 dB	
	(-60 to -50) dB		
Harmonics Measure <sup>1</sup>	(50 to 100) kHz	2 dB	Agilent E4448A Measuring Receiver
	(-40 to 0) dB	2.4 dB	
	(-50 to -40) dB		
	(-80 to -10) dB		
	2 <sup>nd</sup> through 5 <sup>th</sup> Harmonic		
	1 kHz to 600 MHz	0.37 dB	
	(600 to 1 320) MHz	1.1 dB	
	(1 320 to 2 200) MHz	1.4 dB	
	(2 200 to 3 000) MHz	1.4 dB	
	(3 000 to 4 400) MHz	1.7 dB	
	(4 400 to 5 300) MHz	1.9 dB	
	(5 300 to 10 000) MHz	2.1 dB	
	2 <sup>nd</sup> through 4 <sup>th</sup> Harmonic		
(10 000 to 12 500) MHz	2.1 dB		
2 <sup>nd</sup> through 3 <sup>rd</sup> Harmonic			
(12 500 to 16 667) MHz	2.1 dB		
2 <sup>nd</sup> Harmonic			
(16 667 to 25 000) MHz	2.3 dB		

**Length – Dimensional Metrology**

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Calipers <sup>1,2</sup>	Up to 80 in	(280 + 3.4L) μin	Gage blocks (Grade 2)
Micrometers <sup>1,2</sup>	Up to 46 in	(9.3 + 7L) μin	
Height Gages <sup>1,2</sup>	Up to 46 in	(78 + 5.3L) μin	

**Length – Dimensional Metrology**

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Dial Indicators <sup>1,2</sup> Resolution $\geq 50\mu\text{in}$ Resolution $< 50\mu\text{in}$	Up to 10 in Up to 0.1 in	$(26 + 3.7L) \mu\text{in}$ 9.6 $\mu\text{in}$	Gage blocks (Grade 2)
Rulers <sup>1</sup>	Up to 46 in	0.009 1 in	
Bore Micrometers <sup>2</sup> 2 point 3 point	Up to 12 in Up to 12 in	$(56 + 3.2L) \mu\text{in}$ $(64 + 3.1L) \mu\text{in}$	Master gage blocks, P&W universal measuring machine, Master Ring
Spheres	Up to 2 in	30 $\mu\text{in}$	P&W Model C
Metal Tapes and Rules <sup>1,2</sup>	Up to 100 ft	$(0.000 27F + 0.024) \text{in}$	Standard rule
Feeler Gages <sup>1</sup>	Up to 1 in	36 $\mu\text{in}$	Mitutoyo 293-369 Micrometer
Cylindrical gages <sup>1,2</sup> – Plain Pins, Plugs Rings	Up to 10 in Up to 14 in	$(10 + 2.9D) \mu\text{in}$ $(9.8 + 3D) \mu\text{in}$	P&W LabMaster gage blocks (grade 1)
Surface Plates <sup>1</sup> – Overall Flatness Local Area Flatness	$(18 \times 18) \text{in}$ to $(6 \times 6) \text{ft}$ $(-0.001 \text{ to } 0.001) \text{in}$	77 $\mu\text{in}$ 68 $\mu\text{in}$	Rahn Planekator Repeat-o-meter
Gage Blocks <sup>2</sup>	Up to 12 in	$(4.1 + 2.3L) \mu\text{in}$	Universal measuring machine, master gage block set
Micrometer Standards Length Rods	Up to 46 in	$(1 + 7.8L) \mu\text{in}$	Gage blocks (grade 2), P&W universal measuring Machine, MAHR Supramess Dial Comparator
Thread Plugs – Major Diameter Pitch Diameter	Up to 12 in Up to 12 in	40 $\mu\text{in}$ 92 $\mu\text{in}$	Gage blocks, P&W universal measuring machine, Van Keuren thread wire set
Thread Wires	Up to 0.5 in	11 $\mu\text{in}$	Master gage blocks, P&W universal measuring machine

**Length – Dimensional Metrology**

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Optical Comparators <sup>1</sup> – Angle Linearity	Up to 360 ° Up to 20 in (20 to 40) in	0.008 7 ° 320 μin 630 μin	Precision balls, Angle Blocks, Gage Blocks, SI Industries glass scales
Magnification	10x to 100x	420 μin	
Protractors <sup>1</sup>	(0 to 360)°	0.008°	Angle blocks
Coating Thickness Gages <sup>1</sup> Eddy Current & Magnetic Induction, Fixed Point	(0.48 to 38.9) mils	69 μin	Calibration foils, P&W Supermicrometer
Coating Thickness Shims <sup>1</sup>	(0 to 243) mils	57 μin	P&W Supermicrometer

**Mass and Mass Related**

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Scales & Balances <sup>1,2,4</sup>	1 mg to 600 g	(0.014 + 0.001 5X) mg	Class 0 weights
	10 mg to 40 kg (0.022 to 2 000) lb	(7.7 + 0.12X) mg (0.000 08 + 0.000 12W) lb	Class F weights
Pressure <sup>1</sup>	(0 to 30) psia (0 to 61) inHg	0.018 psi 0.037 inHg	Fluke 700GA5 Pressure Gage
	(0 to 30) psi	0.009 psi	Additel 681-GP30 Pressure Gage
	(-15 to 100) psi	0.027 psi	Additel 681- CP100 Pressure Gage
	(0 to 300) psi	0.08 psi	Additel 681- GP300 Pressure Gage
	(0 to 1 000) psi	0.24 psi	Additel 681- GPIK Pressure Gage
	(0 to 3 000) psi	0.7 psi	Additel 681- GP3K Pressure Gage
	(0 to 10 000) psi	2.5 psi	Additel 681- GP10K Pressure Gage
	(-150 to 150) inH <sub>2</sub> O	0.094 inH <sub>2</sub> O	Additel 681-DP150 Pressure Gage

**Mass and Mass Related**

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Pressure <sup>1</sup>	(-20 to 20) inH <sub>2</sub> O	0.001 3 inH <sub>2</sub> O	Additel 681-DP20 Pressure Gage
Torque Tools <sup>1</sup>	(10 to 100) ozf·in (5 to 1 000) lbf·in (25 to 250) lbf·ft (250 to 2 000) lbf·ft	0.59 % of reading 0.35 % of reading 0.35 % of reading 0.6 % of reading	CDI torque system
Torque Analyzers	(5 to 80) ozf·in (5 to 600) lbf·in (50 to 2 000) lbf·ft	0.16 % of reading 0.15 % of reading 0.14 % of reading	Weights and Wheel
Mass - Class F	(0.5 to 10) lb (10 to 50) lb	0.03 g, (0.000 066) lb 0.27 g, (0.000 6) lb	Master balance
Force <sup>1</sup> Tension	(10 to 200) mgf (0.2 to 1) gf (1 to 10) gf (10 to 500) gf (1 to 540) lbf	0.45 mgf 1.2 mgf 0.037 % of reading 0.024 % of reading 0.017 % of reading	Class F weights
Force <sup>1</sup> Tension	(1 000 to 50 000) lbf	0.021 % of reading	Morehouse Load Cells
Force <sup>1</sup> Compression	(200 to 10 000) lbf (500 to 25 000) lbf (1 000 to 50 000) lbf	0.014 % of reading 0.018 % of reading 0.034 % of reading	Morehouse Load Cells
Rockwell Hardness Testers <sup>1</sup>	(< 60) HRBW (≥ 60 to < 80) HRBW  (≥ 35 to < 60) HRC (≥ 60) HRC  < 70 HA > 80 HA	3.1 HRBW 3 HRBW  1.2 HRC 0.7 HRC  1.4 HRA 0.7 HRA	Indirect verification per ASTM E18
Roughness Testers <sup>1</sup>	Ra 117 μin Rmax 370 μin	1.3 μin	Roughness Specimen
Durometers Scale (Force) Accuracy Types A, B, C, D, DO, E, O	(0 to 100) duros	0.06 duros	Direct Verification  Master balance
Indenter Geometry Length Diameter Angle	0.1 in 0.05 in (30 to 35)°	140 μin 140 μin 0.13°	Optical comparator

**Thermodynamic**

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Temperature – Measuring Equipment <sup>1</sup>	(50 to 600) °C	0.17 °C	Hart 1502A Indicator w/5626 PRT and dry block
Temperature – Measure <sup>1</sup>	(-196 to 600) °C	0.027 °C	Hart 1502 Indicator with 5626 PRT
Temperature Uniformity <sup>1</sup> Surveys	Type K (100 to 1 000) °F (1 000 to 2 000) °F (2 000 to 2 200) °F Type T (32 to 400) °F	2 °F 2.1 °F 2.7 °F 1.6 °F	Datalogger and Thermocouples
Relative Humidity – Measure <sup>1</sup>	(0 to 90) % RH	1.2 %RH	Vaisala HM141/HMP46 Humidity Indicator and Probe
Infrared (IR) Thermometry <sup>1</sup>	(20 to 100) °C (100 to 300) °C (300 to 420) °C (420 to 500) °C	1.5 °C 4.3 °C 6 °C 7.7 °C	Fluke 9132 Infrared Calibrator ε= 0.95, λ = (8 to 14) μm

**Time and Frequency**

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency – Generate <sup>1</sup>	10 MHz	1 x 1 <sup>e-9</sup> Hz/Hz	HP Z3801A GPS Receiver
	(1 to 10) Hz	1 x 1 <sup>e-9</sup> Hz/Hz + 0.57 μHz	Agilent 33250A Function Generator / HP Z3801A GPS Receiver
	(10 to 100) Hz	1 x 1 <sup>e-9</sup> Hz/Hz + 5.7 μHz	
	(100 to 1 000) Hz	1 x 1 <sup>e-9</sup> Hz/Hz + 57 μHz	
	(1 to 10) kHz	1 x 1 <sup>e-9</sup> Hz/Hz + 0.57 mHz	
(10 to 100) kHz	1 x 1 <sup>e-9</sup> Hz/Hz + 5.7 mHz	Agilent E8257D Opt 550 Signal Generator / HP Z3801A GPS Receiver	
(0.1 to 1) MHz	1 x 1 <sup>e-9</sup> Hz/Hz + 57 mHz		
(1 to 10) MHz	1 x 1 <sup>e-9</sup> Hz/Hz + 0.57 Hz		
	(10 to 5 0000) MHz	1 x 1 <sup>e-9</sup> Hz/Hz + 0.57 mHz	



ANSI National Accreditation Board

**Time and Frequency**

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency – Measure <sup>1</sup>	(1 to 10) Hz	$5.2 \times 10^{-9}$ Hz/Hz	Agilent 53131A Frequency Counter / HP Z3801A GPS Receiver
	(10 to 100) Hz	$2.5 \times 10^{-9}$ Hz/Hz	
	(100 to 1 000) Hz	$1.6 \times 10^{-9}$ Hz/Hz	
	(1 to 10) kHz	$1.3 \times 10^{-9}$ Hz/Hz	Agilent 5352B Frequency Counter / HP Z3801A GPS Receiver
	(10 to 100) kHz	$1.2 \times 10^{-9}$ Hz/Hz	
	(100 to 200) kHz	$1.2 \times 10^{-9}$ Hz/Hz	
	(0.2 to 3 000) MHz	$1.2 \times 10^{-9}$ Hz/Hz	Agilent E4448A Spectrum Analyzer / HP Z3801A GPS Receiver
	(500 to 46 000) MHz	$1 \times 10^{-9}$ Hz/Hz + 1.0 Hz	
	(10 to 50 000) MHz	$1 \times 10^{-9}$ Hz/Hz + 0.1 Hz	
Type I (digital) Timers	(0 to 19.99) s/day (0 to 599) s/month	0.031 s/day 1.1 s/month	Helmut Klein Timometer 4500
Type II (mechanical) Timers	(0 to 320) s/day	0.6 s/day	Helmut Klein Timometer 4500
Tachometers – RPM <sup>1,2</sup>	Up to 100 000 RPM	0.001 % of reading + 0.6R	HP 33250A Signal Generator & LED

**DIMENSIONAL MEASUREMENT**

**1 Dimensional**

Madison, AL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Length	X Axis (0.01 to 5.0) in Y Axis (0.01 to 3.0) in	340 $\mu$ m	Optical comparator

[Return to Site listing](#) (top)    [Go to Notes](#) (bottom)

**Services performed at satellite laboratory**

**Technical Maintenance, Inc.**

3248 Forest View Road  
Rockford, IL 61109

Brian Schickowski (Branch Manager) Phone: 779-774-3877  
Scott Chamberlain (Quality Manager) Phone: 321-242-0890

**CALIBRATION AND DIMENSIONAL MEASUREMENT**

**CALIBRATION**

**Acoustics and Vibration**

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Accelerometers – Acceleration <sup>1</sup>	(0.01 to 10) g (7 < 10) Hz (10 < 100) Hz (100 to 920) Hz (>920 to 5000) Hz (>5 to 10) kHz	1.7 % of reading 1.2 % of reading 1 % of reading 1.4 % of reading 1.9 % of reading	Accelerometer Calibrator
Sound Pressure Level – Generate	Up to 150 dB (19 to 16 000) Hz (16 000 to 20 000) Hz	0.25 dB 0.32 dB	TMS 9350C Acoustic Calibrator

**Chemical Quantities**

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
pH Meters <sup>1,5</sup>	4 pH 7 pH 10 pH	0.035 pH 0.025 pH 0.069 pH	Buffer Solutions
Conductivity Meters <sup>1,5</sup>	2 µS/cm 10 µS/cm 100 µS/cm 1 000 µS/cm 1 413 µS/cm 10 000 µS/cm	0.19 µS/cm 0.17 µS/cm 0.74 µS/cm 3.6 µS/cm 5.7 µS/cm 36 µS/cm	Conductivity Solutions

**Electrical – DC/Low Frequency**

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Current – Generate <sup>1</sup>	(1 to 2.2) nA (2.2 to 22) nA (22 to 220) nA (0.22 to 2.2) $\mu$ A (2.2 to 10) $\mu$ A	93 $\mu$ A/A + 0.007 nA 92 $\mu$ A/A + 0.007 nA 92 $\mu$ A/A + 0.01 nA 36 $\mu$ A/A + 0.1 nA 15 $\mu$ A/A + 1 nA	Fluke 5730A/03 Multifunction Calibrator, Agilent 3458 Option 002 Precision Multimeter
DC Current – Generate <sup>1</sup>	(10 to 220) $\mu$ A 220 $\mu$ A to 2.2 mA (2.2 to 22) mA (22 to 100) mA (100 to 220) mA (0.22 to 1) A (1 to 2.2) A	38 $\mu$ A/A + 5 nA 30 $\mu$ A/A + 7 nA 30 $\mu$ A/A + 44 nA 38 $\mu$ A/A + 0.7 $\mu$ A 45 $\mu$ A/A + 0.7 $\mu$ A 68 $\mu$ A/A + 12 $\mu$ A 105 $\mu$ A/A + 12 $\mu$ A	Fluke 5730A/03 Multifunction Calibrator
	(2.2 to 11) A	274 $\mu$ A/A + 365 $\mu$ A	Fluke 5730A Multifunction Calibrator /5725A Amplifier
DC Current – Generate <sup>1</sup>	(11 to 20.5) A (20 to 100) A	77 $\mu$ A/A + 761 $\mu$ A 77 $\mu$ A/A + 3.8 mA	Fluke 5522A Multiproduct Calibrator, 52120A Transconductance Amplifier
DC Current Clamp Meters <sup>1</sup> Toroidal-Wound	(0 to 1025) A	0.21 % of output + 0.05A	Fluke 5522A Multiproduct Calibrator, . 5500A/ Coil x50
DC Current Clamp Meters <sup>1</sup> Other	(0 to 1025) A	0.39 % of output + 0.38A	
DC Current – Measure <sup>1</sup>	(1 to 10 nA) (10 to 100 nA) (0.1 to 1 $\mu$ A) (1 to 10 $\mu$ A) (10 to 100 $\mu$ A) 0.1 to 1 mA (1 to 10) mA (10 to 100) mA	18.7 $\mu$ A/A + 0.16 pA 7.1 $\mu$ A/A + 1.6 pA 6.1 $\mu$ A/A + 0.016 nA 5.3 $\mu$ A/A + 0.016 nA 5.3 $\mu$ A/A + 0.16 nA 5.3 $\mu$ A/A + 0.001 6 $\mu$ A 5.3 $\mu$ A/A + 0.016 $\mu$ A 10 $\mu$ A/A + 0.16 $\mu$ A	Agilent 3458A Option 002 Multimeter, 5450A Resistance Calibrator
DC Current – Measure <sup>1</sup>	(10 to 100) $\mu$ A (0.1 to 1) mA (1 to 10) mA (10 to 100) mA (0.1 to 1) A	20 $\mu$ A/A + 0.8 nA 20 $\mu$ A/A + 5 nA 20 $\mu$ A/A + 0.05 $\mu$ A 35 $\mu$ A/A + 0.5 $\mu$ A 110 $\mu$ A/A + 10 $\mu$ A	Agilent 3458A Multimeter
DC Current – Measure <sup>1</sup>	(1 to 10) A (10 to 100) A	41 $\mu$ A/A + 103 $\mu$ A 53 $\mu$ A/A + 103 $\mu$ A	Agilent 3458A Option 002 Precision Multimeter Standard Resistor L&N 4361 Current Shunt

**Electrical – DC/Low Frequency**

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Current – Measure <sup>1</sup>	(1 to 1 000) A	0.25 % of reading	Agilent 3458A Multimeter Empro Current Shunt
DC Voltage – Generate <sup>1</sup>	(0 to 220) mV (0.22 to 2.2) V (2.2 to 11) V (11 to 22) V (22 to 220) V (220 to 1 100) V	6.8 μV/V + 0.8 μV 4.6 μV/V + 0.9 μV 3 μV/V + 2.5 μV 3 μV/V + 3.9 μV 4.6 μV/V + 38 μV 6.1 μV/V + 385 μV	Fluke 5730A Multifunction Calibrator
DC Voltage – Measure <sup>1</sup>	(0 to 100) mV 100 mV to 1V (1 to 10) V (10 to 100) V (100 to 1 000) V	5.0 μV/V + 1.2 μV 4.0 μV/V + 1.2 μV 4.0 μV/V + 2.1 μV 6.0 μV/V + 30 μV 6.0 μV/V + 100 μV + 12 μV/V x (Vin/1 000) ^2	Agilent 3458A Option 002 Multimeter
DC Voltage – Measure <sup>1</sup>	(0 to 1) kV  (1 to 10) kV	0.03 % of reading + 0.000 032 kV  0.03 % of reading + 0.000 12 kV + 1.5 μV/V x (Vin/1 000) ^2	Vitrek 4700 Digital HV Meter
DC Voltage – Measure <sup>1</sup>	(1 to 60) kV (1 to 120) kV	0.1 % of reading 0.1 % of reading	Ross VD60, VD120 High Voltage Divider, HP34401A Multimeter
DC Voltage – Measure <sup>1</sup> Fixed Points	100 mV 1 V 10 V 100 V 1 000 V	3.1 μV/V 0.61 μV/V 0.5 μV/V 0.52 μV/V 0.63 μV/V	Fluke 732B Reference Standard, 752A Divider
Phase Angle - Generate	(-180 to 180)° (10 to 65) Hz (65 to 500) Hz 500 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz	0.08° 0.19° 0.38° 1.9° 3.8° 7.6°	Fluke 5522A Multiproduct Calibrator
Phase Angle – Measure 10 mV to 350 V	(0 to 360)° 5 Hz to 2 kHz (2 to 5) kHz (5 to 10) kHz (10 to 50) kHz (50 to 100) kHz (100 to 200) kHz	0.02° 0.03° 0.04° 0.05° 0.1° 0.2°	Clarke-Hess 6000A Phase Meter



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Generate <sup>1</sup>	(0.22 to 2.2) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	228 $\mu\text{V/V} + 3.9 \mu\text{V}$ 88 $\mu\text{V/V} + 3.9 \mu\text{V}$ 76 $\mu\text{V/V} + 3.9 \mu\text{V}$ 190 $\mu\text{V/V} + 3.9 \mu\text{V}$ 457 $\mu\text{V/V} + 4.6 \mu\text{V}$ 989 $\mu\text{V/V} + 9.2 \mu\text{V}$ 1.3 mV/V + 19 $\mu\text{V}$ 2.6 mV/V + 19 $\mu\text{V}$	Fluke 5730A Multifunction Calibrator
AC Voltage – Generate <sup>1</sup>	(2.2 to 22) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz (22 to 220) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz 220 mV to 2.2 V (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	228 $\mu\text{V/V} + 3.9 \mu\text{V}$ 88 $\mu\text{V/V} + 3.9 \mu\text{V}$ 76 $\mu\text{V/V} + 3.9 \mu\text{V}$ 190 $\mu\text{V/V} + 3.9 \mu\text{V}$ 457 $\mu\text{V/V} + 4.6 \mu\text{V}$ 989 $\mu\text{V/V} + 9.2 \mu\text{V}$ 1.3 mV/V + 19 $\mu\text{V}$ 2.6 mV/V + 19 $\mu\text{V}$ 228 $\mu\text{V/V} + 11.4 \mu\text{V}$ 88 $\mu\text{V/V} + 6.1 \mu\text{V}$ 53 $\mu\text{V/V} + 6.1 \mu\text{V}$ 114 $\mu\text{V/V} + 6.1 \mu\text{V}$ 304 $\mu\text{V/V} + 15.2 \mu\text{V}$ 609 $\mu\text{V/V} + 19 \mu\text{V}$ 1.3 mV /V + 22.8 $\mu\text{V}$ 2.5 mV /V + 45.7 $\mu\text{V}$ 228 $\mu\text{V/V} + 38 \mu\text{V}$ 84 $\mu\text{V/V} + 15 \mu\text{V}$ 37 $\mu\text{V/V} + 8 \mu\text{V}$ 61 $\mu\text{V/V} + 9 \mu\text{V}$ 76 $\mu\text{V/V} + 30 \mu\text{V}$ 304 $\mu\text{V/V} + 76 \mu\text{V}$ 913 $\mu\text{V} /V + 190 \mu\text{V}$ 1.5 mV/V + 304 $\mu\text{V}$	Fluke 5730A Multifunction Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Generate <sup>1</sup>	(2.2 to 22) V (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	228 $\mu$ V/V + 380 $\mu$ V 84 $\mu$ V/V + 152 $\mu$ V 37 $\mu$ V/V + 54 $\mu$ V 61 $\mu$ V/V + 91 $\mu$ V 76 $\mu$ V/V + 190 $\mu$ V 228 $\mu$ V/V + 609 $\mu$ V 913 $\mu$ V/V + 1.9 mV 1.4 mV/V + 3 mV	Fluke 5730A Multifunction Calibrator
AC Voltage – Generate <sup>1</sup>	(22 to 220) V (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (220 to 750) V 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (750 to 1 000) V 40 Hz to 1 kHz (1 to 20) kHz (20 to 30) kHz	228 $\mu$ V/V + 3.8 mV 84 $\mu$ V/V + 1.5 mV 49 $\mu$ V/V + 0.6 mV 76 $\mu$ V/V + 0.9 mV 137 $\mu$ V/V + 2.3 mV 68 $\mu$ V/V + 3 mV 126 $\mu$ V/V + 5 mV 457 $\mu$ V/V + 8 mV 1.8 mV/V + 34 mV 68 $\mu$ V/V + 3 mV 126 $\mu$ V/V + 5 mV 457 $\mu$ V/V + 8 mV	Fluke 5730A Multifunction Calibrator /5725A Amplifier
AC Voltage – Generate <sup>1</sup> Wideband Absolute	(0.3 to 1.1) mV (10 to 30 Hz) 30 Hz to 500 kHz (0.5 to 1.2) MHz (1.2 to 2) MHz (2 to 12) MHz (12 to 20) MHz (20 to 30) MHz (1.1 to 3.3) mV (10 to 30 Hz) 30 Hz to 500 kHz (0.5 to 1.2) MHz (1.2 to 2) MHz (2 to 12) MHz (12 to 20) MHz (20 to 30) MHz	0.65 % of output + 1.5 $\mu$ V 0.61 % of output + 1.5 $\mu$ V 0.63 % of output + 3.8 $\mu$ V 0.63 % of output + 3.8 $\mu$ V 0.68 % of output + 3.8 $\mu$ V 0.76 % of output + 3.8 $\mu$ V 1.3 % of output + 12.9 $\mu$ V 0.58 % of output + 2.3 $\mu$ V 0.53 % of output + 2.3 $\mu$ V 0.54 % of output + 4.6 $\mu$ V 0.54 % of output + 4.6 $\mu$ V 0.58 % of output + 4.6 $\mu$ V 0.65 % of output + 4.6 $\mu$ V 1.3 % of output + 4.6 $\mu$ V	Fluke 5730A Multifunction Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Generate <sup>1</sup> Wideband Absolute	(3.3 to 11) mV		Fluke 5730A Multifunction Calibrator
	(10 to 30 Hz)	0.58 % of output + 6.1 μV	
	30 Hz to 500 kHz	0.53 % of output + 6.1 μV	
	(0.5 to 1.2) MHz	0.54 % of output + 8.4 μV	
	(1.2 to 2) MHz	0.54 % of output + 8.4 μV	
	(2 to 12) MHz	0.55 % of output + 8.4 μV	
	(12 to 20) MHz	0.61 % of output + 8.4 μV	
	(20 to 30) MHz	0.93 % of output + 8.4 μV	
	(11 to 33) mV		
	(10 to 30 Hz)	0.52 % of output + 12 μV	
	30 Hz to 500 kHz	0.46 % of output + 12 μV	
	(0.5 to 1.2) MHz	0.47 % of output + 14 μV	
	(1.2 to 2) MHz	0.47 % of output + 14 μV	
	(2 to 12) MHz	0.49 % of output + 14 μV	
	(12 to 20) MHz	0.55 % of output + 14 μV	
	(20 to 30) MHz	0.89 % of output + 14 μV	
	(33 to 110) mV		
	(10 to 30 Hz)	0.52 % of output + 30 μV	
	30 Hz to 500 kHz	0.46 % of output + 30 μV	
	(0.5 to 1.2) MHz	0.47 % of output + 33 μV	
	(1.2 to 2) MHz	0.47 % of output + 33 μV	
	(2 to 12) MHz	0.49 % of output + 33 μV	
	(12 to 20) MHz	0.55 % of output + 33 μV	
	(20 to 30) MHz	0.89 % of output + 33 μV	
	(110 to 330) mV		
	(10 to 30 Hz)	0.45 % of output + 0.1 mV	
	30 Hz to 500 kHz	0.38 % of output + 0.1 mV	
	(0.5 to 1.2) MHz	0.4 % of output + 0.1 mV	
(1.2 to 2) MHz	0.4 % of output + 0.1 mV		
(2 to 12) MHz	0.42 % of output + 0.1 mV		
(12 to 20) MHz	0.49 % of output + 0.1 mV		
(20 to 30) MHz	0.85 % of output + 0.1 mV		
(0.33 to 1.1) V			
(10 to 30 Hz)	0.45 % of output + 0.3 mV		
30 Hz to 500 kHz	0.38 % of output + 0.3 mV		
(0.5 to 1.2) MHz	0.4 % of output + 0.3 mV		
(1.2 to 2) MHz	0.4 % of output + 0.3 mV		
(2 to 12) MHz	0.42 % of output + 0.3 mV		
(12 to 20) MHz	0.49 % of output + 0.3 mV		
(20 to 30) MHz	0.85 % of output + 0.3 mV		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Generate <sup>1</sup> Wideband Absolute	(1.1 to 3.5) V (10 to 30 Hz) 30 Hz to 500 kHz (0.5 to 1.2) MHz (1.2 to 2) MHz (2 to 12) MHz (12 to 20) MHz (20 to 30) MHz	0.39 % of output + 0.4 mV 0.3 % of output + 0.4 mV 0.32 % of output + 0.4 mV 0.32 % of output + 0.4 mV 0.35 % of output + 0.4 mV 0.44 % of output + 0.4 mV 0.82 % of output + 0.4 mV	Fluke 5730A Multifunction Calibrator
AC Voltage – Measure <sup>1</sup>	(1 to 10) mV (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz 100 kHz to 1 MHz (1 to 4) MHz (4 to 8) MHz (10 to 100) mV (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz (2 to 4) MHz (4 to 8) MHz (8 to 10) MHz (0.1 to 1) V (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz (2 to 4) MHz (4 to 8) MHz (8 to 10) MHz	0.03 % of reading + 30 μV 0.02 % of reading + 1.1 μV 0.03 % of reading + 1.1 μV 0.1 % of reading + 1.1 μV 0.5 % of reading + 1.1 μV 1.2% of reading + 50 μV 7% of reading + 70 μV 20% of reading + 80 μV 0.007 % of reading + 4 μV 0.007 % of reading + 2 μV 0.014 % of reading + 2 μV 0.03 % of reading + 2 μV 0.08 % of reading + 2 μV 0.3 % of reading + 10 μV 1 % of reading + 10 μV 1.5 % of reading + 70 μV 4 % of reading + 70 μV 4 % of reading + 80 μV 15 % of reading + 100 μV 0.007 % of reading + 40 μV 0.07 % of reading + 20 μV 0.014 % of reading + 20 μV 0.03 % of reading + 20 μV 0.08 % of reading + 20 μV 0.3 % of reading + 100 μV 1 % of reading + 100 μV 1.5 % of reading + 0.7 mV 4 % of reading + 0.7 mV 4 % of reading + 0.8 mV 15 % of reading + 1 mV	Agilent 3458A Multimeter



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure <sup>1</sup>	(1 to 10) V		Agilent 3458A Multimeter
	(1 to 40) Hz	0.007 % of reading + 0.4 mV	
	40 Hz to 1 kHz	0.007 % of reading + 0.2 mV	
	(1 to 20) kHz	0.014 % of reading + 0.2 mV	
	(20 to 50) kHz	0.03 % of reading + 0.2 mV	
	(50 to 100) kHz	0.08 % of reading + 0.2 mV	
	(100 to 300) kHz	0.3 % of reading + 1 mV	
	300 kHz to 1 MHz	1 % of reading + 1 mV	
	(1 to 2) MHz	1.5 % of reading + 7 mV	
	(2 to 4) MHz	4 % of reading + 7 mV	
	(4 to 8) MHz	4 % of reading + 8 mV	
	(8 to 10) MHz	15 % of reading + 10 mV	
	(10 to 100) V		
	(1 to 40) Hz	0.02 % of reading + 4 mV	
	40 Hz to 1 kHz	0.02 % of reading + 2 mV	
	(1 to 20) kHz	0.02 % of reading + 2 mV	
	(20 to 50) kHz	0.035 % of reading + 2 mV	
	(50 to 100) kHz	0.12 % of reading + 2 mV	
(100 to 300) kHz	0.4 % of reading + 10 mV		
300 kHz to 1 MHz	1.5 % of reading + 10 mV		
AC Voltage – Measure <sup>1</sup>	(100 to 700) V		Fluke 5790A AC/DC Transfer Standard
	(1 to 40) Hz	0.04 % of reading + 40 mV	
	40 Hz to 1 kHz	0.04 % of reading + 20 mV	
	(1 to 20) kHz	0.06 % of reading + 20 mV	
	(20 to 50) kHz	0.12 % of reading + 20 mV	
	(50 to 100) kHz	0.3 % of reading + 20 mV	



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure <sup>1</sup>	(2.2 to 7) mV		Fluke 5790A AC/DC Transfer Standard
	(10 to 20) Hz	647 $\mu\text{V/V} + 1.2 \mu\text{V}$	
	(20 to 40) Hz	282 $\mu\text{V/V} + 1.2 \mu\text{V}$	
	40 Hz to 20 kHz	160 $\mu\text{V/V} + 1.2 \mu\text{V}$	
	(20 to 50) kHz	304 $\mu\text{V/V} + 1.7 \mu\text{V}$	
	(50 to 100) kHz	457 $\mu\text{V/V} + 2.0 \mu\text{V}$	
	(100 to 300) kHz	913 $\mu\text{V/V} + 3.1 \mu\text{V}$	
	(300 to 500) kHz	989 $\mu\text{V/V} + 6.1 \mu\text{V}$	
	500 kHz to 1 MHz	1.8 mV/V + 6.1 $\mu\text{V}$	
	(7 to 22) mV		
	(10 to 20) Hz	221 $\mu\text{V/V} + 1.2 \mu\text{V}$	
	(20 to 40) Hz	145 $\mu\text{V/V} + 1.2 \mu\text{V}$	
	40 Hz to 20 kHz	84 $\mu\text{V/V} + 1.2 \mu\text{V}$	
	(20 to 50) kHz	160 $\mu\text{V/V} + 1.7 \mu\text{V}$	
	(50 to 100) kHz	236 $\mu\text{V/V} + 2 \mu\text{V}$	
	(100 to 300) kHz	616 $\mu\text{V/V} + 3.1 \mu\text{V}$	
	(300 to 500) kHz	677 $\mu\text{V/V} + 6.1 \mu\text{V}$	
	500 kHz to 1 MHz	1.3 mV/V + 6.1 $\mu\text{V}$	
	(22 to 70) mV		
	(10 to 20) Hz	183 $\mu\text{V/V} + 1.3 \mu\text{V}$	
	(20 to 40) Hz	91 $\mu\text{V/V} + 1.3 \mu\text{V}$	
	40 Hz to 20 kHz	49 $\mu\text{V/V} + 1.3 \mu\text{V}$	
	(20 to 50) kHz	99 $\mu\text{V/V} + 1.7 \mu\text{V}$	
	(50 to 100) kHz	198 $\mu\text{V/V} + 2 \mu\text{V}$	
	(100 to 300) kHz	388 $\mu\text{V/V} + 3.1 \mu\text{V}$	
	(300 to 500) kHz	510 $\mu\text{V/V} + 6.1 \mu\text{V}$	
	500 kHz to 1 MHz	837 $\mu\text{V/V} + 6.1 \mu\text{V}$	
	(70 to 220) mV		
(10 to 20) Hz	160 $\mu\text{V/V} + 1.3 \mu\text{V}$		
(20 to 40) Hz	65 $\mu\text{V/V} + 1.3 \mu\text{V}$		
40 Hz to 20 kHz	29 $\mu\text{V/V} + 1.3 \mu\text{V}$		
(20 to 50) kHz	53 $\mu\text{V/V} + 1.7 \mu\text{V}$		
(50 to 100) kHz	122 $\mu\text{V/V} + 2 \mu\text{V}$		
(100 to 300) kHz	190 $\mu\text{V/V} + 3.1 \mu\text{V}$		
(300 to 500) kHz	289 $\mu\text{V/V} + 6.1 \mu\text{V}$		
500 kHz to 1 MHz	761 $\mu\text{V/V} + 6.1 \mu\text{V}$		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure <sup>1</sup>	(220 to 700) mV		Fluke 5790A AC/DC Transfer Standard
	(10 to 20) Hz	160 $\mu\text{V/V} + 1.3 \mu\text{V}$	
	(20 to 40) Hz	58 $\mu\text{V/V} + 1.3 \mu\text{V}$	
	40 Hz to 20 kHz	25 $\mu\text{V/V} + 1.3 \mu\text{V}$	
	(20 to 50) kHz	39 $\mu\text{V/V} + 1.7 \mu\text{V}$	
	(50 to 100) kHz	60 $\mu\text{V/V} + 2 \mu\text{V}$	
	(100 to 300) kHz	137 $\mu\text{V/V} + 3.1 \mu\text{V}$	
	(300 to 500) kHz	228 $\mu\text{V/V} + 6.1 \mu\text{V}$	
	500 kHz to 1 MHz	730 $\mu\text{V/V} + 6.1 \mu\text{V}$	
	700 mV to 2.2 V		
	(10 to 20) Hz	152 $\mu\text{V/V} + 0.7 \mu\text{V}$	
	(20 to 40) Hz	50 $\mu\text{V/V} + 0.7 \mu\text{V}$	
	40 Hz to 20 kHz	18 $\mu\text{V/V} + 0.7 \mu\text{V}$	
	(20 to 50) kHz	35 $\mu\text{V/V} + 0.7 \mu\text{V}$	
	(50 to 100) kHz	54 $\mu\text{V/V} + 0.7 \mu\text{V}$	
	(100 to 300) kHz	122 $\mu\text{V/V} + 0.7 \mu\text{V}$	
	(300 to 500) kHz	198 $\mu\text{V/V} + 0.7 \mu\text{V}$	
	500 kHz to 1 MHz	685 $\mu\text{V/V} + 0.7 \mu\text{V}$	
	(2.2 to 7) V		
	(10 to 20) Hz	152 $\mu\text{V/V} + 0.89 \mu\text{V}$	
	(20 to 40) Hz	51 $\mu\text{V/V} + 0.89 \mu\text{V}$	
	40 Hz to 20 kHz	18 $\mu\text{V/V} + 0.89 \mu\text{V}$	
	(20 to 50) kHz	37 $\mu\text{V/V} + 0.89 \mu\text{V}$	
	(50 to 100) kHz	62 $\mu\text{V/V} + 0.89 \mu\text{V}$	
	(100 to 300) kHz	145 $\mu\text{V/V} + 0.89 \mu\text{V}$	
	(300 to 500) kHz	304 $\mu\text{V/V} + 0.89 \mu\text{V}$	
	500 kHz to 1 MHz	913 $\mu\text{V/V} + 0.89 \mu\text{V}$	
(7 to 22) V			
(10 to 20) Hz	152 $\mu\text{V/V} + 0.89 \mu\text{V}$		
(20 to 40) Hz	51 $\mu\text{V/V} + 0.89 \mu\text{V}$		
40 Hz to 20 kHz	21 $\mu\text{V/V} + 0.89 \mu\text{V}$		
(20 to 50) kHz	37 $\mu\text{V/V} + 0.89 \mu\text{V}$		
(50 to 100) kHz	62 $\mu\text{V/V} + 0.89 \mu\text{V}$		
(100 to 300) kHz	145 $\mu\text{V/V} + 0.89 \mu\text{V}$		
(200 to 500) kHz	304 $\mu\text{V/V} + 0.89 \mu\text{V}$		
500 kHz to 1 MHz	913 $\mu\text{V/V} + 0.89 \mu\text{V}$		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure <sup>1</sup>	(22 to 70) V		Fluke 5790A AC/DC Transfer Standard
	(10 to 40) Hz	152 μV/V + 5.7 μV	
	40 Hz to 20 kHz	52 μV/V + 5.7 μV	
	(20 to 50) kHz	24 μV/V + 5.7 μV	
	(50 to 100) kHz	43 μV/V + 5.7 μV	
	(100 to 300) kHz	72 μV/V + 5.7 μV	
	(300 to 500) kHz	152 μV/V + 5.7 μV	
	(70 to 220) V		
	(10 to 20) Hz	152 μV/V + 5.7 μV	
	(20 to 40) Hz	52 μV/V + 5.7 μV	
	40 Hz to 20 kHz	24 μV/V + 5.7 μV	
	(20 to 50) kHz	53 μV/V + 5.7 μV	
(50 to 100) kHz	75 μV/V + 5.7 μV		
AC Voltage – Measure <sup>1</sup>	(220 to 700) V		Vitrek 4700 Digital HV Meter
	40 to 20 kHz	31 μV/V + 57 μV	
	(20 to 50) kHz	99 μV/V + 57 μV	
	(50 to 100) kHz	380 μV/V + 57 μV	
	(700 to 1 000) V		
	40 Hz to 20 kHz	29 μV/V + 57 μV	
AC Voltage – Measure <sup>1</sup>	(0 to 1) kV		Ross VD60, VD120 High Voltage Divider, HP 34401A Multimeter
	30 to 200 Hz	0.12 % of reading + 0.000 1 kV	
	200 to 450 Hz	0.12 % of reading + 0.000 15 kV + 1.5 μV/V x (Vin/1 000) ^2	
	(1 to 10) kV		
AC Voltage – Measure <sup>1</sup>	30 to 200 Hz	0.4 % of reading + 0.000 1 kV	
	200 to 450 Hz	0.4 % of reading + 0.000 15 kV + 1.5 μV/V x (Vin/1 000) ^2	
AC Voltage – Measure <sup>1</sup>	60 Hz		
	(1 to 10) kV	0.5 % of reading + 0.002 kV	
AC Voltage – Measure Wideband Flatness Relative to 1 kHz	(1 to 85) kV	0.5 % of reading + 0.02 kV	Fluke 5790A Option 003 AC/DC Transfer Standard
	(0.6 to 2.2) mV		
	(10 to 30) Hz	0.08 % of reading + 0.1 μV	
	30 Hz to 120 kHz	0.04 % of reading + 0.1 μV	
	120 kHz 2 MHz	0.05 % of reading + 0.8 μV	
	(2 to 10) MHz	0.13 % of reading + 0.8 μV	
AC Voltage – Measure <sup>1</sup>	(10 to 20) MHz	0.23 % of reading + 0.8 μV	
	(20 to 30) MHz	0.53 % of reading + 1.6 μV	



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure Wideband Flatness Relative to 1 kHz	(2.2 to 7) mV		Fluke 5790A Option 003 AC/DC Transfer Standard
	(10 to 30) Hz	0.08 % of reading + 0.1 $\mu$ V	
	30 Hz to 120 kHz	0.04 % of reading + 0.1 $\mu$ V	
	120 kHz 2 MHz	0.05 % of reading + 0.8 $\mu$ V	
	(2 to 10) MHz	0.08 % of reading + 0.8 $\mu$ V	
	(10 to 20) MHz	0.13 % of reading + 0.8 $\mu$ V	
	(20 to 30) MHz	0.28 % of reading + 0.8 $\mu$ V	
	(7 to 22) mV		
	(10 to 30) Hz	0.08 % of reading + 0.1 $\mu$ V	
	30 Hz to 120 kHz	0.04 % of reading + 0.1 $\mu$ V	
	120 kHz 2 MHz	0.05 % of reading + 0.1 $\mu$ V	
	(2 to 10) MHz	0.08 % of reading + 0.1 $\mu$ V	
	(10 to 20) MHz	0.13 % of reading + 0.1 $\mu$ V	
	(20 to 30) MHz	0.28 % of reading + 0.1 $\mu$ V	
	(22 to 70) mV		
	(10 to 30) Hz	0.08 % of reading + 0.6 $\mu$ V	
	30 Hz to 2 MHz	0.04 % of reading + 0.6 $\mu$ V	
	(2 to 10) MHz	0.08 % of reading + 0.6 $\mu$ V	
	(10 to 20) MHz	0.11 % of reading + 0.6 $\mu$ V	
	(20 to 30) MHz	0.27 % of reading + 0.6 $\mu$ V	
	(70 to 220) mV		
	(10 to 30) Hz	0.08 % of reading	
	30 Hz to 500 kHz	0.03 % of reading	
	500 kHz to 2 MHz	0.04 % of reading	
	(2 to 10) MHz	0.08 % of reading	
	(10 to 20) MHz	0.11 % of reading	
	(20 to 30) MHz	0.27 % of reading	
	(220 to 700) mV		
(10 to 30) Hz	0.08 % of reading		
30 Hz to 500 kHz	0.02 % of reading		
500 kHz to 2 MHz	0.04 % of reading		
(2 to 10) MHz	0.08 % of reading		
(10 to 20) MHz	0.11 % of reading		
(20 to 30) MHz	0.27 % of reading		
(0.7 to 2) V			
(10 to 30) Hz	0.08 % of reading		
30 Hz to 500 kHz	0.02 % of reading		
500 kHz to 2 MHz	0.04 % of reading		
(2 to 10) MHz	0.08 % of reading		
(10 to 20) MHz	0.11 % of reading		
(20 to 30) MHz	0.27 % of reading		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure Wideband Flatness Relative to 1 kHz	(2 to 7) V (10 to 30) Hz 30 Hz to 500 kHz 500 kHz to 2 MHz (2 to 10) MHz (10 to 20) MHz (20 to 30) MHz	0.08 % of reading 0.02 % of reading 0.04 % of reading 0.08 % of reading 0.11 % of reading 0.27 % of reading	Fluke 5790A Option 003 AC/DC Transfer Standard
AC Current – Generate <sup>1</sup>	(9 to 220) $\mu$ A (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (0.22 to 2.2) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (2.2 to 22) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (22 to 220) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (0.22 to 2.2) A 20 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	228 $\mu$ A/A + 15 nA 152 $\mu$ A/A + 10 nA 91 $\mu$ A/A + 8 nA 266 $\mu$ A/A + 12 nA 989 $\mu$ A/A + 61 nA 228 $\mu$ A/A + 39 nA 152 $\mu$ A/A + 31 nA 91 $\mu$ A/A + 31 nA 183 $\mu$ A/A + 99 nA 989 $\mu$ A/A + 609 nA 228 $\mu$ A/A + 385 nA 152 $\mu$ A/A + 310 nA 91 $\mu$ A/A + 310 nA 183 $\mu$ A/A + 536 nA 989 $\mu$ A/A + 4.6 $\mu$ A 228 $\mu$ A/A + 4 $\mu$ A 152 $\mu$ A/A + 3 $\mu$ A 91 $\mu$ A/A + 2 $\mu$ A 183 $\mu$ A/A + 3 $\mu$ A 989 $\mu$ A/A + 9 $\mu$ A 228 $\mu$ A/A + 31 $\mu$ A 380 $\mu$ A/A + 76 $\mu$ A 6.1 mA/A + 152 $\mu$ A	Fluke 5730A Multifunction Calibrator
AC Current – Generate <sup>1</sup>	(2.2 to 11) A 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	350 $\mu$ A/A + 141 $\mu$ A 723 $\mu$ A/A + 295 $\mu$ A 2.7 mA/A + 573 $\mu$ A	Fluke 5730A Multifunction Calibrator /5725A Amplifier



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Generate <sup>1</sup>	(11 to 20.5) A (45 to 100) Hz 100 Hz to 1 kHz (1 Hz to 5) kHz	0.09 % of output + 5 mA 0.11 % of output + 5 mA 2.3 % of output + 5 mA	Fluke 5522A Multiproduct Calibrator
AC Current – Generate <sup>1</sup>	(11 to 20) A (10 to 300) Hz 300 Hz to 1 kHz (1 to 3) kHz (3 to 6) kHz (6 to 10) kHz (20 to 120) A (10 to 300) Hz 300 Hz to 1 kHz (1 to 3) kHz (3 to 6) kHz (6 to 10) kHz	255 $\mu$ A/A + 10 mA 769 $\mu$ A/A + 11 mA 0.23 % of output + 31 mA 0.76 % of output + 62 mA 2.3 % of output + 92 mA 255 $\mu$ A/A + 28 mA 769 $\mu$ A/A + 92 mA 0.23 % of output + 228 mA 0.76 % of output + 411 mA 3 % of output + 685 mA	Fluke 5522A Multiproduct Calibrator, Fluke 52120A Transconductance Amplifier
AC Current Clamps <sup>1</sup> Toroidal – Wound	(16.5 to 1025) A (45 to 65) Hz (16.5 to 150) A (65 to 440) Hz	0.23 % of output + 0.1 A 0.64 % of output + 0.085 A	Fluke 5522A Multiproduct Calibrator, 5500A/Coil x50
AC Current Clamps <sup>1</sup> Other	(16.5 to 1025) A (45 to 65) Hz (16.5 to 150) A (65 to 440) Hz	0.43 % of output + 0.53 A 0.76 % of output + 0.68 A	Fluke 5522A Multiproduct Calibrator, 5500A/Coil x50
AC Current Clamps - Rogowski	(10 to 100) A (10 to 1000) Hz (1 to 3) kHz (100 to 1 000) A (10 to 1000) Hz (1 to 3) kHz (1 000 to 6 000) A (10 to 1000) Hz (1 000 to 3 500) A (1 to 3) kHz	0.53 % of reading + 0.011 A 0.61 % of reading + 0.015 A 0.53 % of reading + 0.11 A 0.61 % of reading + 0.15 A 0.53 % of reading + 0.69 A 0.61 % of reading + 0.95 A	Fluke 52120 Transconductance Amplifier, Coil6KA Current Coil
AC Current – Measure <sup>1</sup>	(5 to 100) $\mu$ A (10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz	0.4 % of reading + 0.03 $\mu$ A 0.15 % of reading + 0.03 $\mu$ A 0.06 % of reading + 0.03 $\mu$ A	Agilent 3458A/002 Multimeter



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Measure <sup>1</sup>	(0.1 to 1) mA		Agilent 3458A/002 Multimeter
	(10 to 20) Hz	0.4 % of reading + 0.2 $\mu$ A	
	(20 to 45) Hz	0.15 % of reading + 0.2 $\mu$ A	
	(45 to 100) Hz	0.06 % of reading + 0.2 $\mu$ A	
	100 Hz to 5 kHz	0.03 % of reading + 0.2 $\mu$ A	
	(1 to 10) mA		
	(10 to 20) Hz	0.4 % of reading + 2 $\mu$ A	
	(20 to 45) Hz	0.15 % of reading + 2 $\mu$ A	
	(45 to 100) Hz	0.06 % of reading + 2 $\mu$ A	
	100 Hz to 5 kHz	0.03 % of reading + 2 $\mu$ A	
	(10 to 100) mA		
	(10 to 20) Hz	0.4 % of reading + 20 $\mu$ A	
	(20 to 45) Hz	0.15 % of reading + 20 $\mu$ A	
	(45 to 100) Hz	0.06 % of reading + 20 $\mu$ A	
	100 Hz to 5 kHz	0.03 % of reading + 20 $\mu$ A	
AC Current – Measure <sup>1</sup>	(0.1 to 1) A		Fluke 5790A AC/DC Transfer Standard / Holt HCS-1 Current Shunts
	(10 to 20) Hz	0.4 % of reading + 0.2 mA	
	(20 to 45) Hz	0.15 % of reading + 0.2 mA	
	(45 to 100) Hz	0.06 % of reading + 0.2 mA	
	100 Hz to 5 kHz	0.03 % of reading + 0.2 mA	
	(0.22 to 10) mA		
	(10 to 20) Hz	189 $\mu$ A/A + 0.002 $\mu$ A	
	(20 to 40) Hz	90 $\mu$ A/A + 0.002 $\mu$ A	
	40 Hz to 20 kHz	44 $\mu$ A/A + 0.002 $\mu$ A	
	(20 to 50) kHz	76 $\mu$ A/A + 0.002 $\mu$ A	
	(10 to 50) mA		
	(10 to 20) Hz	190 $\mu$ A/A	
	(20 to 40) Hz	87 $\mu$ A/A	
	40 Hz to 20 kHz	42 $\mu$ A/A	
	(20 to 50) kHz	71 $\mu$ A/A	
(50 to 100) mA			
(10 to 20) Hz	190 $\mu$ A/A		
(20 to 40) Hz	88 $\mu$ A/A		
40 Hz to 20 kHz	45 $\mu$ A/A		
(20 to 50) kHz	77 $\mu$ A/A		
(100 to 250) mA			
(10 to 20) Hz	191 $\mu$ A/A		
(20 to 40) Hz	91 $\mu$ A/A		
40 Hz to 20 kHz	42 $\mu$ A/A		
(20 to 50) kHz	78 $\mu$ A/A		

**Electrical – DC/Low Frequency**

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Measure <sup>1</sup>	(250 to 500) mA		Fluke 5790A AC/DC Transfer Standard / Holt HCS-1 Current Shunts
	(10 to 20) Hz	190 $\mu$ A/A	
	(20 to 40) Hz	91 $\mu$ A/A	
	40 Hz to 20 kHz	42 $\mu$ A/A	
	(20 to 50) kHz	82 $\mu$ A/A	
	(0.5 to 1) A		
	(10 to 20) Hz	190 $\mu$ A/A	
	(20 to 40) Hz	92 $\mu$ A/A	
	(0.5 to 1) A		
	40 Hz to 20 kHz	59 $\mu$ A/A	
	(20 to 50) kHz	117 $\mu$ A/A	
	(1 to 2.5) A		
	(10 to 20) Hz	196 $\mu$ A/A	
	(20 to 40) Hz	107 $\mu$ A/A	
	40 Hz to 20 kHz	78 $\mu$ A/A	
	(20 to 50) kHz	129 $\mu$ A/A	
	(2.5 to 5) A		
	(10 to 20) Hz	199 $\mu$ A/A	
	(20 to 40) Hz	110 $\mu$ A/A	
	40 Hz to 20 kHz	85 $\mu$ A/A	
(20 to 50) kHz	177 $\mu$ A/A		
(5 to 10) A			
(10 to 20) Hz	205 $\mu$ A/A		
(20 to 40) Hz	132 $\mu$ A/A		
40 Hz to 20 kHz	103 $\mu$ A/A		
(20 to 50) kHz	149 $\mu$ A/A		
(10 to 20) A			
(10 to 20) Hz	217 $\mu$ A/A		
(20 to 40) Hz	149 $\mu$ A/A		
40 Hz to 20 kHz	127 $\mu$ A/A		
(20 to 50) kHz	184 $\mu$ A/A		
AC Current – Measure <sup>1</sup>	(1 to 3) A		Fluke 8845A Multimeter
	(3 to 5) Hz	0.84 % of reading + 1.4 mA	
	(5 to 10) Hz	0.27 % of reading + 1.4 mA	
	10 Hz to 5 kHz	0.11 % of reading + 1.4 mA	
	(5 to 10) kHz	0.27 % of reading + 16 mA	
	(3 to 10) A		
	(3 to 5) Hz	0.84 % of reading + 4.6 mA	
	(5 to 10) Hz	0.27 % of reading + 4.6 mA	
	10 Hz to 5 kHz	0.11 % of reading + 4.6 mA	
	(5 to 10) kHz	0.27 % of reading + 53 mA	



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Measure <sup>1</sup>	(3 to 30) A 40 Hz to 1 kHz (1 to 5) kHz	0.3 % of reading + 0.07 A 5 % of reading + 0.14 A	Agilent 3458A Multimeter, Keysight 34330A Current Shunt
AC Current – Measure <sup>1</sup>	(10 to 1 000) A (10 to 100) Hz (100 to 500) Hz	1.5 % of reading + 1A 1.9 % of reading+ 1A	Fluke 376 Clamp Meter
Resistance – Generate <sup>1</sup> Fixed Points	(1, 1.9) Ω (10, 19) Ω (100, 190) Ω (1, 1.9) kΩ (10, 19) kΩ 100 kΩ 190 kΩ 1 MΩ 1.9 MΩ 10 MΩ 19 MΩ 100 MΩ	84 μΩ/Ω + 0.1 μΩ 21 μΩ/Ω + 1 μΩ 9.1 μΩ/Ω + 6 μΩ 6.1 μΩ/Ω + 60 μΩ 6.1 μΩ/Ω + 0.6 mΩ 7.6 μΩ/Ω + 6 mΩ 9.1 μΩ/Ω + 6 mΩ 11 μΩ/Ω + 60 mΩ 16 μΩ/Ω + 60 mΩ 35 μΩ/Ω + 0.6 Ω 42 μΩ/Ω + 0.6 Ω 91 μΩ/Ω + 6 Ω	Fluke 5730A Multifunction Calibrator
Resistance – Generate <sup>1</sup> Fixed Points	1 Ω 1.9 Ω 10 Ω 19 Ω 100 Ω 190 Ω 1 kΩ 1.9 kΩ 10 kΩ 19 kΩ 100 kΩ 190 kΩ 1 MΩ 1.9 MΩ 10 MΩ 19 MΩ 100 MΩ	32 μΩ/Ω + 0.6 μΩ 25 μΩ/Ω + 0.6 μΩ 7.5 μΩ/Ω + 6 μΩ 6 μΩ/Ω + 6 μΩ 2.8 μΩ/Ω + 60 μΩ 2.4 μΩ/Ω + 60 μΩ 2.8 μΩ/Ω + 0.6 mΩ 2.4 μΩ/Ω + 0.6 mΩ 2.4 μΩ/Ω + 6 mΩ 2.9 μΩ/Ω + 60 mΩ 2.6 μΩ/Ω + 60 mΩ 4.5 μΩ/Ω + 0.6 Ω 12 μΩ/Ω + 0.6 Ω 14 μΩ/Ω + 6 Ω 17 μΩ/Ω + 6 Ω 82 μΩ/Ω + 60 Ω	Fluke 5450A Resistance Calibrator
Resistance – Generate <sup>1</sup> Fixed Points	1 Ω 10 kΩ	2.6 μΩ/Ω 0.93 μΩ/Ω	Fluke 742A Resistance Standards



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Resistance – Generate <sup>1</sup>	Up to 11 Ω (11 to 33) Ω (33 to 110) Ω (110 to 330) Ω 330 Ω to 1.1 kΩ (1.1 to 3.3) kΩ (3.3 to 11) kΩ (11 to 33) kΩ (33 to 110) kΩ (110 to 330) kΩ (0.33 to 1.1) MΩ (1.1 to 3.3) MΩ (3.3 to 11) MΩ (11 to 33) MΩ (33 to 110) MΩ (110 to 330) MΩ (0.33 to 1.1) GΩ	30 μΩ/Ω + 0.001 Ω 23 μΩ/Ω + 0.001 Ω 21 μΩ/Ω + 0.001 Ω 21 μΩ/Ω + 0.002 Ω 21 μΩ/Ω + 0.002 Ω 21 μΩ/Ω + 0.02 Ω 21 μΩ/Ω + 0.02 Ω 21 μΩ/Ω + 0.2 Ω 21 μΩ/Ω + 0.2 Ω 24 μΩ/Ω + 2 Ω 24 μΩ/Ω + 2 Ω 46 μΩ/Ω + 23 Ω 99 μΩ/Ω + 38 Ω 190 μΩ/Ω + 1.9 kΩ 380 μΩ/Ω + 2.3 kΩ 0.23 % of setting + 76 kΩ 1.1 % of reading + 380 kΩ	Fluke 5522A Multiproduct Calibrator
Resistance – Measure <sup>1</sup>	Up to 12 Ω (10 to 120) Ω 100 Ω to 1.2 kΩ (1 to 12) kΩ (10 to 120) kΩ 100 kΩ to 1.2 MΩ (1 to 12) MΩ (10 to 120) MΩ 100 MΩ to 1.2 GΩ	15 μΩ/Ω + 56 μΩ 12 μΩ/Ω + 0.5 mΩ 10 μΩ/Ω + 0.6 mΩ 10 μΩ/Ω + 5.6 mΩ 10 μΩ/Ω + 56 mΩ 15 μΩ/Ω + 2.2 Ω 50 μΩ/Ω + 120 Ω 500 μΩ/Ω + 1.2 kΩ 0.5 % of reading + 70 kΩ	Agilent 3458A/002 Multimeter
Resistance – Measure <sup>1</sup>	50 Hz (1 to 10) Ω (10 to 100) Ω (100 to 1000) Ω (1 to 10) kΩ (10 to 100) kΩ (0.1 to 1) MΩ (1 to 10) MΩ	0.17 % of reading + 0.003 Ω 0.062 % of reading + 0.006 Ω 0.051 % of reading + 0.06 Ω 0.05 % of reading + 0.000 6 kΩ 0.054 % of reading + 0.006 kΩ 0.09 % of reading + 0.000 06 MΩ 0.45% of reading + 0.000 6 MΩ	IET 1693 RLC Meter



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Resistance – Measure <sup>1</sup>	100/120 Hz		IET 1693 RLC Meter
	(1 to 10) Ω	0.13 % of reading + 0.003 Ω	
	(10 to 100) Ω	0.049 % of reading + 0.006 Ω	
	(100 to 1000) Ω	0.041 % of reading + 0.06 Ω	
	(1 to 10) kΩ	0.04 % of reading + 0.000 6 kΩ	
	(10 to 100) kΩ	0.043 % of reading + 0.006 kΩ	
	(0.1 to 1) MΩ	0.07 % of reading + 0.000 06 MΩ	
	(1 to 10) MΩ	0.34 % of reading + 0.000 6 MΩ	
	1 000 Hz		
	(1 to 10) Ω	0.05 % of reading + 0.003 Ω	
	(10 to 100) Ω	0.023 % of reading + 0.006 Ω	
	(100 to 1000) Ω	0.02 % of reading + 0.06 Ω	
	(1 to 10) kΩ	0.02 % of reading + 0.000 6 kΩ	
	(10 to 100) kΩ	0.021 % of reading + 0.006 kΩ	
	(0.1 to 1) MΩ	0.03 % of reading + 0.000 06 MΩ	
	(1 to 10) MΩ	0.12 % of reading + 0.000 6 MΩ	
	10 kHz		
	(1 to 10) Ω	0.17 % of reading + 0.003 Ω	
	(10 to 100) Ω	0.062 % of reading + 0.006 Ω	
	(100 to 1000) Ω	0.051 % of reading + 0.06 Ω	
	(1 to 10) kΩ	0.05 % of reading + 0.000 6 KΩ	
	(10 to 25.6) kΩ	0.051 % of reading + 0.006 KΩ	
	(25.6 to 100) kΩ	0.19 % of reading + 0.006 KΩ	
	(0.1 to 1) MΩ	0.33 % of reading + 0.000 06 MΩ	
(1 to 10) MΩ	1.8 % of reading + 0.000 6 MΩ		
100 kHz			
(1 to 10) Ω	0.77 % of reading + 0.003 Ω		
(10 to 100) Ω	0.257 % of reading + 0.006 Ω		
(100 to 1000) Ω	0.206 % of reading + 0.06 Ω		
(1 to 10) kΩ	0.23 % of reading + 0.000 6 kΩ		
(10 to 100) kΩ	0.52 % of reading + 0.006 kΩ		
(0.1 to 1) kΩ	3.4 % of reading + 0.000 06 MΩ		
Capacitance – Generate <sup>1</sup> Fixed Points	1 pF	0.17 % of reading	Agilent 16380A Capacitor Set
	10 pF	0.17 % of reading	
	100 pF	0.17 % of reading	
	1 000 pF	0.17 % of reading	



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Capacitance – Generate <sup>1</sup>	(220 to 400) pF (0.4 to 3.299 9) nF (3.3 to 10.999 9) nF (11 to 32.999 9) nF (33 to 109.999) nF (110 to 329.999) nF (0.33 to 1.099 99) μF (1.1 to 3.299 99) μF (3.3 to 10.999 9) μF (11 to 32.999 9) μF (33 to 109.999) μF (110 to 329.999) μF (0.33 to 1.099 99) mF (1.1 to 3.299 99) mF (3.3 to 10.999 9) mF (11 to 32.999 9) mF (33 to 110) mF	0.38 % of output + 7.6pF 0.38 % of output + 0.01 nF 0.19 % of output + 0.01 nF 0.19 % of output + 0.08 nF 0.19 % of output + 0.08 nF 0.19 % of output + 0.23 nF 0.19 % of output + 0.76 nF 0.19 % of output + 2.3 nF 0.19 % of output + 7.6 nF 0.3 % of output + 23 nF 0.34 % of output + 76 nF 0.34 % of output + 228 nF 0.34 % of output + 0.76 μF 0.34 % of output + 2.3 μF 0.34 % of output + 7.6 μF 0.57 % of output + 23 μF 0.84 % of output + 76 μF	Fluke 5522A Multiproduct Calibrator
Capacitance – Measure <sup>1</sup>	50 Hz (50 to 100) pF (100 to 400) pF (400 to 1 000) pF (1 to 10) nF (10 to 100) nF (100 to 1 000) nF (0.1 to 1) μF (1 to 10) μF (10 to 100) μF (100 to 1 000) μF 100 Hz (20 to 100) pF (100 to 400) pF (400 to 1 000) pF (1 to 10) nF (10 to 100) nF (100 to 1 000) nF (0.1 to 1) μF (10 to 100) μF (100 to 1 000) μF	2.6 % of reading + 0.015 pF 1.3 % of reading + 0.06 pF 0.37 % of reading + 0.06 pF 0.18 % of reading + 0.000 6 nF 0.063 % of reading + 0.006 nF 0.051 % of reading + 0.06 nF 0.051 % of reading + 0.000 06 μF 0.05% of reading + 0.000 6 μF 0.054 % of reading + 0.006 μF 0.09 % of reading + 0.06 μF 2.4 % of reading + 0.015 pF 0.52 % of reading + 0.06 pF 0.16 % of reading + 0.06 pF 0.088 % of reading + 0.000 6 nF 0.045 % of reading + 0.006 nF 0.041 % of reading + 0.06 nF 0.041 % of reading + 0.000 06 μF 0.046 % of reading + 0.006 μF 0.1 % of reading + 0.06 μF	IET 1693 LCR Meter



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Capacitance – Measure <sup>1</sup>	120 Hz		IET 1693 LCR Meter
	(20 to 100) pF	2 % of reading + 0.015 pF	
	(100 to 400) pF	0.44 % of reading + 0.06 pF	
	(400 to 1 000) pF	0.14 % of reading + 0.06 pF	
	(1 to 10) nF	0.08 % of reading + 0.000 6 nF	
	(10 to 100) nF	0.044 % of reading + 0.006 nF	
	(100 to 1 000) nF	0.04 % of reading + 0.06 nF	
	(0.1 to 1) μF	0.04 % of reading + 0.000 06 μF	
	(1 to 10) μF	0.041 % of reading + 0.000 6 μF	
	(10 to 100) μF	0.047 % of reading + 0.006 μF	
	(100 to 1 000) μF	0.11 % of reading + 0.06 μF	
	1 000 Hz		
	(1 to 10) pF	1.6 % of reading + 0.014 pF	
	(10 to 20) pF	0.18 % of reading + 0.015 pF	
	(20 to 100) pF	0.1 % of reading + 0.015 pF	
	(100 to 400) pF	0.036 % of reading + 0.06 pF	
	(400 to 1 000) pF	0.024 % of reading + 0.06 pF	
	(1 to 10) nF	0.022 % of reading + 0.000 6 nF	
	(10 to 100) nF	0.02 % of reading + 0.006 nF	
	(100 to 1 000) nF	0.02 % of reading + 0.06 nF	
	(0.1 to 1) μF	0.02 % of reading + 0.000 06 μF	
	(1 to 10) μF	0.022 % of reading + 0.000 6 μF	
	(10 to 100) μF	0.04 % of reading + 0.006 μF	
	(100 to 1 000) μF	0.22 % of reading + 0.06 μF	
	10 kHz		
	(20 to 100) pF	0.3 % of reading + 0.015 pF	
	(100 to 622) pF	0.2 % of reading + 0.06 pF	
	(622 to 1 000) pF	0.051 % of reading + 0.06 pF	
	(1 to 10) nF	0.051 % of reading + 0.000 6 nF	
	(10 to 100) nF	0.051 % of reading + 0.006 nF	
	(100 to 1 000) nF	0.058 % of reading + 0.06 nF	
	(0.1 to 1) μF	0.058 % of reading + 0.000 06 μF	
	(1 to 10) μF	0.13 % of reading + 0.000 6 μF	
	(10 to 100) μF	0.85 % of reading + 0.006 μF	
	(100 to 500) μF	4.1 % of reading + 0.06 μF	



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Capacitance – Measure <sup>1</sup>	100 kHz (20 to 100) pF (100 to 400) pF (400 to 1 000) pF (1 to 10) nF (10 to 100) nF (100 to 1 000) nF (0.1 to 1) μF (1 to 10) μF	0.49 % of reading + 0.015 pF 0.26 % of reading + 0.06 pF 0.21 % of reading + 0.06 pF 0.21 % of reading + 0.000 6 nF 0.24 % of reading + 0.006 nF 0.58 % of reading + 0.06 nF 0.58 % of reading + 0.000 06 μF 4 % of reading + 0.000 6 μF	IET 1693 LCR Meter
Inductance – Measure <sup>1</sup>	100 Hz (40 to 100) μH (0.1 to 0.4) mH (0.4 to 1) mH (1 to 10) mH (10 to 100) mH (0.1 to 1) H (1 to 10) H (10 to 100) H 1 000 Hz (4 to 10) μH (10 to 40) μH (40 to 100) μH (0.1 to 0.4) mH (0.4 to 1) mH (1 to 10) mH (10 to 100) mH (0.1 to 1) H (1 to 10) H (10 to 100) H 10 kHz (1 to 4) μH (4 to 10) μH (10 to 40) μH (40 to 100) μH (0.1 to 0.4) mH (0.4 to 1) mH (1 to 10) mH (10 to 100) mH (0.1 to 0.407) H (0.407 to 1) H (1 to 10) H (10 to 100) H	3.8 % of reading + 0.006 μH 1.6 % of reading + 0.000 06 mH 0.42 % of reading + 0.000 06 mH 0.19 % of reading + 0.0006 mH 0.056 % of reading + 0.006 mH 0.042 % of reading + 0.000 06 H 0.041 % of reading + 0.000 6 H 0.042 % of reading + 0.006 H 1.3 % of reading + 0.003 μH 0.53 % of reading + 0.006 μH 0.15 % of reading + 0.006 μH 0.071 % of reading + 0.000 06 mH 0.033 % of reading + 0.000 06 mH 0.025 % of reading + 0.000 6 mH 0.02 % of reading + 0.006 mH 0.02% of reading + 0.000 06 H 0.02 % of reading + 0.000 6 H 0.026 % of reading + 0.006 H 2.1 % of reading + 0.003 μH 0.56 % of reading + 0.003 μH 0.25 % of reading + 0.006 μH 0.1 % of reading + 0.006 μH 0.071 % of reading + 0.000 06 mH 0.056 % of reading + 0.000 06 mH 0.052 % of reading + 0.000 6 mH 0.051 % of reading + 0.006 mH 0.051 % of reading + 0.006 mH 0.18 % of reading + 0.000 06 H 0.27 % of reading + 0.000 6 H 1.2 % of reading + 0.006 H	IET 1693 LCR Meter



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Inductance – Measure <sup>1</sup>	100 kHz (1 to 4) $\mu$ H (4 to 10) $\mu$ H (10 to 40) $\mu$ H (40 to 100) $\mu$ H (0.1 to 0.4) mH (0.4 to 1) mH (1 to 10) mH (10 to 100) mH (0.1 to 1) H	1.16 % of reading + 0.003 $\mu$ H 0.44 % of reading + 0.003 $\mu$ H 0.3 % of reading + 0.006 $\mu$ H 0.23 % of reading + 0.006 $\mu$ H 0.21 % of reading + 0.000 06 mH 0.21 % of reading + 0.000 06 mH 0.22 % of reading + 0.000 6 mH 0.39 % of reading + 0.006 mH 2.1 % of reading + 0.000 06 H	IET 1693 LCR Meter
Inductance – Generate <sup>1</sup> Fixed Point	100 $\mu$ H 10 kHz	0.25 $\mu$ H	GR 1482-B Standard Inductor
Inductance – Generate <sup>1</sup> Fixed Point	100 mH 100 Hz	0.047mH	GR 1482-L Standard Inductor
Low Frequency Power – Generate	1 mW to 20 kW (45 to 65) Hz, 1 PF DC	0.28 % of reading 0.14 % of reading	Fluke 5522A Multiproduct Calibrator
Electrical Calibration of Thermocouple Indicating Devices <sup>1</sup>	Type B (600 to 800) °C (800 to 1 550) °C (1 550 to 1 820) °C Type C (0 to 1 000) °C (1 000 to 1 800) °C (1 800 to 2 000) °C (2 000 to 2 316) °C Type E (-250 to -200) °C (-200 to -100) °C (-100 to 0) °C (0 to 600) °C (600 to 1 000) °C Type J (-210 to -100) °C (-100 to 800) °C (800 to 1 200) °C Type K (-250 to -200) °C (-200 to -100) °C (-100 to 800) °C (800 to 1 372) °C	0.27 °C 0.22 °C 0.17 °C 0.13 °C 0.18 °C 0.2 °C 0.27 °C 0.19 °C 0.1 °C 0.07 °C 0.07 °C 0.08 °C 0.11 °C 0.07 °C 0.08 °C 0.35 °C 0.13 °C 0.08 °C 0.1 °C	Fluke 7526A Process Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Calibration of Thermocouple Indicating Devices <sup>1</sup>	Type L (-200°C to -100) °C	0.08 °C	Fluke 7526A Process Calibrator
	(-100 to 900) °C	0.07 °C	
	Type N (-250 to -200) °C	0.56 °C	
	(-200 to -100) °C	0.18 °C	
	(-100 to 0) °C	0.1 °C	
	(0 to 100) °C	0.09 °C	
	(100 to 800) °C	0.08 °C	
	(800 to 1 300) °C	0.1 °C	
	Type R (-50 to -25) °C	0.42 °C	
	(-25 to 0) °C	0.34 °C	
	(0 to 100) °C	0.3 °C	
	(100 to 400) °C	0.22 °C	
	(400 to 600) °C	0.17 °C	
	(600 to 1 000) °C	0.16 °C	
	(1 000 to 1 600) °C	0.15 °C	
	(1 600 to 1 767) °C	0.18 °C	
	Type S (-50 to -25) °C	0.39 °C	
	(-25 to 0) °C	0.33 °C	
	(0 to 100) °C	0.29 °C	
	(100 to 400) °C	0.22 °C	
	(400 to 600) °C	0.18 °C	
(600 to 1 600) °C	0.17 °C		
(1 600 to 1 767) °C	0.2 °C		
Type T (-250 to -200) °C	0.26 °C		
(-200 to -100) °C	0.13 °C		
(-100 to 0) °C	0.09 °C		
(0 to 400) °C	0.07 °C		
Type U (-200 to 0) °C	0.13 °C		
(0 to 600) °C	0.08 °C		
Electrical Calibration of RTD Indicators <sup>1</sup>	Pt 385, 100 Ω (-200 to 800) °C	0.05 °C	Fluke 5522A Multiproduct Calibrator
	Pt 3926, 100 Ω (-200 to 630) °C	0.05 °C	



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment		
Electrical Calibration of RTD Indicators <sup>1</sup>	Pt 3916, 100 Ω (-200 to 630) °C	0.05 °C	Fluke 7526A Process Calibrator		
	Pt 385, 200 Ω (-200 to 400) °C (400 to 630) °C	0.4 °C 0.5 °C			
	Pt 385, 500 Ω (-200 to 630) °C	0.17 °C			
	Pt 385, 1 000 Ω (-200 to 630) °C	0.09 °C			
	PtNi 385, 120 Ω (Ni120) (-80 to 0) °C (0 to 100) °C (100 to 260) °C	0.093 °C 0.093 °C 0.16 °C			
	Cu 427, 10 Ω 100 °C to 260) °C	0.35 °C			
	Oscilloscopes Calibration – Generate <sup>1</sup>			Fluke 5820A Oscilloscope Calibrator w/ GHz Option	
	Voltage DC – 50 Ω	(1 to 24.999) mV (25 to 109.99) mV (110mV to 2.1999) V (2.2 to 6.6) V			0.19 % of Output + 31 μV 0.19 % of Output + 36 μV 0.19 % of Output + 87 μV 0.19 % of Output + 0.6 mV
DC – 1 MΩ	(1 to 24.999) mV (25 to 109.99) mV (110mV to 2.1999) V (2.2 to 10.999) V (11 to 130) V	0.019 % of reading + 20 μV 0.019 % of reading + 25 μV 0.019 % of reading + 76 μV 0.019 % of reading + 0.6 mV 0.019 % of reading + 6.0 mV			
Square Wave 10 Hz to 10 kHz – 50 Ω	(1 to 24.999) mVpp (25 to 109.99) mVpp (110mV to 2.1999) Vpp (2.2 to 6.6) Vpp	0.19 % of Output + 31 μV 0.19 % of Output + 36 μV 0.19 % of Output + 87 μV 0.19 % of Output + 0.6 mV			
Square Wave 10 Hz to 1 kHz – 1 MΩ	(1 to 24.999) mV (25 to 109.99) mV (110mV to 2.1999) V (2.2 to 10.999) V (11 to 130) V	0.038 % of reading + 4 μV 0.038 % of reading + 9 μV 0.038 % of reading + 60 μV 0.038 % of reading + 0.6 mV 0.038 % of reading + 6 mV			



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Oscilloscopes Calibration – Generate <sup>1</sup>  Square Wave 1 to 10 kHz - 1MΩ	(1 to 24.999) mV (25 to 109.99) mV (110mV to 2.1999) V (2.2 to 10.999) V (11 to 130) V	0.19 % of reading + 31 μV 0.19 % of reading + 36 μV 0.19 % of reading + 87 μV 0.19 % of reading + 0.6 mV 0.19 % of reading + 6 mV	Fluke 5820A Oscilloscope Calibrator w/ GHz Option
Oscilloscopes Calibration – Generate <sup>1</sup> Leveled Sine Flatness 50 kHz to 10 MHz Reference	3 dB Bandwidth (5 to 50) mVpp 50 kHz to 100 MHz (100 to 300) MHz (300 to 500) MHz (500 to 600) MHz (600 to 1 600) MHz (1 600 to 2 100) MHz 50 mV to 3.5 Vpp 50 kHz to 100 MHz (100 to 300) MHz (300 to 500) MHz (500 to 600) MHz (600 to 1 600) MHz (1 600 to 2 100) MHz 3.5 to 5 Vpp 50 kHz to 100 MHz (100 to 300) MHz (300 to 500) MHz (500 to 600) MHz	0.34 dB 0.36 dB 0.42 dB 0.46 dB 0.5 dB 0.56 dB 0.24 dB 0.24 dB 0.32 dB 0.34 dB 0.4 dB 0.44 dB 0.24 dB 0.24 dB 0.32 dB 0.34 dB	Fluke 5820A Oscilloscope Calibrator w/ GHz Option
Oscilloscopes Calibration – Generate <sup>1</sup> Leveled Sine Flatness	3 dB Bandwidth 50 mV to 3.5Vpp (2 100 to 4 000) MHz (4 000 to 8 000) MHz (8 000 to 18 000) MHz	0.25 dB 0.35 dB 0.46 dB	EPM Power Meter w/ E Series Power Sensors
Oscilloscopes Calibration – Generate Time Marker	500 ps to 20 ms 50 ms to 5 s	0.25 μs/s 1.9 μs/s + 3.8 μHz	Fluke 5820A Oscilloscope Calibrator w/ GHz Option
Oscilloscopes Calibration – Measure <sup>1</sup>  Input Impedance  Resistance  Leakage	(40 to 60) Ω  500 kΩ to 1.5MΩ  (0 to 5.99) V	0.08 % of reading  0.08 % of reading  0.038 % of reading + 0.8 mV	Fluke 5820A Oscilloscope Calibrator w/ GHz Option



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Power Meter Range Calibration <sup>1</sup>	3 $\mu$ W to 100 mW	0.3 % of reading	HP 11683A Range Calibrator

Electrical – RF/Microwave

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency Modulation – Measure <sup>1</sup>  $\beta$ = deviation / rate	250 kHz to 10 MHz Rates 20 Hz to 10 kHz Peak Dev 200 Hz to 40 kHz 10 MHz to 6.6 GHz Rates 50 Hz to 200 kHz Peak Dev 250 Hz to 400 kHz (6.6 to 13.2) GHz Rates 50 Hz to 200 kHz Peak Dev 250 Hz to 400 kHz (13.2 to 26.5) GHz Rates 50 Hz to 200 kHz Peak Dev 250 Hz to 400 kHz (31.15 to 50.0) GHz Rates 50 Hz to 200 kHz Peak Dev 250 Hz to 400 kHz	$\beta > 0.2$ - 1.5 % of reading + 2 Hz $\beta > 1.2$ - 1 % of reading + 2 Hz $\beta > 0.2$ - 1.5 % of reading + 2 Hz $\beta > 0.45$ - 1 % of reading + 2 Hz $\beta > 0.2$ - 2.5 % of reading + 4 Hz $\beta > 8.0$ - 1 % of reading + 4 Hz $\beta > 0.2$ - 3.8 % of reading + 9 Hz $\beta > 16$ - 1 % of reading + 9 Hz $\beta > 0.2$ - 8.5 % of reading + 17 Hz $\beta > 32$ - 1 % of reading + 17 Hz	Agilent N5531S Measuring Receiver System
Amplitude Modulation – Measure <sup>1</sup>	100 kHz to 10 MHz Rate 50 Hz to 10 kHz (5 to 99) % Depth 10 MHz to 3 GHz Rate 50 Hz to 100 kHz (5 to 20) % Depth (20 to 99) % Depth (3 to 26.5) GHz Rate 50 Hz to 100 kHz (5 to 20) % Depth (20 to 99) % Depth (26.5 to 31.15) GHz Rate 50 Hz to 100 kHz (5 to 20) % Depth (20 to 99) % Depth (31.15 to 50) GHz Rate 50 Hz to 100 kHz (5 to 20) % Depth (20 to 99) % Depth	0.75 % of reading + 0.3 digits  2.5 % of reading + 0.4 digits 1.5 % of reading + 0.4 digits  4.5 % of reading + 0.4 digits 1.5 % of reading + 0.4 digits  6.8 % of reading + 0.4 digits 1.9 % of reading + 0.4 digits  26 % of reading + 0.4 digits 6 % of reading + 0.4 digits	Agilent N5531S Measuring Receiver System



ANSI National Accreditation Board

Electrical – RF/Microwave

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Phase Modulation – Measure <sup>1</sup>	100 kHz to 6.6 GHz Deviations > 0.3 rad Deviations > 0.7 rad (6.6 to 13.2) GHz Deviations > 0.6 rad Deviations > 2.0 rad (13.2 to 26.5) GHz Deviations: > 1.2 rad Deviations > 4.0 rad (26.5 to 31.15) GHz Deviations: > 1.3 rad Deviations > 4.0 rad (31.15 to 50) GHz Deviations: > 2.4 rad Deviations > 8.0 rad	3 % of reading + 0.002 rad 1 % of reading + 0.002 rad 3 % of reading + 0.005 rad 1 % of reading + 0.005 rad 3 % of reading + 0.009 rad 1 % of reading + 0.009 rad 3 % of reading + 0.009 rad 1 % of reading + 0.009 rad 3 % of reading + 0.018 rad 1 % of reading + 0.018 rad	Agilent N5531S Measuring Receiver System
RF Flatness – Measure <sup>1</sup>	9 kHz to 2 000MHz (20 to -10) dBm (-10 to -30) dBm (-30 to -40) dBm (-40 to -42) dBm (2 to 14) GHz (20 to -10) dBm (-10 to -30) dBm (-30 to -40) dBm (-40 to -42) dBm (14 to 18) GHz (20 to -10) dBm (-10 to -30) dBm (-30 to -40) dBm (-40 to -42) dBm	0.1 dB 0.1 dB 0.11 dB 0.12 dB 0.10 dB 0.09 dB 0.1 dB 0.11 dB 0.11 dB 0.12 dB 0.12 dB 0.13 dB	Agilent EPM Series Power Meter w/E9304A H18 Power Sensor
RF Attenuation – Measure <sup>1</sup>	(30 to 3 050) MHz (0 to 10) dB (10 to 20) dB (20 to 30) dB (30 to 40) dB (40 to 50) dB (50 to 60) dB (60 to 70) dB (70 to 80) dB (80 to 90) dB (90 to 100) dB (100 to 110) dB	0.02 dB 0.025 dB 0.03 dB 0.035 dB 0.04 dB 0.076 dB 0.081 dB 0.12 dB 0.12 dB 0.13 dB 0.13 dB	Agilent N5531S Measuring Receiver w/N5532B Opt 550 Power Sensor



ANSI National Accreditation Board

Electrical – RF/Microwave

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Attenuation – Measure <sup>1</sup>	(3 050 to 6 600) MHz		Agilent N5531S Measuring Receiver w/N5532B Opt 550 Power Sensor
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.12 dB	
	(90 to 100) dB	0.13 dB	
	(100 to 110) dB	0.19 dB	
	(6 600 to 13 200) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.12 dB	
	(90 to 100) dB	0.13 dB	
	(100 to 110) dB	0.25 dB	
	(13 200 to 19 200) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
(50 to 60) dB	0.076 dB		
(60 to 70) dB	0.081 dB		
(70 to 80) dB	0.12 dB		
(80 to 90) dB	0.12 dB		
(90 to 100) dB	0.13 dB		
(100 to 110) dB	0.31 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Attenuation – Measure 1	(19 200 to 26 500) MHz		Agilent N5531S Measuring Receiver w/N5532B Opt 550 Power Sensor
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.14 dB	
	(90 to 100) dB	0.36 dB	
	(100 to 110) dB	0.82 dB	
	(26 500 to 31 150) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.13 dB	
	(90 to 100) dB	0.33 dB	
	(100 to 110) dB	0.77 dB	
	(31 150 to 41 000) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
(50 to 60) dB	0.076 dB		
(60 to 70) dB	0.081 dB		
(70 to 80) dB	0.14 dB		
(80 to 90) dB	0.36 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Attenuation – Measure <sup>1</sup>	(41 000 to 45 000) MHz		Agilent N5531S Measuring Receiver w/N5532B Opt 550 Power Sensor
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.11 dB	
	(70 to 80) dB	0.24 dB	
	(80 to 90) dB	0.6 dB	
	(45 000 to 50 000) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.11 dB	
	(60 to 70) dB	0.29 dB	
(70 to 80) dB	0.7 dB		
(80 to 90) dB	1.4 dB		
RF Power – Measure <sup>1</sup>	9 kHz to 14 000 MHz		Agilent EPM Series Power Meter w/E9304A H18 Power Sensor
	(20 to 0) dB	0.13 dB	
	(0 to -40) dB	0.15 dB	
	(-40 to -50) dB	0.35 dB	
	(-50 to -55) dB	0.93 dB	
	(14 000 to 18 000) MHz		
	(20 to 0) dB	0.12 dB	
	(0 to -40) dB	0.16 dB	
	(-40 to -50) dB	0.35 dB	
	(-50 to -55) dB	0.93 dB	
RF Power Measure <sup>1</sup>	(10 to 100) MHz		Agilent EPM Series Power Meter RF Power / Keysight N8485A Power Sensor
	(20 to 10) dB	0.07 dB	
	(10 to 0) dB	0.06 dB	
	(0 to -10) dB	0.06 dB	
	(-10 to -20) dB	0.06 dB	
	(-20 to -25) dB	0.11 dB	



ANSI National Accreditation Board

Electrical – RF/Microwave

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power Measure <sup>1</sup>	(100 to 2 000) MHz		Agilent EPM Series Power Meter RF Power / Keysight N8485A Power Sensor
	(20 to 10) dB	0.07 dB	
	(10 to 0) dB	0.07 dB	
	(0 to -10) dB	0.06 dB	
	(-10 to -20) dB	0.07 dB	
	(-20 to -25) dB	0.11 dB	
	(2 000 to 12 400) MHz		
	(20 to 10) dB	0.08 dB	
	(10 to 0) dB	0.08 dB	
	(0 to -10) dB	0.08 dB	
	(-10 to -20) dB	0.08 dB	
	(-20 to -25) dB	0.12 dB	
	(12 400 to 18 000) MHz		
	(20 to 10) dB	0.09 dB	
	(10 to 0) dB	0.08 dB	
	(0 to -10) dB	0.08 dB	
	(-10 to -20) dB	0.09 dB	
	(-20 to -25) dB	0.12 dB	
(18 000 to 26 500) MHz			
(20 to 10) dB	0.12 dB		
(10 to 0) dB	0.12 dB		
(0 to -10) dB	0.12 dB		
(-10 to -20) dB	0.12 dB		
(-20 to -25) dB	0.15 dB		
RF Power Measure <sup>1</sup>	(100 to 2 000) MHz		Agilent EPM Series Power Meter 8487A Power Sensor
	(20 to 10) dB	0.14 dB	
	(10 to 0) dB	0.06 dB	
	(0 to -10) dB	0.07 dB	
	(-10 to -20) dB	0.08 dB	
	(-20 to -25) dB	0.15 dB	
	(2 000 to 12 400) MHz		
	(20 to 10) dB	0.15 dB	
	(10 to 0) dB	0.07 dB	
	(0 to -10) dB	0.07 dB	
	(-10 to -20) dB	0.08 dB	
	(-20 to -25) dB	0.15 dB	



ANSI National Accreditation Board

Electrical – RF/Microwave

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power Measure <sup>1</sup>	(12 400 to 18 000) MHz (20 to 10) dB (10 to 0) dB (0 to -10) dB (-10 to -20) dB (-20 to -25) dB (18 000 to 26 500) MHz (20 to 10) dB (10 to 0) dB (0 to -10) dB (-10 to -20) dB (-20 to -25) dB (26 5000 to 40 000) MHz (20 to 10) dB (10 to 0) dB (0 to -10) dB (-10 to -20) dB (-20 to -25) dB (40 000 to 50 000) MHz (20 to 10) dB (10 to 0) dB (0 to -10) dB (-10 to -20) dB (-20 to -25) dB	0.15 dB 0.08 dB 0.08 dB 0.09 dB 0.15 dB 0.16 dB 0.1 dB 0.1 dB 0.11 dB 0.16 dB 0.17 dB 0.12 dB 0.12 dB 0.13 dB 0.18 dB 0.22 dB 0.19 dB 0.19 dB 0.19 dB 0.23 dB	Agilent EPM Series Power Meter 8487A Power Sensor
RF Power – Measure <sup>1</sup>	100 kHz to 30 MHz (20 to 0) dB (0 to -58) dB (-58 to -78) dB (-78 to -110) dB (-110 to -115) dB (-115 to -120) dB (-120 to -125) dB	0.12 dB 0.13 dB 0.14 dB 0.18 dB 0.2 dB 0.28 dB 0.43 dB	Agilent N5531S Measuring Receiver, N1912A w/E9304A Power Sensor
RF Power – Measure <sup>1</sup>	(30 to 2 000) MHz (30 to 20) dB (20 to 0) dB (0 to -58) dB (-58 to -78) dB (-78 to -110) dB (-110 to -115) dB (-115 to -120) dB (-120 to -125) dB	0.36 dB 0.2 dB 0.22 dB 0.23 dB 0.25 dB 0.27 dB 0.33 dB 0.46 dB	Agilent N5531S Measuring Receiver w/N5532B Opt 550 Power Sensor



ANSI National Accreditation Board

Electrical – RF/Microwave

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Measure <sup>1</sup>	(2 000 to 3 050) MHz		Agilent N5531S Measuring Receiver w/N5532B Opt 550 Power Sensor
	(30 to 20) dB	0.37 dB	
	(20 to 0) dB	0.21 dB	
	(0 to -58) dB	0.23 dB	
	(-58 to -78) dB	0.24 dB	
	(-78 to -110) dB	0.26 dB	
	(-110 to -115) dB	0.28 dB	
	(-115 to -120) dB	0.34 dB	
	(-120 to -125) dB	0.7 dB	
	(3 050 to 6 600) MHz		
	(30 to 20) dB	0.37 dB	
	(20 to 0) dB	0.21 dB	
	(0 to -58) dB	0.23 dB	
	(-58 to -78) dB	0.24 dB	
	(-78 to -110) dB	0.29 dB	
	(-110 to -115) dB	0.38 dB	
	(-115 to -120) dB	0.38 dB	
	(-120 to -125) dB	0.53 dB	
	(6 600 to 13 200) MHz		
	(30 to 20) dB	0.37 dB	
	(20 to 0) dB	0.21 dB	
	(0 to -58) dB	0.23 dB	
	(-58 to -78) dB	0.24 dB	
	(-78 to -110) dB	0.34 dB	
	(-110 to -115) dB	0.46 dB	
	(-115 to -120) dB	0.65 dB	
	(13 200 to 18 000) MHz		
	(30 to 20) dB	0.37 dB	
	(20 to 0) dB	0.21 dB	
	(0 to -58) dB	0.23 dB	
	(-58 to -78) dB	0.24 dB	
	(-78 to -90) dB	0.26 dB	
	(-90 to -95) dB	0.26 dB	
	(-95 to -100) dB	0.26 dB	
	(-100 to -105) dB	0.29 dB	
	(-105 to -110) dB	0.38 dB	
(-110 to -115) dB	0.53 dB		
(-115 to -120) dB	0.75 dB		

**Electrical – RF/Microwave**

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Measure <sup>1</sup>	(18 000 to 19 200) MHz		Agilent N5531S Measuring Receiver w/N5532B Opt 550 Power Sensor
	(30 to 20) dB	0.39 dB	
	(20 to 0) dB	0.25 dB	
	(0 to -58) dB	0.24 dB	
	(-58 to -78) dB	0.25 dB	
	(-78 to -90) dB	0.27 dB	
	(-90 to -95) dB	0.27 dB	
	(-95 to -100) dB	0.27 dB	
	(-100 to -105) dB	0.3 dB	
	(-105 to -110) dB	0.38 dB	
	(-110 to -115) dB	0.53 dB	
	(-115 to -120) dB	0.75 dB	
	(19 200 to 26 500) MHz		
	(30 to 20) dB	0.39 dB	
	(20 to 0) dB	0.25 dB	
	(0 to -58) dB	0.24 dB	
	(-58 to -78) dB	0.25 dB	
	(-78 to -90) dB	0.28 dB	
	(-90 to -95) dB	0.33 dB	
	(-95 to -100) dB	0.43 dB	
	(-100 to -105) dB	0.61 dB	
	(-105 to -110) dB	0.85 dB	
	(-110 to -115) dB	1.2 dB	
	(-115 to -120) dB	1.5 dB	
	(26 500 to 31 150) MHz		
	(30 to 20) dB	0.42 dB	
	(20 to 0) dB	0.3 dB	
	(0 to -58) dB	0.34 dB	
	(-58 to -78) dB	0.34 dB	
	(-78 to -90) dB	0.36 dB	
(-90 to -95) dB	0.39 dB		
(-95 to -100) dB	0.46 dB		
(-100 to -105) dB	0.61 dB		
(-105 to -110) dB	0.82 dB		
(-110 to -115) dB	1.1 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Measure <sup>1</sup>	(31 150 to 41 000) MHz		Agilent N5531S Measuring Receiver w/N5532B Opt 550 Power Sensor
	(30 to 20) dB	0.42 dB	
	(20 to 0) dB	0.3 dB	
	(0 to -58) dB	0.34 dB	
	(-58 to -78) dB	0.35 dB	
	(-78 to -90) dB	0.48 dB	
	(-90 to -95) dB	0.64 dB	
	(-95 to -100) dB	0.87 dB	
	(-100 to -105) dB	1.2 dB	
	(-105 to -110) dB	1.5 dB	
	(41 000 to 45 000) MHz		
	(30 to 20) dB	0.42 dB	
	(20 to 0) dB	0.3 dB	
	(0 to -58) dB	0.34 dB	
	(-58 to -78) dB	0.38 dB	
	(-78 to -90) dB	0.68 dB	
	(-90 to -95) dB	0.93 dB	
	(-95 to -100) dB	1.2 dB	
(-100 to -105) dB	1.6 dB		
RF Power – Generate <sup>1</sup>	(45 000 to 50 000) MHz		
	(30 to 20) dB	0.42 dB	
	(20 to 0) dB	0.3 dB	
	(0 to -58) dB	0.34 dB	
	(-58 to -78) dB	0.67 dB	
	(-78 to -90) dB	1.4 dB	
	10 Hz to 100 kHz		
	(24 to -48) dBm	0.06 dB	
	100 kHz to 10 MHz		
	(24 to -48) dBm	0.07 dB	
	(-48 to -74) dBm	0.16 dB	
	(-74 to -94) dBm	0.39 dB	
(10 to 128) MHz			
(24 to -48) dBm	0.07 dB		
(-48 to -84) dBm	0.09 dB		
(-84 to -94) dBm	0.24 dB		
(-94 to -124) dBm	0.54 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Generate <sup>1</sup>	(128 to 300) MHz		Fluke 9640A RF Reference Source
	(20 to -48) dBm	0.08 dB	
	(-48 to -74) dBm	0.09 dB	
	(-74 to -84) dBm	0.24 dB	
	(-84 to -94) dBm	0.39 dB	
	(-94 to -124) dBm	1.2 dB	
	(300 to 1.4) GHz		
	(20 to -48) dBm	0.16 dB	
	(-48 to -74) dBm	0.31 dB	
	(-74 to -84) dBm	0.39 dB	
	(-84 to -94) dBm	0.77 dB	
	(-94 to -124) dBm	1.2 dB	
	(1.4 to 3) GHz		
	(14 to -48) dBm	0.24 dB	
	(-48 to -74) dBm	0.39 dB	
(-74 to -94) dBm	0.77 dB		
(-94 to -124) dBm	1.2 dB		
RF Power – Generate <sup>1</sup>	(30 to 2 000) MHz		Agilent N5531S Measuring Receiver w/N5532B Opt 550 Power Sensor, 83650B Signal Generator
	(20 to 0) dB	0.29 dB	
	(0 to -58) dB	0.3 dB	
	(-58 to -78) dB	0.31 dB	
	(-78 to -110) dB	0.32 dB	
	(2 000 to 3 050) MHz		
	(20 to 0) dB	0.34 dB	
	(0 to -58) dB	0.35 dB	
	(-58 to -78) dB	0.36 dB	
	(-78 to -110) dB	0.39 dB	
	(3 050 to 6 600) MHz		
	(20 to 0) dB	0.34 dB	
	(0 to -58) dB	0.35 dB	
	(-58 to -78) dB	0.36 dB	
	(-78 to -110) dB	0.39 dB	
	(6 600 to 13 200) MHz		
	(20 to 0) dB	0.34 dB	
	(0 to -58) dB	0.35 dB	
(-58 to -78) dB	0.36 dB		
(-78 to -110) dB	0.42 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Generate <sup>1</sup>	(13 200 to 18 000) MHz		Agilent N5531S Measuring Receiver w/N5532B Opt 550 Power Sensor, 83650B Signal Generator
	(15 to 0) dB	0.34 dB	
	(0 to -58) dB	0.35 dB	
	(-58 to -78) dB	0.36 dB	
	(-78 to -110) dB	0.46 dB	
	(18 000 to 19 200) MHz		
	(15 to 0) dB	0.41 dB	
	(0 to -58) dB	0.41 dB	
	(-58 to -78) dB	0.42 dB	
	(-78 to -110) dB	0.5 dB	
	(19 200 to 26 500) MHz		
	(15 to 0) dB	0.41 dB	
	(0 to -58) dB	0.41 dB	
	(-58 to -78) dB	0.42 dB	
	(-78 to -110) dB	0.9 dB	
	(26 500 to 31 150) MHz		
	(15 to 0) dB	0.62 dB	
	(0 to -58) dB	0.63 dB	
	(-58 to -78) dB	0.64 dB	
	(-78 to -110) dB	0.96 dB	
	(31 150 to 41 000) MHz		
	(10 to 0) dB	0.82 dB	
	(0 to -58) dB	0.83 dB	
	(-58 to -78) dB	0.84 dB	
(-78 to -100) dB	1.1 dB		
(41 000 to 45 000) MHz			
(10 to 0) dB	0.82 dB		
(0 to -58) dB	0.83 dB		
(-58 to -78) dB	0.85 dB		
(-78 to -100) dB	1.4 dB		
(45 000 to 50 000) MHz			
(10 to 0) dB	0.82 dB		
(0 to -58) dB	0.83 dB		
(-58 to -78) dB	1 dB		
(-78 to -90) dB	1.5 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Power Sensor (Cal Factor)	Type-N (50 Ω)		M is the mismatch Uncertainty  Agilent EPM Series Power Meter RF Power / Keysight N8482A H84 Power Sensor
	0.1 MHz	1.2 % of reading + M	
	0.3 MHz	1 % of reading + M	
	0.5 MHz	1 % of reading + M	
	1 MHz	1 % of reading + M	
	3 MHz	1 % of reading + M	
	5 MHz	1 % of reading + M	
	10 MHz	0.97 % of reading + M	
	30 MHz	0.95 % of reading + M	
	50 MHz	0.85 % of reading + M	
	100 MHz	0.85 % of reading + M	
	300 MHz	0.89 % of reading + M	
	500 MHz	0.89 % of reading + M	
	1 000 MHz	0.89 % of reading + M	
	1 500 MHz	0.89 % of reading + M	
	2 000 MHz	0.89 % of reading + M	
	2 500 MHz	0.88 % of reading + M	
	3 000 MHz	0.88 % of reading + M	
3 500 MHz	0.9 % of reading + M		
3 700 MHz	0.9 % of reading + M		
4 000 MHz	0.92 % of reading + M		
4 200 MHz	0.93 % of reading + M		
5 000 MHz	0.93 % of reading + M		
6 000 MHz	0.93 % of reading + M		
RF Power – Power Sensor (Cal Factor)	Type-N (50 Ω)		M is the mismatch Uncertainty  Agilent EPM Series Power Meter RF Power / Keysight N8481A H84 Power Sensor
	10 MHz	1.3 % of reading + M	
	30 MHz	1.3 % of reading + M	
	50 MHz	1.2 % of reading + M	
	100 MHz	1.2 % of reading + M	
	300 MHz	1.2 % of reading + M	
	500 MHz	1.2 % of reading + M	
	800 MHz	1.2 % of reading + M	
	1 000 MHz	1.2 % of reading + M	
	1 200 MHz	1.2 % of reading + M	
	1 500 MHz	1.2 % of reading + M	
2 000 MHz	1.2 % of reading + M		



ANSI National Accreditation Board

Electrical – RF/Microwave

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Power Sensor (Cal Factor)	Type-N (50 Ω)		M is the mismatch Uncertainty  Agilent EPM Series Power Meter RF Power / Keysight N8481A H84 Power Sensor
	3 000 MHz	1.2 % of reading + M	
	4 000 MHz	1.2 % of reading + M	
	5 000 MHz	1.3 % of reading + M	
	6 000 MHz	1.3 % of reading + M	
	7 000 MHz	1.3 % of reading + M	
	8 000 MHz	1.3 % of reading + M	
	9 000 MHz	1.4 % of reading + M	
	10 000 MHz	1.4 % of reading + M	
	11 000 MHz	1.4 % of reading + M	
	12 000 MHz	1.4 % of reading + M	
	12 400 MHz	1.4 % of reading + M	
	13 000 MHz	1.5 % of reading + M	
	14 000 MHz	1.5 % of reading + M	
	15 000 MHz	1.5 % of reading + M	
	16 000 MHz	1.6 % of reading + M	
	17 000 MHz	1.6 % of reading + M	
18 000 MHz	1.6 % of reading + M		
AM Distortion Measure <sup>1</sup> Rate 20 Hz to 1 kHz	(0.1 to 10) MHz		Agilent N5531S Measuring Receiver
	AM Depth > 1% (0 to -20) dB	1.2 dB	
	(-20 to -30) dB	2.2 dB	
	AM Depth > 3% (0 to -20) dB	1 dB	
	(-20 to -30) dB	1.3 dB	
	(-30 to -40) dB	2.4 dB	
	10 MHz to 26.5 GHz		
	AM Depth > 1% (0 to -20) dB	1.3 dB	
	(-20 to -30) dB	2.5 dB	
	AM Depth > 3% (0 to -20) dB	1.1 dB	
	(-20 to -30) dB	1.4 dB	
	(-30 to -40) dB	3 dB	
	26.5 MHz to 50.0 GHz		
	AM Depth > 3% (0 to -20) dB	1.8 dB	
	AM Depth > 5% (0 to -20) dB	1.5 dB	
(-20 to -30) dB	3.5 dB		

**Electrical – RF/Microwave**

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
FM Distortion Measure <sup>1</sup> Rate 20 Hz to 1 kHz	(1 to 6 600) MHz Dev 500 Hz to 2 kHz (0 to -20) dB (-20 to -30) dB (-30 to -40) dB Dev > 2 kHz (0 to -20) dB (-20 to -30) dB (-30 to -40) dB (-40 to -50) dB	0.26 dB 0.79 dB 2.28 dB 0.09 dB 0.27 dB 0.83 dB 2.39 dB	Agilent N5531S Measuring Receiver
FM Distortion Measure <sup>1</sup> Rate 20 Hz to 1 kHz	(6.6 to 13.2) GHz Dev > 2.3 kHz (0 to -20) dB (-20 to -30) dB (-30 to -40) dB Dev > 4.5 kHz (0 to -20) dB (-20 to -30) dB (-30 to -40) dB (-40 to -50) dB (13.2 to 31.15) GHz Dev > 2.7 kHz (0 to -20) dB (-20 to -30) dB (-30 to -40) dB Dev > 6.0 kHz (0 to -20) dB (-20 to -30) dB (-30 to -40) dB (-40 to -50) dB (31.15 to 50.0) GHz Dev > 4 kHz (0 to -20) dB (-20 to -30) dB (-30 to -40) dB Dev > 12.0 kHz (0 to -20) dB (-20 to -30) dB (-30 to -40) dB (-40 to -50) dB	0.26 dB 0.79 dB 2.3 dB 0.09 dB 0.27 dB 0.83 dB 2.4 dB 0.26 dB 0.79 dB 2.3 dB 0.09 dB 0.27 dB 0.83 dB 2.4 dB 0.26 dB 0.79 dB 2.3 dB 0.09 dB 0.27 dB 0.83 dB 2.3 dB	Agilent N5531S Measuring Receiver



ANSI National Accreditation Board

Electrical – RF/Microwave

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
PM Distortion Measure <sup>1</sup>	(1 to 6 600) MHz Rate 20 to 500 Hz Dev > 0.8 rad (0 to -20) dB (-20 to -30) dB (-30 to -40) dB Dev > 2.5 rad (0 to -20) dB (-20 to -30) dB (-30 to -40) dB (-40 to -50) dB	0.26 dB 0.79 dB 2.3 dB 0.09 dB 0.27 dB 0.83 dB 2.3 dB	Agilent N5531S Measuring Receiver
PM Distortion Measure <sup>1</sup>	(1 to 6 600) MHz Rate (500 to 1 000) Hz Dev > 0.4 rad (0 to -20) dB (-20 to -30) dB (-30 to -40) dB Dev > 1.0 rad (0 to -20) dB (-20 to -30) dB (-30 to -40) dB (-40 to -50) dB (6.6 to 13.2) GHz Rate (20 to 500) Hz Dev > 1.8 rad (0 to -20) dB (-20 to -30) dB (-30 to -40) dB Dev > 5.5 rad (0 to -20) dB (-20 to -30) dB (-30 to -40) dB (-40 to -50) dB	0.26 dB 0.79 dB 2.3 dB 0.09 dB 0.27 dB 0.83 dB 2.3 dB 0.26 dB 0.79 dB 2.3 dB 0.09 dB 0.27 dB 0.83 dB 2.3 dB	Agilent N5531S Measuring Receiver

Electrical – RF/Microwave

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
PM Distortion Measure <sup>1</sup>	(6.6 to 13.2) GHz		
	Rate 500 to 1 000 Hz		
	Dev > 0.8 rad		
	(0 to -20) dB	0.26 dB	
	(-20 to -30) dB	0.79 dB	
	(-30 to -40) dB	2.28 dB	
	Dev > 2.5 rad		
	(0 to -20) dB	0.09 dB	
	(-20 to -30) dB	0.27 dB	
	(-30 to -40) dB	0.83 dB	
	(-40 to -50) dB	2.3 dB	
	(13.2 to 31.15) GHz		
	Rate 20 to 500 Hz		
	Dev > 3.5 rad		
	(0 to -20) dB	0.26 dB	
	(-20 to -30) dB	0.79 dB	
	(-30 to -40) dB	2.3 dB	
	Dev > 10.0 rad		
	(0 to -20) dB	0.09 dB	
	(-20 to -30) dB	0.27 dB	
	(-30 to -40) dB	0.83 dB	
(-40 to -50) dB	2.3 dB		
Rate 500 Hz to 1000 Hz			
Dev > 1.2 rad			
(0 to -20) dB	0.26 dB		
(-20 to -30) dB	0.79 dB		
(-30 to -40) dB	2.3 dB		
Dev > 4.0 rad			
(0 to -20) dB	0.09 dB		
(-20 to -30) dB	0.27 dB		
(-30 to -40) dB	0.83 dB		
(-40 to -50) dB	2.3 dB		
Agilent N5531S Measuring Receiver			



ANSI National Accreditation Board

Electrical – RF/Microwave

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
PM Distortion Measure <sup>1</sup>	(31.15 to 50.0) GHz		Agilent N5531S Measuring Receiver
	Rate 20 to 500 Hz		
	Dev > 7.5 rad	0.26 dB	
	(0 to -20) dB	0.79 dB	
	(-20 to -30) dB	2.3 dB	
	(-30 to -40) dB		
	Dev > 19.0 rad	0.09 dB	
	(0 to -20) dB	0.27 dB	
	(-20 to -30) dB	0.83 dB	
	(-30 to -40) dB	2.3 dB	
	(-40 to -50) dB		
	Rate 500 Hz to 1000 Hz		
	Dev > 3.0 rad	0.26 dB	
	(0 to -20) dB	0.79 dB	
(-20 to -30) dB	2.3 dB		
(-30 to -40) dB			
Dev > 8.0 rad	0.09 dB		
(0 to -20) dB	0.27 dB		
(-20 to -30) dB	0.83 dB		
(-30 to -40) dB	2.3 dB		
(-40 to -50) dB			
Total Harmonic Distortion (THD)	(0 to -60) dB		HP 8903B Audio Analyzer
	20 Hz to 20 kHz		
	(0 to -40) dB	1 dB	
	(-40 to -50) dB	1 dB	
	(-50 to -60) dB	1.3 dB	
	(-60 to -65) dB	1.7 dB	
	(20 to 50) kHz		
	(0 to -40) dB	2 dB	
	(-40 to -50) dB	2.1 dB	
	(-50 to -60) dB	3 dB	
(50 to 100) kHz			
(0 to -40) dB	2 dB		
(-40 to -50) dB	2.4 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Harmonics Measure <sup>1</sup>	(-80 to -10) dB 2 <sup>nd</sup> through 5 <sup>th</sup> Harmonic 1 kHz to 600 MHz (600 to 1 320) MHz (1 320 to 2 200) MHz (2 200 to 3 000) MHz (3 000 to 4 400) MHz (4 400 to 5 300) MHz (5 300 to 10 000) MHz 2 <sup>nd</sup> through 4 <sup>th</sup> Harmonic (10 000 to 12 500) MHz 2 <sup>nd</sup> through 3 <sup>rd</sup> Harmonic (12 500 to 16 667) MHz 2 <sup>nd</sup> Harmonic (16 667 to 25 000) MHz	0.37 dB 1.1 dB 1.4 dB 1.4 dB 1.7 dB 1.9 dB 2.1 dB 2.1 dB 2.1 dB 2.3 dB	Agilent E4440A Measuring Receiver
Phase Noise – SSB Measure <sup>1</sup>	1 MHz to 26.5 GHz	0.45 dB	Agilent E4440A Spectrum Analyzer

Length – Dimensional Metrology

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Micrometers <sup>2,3</sup>	Up to 46 in	(29 + 4.5L) μin	Gage blocks (Grade 0)
Gage Blocks <sup>2</sup>	Up to 13 in	(1.2+ 2.5L) μin	Master gage blocks, P&W universal measuring machine
Bore Micrometers <sup>2</sup> 2 point 3 point	Up to 8 in	(7 + 2L) μin (42 + 1.4L) μin	Master gage blocks, P&W universal measuring machine, Master Ring
Calipers <sup>1,2</sup>	Up to 46 in	(280 + 1.7L) μin	Gage blocks (Grade 0)
Dial Indicators <sup>1,2</sup> Resolution ≥ 50μin Resolution < 50μin	Up to 10 in Up to 0.1 in	(25 + 3.6L) μin 7 μin	716 Starrett Calibrator, gage blocks (Grade 0)
Micrometer Rods	Up to 40 in	(1.1 + 5.7L) μin	Gage blocks (grade 0)
Height Gages <sup>1,2</sup>	Up to 40 in	(95 + 3L) μin	Gage blocks (grade 0)

**Length – Dimensional Metrology**

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Roughness Measure	123 $\mu\text{in Ra}$ 117 $\mu\text{in Ra}$ 116 $\mu\text{in Ra}$ 3 $\mu\text{m Ra}$	3.2 $\mu\text{in}$ 2.5 $\mu\text{in}$ 2.9 $\mu\text{in}$ 0.06 $\mu\text{m}$	Mitutoyo SV-3200H4 Surface Measuring System and Roughness Specimens
Cylindrical Gages <sup>2</sup>			
Ring Gages	(0.04 to 14) in	(9.6 + 2.1D) $\mu\text{in}$	Master gage blocks, P&W universal measuring machine
Plain Plugs/Pins	Up to 13 in	(4.2 + 3.4D) $\mu\text{in}$	
Thread Wires	Up to 0.5D	10 $\mu\text{in}$	Master gage blocks, P&W universal measuring machine
Thread Plugs –			
Major Diameter	Up to 12 in	44 $\mu\text{in}$	P & W Model C Bench Micrometer, Van Keuren thread wire set
Pitch Diameter	Up to 12 in	79 $\mu\text{in}$	
Solid Thread Rings	(0.625 to 12) in	106 $\mu\text{in}$	Master gage blocks, P&W universal measuring machine
Adjustable Thread Rings <sup>2,3</sup> Pitch Diameter (Tactile Fit)	Up to 12 in	(350 + 47D) $\mu\text{in}$	Thread setting plug gages
Surface Plates <sup>1,2</sup>			
Overall Flatness	Up to 6 ft $\times$ 6 ft	77 $\mu\text{in}$	Planekator
Local Area Flatness	$\pm$ 0.001 in	68 $\mu\text{in}$	Repeat-O-Meter
Parallelism & Straightness	(0 to 0.01) in	13 $\mu\text{in}$	Gage Amplifier, Surface Plate
Protractors <sup>1</sup>	(0 to 360) $^{\circ}$	0.075 $^{\circ}$	Angle Blocks
Radius Gages	(0.016 to 1.00) in	240 $\mu\text{in}$	Optical Comparator
Rulers <sup>1</sup>	Up to 72 in	0.009 in	Gage blocks (grade 0)
Tape Measures <sup>1</sup>	Up to 100 ft	(0.000 27F + 0.024) in	Standard rule
Feeler Gage <sup>1</sup>	Up to 1 in	31 $\mu\text{in}$	Pratt & Whitney Model C Bench Micrometer



ANSI National Accreditation Board

Length – Dimensional Metrology

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Optical Comparators <sup>1</sup> Angle Linearity  Magnification	(0 to 360) ° Up to 20 in (20 to 40) in  (10 to 100) x	0.042 ° 320 μin 630 μin  420 μin	Gage blocks, Angle blocks, SI Industries glass scales

Mass and Mass Related

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Scales & Balances <sup>1,2,4</sup>	1 mg to 5 000 g (0.001 to 100) lb	(0.049 + 0.003 <i>X</i> ) mg (3.2 E <sup>-6</sup> + 3.1 E <sup>-6</sup> <i>W</i> ) lb	Class 1 weights
Scales & Balances <sup>1,2,4</sup>	Up to 1 000 lb	(0.00 02 + 0.000 12 <i>W</i> ) lb	Class F weights
Mass - Measure	(1 to 200) g (200 to 2 500) g (1 to 34) kg	0.69 mg 0.026 g 0.25 g	Precision Balance
Mass - Fixed Points Metric	200 g 500 g 1 kg 5 kg	0.73 mg 29 mg 30 mg 30 mg	Comparison to ASTM E617 Class 1 weights
Pressure – Generate <sup>1</sup>	(10 to 16 000) psi	0.019 % of reading	Dead Weight Tester Fluke P3125
Pressure – Measure <sup>1</sup>	(-1 to 1) inH <sub>2</sub> O (-5 to 5) inH <sub>2</sub> O (-50 to 50) inH <sub>2</sub> O (-15 to 0) psi (0 to 15) psi (0 to 30) psi (0 to 100) psi (0 to 300) psi (0 to 1 000) psi (0 to 3 000) psi (0 to 10 000) psi	0.001 5 inH <sub>2</sub> O 0.008 inH <sub>2</sub> O 0.061 inH <sub>2</sub> O 0.008 9 psi 0.004 psi 0.007 8 psi 0.028 psi 0.064 psi 0.23 psi 0.97 psi 3.3 psi	Pressure Gauges ADT681-05-DP1-inH <sub>2</sub> O ADT681-05-DP5-inH <sub>2</sub> O ADT681-05-DP50-inH <sub>2</sub> O Fluke 2700G-BG100K Fluke 2700G-BG100K Additel ADT681-GP30 Fluke 2700G-BG700K Fluke 2700G-BG2M Fluke 2700G-BG7M Additel ADT681-GP3K Additel ADT681-GP10K
Force <sup>1</sup> Tension and Compression	(0.01 to 500) lbf	0.063 % of reading	Class F weights



ANSI National Accreditation Board

Mass and Mass Related

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Force <sup>1</sup> Tension and Compression	(50 to 2 000) lbf (200 to 10 000) lbf (500 to 25 000) lbf	0.068 % of reading 0.044 % of reading 0.087 % of reading	Morehouse Press with Load Cells
Torque Tools <sup>1</sup>	(10 to 100) ozf-in 4 lbf-in to 1 000 lbf-ft	0.59 % of reading 0.32 % of reading	CDI 1001 Torque Tester CDI 5000 ST torque tester
Torque Tools <sup>1</sup>	(0.5 to 2.5) ozf-in (2 to 10) ozf-in (6 to 43) ozf-in (30 to 215) ozf-in	0.18 % of reading 0.18 % of reading 0.18 % of reading 0.18 % of reading	Waters Torque Analyzer
Torque Analyzers	0.4 ozf-in to 1 000 lbf-ft	0.065 % of reading	Class F weights and arm
Durometers Scale (Force) Accuracy Types A, B, E, O,C, D, DO Type M Types OO, OOO	(0 to 100) duros	0.06 duros 0.07 duros 0.08 duros	Direct Verification  Master balance
Indenter Geometry Length Diameter Angle	0.1 in 0.05 in (30 to 35) °	130 μin 130 μin 0.085 °	Optical comparator
Rockwell Hardness Testers <sup>1</sup>	(< 60) HRBW (≥ 60 to < 80) HRBW (≥ 80) HRBW  (< 35) HRC (≥ 35 to < 60) HRC (≥ 60) HRC  (< 81) HR15TW (≥ 81 to < 87) HR15TW (> 87) HR15TW	3 HRBW 3 HRBW 1.4 HRBW  1.3 HRC 1.3 HRC 0.73 HRC  1.8 HR15TW 1.3 HR15TW 1.3 HR15TW	Indirect verification per ASTM E18

Thermodynamic

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Relative Humidity – Generate	(10 to 95) %RH	0.5 %RH	Thunder Scientific 1200 Humidity Chamber

**Thermodynamic**

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Relative Humidity – Measure	(0 to 90) %RH (90 to 100) %RH	1.2 %RH 2 %RH	Vaisala MI70/HMP76 Humidity Indicator and Probe
Temperature – Measuring Equipment <sup>1</sup>	(-25 to 350) °C	0.087 °C	Hart 1502A Indicator with Burnes Engineering 12005 PRT and dry block
Temperature – Measure <sup>1</sup>	(-200 to 420) °C	0.036 °C	Hart 1502A Indicator with Burnes Engineering 12005 PRT
Infrared (IR) Thermometers	(20 to 100) °C (100 to 300) °C (300 to 420) °C (420 to 500) °C	1.5 °C 4.3 °C 6 °C 7.7 °C	Fluke 9132 IR Calibrator $\epsilon = 0.95, \lambda = (8 \text{ to } 14) \mu\text{m}$

**Time and Frequency**

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency – Generate <sup>1</sup>	1 to 10 Hz 10 to 100 Hz 100 to 1 000 Hz 1 to 10 kHz 10 to 100 kHz 0.1 to 1 MHz 1 to 10 MHz	$1 \times 10^{-12}$ Hz/Hz + 0.57 $\mu$ Hz $1 \times 10^{-12}$ Hz/Hz + 5.7 $\mu$ Hz $1 \times 10^{-12}$ Hz/Hz + 57 $\mu$ Hz $1 \times 10^{-12}$ Hz/Hz + 0.57 mHz $1 \times 10^{-12}$ Hz/Hz + 5.7 mHz $1 \times 10^{-12}$ Hz/Hz + 57 mHz $1 \times 10^{-12}$ Hz/Hz + 0.57 Hz	Agilent 33250A Function Generator / HP 58503A GPS Receiver
Frequency – Generate <sup>1</sup>	10 to 40 000 MHz	$1 \times 10^{-12}$ Hz/Hz + 0.57 mHz	Agilent E8257D Opt 540 Signal Generator / HP 58503A GPS Receiver
Frequency – Generate <sup>1</sup>	10 to 50 000 MHz	$1 \times 10^{-12}$ Hz/Hz + 0.57 kHz	HP 83650B Signal Generator / HP 58503A GPS Receiver
Frequency – Measure <sup>1</sup>	1 to 10 Hz 10 to 100 Hz 100 to 1 000 Hz 1 to 10 kHz 10 to 100 kHz 100 to 200 kHz 0.2 to 3 000 MHz	$4.20 \times 10^{-9}$ Hz/Hz $1.47 \times 10^{-9}$ Hz/Hz $0.60 \times 10^{-9}$ Hz/Hz $0.33 \times 10^{-9}$ Hz/Hz $0.24 \times 10^{-9}$ Hz/Hz $0.21 \times 10^{-9}$ Hz/Hz $0.21 \times 10^{-9}$ Hz/Hz	Agilent 53131A Opt 030 Frequency Counter / HP 58503A GPS Receiver

**Time and Frequency**

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency – Measure <sup>1</sup>	10 to 50 000 MHz	$1 \times 10^{-12}$ Hz/Hz + 0.1 Hz	Agilent E4448A Spectrum Analyzer / HP 58503A GPS Receiver
Time – Generate	1 pps	$1 \times 10^{-12}$ s/s + 750 ps	HP 58503A GPS Receiver
Type I (digital) Timers	(0 to 19.99) sec/day (0 to 599) sec/month	0.031 sec/day 1.1 sec/month	Helmut Klein Timometer 4500
Type II (mechanical) Timers	(0 to 320) sec/day	0.6 sec/day	Helmut Klein Timometer 4500
Timer, Stopwatch	10 s to 24 hr	34 ms	Totalize method with counter
Tachometers – RPM <sup>1</sup>	Up to 100 000 RPM	0.001 % of reading + 0.6R	HP 33250A Signal Generator & LED

**DIMENSIONAL MEASUREMENT**

**1 Dimensional**

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Length	X Axis (0.01 to 10) in Y Axis (0.01 to 6) in	162 $\mu$ m 123 $\mu$ m	Optical comparator

**2 Dimensional**

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Angle	Up to 360°	0.004°	Optical comparator



ANSI National Accreditation Board


3 Dimensional

Rockford, IL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
General Inspection	X axis (0 to 33) in (0 to 838) mm Y axis (0 to 60) in (0 to 1 524) mm Z axis (0 to 20) in (0 to 508) mm	(80+ 3.6L) $\mu$ m (2.1 + 0.003 6L) $\mu$ m (200 + 0.13L) $\mu$ m (5.1 + 0.000 13L) $\mu$ m (200 + 0.13L) $\mu$ m (5.1 + 0.000 13L) $\mu$ m	CMM

[Return to Site listing](#) (top)

[Go to Notes](#) (bottom)





ANSI National Accreditation Board

### Services performed at satellite laboratory

#### Technical Maintenance, Inc.

425 Hayden Station Road, Suite B  
Windsor, CT 06095

Matt Nicewicz (Branch Manager) Phone: 860-219-0046  
Scott Chamberlain (Quality Manager) Phone: 321-242-0890

### CALIBRATION

#### Acoustics and Vibration

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Accelerometers – Acceleration	(0.01 to 10) g (7 < 10) Hz (10 < 30) Hz (30 < 2 000) Hz (2 to 10) kHz	4 % of reading 3 % of reading 1.5 % of reading 4 % of reading	Accelerometer Calibrator

#### Chemical Quantities

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
pH Meters <sup>1,5</sup>	4 pH 7 pH 10 pH	0.035 pH 0.025 pH 0.069 pH	pH buffer solutions

#### Electrical – DC/Low Frequency

Windsor, CT

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Voltage – Generate <sup>1</sup>	(0 to 220) mV 220 mV to 2.2 V (2.2 to 11) V (11 to 22) V (22 to 220) V (220 to 1 000) V	6.8 μV/V + 0.9 μV 6.1 μV/V + 0.9 μV 6.1 μV/V + 3.2 μV 6.1 μV/V + 6.2 μV 6.8 μV/V + 76 μV 8.4 μV/V + 460 μV	Fluke 5700A Multifunction Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Windsor, CT

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Voltage – Measure <sup>1</sup>	(0 to 100) mV 100 mV to 1 V (1 to 10) V (10 to 100) V (100 to 1 000) V	5.0 $\mu\text{V/V} + 1.2 \mu\text{V}$ 4.0 $\mu\text{V/V} + 1.2 \mu\text{V}$ 4.0 $\mu\text{V/V} + 2.1 \mu\text{V}$ 6.0 $\mu\text{V/V} + 30 \mu\text{V}$ 6.0 $\mu\text{V/V} + 100 \mu\text{V} + 12 \mu\text{V/V} \times (\text{Vin}/1\ 000) ^2$	Agilent 3458A Option 002 Multimeter
DC Voltage – Measure <sup>1</sup>	(1 to 60) kV	0.1 % of reading	Ross VD60 High Voltage Divider, HP 34401A Multimeter
DC Current – Generate <sup>1</sup>	(1 to 2.2) nA (2.2 to 22) nA (22 to 220) nA (0.22 to 2.2) $\mu\text{A}$ (2.2 to 10) $\mu\text{A}$	100 $\mu\text{A/A} + 0.007 \text{ nA}$ 99 $\mu\text{A/A} + 0.007 \text{ nA}$ 99 $\mu\text{A/A} + 0.01 \text{ nA}$ 36 $\mu\text{A/A} + 0.1 \text{ nA}$ 20 $\mu\text{A/A} + 1 \text{ nA}$	Fluke 5700A & 5522A Multifunction Calibrator
DC Current – Generate <sup>1</sup>	(10 to 220) $\mu\text{A}$ (0.22 to 2.2) mA (2.2 to 22) mA (22 to 100) mA (100 to 220) mA (0.22 to 1) A (1 to 2.2) A	46 $\mu\text{A/A} + 8 \text{ nA}$ 46 $\mu\text{A/A} + 8 \text{ nA}$ 46 $\mu\text{A/A} + 82 \text{ nA}$ 53 $\mu\text{A/A} + 0.9 \mu\text{A}$ 61 $\mu\text{A/A} + 0.9 \mu\text{A}$ 72 $\mu\text{A/A} + 23 \mu\text{A}$ 109 $\mu\text{A/A} + 23 \mu\text{A}$	Fluke 5700A Multifunction Calibrator
DC Current – Generate <sup>1</sup>	(2.2 to 11) A	274 $\mu\text{A/A} + 371 \mu\text{A}$	Fluke 5700A Multifunction Calibrator /5725A Amplifier
DC Current – Generate <sup>1</sup>	(11 to 20.5) A	761 $\mu\text{A/A} + 631 \mu\text{A}$	Fluke 5522A Multiproduct Calibrator
DC Current – Generate <sup>1</sup> Clamp Meters	(0 to 200) A	0.21 % of output + 0.028 A	Fluke 5522A Multiproduct Calibrator / 9100-200 x10 Coil
DC Current – Generate <sup>1</sup> Clamp Meters	(0 to 1 000) A	0.21 % of output + 0.04 A	Fluke 5522A Multiproduct Calibrator / 9100-200 x50 Coil
DC Current – Measure	(1 to 10) nA (10 to 100) nA (0.1 to 1) $\mu\text{A}$ (1 to 10) $\mu\text{A}$	35 $\mu\text{A/A} + 0.12 \text{ pA}$ 18 $\mu\text{A/A} + 1.2 \text{ pA}$ 13 $\mu\text{A/A} + 0.01 \text{ nA}$ 12 $\mu\text{A/A} + 0.12 \text{ nA}$	Fluke 5700A/03 Multifunction Calibrator Agilent 3458A Option 002 Multimeter
DC Current – Measure	(10 to 100) $\mu\text{A}$ (0.1 to 1) mA (1 to 10) mA (10 to 100) mA (0.1 to 1) A	20 $\mu\text{A/A} + 0.8 \text{ nA}$ 20 $\mu\text{A/A} + 5 \text{ nA}$ 20 $\mu\text{A/A} + 0.05 \mu\text{A}$ 35 $\mu\text{A/A} + 0.5 \mu\text{A}$ 110 $\mu\text{A/A} + 10 \mu\text{A}$	Agilent 3458A Multimeter



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Windsor, CT

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Current – Measure	(1 to 3) A (3 to 10) A	761 $\mu$ A/A + 462 $\mu$ A 1.1 mA/A + 614 $\mu$ A	Fluke 8845A Multimeter
DC Current – Measure	(10 to 100) A	0.25 % of reading	Agilent 3458A Option 002 Multimeter Empro Current Shunt
AC Voltage – Generate <sup>1</sup>	(0.22 to 2.2) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz (2.2 to 22) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz (22 to 220) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	457 $\mu$ V/V + 3.9 $\mu$ V 183 $\mu$ V/V + 3.9 $\mu$ V 91 $\mu$ V/V + 3.9 $\mu$ V 312 $\mu$ V/V + 3.9 $\mu$ V 723 $\mu$ V/V + 6.1 $\mu$ V 989 $\mu$ V/V + 11.4 $\mu$ V 1.4 mV/V + 23 $\mu$ V 3.7 mV/V + 30 $\mu$ V 457 $\mu$ V/V + 4.6 $\mu$ V 183 $\mu$ V/V + 4.6 $\mu$ V 91 $\mu$ V/V + 4.6 $\mu$ V 312 $\mu$ V/V + 4.6 $\mu$ V 723 $\mu$ V/V + 6.1 $\mu$ V 989 $\mu$ V/V + 11.4 $\mu$ V 1.4 mV/V + 23 $\mu$ V 3.7 mV/V + 30 $\mu$ V 457 $\mu$ V/V + 12 $\mu$ V 183 $\mu$ V/V + 7.6 $\mu$ V 84 $\mu$ V/V + 7.6 $\mu$ V 274 $\mu$ V/V + 7.6 $\mu$ V 685 $\mu$ V/V + 23 $\mu$ V 837 $\mu$ V/V + 23 $\mu$ V 1.4 mV/V + 30 $\mu$ V 2.7 mV/V + 76 $\mu$ V	Fluke 5700A/03 Multifunction Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Windsor, CT

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Generate <sup>1</sup>	220 mV to 2.2 V		Fluke 5700A/03 Multifunction Calibrator
	(10 to 20) Hz	457 $\mu$ V/V + 8 $\mu$ V	
	(20 to 40) Hz	137 $\mu$ V/V + 23 $\mu$ V	
	40 Hz to 20 kHz	65 $\mu$ V/V + 5 $\mu$ V	
	(20 to 50) kHz	107 $\mu$ V/V + 15 $\mu$ V	
	(50 to 100) kHz	213 $\mu$ V/V + 61 $\mu$ V	
	(100 to 300) kHz	365 $\mu$ V/V + 114 $\mu$ V	
	(300 to 500) kHz	913 $\mu$ V /V + 304 $\mu$ V	
	500 kHz to 1 MHz	1.8 mV/V + 761 $\mu$ V	
	(2.2 to 22) V		
	(10 to 20) Hz	457 $\mu$ V/V + 761 $\mu$ V	
	(20 to 40) Hz	137 $\mu$ V/V + 228 $\mu$ V	
	40 Hz to 20 kHz	65 $\mu$ V/V + 54 $\mu$ V	
	(20 to 50) kHz	107 $\mu$ V/V + 152 $\mu$ V	
	(50 to 100) kHz	213 $\mu$ V/V + 304 $\mu$ V	
(100 to 300) kHz	457 $\mu$ V/V + 1294 $\mu$ V		
(300 to 500) kHz	1.1 mV/V + 3805 $\mu$ V		
500 kHz to 1 MHz	2.3 mV/V + 6848 $\mu$ V		
AC Voltage – Generate <sup>1</sup>	(22 to 220) V		Fluke 5700A/03 Multifunction Calibrator, 5725A Amplifier
	(10 to 20) Hz	457 $\mu$ V/V + 7.6 mV	
	(20 to 40) Hz	137 $\mu$ V/V + 2.4 mV	
	40 Hz to 20 kHz	68 $\mu$ V/V + 1.0 mV	
	(20 to 50) kHz	190 $\mu$ V/V + 3.1 mV	
	(50 to 100) kHz	457 $\mu$ V/V + 7.6 mV	
	(220 to 750) V		
40 Hz to 1 kHz	68 $\mu$ V/V + 3 mV		
(1 to 20) kHz	126 $\mu$ V/V + 5 mV		
(20 to 50) kHz	457 $\mu$ V/V + 9 mV		
(50 to 100) kHz	1.8 mV/V + 34 mV		
AC Voltage – Generate <sup>1</sup> Wideband Absolute	(750 to 1 000) V		Fluke 5700A Option 003 Multifunction Calibrator
	40 Hz to 1 kHz	68 $\mu$ V/V + 3 mV	
	(1 to 20) kHz	126 $\mu$ V/V + 5 mV	
	(20 to 30) kHz	457 $\mu$ V/V + 8 mV	
	(0.3 to 1.1) mV		
(10 to 30 Hz)	0.65 % of output + 1.5 $\mu$ V		
30 to 500 kHz	0.61% of output + 1.5 $\mu$ V		
(0.5 to 1.2) MHz	0.63 % of output + 3.8 $\mu$ V		
(1.2 to 2) MHz	0.63 % of output + 3.8 $\mu$ V		
(2 to 12) MHz	0.68 % of output + 3.8 $\mu$ V		
(12 to 20) MHz	0.76 % of output + 3.8 $\mu$ V		
(20 to 30) MHz	1.3 % of output + 13 $\mu$ V		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Windsor, CT

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Generate <sup>1</sup> Wideband Absolute	(1.1 to 3.3) mV		Fluke 5700A Option 003 Multifunction Calibrator
	(10 to 30 Hz)	0.58 % of output + 2.3 $\mu$ V	
	30 Hz to 500 kHz	0.53 % of output + 2.3 $\mu$ V	
	(0.5 to 1.2) MHz	0.54 % of output + 4.6 $\mu$ V	
	(1.2 to 2) MHz	0.54 % of output + 4.6 $\mu$ V	
	(2 to 12) MHz	0.58 % of output + 4.6 $\mu$ V	
	(12 to 20) MHz	0.65 % of output + 4.6 $\mu$ V	
	(20 to 30) MHz	1.3 % of output + 4.6 $\mu$ V	
	(3.3 to 11) mV		
	(10 to 30 Hz)	0.58 % of output + 6.1 $\mu$ V	
	30 Hz to 500 kHz	0.53 % of output + 6.1 $\mu$ V	
	(0.5 to 1.2) MHz	0.54 % of output + 8.4 $\mu$ V	
	(1.2 to 2) MHz	0.54 % of output + 8.4 $\mu$ V	
	(2 to 12) MHz	0.55 % of output + 8.4 $\mu$ V	
	(12 to 20) MHz	0.61 % of output + 8.4 $\mu$ V	
	(20 to 30) MHz	0.93 % of output + 8.4 $\mu$ V	
	(11 to 33) mV		
	(10 to 30 Hz)	0.52 % of output + 12 $\mu$ V	
	30 Hz to 500 kHz	0.46 % of output + 12 $\mu$ V	
	(0.5 to 1.2) MHz	0.47 % of output + 14 $\mu$ V	
	(1.2 to 2) MHz	0.47 % of output + 14 $\mu$ V	
	(2 to 12) MHz	0.49 % of output + 14 $\mu$ V	
	(12 to 20) MHz	0.55 % of output + 14 $\mu$ V	
	(20 to 30) MHz	0.89 % of output + 14 $\mu$ V	
	(33 to 110) mV		
	(10 to 30 Hz)	0.52 % of output + 30 $\mu$ V	
	30 Hz to 500 kHz	0.46 % of output + 30 $\mu$ V	
	(0.5 to 1.2) MHz	0.47 % of output + 33 $\mu$ V	
	(1.2 to 2) MHz	0.47 % of output + 33 $\mu$ V	
	(2 to 12) MHz	0.49 % of output + 33 $\mu$ V	
(12 to 20) MHz	0.55 % of output + 33 $\mu$ V		
(20 to 30) MHz	0.89 % of output + 33 $\mu$ V		
(110 to 330) mV			
(10 to 30 Hz)	0.45 % of output + 0.1 mV		
30 Hz to 500 kHz	0.38 % of output + 0.1 mV		
(0.5 to 1.2) MHz	0.4 % of output + 0.1 mV		
(1.2 to 2) MHz	0.4 % of output + 0.1 mV		
(2 to 12) MHz	0.42 % of output + 0.1 mV		
(12 to 20) MHz	0.49 % of output + 0.1 mV		
(20 to 30) MHz	0.85 % of output + 0.1 mV		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Windsor, CT

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Generate <sup>1</sup> Wideband Absolute	(0.33 to 1.1) V		Fluke 5700A Multifunction Calibrator Option 003
	(10 to 30 Hz)	0.45 % of output + 0.3 mV	
	30 Hz to 500 kHz	0.38 % of output + 0.3 mV	
	(0.5 to 1.2) MHz	0.4 % of output + 0.3 mV	
	(1.2 to 2) MHz	0.4 % of output + 0.3 mV	
	(2 to 12) MHz	0.42 % of output + 0.3 mV	
	(12 to 20) MHz	0.49 % of output + 0.3 mV	
	(20 to 30) MHz	0.85 % of output + 0.3 mV	
	(1.1 to 3.5) V		
	(10 to 30 Hz)	0.39 % of output + 0.4 mV	
	30 Hz to 500 kHz	0.3 % of output + 0.4 mV	
	(0.5 to 1.2) MHz	0.32 % of output + 0.4 mV	
	(1.2 to 2) MHz	0.32 % of output + 0.4 mV	
	(2 to 12) MHz	0.35 % of output + 0.4 mV	
(12 to 20) MHz	0.44 % of output + 0.4 mV		
(20 to 30) MHz	0.82 % of output + 0.4 mV		
AC Voltage – Measure <sup>1</sup>	(1 to 10) mV		Agilent 3458A Multimeter
	(1 to 40) Hz	0.03 % of reading + 3 μV	
	40 Hz to 1 kHz	0.02 % of reading + 1.1 μV	
	(1 to 20) kHz	0.03 % of reading + 1.1 μV	
	(20 to 50) kHz	0.1 % of reading + 1.1 μV	
	(50 to 100) kHz	0.5 % of reading + 1.1 μV	
	100 kHz to 1 MHz	1.2 % of reading + 5 μV	
	(1 to 4) MHz	7 % of reading + 7 μV	
	(4 to 8) MHz	20 % of reading + 8 μV	
	(10 to 100) mV		
	(1 to 40) Hz	0.007 % of reading + 4 μV	
	40 Hz to 1 kHz	0.007 % of reading + 2 μV	
	(1 to 20) kHz	0.014 % of reading + 2 μV	
	(20 to 50) kHz	0.03 % of reading + 2 μV	
	(50 to 100) kHz	0.08 % of reading + 2 μV	
	(100 to 300) kHz	0.3 % of reading + 10 μV	
	300 kHz to 1 MHz	1 % of reading + 10 μV	
	(1 to 2) MHz	1.5 % of reading + 70 μV	
(2 to 4) MHz	4 % of reading + 70 μV		
(4 to 8) MHz	4 % of reading + 80 μV		
(8 to 10) MHz	15 % of reading + 100 μV		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Windsor, CT

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure <sup>1</sup>	100 mV to 1 V		Agilent 3458A Multimeter
	(1 to 40) Hz	0.007 % of reading + 40 μV	
	40 Hz to 1 kHz	0.07 % of reading + 20 μV	
	(1 to 20) kHz	0.014 % of reading + 20 μV	
	(20 to 50) kHz	0.03 % of reading + 20 μV	
	(50 to 100) kHz	0.08 % of reading + 20 μV	
	(100 to 300) kHz	0.3 % of reading + 100 μV	
	300 kHz to 1 MHz	1 % of reading + 100 μV	
	(1 to 2) MHz	1.5 % of reading + 0.7 mV	
	(2 to 4) MHz	4 % of reading + 0.7 mV	
	(4 to 8) MHz	4 % of reading + 0.8 mV	
	(8 to 10) MHz	15 % of reading + 1 mV	
	(1 to 10) V		
	(1 to 40) Hz	0.007 % of reading + 0.4 mV	
	40 Hz to 1 kHz	0.007 % of reading + 0.2 mV	
	(1 to 20) kHz	0.014 % of reading + 0.2 mV	
	(20 to 50) kHz	0.03 % of reading + 0.2 mV	
	(50 to 100) kHz	0.08 % of reading + 0.2 mV	
	(100 to 300) kHz	0.3 % of reading + 1 mV	
	300 kHz to 1 MHz	1 % of reading + 1 mV	
(1 to 2) MHz	1.5 % of reading + 7 mV		
(2 to 4) MHz	4 % of reading + 7 mV		
(4 to 8) MHz	4 % of reading + 8 mV		
(8 to 10) MHz	15 % of reading + 10 mV		
(10 to 100) V	(1 to 40) Hz	0.02 % of reading + 4 mV	
	40 Hz to 1 kHz	0.02 % of reading + 2 mV	
	(1 to 20) kHz	0.02 % of reading + 2 mV	
	(20 to 50) kHz	0.035 % of reading + 2 mV	
	(50 to 100) kHz	0.12 % of reading + 2 mV	
	(100 to 300) kHz	0.4 % of reading + 10 mV	
	300 kHz to 1 MHz	1.5 % of reading + 10 mV	
	(100 to 700) V	(1 to 40) Hz	0.04 % of reading + 40 mV
		40 Hz to 1 kHz	0.04 % of reading + 20 mV
		(1 to 20) kHz	0.06 % of reading + 20 mV
		(20 to 50) kHz	0.12 % of reading + 20 mV
		(50 to 100) kHz	0.3 % of reading + 20 mV
AC Voltage – Measure <sup>1</sup>	(60) Hz		Ross VD60 High Voltage Divider
	(1 to 10) kV	0.5 % of reading + 0.002 kV	
	(10 to 42) kV	0.5 % of reading + 0.02 kV	

Electrical – DC/Low Frequency

Windsor, CT

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Generate <sup>1</sup>	(9 to 220) $\mu$ A		Fluke 5700A/03 Multifunction Calibrator
	(10 to 20) Hz	228 $\mu$ A/A + 15 nA	
	(20 to 40) Hz	152 $\mu$ A/A + 10 nA	
	40 Hz to 1 kHz	107 $\mu$ A/A + 8 nA	
	(1 to 5) kHz	266 $\mu$ A/A + 12 nA	
	(5 to 10) kHz	989 $\mu$ A/A + 61 nA	
	(0.22 to 2.2) mA		
	(10 to 20) Hz	228 $\mu$ A/A + 39 nA	
	(20 to 40) Hz	152 $\mu$ A/A + 31 nA	
	40 Hz to 1 kHz	107 $\mu$ A/A + 31 nA	
	(1 to 5) kHz	183 $\mu$ A/A + 99 nA	
	(5 to 10) kHz	989 $\mu$ A/A + 609 nA	
	(2.2 to 22) mA		
	(10 to 20) Hz	228 $\mu$ A/A + 385 nA	
	(20 to 40) Hz	152 $\mu$ A/A + 310 nA	
	40 Hz to 1 kHz	107 $\mu$ A/A + 310 nA	
(1 to 5) kHz	183 $\mu$ A/A + 536 nA		
(5 to 10) kHz	989 $\mu$ A/A + 4.6 $\mu$ A		
(22 to 220) mA			
(10 to 20) Hz	228 $\mu$ A/A + 4 $\mu$ A		
(20 to 40) Hz	152 $\mu$ A/A + 3 $\mu$ A		
40 Hz to 1 kHz	107 $\mu$ A/A + 2 $\mu$ A		
(1 to 5) kHz	183 $\mu$ A/A + 3 $\mu$ A		
AC Current – Generate <sup>1</sup>	(22 to 220) mA		Fluke 5700A/03 Multifunction Calibrator
	(5 to 10) kHz	989 $\mu$ A/A + 9 $\mu$ A	
	(0.22 to 2.2) A		
	20 Hz to 1 kHz	243 $\mu$ A/A + 31 $\mu$ A	
AC Current – Generate <sup>1</sup>	(1 to 5) kHz	380 $\mu$ A/A + 76 $\mu$ A	Fluke 5700A/03 Multifunction Calibrator / 5725A Amplifier
	(5 to 10) kHz	6.1 mA/A + 152 $\mu$ A	
	(2.2 to 11) A		
	40 Hz to 1 kHz	350 $\mu$ A /A + 141 $\mu$ A	
AC Current – Generate <sup>1</sup>	(1 to 5) kHz	723 $\mu$ A /A + 295 $\mu$ A	Fluke 5522A Multiproduct Calibrator
	(5 to 10) kHz	2.8 mA /A + 573 $\mu$ A	
	(11 to 20.5) A		
	(45 to 100) Hz	0.09 % of reading + 5 mA	
AC Current – Generate <sup>1</sup>	100 Hz to 1 kHz	0.11 % of reading + 5 mA	
	(1 to 5) kHz	2.28 % of reading + 5 mA	



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Windsor, CT

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Generate <sup>1</sup> Clamp Meters	(3.3 to 30) A (10 to 100) Hz (100 to 440) Hz (30 to 200) A 10 to 100 Hz (100 to 440) Hz	0.22 % of output + 0.028 A 0.30 % of output + 0.07 A 0.22 % of output + 0.032 A 0.79 % of output + 0.08 A	Fluke 5522A Multiproduct Calibrator / 9100-200 x10 Coil
AC Current – Generate <sup>1</sup> Clamp Meters	(16.5 to 150) A (10 to 100) Hz (100 to 440) Hz (150 to 1 000) A (10 to 100) Hz (100 to 440) Hz	0.22 % of output + 0.029 A 0.30 % of output + 0.08 A 0.22 % of output + 0.081 A 0.79 % of output + 0.20 A	Fluke 5522A Multiproduct Calibrator / 9100-200 x50 Coil
AC Current – Measure <sup>1</sup>	(5 to 100) $\mu$ A (10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz (0.1 to 1) mA (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz (1 to 10) mA (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz (10 to 100) mA (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz (0.1 to 1) A (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz	0.4 % of reading + 0.03 $\mu$ A 0.15 % of reading + 0.03 $\mu$ A 0.06 % of reading + 0.03 $\mu$ A 0.4 % of reading + 0.2 $\mu$ A 0.15 % of reading + 0.2 $\mu$ A 0.06 % of reading + 0.2 $\mu$ A 0.03 % of reading + 0.2 $\mu$ A 0.4 % of reading + 2 $\mu$ A 0.15 % of reading + 2 $\mu$ A 0.06 % of reading + 2 $\mu$ A 0.03 % of reading + 2 $\mu$ A 0.4 % of reading + 20 $\mu$ A 0.15 % of reading + 20 $\mu$ A 0.06 % of reading + 20 $\mu$ A 0.03 % of reading + 20 $\mu$ A 0.4 % of reading + 0.2 mA 0.15 % of reading + 0.2 mA 0.06 % of reading + 0.2 mA 0.03 % of reading + 0.2 mA	Agilent 3458A Multimeter



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Windsor, CT

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Measure <sup>1</sup>	(1 to 3) A (3 to 5) Hz (5 to 10) Hz 10 Hz to 5 kHz (5 to 10) kHz	0.84 % of reading + 1.4 mA 0.27 % of reading + 1.4 mA 0.11 % of reading + 1.4 mA 0.27 % of reading + 16 mA	Fluke 8845A Multimeter
	(3 to 10) A (3 to 5) Hz (5 to 10) Hz 10 Hz to 5 kHz (5 to 10) kHz	0.84 % of reading + 4.6 mA 0.27 % of reading + 4.6 mA 0.11 % of reading + 4.6 mA 0.27 % of reading + 53 mA	
AC Current – Measure <sup>1</sup>	(10 to 1 000) A (10 to 100) Hz (100 to 500) Hz	1.5 % of reading + 1A 1.9 % of reading+ 1A	Fluke 376 Clamp Meter
AC Current – Measure <sup>1</sup>	(100 to 2 500) A (10 to 500) Hz	2.3 % of reading + 5A	Fluke 376 Clamp Meter W/i2500 flex probe
Resistance – Generate <sup>1</sup>	Up to 11 Ω (11 to 33) Ω (33 to 110) Ω (110 to 330) Ω 330 Ω to 1.1 kΩ (1.1 to 3.3) kΩ (3.3 to 11) kΩ (11 to 33) kΩ (33 to 110) kΩ (110 to 330) kΩ (0.33 to 1.1) MΩ (1.1 to 3.3) MΩ (3.3 to 11) MΩ	30 μΩ/Ω + 0.001 Ω 23 μΩ/Ω + 0.001 Ω 21 μΩ/Ω + 0.001 Ω 21 μΩ/Ω + 0.002 Ω 21 μΩ/Ω + 0.002 Ω 21 μΩ/Ω + 0.02 Ω 21 μΩ/Ω + 0.02 Ω 21 μΩ/Ω + 0.2 Ω 21 μΩ/Ω + 0.2 Ω 24 μΩ/Ω + 2 Ω 24 μΩ/Ω + 2 Ω 46 μΩ/Ω + 23 Ω 99 μΩ/Ω + 38 Ω	Fluke 5522A Multiproduct Calibrator
Resistance – Generate <sup>1</sup>	(11 to 33) MΩ (33 to 110) MΩ (110 to 330) MΩ 330 MΩ to 1.1 GΩ	190 μΩ/Ω + 1.9 KΩ 380 μΩ/Ω + 2.3 KΩ 0.23 % of setting + 76 kΩ 1.1 % of reading + 380 kΩ	Fluke 5522A Multiproduct Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Windsor, CT

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Resistance – Generate <sup>1</sup> Fixed Points	(1, 1.9) Ω (10, 19) Ω (100, 190) Ω (1, 1.9) kΩ (10, 19) kΩ 100 kΩ 190 kΩ 1 MΩ 1.9 MΩ 10 MΩ 19 MΩ 100 MΩ	84 μΩ/Ω + 0.1 μΩ 25 μΩ/Ω + 1 μΩ 15 μΩ/Ω + 6 μΩ 11 μΩ/Ω + 60 μΩ 11 μΩ/Ω + 0.6 mΩ 12 μΩ/Ω + 6 mΩ 12 μΩ/Ω + 6 mΩ 18 μΩ/Ω + 60 mΩ 18 μΩ/Ω + 60 mΩ 35 μΩ/Ω + 0.6Ω 42 μΩ/Ω + 0.6Ω 99 μΩ/Ω + 6Ω	Fluke 5700A Multifunction Calibrator
Resistance – Generate <sup>1</sup> Fixed Points	100 V 100 kΩ (100 to 1 000) V 1 MΩ 10 MΩ 100 MΩ 1 GΩ 10 GΩ	1 % of output 1 % of output 1 % of output 1 % of output 1 % of output 1.2 % of output	TMI RB Resistance Standard
Resistance – Measure	Up to 12 Ω (10 to 120) Ω (0.1 to 1.2) kΩ (1 to 12) kΩ (10 to 120) kΩ (0.1 to 1.2) MΩ (1 to 12) MΩ (10 to 120) MΩ (0.1 to 1.2) GΩ	15 μΩ/Ω + 56 μΩ 12 μΩ/Ω + 0.5 mΩ 10 μΩ/Ω + 0.6 mΩ 10 μΩ/Ω + 5.6 mΩ 10 μΩ/Ω + 56 mΩ 15 μΩ/Ω + 2.2 Ω 50 μΩ/Ω + 120 Ω 500 μΩ/Ω + 1.2 kΩ 0.5 % of reading + 70 kΩ	Agilent 3458A Multimeter
Electrical Calibration of Thermocouple Indicating Devices <sup>1</sup>	Type B (600 to 800) °C (800 to 1 000) °C (1 000 to 1 550) °C (1 550 to 1 820) °C	0.34 °C 0.26 °C 0.23 °C 0.25 °C	Fluke 5522A Multiproduct Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Windsor, CT

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Calibration of Thermocouple Indicating Devices <sup>1</sup>	Type C		Fluke 5522A Multiproduct Calibrator
	(0 to 150) °C	0.23 °C	
	(150 to 650) °C	0.2 °C	
	(650 to 1 000) °C	0.24 °C	
	(1 000 to 1 800) °C	0.39 °C	
	(1 800 to 2 316) °C	0.64 °C	
	Type E		
	(-250 to -100) °C	0.38 °C	
	(-100 to -25) °C	0.13 °C	
	(-25 to 350) °C	0.11 °C	
	(350 to 650) °C	0.13 °C	
	(650 to 1 000) °C	0.16 °C	
	Type J		
	(-210 to -100) °C	0.21 °C	
	(-100 to -30) °C	0.13 °C	
	(-30 to 150) °C	0.11 °C	
	(150 to 760) °C	0.13 °C	
	(760 to 1 200) °C	0.18 °C	
	Type K		
	(-200 to -100) °C	0.25 °C	
	(-100 to -25) °C	0.14 °C	
	(-25 to 120) °C	0.13 °C	
	(120 to 1 000) °C	0.2 °C	
	(1 000 to 1 372) °C	0.31 °C	
Type L			
(-200 to -100) °C	0.28 °C		
(-100 to 800) °C	0.2 °C		
(800 to 900) °C	0.13 °C		
Type N			
(-200 to -100) °C	0.31 °C		
(-100 to -25) °C	0.17 °C		
(-25 to 120) °C	0.15 °C		
(120 to 410) °C	0.14 °C		
(410 to 1 300) °C	0.21 °C		
Type R			
(0 to 250) °C	0.43 °C		
(250 to 1 000) °C	0.27 °C		
(1 000 to 1 400) °C	0.25 °C		
(1 400 to 1 767) °C	0.31 °C		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Windsor, CT

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment		
Electrical Calibration of Thermocouple Indicating Devices <sup>1</sup>	Type S (0 to 250) °C	0.36 °C	Fluke 5522A Multiproduct Calibrator		
	(250 to 1 000) °C	0.28 °C			
	(1 000 to 1 400) °C	0.28 °C			
	(1 400 to 1 767) °C	0.35 °C			
	Type T (-250 to -150) °C	0.48 °C			
	(-150 to 0) °C	0.18 °C			
	(0 to 120) °C	0.13 °C			
	(120 to 400) °C	0.11 °C			
	Type U (-200 to 0) °C	0.43 °C			
	(0 to 600) °C	0.21 °C			
	Capacitance – Generate <sup>1</sup>	(220 to 400) pF		0.38 % of output + 7.6pF	Fluke 5522A Multiproduct Calibrator
		(0.4 to 3.299 9) nF		0.38 % of output + 0.01 nF	
(3.3 to 10.999 9) nF		0.19 % of output + 0.01 nF			
(11 to 32.999 9) nF		0.19 % of output + 0.08 nF			
(33 to 109.999) nF		0.19 % of output + 0.08 nF			
(110 to 329.999) nF		0.19 % of output + 0.23 nF			
(0.33 to 1.099 99) μF		0.19 % of output + 0.76 nF			
(1.1 to 3.299 99) μF		0.19 % of output + 2.3 nF			
(3.3 to 10.999 9) μF		0.19 % of output + 7.6 nF			
(11 to 32.999 9) μF		0.3 % of output + 23 nF			
(33 to 109.999) μF		0.34 % of output + 76 nF			
(110 to 329.999) μF		0.34 % of output + 228 nF			
(0.33 to 1.099 99) mF		0.34 % of output + 0.76 μF			
(1.1 to 3.299 99) mF		0.34 % of output + 2.3 μF			
(3.3 to 10.999 9) mF	0.34 % of output + 7.6 μF				
(11 to 32.999 9) mF	0.57 % of output + 22.8 μF				
(33 to 110) mF	0.84 % of output + 76 μF				
Oscilloscopes Calibration – Generate <sup>1</sup> Voltage DC - 50Ω	(1 to 24.999) mV	0.19 % of Output + 31 μV	Fluke 5820A Oscilloscope Calibrator w/ GHz Option		
	(25 to 109.99) mV	0.19 % of Output + 36 μV			
	(110mV to 2.1999) V	0.19 % of Output + 87 μV			
	(2.2 to 6.6) V	0.19 % of Output + 0.6 mV			
	(1 to 24.999) mV	0.019 % of reading + 20 μV			
	(25 to 109.99) mV	0.019 % of reading + 25 μV			
	(110mV to 2.1999) V	0.019 % of reading + 76 μV			
	(2.2 to 10.999) V	0.019 % of reading + 0.6 mV			
DC - 1MΩ	(11 to 130) V	0.019 % of reading + 6 mV			



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Windsor, CT

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Oscilloscopes Calibration – Generate <sup>1</sup> Square Wave 10Hz to 10kHz - 50Ω  Square Wave 10 Hz to 1KHz - 1MΩ  Square Wave 1 to 10KHz - 1MΩ	(1 to 24.999) mVpp (25 to 109.99) mVpp (110mV to 2.1999) Vpp (2.2 to 6.6) Vpp  (1 to 24.999) mV (25 to 109.99) mV (110mV to 2.1999) V (2.2 to 10.999) V (11 to 130) V  (1 to 24.999) mV (25 to 109.99) mV (110mV to 2.1999) V (2.2 to 10.999) V (11 to 130) V	0.19 % of Output + 31 μV 0.19 % of Output + 36 μV 0.19 % of Output + 87 μV 0.19 % of Output + 0.6 mV  0.038 % of reading + 4 μV 0.038 % of reading + 9 μV 0.038 % of reading + 60 μV 0.038 % of reading + 0.6 mV 0.038 % of reading + 6 mV  0.19 % of reading + 31 μV 0.19 % of reading + 36 μV 0.19 % of reading + 87 μV 0.19 % of reading + 0.6 mV 0.19 % of reading + 6 mV	Fluke 5820A Oscilloscope Calibrator w/ GHz Option
Oscilloscopes Calibration – Generate <sup>1</sup> Leveled Sine Flatness 50 kHz to 10 MHz Reference	3 dB Bandwidth (5 to 50) mVpp 50 kHz to 100) MHz (100 to 300) MHz (300 to 500) MHz (500 to 600) MHz (600 to 16 00) MHz (1 600 to 2 100) MHz 50 mV to 3.5 Vpp 50 kHz to 100 MHz (100 to 300) MHz (300 to 500) MHz (500 to 600) MHz (600 to 16 00) MHz (1 600 to 2 100) MHz (3.5 to 5) Vpp 50 kHz to 100 MHz (100 to 300) MHz (300 to 500) MHz (500 to 600) MHz	0.34 dB 0.36 dB 0.42 dB 0.46 dB 0.5 dB 0.56 dB  0.24 dB 0.24 dB 0.32 dB 0.34 dB 0.4 dB 0.44 dB  0.24 dB 0.24 dB 0.32 dB 0.34 dB	Fluke 5820A Oscilloscope Calibrator w/ GHz Option
Oscilloscopes Calibration – Generate <sup>1</sup> Leveled Sine Flatness	3 dB Bandwidth 50 mV to 3.5Vpp (2 100 to 4 000) MHz (4 000 to 8 000) MHz (8 000 to 18 000) MHz	0.25 dB 0.35 dB 0.46 dB	EPM Power Meter w/ E Series Power Sensors



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Windsor, CT

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Oscilloscopes Calibration – Generate <sup>1</sup> Time Marker	500 ps to 20 ms 50 ms to 5 s	0.25 $\mu$ s/s 1.9 $\mu$ s/s + 3.8 $\mu$ Hz	Fluke 5820A Oscilloscope Calibrator w/ GHz Option
Oscilloscopes Calibration – Measure <sup>1</sup>			
Input Impedance	(40 to 60) $\Omega$	0.08 % of reading	Fluke 5820A Oscilloscope Calibrator w/ GHz Option
Resistance	500 k $\Omega$ to 1.5M $\Omega$	0.08 % of reading	
Leakage	(0 to 5.99) V	0.038 % of reading + 0.8 mV	

Electrical – RF/Microwave

Windsor, CT

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Flatness – Measure <sup>1</sup>	9 kHz to 2 000MHz		Agilent EPM Series Power Meter w/E9304A H18 Power Sensor
	(20 to -10) dBm	0.1 dB	
	(-10 to -30) dBm	0.1 dB	
	(-30 to -40) dBm	0.11 dB	
	(-40 to -42) dBm	0.12 dB	
	(2 to 14) GHz		
	(20 to -10) dBm	0.1 dB	
	(-10 to -30) dBm	0.09 dB	
	(-30 to -40) dBm	0.1 dB	
	(-40 to -42) dBm	0.11 dB	
	(14 to 18) GHz		
	(20 to -10) dBm	0.11 dB	
	(-10 to -30) dBm	0.12 dB	
(-30 to -40) dBm	0.12 dB		
(-40 to -42) dBm	0.13 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Windsor, CT

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Measure <sup>1</sup>	9 kHz to 14 000 MHz	0.13 dB	Agilent EPM Series Power Meter w/E9304A H18 Power Sensor
	(20 to 0) dB	0.15 dB	
	(0 to -40) dB	0.34 dB	
	(-40 to -50) dB	0.93 dB	
	(-50 to -55) dB		
	(14 000 to 18 000) MHz	0.12 dB	
	(20 to 0) dB	0.15 dB	
	(0 to -40) dB	0.34 dB	
RF Power – Measure <sup>1</sup>	100 kHz to 30 MHz	0.12 dB	Agilent N5531S Measuring Receiver N1912A w/E9304A Power Sensor
	(20 to 0) dB	0.13 dB	
	(0 to -58) dB	0.15 dB	
	(-58 to -78) dB	0.18 dB	
	(-78 to -110) dB	0.21 dB	
	(-110 to -115) dB	0.28 dB	
	(-115 to -120) dB	0.43 dB	
	(-120 to -125) dB		
RF Power – Measure <sup>1</sup>	(30 to 2 000) MHz	0.36 dB	Agilent N5531S Measuring Receiver w/N5532A Opt 526 Power Sensor
	(30 to 20) dB	0.2 dB	
	(20 to 0) dB	0.22 dB	
	(0 to -58) dB	0.23 dB	
	(-58 to -78) dB	0.25 dB	
	(-78 to -110) dB	0.27 dB	
	(-110 to -115) dB	0.33 dB	
	(-115 to -120) dB	0.46 dB	
	(-120 to -125) dB		
	(2 000 to 3 050) MHz	0.42 dB	
	(30 to 20) dB	0.3 dB	
	(20 to 0) dB	0.31 dB	
	(0 to -58) dB	0.32 dB	
	(-58 to -78) dB	0.34 dB	
	(-78 to -110) dB	0.35 dB	
	(-110 to -115) dB	0.4 dB	
	(-115 to -120) dB	0.51 dB	
	(-120 to -125) dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Windsor, CT

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Measure <sup>1</sup>	(3 050 to 6 600) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 526 Power Sensor
	(30 to 20) dB	0.42 dB	
	(20 to 0) dB	0.3 dB	
	(0 to -58) dB	0.31 dB	
	(-58 to -78) dB	0.32 dB	
	(-78 to -110) dB	0.34 dB	
	(-110 to -115) dB	0.38 dB	
	(-115 to -120) dB	0.48 dB	
	(-120 to -125) dB	0.64 dB	
	(6 600 to 13 200) MHz		
	(30 to 20) dB	0.42 dB	
	(20 to 0) dB	0.3 dB	
	(0 to -58) dB	0.31 dB	
	(-58 to -78) dB	0.32 dB	
	(-78 to -100) dB	0.34 dB	
	(-100 to -105) dB	0.37 dB	
	(-105 to -110) dB	0.45 dB	
	(-110 to -115) dB	0.6 dB	
	(-115 to -120) dB	0.82 dB	
	(13 200 to 18 000) MHz		
	(30 to 20) dB	0.42 dB	
	(20 to 0) dB	0.3 dB	
	(0 to -58) dB	0.31 dB	
	(-58 to -78) dB	0.32 dB	
	(-78 to -90) dB	0.33 dB	
	(-90 to -95) dB	0.35 dB	
	(-95 to -100) dB	0.41 dB	
	(-100 to -105) dB	0.53 dB	
	(-105 to -110) dB	0.72 dB	
	(18 000 to 19 200) MHz		
(30 to 20) dB	0.48 dB		
(20 to 0) dB	0.38 dB		
(0 to -58) dB	0.39 dB		
(-58 to -78) dB	0.4 dB		
(-78 to -90) dB	0.41 dB		
(-90 to -95) dB	0.42 dB		
(-95 to -100) dB	0.47 dB		
(-100 to -105) dB	0.58 dB		
(-105 to -110) dB	0.75 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Windsor, CT

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Measure <sup>1</sup>	(19 200 to 26 500) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 526 Power Sensor
	(30 to 20) dB	0.48 dB	
	(20 to 0) dB	0.38 dB	
	(0 to -58) dB	0.39 dB	
	(-58 to -78) dB	0.4 dB	
	(-78 to -90) dB	0.43 dB	
	(-90 to -95) dB	0.50 dB	
	(-95 to -100) dB	0.63 dB	
	(-100 to -105) dB	0.84 dB	
(-105 to -110) dB	1.1 dB		
Attenuation – Measure <sup>1</sup>	100 kHz to 3 050 MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 526 Power Sensor
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.12 dB	
	(90 to 100) dB	0.13 dB	
	(100 to 110) dB	0.13 dB	
	(110 to 120) dB	0.26 dB	
	(3 050 to 6 600) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.12 dB	
	(90 to 100) dB	0.13 dB	
(100 to 110) dB	0.15 dB		
(110 to 120) dB	0.37 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Windsor, CT

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Attenuation – Measure <sup>1</sup>	(6 600 to 13 200) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 526 Power Sensor
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.12 dB	
	(90 to 100) dB	0.14 dB	
	(100 to 110) dB	0.34 dB	
	(110 to 120) dB	0.77 dB	
	(13 200 to 19 200) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.12 dB	
	(90 to 100) dB	0.27 dB	
	(100 to 110) dB	0.66 dB	
	(19 200 to 26 500) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
(70 to 80) dB	0.12 dB		
(80 to 90) dB	0.2 dB		
(90 to 100) dB	0.5 dB		
(100 to 110) dB	1.1 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Windsor, CT

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Amplitude Modulation – Measure <sup>1</sup>	100 kHz to 10 MHz Rate 50 Hz to 10 kHz (5 to 99) % Depth 10 MHz to 3 GHz Rate 50 Hz to 100 kHz (5 to 20) % Depth (20 to 99) % Depth (3 to 26.5) GHz Rate 50 Hz to 100 kHz (5 to 20) % Depth (20 to 99) % Depth	0.75 % of reading + 0.3 digits  2.5 % of reading + 0.4 digits 1.5 % of reading + 0.4 digits  4.5 % of reading + 0.4 digits 1.5 % of reading + 0.4 digits	Agilent N5531S Measuring Receiver
Frequency Modulation – Measure <sup>1</sup>  $\beta$ = deviation / rate	250 kHz to 10 MHz Rates 20 Hz to 10 kHz Peak Dev 200 to 40KHz 10 MHz to 6.6 GHz Rates 50 Hz to 200 kHz Peak Dev 250 to 400KHz (6.6 to 13.2) GHz Rates 50 Hz to 200 kHz Peak Dev 250 to 400KHz (13.2 to 26.5) GHz Rates 50 Hz to 200 kHz Peak Dev 250 to 400KHz	$\beta > 0.2$ - 1.5 % of reading + 2 Hz $\beta > 1.2$ - 1 % of reading + 2 Hz  $\beta > 0.2$ - 1.5 % of reading + 2 Hz $\beta > 0.45$ - 1 % of reading + 2 Hz  $\beta > 0.2$ - 2.5 % of reading + 4 Hz $\beta > 8.0$ - 1 % of reading + 4 Hz  $\beta > 0.2$ - 3.8 % of reading + 9 Hz $\beta > 16$ - 1 % of reading + 9 Hz	Agilent N5531S Measuring Receiver
Phase Modulation – Measure <sup>1</sup>	100 kHz to 6.6 GHz Deviations > 0.3 rad Deviations > 0.7 rad (6.6 to 13.2) GHz Deviations > 0.6 rad Deviations > 2.0 rad (13.2 to 26.5) GHz Deviations: > 1.2 rad Deviations > 4.0 rad	3 % of reading + 0.002 rad 1 % of reading + 0.002 rad  3 % of reading + 0.005 rad 1 % of reading + 0.005 rad  3 % of reading + 0.009 rad 1 % of reading + 0.009 rad	Agilent N5531S Measuring Receiver
RF Power – Generate	(30 to 2 000) MHz (20 to 0) dB (0 to -58) dB (-58 to -78) dB (-78 to -110) dB (-110 to -120) dB	0.26 dB 0.28 dB 0.29 dB 0.3 dB 0.37 dB	Agilent N5531S Measuring Receiver w/N5532A Opt 526 Power Sensor, 83630B Signal Generator



ANSI National Accreditation Board

Electrical – RF/Microwave

Windsor, CT

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Generate	(2 000 to 3 050) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 526 Power Sensor, 83630B Signal Generator
	(20 to 0) dB	0.37 dB	
	(0 to -58) dB	0.38 dB	
	(-58 to -78) dB	0.38 dB	
	(-78 to -110) dB	0.4 dB	
	(-110 to -120) dB	0.52 dB	
	(3 050 to 6 600) MHz		
	(20 to 0) dB	0.37 dB	
	(0 to -58) dB	0.38 dB	
	(-58 to -78) dB	0.38 dB	
	(-78 to -110) dB	0.4 dB	
	(-110 to -120) dB	0.52 dB	
	(6 600 to 13 200) MHz		
	(20 to 0) dB	0.37 dB	
	(0 to -58) dB	0.38 dB	
	(-58 to -78) dB	0.38 dB	
	(-78 to -110) dB	0.49 dB	
	(-110 to -120) dB	0.84 dB	
	(13 200 to 18 000) MHz		
	(20 to 0) dB	0.37 dB	
	(0 to -58) dB	0.38 dB	
	(-58 to -78) dB	0.38 dB	
	(-78 to -110) dB	0.75 dB	
	(-110 to -120) dB	1.3 dB	
(18 000 to 19 200) MHz			
(20 to 0) dB	0.49 dB		
(0 to -58) dB	0.5 dB		
(-58 to -78) dB	0.5 dB		
(-78 to -110) dB	0.81 dB		
(-110 to -120) dB	1.4 dB		
(19 200 to 26 500) MHz			
(20 to 0) dB	0.49 dB		
(0 to -58) dB	0.5 dB		
(-58 to -78) dB	0.5 dB		
(-78 to -110) dB	1.2 dB		
(-110 to -120) dB	1.9 dB		

Electrical – RF/Microwave

Windsor, CT

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Total Harmonic Distortion (THD)	(0 to 60) dB 20 Hz to 20 kHz (-40 to 0) dB (-50 to -40) dB (-60 to -50) dB (-65 to -60) dB 20 to 50 kHz (-40 to 0) dB (-50 to -40) dB (-60 to -50) dB 50 to 100 kHz (-40 to 0) dB (-50 to -40) dB	1 dB 1 dB 1.3 dB 1.7 dB 2 dB 2.1 dB 3 dB 2 dB 2.4 dB	HP 8903B Audio Analyzer
AM Distortion Measure <sup>1</sup> Rate 20 Hz to 1 kHz	(0.1 to 10) MHz AM Depth > 1% (0 to -20) dB (-20 to -30) dB AM Depth > 3% (0 to -20) dB (-20 to -30) dB (-30 to -40) dB 10 MHz to 26.5 GHz AM Depth > 1% (0 to -20) dB (-20 to -30) dB AM Depth > 3% (0 to -20) dB (-20 to -30) dB (-30 to -40) dB	1.2 dB 2.2 dB 1 dB 1.3 dB 2.4 dB 1.3 dB 2.5 dB 1.1 dB 1.4 dB 3 dB	Agilent N5531S Measuring Receiver
FM Distortion Measure <sup>1</sup> Rate 20 Hz to 1 kHz	(1 to 6 600) MHz Dev 500 Hz to 2 kHz (0 to -20) dB (-20 to -30) dB (-30 to -40) dB Dev > 2 kHz (0 to -20) dB (-20 to -30) dB (-30 to -40) dB (-40 to -50) dB	0.26 dB 0.79 dB 2.3 dB 0.09 dB 0.27 dB 0.83 dB 2.4 dB	Agilent N5531S Measuring Receiver



ANSI National Accreditation Board

Electrical – RF/Microwave

Windsor, CT

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
FM Distortion Measure <sup>1</sup> Rate 20 Hz to 1 kHz	(6.6 to 13.2) GHz Dev > 2.3 kHz (0 to -20) dB (-20 to -30) dB (-30 to -40) dB Dev > 4.5 kHz (0 to -20) dB (-20 to -30) dB (-30 to -40) dB (-40 to -50) dB (13.2 to 26.5) GHz Dev > 2.7 kHz (0 to -20) dB (-20 to -30) dB (-30 to -40) dB Dev > 6.0 kHz (0 to -20) dB (-20 to -30) dB (-30 to -40) dB (-40 to -50) dB	0.26 dB 0.79 dB 2.3 dB 0.09 dB 0.27 dB 0.83 dB 2.4 dB 0.26 dB 0.79 dB 2.3 dB 0.09 dB 0.27 dB 0.83 dB 2.4 dB	Agilent N5531S Measuring Receiver
PM Distortion Measure <sup>1</sup>	(1 to 6 600) MHz Rate (20 to 500) Hz Dev > 0.8 rad (0 to -20) dB (-20 to -30) dB (-30 to -40) dB Dev > 2.5 rad (0 to -20) dB (-20 to -30) dB (-30 to -40) dB (-40 to -50) dB Rate (500 to 1 000) Hz Dev > 0.4 rad (0 to -20) dB (-20 to -30) dB (-30 to -40) dB Dev > 1.0 rad (0 to -20) dB (-20 to -30) dB (-30 to -40) dB (-40 to -50) dB	0.26 dB 0.79 dB 2.3 dB 0.09 dB 0.27 dB 0.83 dB 2.4 dB 0.26 dB 0.79 dB 2.3 dB 0.09 dB 0.27 dB 0.83 dB 2.4 dB	Agilent N5531S Measuring Receiver



ANSI National Accreditation Board

Electrical – RF/Microwave

Windsor, CT

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
PM Distortion Measure <sup>1</sup>	(6.6 to 13.2) GHz		
	Rate (20 to 500) Hz		
	Dev > 1.8 rad		
	(0 to -20) dB		0.26 dB
	(-20 to -30) dB		0.79 dB
	(-30 to -40) dB		2.3 dB
	Dev > 5.5 rad		
	(0 to -20) dB		0.09 dB
	(-20 to -30) dB		0.27 dB
	(-30 to -40) dB		0.83 dB
	(-40 to -50) dB		2.4 dB
	Rate (500 to 1 000) Hz		
	Dev > 0.8 rad		
	(0 to -20) dB		0.26 dB
	(-20 to -30) dB		0.79 dB
	(-30 to -40) dB		2.3 dB
	Dev > 2.5 rad		
	(0 to -20) dB		0.09 dB
(-20 to -30) dB		0.27 dB	
(-30 to -40) dB		0.83 dB	
(-40 to -50) dB		2.4 dB	
(13.2 to 26.5) GHz			
Rate (20 to 500) Hz			
Dev > 3.5 rad			
(0 to -20) dB		0.26 dB	
(-20 to -30) dB		0.79 dB	
(-30 to -40) dB		2.3 dB	

Agilent N5531S  
Measuring Receiver



ANSI National Accreditation Board

Electrical – RF/Microwave

Windsor, CT

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
PM Distortion Measure <sup>1</sup>	(13.2 to 26.5) GHz		Agilent N5531S Measuring Receiver
	Dev > 10.0 rad	0.09 dB	
	(0 to -20) dB	0.27 dB	
	(-20 to -30) dB	0.83 dB	
	(-30 to -40) dB	2.4 dB	
	(-40 to -50) dB		
	Rate (500 to 1 000) Hz		
	Dev > 1.2 rad	0.26 dB	
	(0 to -20) dB	0.79 dB	
	(-20 to -30) dB	2.3 dB	
	(-30 to -40) dB		
	Dev > 4.0 rad	0.09 dB	
(0 to -20) dB	0.27 dB		
(-20 to -30) dB	0.83 dB		
(-30 to -40) dB	2.4 dB		
(-40 to -50) dB			
Harmonics Measure <sup>1</sup>	(-80 to -10) dB		Agilent E4440A Measuring Receiver
	2 <sup>nd</sup> through 5 <sup>th</sup> Harmonic	0.37 dB	
	1 kHz to 600 MHz	1.1 dB	
	(600 to 1 320) MHz	1.4 dB	
	(1 320 to 2 200) MHz	1.4 dB	
	(2 200 to 3 000) MHz	1.7 dB	
(3 000 to 4 400) MHz	1.9 dB		
(4 400 to 5 300) MHz			
Harmonics Measure <sup>1</sup>	(-80 to -10) dB		Agilent E4440A Measuring Receiver
	2 <sup>nd</sup> through 4 <sup>th</sup> Harmonic	2.1 dB	
	(5 300 to 6 625) MHz	2.1 dB	
	2 <sup>nd</sup> through 3 <sup>rd</sup> Harmonic	2.1 dB	
(6 625 to 8 833) MHz			
2 <sup>nd</sup> Harmonic	2.1 dB		
(8 833 to 13 250) MHz			

Length – Dimensional metrology

Windsor, CT

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Gage Blocks <sup>2</sup>	Up to 12 in	$(3.7 + 2.4L) \mu\text{in}$	Master gage blocks, P&W Universal Measuring Machine
Micrometers <sup>1,2</sup>	Up to 18 in	$(26 + 4.7L) \mu\text{in}$	Gage blocks

**Length – Dimensional metrology**

Windsor, CT

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Laser Micrometer <sup>1</sup>	(0.5 to 1.5) in	88 μin	Pin Gages
Calipers <sup>1,2</sup>	Up to 40 in	(281 + 1.9L) μin	Gage blocks
Dial Indicators <sup>1,2</sup> Resolution: ≥ 50μin < 50μin	Up to 10 in Up to 0.1 in	(25 + 4L) μin 10 μin	Gage blocks
Height Gages <sup>1,2</sup>	Up to 40 in	(94 + 3.1L) μin	Gage blocks
Cylindrical Gages <sup>2</sup> – Rings Plain Pins, Plugs	(0.02 to 8) in (0 to 8) in	(10 + 2.7D) μin (10 + 3D) μin	Master gage blocks, P&W Universal Measuring Machine
Thread Plugs <sup>1</sup> – Major Diameter Pitch Diameter	Up to 10 in Up to 10 in	43 μin 94 μin	Fisher Machine Shop Model 1962 thread wire set, Gage blocks, Pratt & Whitney Model C
Surface Plates <sup>1</sup> Overall Flatness Local Area Flatness	(18 x 18) in to (10 x 10) ft (-0.001 to 0.001) in	66 μin 68 μin	Mahr Leveling System Repeat-O-Meter
Protractors <sup>1</sup>	(0 to 360)°	0.015°	Angle blocks
Linear Scales – Rulers <sup>1</sup>	Up to 36 in	0.009 1 in	Gage blocks
Pi Tapes <sup>1</sup>	Up to 3 in	450 μin	Plug Gages
Tape Measures <sup>1</sup>	Up to 150 ft	(0.000 28F + 0.024) in	Standard rule
Microscopes <sup>1</sup>	Up to 2 in	100 μin	Stage Micrometer

**Mass and Mass Related**

Windsor, CT

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Scales & Balances <sup>1,4</sup>	1 mg to 220 g	(0.048 + 0.003 1X) mg	ASTM E617 Class 1 weights
	Up to 1 000 lb Up to 454 kg	(0.000 12W) lb (0.000 12X) g	NIST Class F weights

**Mass and Mass Related**

Windsor, CT

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Pressure <sup>1</sup>	(-1 to 1) inH <sub>2</sub> O -15 to 100 psig Up to 500 psig Up to 1 000 psig Up to 10 000 psig	0.004 inH <sub>2</sub> O 0.027 psi 0.12 psi 0.26 psi 2.5 psi	Pressure Gauge/Module Additel ADT155-DPI-760, ADT681-02-GP100-PSI, ADT681-02-GP500-PSI ADT681-02-GP1K-PSI, ADT681-02-GP10K-PSI
Torque Tools <sup>1</sup>	(10 to 100) ozf·in 4 lbf·in to 250 lbf·ft	0.6 % of reading 0.32 % of reading	CDI 1001-O-DDT CDI 5000 ST Torque Tester
Force <sup>1,2</sup> Tension and Compression	(0.5 to 500) lbf 220 gf to 23 kgf	(0.000 06 + 0.000 12 <i>W</i> ) lbf <sup>2</sup> (0.03 + 0.000 12 <i>M</i> ) grams <sup>2</sup>	Class F weights

**Thermodynamic**

Windsor, CT

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Relative Humidity Generate	(10 to 95) %RH	0.5 %RH	Thunder Scientific 2500 Humidity Chamber
Relative Humidity Measure	(0 to 90) %RH	1.6 %RH	Vaisala M170/HMP75 Humidity Indicator and Probe
Temperature – Measure <sup>1</sup>	(-196 to 0.01) °C (0.01 to 660) °C	0.004 3 °C (0.004 3 + 0.000 001 <i>T</i> ) °C	Fluke 5699 SPRT & Fluke 1594A Super-Thermometer
Temperature – Source	(-196 to 300) °C (>300 to 500) °C	0.014 °C 0.02 °C	Fluke 5699 SPRT, Fluke 1594A Super-Thermometer & Temperature Baths
Temperature – Source Fix Point	0.01 °C	0.002 6 °C	Triple Point of Water

**Time and Frequency**

Windsor, CT

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency – Generate	10 MHz	1 x 1 <sup>e-9</sup> Hz/Hz	HP Z3801A GPS Receiver



ANSI National Accreditation Board

Time and Frequency

Windsor, CT

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency – Generate	(1 to 10) Hz (10 to 100) Hz (100 to 1 000) Hz (1 to 10) kHz (10 to 100) kHz (0.1 to 1) MHz (1 to 10) MHz	$1 \times 10^{-9}$ Hz/Hz + 0.57 $\mu$ Hz $1 \times 10^{-9}$ Hz/Hz + 5.7 $\mu$ Hz $1 \times 10^{-9}$ Hz/Hz + 57 $\mu$ Hz $1 \times 10^{-9}$ Hz/Hz + 0.57 mHz $1 \times 10^{-9}$ Hz/Hz + 5.7 mHz $1 \times 10^{-9}$ Hz/Hz + 57 mHz $1 \times 10^{-9}$ Hz/Hz + 0.57 Hz	Agilent 33250A Function Generator / HP Z3801A GPS Receiver
Frequency – Generate	(10 to 26 500) MHz	$1 \times 10^{-9}$ Hz/Hz + 0.57 kHz	HP 83630B Signal Generator / HP Z3801A GPS Receiver
Frequency – Measure <sup>1</sup>	1 to 10 Hz 10 to 100 Hz 100 to 1000 Hz 1 to 10 kHz 10 to 100 kHz 100 to 200 kHz 0.2 to 3 000 MHz	$5.20 \times 10^{-9}$ Hz/Hz $2.46 \times 10^{-9}$ Hz/Hz $1.60 \times 10^{-9}$ Hz/Hz $1.33 \times 10^{-9}$ Hz/Hz $1.24 \times 10^{-9}$ Hz/Hz $1.21 \times 10^{-9}$ Hz/Hz $1.21 \times 10^{-9}$ Hz/Hz	Agilent 53131A Frequency Counter / HP Z3801A GPS Receiver
	10 to 26 500 MHz	$1 \times 10^{-9}$ Hz/Hz + 0.1 Hz	Agilent E4440A Spectrum Analyzer / HP Z3801A GPS Receiver
Timer, Stopwatch <sup>1</sup>	10 s to 24 hr	34 ms	Totalize method with counter
Tachometers – RPM <sup>1</sup>	Up to 100 000 RPM	$\pm (0.001\% + 0.6R)$	HP 33250A Signal Generator & LED

[Return to Site listing](#) (top) [Go to Notes](#) (bottom)



ANSI National Accreditation Board

Services performed at satellite laboratory

**Technical Maintenance, Inc.**

4613 Northwest Parkway  
Hilliard, OH 43026

Matthew Ayres (Branch Manager) Phone: 614-850-9940  
Scott Chamberlain (Quality Manager) Phone: 321-242-0890

**CALIBRATION AND DIMENSIONAL MEASUREMENT**

**CALIBRATION**

**Acoustics and Vibration**

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Accelerometers – Acceleration	(0.01 to 10) g (7 < 10) Hz (10 < 30) Hz (30 < 2 000) Hz (2 to 10) kHz	4 % of reading 3 % of reading 1.5 % of reading 4 % of reading	Accelerometer Calibrator

**Chemical Quantities**

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
pH Meters <sup>1,5</sup>	(4, 7, 10) pH	0.029 pH	Standard pH buffers
Conductivity Meters <sup>1,5</sup>	1 µS/cm 10 µS/cm 100 µS/cm 1 000 µS/cm 10 000 µS/cm	0.5 µS/cm 0.41 µS/cm 2.7 µS/cm 15 µS/cm 140 µS/cm	Standard conductivity solutions

**Electrical – DC/Low Frequency**

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Voltage – Generate <sup>1</sup>	(0 to 220) mV (0.22 to 2.2) V (2.2 to 11) V (11 to 22) V (22 to 220) V (220 to 1 100) V	6.8 $\mu\text{V/V} + 0.8 \mu\text{V}$ 4.6 $\mu\text{V/V} + 0.9 \mu\text{V}$ 3 $\mu\text{V/V} + 2.5 \mu\text{V}$ 3 $\mu\text{V/V} + 3.9 \mu\text{V}$ 4.6 $\mu\text{V/V} + 38 \mu\text{V}$ 6.1 $\mu\text{V/V} + 385 \mu\text{V}$	Fluke 5720A Multifunction Calibrator
DC Voltage – Generate <sup>1</sup>	(1 to 6) kV	2.4 % of output + 10 V	Associated Research 3565D Hypot
DC Voltage – Measure <sup>1</sup>	(0 to 100) mV 100 mV to 1V (1 to 10) V (10 to 100) V (100 to 1 000) V	5.0 $\mu\text{V/V} + 1.2 \mu\text{V}$ 4.0 $\mu\text{V/V} + 1.2 \mu\text{V}$ 4.0 $\mu\text{V/V} + 2.1 \mu\text{V}$ 6.0 $\mu\text{V/V} + 30 \mu\text{V}$ 6 $\mu\text{V/V} + 100 \mu\text{V} + 12 \mu\text{V/V} \times (\text{Vin}/1\ 000)^2$	HP 3458A Multimeter
DC Voltage – Measure <sup>1</sup>	(1 to 60) kV	0.1 % of reading	Ross VD60 High Voltage Divider, HP 34401A Multimeter
DC Current – Generate <sup>1</sup>	(1 to 2.2) nA (2.2 to 22) nA (22 to 220) nA (0.22 to 2.2) $\mu\text{A}$ (2.2 to 10) $\mu\text{A}$	93 $\mu\text{A/A} + 0.007 \text{ nA}$ 92 $\mu\text{A/A} + 0.007 \text{ nA}$ 92 $\mu\text{A/A} + 0.01 \text{ nA}$ 36 $\mu\text{A/A} + 0.1 \text{ nA}$ 20 $\mu\text{A/A} + 1 \text{ nA}$	Fluke 5720A Multifunction Calibrator & Fluke 5522A Multifunction Calibrator
DC Current – Generate <sup>1</sup>	(10 to 220) $\mu\text{A}$ (0.22 to 2.2) mA (2.2 to 22) mA (22 to 100) mA (100 to 220) mA (0.22 to 1) A (1 to 2.2) A	38 $\mu\text{A/A} + 5 \text{ nA}$ 30 $\mu\text{A/A} + 7 \text{ nA}$ 30 $\mu\text{A/A} + 44 \text{ nA}$ 38 $\mu\text{A/A} + 0.7 \mu\text{A}$ 45 $\mu\text{A/A} + 0.7 \mu\text{A}$ 68 $\mu\text{A/A} + 12 \mu\text{A}$ 105 $\mu\text{A/A} + 12 \mu\text{A}$	Fluke 5720A Multifunction Calibrator
DC Current – Generate <sup>1</sup>	(2.2 to 11) A	274 $\mu\text{A/A} + 365 \mu\text{A}$	Fluke 5720A Multifunction Calibrator, 5725A Amplifier
DC Current – Generate <sup>1</sup>	(11 to 20.5) A	761 $\mu\text{A/A} + 578 \mu\text{A}$	Fluke 5522A Multiproduct Calibrator
DC Current – Generate <sup>1</sup> Clamp Meters	(0 to 200) A	0.21% of output + 0.028A	Fluke 5522A Multiproduct Calibrator / 9100-200 x10 Coil
DC Current – Generate <sup>1</sup> Clamp Meters	(0 to 1000) A	0.21% of output + 0.040A	Fluke 5522A Multiproduct Calibrator / 9100-200 x50 Coil



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
High Voltage -Generate DC Voltage	(1 to 6) kV	2.4 % of reading + 10 V	Associated Research 3565D High Voltage Tester
DC Current – Measure <sup>1</sup>	(1 to 10) nA (10 to 100) nA (0.1 to 1) $\mu$ A (1 to 10) $\mu$ A	34 $\mu$ A/A + 0.12 pA 18 $\mu$ A/A + 1.2 pA 11 $\mu$ A/A + 0.01 nA 9.1 $\mu$ A/A + 0.12 nA	Fluke 5720A Multifunction Calibrator, Agilent 3458A Multimeter Option 002
DC Current – Measure <sup>1</sup>	(10 to 100) $\mu$ A (0.1 to 1) mA (1 to 10) mA (10 to 100) mA (0.1 to 1) A	20 $\mu$ A/A + 0.8 nA 20 $\mu$ A/A + 5 nA 20 $\mu$ A/A + 0.05 $\mu$ A 35 $\mu$ A/A + 0.5 $\mu$ A 110 $\mu$ A/A + 10.1 $\mu$ A	Agilent 3458A Multimeter
DC Current – Measure <sup>1</sup>	(1 to 1 000) A	0.26 % of reading	Current shunt
Resistance – Measure <sup>1</sup>	Up to 12 $\Omega$ (10 to 120) $\Omega$ (0.1 to 1.2 k $\Omega$ (1 to 12) k $\Omega$ (10 to 120) k $\Omega$ (0.1 to 1.2) M $\Omega$ (1 to 12) M $\Omega$ (10 to 120) M $\Omega$ (0.1 to 1.2) G $\Omega$	15 $\mu\Omega/\Omega$ + 56 $\mu\Omega$ 12 $\mu\Omega/\Omega$ + 0.5 m $\Omega$ 10 $\mu\Omega/\Omega$ + 0.6 m $\Omega$ 10 $\mu\Omega/\Omega$ + 5.6 m $\Omega$ 10 $\mu\Omega/\Omega$ + 56 m $\Omega$ 15 $\mu\Omega/\Omega$ + 2.2 $\Omega$ 50 $\mu\Omega/\Omega$ + 120 $\Omega$ 500 $\mu\Omega/\Omega$ + 1.2 k $\Omega$ 0.5 % of reading + 70 k $\Omega$	Agilent 3458A Multimeter
Resistance <sup>1</sup> Fixed Points	(1, 1.9) $\Omega$ (10, 19) $\Omega$ (100, 190) $\Omega$ (1, 1.9) k $\Omega$ (10, 19) k $\Omega$ (100, 190) k $\Omega$ 1 M $\Omega$ 1.9 M $\Omega$ 10 M $\Omega$ 19 M $\Omega$ 100 M $\Omega$	84 $\mu\Omega/\Omega$ + 0.1 $\mu\Omega$ 21 $\mu\Omega/\Omega$ + 1 $\mu\Omega$ 9.1 $\mu\Omega/\Omega$ + 6 $\mu\Omega$ 7.6 $\mu\Omega/\Omega$ + 60 $\mu\Omega$ 7.6 $\mu\Omega/\Omega$ + 0.6 m $\Omega$ 9.9 $\mu\Omega/\Omega$ + 6 m $\Omega$ 18 $\mu\Omega/\Omega$ + 60 m $\Omega$ 18 $\mu\Omega/\Omega$ + 60 m $\Omega$ 35 $\mu\Omega/\Omega$ + 0.6 $\Omega$ 42 $\mu\Omega/\Omega$ + 0.6 $\Omega$ 91 $\mu\Omega/\Omega$ + 6 $\Omega$	Fluke 5720A Multifunction Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Resistance – Generate <sup>1</sup>	Up to 11 Ω (11 to 33) Ω (33 to 110) Ω (110 to 330) Ω (0.33 to 1.1) kΩ (1.1 to 3.3) kΩ (3.3 to 11) kΩ (11 to 33) kΩ (33 to 110) kΩ (110 to 330) kΩ 0.33 kΩ to 1.1 MΩ (1.1 to 3.3) MΩ (3.3 to 11) MΩ (11 to 33) MΩ (33 to 110) MΩ (110 to 330) MΩ 330 MΩ to 1.1 GΩ	30 μΩ/Ω + 0.001 Ω 23 μΩ/Ω + 0.001 Ω 21 μΩ/Ω + 0.001 Ω 21 μΩ/Ω + 0.002 Ω 21 μΩ/Ω + 0.002 Ω 21 μΩ/Ω + 0.02 Ω 21 μΩ/Ω + 0.02 Ω 21 μΩ/Ω + 0.2 Ω 21 μΩ/Ω + 0.2 Ω 24 μΩ/Ω + 2 Ω 24 μΩ/Ω + 2 Ω 46 μΩ/Ω + 23 Ω 99 μΩ/Ω + 38 Ω 190 μΩ/Ω + 1.9 k Ω 380 μΩ/Ω + 2.3 k Ω 0.23 % of setting + 76 k Ω 1.1 % of reading + 380 k Ω	Fluke 5522A Multiproduct Calibrator
Resistance – Generate <sup>1</sup> Fixed Points	100 V 100 kΩ (100 to 1 000) V 1 MΩ 10 MΩ 100 MΩ 1 GΩ 10 GΩ	1 % of output 1 % of output 1 % of output 1 % of output 1 % of output 1.2 % of output	TMI RB Resistance Standard
Resistance – Measure <sup>1</sup>	Up to 12 Ω (10 to 120) Ω (0.1 to 1.2 kΩ) (1 to 12) kΩ (10 to 120) kΩ (0.1 to 1.2) MΩ (1 to 12) MΩ (10 to 120) MΩ (0.1 to 1.2) GΩ	15 μΩ/Ω + 56 μΩ 12 μΩ/Ω + 0.5 mΩ 10.0 μΩ/Ω + 0.6 mΩ 10.0 μΩ/Ω + 5.6 mΩ 10.0 μΩ/Ω + 56 mΩ 15 μΩ/Ω + 2.2 Ω 50 μΩ/Ω + 120 Ω 500 μΩ/Ω + 1.2 kΩ 0.5 % of reading + 70 kΩ	Agilent 3458A Multimeter



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Capacitance – Generate <sup>1</sup>	(220 to 400) pF (0.4 to 3.299 9) nF (3.3 to 10.999 9) nF (11 to 32.999 9) nF (33 to 109.999) nF (110 to 329.999) nF (0.33 to 1.099 99) μF (1.1 to 3.299 99) μF (3.3 to 10.999 9) μF (11 to 32.999 9) μF (33 to 109.999) μF (110 to 329.999) μF (0.33 to 1.099 99) mF (1.1 to 3.299 99) mF (3.3 to 10.999 9) mF (11 to 32.999 9) mF (33 to 110) mF	0.38 % of output + 7.6 pF 0.38 % of output + 0.01 nF 0.19 % of output + 0.01 nF 0.19 % of output + 0.08 nF 0.19 % of output + 0.08 nF 0.19 % of output + 0.23 nF 0.19 % of output + 0.76 nF 0.19 % of output + 2.3 nF 0.19 % of output + 7.6 nF 0.3 % of output + 23 nF 0.34 % of output + 76 nF 0.34 % of output + 228 nF 0.34 % of output + 0.76 μF 0.34 % of output + 2.3 μF 0.34 % of output + 7.6 μF 0.57 % of output + 23 μF 0.84 % of output + 76 μF	Fluke 5522A Multiproduct Calibrator
AC Voltage – Generate <sup>1</sup>	(0.22 to 2.2) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz (2.2 to 22) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	228 μV/V + 3.9 μV 88 μV/V + 3.9 μV 76 μV/V + 3.9 μV 190 μV/V + 3.9 μV 457 μV/V + 4.6 μV 989 μV/V + 9.2 μV 1.3 mV/V + 19 μV 2.6 mV/V + 19 μV 228 μV/V + 3.9 μV 88 μV/V + 3.9 μV 76 μV/V + 3.9 μV 190 μV/V + 3.9 μV 457 μV/V + 4.6 μV 989 μV/V + 9.2 μV 1.3 mV/V + 19 μV 2.6 mV/V + 19 μV	Fluke 5720A Multifunction Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Generate <sup>1</sup>	(22 to 220) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	228 $\mu$ V/V + 11.4 $\mu$ V 88 $\mu$ V/V + 6.1 $\mu$ V 76 $\mu$ V/V + 6.1 $\mu$ V 190 $\mu$ V/V + 6.1 $\mu$ V 457 $\mu$ V/V + 15.2 $\mu$ V 837 $\mu$ V/V + 19 $\mu$ V 1.3 mV/V + 23 $\mu$ V 2.5 mV/V + 46 $\mu$ V	Fluke 5720A Multifunction Calibrator
	220 mV to 2.2 V (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	228 $\mu$ V/V + 38 $\mu$ V 84 $\mu$ V/V + 15 $\mu$ V 40 $\mu$ V/V + 8 $\mu$ V 68 $\mu$ V/V + 9 $\mu$ V 99 $\mu$ V/V + 30 $\mu$ V 380 $\mu$ V/V + 76 $\mu$ V 913 $\mu$ V/V + 190 $\mu$ V 1.5 mV/V + 304 $\mu$ V	
AC Voltage – Generate <sup>1</sup>	(2.2 to 22) V (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	228 $\mu$ V/V + 380 $\mu$ V 84 $\mu$ V/V + 152 $\mu$ V 37 $\mu$ V/V + 54 $\mu$ V 61 $\mu$ V/V + 91 $\mu$ V 76 $\mu$ V/V + 190 $\mu$ V 228 $\mu$ V/V + 609 $\mu$ V 913 $\mu$ V/V + 1.9 mV 1.4 mV/V + 3 mV	Fluke 5720A Multifunction Calibrator / Fluke 5725A Amplifier
	(22 to 220) V (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz	228 $\mu$ V/V + 3.8 mV 84 $\mu$ V/V + 1.5 mV 49 $\mu$ V/V + 0.6 mV 76 $\mu$ V/V + 0.9 mV 137 $\mu$ V/V + 2.3 mV	
AC Voltage – Generate <sup>1</sup>	(220 to 750) V 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz	68 $\mu$ V/V + 3 mV 126 $\mu$ V/V + 5 mV 457 $\mu$ V/V + 8 mV 1.8 mV/V + 34 mV	Fluke 5720A Multifunction Calibrator / Fluke 5725A Amplifier
	(750 to 1 000) V 40 Hz to 1 kHz (1 to 20) kHz (20 to 30) kHz	68 $\mu$ V/V + 3 mV 126 $\mu$ V/V + 5 mV 457 $\mu$ V/V + 8 mV	



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Generate <sup>1</sup> Wideband Absolute	(0.3 to 1.1) mV		Fluke 5720A Option 003 Multifunction Calibrator
	(10 to 30 Hz)	0.65 % of output + 1.5 $\mu$ V	
	30 Hz to 500 kHz	0.61 % of output + 1.5 $\mu$ V	
	(0.5 to 1.2) MHz	0.63 % of output + 3.8 $\mu$ V	
	(1.2 to 2) MHz	0.63 % of output + 3.8 $\mu$ V	
	(2 to 12) MHz	0.68 % of output + 3.8 $\mu$ V	
	(12 to 20) MHz	0.76 % of output + 3.8 $\mu$ V	
	(20 to 30) MHz	1.3 % of output + 12.9 $\mu$ V	
	(1.1 to 3.3) mV		
	(10 to 30 Hz)	0.58 % of output + 2.3 $\mu$ V	
	30 Hz to 500 kHz	0.53 % of output + 2.3 $\mu$ V	
	(0.5 to 1.2) MHz	0.54 % of output + 4.6 $\mu$ V	
	(1.2 to 2) MHz	0.54 % of output + 4.6 $\mu$ V	
	(2 to 12) MHz	0.58 % of output + 4.6 $\mu$ V	
	(12 to 20) MHz	0.65 % of output + 4.6 $\mu$ V	
	(20 to 30) MHz	1.3 % of output + 4.6 $\mu$ V	
	(3.3 to 11) mV		
	(10 to 30 Hz)	0.58 % of output + 6.1 $\mu$ V	
	30 Hz to 500 kHz	0.53 % of output + 6.1 $\mu$ V	
	(0.5 to 1.2) MHz	0.54 % of output + 8.4 $\mu$ V	
	(1.2 to 2) MHz	0.54 % of output + 8.4 $\mu$ V	
	(2 to 12) MHz	0.55 % of output + 8.4 $\mu$ V	
	(12 to 20) MHz	0.61 % of output + 8.4 $\mu$ V	
	(20 to 30) MHz	0.93 % of output + 8.4 $\mu$ V	
	(11 to 33) mV		
	(10 to 30 Hz)	0.52 % of output + 12 $\mu$ V	
	30 Hz to 500 kHz	0.46 % of output + 12 $\mu$ V	
	(0.5 to 1.2) MHz	0.47 % of output + 14 $\mu$ V	
(1.2 to 2) MHz	0.47 % of output + 14 $\mu$ V		
(2 to 12) MHz	0.49 % of output + 14 $\mu$ V		
(12 to 20) MHz	0.55 % of output + 14 $\mu$ V		
(20 to 30) MHz	0.89% of output + 14 $\mu$ V		
(33 to 110) mV			
(10 to 30 Hz)	0.52 % of output + 30 $\mu$ V		
30 Hz to 500 kHz	0.46 % of output + 30 $\mu$ V		
(0.5 to 1.2) MHz	0.47 % of output + 33 $\mu$ V		
(1.2 to 2) MHz	0.47 % of output + 33 $\mu$ V		
(2 to 12) MHz	0.49 % of output + 33 $\mu$ V		
(12 to 20) MHz	0.55 % of output + 33 $\mu$ V		
(20 to 30) MHz	0.89 % of output + 33 $\mu$ V		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Generate <sup>1</sup> Wideband Absolute	(110 to 330) mV (10 to 30 Hz) 30 Hz to 500 kHz (0.5 to 1.2) MHz (1.2 to 2) MHz (2 to 12) MHz (12 to 20) MHz (20 to 30) MHz 0.33 to 1.1 V (10 to 30 Hz) 30 Hz to 500 kHz (0.5 to 1.2) MHz (1.2 to 2) MHz (2 to 12) MHz (12 to 20) MHz (20 to 30) MHz (1.1 to 3.5) V (10 to 30 Hz) 30 Hz to 500 kHz (0.5 to 1.2) MHz (1.2 to 2) MHz (2 to 12) MHz (12 to 20) MHz (20 to 30) MHz	0.45 % of output + 0.1 mV 0.38 % of output + 0.1 mV 0.4 % of output + 0.1 mV 0.4 % of output + 0.1 mV 0.42 % of output + 0.1 mV 0.49 % of output + 0.1 mV 0.85 % of output + 0.1 mV 0.45 % of output + 0.3 mV 0.38 % of output + 0.3 mV 0.4 % of output + 0.3 mV 0.4 % of output + 0.3 mV 0.42 % of output + 0.3 mV 0.49 % of output + 0.3 mV 0.85 % of output + 0.3 mV 0.39 % of output + 0.4 mV 0.3 % of output + 0.4 mV 0.32 % of output + 0.4 mV 0.32 % of output + 0.4 mV 0.35 % of output + 0.4 mV 0.44 % of output + 0.4 mV 0.82 % of output + 0.4 mV	Fluke 5720A Option 003 Multifunction Calibrator
AC Voltage – Measure <sup>1</sup>	(0 to 10) mV (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz 100 kHz to 1 MHz (1 to 4) MHz (4 to 8) MHz	0.03 % of reading + 3 μV 0.02 % of reading + 1.1 μV 0.03 % of reading + 1.1 μV 0.1 % of reading + 1.1 μV 0.5 % of reading + 1.1 μV 1.2 % of reading + 5 μV 7 % of reading + 7 μV 20 % of reading + 8 μV	HP 3458A Multimeter



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure <sup>1</sup>	(10 to 100) mV		HP 3458A Multimeter
	(1 to 40) Hz	0.007 % of reading + 4 μV	
	40 Hz to 1 kHz	0.007 % of reading + 2 μV	
	(1 to 20) kHz	0.014 % of reading + 2 μV	
	(20 to 50) kHz	0.03 % of reading + 2 μV	
	(50 to 100) kHz	0.08 % of reading + 2 μV	
	(100 to 300) kHz	0.3 % of reading + 10 μV	
	300 kHz to 1 MHz	1 % of reading + 10 μV	
	(1 to 2) MHz	1.5 % of reading + 70 μV	
	(2 to 4) MHz	4 % of reading + 70 μV	
	(4 to 8) MHz	4 % of reading + 80 μV	
	(8 to 10) MHz	15 % of reading + 100 μV	
	100 mV to 1 V		
	(1 to 40) Hz	0.007 % of reading + 40 μV	
	40 Hz to 1 kHz	0.07 % of reading + 20 μV	
	(1 to 20) kHz	0.014 % of reading + 20 μV	
	(20 to 50) kHz	0.03 % of reading + 20 μV	
	(50 to 100) kHz	0.08 % of reading + 20 μV	
	(100 to 300) kHz	0.3 % of reading + 100 μV	
	300 kHz to 1 MHz	1 % of reading + 100 μV	
	(1 to 2) MHz	1.5 % of reading + 0.7 mV	
	(2 to 4) MHz	4 % of reading + 0.7 mV	
	(4 to 8) MHz	4 % of reading + 0.8 mV	
	(8 to 10) MHz	15 % of reading + 1 mV	
	(1 to 10) V		
	(1 to 40) Hz	0.007 % of reading + 0.4 mV	
	40 Hz to 1 kHz	0.007 % of reading + 0.2 mV	
(1 to 20) kHz	0.014 % of reading + 0.2 mV		
(20 to 50) kHz	0.03 % of reading + 0.2 mV		
(50 to 100) kHz	0.08 % of reading + 0.2 mV		
(100 to 300) kHz	0.3 % of reading + 1 mV		
300 kHz to 1 MHz	1 % of reading + 1 mV		
(1 to 2) MHz	1.5 % of reading + 7 mV		
(2 to 4) MHz	4 % of reading + 7 mV		
(4 to 8) MHz	4 % of reading + 8 mV		
(8 to 10) MHz	15 % of reading + 10 mV		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure <sup>1</sup>	(10 to 100) V		HP 3458A Multimeter
	(1 to 40) Hz	0.02 % of reading + 4 mV	
	40 Hz to 1 kHz	0.02 % of reading + 2 mV	
	(1 to 20) kHz	0.02 % of reading + 2 mV	
	(20 to 50) kHz	0.035 % of reading + 2 mV	
	(50 to 100) kHz	0.12 % of reading + 2 mV	
	(100 to 300) kHz	0.4 % of reading + 10 mV	
	300 kHz to 1 MHz	1.5 % of reading + 10 mV	
	(100 to 700) V		
	(1 to 40) Hz	0.04 % of reading + 40 mV	
	40 Hz to 1 kHz	0.04 % of reading + 20 mV	
	(1 to 20) kHz	0.06 % of reading + 20 mV	
	(20 to 50) kHz	0.12 % of reading + 20 mV	
	(50 to 100) kHz	0.3 % of reading + 20 mV	
AC Voltage – Measure <sup>1</sup>	(1 to 10) kV	0.5 % of reading + 0.002 kV	Ross VD60 High Voltage Divider, HP 34401A Multimeter
	(60) Hz		
	(10 to 42) kV	0.5 % of reading + 0.02 kV	
	(60) Hz		
AC Current – Generate <sup>1</sup>	(9 to 200) $\mu$ A		Fluke 5720A Multifunction Calibrator
	(10 to 20) Hz	228 $\mu$ A/A + 15 nA	
	(20 to 40) Hz	152 $\mu$ A/A + 10 nA	
	40 Hz to 1 kHz	91 $\mu$ A/A + 8 nA	
	(1 to 5) kHz	266 $\mu$ A/A + 12 nA	
	(5 to 10) kHz	989 $\mu$ A/A + 61 nA	
	(0.22 to 2.2) mA		
	(10 to 20) Hz	228 $\mu$ A/A + 39 nA	
	(20 to 40) Hz	152 $\mu$ A/A + 31 nA	
	40 Hz to 1 kHz	107 $\mu$ A/A + 31 nA	
	(1 to 5) kHz	183 $\mu$ A/A + 99 nA	
	(5 to 10) kHz	989 $\mu$ A/A + 609 nA	
	(2.2 to 22) mA		
	(10 to 20) Hz	228 $\mu$ A/A + 385 nA	
	(20 to 40) Hz	152 $\mu$ A/A + 310 nA	
	40 Hz to 1 kHz	107 $\mu$ A/A + 310 nA	
	(1 to 5) kHz	183 $\mu$ A/A + 536 nA	
	(5 to 10) kHz	989 $\mu$ A/A + 4.6 $\mu$ A	



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Generate <sup>1</sup>	(22 to 220) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (0.22 to 2.2) A 20 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	228 $\mu$ A/A + 4 $\mu$ A 152 $\mu$ A/A + 3 $\mu$ A 107 $\mu$ A/A + 2 $\mu$ A 183 $\mu$ A/A + 3 $\mu$ A 989 $\mu$ A/A + 9 $\mu$ A 243 $\mu$ A/A + 31 $\mu$ A 380 $\mu$ A/A + 76 $\mu$ A 6.1 mA/A + 152 $\mu$ A	Fluke 5720A Multifunction Calibrator
AC Current – Generate <sup>1</sup>	(2.2 to 11) A 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	350 $\mu$ A/A + 141 $\mu$ A 723 $\mu$ A/A + 295 $\mu$ A 2.7 mA/A + 573 $\mu$ A	Fluke 5720A Multifunction Calibrator / Fluke 5725A Amplifier
AC Current – Generate <sup>1</sup>	(11 to 20.5) A (45 to 100) Hz 100 Hz to 1 kHz (1 Hz to 5) kHz	0.09 % of output + 5 mA 0.11 % of output + 5 mA 2.3 % of output + 5 mA	Fluke 5522A Multiproduct Calibrator
AC Current – Generate <sup>1</sup> Clamp Meters	(3.3 to 30) A (10 to 100) Hz (100 to 440) Hz (30 to 200) A 10 to 100 Hz (100 to 440) Hz	0.22 % of output + 0.028 A 0.3 % of output + 0.07 A 0.22 % of output + 0.032 A 0.79 % of output + 0.08 A	Fluke 5522A Multiproduct Calibrator / 9100-200 x10 Coil
AC Current – Generate <sup>1</sup> Clamp Meters	(16.5 to 150) A (10 to 100) Hz (100 to 440) Hz (150 to 1 000) A (10 to 100) Hz (100 to 440) Hz	0.22 % of output + 0.029 A 0.3 % of output + 0.08 A 0.22 % of output + 0.081 A 0.79 % of output + 0.20 A	Fluke 5522A Multiproduct Calibrator / 9100-200 x50 Coil
AC Current – Measure <sup>1</sup>	(5 to 100) $\mu$ A (10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz (0.1 to 1) mA (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz	0.4 % of reading + 0.03 $\mu$ A 0.15 % of reading + 0.03 $\mu$ A 0.06 % of reading + 0.03 $\mu$ A 0.4 % of reading + 0.2 $\mu$ A 0.15 % of reading + 0.2 $\mu$ A 0.06 % of reading + 0.2 $\mu$ A 0.03 % of reading + 0.2 $\mu$ A	Agilent 3458A Multimeter



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Measure <sup>1</sup>	(1 to 10) mA		Agilent 3458A Multimeter
	(10 to 20) Hz	0.4 % of reading + 2 μA	
	(20 to 45) Hz	0.15 % of reading + 2 μA	
	(45 to 100) Hz	0.06 % of reading + 2 μA	
	100 Hz to 5 kHz	0.03 % of reading + 2 μA	
	(10 to 100) mA		
	(10 to 20) Hz	0.4 % of reading + 20 μA	
	(20 to 45) Hz	0.15 % of reading + 20 μA	
	(45 to 100) Hz	0.06 % of reading + 20 μA	
	100 Hz to 5 kHz	0.03 % of reading + 20 μA	
	(0.1 to 1 A)		
	(10 to 20) Hz	0.4 % of reading + 0.2 mA	
	(20 to 45) Hz	0.15 % of reading + 0.2 mA	
	(45 to 100) Hz	0.06 % of reading + 0.2 mA	
	100 Hz to 5 kHz	0.03 % of reading + 0.2 mA	
AC Current – Measure <sup>1</sup>	(1 to 3) A		Fluke 8845A Multimeter
	(3 to 5) Hz	0.84 % of reading + 1.4 mA	
	(5 to 10) Hz	0.27 % of reading + 1.4 mA	
	10 Hz to 5 kHz	0.11 % of reading + 1.4 mA	
	(5 to 10) kHz	0.27 % of reading + 16 mA	
	(3 to 10) A		
	(3 to 5) Hz	0.84 % of reading + 4.6 mA	
	(5 to 10) Hz	0.27 % of reading + 4.6 mA	
	10 Hz to 5 kHz	0.11 % of reading + 4.6 mA	
	(5 to 10) kHz	0.27 % of reading + 53 mA	
AC Current – Measure <sup>1</sup>	(10 to 30) A		Agilent 3458A Multimeter, Keysight, 34330A Current Shunt
	40 Hz to 1 kHz	0.3 % of reading + 0.07 A	
	(1 to 5) kHz	5 % of reading + 0.14 A	
Low Frequency Power – Generate <sup>1</sup>	Up to 20 kW (45 to 65) Hz, 0 PF DC	0.44 % of reading 0.13 % of reading	Fluke 5522A Multiproduct Calibrator
Electrical Calibration of Thermocouple Indicators <sup>1</sup>	Type B		Fluke 7526A Process Calibrator
	(600 to 800) °C	0.27 °C	
	(800 to 1 550) °C	0.22 °C	
	(1 550 to 1 820) °C	0.17 °C	
	Type C		
	(0 to 1 000) °C	0.13 °C	
	(1 000 to 1 800) °C	0.18 °C	
	(1 800 to 2 000) °C	0.2 °C	
(2 000 to 2 316) °C	0.27 °C		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Calibration of Thermocouple Indicators <sup>1</sup>	Type E		Fluke 7526A Process Calibrator
	(-250 to -200) °C	0.19 °C	
	(-200 to -100) °C	0.1 °C	
	(-100 to 0) °C	0.07 °C	
	(0 to 600) °C	0.07 °C	
	(600 to 1 000) °C	0.08 °C	
	Type J		
	(-210 to -100) °C	0.11 °C	
	(-100 to 800) °C	0.07 °C	
	(800 to 1 200) °C	0.08 °C	
	Type K		
	(-250 to -200) °C	0.35 °C	
	(-200 to -100) °C	0.13 °C	
	(-100 to 800) °C	0.08 °C	
	(800 to 1 372) °C	0.1 °C	
	Type L		
	(-200 to -100) °C	0.08 °C	
	(-100 to 900) °C	0.07 °C	
	Type N		
	(-250 to -200) °C	0.56 °C	
	(-200 to -100) °C	0.18 °C	
	(-100 to 0) °C	0.1 °C	
	(0 to 100) °C	0.09 °C	
	(100 to 800) °C	0.08 °C	
	(800 to 1 300) °C	0.1 °C	
	Type R		
	(-50 to -25) °C	0.42 °C	
	(-25 to 0) °C	0.34 °C	
(0 to 100) °C	0.3 °C		
(100 to 400) °C	0.22 °C		
(400 to 600) °C	0.17 °C		
(600 to 1 000) °C	0.16 °C		
(1 000 to 1 600) °C	0.15 °C		
(1 600 to 1 767) °C	0.18 °C		
Type S			
(50 to -25) °C	0.39 °C		
(-25 to 0) °C	0.33 °C		
(0 to 100) °C	0.29 °C		
(100 to 400) °C	0.22 °C		
(400 to 600) °C	0.18 °C		
(600 to 1 600) °C	0.17 °C		
(1 600 to 1 767) °C	0.2 °C		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Calibration of Thermocouple Indicators <sup>1</sup>	Type T (-250 to -200) °C (-200 to -100) °C (-100 to 0) °C (0 to 400) °C Type U (-200 to 0) °C (0 to 600) °C	0.26 °C 0.13 °C 0.09 °C 0.07 °C 0.13 °C 0.08 °C	Fluke 7526A Process Calibrator
Electrical Calibration of RTD Indicating Devices <sup>1</sup>	Pt 385, 100 Ω (-200 to 800) °C Pt 3926, 100 Ω (-200 to 630) °C Pt 3916, 100 Ω (-200 to 630) °C Pt 385, 200 Ω (-200 to 400) °C (400 to 630) °C Pt 385, 500 Ω (-200 to 630) °C Pt 385, 1 000 Ω (-200 to 630) °C	0.05 °C 0.05 °C 0.05 °C 0.4 °C 0.5 °C 0.17 °C 0.09 °C	Fluke 7526A Process Calibrator
Oscilloscopes Calibration <sup>1</sup> – Generate			
Voltage DC - 50Ω	(1 to 24.999) mV (25 to 109.99) mV (110mV to 2.1999) V (2.2 to 6.6) V	0.19 % of Output + 31 μV 0.19 % of Output + 36 μV 0.19 % of Output + 87 μV 0.19 % of Output + 0.6 mV	Fluke 5820A Oscilloscope Calibrator w/ GHz Option
DC - 1MΩ	(1 to 24.999) mV (25 to 109.99) mV (110mV to 2.1999) V (2.2 to 10.999) V (11 to 130) V	0.019 % of reading + 20 μV 0.019 % of reading + 25 μV 0.019 % of reading + 76 μV 0.019 % of reading + 0.6 mV 0.019 % of reading + 6 mV	
Square Wave 10 Hz to 10 kHz - 50Ω	(1 to 24.999) mVpp (25 to 109.99) mVpp (110mV to 2.1999) Vpp (2.2 to 6.6) Vpp	0.19 % of Output + 31 μV 0.19 % of Output + 36 μV 0.19 % of Output + 87 μV 0.19 % of Output + 0.6 mV	



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Oscilloscopes Calibration <sup>1</sup> – Generate  Square Wave 10 Hz to 1 kHz – 1 MΩ  Square Wave (1 to 10) kHz – 1 MΩ	(1 to 24.999) mV (25 to 109.99) mV 110 mV to 2.199 9 V (2.2 to 10.999) V (11 to 130) V  (1 to 24.999) mV (25 to 109.99) mV 110mV to 2.199 9 V (2.2 to 10.999) V (11 to 130) V	0.038 % of reading + 4 μV 0.038 % of reading + 9 μV 0.038 % of reading + 60 μV 0.038 % of reading + 0.6 mV 0.038 % of reading + 6 mV  0.19 % of reading + 31 μV 0.19 % of reading + 36 μV 0.19 % of reading + 87 μV 0.19 % of reading + 0.6 mV 0.19 % of reading + 6 mV	Fluke 5820A Oscilloscope Calibrator w/ GHz Option
Oscilloscopes Calibration – Generate <sup>1</sup>  Leveled Sine Flatness 50 kHz to 10 MHz Reference	3 dB Bandwidth (5 to 50) mVpp 50 kHz to 100) MHz (100 to 300) MHz (300 to 500) MHz (500 to 600) MHz (600 to 16 00) MHz (1 600 to 2 100) MHz	0.34 dB 0.36 dB 0.42 dB 0.46 dB 0.5 dB 0.56 dB	Fluke 5820A Oscilloscope Calibrator w/ GHz Option
Oscilloscopes Calibration – Generate <sup>1</sup>  Leveled Sine Flatness 50 kHz to 10 MHz Reference	3 dB Bandwidth 50 mV to 3.5 Vpp 50 kHz to 100 MHz (100 to 300) MHz (300 to 500) MHz (500 to 600) MHz (600 to 16 00) MHz (1 600 to 2 100) MHz (3.5 to 5) Vpp 50 kHz to 100 MHz (100 to 300) MHz (300 to 500) MHz (500 to 600) MHz	0.24 dB 0.24 dB 0.32 dB 0.34 dB 0.4 dB 0.44 dB  0.24 dB 0.24 dB 0.32 dB 0.34 dB	Fluke 5820A Oscilloscope Calibrator w/ GHz Option
Oscilloscopes Calibration – Generate <sup>1</sup>  Leveled Sine Flatness	3 dB Bandwidth 50 mV to 3.5Vpp (2 100 to 4 000) MHz (4 000 to 8 000) MHz (8 000 to 18 000) MHz	0.25 dB 0.35 dB 0.46 dB	EPM Power Meter w/ E Series Power Sensors
Oscilloscopes Calibration – Generate <sup>1</sup>  Time Marker	500 ps to 20 ms 50 ms to 5 s	0.25 μs/s 1.9 μs/s + 3.8 μHz	Fluke 5820A Oscilloscope Calibrator w/ GHz Option



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Oscilloscopes Calibration – Measure <sup>1</sup> Input Impedance Resistance Capacitance Leakage	(40 to 60) Ω 500 kΩ to 1.5MΩ (5 to 50) pF (0 to 5.99) V	0.08 % of reading 0.08 % of reading 3.8 % of reading + 0.4pF 0.038 % of reading + 0.8 mV	Fluke 5820A Oscilloscope Calibrator w/ GHz Option
Fundamental AC Voltage – Generate <sup>1</sup>	(15 to 17) V (45 to 65) Hz (0 to 23) V (45 to 65) Hz (16 to 850) Hz (28 to 32) V (45 to 65) Hz (23 to 45) V (45 to 65) Hz (16 to 850) Hz (56 to 64) V (45 to 65) Hz (45 to 90) V (45 to 65) Hz (16 to 850) Hz (110 to 128) V (45 to 65) Hz (90 to 180) V (45 to 65) Hz (16 to 850) Hz (215 to 246) V (45 to 65) Hz (180 to 360) V (45 to 65) Hz (16 to 850) Hz (425 to 490) V (45 to 65) Hz (360 to 650) V (45 to 65) Hz (16 to 850) Hz	42 μV/V 42 μV/V + 0.2 mV 60 μV/V + 0.2 mV 42 μV/V 42 μV/V + 0.4 mV 60 μV/V + 0.4 mV 42 μV/V 42 μV/V + 0.8 mV 60 μV/V + 0.8 mV 44 μV/V 44 μV/V + 1.6 mV 60 μV/V + 1.6 mV 44 μV/V 60 μV/V + 3.2 mV 61 μV/V + 3.2 mV 44 μV/V 60 μV/V + 5.8 mV 61 μV/V + 5.8 mV	Fluke 6105A Power Quality Calibrator
Fundamental AC Voltage – Generate <sup>1</sup>	(740 to 850) V (45 to 65) Hz (650 to 1 008) V (45 to 65) Hz (16 to 850) Hz	44 μV/V 60 μV/V + 10 mV 61 μV/V + 10 mV	Fluke 6105A Power Quality Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Additive DC Offset Voltage – Generate <sup>1</sup>  Additive signal to the Fundamental AC Voltage, up to 50% of range	(0 to 11.5) V (0 to 22.5) V (0 to 45) V (0 to 90) V (0 to 180) V (0 to 325) V (0 to 504) V	91 $\mu\text{V/V} + 2 \text{ mV}$ 91 $\mu\text{V/V} + 4 \text{ mV}$ 91 $\mu\text{V/V} + 8 \text{ mV}$ 91 $\mu\text{V/V} + 16 \text{ mV}$ 91 $\mu\text{V/V} + 32 \text{ mV}$ 92 $\mu\text{V/V} + 60 \text{ mV}$ 92 $\mu\text{V/V} + 100 \text{ mV}$	Fluke 6105A Power Quality Calibrator
Additive AC Voltage Harmonics – Generate <sup>1</sup>  Additive signal to the Fundamental AC Voltage, up to 30% of range	(0 to 6.9) V (16 to 850) Hz (850 to 6000) Hz (0 to 13.5) V (16 to 850) Hz (850 to 6 000) Hz	58 $\mu\text{V/V} + 1 \text{ mV}$ 451 $\mu\text{V/V} + 1 \text{ mV}$  58 $\mu\text{V/V} + 2 \text{ mV}$ 451 $\mu\text{V/V} + 2 \text{ mV}$	Fluke 6105A Power Quality Calibrator
Additive AC Voltage Harmonics – Generate <sup>1</sup>  Additive signal to the Fundamental AC Voltage, up to 30% of range	(0 to 27) V (16 to 850) Hz (850 to 6 000) Hz (0 to 54) V (16 to 850) Hz (850 to 6 000) Hz (0 to 108) V (16 to 850) Hz (850 to 6 000) Hz (0 to 195) V (16 to 850) Hz (850 to 6 000) Hz (0 to 302) V (16 to 850) Hz (850 to 6 000) Hz	60 $\mu\text{V/V} + 2.2 \text{ mV}$ 451 $\mu\text{V/V} + 2.2 \text{ mV}$  60 $\mu\text{V/V} + 4.4 \text{ mV}$ 451 $\mu\text{V/V} + 4.4 \text{ mV}$  60 $\mu\text{V/V} + 12 \text{ mV}$ 451 $\mu\text{V/V} + 12 \text{ mV}$  61 $\mu\text{V/V} + 22 \text{ mV}$ 451 $\mu\text{V/V} + 22 \text{ mV}$  61 $\mu\text{V/V} + 33 \text{ mV}$ 451 $\mu\text{V/V} + 33 \text{ mV}$	Fluke 6105A Power Quality Calibrator
Fundamental AC Current – Generate <sup>1</sup>	Up to 0.25 A (45 to 65) Hz (16 to 850) Hz (0.25 to 0.5) A (45 to 65) Hz (16 to 850) Hz	46 $\mu\text{A/A} + 2.5 \mu\text{A}$ 60 $\mu\text{A/A} + 5 \mu\text{A}$  46 $\mu\text{A/A} + 5 \mu\text{A}$ 61 $\mu\text{A/A} + 10 \mu\text{A}$	Fluke 6105A Power Quality Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment		
Fundamental AC Current – Generate <sup>1</sup>	(0.5 to 1) A (45 to 65) Hz (16 to 850) Hz	47 $\mu$ A/A + 10 $\mu$ A 61 $\mu$ A/A + 20 $\mu$ A	Fluke 6105A Power Quality Calibrator		
	(1 to 2) A (45 to 65) Hz (16 to 850) Hz	46 $\mu$ A/A + 20 $\mu$ A 61 $\mu$ A/A + 40 $\mu$ A			
	(2 to 5) A (45 to 65) Hz (16 to 850) Hz	49 $\mu$ A/A + 50 $\mu$ A 64 $\mu$ A/A + 100 $\mu$ A			
	(5 to 10) A (45 to 65) Hz (16 to 850) Hz	49 $\mu$ A/A + 100 $\mu$ A 65 $\mu$ A/A + 200 $\mu$ A			
	(10 to 21) A (45 to 65) Hz (16 to 850) Hz	49 $\mu$ A/A + 200 $\mu$ A 69 $\mu$ A/A + 400 $\mu$ A			
	(21 to 50) A (45 to 65) Hz (40 to 850) Hz	49 $\mu$ A/A + 500 $\mu$ A 74 $\mu$ A/A + 1 mA			
	(50 to 80) A (40 to 450) Hz (450 to 850) Hz	106 $\mu$ A/A + 2.8 mA 118 $\mu$ A/A + 2.8 mA			
	Additive DC Offset Current – Generate <sup>1</sup>  Additive signal to the Fundamental AC Current, up to 50% of range	(0 to 0.125) A		89 $\mu$ A/A + 25 $\mu$ A	Fluke 6105A Power Quality Calibrator
		(0 to 0.25) A		89 $\mu$ A/A + 50 $\mu$ A	
		(0 to 0.5) A		89 $\mu$ A/A + 100 $\mu$ A	
		(0 to 1) A		89 $\mu$ A/A + 200 $\mu$ A	
		(0 to 2.5) A		89 $\mu$ A/A + 500 $\mu$ A	
		(0 to 5) A (0 to 10) A		89 $\mu$ A/A + 1 mA 90 $\mu$ A/A + 2 mA	
	Additive AC Current Harmonics – Generate <sup>1</sup>  Additive signal to the Fundamental AC Current, up to 30% of range	Up to 0.075 A (16 to 850) Hz (850 to 6 000) Hz		61 $\mu$ A/A + 5 $\mu$ A 400 $\mu$ A/A + 5 $\mu$ A	Fluke 6105A Power Quality Calibrator
Up to 0.15 A (16 to 850) Hz (850 to 6 000) Hz		61 $\mu$ A/A + 10 $\mu$ A 400 $\mu$ A/A + 10 $\mu$ A			
Up to 0.3 A (16 to 850) Hz (850 to 6 000) Hz		61 $\mu$ A/A + 20 $\mu$ A 400 $\mu$ A/A + 20 $\mu$ A			
Up to 0.6 A (16 to 850) Hz (850 to 6 000) Hz		61 $\mu$ A/A + 40 $\mu$ A 400 $\mu$ A/A + 40 $\mu$ A			



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment		
Additive AC Current Harmonics – Generate <sup>1</sup>  Additive signal to the Fundamental AC Current, up to 30% of range	Up to 1.5 A (16 to 850) Hz (850 to 6 000) Hz	61 $\mu\text{A}/\text{A}$ + 100 $\mu\text{A}$ 400 $\mu\text{A}/\text{A}$ + 100 $\mu\text{A}$	Fluke 6105A Power Quality Calibrator		
	Up to 3 A (16 to 850) Hz (850 to 6 000) Hz	64 $\mu\text{A}/\text{A}$ + 200 $\mu\text{A}$ 400 $\mu\text{A}/\text{A}$ + 200 $\mu\text{A}$			
	Up to 6 A (16 to 850) Hz (850 to 6 000) Hz	65 $\mu\text{A}/\text{A}$ + 400 $\mu\text{A}$ 400 $\mu\text{A}/\text{A}$ + 400 $\mu\text{A}$			
	Up to 15 A (16 to 850) Hz (850 to 3000) Hz	69 $\mu\text{A}/\text{A}$ + 1 mA 400 $\mu\text{A}/\text{A}$ + 1 mA			
	Up to 24 A (16 to 850) Hz (850 to 3 000) Hz	112 $\mu\text{A}/\text{A}$ + 2 mA 400 $\mu\text{A}/\text{A}$ + 2 mA			
	Current to Voltage Phase Offsets (regardless of voltage range)	(-180 to +180) <sup>o</sup> or (0 to 360) <sup>o</sup>			Fluke 6105A Power Quality Calibrator
		Up to 50 A (45 to 65) Hz		0.002 3 <sup>o</sup>	
(16 to 69) Hz		0.003 <sup>o</sup>			
(69 to 180) Hz		0.007 <sup>o</sup>			
(180 to 450) Hz		0.018 <sup>o</sup>			
(450 to 850) Hz		0.033 <sup>o</sup>			
(850 to 3 000) Hz		0.12 <sup>o</sup>			
(3 000 to 6 000) kHz		0.23 <sup>o</sup>			
Up to 80 A (45 to 65) Hz		0.003 <sup>o</sup>			
(16 to 69) Hz		0.003 <sup>o</sup>			
(69 to 180) Hz		0.008 <sup>o</sup>			
(180 to 450) Hz		0.025 <sup>o</sup>			
(450 to 850) Hz		0.05 <sup>o</sup>			
(850 to 3 000) Hz		0.25 <sup>o</sup>			



ANSI National Accreditation Board

Electrical – RF/Microwave

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Attenuation – Measure <sup>1</sup>	(30 to 3 050) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 526 Power Sensor
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.12 dB	
	(90 to 100) dB	0.13 dB	
	(100 to 110) dB	0.13 dB	
	(110 to 120) dB	0.26 dB	
	(3 050 to 6 600) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.12 dB	
	(90 to 100) dB	0.13 dB	
	(100 to 110) dB	0.15 dB	
	(110 to 120) dB	0.37 dB	
	(6 600 to 13 200) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.12 dB	
(90 to 100) dB	0.14 dB		
(100 to 110) dB	0.34 dB		
(110 to 120) dB	0.77 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment	
RF Attenuation – Measure <sup>1</sup>	(13 200 to 19 200) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 526 Power Sensor	
	(0 to 10) dB	0.02 dB		
	(10 to 20) dB	0.025 dB		
	(20 to 30) dB	0.03 dB		
	(30 to 40) dB	0.035 dB		
	(40 to 50) dB	0.04 dB		
	(50 to 60) dB	0.076 dB		
	(60 to 70) dB	0.081 dB		
	(70 to 80) dB	0.12 dB		
	(80 to 90) dB	0.12 dB		
	(90 to 100) dB	0.27 dB		
	(100 to 110) dB	0.66 dB		
	(19 200 to 26 500) MHz			
	(0 to 10) dB	0.02 dB		
	(10 to 20) dB	0.025 dB		
	(20 to 30) dB	0.03 dB		
	(30 to 40) dB	0.035 dB		
	(40 to 50) dB	0.04 dB		
(50 to 60) dB	0.076 dB			
(60 to 70) dB	0.081 dB			
(19 200 to 26 500) MHz	(70 to 80) dB	0.12 dB		
	(80 to 90) dB	0.2 dB		
	(90 to 100) dB	0.5 dB		
	(100 to 110) dB	1.1 dB		
	RF Flatness – Measure <sup>1</sup>	9 kHz to 2 000 MHz		Agilent EPM Series Power Meter w/E9304A H18 Power Sensor
		(20 to -10) dBm	0.1 dB	
(-10 to -30) dBm		0.1 dB		
(-30 to -40) dBm		0.11 dB		
(-40 to -42) dBm		0.12 dB		
(2 to 14) GHz				
(20 to -10) dBm		0.1 dB		
(-10 to -30) dBm		0.09 dB		
(-30 to -40) dBm		0.1 dB		
(-40 to -42) dBm		0.11 dB		
(14 to 18) GHz				
(20 to -10) dBm		0.11 dB		
(-10 to -30) dBm	0.12 dB			
(-30 to -40) dBm	0.12 dB			
(-40 to -42) dBm	0.13 dB			



ANSI National Accreditation Board

Electrical – RF/Microwave

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Measure <sup>1</sup>	100 kHz to 30 MHz		Agilent N5531S Measuring Receiver N1912A w/E9304A Power Sensor
	(20 to 0) dB	0.12 dB	
	(0 to -58) dB	0.13 dB	
	(-58 to -78) dB	0.15 dB	
	(-78 to -110) dB	0.18 dB	
	(-110 to -115) dB	0.21 dB	
	(-115 to -120) dB	0.28 dB	
	(-120 to -125) dB	0.43 dB	
	(30 to 2 000) MHz		
	(30 to 20) dB	0.36 dB	
	(20 to 0) dB	0.2 dB	
	(0 to -58) dB	0.22 dB	
	(-58 to -78) dB	0.23 dB	
	(-78 to -110) dB	0.25 dB	
	(-110 to -115) dB	0.27 dB	
	(-115 to -120) dB	0.33 dB	
	(-120 to -125) dB	0.46 dB	
	(2 000 to 3 050) MHz		
	(30 to 20) dB	0.42 dB	
	(20 to 0) dB	0.3 dB	
	(0 to -58) dB	0.31 dB	
	(-58 to -78) dB	0.32 dB	
	(-78 to -110) dB	0.34 dB	
	(-110 to -115) dB	0.35 dB	
	(-115 to -120) dB	0.4 dB	
	(-120 to -125) dB	0.51 dB	
	(3 050 to 6 600) MHz		
	(30 to 20) dB	0.42 dB	
	(20 to 0) dB	0.3 dB	
	(0 to -58) dB	0.31 dB	
(-58 to -78) dB	0.32 dB		
(-78 to -110) dB	0.34 dB		
(-110 to -115) dB	0.38 dB		
(-115 to -120) dB	0.48 dB		
(-120 to -125) dB	0.64 dB		

Electrical – RF/Microwave

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Measure <sup>1</sup>	(6 600 to 13 200) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 526 Power Sensor
	(30 to 20) dB	0.42 dB	
	(20 to 0) dB	0.3 dB	
	(0 to -58) dB	0.31 dB	
	(-58 to -78) dB	0.32 dB	
	(-78 to -100) dB	0.34 dB	
	(-100 to -105) dB	0.37 dB	
	(-105 to -110) dB	0.45 dB	
	(-110 to -115) dB	0.60 dB	
	(-115 to -120) dB	0.82 dB	
	(13 200 to 18 000) MHz		
	(30 to 20) dB	0.42 dB	
	(20 to 0) dB	0.3 dB	
	(0 to -58) dB	0.31 dB	
	(-58 to -78) dB	0.32 dB	
	(-78 to -90) dB	0.33 dB	
	(-90 to -95) dB	0.35 dB	
	(-95 to -100) dB	0.41 dB	
	(-100 to -105) dB	0.53 dB	
	(-105 to -110) dB	0.72 dB	
	(18 000 to 19 200) MHz		
	(30 to 20) dB	0.48 dB	
	(20 to 0) dB	0.38 dB	
	(0 to -58) dB	0.39 dB	
	(-58 to -78) dB	0.4 dB	
	(-78 to -90) dB	0.41 dB	
	(-90 to -95) dB	0.42 dB	
	(-95 to -100) dB	0.47 dB	
	(-100 to -105) dB	0.58 dB	
	(-105 to -110) dB	0.75 dB	
	(19 200 to 26 500) MHz		
	(30 to 20) dB	0.48 dB	
	(20 to 0) dB	0.38 dB	
	(0 to -58) dB	0.39 dB	
	(-58 to -78) dB	0.4 dB	
	(-78 to -90) dB	0.43 dB	
(-90 to -95) dB	0.50 dB		
(-95 to -100) dB	0.63 dB		
(-100 to -105) dB	0.84 dB		
(-105 to -110) dB	1.1 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Amplitude Modulation – Measure <sup>1</sup>	100 kHz to 10 MHz Rate 50 Hz to 10 kHz (5 to 99) % Depth	0.75 % of reading + 0.3 digits	Agilent N5531S Measuring Receiver
	10 MHz to 3 GHz Rate 50 Hz to 100 kHz (5 to 20) % Depth (20 to 99) % Depth	2.5 % of reading + 0.4 digits 0.5 % of reading + 0.4 digits	
	(3 to 26.5) GHz Rate 50 Hz to 100 kHz (5 to 20) % Depth (20 to 99) % Depth	4.5 % of reading + 0.4 digits 1.5 % of reading + 0.4 digits	
Frequency Modulation – Measure <sup>1</sup>  $\beta$ = deviation / rate	250 kHz to 10 MHz Rates 20 Hz to 10 kHz Peak Dev 200 to 40 kHz	$\beta > 0.2$ - 1.5 % of reading + 2 Hz $\beta > 1.2$ - 1 % of reading + 2 Hz	Agilent N5531S Measuring Receiver
	10 MHz to 6.6 GHz Rates 50 Hz to 200 kHz Peak Dev 250 to 400 kHz (6.6 to 13.2) GHz	$\beta > 0.20$ - 1.5 % of reading + 2 Hz $\beta > 0.45$ - 1 % of reading + 2 Hz	
	Rates 50 Hz to 200 kHz Peak Dev 250 to 400 kHz (13.2 to 26.5) GHz	$\beta > 0.2$ - 2.5 % of reading + 4 Hz $\beta > 8.0$ - 1 % of reading + 4 Hz	
	Rates 50 Hz to 200 kHz Peak Dev 250 to 400 kHz	$\beta > 0.2$ - 3.8 % of reading + 9 Hz $\beta > 16$ - 1 % of reading + 9 Hz	
Phase Modulation – Measure <sup>1</sup>	100 kHz to 6.6 GHz Deviations > 0.3 rad Deviations > 0.7 rad	3 % of reading + 0.002 rad 1 % of reading + 0.002 rad	Agilent N5531S Measuring Receiver
	(6.6 to 13.2) GHz Deviations > 0.6 rad Deviations > 2.0 rad	3 % of reading + 0.005 rad 1 % of reading + 0.005 rad	
	(13.2 to 26.5) GHz Deviations: > 1.2 rad Deviations > 4.0 rad	3 % of reading + 0.009 rad 1 % of reading + 0.009 rad	
RF Power – Generate <sup>1</sup>	(30 to 2 000) MHz (20 to 0) dB (0 to -58) dB (-58 to -78) dB (-78 to -110) dB (-110 to -120) dB	0.26 dB 0.28 dB 0.29 dB 0.3 dB 0.37 dB	Agilent N5531S Measuring Receiver w/N5532A Opt 526 Power Sensor, 83630B Signal Generator



ANSI National Accreditation Board

Electrical – RF/Microwave

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Generate <sup>1</sup>	(2 000 to 3 050) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 526 Power Sensor, 83630B Signal Generator
	(20 to 0) dB	0.37 dB	
	(0 to -58) dB	0.38 dB	
	(-58 to -78) dB	0.38 dB	
	(-78 to -110) dB	0.4 dB	
	(-110 to -120) dB	0.52 dB	
	(3 050 to 6 600) MHz		
	(20 to 0) dB	0.37 dB	
	(0 to -58) dB	0.38 dB	
	(-58 to -78) dB	0.38 dB	
	(-78 to -110) dB	0.4 dB	
	(-110 to -120) dB	0.52 dB	
	(6 600 to 13 200) MHz		
	(20 to 0) dB	0.37 dB	
	(0 to -58) dB	0.38 dB	
	(-58 to -78) dB	0.38 dB	
	(-78 to -110) dB	0.49 dB	
	(13 200 to 18 000) MHz		
	(20 to 0) dB	0.37 dB	
	(0 to -58) dB	0.38 dB	
(-58 to -78) dB	0.38 dB		
(-78 to -110) dB	0.75 dB		
(18 000 to 19 200) MHz			
(20 to 0) dB	0.49 dB		
(0 to -58) dB	0.5 dB		
(-58 to -78) dB	0.5 dB		
(-78 to -110) dB	0.81 dB		
(19 200 to 26 500) MHz			
(20 to 0) dB	0.49 dB		
(0 to -58) dB	0.5 dB		
(-58 to -78) dB	0.5 dB		
(-78 to -110) dB	1.2 dB		
AM Distortion Measure <sup>1</sup> Rate 20 Hz to 1 kHz	(0.1 to 10) MHz		Agilent N5531S Measuring Receiver
	AM Depth > 1%		
	(0 to -20) dB	1.2 dB	
	(-20 to -30) dB	2.2 dB	
	AM Depth > 3%		
	(0 to -20) dB	1 dB	
(-20 to -30) dB	1.3 dB		
(-30 to -40) dB	2.4 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AM Distortion Measure <sup>1</sup> Rate 20 Hz to 1 kHz	10 MHz to 26.5 GHz AM Depth > 1% (0 to -20) dB (-20 to -30) dB AM Depth > 3% (0 to -20) dB (-20 to -30) dB (-30 to -40) dB	1.3 dB 2.5 dB 1.1 dB 1.4 dB 3 dB	Agilent N5531S Measuring Receiver
FM Distortion Measure <sup>1</sup> Rate 20 Hz to 1 kHz	(1 to 6 600) MHz Dev 500 Hz to 2KHz (0 to -20) dB (-20 to -30) dB (-30 to -40) dB Dev > 2 kHz (0 to -20) dB (-20 to -30) dB (-30 to -40) dB (-40 to -50) dB (6.6 to 13.2) GHz Dev > 2.3 kHz (0 to -20) dB (-20 to -30) dB (-30 to -40) dB Dev > 4.5 kHz (0 to -20) dB (-20 to -30) dB (-30 to -40) dB (-40 to -50) dB (13.2 to 26.5) GHz Dev > 2.7 kHz (0 to -20) dB (-20 to -30) dB (-30 to -40) dB Dev > 6.0 kHz (0 to -20) dB (-20 to -30) dB (-30 to -40) dB (-40 to -50) dB	0.26 dB 0.79 dB 2.3 dB 0.09 dB 0.27 dB 0.83 dB 2.4 dB 0.26 dB 0.79 dB 2.3 dB 0.09 dB 0.27 dB 0.83 dB 2.4 dB 0.26 dB 0.79 dB 2.3 dB 0.09 dB 0.27 dB 0.83 dB 2.4 dB	Agilent N5531S Measuring Receiver

Electrical – RF/Microwave

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
PM Distortion Measure <sup>1</sup>	(1 to 6 600) MHz		
	Rate (20 to 500) Hz		
	Dev > 0.8 rad		
	(0 to -20) dB	0.26 dB	
	(-20 to -30) dB	0.79 dB	
	(-30 to -40) dB	2.3 dB	
	Dev > 2.5 rad		
	(0 to -20) dB	0.09 dB	
	(-20 to -30) dB	0.27 dB	
	(-30 to -40) dB	0.83 dB	
	(-40 to -50) dB	2.3 dB	
	Rate (500 to 1 000) Hz		
	Dev > 0.4 rad		
	(0 to -20) dB	0.26 dB	
	(-20 to -30) dB	0.79 dB	
	(-30 to -40) dB	2.3 dB	
	Dev > 1.0 rad		
	(0 to -20) dB	0.09 dB	
	(-20 to -30) dB	0.27 dB	
	(-30 to -40) dB	0.83 dB	
	(-40 to -50) dB	2.3 dB	
	(6.6 to 13.2) GHz		
	Rate (20 to 500) Hz		
	Dev > 1.8 rad		
(0 to -20) dB	0.26 dB		
(-20 to -30) dB	0.79 dB		
(-30 to -40) dB	2.3 dB		
Dev > 5.5 rad			
(0 to -20) dB	0.09 dB		
(-20 to -30) dB	0.27 dB		
(-30 to -40) dB	0.83 dB		
(-40 to -50) dB	2.3 dB		
Rate (500 to 1 000) Hz			
Dev > 0.8 rad			
(0 to -20) dB	0.26 dB		
(-20 to -30) dB	0.79 dB		
(-30 to -40) dB	2.3 dB		
Dev > 2.5 rad			
(0 to -20) dB	0.09 dB		
(-20 to -30) dB	0.27 dB		
(-30 to -40) dB	0.83 dB		
(-40 to -50) dB	2.3 dB		

Agilent N5531S  
Measuring Receiver



ANSI National Accreditation Board

Electrical – RF/Microwave

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
PM Distortion Measure <sup>1</sup>	(13.2 to 26.5) GHz		Agilent N5531S Measuring Receiver
	Rate (20 to 500) Hz		
	Dev > 3.5 rad	0.26 dB	
	(0 to -20) dB	0.79 dB	
	(-20 to -30) dB	2.3 dB	
	(-30 to -40) dB		
	Dev > 10.0 rad	0.09 dB	
	(0 to -20) dB	0.27 dB	
	(-20 to -30) dB	0.83 dB	
	(-30 to -40) dB	2.3 dB	
	Rate (500 to 1 000) Hz		
	Dev > 3.0 rad	0.26 dB	
	(0 to -20) dB	0.79 dB	
	(-20 to -30) dB	2.3 dB	
(-30 to -40) dB			
Dev > 8.0 rad	0.09 dB		
(0 to -20) dB	0.27 dB		
(-20 to -30) dB	0.83 dB		
(-30 to -40) dB	2.3 dB		
(-40 to -50) dB			
Total Harmonic Distortion (THD)	(0 to -60) dB		HP 8903B Audio Analyzer
	20 Hz to 20 kHz		
	(0 to -40) dB	1 dB	
	(-40 to -50) dB	1 dB	
	(-50 to -60) dB	1.3 dB	
	(-60 to -65) dB	1.7 dB	
	(20 to 50) kHz		
	(0 to -40) dB	2 dB	
	(-40 to -50) dB	2.1 dB	
	(-50 to -60) dB	3 dB	
	(50 to 100) kHz		
	(0 to -40) dB	2 dB	
	(-40 to -50) dB	2.4 dB	

**Electrical – RF/Microwave**

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Harmonics Measure <sup>1</sup>	(-80 to -10) dB		Agilent E4440A Measuring Receiver
	2 <sup>nd</sup> through 5 <sup>th</sup> Harmonic	0.37 dB	
	1 kHz to 600 MHz	1.1 dB	
	(600 to 1 320) MHz	1.4 dB	
	(1 320 to 2 200) MHz	1.4 dB	
	(2 200 to 3 000) MHz	1.7 dB	
	(3 000 to 4 400) MHz	1.9 dB	
	(4 400 to 5 300) MHz		
(-80 to -10) dB			
2 <sup>nd</sup> through 4 <sup>th</sup> Harmonic	2.1 dB		
(5 300 to 6 625) MHz			
2 <sup>nd</sup> through 3 <sup>rd</sup> Harmonic	2.1 dB		
(6 625 to 8 833) MHz			
2 <sup>nd</sup> Harmonic			
(8 833 to 13 250) MHz	2.1 dB		

**Length – Dimensional Metrology**

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Calipers <sup>1,2</sup>	Up to 24 in (24 to 80) in	(283 + 1.7L) μin (231 + 3.6L) μin	Grade 2 gage blocks
Micrometers <sup>1,2</sup>	Up to 46 in	(19 + 9L) μin	
Dial Indicators <sup>1,2</sup> Resolution: ≥ 50μin < 50μin	Up to 10 in Up to 0.1 in	(27 + 3.3L) μin 7 μin	
Feeler Gages <sup>1</sup>	Up to 1 in	52 μin	Pratt & Whitney Supermicrometer Model C
Height Gages <sup>1,2</sup>	Up to 46 in	(137 + 2.2L) μin	Grade 2 gage blocks
Protractors <sup>1</sup>	(0 to 360)°	0.011°	Angle blocks
Surface Plates <sup>1</sup> Overall Flatness Local Area Flatness	Up to 6 ft × 6 ft (-0.001 to 0.001) in	125 μin 68 μin	Leveling system Repeat-o-Meter
Cylindrical Gages <sup>1,2</sup> – Plain Pin, Plug Gages Ring Gages	(0 to 13) in (0 to 14) in	(4.8 + 2.3D) μin (9.3 + 1.5D) μin	Universal measuring machine, gage blocks (Grade 1)

**Length – Dimensional Metrology**

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Coating Thickness Gages <sup>1</sup> Eddy Current & Magnetic Induction, Fixed Point	(0.737 to 100) mils (100 to 243) mils	53 $\mu$ in 280 $\mu$ in	Coating thickness standards
Coating Thickness Shims <sup>1,3</sup>	Up to 243 mils	98 $\mu$ in	Micrometer
Metal Tapes and Rules <sup>1</sup>	Up to 100 ft	(0.023 + 0.000 27 <i>F</i> ) in	Standard rule
Gage Blocks <sup>2</sup>	(0 to 13) in	(2.6 + 2.4 <i>L</i> ) $\mu$ in	Universal measuring machine, master gage block set
Thread Plugs – Major Diameter Pitch Diameter, (6 to 72) TPI	Up to 12 in Up to 12 in	46 $\mu$ in 92 $\mu$ in	Gage Blocks, P&W Model C Bench Micrometer, Van Keuren thread wire set
Adjustable Thread Rings <sup>2,3</sup> Pitch Diameter (Tactile Fit)	Up to 12 in	(350 + 47 <i>D</i> ) $\mu$ in	Thread Setting Plug Gages
Radius Gages	(0.01 to 1) in	340 $\mu$ in	Optical comparator
Micrometer Standards Length Rods <sup>2</sup>	(1 to 10) in	(21 + 2.2 <i>L</i> ) $\mu$ in	Gage blocks (grade 2), P&W Model C Bench Micrometer

**Mass and Mass Related**

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Scales & Balances <sup>1,4</sup>	(1 to 500) mg 500 mg to 200 g (200 to 300) g 300 g to 3 kg	0.01 mg 10 mg 19 mg 47 mg	Class 1 weights
	(3 to 18) kg (18 to 30) kg	0.65 g 2.5 g	Class F weights & Class 2 weights
	(1 to 1 550) lb	0.01 % of reading	Class F weights
Torque Wrenches <sup>1</sup>	5 lbf·in to 1 000 lbf·ft	0.31 % of reading	CDI torque system
Torque Analyzers	5 lbf·in to 1 000 lbf·ft	0.06 % of reading	Class F weights, torque arm



ANSI National Accreditation Board

Mass and Mass Related

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Pressure Gauges	(-14.75 to 1 015) psig	0.001 7 % of reading	Fluke PG7601 Piston Gauge w/ PC-7100/7600-1, PC-7200-2 and PC-7300-5 Piston-Cylinder
	(0.75 to 1 015) psia	0.003 5 % of reading	
	(60 to 13 000) psig (72 to 13 000) psia (145 to 29 000) psig (160 to 29 000) psia	0.004 % of reading	
Pressure Gauges	(0 to 12) inH <sub>2</sub> O	0.002 5 inH <sub>2</sub> O	Dwyer 1425-25 Hook Gage
Pressure Gauges <sup>1</sup>	(-15 to 100) psi (0 to 10 000) psi	0.08 psi 3 psi	Digital gage Additel 681
Force Gages, Load Cells & Dynamometers Tension/Compression	10 mg to 10 kg (20 to 500) lb	0.028 % of reading 0.024 % of reading	Class F weights
Rockwell Hardness Testers <sup>1</sup>	(20 to 65) HRA	0.62 HRA	Indirect Verification per ASTM E18-17e1
	(70 to 78) HRA	0.58 HRA	
	(80 to 84) HRA	0.41 HRA	
	(40 to 59) HRBW	0.84 HRBW	
	(60 to 79) HRBW	0.79 HRBW	
	(80 to 100) HRBW	0.68 HRBW	
Mass - Fixed Points Metric	(20 to 30) HRC	0.57 HRC	ASTM E617 Class 0 Weights OIML Class E2 Weights Precision Mass Comparators Balances
	(35 to 55) HRC	0.5 HRC	
	(56 to 65) HRC	0.44 HRC	
	(1, 2, 3, 5, 10, 20, 30) mg	1.8 µg	
	(50, 100, 200) mg	1.8 µg	
	300 mg	2 µg	
	500 mg	1.8 µg	
	(1, 2, 5) g	5.2 µg	
	10 g	8.2 µg	
	20 g	10 µg	
	50 g	21 µg	
	100 g	41 µg	
200 g	61 µg		
500 g	0.16 mg		
1 kg	0.31 mg		
2 kg	0.61 mg		
5 kg	2.1 mg		
10 kg	3.1 mg		
20 kg	8.1 mg		
25 kg	8.1 mg		

**Mass and Mass Related**

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Mass - Fixed Points Avoirdupois	(0.001, 0.002) lb	2.5 µg	Comparison to ASTM E617 Class 1 weights  Precision Mass Comparators Balances
	0.005 lb	6.1 µg	
	0.01 lb	6.2 µg	
	0.02 lb	7.3 µg	
	0.05 lb	15 µg	
	0.1 lb	20 µg	
	0.2 lb	38 µg	
	(0.5, 1) lb	0.25 mg	
	2 lb	0.33 mg	
	5 lb	1.2 mg	
	10 lb	2.1 mg	
	20 lb	4.1 mg	
	25 lb	4.2 mg	
	50 lb	14 mg	
Durometers Scale (Force) Accuracy Types A, B, C, D, E, DO Type M, Types O, OO Indenter Geometry Length Diameter Angle	(0 to 100) duros	0.06 duros	Direct Verification
		0.09 duros	Master balance
		0.07 duros	
	0.1 in	180 µin	Optical comparator
	0.05 in	180 µin	
	(30 to 35)°	0.004°	

**Photometry and Radiometry**

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Gloss Meters <sup>1</sup>	(20, 60, 85) ° (0 to 100) GU	0.71 GU	Gloss standards

**Thermodynamic**

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Temperature – Measure	(-196 to 420) °C	0.044 °C	Hart 1502 Indicator, Burns Engineering 12005 PRT

**Thermodynamic**

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Infrared (IR) Thermometry <sup>1</sup>	(20 to 100) °C (100 to 300) °C (300 to 420) °C (420 to 500) °C	1.5 °C 4.3 °C 6 °C 7.7 °C	Fluke 9132 Infrared Calibrator $\lambda = (8 \text{ to } 14) \mu\text{m}, \epsilon = 0.95$
Humidity – Measure <sup>1</sup>	(10 to 90) %RH	1.4 %RH	Vaisala MI70 Indicator and HMP77B Probe
Relative Humidity Measuring Equipment <sup>1</sup>	(10 to 95) %RH	0.5 %RH	Thunder Scientific 2500 Humidity Chamber

**Time and Frequency**

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency – Generate <sup>1</sup>	10 MHz	$1 \times 10^{-9}$ Hz/Hz	HP Z3801A GPS Receiver
Frequency – Generate <sup>1</sup>	(1 to 10) Hz (10 to 100) Hz (100 to 1 000) Hz (1 to 10) kHz (10 to 100) kHz (0.1 to 1) MHz (1 to 10) MHz	$1 \times 10^{-9}$ Hz/Hz + 0.57 $\mu$ Hz $1 \times 10^{-9}$ Hz/Hz + 5.7 $\mu$ Hz $1 \times 10^{-9}$ Hz/Hz + 57 $\mu$ Hz $1 \times 10^{-9}$ Hz/Hz + 0.57 mHz $1 \times 10^{-9}$ Hz/Hz + 5.7 mHz $1 \times 10^{-9}$ Hz/Hz + 57 mHz $1 \times 10^{-9}$ Hz/Hz + 0.57 Hz	Agilent 3325B Function Generator / HP Z3801A GPS Receiver
Frequency – Generate <sup>1</sup>	(10 to 26 500) MHz	$1 \times 10^{-9}$ Hz/Hz + 0.57 kHz	HP 83630B Signal Generator / HP Z3801A GPS Receiver
Frequency – Measure <sup>1</sup>	(1 to 10) Hz (10 to 100) Hz (100 to 1 000) Hz (1 to 10) kHz (10 to 100) kHz (100 to 200) kHz (0.2 to 3 000) MHz	$5.2 \times 10^{-9}$ Hz/Hz $2.5 \times 10^{-9}$ Hz/Hz $1.6 \times 10^{-9}$ Hz/Hz $1.3 \times 10^{-9}$ Hz/Hz $1.2 \times 10^{-9}$ Hz/Hz $1. \times 10^{-9}$ Hz/Hz $1.21 \times 10^{-9}$ Hz/Hz	Agilent 53131A Frequency Counter / HP Z3801A GPS Receiver
	(10 to 26 500) MHz	$1 \times 10^{-9}$ Hz/Hz + 0.1 Hz	Agilent E4440A Spectrum Analyzer / HP Z3801A GPS Receiver
Timer, Stopwatch <sup>1</sup>	10 s to 24 hr	34 ms	Totalize method with counter

**Time and Frequency**

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Tachometers – RPM <sup>1,2</sup>	Up to 100 000 RPM	0.001 % of reading + 0.6R	HP 3325B Signal Generator & LED

**DIMENSIONAL MEASUREMENT**

**1 Dimensional**

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Length	X Axis (0.01 to 8) in Y Axis (0.01 to 4) in	180 μin 180 μin	Optical comparator

**2 Dimensional**

Hilliard, OH

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Angle	Up to 360°	0.004°	Optical comparator

[Return to Site listing](#) (top)   [Go to Notes](#) (bottom)



ANSI National Accreditation Board

### Services performed at satellite laboratory

#### Technical Maintenance, Inc.

12240 SW 53<sup>rd</sup> Street, Suite 506

Fort Lauderdale, FL 33330

Jon Parmenter (Branch Manager) Phone: (954)-252-2223

Scott Chamberlain (Quality Manager) Phone: (321) 242-0890

### CALIBRATION

#### Acoustics and Vibration

Fort Lauderdale, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Accelerometers – Acceleration	(0.01 to 10) g (7 to < 10) Hz (10 to < 30) Hz (30 to < 2 000) Hz (2 to 10) kHz	4 % of reading 3 % of reading 1.5 % of reading 4 % of reading	Accelerometer Calibrator

#### Chemical Quantities

Ft Lauderdale, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
pH meters <sup>1,5</sup>	(4, 7, 10) pH	0.034 pH	pH buffer solutions
Conductivity Meters <sup>1,5</sup>	2 μS/cm 5 μS/cm 10 μS/cm 100 μS/cm 500 μS/cm 1 000 μS/cm 10 000 μS/cm 100 000 μS/cm	0.2 μS/cm 0.2 μS/cm 0.17 μS/cm 0.74 μS/cm 2.3 μS/cm 3.6 μS/cm 36 μS/cm 350 μS/cm	Conductivity solutions

**Electrical – DC/Low Frequency**

Fort Lauderdale, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Voltage – Generate <sup>1</sup>	(0 to 220) mV (0.22 to 2.2) V (2.2 to 11) V (11 to 22) V (22 to 220) V (220 to 1 100) V	6.8 $\mu\text{V/V} + 0.8 \mu\text{V}$ 4.6 $\mu\text{V/V} + 0.9 \mu\text{V}$ 3 $\mu\text{V/V} + 2.5 \mu\text{V}$ 3 $\mu\text{V/V} + 3.9 \mu\text{V}$ 4.6 $\mu\text{V/V} + 38 \mu\text{V}$ 6.1 $\mu\text{V/V} + 385 \mu\text{V}$	Fluke 5720A Multifunction Calibrator
DC Voltage – Measure <sup>1</sup>	Up to 100 mV 100 mV to 1V (1 to 10) V (10 to 100) V (100 to 1 000) V	5.0 $\mu\text{V/V} + 1.2 \mu\text{V}$ 4.0 $\mu\text{V/V} + 1.2 \mu\text{V}$ 4.0 $\mu\text{V/V} + 2.1 \mu\text{V}$ 6.0 $\mu\text{V/V} + 30 \mu\text{V}$ 6.0 $\mu\text{V/V} + 100 \mu\text{V} + 12 \mu\text{V/V} \times (\text{Vin}/1\ 000) ^2$	Agilent 3458A Multimeter
DC Voltage – Measure <sup>1</sup>	(1 to 40) kV	2.3 % of reading	Agilent 34401A Multimeter, Fluke 80k-40 High Voltage Probe
DC Current – Generate <sup>1</sup>	(1 to 2.2) nA (2.2 to 22) nA (22 to 220) nA (0.22 to 2.2) $\mu\text{A}$ (2.2 to 10) $\mu\text{A}$	93 $\mu\text{A/A} + 0.007 \text{ nA}$ 92 $\mu\text{A/A} + 0.007 \text{ nA}$ 92 $\mu\text{A/A} + 0.01 \text{ nA}$ 36 $\mu\text{A/A} + 0.1 \text{ nA}$ 20 $\mu\text{A/A} + 1 \text{ nA}$	Fluke 5720A Multifunction Calibrator & Fluke 5522A Multifunction Calibrator
DC Current – Generate <sup>1</sup>	(10 to 220) $\mu\text{A}$ 0.22 to 2.2 mA (2.2 to 22) mA (22 to 100) mA (100 to 220) mA (0.22 to 1) A (1 to 2.2) A	38 $\mu\text{A/A} + 5 \text{ nA}$ 30 $\mu\text{A/A} + 7 \text{ nA}$ 30 $\mu\text{A/A} + 44 \text{ nA}$ 38 $\mu\text{A/A} + 0.7 \mu\text{A}$ 45 $\mu\text{A/A} + 0.7 \mu\text{A}$ 68 $\mu\text{A/A} + 12 \mu\text{A}$ 105 $\mu\text{A/A} + 12 \mu\text{A}$	Fluke 5720A Multifunction Calibrator
DC Current – Generate <sup>1</sup>	(2.2 to 11) A	274 $\mu\text{A/A} + 365 \mu\text{A}$	Fluke 5720A Multifunction Calibrator, 5725A Amplifier
DC Current – Generate <sup>1</sup>	(11 to 20.5) A	761 $\mu\text{A/A} + 578 \mu\text{A}$	Fluke 5522A Multiproduct Calibrator
DC Current – Generate <sup>1</sup> Clamps Only	(0 to 200) A	0.21 % of output + 0.028 A	Fluke 5522A Multiproduct Calibrator / 9100-200 x10 Coil
DC Current – Generate <sup>1</sup> Clamps Only	(0 to 1 000) A	0.21 % of output + 0.04 A	Fluke 5522A Multiproduct Calibrator / 9100-200 x50 Coil



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Fort Lauderdale, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Current – Measure <sup>1</sup>	(1 to 10) nA (10 to 100) nA (0.1 to 1) $\mu$ A (1 to 10) $\mu$ A	35 $\mu$ A/A + 0.12 pA 18 $\mu$ A/A + 1.2 pA 11 $\mu$ A/A + 0.01 nA 9.1 $\mu$ A/A + 0.12 nA	Fluke 5720A Multifunction Calibrator, Agilent 3458A Multimeter Option 002
DC Current – Measure <sup>1</sup>	(10 to 100) $\mu$ A (0.1 to 1) mA (1 to 10) mA (10 to 100) mA (0.1 to 1) A	20 $\mu$ A/A + 0.8 nA 20 $\mu$ A/A + 5 nA 20 $\mu$ A/A + 0.05 $\mu$ A 35 $\mu$ A/A + 0.5 $\mu$ A 110 $\mu$ A/A + 10 $\mu$ A	Agilent 3458A Multimeter
DC Current – Measure <sup>1</sup>	(1 to 500) A	0.26 % of reading	Empro current shunts, Multimeter
Resistance – Generate <sup>1</sup> Fixed Points	100 V 100 k $\Omega$ (100 to 1 000) V 1 M $\Omega$ 10 M $\Omega$ 100 M $\Omega$ 1 G $\Omega$ 10 G $\Omega$	1 % of output 1 % of output 1 % of output 1 % of output 1 % of output 1.2 % of output	TMI RB Resistance Standard
Resistance – Generate <sup>1</sup>	Up to 11 $\Omega$ (11 to 33) $\Omega$ (33 to 110) $\Omega$ (110 to 330) $\Omega$ (0.33 to 1.1) k $\Omega$ (1.1 to 3.3) k $\Omega$ (3.3 to 11) k $\Omega$ (11 to 33) k $\Omega$ (33 to 110) k $\Omega$ (110 to 330) k $\Omega$ (0.33 to 1.1) M $\Omega$ (1.1 to 3.3) M $\Omega$ (3.3 to 11) M $\Omega$ (11 to 33) M $\Omega$ (33 to 110) M $\Omega$ (110 to 330) M $\Omega$ (0.33 to 1.1) G $\Omega$	30 $\mu\Omega/\Omega$ + 0.001 $\Omega$ 23 $\mu\Omega/\Omega$ + 0.001 $\Omega$ 21 $\mu\Omega/\Omega$ + 0.001 $\Omega$ 21 $\mu\Omega/\Omega$ + 0.002 $\Omega$ 21 $\mu\Omega/\Omega$ + 0.002 $\Omega$ 21 $\mu\Omega/\Omega$ + 0.02 $\Omega$ 21 $\mu\Omega/\Omega$ + 0.02 $\Omega$ 21 $\mu\Omega/\Omega$ + 0.2 $\Omega$ 21 $\mu\Omega/\Omega$ + 0.2 $\Omega$ 24 $\mu\Omega/\Omega$ + 2 $\Omega$ 24 $\mu\Omega/\Omega$ + 2 $\Omega$ 46 $\mu\Omega/\Omega$ + 23 $\Omega$ 99 $\mu\Omega/\Omega$ + 38 $\Omega$ 190 $\mu\Omega/\Omega$ + 1.9 k $\Omega$ 380 $\mu\Omega/\Omega$ + 2.3 k $\Omega$ 0.23 % of setting + 76 k $\Omega$ 1.1 % of reading + 380 k $\Omega$	Fluke 5522A Multiproduct Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Fort Lauderdale, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Resistance – Fixed Points	(1, 1.9) Ω (10, 19) Ω (100, 190) Ω (1, 1.9) kΩ (10, 19) kΩ (100, 190) kΩ 1 MΩ 1.9 MΩ 10 MΩ 19 MΩ 100 MΩ	84 μΩ/Ω + 0.1 μΩ 21 μΩ/Ω + 1 μΩ 9.1 μΩ/Ω + 6 μΩ 7.6 μΩ/Ω + 60 μΩ 7.6 μΩ/Ω + 0.6 mΩ 9.9 μΩ/Ω + 6 mΩ 18 μΩ/Ω + 60 mΩ 18 μΩ/Ω + 60 mΩ 35 μΩ/Ω + 0.6 Ω 42 μΩ/Ω + 0.6 Ω 91 μΩ/Ω + 6 Ω	Fluke 5720A Multifunction Calibrator
Resistance – Measure <sup>1</sup>	Up to 12 Ω (10 to 120) Ω (0.1 to 1.2 kΩ (1 to 12) kΩ (10 to 120) kΩ (0.1 to 1.2) MΩ (1 to 12) MΩ (10 to 120) MΩ (0.1 to 1.2) GΩ	15 μΩ/Ω + 56 μΩ 12 μΩ/Ω + 0.5 mΩ 10 μΩ/Ω + 0.6 mΩ 10 μΩ/Ω + 5.6 mΩ 10 μΩ/Ω + 56 mΩ 15 μΩ/Ω + 2.2 Ω 50 μΩ/Ω + 120 Ω 500 μΩ/Ω + 1.2 kΩ 0.5 % of reading + 70 kΩ	Agilent 3458A Multimeter
Capacitance – Generate <sup>1</sup>	(220 to 400) pF (0.4 to 3.299 9) nF (3.3 to 10.999 9) nF (11 to 32.999 9) nF (33 to 109.999) nF (110 to 329.999) nF (0.33 to 1.099 99) μF (1.1 to 3.299 99) μF (3.3 to 10.999 9) μF (11 to 32.999 9) μF (33 to 109.999) μF (110 to 329.999) μF (0.33 to 1.099 99) mF (1.1 to 3.299 99) mF (3.3 to 10.999 9) mF (11 to 32.999 9) mF (33 to 110) mF	0.38% of output + 7.6 pF 0.38 % of output + 0.01 nF 0.19 % of output + 0.01 nF 0.19 % of output + 0.08 nF 0.19 % of output + 0.08 nF 0.19 % of output + 0.23 nF 0.19 % of output + 0.76 nF 0.19 % of output + 2.3 nF 0.19 % of output + 7.6 nF 0.3 % of output + 23 nF 0.34 % of output + 76 nF 0.34 % of output + 228 nF 0.34 % of output + 0.76 μF 0.34 % of output + 2.3 μF 0.34 % of output + 7.6 μF 0.57 % of output + 23 μF 0.84 % of output + 76 μF	Fluke 5522A Multiproduct Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Fort Lauderdale, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Capacitance – Measure <sup>1</sup>	100 Hz /120 Hz		Agilent 4263B LCR Meter
	(16 to 400) pF	2 % of reading + 0.3 pF	
	(0.4 to 1) nF	0.17 % of reading	
	(1 to 100) nF	0.23 % of reading	
	(0.1 to 1) μF	0.23 % of reading	
	(1 to 100) μF	0.23 % of reading	
	(0.1 to 1) mF	0.23 % of reading	
	1 000 Hz		
	(16 to 400) pF	0.43 % of reading + 0.3 pF	
	(0.4 to 1) nF	0.1 % of reading	
	(1 to 100) nF	0.13 % of reading	
	(0.1 to 1) μF	0.12 % of reading	
	(1 to 100) μF	0.13 % of reading	
	(0.1 to 1) mF	0.45 % of reading	
	10 kHz		
	(16 to 400) pF	0.56 % of reading + 0.3 pF	
	(0.4 to 1) nF	0.15 % of reading	
	(1 to 100) nF	0.2 % of reading	
	(0.1 to 1) μF	0.17 % of reading	
	(1 to 100) μF	0.69 % of reading	
	(0.1 to 1) mF	3.5 % of reading	
20 kHz			
(16 to 400) pF	1.2 % of reading + 0.3 pF		
(0.4 to 1) nF	0.66 % of reading		
(1 to 100) nF	0.65 % of reading		
(0.1 to 1) μF	0.53 % of reading		
(1 to 100) μF	1.7 % of reading		
Inductance – Measure <sup>1</sup>	100 Hz /120 Hz		Agilent 4263B LCR Meter
	(4 to 10) μH	5.2 % of reading + 0.03 μH	
	(10 to 40) μH	2.3 % of reading + 0.03 μH	
	(40 to 100) μH	0.57 % of reading + 0.03 μH	
	(0.1 to 0.4) mH	0.4 % of reading	
	(0.4 to 1) mH	0.32 % of reading	
	(1 to 10) mH	0.28 % of reading	
	(10 to 100) mH	0.23 % of reading	
	(0.1 to 1) H	0.18 % of reading	
	(1 to 10) H	0.22 % of reading	



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Fort Lauderdale, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Inductance – Measure <sup>1</sup>	1 000 Hz		Agilent 4263B LCR Meter
	(1 to 4) $\mu$ H	1 % of reading + 0.03 $\mu$ H	
	(4 to 10) $\mu$ H	0.46 % of reading + 0.03 $\mu$ H	
	(10 to 40) $\mu$ H	0.74 % of reading + 0.03 $\mu$ H	
	(40 to 100) $\mu$ H	0.35 % of reading + 0.03 $\mu$ H	
	(0.01 to 0.4) mH	0.3 % of reading	
	(0.4 to 1) mH	0.11 % of reading	
	(1 to 10) mH	0.12 % of reading	
	(10 to 100) mH	0.1 % of reading	
	(0.1 to 1) H	0.1 % of reading	
	(1 to 10) H	0.12 % of reading	
	10 kHz		
	(1 to 4) $\mu$ H	0.62 % of reading + 0.03 $\mu$ H	
	(4 to 10) $\mu$ H	0.4 % of reading + 0.03 $\mu$ H	
	(10 to 40) $\mu$ H	0.32 % of reading + 0.03 $\mu$ H	
	(40 to 100) $\mu$ H	0.18 % of reading + 0.03 $\mu$ H	
	(0.1 to 0.4) mH	0.2 % of reading	
	(0.4 to 1) mH	0.14 % of reading	
	(1 to 10) mH	0.17 % of reading	
	(10 to 100) mH	0.2 % of reading	
	(0.1 to 1) H	0.2 % of reading	
	(1 to 10) H	0.44 % of reading	
	20 kHz		
	(1 to 4) $\mu$ H	1.3 % of reading + 0.03 $\mu$ H	
	(4 to 10) $\mu$ H	0.68 % of reading + 0.03 $\mu$ H	
	(10 to 40) $\mu$ H	0.63 % of reading + 0.03 $\mu$ H	
	(40 to 100) $\mu$ H	0.53 % of reading + 0.03 $\mu$ H	
(0.1 to 0.4) mH	0.49 % of reading		
(0.4 to 1) mH	0.48 % of reading		
(1 to 10) mH	0.65 % of reading		
(10 to 100) mH	0.66 % of reading		
(0.1 to 1) H	1.2 % of reading		
(1 to 10) H	3.3 % of reading		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Fort Lauderdale, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Inductance – Measure <sup>1</sup>	100 kHz (1 to 4) $\mu$ H (4 to 10) $\mu$ H (10 to 40) $\mu$ H (40 to 100) $\mu$ H (0.1 to 0.4) mH (0.4 to 1) mH (1 to 10) mH (10 to 100) mH (0.1 to 1) H	2.7 % of reading + 0.03 $\mu$ H 0.86 % of reading + 0.03 $\mu$ H 1.2 % of reading + 0.03 $\mu$ H 0.68 % of reading + 0.03 $\mu$ H 0.9 % of reading 1.1 % of reading 1.4 % of reading 1.4 % of reading 6.3 % of reading	Agilent 4263B LCR Meter
AC Voltage – Generate <sup>1</sup>	(0.22 to 22) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz (22 to 220) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz (0.22 to 2.2) V (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	228 $\mu$ V/V + 3.9 $\mu$ V 88 $\mu$ V/V + 3.9 $\mu$ V 76 $\mu$ V/V + 3.9 $\mu$ V 190 $\mu$ V/V + 3.9 $\mu$ V 457 $\mu$ V/V + 4.6 $\mu$ V 989 $\mu$ V/V + 9.2 $\mu$ V 1.3 mV/V + 19 $\mu$ V 2.6 mV/V + 19 $\mu$ V 228 $\mu$ V/V + 11 $\mu$ V 88 $\mu$ V/V + 6.1 $\mu$ V 76 $\mu$ V/V + 6.1 $\mu$ V 190 $\mu$ V/V + 6.1 $\mu$ V 457 $\mu$ V/V + 15 $\mu$ V 837 $\mu$ V/V + 19 $\mu$ V 1.3 mV/V + 23 $\mu$ V 2.5 mV/V + 46 $\mu$ V 228 $\mu$ V/V + 38 $\mu$ V 84 $\mu$ V/V + 15 $\mu$ V 40 $\mu$ V/V + 8 $\mu$ V 68 $\mu$ V/V + 9 $\mu$ V 99 $\mu$ V/V + 30 $\mu$ V 380 $\mu$ V/V + 76 $\mu$ V 913 $\mu$ V/V + 190 $\mu$ V 1.5 mV/V + 304 $\mu$ V	Fluke 5720A Multifunction Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Fort Lauderdale, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Generate <sup>1</sup>	(2.2 to 22) V		Fluke 5720A Multifunction Calibrator
	(10 to 20) Hz	228 $\mu$ V/V + 380 $\mu$ V	
	(20 to 40) Hz	84 $\mu$ V/V + 152 $\mu$ V	
	40 Hz to 20 kHz	37 $\mu$ V/V + 54 $\mu$ V	
	(20 to 50) kHz	61 $\mu$ V/V + 91 $\mu$ V	
	(50 to 100) kHz	76 $\mu$ V/V + 190 $\mu$ V	
	(100 to 300) kHz	228 $\mu$ V/V + 609 $\mu$ V	
	(300 to 500) kHz	913 $\mu$ V/V + 1.9 mV	
	500 kHz to 1 MHz	1.4 mV/V + 3 mV	
	(22 to 220) V		
	(10 to 20) Hz	228 $\mu$ V/V + 3.8 mV	
	(20 to 40) Hz	84 $\mu$ V/V + 1.5 mV	
	40 Hz to 20 kHz	49 $\mu$ V/V + 0.6 mV	
	(20 to 50) kHz	76 $\mu$ V/V + 0.9 mV	
(50 to 100) kHz	137 $\mu$ V/V + 2.3 mV		
AC Voltage – Generate <sup>1</sup>	(220 to 750) V		Fluke 5720A Multifunction Calibrator, 5725A Amplifier
	40 Hz to 1 kHz	68 $\mu$ V/V + 3 mV	
	(1 to 20) kHz	126 $\mu$ V/V + 5 mV	
	(20 to 50) kHz	457 $\mu$ V/V + 8 mV	
	(50 to 100) kHz	1.8 mV/V + 34 mV	
	(750 to 1 000) V		
	40 Hz to 1 kHz	68 $\mu$ V/V + 3 mV	
	(1 to 20) kHz	126 $\mu$ V/V + 5 mV	
	(20 to 30) kHz	457 $\mu$ V/V + 8 mV	
	AC Voltage – Generate <sup>1</sup> Wideband Absolute	(0.3 to 1.1) mV	
(10 to 30 Hz)		0.65 % of output + 1.5 $\mu$ V	
30 Hz to 500 kHz		0.61 % of output + 1.5 $\mu$ V	
(0.5 to 1.2) MHz		0.63 % of output + 3.8 $\mu$ V	
(1.2 to 2) MHz		0.63 % of output + 3.8 $\mu$ V	
(2 to 12) MHz		0.68 % of output + 3.8 $\mu$ V	
(12 to 20) MHz		0.76 % of output + 3.8 $\mu$ V	
(20 to 30) MHz		1.3 % of output + 13 $\mu$ V	
(1.1 to 3.3) mV			
(10 to 30 Hz)		0.58 % of output + 2.3 $\mu$ V	
30 Hz to 500 kHz		0.53 % of output + 2.3 $\mu$ V	
(0.5 to 1.2) MHz		0.54 % of output + 4.6 $\mu$ V	
(1.2 to 2) MHz		0.54 % of output + 4.6 $\mu$ V	
(2 to 12) MHz		0.58 % of output + 4.6 $\mu$ V	
(12 to 20) MHz		0.65 % of output + 4.6 $\mu$ V	
(20 to 30) MHz		1.3 % of output + 4.6 $\mu$ V	



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Fort Lauderdale, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Generate <sup>1</sup> Wideband Absolute	(3.3 to 11) mV		Fluke 5720A Option 003 Multifunction Calibrator
	(10 to 30 Hz)	0.58 % of output + 6.1 μV	
	30 Hz to 500 kHz	0.53 % of output + 6.1 μV	
	(0.5 to 1.2) MHz	0.54 % of output + 8.4 μV	
	(1.2 to 2) MHz	0.54 % of output + 8.4 μV	
	(2 to 12) MHz	0.55 % of output + 8.4 μV	
	(12 to 20) MHz	0.61 % of output + 8.4 μV	
	(20 to 30) MHz	0.93 % of output + 8.4 μV	
	(11 to 33) mV		
	(10 to 30 Hz)	0.52 % of output + 12 μV	
	30 Hz to 500 kHz	0.46 % of output + 12 μV	
	(0.5 to 1.2) MHz	0.47 % of output + 14 μV	
	(1.2 to 2) MHz	0.47 % of output + 14 μV	
	(2 to 12) MHz	0.49 % of output + 14 μV	
	(12 to 20) MHz	0.55 % of output + 14 μV	
	(20 to 30) MHz	0.89 % of output + 14 μV	
	(33 to 110) mV		
	(10 to 30 Hz)	0.52 % of output + 30 μV	
	30 Hz to 500 kHz	0.46 % of output + 30 μV	
	(0.5 to 1.2) MHz	0.47 % of output + 33 μV	
	(1.2 to 2) MHz	0.47 % of output + 33 μV	
	(2 to 12) MHz	0.49 % of output + 33 μV	
	(12 to 20) MHz	0.55 % of output + 33 μV	
	(20 to 30) MHz	0.89 % of output + 33 μV	
(110 to 330) mV			
(10 to 30 Hz)	0.45 % of output + 0.1 mV		
30 Hz to 500 kHz	0.38 % of output + 0.1 mV		
(0.5 to 1.2) MHz	0.4 % of output + 0.1 mV		
(1.2 to 2) MHz	0.4 % of output + 0.1 mV		
(2 to 12) MHz	0.42 % of output + 0.1 mV		
(12 to 20) MHz	0.49 % of output + 0.1 mV		
(20 to 30) MHz	0.85 % of output + 0.1 mV		
0.33 to 1.1 V			
(10 to 30 Hz)	0.45% of output + 0.3 mV		
30 Hz to 500 kHz	0.38 % of output + 0.3 mV		
(0.5 to 1.2) MHz	0.4 % of output + 0.3 mV		
(1.2 to 2) MHz	0.4 % of output + 0.3 mV		
(2 to 12) MHz	0.42 % of output + 0.3 mV		
(12 to 20) MHz	0.49 % of output + 0.3 mV		
(20 to 30) MHz	0.85 % of output + 0.3 mV		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Fort Lauderdale, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Generate <sup>1</sup> Wideband Absolute	(1.1 to 3.5) V (10 to 30 Hz) 30 Hz to 500 kHz (0.5 to 1.2) MHz (1.2 to 2) MHz (2 to 12) MHz (12 to 20) MHz (20 to 30) MHz	0.39 % of output + 0.4 mV 0.3 % of output + 0.4 mV 0.32 % of output + 0.4 mV 0.32 % of output + 0.4 mV 0.35 % of output + 0.4 mV 0.44 % of output + 0.4 mV 0.82 % of output + 0.4 mV	Fluke 5720A Multifunction Calibrator Option 003
AC Voltage – Measure <sup>1</sup>	(0 to 10) mV (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz 100 kHz to 1 MHz (1 to 4) MHz (4 to 8) MHz (10 to 100) mV (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz (2 to 4) MHz (4 to 8) MHz (8 to 10) MHz (0.1 to 1) V (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz (2 to 4) MHz (4 to 8) MHz (8 to 10) MHz	0.03 % of reading + 3 μV 0.02 % of reading + 1.1 μV 0.03 % of reading + 1.1 μV 0.1 % of reading + 1.1 μV 0.5 % of reading + 1.1 μV 1.2 % of reading + 5 μV 7 % of reading + 7 μV 20 % of reading + 8 μV 0.007 % of reading + 4 μV 0.007 % of reading + 2 μV 0.014 % of reading + 2 μV 0.03 % of reading + 2 μV 0.08 % of reading + 2 μV 0.3 % of reading + 10 μV 1 % of reading + 10 μV 1.5 % of reading + 70 μV 4 % of reading + 70 μV 4 % of reading + 80 μV 15 % of reading + 100 μV 0.007 % of reading + 40 μV 0.07 % of reading + 20 μV 0.014 % of reading + 20 μV 0.03 % of reading + 20 μV 0.08 % of reading + 20 μV 0.3 % of reading + 100 μV 1 % of reading + 100 μV 1.5 % of reading + 0.7 mV 4 % of reading + 0.7 mV 4 % of reading + 0.8 mV 15 % of reading + 1 mV	Agilent 3458A Multimeter



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Fort Lauderdale, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure <sup>1</sup>	(1 to 10) V		Agilent 3458A Multimeter
	(1 to 40) Hz	0.007 % of reading + 0.4 mV	
	40 Hz to 1 kHz	0.007 % of reading + 0.2 mV	
	(1 to 20) kHz	0.014 % of reading + 0.2 mV	
	(20 to 50) kHz	0.03 % of reading + 0.2 mV	
	(50 to 100) kHz	0.08 % of reading + 0.2 mV	
	(100 to 300) kHz	0.3 % of reading + 1 mV	
	300 kHz to 1 MHz	1 % of reading + 1 mV	
	(1 to 2) MHz	1.5 % of reading + 7 mV	
	(2 to 4) MHz	4 % of reading + 7 mV	
	(4 to 8) MHz	4 % of reading + 8 mV	
	(8 to 10) MHz	15 % of reading + 10 mV	
	(10 to 100) V		
	(1 to 40) Hz	0.02 % of reading + 4 mV	
	40 Hz to 1 kHz	0.02 % of reading + 2 mV	
(1 to 20) kHz	0.02 % of reading + 2 mV		
(20 to 50) kHz	0.035 % of reading + 2 mV		
(50 to 100) kHz	0.1 % of reading + 2 mV		
(100 to 300) kHz	0.4 % of reading + 10 mV		
300 kHz to 1 MHz	1.5 % of reading + 10 mV		
(100 to 700) V			
(1 to 40) Hz	0.04 % of reading + 40 mV		
40 Hz to 1 kHz	0.04 % of reading + 20 mV		
(1 to 20) kHz	0.06 % of reading + 20 mV		
(20 to 50) kHz	0.12 % of reading + 20 mV		
(50 to 100) kHz	0.3 % of reading + 20 mV		
AC Voltage – Measure <sup>1</sup>	(1 to 28) kV 60 Hz	5.8 % of reading	Agilent 34401A Multimeter / Fluke 80k-40 High Voltage Probe
AC Current – Generate <sup>1</sup>	(9 to 220) $\mu$ A		Fluke 5720A Multifunction Calibrator
	(10 to 20) Hz	228 $\mu$ A/A + 15 nA	
	(20 to 40) Hz	152 $\mu$ A/A + 10 nA	
	40 Hz to 1 kHz	91 $\mu$ A/A + 8 nA	
	(1 to 5) kHz	266 $\mu$ A/A + 12 nA	
	(5 to 10) kHz	989 $\mu$ A/A + 61 nA	
	(0.22 to 2.2) mA		
	(10 to 20) Hz	228 $\mu$ A/A + 39 nA	
	(20 to 40) Hz	152 $\mu$ A/A + 31 nA	
	40 Hz to 1 kHz	107 $\mu$ A/A + 31 nA	
	(1 to 5) kHz	183 $\mu$ A/A + 99 nA	
	(5 to 10) kHz	989 $\mu$ A/A + 609 nA	



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Fort Lauderdale, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Generate <sup>1</sup>	(2.2 to 22) mA		Fluke 5720A Multifunction Calibrator
	(10 to 20) Hz	228 $\mu$ A/A + 385 nA	
	(20 to 40) Hz	152 $\mu$ A/A + 310 nA	
	40 Hz to 1 kHz	107 $\mu$ A/A + 310 nA	
	(1 to 5) kHz	183 $\mu$ A/A + 536 nA	
	(5 to 10) kHz	989 $\mu$ A/A + 4.6 $\mu$ A	
	(22 to 220) mA		
	(10 to 20) Hz	228 $\mu$ A/A + 4 $\mu$ A	
	(20 to 40) Hz	152 $\mu$ A/A + 3 $\mu$ A	
	40 Hz to 1 kHz	107 $\mu$ A/A + 2 $\mu$ A	
	(1 to 5) kHz	183 $\mu$ A/A + 3 $\mu$ A	
	(5 to 10) kHz	989 $\mu$ A/A + 9 $\mu$ A	
AC Current – Generate <sup>1</sup>	(0.22 to 2.2) A		Fluke 5720A Multifunction Calibrator, 5725A Amplifier
	20 Hz to 1 kHz	243 $\mu$ A/A + 31 $\mu$ A	
	(1 to 5) kHz	380 $\mu$ A/A + 76 $\mu$ A	
	(5 to 10) kHz	6.1 mA/A + 152 $\mu$ A	
AC Current – Generate <sup>1</sup>	(2.2 to 11) A		Fluke 5720A Multifunction Calibrator, 5725A Amplifier
	40 Hz to 1 kHz	350 $\mu$ A/A + 141 $\mu$ A	
	(1 to 5) kHz	723 $\mu$ A/A + 295 $\mu$ A	
	(5 to 10) kHz	2.7 mA/A + 573 $\mu$ A	
AC Current – Generate <sup>1</sup>	(11 to 20.5) A		Fluke 5522A Multiproduct Calibrator
	(45 to 100) Hz	0.09 % of output + 5 mA	
	100 Hz to 1 kHz	0.11 % of output + 5 mA	
	(1 to 5) kHz	2.28 % of output + 5 mA	
AC Current – Generate <sup>1</sup> Clamp Meters Only	(3.3 to 30) A		Fluke 5522A Multiproduct Calibrator / 9100-200 x10 Coil
	(10 to 100) Hz	0.22 % of output + 0.028 A	
	(100 to 440) Hz	0.3 % of output + 0.07 A	
	(30 to 200) A		
	10 to 100 Hz	0.22 % of output + 0.032 A	
	(100 to 440) Hz	0.79 % of output + 0.08 A	
AC Current – Generate <sup>1</sup> Clamp Meters Only	(16.5 to 150) A		Fluke 5522A Multiproduct Calibrator /9100-200 x50 Coil
	(10 to 100) Hz	0.22 % of output + 0.029 A	
	(100 to 440) Hz	0.3 % of output + 0.08 A	
	(150 to 1 000) A		
	(10 to 100) Hz	0.22 % of output + 0.081 A	
	(100 to 440) Hz	0.79 % of output + 0.2 A	



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Fort Lauderdale, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Measure <sup>1</sup>	(5 to 100) $\mu$ A		Agilent 3458A Multimeter
	(10 to 20) Hz	0.4 % of reading + 0.03 $\mu$ A	
	(20 to 45) Hz	0.15 % of reading + 0.03 $\mu$ A	
	45 Hz to 1 kHz	0.06 % of reading + 0.03 $\mu$ A	
	(0.1 to 1) mA		
	(10 to 20) Hz	0.4 % of reading + 0.2 $\mu$ A	
	(20 to 45) Hz	0.15 % of reading + 0.2 $\mu$ A	
	(45 to 100) Hz	0.06 % of reading + 0.2 $\mu$ A	
	100 Hz to 5 kHz	0.03 % of reading + 0.2 $\mu$ A	
	(1 to 10) mA		
	(10 to 20) Hz	0.4 % of reading + 2 $\mu$ A	
	(20 to 45) Hz	0.15 % of reading + 2 $\mu$ A	
	(45 to 100) Hz	0.06 % of reading + 2 $\mu$ A	
	100 Hz to 5 kHz	0.03 % of reading + 2 $\mu$ A	
	(10 to 100) mA		
	(10 to 20) Hz	0.4 % of reading + 20 $\mu$ A	
(20 to 45) Hz	0.15 % of reading + 20 $\mu$ A		
(45 to 100) Hz	0.06 % of reading + 20 $\mu$ A		
100 Hz to 5 kHz	0.03 % of reading + 20 $\mu$ A		
(0.1 to 1) A			
(10 to 20) Hz	0.4 % of reading + 0.2 mA		
(20 to 45) Hz	0.15 % of reading + 0.2 mA		
(45 to 100) Hz	0.06 % of reading + 0.2 mA		
100 Hz to 5 Hz	0.03 % of reading + 0.2 mA		
AC Current – Measure <sup>1</sup>	(1 to 3) A (3 to 5) Hz (5 to 10) Hz 10 Hz to 5 kHz	0.72 % of reading + 1.2 mA 0.23 % of reading + 1.2 mA 0.1 % of reading + 1.2 mA	Agilent 34401A Multimeter
AC Current – Measure <sup>1</sup>	(3 to 30) A 40 Hz to 1 kHz (1 to 5) kHz	0.3 % of reading + 0.07 A 5 % of reading + 0.14 A	Agilent 3458A Multimeter, Keysight 34330A Current Shunt
AC Current – Measure <sup>1</sup>	(30 to 600) A (10 to 100) Hz (100 to 400) Hz	1.5 % of reading + 1.1 A 4.6 % of reading + 1.1 A	Fluke 336 Clamp Meter



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Fort Lauderdale, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Oscilloscopes Calibration <sup>1</sup> – Generate			
Voltage DC - 50Ω	(1 to 24.999) mV (25 to 109.99) mV (110mV to 2.199 9) V (2.2 to 6.6) V	0.19 % of Output + 31 μV 0.19 % of Output + 36 μV 0.19 % of Output + 87 μV 0.19 % of Output + 0.6 mV	Fluke 5820A Oscilloscope Calibrator w/ GHz Option
DC - 1MΩ	(1 to 24.999) mV (25 to 109.99) mV (110mV to 2.199 9) V (2.2 to 10.999) V (11 to 130) V	0.019 % of reading + 20 μV 0.019 % of reading + 25 μV 0.019 % of reading + 76 μV 0.019 % of reading + 0.6 mV 0.019 % of reading + 6 mV	
Square Wave 10Hz to 10kHz - 50Ω	(1 to 24.999) mVpp (25 to 109.99) mVpp (110mV to 2.199 9) Vpp (2.2 to 6.6) Vpp	0.19 % of Output + 31 μV 0.19 % of Output + 36 μV 0.19 % of Output + 87 μV 0.19 % of Output + 0.6 mV	
Square Wave 10 Hz to 1KHz - 1MΩ	(1 to 24.999) mV (25 to 109.99) mV (110mV to 2.1999) V (2.2 to 10.999) V (11 to 130) V	0.038 % of reading + 4 μV 0.038 % of reading + 9 μV 0.038 % of reading + 60 μV 0.038 % of reading + 0.6 mV 0.038 % of reading + 6 mV	
Square Wave 1 to 10KHz - 1MΩ	(1 to 24.999) mV (25 to 109.99) mV (110mV to 2.199 9) V (2.2 to 10.999) V (11 to 130) V	0.19 % of reading + 31 μV 0.19 % of reading + 36 μV 0.19 % of reading + 87 μV 0.19 % of reading + 0.6 mV 0.19 % of reading + 6 mV	



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Fort Lauderdale, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Oscilloscopes Calibration – Generate <sup>1</sup> Leveled Sine Flatness 50 kHz to 10 MHz Reference	3 dB Bandwidth (5 to 50) mVpp 50 kHz to 100) MHz (100 to 300) MHz (300 to 500) MHz (500 to 600) MHz (600 to 16 00) MHz (1 600 to 2 100) MHz 50 mV to 3.5 Vpp 50 kHz to 100 MHz (100 to 300) MHz (300 to 500) MHz (500 to 600) MHz (600 to 16 00) MHz (1 600 to 2 100) MHz (3.5 to 5) Vpp 50 kHz to 100 MHz (100 to 300) MHz (300 to 500) MHz (500 to 600) MHz	0.34 dB 0.36 dB 0.42 dB 0.46 dB 0.5 dB 0.56 dB 0.24 dB 0.24 dB 0.32 dB 0.34 dB 0.4 dB 0.44 dB 0.24 dB 0.24 dB 0.32 dB 0.34 dB	Fluke 5820A Oscilloscope Calibrator w/ GHz Option
Oscilloscopes Calibration – Generate <sup>1</sup> Leveled Sine Flatness 50 kHz to 10 MHz Reference	3 dB Bandwidth 50 mVpp to 3.5Vpp (2 100 to 4 000) MHz (4 000 to 8 000) MHz (8 000 to 18 000) MHz	0.25 dB 0.35 dB 0.46 dB	EPM Power Meter w/ E Series Power Sensors
Oscilloscopes Calibration – Generate <sup>1</sup> Time Marker Input Impedance Resistance Leakage	500 ps to 20 ms 50 ms to 5 s (40 to 60) Ω 500 kΩ to 1.5MΩ (0 to 5.99) V	0.25 μs/s 1.9 μs/s + 3.8 μHz 0.08 % of reading 0.08 % of reading 0.038 % of reading + 0.8 mV	Fluke 5820A Oscilloscope Calibrator w/ GHz Option



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Fort Lauderdale, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Calibration of Thermocouple Indicators <sup>1</sup>	Type B		Fluke 7526A Process Calibrator
	(600 to 800) °C	0.27 °C	
	(800 to 1 550) °C	0.22 °C	
	(1 550 to 1 820) °C	0.17 °C	
	Type C		
	(0 to 1 000) °C	0.13 °C	
	(1 000 to 1 800) °C	0.18 °C	
	(1 800 to 2 000) °C	0.2 °C	
	(2 000 to 2 316) °C	0.27 °C	
	Type E		
	(-250 to -200) °C	0.19 °C	
	(-200 to -100) °C	0.1 °C	
	(-100 to 0) °C	0.07 °C	
	(0 to 600) °C	0.07 °C	
	(600 to 1 000) °C	0.08 °C	
	Type J		
	(-210 to -100) °C	0.11 °C	
	(-100 to 800) °C	0.07 °C	
	(800 to 1 200) °C	0.08 °C	
	Type K		
	(-250 to -200) °C	0.35 °C	
(-200 to -100) °C	0.13 °C		
(-100 to 800) °C	0.08 °C		
(800 to 1 372) °C	0.1 °C		
Type L			
(-200 to -100) °C	0.08 °C		
(-100 to 900) °C	0.07 °C		
Type N			
(-250 to -200) °C	0.56 °C		
(-200 to -100) °C	0.18 °C		
(-100 to 0) °C	0.1 °C		
(0 to 100) °C	0.09 °C		
(100 to 800) °C	0.08 °C		
(800 to 1 300) °C	0.1 °C		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Fort Lauderdale, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Calibration of Thermocouple Indicators <sup>1</sup>	Type R		Fluke 7526A Process Calibrator
	(-50 to -25) °C	0.42 °C	
	(-25 to 0) °C	0.34 °C	
	(0 to 100) °C	0.3 °C	
	(100 to 400) °C	0.22 °C	
	(400 to 600) °C	0.17 °C	
	(600 to 1 000) °C	0.16 °C	
	(1 000 to 1 600) °C	0.15 °C	
	(1 600 to 1 767) °C	0.18 °C	
	Type S		
	(50 to -25) °C	0.39 °C	
	(-25 to 0) °C	0.33 °C	
	(0 to 100) °C	0.29 °C	
	(100 to 400) °C	0.22 °C	
	(400 to 600) °C	0.18 °C	
	(600 to 1 600) °C	0.17 °C	
	(1 600 to 1 767) °C	0.2 °C	
	Type T		
(-250 to -200) °C	0.26 °C		
(-200 to -100) °C	0.13 °C		
(-100 to 0) °C	0.09 °C		
(0 to 400) °C	0.07 °C		
Type U			
(-200 to 0) °C	0.13 °C		
(0 to 600) °C	0.08 °C		
Electrical Calibration of RTD Indicators <sup>1</sup>	Pt 385, 100 Ω		Fluke 7526A Process Calibrator
	(-200 to 800) °C	0.05 °C	
	Pt 3926, 100 Ω		
	(-200 to 630) °C	0.05 °C	
	Pt 3916, 100 Ω		
	(-200 to 630) °C	0.05 °C	
	Pt 385, 200 Ω		
	(-200 to 400) °C	0.4 °C	
(400 to 630) °C	0.5 °C		
Pt 385, 500 Ω			
(-200 to 630) °C	0.17 °C		
Pt 385, 1 000 Ω			
(-200 to 630) °C	0.09 °C		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Fort Lauderdale, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Calibration of RTD Indicators <sup>1</sup>	Cu 427, 10 Ω (100 °C to 260) °C	0.35 °C	Fluke 5522A Multiproduct Calibrator
	PtNi 385, 120 Ω (Ni120) (-80 to 0) °C	0.093 °C	
	(0 to 100) °C	0.093 °C	
	(100 to 260) °C	0.16 °C	
Phase Angle	(-180 to 180)°		Fluke 5522A Multiproduct Calibrator
	(10 to 65) Hz	0.14 % of reading	
	(65 to 500) Hz	0.3 % of reading	
	500 Hz to 1 kHz	0.61 % of reading	
	(1 to 5) kHz	2.9 % of reading	
	(5 to 10) kHz	5.8% of reading	
(10 to 30) kHz	12% of reading		
Power Meters <sup>1</sup>	3 μW to 100 mW	0.32 % of reading	HP 11683A Range Calibrator

Electrical – RF/Microwave

Fort Lauderdale, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Attenuation –Measure <sup>1</sup>	(30 to 3 050) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 526 Power Sensor
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.12 dB	
	(90 to 100) dB	0.13 dB	
	(100 to 110) dB	0.13 dB	
(110 to 120) dB	0.26 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Fort Lauderdale, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Attenuation – Measure <sup>1</sup>	(3 050 to 6 600) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 526 Power Sensor
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.12 dB	
	(90 to 100) dB	0.13 dB	
	(100 to 110) dB	0.15 dB	
	(110 to 120) dB	0.37 dB	
	(6 600 to 13 200) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.12 dB	
	(90 to 100) dB	0.14 dB	
	(100 to 110) dB	0.34 dB	
	(110 to 120) dB	0.77 dB	
	(13 200 to 19 200) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.12 dB	
(90 to 100) dB	0.27 dB		
(100 to 110) dB	0.66 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Fort Lauderdale, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Attenuation – Measure <sup>1</sup>	(19 200 to 26 500) MHz (0 to 10) dB (10 to 20) dB (20 to 30) dB (30 to 40) dB (40 to 50) dB (50 to 60) dB (60 to 70) dB (70 to 80) dB (80 to 90) dB (90 to 100) dB (100 to 110) dB	0.02 dB 0.025 dB 0.03 dB 0.035 dB 0.04 dB 0.076 dB 0.081 dB 0.12 dB 0.2 dB 0.52 dB 1.1 dB	Agilent N5531S Measuring Receiver w/N5532A Opt 526 Power Sensor
RF Power – Measure <sup>1</sup> 50 MHz	1.0 mW	0.0032 mW	HP 478A, 432A Power Meter
RF Flatness – Measure <sup>1</sup>	9 kHz to 2 000 MHz (20 to -10) dBm (-10 to -30) dBm (-30 to -40) dBm (-40 to -42) dBm (2 to 14) GHz (20 to -10) dBm (-10 to -30) dBm (-30 to -40) dBm (-40 to -42) dBm (14 to 18) GHz (20 to -10) dBm (-10 to -30) dBm (-30 to -40) dBm (-40 to -42) dBm	0.1 dB 0.1 dB 0.11 dB 0.12 dB 0.1 dB 0.09 dB 0.1 dB 0.11 dB 0.11 dB 0.12 dB 0.12 dB 0.12 dB 0.13 dB	Agilent EPM Series Power Meter w/E9304A H18 Power Sensor
RF Power – Measure <sup>1</sup>	9 kHz to 14 000 MHz (20 to 0) dB (0 to -40) dB (-40 to -50) dB (-50 to -55) dB (14 000 to 18 000) MHz (20 to 0) dB (0 to -40) dB (-40 to -50) dB (-50 to -55) dB	0.13 dB 0.15 dB 0.35 dB 0.93 dB 0.12 dB 0.16 dB 0.35 dB 0.93 dB	Agilent EPM Series Power Meter w/E9304A H18 Power Sensor



ANSI National Accreditation Board

Electrical – RF/Microwave

Fort Lauderdale, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power Measure <sup>1</sup>	100 kHz to 30 MHz (20 to 0) dB (0 to -58) dB (-58 to -78) dB (-78 to -110) dB (-110 to -115) dB (-115 to -120) dB (-120 to -125) dB	0.12 dB 0.13 dB 0.15 dB 0.19 dB 0.26 dB 0.39 dB 0.59 dB	Agilent EPM Series Power Meter w/E9304A H18 Power Sensor
RF Power Measure <sup>1</sup>	(30 to 2 000) MHz (30 to 20) dB (20 to 0) dB (0 to -58) dB (-58 to -78) dB (-78 to -110) dB (-110 to -115) dB (-115 to -120) dB (-120 to -125) dB (2 000 to 3 050) MHz (30 to 20) dB (20 to 0) dB (0 to -58) dB (-58 to -78) dB (-78 to -110) dB (-110 to -115) dB (-115 to -120) dB (-120 to -125) dB (3 050 to 6 600) MHz (30 to 20) dB (20 to 0) dB (0 to -58) dB (-58 to -78) dB (-78 to -110) dB (-110 to -115) dB (-115 to -120) dB (-120 to -125) dB	0.36 dB 0.2 dB 0.22 dB 0.23 dB 0.25 dB 0.27 dB 0.33 dB 0.46 dB 0.42 dB 0.3 dB 0.31 dB 0.32 dB 0.34 dB 0.35 dB 0.4 dB 0.51 dB 0.42 dB 0.3 dB 0.31 dB 0.32 dB 0.34 dB 0.38 dB 0.48 dB 0.64 dB	Agilent N5531S Measuring Receiver w/N5532A Opt 526 Power Sensor



ANSI National Accreditation Board

Electrical – RF/Microwave

Fort Lauderdale, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power Measure <sup>1</sup>	(6 600 to 13 200) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 526 Power Sensor
	(30 to 20) dB	0.42 dB	
	(20 to 0) dB	0.3 dB	
	(0 to -58) dB	0.31 dB	
	(-58 to -78) dB	0.32 dB	
	(-78 to -100) dB	0.34 dB	
	(-100 to -105) dB	0.37 dB	
	(-105 to -110) dB	0.45 dB	
	(-110 to -115) dB	0.6 dB	
	(-115 to -120) dB	0.82 dB	
	(13 200 to 18 000) MHz		
	(30 to 20) dB	0.42 dB	
	(20 to 0) dB	0.3 dB	
	(0 to -58) dB	0.31 dB	
	(-58 to -78) dB	0.32 dB	
	(-78 to -90) dB	0.33 dB	
	(-90 to -95) dB	0.35 dB	
	(-95 to -100) dB	0.41 dB	
	(-100 to -105) dB	0.53 dB	
	(-105 to -110) dB	0.72 dB	
	(18 000 to 19 200) MHz		
	(30 to 20) dB	0.48 dB	
	(20 to 0) dB	0.38 dB	
	(0 to -58) dB	0.39 dB	
	(-58 to -78) dB	0.4 dB	
	(-78 to -90) dB	0.41 dB	
	(-90 to -95) dB	0.42 dB	
	(-95 to -100) dB	0.47 dB	
	(-100 to -105) dB	0.58 dB	
	(-105 to -110) dB	0.75 dB	
	(19 200 to 26 500) MHz		
	(30 to 20) dB	0.48 dB	
	(20 to 0) dB	0.38 dB	
	(0 to -58) dB	0.39 dB	
	(-58 to -78) dB	0.4 dB	
	(-78 to -90) dB	0.43 dB	
(-90 to -95) dB	0.5 dB		
(-95 to -100) dB	0.63 dB		
(-100 to -105) dB	0.84 dB		
(-105 to -110) dB	1.1 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Fort Lauderdale, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Amplitude Modulation – Measure <sup>1</sup>	100 kHz to 10 MHz Rate 50 Hz to 10 kHz (5 to 99) % Depth	0.75 % of reading + 0.3 digits	Agilent N5531S Measuring Receiver
	10 MHz to 3 GHz Rate 50 Hz to 100 kHz (5 to 20) % Depth (20 to 99) % Depth	2.5 % of reading + 0.4 digits 0.5 % of reading + 0.4 digits	
	(3 to 26.5) GHz Rate 50 Hz to 100 kHz (5 to 20) % Depth (20 to 99) % Depth	4.5 % of reading + 0.4 digits 1.5 % of reading + 0.4 digits	
Frequency Modulation Measure <sup>1</sup>  $\beta$ = deviation / rate	250 kHz to 10 MHz Rates 20 Hz to 10 kHz Peak Dev 200 to 40 kHz	$\beta > 0.2$ - 1.5 % of reading + 2 Hz $\beta > 1.2$ - 1 % of reading + 2 Hz	Agilent N5531S Measuring Receiver
	10 MHz to 6.6 GHz Rates 50 Hz to 200 kHz Peak Dev 250 to 400 kHz (6.6 to 13.2) GHz	$\beta > 0.2$ - 1.5 % of reading + 2 Hz $\beta > 0.45$ - 1 % of reading + 2 Hz	
	Rates 50 Hz to 200 kHz Peak Dev 250 to 400 kHz (13.2 to 26.5) GHz	$\beta > 0.2$ - 2.5 % of reading + 4 Hz $\beta > 8$ - 1 % of reading + 4 Hz	
	Rates 50 Hz to 200 kHz Peak Dev 250 to 400kHz	$\beta > 0.2$ - 3.8 % of reading + 9 Hz $\beta > 16$ - 1 % of reading + 9 Hz	
Phase Modulation – Measure <sup>1</sup>	100 kHz to 6.6 GHz Deviations > 0.3 rad Deviations > 0.7 rad	3 % of reading + 0.002 rad 1 % of reading + 0.002 rad	Agilent N5531S Measuring Receiver
	(6.6 to 13.2) GHz Deviations > 0.6 rad Deviations > 2.0 rad	3 % of reading + 0.005 rad 1 % of reading + 0.005 rad	
	(13.2 to 26.5) GHz Deviations: > 1.2 rad Deviations > 4.0 rad	3 % of reading + 0.009 rad 1 % of reading + 0.009 rad	
AM Distortion Measure <sup>1</sup> Rate 20 Hz to 1 kHz	(0.1 to 10) MHz AM Depth > 1% (0 to -20) dB	1.2 dB	Agilent N5531S Measuring Receiver
	(-20 to -30) dB AM Depth > 3% (0 to -20) dB	2.2 dB 1 dB	
	(-20 to -30) dB (-30 to -40) dB	1.3 dB 2.4 dB	



ANSI National Accreditation Board

Electrical – RF/Microwave

Fort Lauderdale, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AM Distortion Measure <sup>1</sup> Rate 20 Hz to 1 kHz	10 MHz to 26.5 GHz		Agilent N5531S Measuring Receiver
	AM Depth > 1%		
	(0 to -20) dB	1.3 dB	
	(-20 to -30) dB	2.5 dB	
	AM Depth > 3%		
FM Distortion Measure <sup>1</sup> Rate 20 Hz to 1 kHz	(0 to -20) dB	1.1 dB	Agilent N5531S Measuring Receiver
	(-20 to -30) dB	1.4 dB	
	(-30 to -40) dB	3 dB	
	(1 to 6 600) MHz		
	Dev 500 Hz to 2 kHz		
	(0 to -20) dB	0.26 dB	
	(-20 to -30) dB	0.79 dB	
	(-30 to -40) dB	2.3 dB	
	Dev > 2 kHz		
	(0 to -20) dB	0.09 dB	
	(-20 to -30) dB	0.27 dB	
	(-30 to -40) dB	0.83 dB	
	(-40 to -50) dB	2.4 dB	
	(6.6 to 13.2) GHz		
	Dev > 2.3 kHz		
	(0 to -20) dB	0.26 dB	
	(-20 to -30) dB	0.79 dB	
	(-30 to -40) dB	2.3 dB	
	Dev > 4.5 kHz		
	(0 to -20) dB	0.09 dB	
(-20 to -30) dB	0.27 dB		
(-30 to -40) dB	0.83 dB		
(-40 to -50) dB	2.4 dB		
(13.2 to 26.5) GHz			
Dev > 2.7 kHz			
(0 to -20) dB	0.26 dB		
(-20 to -30) dB	0.79 dB		
(-30 to -40) kHz	2.3 dB		
Dev > 6.0 kHz			
(0 to -20) dB	0.09 dB		
(-20 to -30) dB	0.27 dB		
(-30 to -40) dB	0.83 dB		
(-40 to -50) dB	2.4 dB		

**Electrical – RF/Microwave**

Fort Lauderdale, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Total Harmonic Distortion (THD)	(0 to -60) dB		HP 8903B Audio Analyzer
	20 Hz to 20 kHz		
	(0 to -40) dB	1 dB	
	(-40 to -50) dB	1 dB	
	(-50 to -60) dB	1.3 dB	
	(-60 to -65) dB	1.7 dB	
	(20 to 50) kHz		
	(0 to -40) dB	2 dB	
	(-40 to -50) dB	2.1 dB	
	(-50 to -60) dB	3 dB	
Harmonics Measure <sup>1</sup>	(50 to 100) kHz		Agilent E4440A Measuring Receiver
	(0 to -40) dB	2 dB	
	(-40 to -50) dB	2.1 dB	
	(-50 to -60) dB	3 dB	
	(-80 to -10) dB		
	2 <sup>nd</sup> through 5 <sup>th</sup> Harmonic		
	1 kHz to 600 MHz	0.37 dB	
	(600 to 1 320) MHz	1.1 dB	
	(1 320 to 2 200) MHz	1.4 dB	
	(2 200 to 3 000) MHz	1.4 dB	
(3 000 to 4 400) MHz	1.7 dB		
(4 400 to 5 300) MHz	1.9 dB		
2 <sup>nd</sup> through 4 <sup>th</sup> Harmonic			
(5 300 to 6 625) MHz	2.1 dB		
2 <sup>nd</sup> through 3 <sup>rd</sup> Harmonic			
(6 625 to 8 833) MHz	2.1 dB		
2 <sup>nd</sup> Harmonic			
(8 833 to 13 250) MHz	2.1 dB		

**Length – Dimensional Metrology**

Fort Lauderdale, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Calipers <sup>1,2</sup>	Up to 46 in	$(281 + 1.8L) \mu\text{in}$	Gage blocks (Grade 2)
Micrometers <sup>1,2</sup>	Up to 46 in	$(31 + 4.2L) \mu\text{in}$	
Dial Indicators <sup>1,2</sup> Resolution: $\geq 50\mu\text{in}$ $< 50\mu\text{in}$	Up to 10 in	$(26 + 3.3L) \mu\text{in}$	
	Up to 0.1 in	12 $\mu\text{in}$	
Height Gages <sup>1,2</sup>	Up to 46 in	$(236 + 1.5L) \mu\text{in}$	



ANSI National Accreditation Board

**Length – Dimensional Metrology**

Fort Lauderdale, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Scales – Rulers <sup>1</sup>	Up to 46 in	0.009 1 in	Gage blocks (Grade 2)
Tape Measures	Up to 100 ft	(0.025 + 0.000 26 <i>F</i> ) in	Standard rule
Protractors <sup>1</sup>	(0 to 360) <sup>o</sup>	0.014 <sup>o</sup>	Angle blocks
Feeler Gages <sup>1</sup>	Up to 1 in	30 μin	P&W Model C Bench Micrometer
Cylindrical Gages <sup>2</sup> – Inside Outside	Up to 11 in Up to 10 in	(14 + 2 <i>D</i> ) μin (4.3 + 2.3 <i>D</i> ) μin	Gage blocks, P&W universal measuring machine
Gage Blocks <sup>2</sup> – Length Only	Up to 10 in	(3.6 + 2.9 <i>L</i> ) μin	Master gage blocks, P&W universal measuring machine
Thread Plugs – Major Diameter Pitch Diameter	Up to 10 in Up to 10 in	47 μin 80 μin	P & W Model C Bench Micrometer, Van Keuren thread wire set, gage blocks

**Mass and Mass Related**

Fort Lauderdale, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Scales & Balances <sup>1,2,4</sup>	0.001 g to 10 kg	(0.02 + 0.003 <i>X</i> ) mg	Class 1 weights
	Up to 400 lb Up to 181 kg	(0.000 5 + 0.000 13 <i>M</i> ) lb (0.22 + 0.000 13 <i>X</i> ) g	Class F weights
Torque – Measure <sup>1</sup>	(10 to 100) ozf·in 5 lbf·in to 250 lbf·ft (250 to 1 000) lbf·ft	0.59 % of reading 0.32 % of reading 0.32 % of reading	Torque Testers CDI 1001-0-DDT CDI 5000-ST CDI 2000-13-02
Pressure <sup>1</sup>	(0.2 to 1 000) psi	0.002 2 % of reading	Ruska 2468A Deadweight Tester
Pressure <sup>1</sup>	(-15 to 100) psi	0.082 psi	Fluke 74x Series Process Calibrator w/700PD6 Pressure Module
Pressure <sup>1</sup>	(100 to 500) psi	0.31 psi	Fluke 700G07 Pressure Gauge



ANSI National Accreditation Board

**Mass and Mass Related**

Fort Lauderdale, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Pressure <sup>1</sup>	(-150 to 150) inH <sub>2</sub> O (0 to 15) psi (500 to 1 000) psi (1 000 to 10 000) psi	0.09 inH <sub>2</sub> O 0.004 psi 0.6 psi 6.1 psi	Additel 681 Pressure Gage
Force <sup>1</sup> Tension and Compression	Up to 10 kgf (0.1 to 300) lbf	(5.7 + 0.002 9X) mg 0.086 % of reading	Class 1 weights Class F weights

**Thermodynamic**

Fort Lauderdale, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Temperature – Measure <sup>1</sup>	(-200 to 300) °C	0.031 °C	Hart 1502 Indicator with 5622 PRT
Temperature Measuring Equipment <sup>1</sup>	(-25 to 140) °C (140 to 300) °C	0.13 °C 0.42 °C	Hart 1502 Indicator with 5622 PRT and dry block
Infrared (IR) Temperature <sup>1</sup>	(20 to 100) °C (100 to 300) °C (300 to 420) °C (420 to 500) °C	1.5 °C 4.3 °C 6 °C 7.7 °C	Fluke 9132 Portable Infrared Calibrator ε= 0.95, λ = (8 to 14) μm
Relative Humidity Measuring Equipment <sup>1</sup>	11 %RH 33 %RH 75 %RH 97 %RH	1.2 %RH 1.1 %RH 1.3 %RH 2.8 %RH	Standard salts
Relative Humidity Generate	(10 to 95) %RH	0.5 %RH	Thunder Scientific 2500 Humidity Chamber
Relative Humidity – Measure <sup>1</sup>	Up to 90 %RH (90 to 97) %RH	1.4 %RH 2.4 %RH	Vaisala HM141/HMP46 Indicator and Probe

**Time and Frequency**

Fort Lauderdale, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency – Generate <sup>1</sup>	10 MHz	1 x 1 <sup>-12</sup> Hz/Hz	HP 58503A GPS Receiver



ANSI National Accreditation Board

Time and Frequency

Fort Lauderdale, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency – Generate <sup>1</sup>	(1 to 10) Hz (10 to 100) Hz (100 to 1 000) Hz (1 to 10) kHz (10 to 100) kHz (0.1 to 1) MHz (1 to 10) MHz	$1 \times 10^{-12}$ Hz/Hz + 0.57 $\mu$ Hz $1 \times 10^{-12}$ Hz/Hz + 5.7 $\mu$ Hz $1 \times 10^{-12}$ Hz/Hz + 57 $\mu$ Hz $1 \times 10^{-12}$ Hz/Hz + 0.57 mHz $1 \times 10^{-12}$ Hz/Hz + 5.7 mHz $1 \times 10^{-12}$ Hz/Hz + 57 mHz $1 \times 10^{-12}$ Hz/Hz + 0.57 Hz	Agilent 33250A Function Generator / HP 58503A GPS Receiver
Frequency – Generate <sup>1</sup>	(10 to 20 000) MHz	$1 \times 10^{-12}$ Hz/Hz + 0.57 mHz	Agilent E8257D Opt 520 Signal Generator / HP 58503A GPS Receiver
Frequency – Generate <sup>1</sup>	(10 to 26 500) MHz	$1 \times 10^{-12}$ Hz/Hz + 0.57 kHz	HP 83630B Signal Generator / HP 58503A GPS Receiver
Time – Generate	1 pps	$1 \times 10^{-12}$ s/s + 750 ps	HP 58503A GPS Receiver
Frequency – Measure <sup>1</sup>	(1 to 10) Hz (10 to 100) Hz (100 to 1 000) Hz (1 to 10) kHz (10 to 100) kHz (100 to 200) kHz (0.2 to 3 000) MHz	$4.2 \times 10^{-9}$ Hz/Hz $1.5 \times 10^{-9}$ Hz/Hz $0.6 \times 10^{-9}$ Hz/Hz $0.33 \times 10^{-9}$ Hz/Hz $0.24 \times 10^{-9}$ Hz/Hz $0.21 \times 10^{-9}$ Hz/Hz $0.21 \times 10^{-9}$ Hz/Hz	Agilent 53131A Opt 030 Frequency Counter / HP 58503A GPS Receiver
Frequency – Measure <sup>1</sup>	(10 to 26 500) MHz	$1 \times 10^{-12}$ Hz/Hz + 0.1 Hz	Agilent E4440A Spectrum Analyzer / HP 58503A GPS Receiver
Timers and Stopwatches <sup>1</sup>	Up to 24 hr	346 ms	GPS Reference Frequency Counter
Tachometers – RPM	Up to 100 000 RPM	$\pm (0.001\% + 0.6R)$	HP 33250A Signal Generator & LED

[Return to Site listing](#) (top) [Go to Notes](#) (bottom)



ANSI National Accreditation Board

## Services performed at satellite laboratory

### Technical Maintenance, Inc.

3060 Venture Lane, Suite 106  
Melbourne, FL 32934

Schuyler Cournoyer (Branch Manager) Phone: 321-242-0890  
Scott Chamberlain (Quality Manager) Phone: 321-242-0890

## CALIBRATION

### Chemical Quantities

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
pH Meters <sup>1,5</sup>	(4, 7) pH (10) pH	0.02 pH 0.03 pH	pH buffer solutions
Conductivity Meters <sup>1,5</sup>	84 $\mu$ S/cm 1 413 $\mu$ S/cm 10 000 $\mu$ S/cm	0.83 $\mu$ S/cm 5.7 $\mu$ S/cm 37 $\mu$ S/cm	Conductivity solutions

### Electrical – DC/Low Frequency

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Current – Generate <sup>1</sup>	(1 to 2.2) nA (2.2 to 22) nA (22 to 220) nA (0.22 to 2.2) $\mu$ A (2.2 to 10) $\mu$ A	93 $\mu$ A/A + 0.007 nA 92 $\mu$ A/A + 0.007 nA 92 $\mu$ A/A + 0.01 nA 36 $\mu$ A/A + 0.1 nA 15 $\mu$ A/A + 1 nA	Fluke 5730A/03 Multifunction Calibrator, Agilent 3458 Option 002 Precision Multimeter
DC Current – Generate <sup>1</sup>	(10 to 220) $\mu$ A (0.22 to 2.2) mA (2.2 to 22) mA (22 to 100) mA (100 to 220) mA (0.22 to 1) A (1 to 2.2) A	38 $\mu$ A/A + 5 nA 30 $\mu$ A/A + 7 nA 30 $\mu$ A/A + 44 nA 38 $\mu$ A/A + 0.7 $\mu$ A 45 $\mu$ A/A + 0.7 $\mu$ A 68 $\mu$ A/A + 12 $\mu$ A 105 $\mu$ A/A + 12 $\mu$ A	Fluke 5730A/03 Multifunction Calibrator
DC Current – Generate <sup>1</sup>	(2.2 to 11) A	274 $\mu$ A/A + 365 $\mu$ A	Fluke 5730A/03 Multifunction Calibrator and 5725A Amplifier
DC Current – Generate <sup>1</sup>	(11 to 20.5) A	761 $\mu$ A/A + 578 $\mu$ A	Fluke 5522A Multifunction Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Current Clamp Meters Toroidal-Wound	(0 to 1025) A	0.21 % of output + 0.05 A	Fluke 5522A Multiproduct Calibrator / 5500A/Coil x50
DC Current Clamp Meters Other	(0 to 1 025) A	0.39 % of output + 0.38A	Fluke 5522A Multiproduct Calibrator / 5500A/Coil x50
DC Current – Measure <sup>1</sup>	(1 to 10) nA (10 to 100) nA (0.1 to 1) μA (1 to 10) μA	35 μA/A + 0.12 pA 13 μA/A + 1.2 pA 9.1 μA/A + 0.01 nA 7.9 μA/A + 0.12 nA	Fluke 5730A Multifunction Calibrator Agilent 3458A Option 002 Precision Multimeter
DC Current – Measure <sup>1</sup>	(10 to 100) μA (0.1 to 1) mA (1 to 10) mA (10 to 100) mA (0.1 to 1) A	20 μA/A + 0.8 nA 20 μA/A + 5 nA 20 μA/A + 0.05 μA 35 μA/A + 0.5 μA 110 μA/A + 10 μA	Agilent 3458A Multimeter
DC Current – Measure <sup>1</sup>	(1 to 10) A (10 to 100) A	100 μA/A + 72 μA 100 μA/A + 1 mA	Agilent 3458A Option 002 Multimeter Standard Resistor L&N 4361 Current Shunt
DC Current – Measure <sup>1</sup>	(100 to 1 000) A	0.25% of reading	Agilent 3458A Multimeter Empro Current Shunt
DC Voltage – Generate <sup>1</sup>	(0 to 220) mV (0.22 to 2.2) V (2.2 to 11) V (11 to 22) V (22 to 220) V (220 to 1 100) V	6.8 μV/V + 0.8 μV 4.6 μV/V + 0.9 μV 3.0 μV/V + 2.5 μV 3.0 μV/V + 3.9 μV 4.6 μV/V + 38 μV 6.1 μV/V + 385 μV	Fluke 5730A Multifunction Calibrator
DC Voltage – Measure <sup>1</sup>	(0 to 100) mV (0.1 to 1) V (1 to 10) V (10 to 100) V (100 to 1 000) V	5 μV/V + 1.2 μV 4 μV/V + 1.2 μV 4 μV/V + 2.1 μV 6 μV/V + 30 μV 6 μV/V + 100 μV + 12 μV/V x (Vin/1 000) ^2	Agilent 3458A Option 002 Precision Multimeter
DC Voltage – Measure <sup>1</sup>	(1 to 60) kV	0.1 % of reading	Ross VD60 High Voltage Divider, HP 34401A Precision Multimeter



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Generate <sup>1</sup>	(0.22 to 2.2) mV		Fluke 5730A Multifunction Calibrator
	(10 to 20) Hz	228 $\mu\text{V}/\text{V} + 3.9 \mu\text{V}$	
	(20 to 40) Hz	88 $\mu\text{V}/\text{V} + 3.9 \mu\text{V}$	
	40 Hz to 20 kHz	76 $\mu\text{V}/\text{V} + 3.9 \mu\text{V}$	
	(20 to 50) kHz	190 $\mu\text{V}/\text{V} + 3.9 \mu\text{V}$	
	(50 to 100) kHz	457 $\mu\text{V}/\text{V} + 4.6 \mu\text{V}$	
	(100 to 300) kHz	989 $\mu\text{V}/\text{V} + 9.2 \mu\text{V}$	
	(300 to 500) kHz	1.3 mV/V + 19 $\mu\text{V}$	
	500 kHz to 1 MHz	2.6 mV/V + 19 $\mu\text{V}$	
	(2.2 to 22) mV		
	(10 to 20) Hz	228 $\mu\text{V}/\text{V} + 3.9 \mu\text{V}$	
	(20 to 40) Hz	88 $\mu\text{V}/\text{V} + 3.9 \mu\text{V}$	
	40 Hz to 20 kHz	76 $\mu\text{V}/\text{V} + 3.9 \mu\text{V}$	
	(20 to 50) kHz	190 $\mu\text{V}/\text{V} + 3.9 \mu\text{V}$	
	(50 to 100) kHz	457 $\mu\text{V}/\text{V} + 4.6 \mu\text{V}$	
	(100 to 300) kHz	989 $\mu\text{V}/\text{V} + 9.2 \mu\text{V}$	
	(300 to 500) kHz	1.3 mV/V + 19 $\mu\text{V}$	
	500 kHz to 1 MHz	2.6 mV/V + 19 $\mu\text{V}$	
	(22 to 220) mV		
	(10 to 20) Hz	228 $\mu\text{V}/\text{V} + 11.4 \mu\text{V}$	
	(20 to 40) Hz	88 $\mu\text{V}/\text{V} + 6.1 \mu\text{V}$	
	40 Hz to 20 kHz	53 $\mu\text{V}/\text{V} + 6.1 \mu\text{V}$	
	(20 to 50) kHz	114 $\mu\text{V}/\text{V} + 6.1 \mu\text{V}$	
	(50 to 100) kHz	304 $\mu\text{V}/\text{V} + 15 \mu\text{V}$	
	(100 to 300) kHz	609 $\mu\text{V}/\text{V} + 19 \mu\text{V}$	
	(300 to 500) kHz	1.3 mV /V + 22.8 $\mu\text{V}$	
	500 kHz to 1 MHz	2.5 mV /V + 45.7 $\mu\text{V}$	
220 mV to 2.2 V			
(10 to 20) Hz	228 $\mu\text{V}/\text{V} + 38 \mu\text{V}$		
(20 to 40) Hz	84 $\mu\text{V}/\text{V} + 15 \mu\text{V}$		
40 Hz to 20 kHz	37 $\mu\text{V}/\text{V} + 8 \mu\text{V}$		
(20 to 50) kHz	61 $\mu\text{V}/\text{V} + 9 \mu\text{V}$		
(50 to 100) kHz	76 $\mu\text{V}/\text{V} + 30 \mu\text{V}$		
(100 to 300) kHz	304 $\mu\text{V}/\text{V} + 76 \mu\text{V}$		
(300 to 500) kHz	913 $\mu\text{V} /\text{V} + 190 \mu\text{V}$		
500 kHz to 1 MHz	1.5 mV/V + 304 $\mu\text{V}$		

**Electrical – DC/Low Frequency**

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Generate <sup>1</sup>	(2.2 to 22) V (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz (22 to 220) V (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz	228 $\mu$ V/V + 380 $\mu$ V 84 $\mu$ V/V + 152 $\mu$ V 37 $\mu$ V/V + 54 $\mu$ V 61 $\mu$ V/V + 91 $\mu$ V 76 $\mu$ V/V + 190 $\mu$ V 228 $\mu$ V/V + 609 $\mu$ V 913 $\mu$ V/V + 1.9 mV 1.4 mV/V + 3 mV 228 $\mu$ V/V + 3.8 mV 84 $\mu$ V/V + 1.5 mV 49 $\mu$ V/V + 0.6 mV 76 $\mu$ V/V + 0.9 mV 137 $\mu$ V/V + 2.3 mV	Fluke 5730A Multifunction Calibrator
AC Voltage – Generate <sup>1</sup>	(220 to 750) V 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (750 to 1 000) V 40 Hz to 1 kHz (1 to 20) kHz (20 to 30) kHz	68 $\mu$ V/V + 3 mV 126 $\mu$ V/V + 5 mV 457 $\mu$ V/V + 8 mV 1.8 mV/V + 34 mV 68 $\mu$ V/V + 3 mV 126 $\mu$ V/V + 5 mV 457 $\mu$ V/V + 8 mV	Fluke 5730A Multifunction Calibrator, 5725A Amplifier
AC Voltage – Generate <sup>1</sup> Wideband Absolute	(0.3 to 1.1) mV (10 to 30 Hz) 30 Hz to 500 kHz (0.5 to 1.2) MHz (1.2 to 2) MHz (2 to 12) MHz (12 to 20) MHz (20 to 30) MHz (1.1 to 3.3) mV (10 to 30 Hz) 30 Hz to 500 kHz (0.5 to 1.2) MHz (1.2 to 2) MHz (2 to 12) MHz (12 to 20) MHz (20 to 30) MHz	0.65 % of output + 1.5 $\mu$ V 0.61 % of output + 1.5 $\mu$ V 0.63 % of output + 3.8 $\mu$ V 0.63 % of output + 3.8 $\mu$ V 0.68 % of output + 3.8 $\mu$ V 0.76 % of output + 3.8 $\mu$ V 1.3 % of output + 13 $\mu$ V 0.58 % of output + 2.3 $\mu$ V 0.53 % of output + 2.3 $\mu$ V 0.54 % of output + 4.6 $\mu$ V 0.54 % of output + 4.6 $\mu$ V 0.58 % of output + 4.6 $\mu$ V 0.65 % of output + 4.6 $\mu$ V 1.3 % of output + 4.6 $\mu$ V	Fluke 5730A Option 003 Multifunction Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Generate <sup>1</sup> Wideband Absolute	(3.3 to 11) mV		Fluke 5730A Option 003 Multifunction Calibrator
	(10 to 30 Hz)	0.58 % of output + 6.1 μV	
	30 Hz to 500 kHz	0.53 % of output + 6.1 μV	
	(0.5 to 1.2) MHz	0.54 % of output + 8.4 μV	
	(1.2 to 2) MHz	0.54 % of output + 8.4 μV	
	(2 to 12) MHz	0.55 % of output + 8.4 μV	
	(12 to 20) MHz	0.61 % of output + 8.4 μV	
	(20 to 30) MHz	0.93 % of output + 8.4 μV	
	(11 to 33) mV		
	(10 to 30 Hz)	0.52 % of output + 12 μV	
	30 Hz to 500 kHz	0.46 % of output + 12 μV	
	(0.5 to 1.2) MHz	0.47 % of output + 14 μV	
	(1.2 to 2) MHz	0.47 % of output + 14 μV	
	(2 to 12) MHz	0.49 % of output + 14 μV	
	(12 to 20) MHz	0.55 % of output + 14 μV	
	(20 to 30) MHz	0.89 % of output + 14 μV	
	(33 to 110) mV		
	(10 to 30 Hz)	0.52 % of output + 30 μV	
	30 Hz to 500 kHz	0.46 % of output + 30 μV	
	(0.5 to 1.2) MHz	0.47 % of output + 33 μV	
	(1.2 to 2) MHz	0.47 % of output + 33 μV	
	(2 to 12) MHz	0.49 % of output + 33 μV	
	(12 to 20) MHz	0.55 % of output + 33 μV	
	(20 to 30) MHz	0.89 % of output + 33 μV	
	(110 to 330) mV		
	(10 to 30 Hz)	0.45 % of output + 0.1 mV	
	30 Hz to 500 kHz	0.38 % of output + 0.1 mV	
	(0.5 to 1.2) MHz	0.4 % of output + 0.1 mV	
(1.2 to 2) MHz	0.4 % of output + 0.1 mV		
(2 to 12) MHz	0.42 % of output + 0.1 mV		
(12 to 20) MHz	0.49 % of output + 0.1 mV		
(20 to 30) MHz	0.85 % of output + 0.1 mV		
0.33 to 1.1 V			
(10 to 30 Hz)	0.45 % of output + 0.3 mV		
30 Hz to 500 kHz	0.38 % of output + 0.3 mV		
(0.5 to 1.2) MHz	0.4 % of output + 0.3 mV		
(1.2 to 2) MHz	0.4 % of output + 0.3 mV		
(2 to 12) MHz	0.42 % of output + 0.3 mV		
(12 to 20) MHz	0.49 % of output + 0.3 mV		
(20 to 30) MHz	0.85 % of output + 0.3 mV		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Generate <sup>1</sup> Wideband Absolute	(1.1 to 3.5) V (10 to 30) Hz 30 Hz to 500 kHz (0.5 to 1.2) MHz (1.2 to 2) MHz (2 to 12) MHz (12 to 20) MHz (20 to 30) MHz	0.39 % of output + 0.4 mV 0.3 % of output + 0.4 mV 0.32 % of output + 0.4 mV 0.32 % of output + 0.4 mV 0.35 % of output + 0.4 mV 0.44 % of output + 0.4 mV 0.82 % of output + 0.4 mV	Fluke 5730A Option 003 Multifunction Calibrator
AC Voltage – Measure <sup>1</sup>	(1 to 10) mV (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz 100 kHz to 1 MHz (1 to 4) MHz (4 to 8) MHz (10 to 100) mV (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz (2 to 4) MHz (4 to 8) MHz (8 to 10) MHz (0.1 to 1) V (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz (2 to 4) MHz (4 to 8) MHz (8 to 10) MHz	0.03 % of reading + 3 μV 0.02 % of reading + 1.1 μV 0.03 % of reading + 1.1 μV 0.1 % of reading + 1.1 μV 0.5 % of reading + 1.1 μV 1.2 % of reading + 5 μV 7 % of reading + 7 μV 20 % of reading + 8 μV 0.007 % of reading + 4 μV 0.007 % of reading + 2 μV 0.014 % of reading + 2 μV 0.03 % of reading + 2 μV 0.08 % of reading + 2 μV 0.3 % of reading + 10 μV 1 % of reading + 10 μV 1.5 % of reading + 70 μV 4 % of reading + 70 μV 4 % of reading + 80 μV 15 % of reading + 100 μV 0.007 % of reading + 40 μV 0.07 % of reading + 20 μV 0.014 % of reading + 20 μV 0.03 % of reading + 20 μV 0.08 % of reading + 20 μV 0.3 % of reading + 100 μV 1 % of reading + 100 μV 1.5 % of reading + 0.7 mV 4 % of reading + 0.7 mV 4 % of reading + 0.8 mV 15 % of reading + 1 mV	Agilent 3458A/002 Precision Multimeter



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure <sup>1</sup>	(1 to 10) V		Agilent 3458A Precision Multimeter
	(1 to 40) Hz	0.007 % of reading + 0.4 mV	
	40 Hz to 1 kHz	0.007 % of reading + 0.2 mV	
	(1 to 20) kHz	0.014 % of reading + 0.2 mV	
	(20 to 50) kHz	0.03 % of reading + 0.2 mV	
	(50 to 100) kHz	0.08 % of reading + 0.2 mV	
	(100 to 300) kHz	0.3 % of reading + 1 mV	
	300 kHz to 1 MHz	1 % of reading + 1 mV	
	(1 to 2) MHz	1.5 % of reading + 7 mV	
	(2 to 4) MHz	4 % of reading + 7 mV	
	(4 to 8) MHz	4 % of reading + 8 mV	
	(8 to 10) MHz	15 % of reading + 10 mV	
	(10 to 100) V		
	(1 to 40) Hz	0.02 % of reading + 4 mV	
	40 Hz to 1 kHz	0.02 % of reading + 2 mV	
(1 to 20) kHz	0.02 % of reading + 2 mV		
(20 to 50) kHz	0.035 % of reading + 2 mV		
(50 to 100) kHz	0.12 % of reading + 2 mV		
(100 to 300) kHz	0.4 % of reading + 10 mV		
300 kHz to 1 MHz	1.5 % of reading + 10 mV		
(100 to 700) V	(1 to 40) Hz	0.04 % of reading + 40 mV	
	40 Hz to 1 kHz	0.04 % of reading + 20 mV	
	(1 to 20) kHz	0.06 % of reading + 20 mV	
	(20 to 50) kHz	0.12 % of reading + 20 mV	
	(50 to 100) kHz	0.3 % of reading + 20 mV	
AC Voltage – Measure <sup>1</sup>	(1 to 10) kV 60 Hz	0.5 % of reading + 0.002 kV	Ross VD60 High Voltage Divider, HP 34401A Multimeter
	(10 to 42) kV 60 Hz	0.5 % of reading + 0.02 kV	
AC Current – Generate <sup>1</sup>	(9 to 220) $\mu$ A		Fluke 5730A Multifunction Calibrator
	(10 to 20) Hz	228 $\mu$ A/A + 15 nA	
	(20 to 40) Hz	152 $\mu$ A/A + 10 nA	
	40 Hz to 1 kHz	91 $\mu$ A/A + 8 nA	
	(1 to 5) kHz	266 $\mu$ A/A + 12 nA	
(5 to 10) kHz	989 $\mu$ A/A + 61 nA		

**Electrical – DC/Low Frequency**

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment						
AC Current – Generate <sup>1</sup>	0.22 to 2.2 mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	228 $\mu$ A/A + 39 nA 152 $\mu$ A/A + 31 nA 91 $\mu$ A/A + 31 nA 183 $\mu$ A/A + 99 nA 989 $\mu$ A/A + 609 nA	Fluke 5730A Multifunction Calibrator						
	2.2 to 22 mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	228 $\mu$ A/A + 385 nA 152 $\mu$ A/A + 310 nA 91 $\mu$ A/A + 310 nA 183 $\mu$ A/A + 536 nA 989 $\mu$ A/A + 4566 nA							
	22 to 220 mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	228 $\mu$ A/A + 4 $\mu$ A 152 $\mu$ A/A + 3 $\mu$ A 91 $\mu$ A/A + 2 $\mu$ A 183 $\mu$ A/A + 3 $\mu$ A 989 $\mu$ A/A + 9 $\mu$ A							
	(0.22 to 2.2) A 20 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	228 $\mu$ A/A + 31 $\mu$ A 380 $\mu$ A/A + 76 $\mu$ A 6.1 mA/A + 152 $\mu$ A							
	AC Current – Generate <sup>1</sup>	(2.2 to 11) A 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz		350 $\mu$ A/A + 141 $\mu$ A 723 $\mu$ A/A + 295 $\mu$ A 2.7 mA/A + 573 $\mu$ A	Fluke 5730A Multifunction Calibrator 5725A Amplifier				
		AC Current – Generate <sup>1</sup>		(11 to 20.5) A (45 to 100) Hz 100 Hz to 1 kHz (1 to 5) kHz		0.09 % of output + 5 mA 0.11 % of output + 5 mA 2.3 % of output + 5 mA	Fluke 5522A Multiproduct Calibrator		
				AC Current Clamps – Toroidal-Wound		(16.5 to 1025) A (45 to 65) Hz		0.23 % of reading + 0.1 A	Fluke 5522A Multiproduct Calibrator /5500A/Coil x50
						(16.5 to 150) A (65 to 440) Hz		0.6 % of reading + 0.085 A	
	AC Current Clamps – Other	(16.5 to 1025) A (45 to 65) Hz		0.44 % of reading +0.53 A	Fluke 5522A Multiproduct Calibrator /5500A/Coil x50				
		(16.5 to 150) A (65 to 440) Hz		0.79 % of reading +0.68 A					



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment		
AC Current – Measure <sup>1</sup>	Up to 100 $\mu$ A (10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz	0.4 % of reading + 0.03 $\mu$ A 0.15 % of reading + 0.03 $\mu$ A 0.06 % of reading + 0.03 $\mu$ A	Agilent 3458A Precision Multimeter		
	100 $\mu$ A to 1 mA (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz	0.4 % of reading + 0.2 $\mu$ A 0.15 % of reading + 0.2 $\mu$ A 0.06 % of reading + 0.2 $\mu$ A 0.03 % of reading + 0.2 $\mu$ A			
	(1 to 10) mA (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz	0.4 % of reading + 2 $\mu$ A 0.15 % of reading + 2 $\mu$ A 0.06 % of reading + 2 $\mu$ A 0.03 % of reading + 2 $\mu$ A			
	(10 to 100) mA (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz	0.4 % of reading + 20 $\mu$ A 0.15 % of reading + 20 $\mu$ A 0.06 % of reading + 20 $\mu$ A 0.03 % of reading + 20 $\mu$ A			
	100 mA to 1 A (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 Hz	0.4 % of reading + 0.2 mA 0.15 % of reading + 0.2 mA 0.06 % of reading + 0.2 mA 0.03 % of reading + 0.2 mA			
	(1 to 3) A (3 to 5) Hz (5 to 10) Hz 10 Hz to 5 kHz	0.72 % of reading + 1.2 mA 0.23 % of reading + 1.2 mA 0.1 % of reading + 1.2 mA		Agilent 34401A Multimeter	
	(3 to 30) A 40 Hz to 1 kHz (1 to 5) kHz	0.3 % of reading + 0.07 A 5 % of reading + 0.14 A		Agilent 3458A Multimeter Keysight 34330A Current Shunt	
	(100 to 1 000) A (10 to 100) Hz (100 to 500) Hz	1.5 % of reading + 1A 1.9 % of reading + 1A		Fluke 376 Clamp Meter	
	(100 to 2 500) A (10 to 500) Hz	2.3 % of reading + 5A		Fluke 376 Clamp Meter W/i2500 flex probe	
	Resistance – Generate <sup>1</sup> Fixed Points	(0.001, 0.01, 0.1) $\Omega$		100 $\mu\Omega/\Omega$	Standard Resistors



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Resistance – Generate <sup>1</sup> Fixed Points	(1, 1.9) Ω (10, 19) Ω (100, 190) Ω (1, 1.9) kΩ (10, 19) kΩ 100 kΩ 190 kΩ 1 MΩ 1.9 MΩ 10 MΩ 19 MΩ 100 MΩ	84 μΩ/Ω + 0.1 μΩ 21 μΩ/Ω + 1 μΩ 9.1 μΩ/Ω + 6 μΩ 6.1 μΩ/Ω + 60 μΩ 6.1 μΩ/Ω + 0.6 mΩ 7.6 μΩ/Ω + 6 mΩ 9.1 μΩ/Ω + 6 mΩ 11 μΩ/Ω + 60 mΩ 16 μΩ/Ω + 60 mΩ 35 μΩ/Ω + 0.6 Ω 42 μΩ/Ω + 0.6 Ω 91 μΩ/Ω + 6 Ω	Fluke 5730A/03 Multifunction Calibrator
Resistance – Generate <sup>1</sup>	Up to 11 Ω (11 to 33) Ω (33 to 110) Ω (110 to 330) Ω (0.33 to 1.1) kΩ (1.1 to 3.3) kΩ (3.3 to 11) kΩ (11 to 33) kΩ (33 to 110) kΩ (110 to 330) kΩ (0.33 to 1.1) MΩ (1.1 to 3.3) MΩ (3.3 to 11) MΩ (11 to 33) MΩ (33 to 110) MΩ (110 to 330) MΩ (0.33 to 1.1) GΩ	30 μΩ/Ω + 0.001 Ω 23 μΩ/Ω + 0.001 Ω 21 μΩ/Ω + 0.001 Ω 21 μΩ/Ω + 0.002 Ω 21 μΩ/Ω + 0.002 Ω 21 μΩ/Ω + 0.02 Ω 21 μΩ/Ω + 0.02 Ω 21 μΩ/Ω + 0.2 Ω 21 μΩ/Ω + 0.2 Ω 24 μΩ/Ω + 2 Ω 24 μΩ/Ω + 2 Ω 46 μΩ/Ω + 23 Ω 99 μΩ/Ω + 38 Ω 190 μΩ/Ω + 1.9 kΩ 380 μΩ/Ω + 2.3 kΩ 0.23 % of setting + 76 kΩ 1.1 % of reading + 380 kΩ	Fluke 5522A Multifunction Calibrator
Resistance – Generate <sup>1</sup> Fixed Points	100V 100 kΩ	1 % of reading	Standard Resistors
Resistance – Generate <sup>1</sup> Fixed Points	(100 to 1 000) V 1 MΩ 10 MΩ 100 MΩ 1 GΩ 10 GΩ 100 GΩ	1 % of reading 1 % of reading 1 % of reading 1 % of reading 1.2 % of reading 3.2 % of reading	Standard Resistors



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Resistance – Measure <sup>1</sup>	Up to 12 Ω (10 to 120) Ω (0.1 to 1.2) kΩ (1 to 12) kΩ (10 to 120) kΩ (0.1 to 1.2) MΩ (1 to 12) MΩ (10 to 120) MΩ (0.1 to 1.2) GΩ	15 μΩ/Ω + 56 μΩ 12 μΩ/Ω + 0.5 mΩ 10 μΩ/Ω + 0.6 mΩ 10 μΩ/Ω + 5.6 mΩ 10 μΩ/Ω + 56 mΩ 15 μΩ/Ω + 2.2 Ω 50 μΩ/Ω + 120 Ω 500 μΩ/Ω + 1.2 kΩ 0.5 % of reading + 70 kΩ	Agilent 3458A Multimeter
Capacitance – Generate <sup>1</sup>	(220 to 400) pF (0.4 to 3.299 9) nF (3.3 to 10.999 9) nF (11 to 32.999 9) nF (33 to 109.999) nF (110 to 329.999) nF (0.33 to 1.099 99) μF (1.1 to 3.299 99) μF (3.3 to 10.999 9) μF (11 to 32.999 9) μF (33 to 109.999) μF (110 to 329.999) μF (0.33 to 1.099 99) mF (1.1 to 3.299 99) mF (3.3 to 10.999 9) mF (11 to 32.999 9) mF (33 to 110) mF	0.38% of output + 7.6 pF 0.38 % of output + 0.01 nF 0.19 % of output + 0.01 nF 0.19 % of output + 0.08 nF 0.19 % of output + 0.08 nF 0.19 % of output + 0.23 nF 0.19 % of output + 0.76 nF 0.19 % of output + 2.3 nF 0.19 % of output + 7.6 nF 0.3 % of output + 23 nF 0.34 % of output + 76 nF 0.34 % of output + 228 nF 0.34 % of output + 0.76 μF 0.34 % of output + 2.3 μF 0.34 % of output + 7.6 μF 0.57 % of output + 23 μF 0.84 % of output + 76 μF	Fluke 5522A Multiproduct Calibrator
Capacitance – Measure <sup>1</sup> 1V @ 1 kHz	(400 to 6 400) pF (6.4 to 100) nF (100 to 1 000) nF (1 000 to 1 600) nF (1.6 to 10) μF (10 to 25) μF	0.1 % of reading + 0.4 pF 0.1 % of reading + 0.003 nF 0.1 % of reading + 0.012 nF 0.1 % of reading + 0.11 nF 0.1 % of reading + 0.001 μF 0.1 % of reading + 0.01 μF	GW Instek LCR-816 LCR Meter

**Electrical – DC/Low Frequency**

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
<p>Oscilloscopes Calibration<sup>1</sup> – Generate</p> <p>Voltage DC - 50Ω</p> <p>DC - 1MΩ</p> <p>Square Wave 10 Hz to 10 kHz - 50Ω</p> <p>Square Wave 10 Hz to 1 kHz – 1 MΩ</p> <p>Square Wave (1 to 10) kHz – 1 MΩ</p>	<p>(1 to 24.999) mV (25 to 109.99) mV (110mV to 2.1999) V (2.2 to 6.6) V</p> <p>(1 to 24.999) mV (25 to 109.99) mV (110mV to 2.1999) V (2.2 to 10.999) V (11 to 130) V</p> <p>(1 to 24.999) mVpp (25 to 109.99) mVpp (110mV to 2.1999) Vpp (2.2 to 6.6) Vpp</p> <p>(1 to 24.999) mV (25 to 109.99) mV (110mV to 2.1999) V (2.2 to 10.999) V (11 to 130) V</p> <p>(1 to 24.999) mV (25 to 109.99) mV (110mV to 2.1999) V (2.2 to 10.999) V (11 to 130) V</p>	<p>0.19 % of Output + 31 μV 0.19 % of Output + 36 μV 0.19 % of Output + 87 μV 0.19 % of Output + 0.6 mV</p> <p>0.019 % of reading + 20 μV 0.019 % of reading + 25 μV 0.019 % of reading + 76 μV 0.019 % of reading + 0.6 mV 0.019 % of reading + 6.0 mV</p> <p>0.19 % of Output + 31 μV 0.19 % of Output + 36 μV 0.19 % of Output + 87 μV 0.19 % of Output + 0.6 mV</p> <p>0.038 % of reading + 4 μV 0.038 % of reading + 9 μV 0.038 % of reading + 60 μV 0.038 % of reading + 0.6 mV 0.038 % of reading + 6.0 mV</p> <p>0.19 % of reading + 31 μV 0.19 % of reading + 36 μV 0.19 % of reading + 87 μV 0.19 % of reading + 0.6 mV 0.19 % of reading + 6.0 mV</p>	<p>Fluke 5820A Oscilloscope Calibrator w/ GHz Option</p>
<p>Leveled Sine Flatness 50 kHz to 10 MHz Reference</p>	<p>3 dB Bandwidth (5 to 50) mVpp 50 kHz to 100 MHz (100 to 300) MHz (300 to 500) MHz (500 to 600) MHz (600 to 1 600) MHz (1 600 to 2 100) MHz</p>	<p>0.34 dB 0.36 dB 0.42 dB 0.46 dB 0.5 dB 0.56 dB</p>	<p>Fluke 5820A Oscilloscope Calibrator w/ GHz Option</p>



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Leveled Sine Flatness 50 kHz to 10 MHz Reference	3 dB Bandwidth 50 mV to 3.5 Vpp 50 kHz to 100 MHz (100 to 300) MHz (300 to 500) MHz (500 to 600) MHz (600 to 1 600) MHz (1 600 to 2 100) MHz (3.5 to 5) Vpp 50 kHz to 100 MHz (100 to 300) MHz (300 to 500) MHz (500 to 600) MHz	0.24 dB 0.24 dB 0.32 dB 0.34 dB 0.4 dB 0.44 dB 0.24 dB 0.24 dB 0.32 dB 0.34 dB	Fluke 5820A Oscilloscope Calibrator w/ GHz Option
Oscilloscope Calibration Generate <sup>1</sup> Leveled Sine Flatness	3 dB Bandwidth 50 mV to 3.5 Vpp (2 100 to 4 000) MHz (4 000 to 8 000) MHz (8 000 to 18 000) MHz	0.25 dB 0.35 dB 0.46 dB	Agilent EPM Series Power Meter w/E9304A H18 Power Sensor
Oscilloscope Calibration Generate <sup>1</sup> Time Marker	500 ps to 20 ms 50 ms to 5 s	0.25 μs/s 1.9 μs/s + 3.8 μHz	Fluke 5820A Oscilloscope Calibrator w/ GHz Option
Oscilloscope Calibration Measure <sup>1</sup> Resistance Leakage	(40 to 60) Ω 500 kΩ to 1.5 MΩ (0 to 5.99) V	0.08 % of reading 0.08 % of reading 0.038 % of reading + 0.8 mV	Fluke 5820A Oscilloscope Calibrator w/ GHz Option
Power Meter Range Calibration <sup>1</sup>	3 μW to 100 mW	0.25 % of reading	HP 11683A Power Meter Calibrator
Electrical Calibration of Thermocouple Indicating Devices <sup>1</sup>	Type B (600 to 800) °C (800 to 1 550) °C (1 550 to 1 820) °C Type C (0 to 1 000) °C (1 000 to 1 800) °C (1 800 to 2 000) °C (2 000 to 2 316) °C	0.27 °C 0.22 °C 0.17 °C 0.13 °C 0.18 °C 0.2 °C 0.27 °C	Fluke 7526A Process Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Calibration of Thermocouple Indicating Devices <sup>1</sup>	Type E		Fluke 7526A Process Calibrator
	(-250 to -200) °C	0.19 °C	
	(-200 to -11) °C	0.1 °C	
	(-100 to 0) °C	0.07 °C	
	(0 to 600) °C	0.07 °C	
	(600 to 1 000) °C	0.08 °C	
	Type J		
	(-210 to -100) °C	0.11 °C	
	(-100 to 800) °C	0.07 °C	
	(800 to 1 200) °C	0.08 °C	
	Type K		
	(-250 to -200) °C	0.35 °C	
	-200 to -100) °C	0.13 °C	
	(-100 to 800) °C	0.08 °C	
	(800 to 1 372) °C	0.1 °C	
	Type L		
	(-200 to -100) °C	0.08 °C	
	(-100 to 900) °C	0.07 °C	
	Type N		
	(-250 to -200) °C	0.56 °C	
	(-200 to -100) °C	0.18 °C	
	(-100 to 0) °C	0.1 °C	
	(0 to 100) °C	0.09 °C	
	(100 to 800) °C	0.08 °C	
	(800 to 1 300) °C	0.1 °C	
	Type R		
	(-50 to -25) °C	0.42 °C	
	(-25 to 0) °C	0.34 °C	
(0 to 100) °C	0.3 °C		
(100 to 400) °C	0.22 °C		
(400 to 600) °C	0.17 °C		
(600 to 1 000) °C	0.16 °C		
(1 000 to 1 600) °C	0.15 °C		
(1 600 to 1 767) °C	0.18 °C		
Type S			
(-50 to -25) °C	0.39 °C		
(-25 to 0) °C	0.33 °C		
(0 to 100) °C	0.29 °C		
(100 to 400) °C	0.22 °C		
(400 to 600) °C	0.18 °C		
(600 to 1 600) °C	0.17 °C		
(1 600 to 1 767) °C	0.2 °C		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Calibration of Thermocouple Indicating Devices <sup>1</sup>	Type T (-250 to -200) °C	0.26 °C	Fluke 7526A Process Calibrator
	(-200 to -100) °C	0.13 °C	
	(-100 to 0) °C	0.09 °C	
	(0 to 400) °C	0.07 °C	
	Type U (-200 to 0) °C	0.13 °C	
	(0 to 600) °C	0.08 °C	
Electrical Calibration of RTD Indicating Devices <sup>1</sup>	Pt 385, 100 Ω (-200 to 800) °C	0.05 °C	Fluke 7526A Process Calibrator
	Pt 3926, 100 Ω (-200 to 630) °C	0.05 °C	
	Pt 3916, 100 Ω (-200 to 630) °C	0.05 °C	
	Pt 385, 200 Ω (-200 to 400) °C	0.4 °C	
	(400 to 630) °C	0.5 °C	
	Pt 385, 500 Ω (-200 to 630) °C	0.17 °C	
	Pt 385, 1 000 Ω (-200 to 630) °C	0.09 °C	

Electrical – RF/Microwave

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Measure <sup>1</sup>	(50 to 26 500) MHz	0.25 dB	Agilent EPM Series Power Meter w/E4413A Power Sensor
	(20 to 10) dB	0.22 dB	
	(10 to -50) dB	0.37 dB	
	(-50 to -60) dB	0.94 dB	
	(-60 to -65) dB		
RF Flatness – Measure <sup>1</sup>	9 kHz to 2 000 MHz	0.1 dB	Agilent EPM Series Power Meter w/E9304A H18 Power Sensor
	(20 to -10) dBm	0.1 dB	
	(-10 to -30) dBm	0.11 dB	
	(-30 to -40) dBm	0.12 dB	
	(-40 to -42) dBm		



ANSI National Accreditation Board

Electrical – RF/Microwave

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Flatness – Measure <sup>1</sup>	(2 to 14) GHz (20 to -10) dBm (-10 to -30) dBm (-30 to -40) dBm (-40 to -42) dBm	0.1 dB 0.09 dB 0.1 dB 0.11 dB	Agilent EPM Series Power Meter w/E9304A H18 Power Sensor
	(14 to 18) GHz (20 to -10) dBm (-10 to -30) dBm (-30 to -40) dBm (-40 to -42) dBm	0.11 dB 0.12 dB 0.12 dB 0.13 dB	
RF Power – Measure <sup>1</sup>	9 kHz to 14 000 MHz (20 to 0) dB (0 to -40) dB (-40 to -50) dB (-50 to -55) dB	0.13 dB 0.15 dB 0.35 dB 0.93 dB	Agilent EPM Series Power Meter w/E9304A H18 Power Sensor
	(14 000 to 18 000) MHz (20 to 0) dB (0 to -40) dB (-40 to -50) dB (-50 to -55) dB	0.12 dB 0.16 dB 0.35 dB 0.93 dB	
RF Power – Measure <sup>1</sup>	(10 to 100) MHz (20 to 10) dB (10 to 0) dB (0 to -10) dB (-10 to -20) dB (-20 to -25) dB	0.07 dB 0.06 dB 0.06 dB 0.06 dB 0.11 dB	Agilent EPM Series Power Meter w/N8485A Power Sensor
RF Power – Measure <sup>1</sup>	(100 to 2 000) MHz (20 to 10) dB (10 to 0) dB (0 to -10) dB (-10 to -20) dB (-20 to -25) dB	0.07 dB 0.07 dB 0.06 dB 0.07 dB 0.11 dB	Agilent EPM Series Power Meter w/N8485A Power Sensor
	(2 000 to 12 400) MHz (20 to 10) dB (10 to 0) dB (0 to -10) dB (-10 to -20) dB (-20 to -25) dB	0.08 dB 0.08 dB 0.08 dB 0.08 dB 0.12 dB	



ANSI National Accreditation Board

Electrical – RF/Microwave

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Measure <sup>1</sup>	(12 400 to 18 000) MHz		Agilent EPM Series Power Meter w/N8485A Power Sensor
	(20 to 10) dB	0.09 dB	
	(10 to 0) dB	0.08 dB	
	(0 to -10) dB	0.08 dB	
	(-10 to -20) dB	0.09 dB	
	(-20 to -25) dB	0.12 dB	
	(18 000 to 26 500) MHz		
	(20 to 10) dB	0.12 dB	
	(10 to 0) dB	0.12 dB	
	(0 to -10) dB	0.12 dB	
	(-10 to -20) dB	0.12 dB	
	(-20 to -25) dB	0.15 dB	
RF Power – Measure <sup>1</sup>	(50 to 100) MHz		Agilent EPM Series Power Meter w/N8487A Power Sensor
	(20 to 10) dB	0.07 dB	
	(10 to 0) dB	0.07 dB	
	(0 to -10) dB	0.06 dB	
	(-10 to -20) dB	0.07 dB	
	(-20 to -25) dB	0.11 dB	
	(100 to 6 000) MHz		
	(20 to 10) dB	0.07 dB	
	(10 to 0) dB	0.07 dB	
	(0 to -10) dB	0.07 dB	
	(-10 to -20) dB	0.07 dB	
	(-20 to -25) dB	0.11 dB	
	(6 000 to 12 400) MHz		
	(20 to 10) dB	0.08 dB	
	(10 to 0) dB	0.08 dB	
	(0 to -10) dB	0.07 dB	
	(-10 to -20) dB	0.08 dB	
	(-20 to -25) dB	0.12 dB	
	(12 400 to 18 000) MHz		
	(20 to 10) dB	0.09 dB	
	(10 to 0) dB	0.08 dB	
	(0 to -10) dB	0.08 dB	
	(-10 to -20) dB	0.08 dB	
	(-20 to -25) dB	0.12 dB	



ANSI National Accreditation Board

Electrical – RF/Microwave

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Measure <sup>1</sup>	(18 000 to 26 500) MHz		Agilent EPM Series Power Meter w/N8487A Power Sensor
	(20 to 10) dB	0.11 dB	
	(10 to 0) dB	0.10 dB	
	(0 to -10) dB	0.10 dB	
	(-10 to -20) dB	0.10 dB	
	(-20 to -25) dB	0.13 dB	
	(26 5000 to 33 000) MHz		
	(20 to 10) dB	0.12 dB	
	(10 to 0) dB	0.12 dB	
	(0 to -10) dB	0.11 dB	
	(-10 to -20) dB	0.12 dB	
	(-20 to -25) dB	0.14 dB	
	(33 000 to 40 000) MHz		
	(20 to 10) dB	0.13 dB	
	(10 to 0) dB	0.12 dB	
	(0 to -10) dB	0.12 dB	
	(-10 to -20) dB	0.13 dB	
	(-20 to -25) dB	0.15 dB	
(40 000 to 50 000) MHz			
(20 to 10) dB	0.18 dB		
(10 to 0) dB	0.18 dB		
(0 to -10) dB	0.18 dB		
(-10 to -20) dB	0.18 dB		
(-20 to -25) dB	0.2 dB		
RF Power – Measure <sup>1</sup>	100 kHz to 30 MHz		Agilent N5531S Measuring Receiver N1912A w/E9304A Power Sensor
	(20 to 0) dB	0.12 dB	
	(0 to -58) dB	0.13 dB	
	(-58 to -78) dB	0.14 dB	
	(-78 to -110) dB	0.18 dB	
	(-110 to -115) dB	0.2 dB	
	(-115 to -120) dB	0.28 dB	
	(-120 to -125) dB	0.43 dB	
RF Power – Measure <sup>1</sup>	(30 to 2 000) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 526 Power Sensor
	(30 to 20) dB	0.36 dB	
	(20 to 0) dB	0.2 dB	
	(0 to -58) dB	0.22 dB	
	(-58 to -78) dB	0.23 dB	
	(-78 to -110) dB	0.25 dB	
	(-110 to -115) dB	0.27 dB	
	(-115 to -120) dB	0.33 dB	
(-120 to -125) dB	0.46 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Measure <sup>1</sup>	(2 000 to 3 050) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 526 Power Sensor
	(30 to 20) dB	0.42 dB	
	(20 to 0) dB	0.3 dB	
	(0 to -58) dB	0.31 dB	
	(-58 to -78) dB	0.32 dB	
	(-78 to -110) dB	0.34 dB	
	(-110 to -115) dB	0.35 dB	
	(-115 to -120) dB	0.4 dB	
	(-120 to -125) dB	0.51 dB	
	(3 050 to 6 600) MHz		
	(30 to 20) dB	0.42 dB	
	(20 to 0) dB	0.3 dB	
	(0 to -58) dB	0.31 dB	
	(-58 to -78) dB	0.32 dB	
	(-78 to -110) dB	0.34 dB	
	(-110 to -115) dB	0.38 dB	
	(-115 to -120) dB	0.48 dB	
	(-120 to -125) dB	0.64 dB	
	(6 600 to 13 200) MHz		
	(30 to 20) dB	0.42 dB	
	(20 to 0) dB	0.3 dB	
	(0 to -58) dB	0.31 dB	
	(-58 to -78) dB	0.32 dB	
	(-78 to -100) dB	0.34 dB	
	(-100 to -105) dB	0.37 dB	
	(-105 to -110) dB	0.45 dB	
	(-110 to -115) dB	0.6 dB	
	(-115 to -120) dB	0.82 dB	
	(13 200 to 18 000) MHz		
	(30 to 20) dB	0.42 dB	
(20 to 0) dB	0.3 dB		
(0 to -58) dB	0.31 dB		
(-58 to -78) dB	0.32 dB		
(-78 to -90) dB	0.33 dB		
(-90 to -95) dB	0.35 dB		
(-95 to -100) dB	0.41 dB		
(-100 to -105) dB	0.53 dB		
(-105 to -110) dB	0.72 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Measure <sup>1</sup>	(18 000 to 19 200) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 526 Power Sensor
	(30 to 20) dB	0.48 dB	
	(20 to 0) dB	0.38 dB	
	(0 to -58) dB	0.39 dB	
	(-58 to -78) dB	0.4 dB	
	(-78 to -90) dB	0.41 dB	
	(-90 to -95) dB	0.42 dB	
	(-95 to -100) dB	0.47 dB	
	(-100 to -105) dB	0.58 dB	
	(-105 to -110) dB	0.75 dB	
	(19 200 to 26 500) MHz		
	(30 to 20) dB	0.48 dB	
	(20 to 0) dB	0.38 dB	
	(0 to -58) dB	0.39 dB	
	(-58 to -78) dB	0.4 dB	
	(-78 to -90) dB	0.43 dB	
	(-90 to -95) dB	0.5 dB	
	(-95 to -100) dB	0.63 dB	
(-100 to -105) dB	0.84 dB		
(-105 to -110) dB	1.1 dB		
(30 to 3 050) MHz	(30 to 3 050) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 526 Power Sensor
(0 to 10) dB	(0 to 10) dB	0.02 dB	
(10 to 20) dB	(10 to 20) dB	0.025 dB	
(20 to 30) dB	(20 to 30) dB	0.03 dB	
(30 to 40) dB	(30 to 40) dB	0.035 dB	
(40 to 50) dB	(40 to 50) dB	0.04 dB	
(50 to 60) dB	(50 to 60) dB	0.076 dB	
(60 to 70) dB	(60 to 70) dB	0.081 dB	
(70 to 80) dB	(70 to 80) dB	0.12 dB	
(80 to 90) dB	(80 to 90) dB	0.12 dB	
(90 to 100) dB	(90 to 100) dB	0.13 dB	
(100 to 110) dB	(100 to 110) dB	0.13 dB	
(110 to 120) dB	(110 to 120) dB	0.26 dB	



ANSI National Accreditation Board

Electrical – RF/Microwave

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Attenuation – Measure <sup>1</sup>	(3 050 to 6 600) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 526 Power Sensor
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.12 dB	
	(90 to 100) dB	0.13 dB	
	(100 to 110) dB	0.15 dB	
	(110 to 120) dB	0.37 dB	
	(6 600 to 13 200) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.12 dB	
	(90 to 100) dB	0.14 dB	
	(100 to 110) dB	0.34 dB	
	(110 to 120) dB	0.77 dB	
	(13 200 to 19 200) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
(30 to 40) dB	0.035 dB		
(40 to 50) dB	0.04 dB		
(50 to 60) dB	0.076 dB		
(60 to 70) dB	0.081 dB		
(70 to 80) dB	0.12 dB		
(80 to 90) dB	0.12 dB		
(90 to 100) dB	0.27 dB		
(100 to 110) dB	0.66 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Attenuation – Measure <sup>1</sup>	(19 200 to 26 500) MHz (0 to 10) dB (10 to 20) dB (20 to 30) dB (30 to 40) dB (40 to 50) dB (50 to 60) dB (60 to 70) dB (70 to 80) dB (80 to 90) dB (90 to 100) dB (100 to 110) dB	0.02 dB 0.025 dB 0.03 dB 0.035 dB 0.04 dB 0.076 dB 0.081 dB 0.12 dB 0.2 dB 0.52 dB 1.1 dB	Agilent N5531S Measuring Receiver w/N5532A Opt 526 Power Sensor
Amplitude Modulation – Measure <sup>1</sup>	100 kHz to 10 MHz Rate 50 Hz to 10 kHz (5 to 99) % Depth 10 MHz to 3 GHz Rate 50 Hz to 100 kHz (5 to 20) % Depth (20 to 99) % Depth (3 to 26.5) GHz Rate 50 Hz to 100 kHz (5 to 20) % Depth (20 to 99) % Depth	0.75 % of reading + 0.3 digits  2.5 % of reading + 0.4 digits 0.5 % of reading + 0.4 digits  4.5 % of reading + 0.4 digits 1.5 % of reading + 0.4 digits	Agilent N5531S Measuring Receiver
Frequency Modulation – Measure <sup>1</sup>  $\beta$ = deviation / rate	250 kHz to 10 MHz Rates 20 Hz to 10 kHz Peak Dev 200 to 40 kHz  10 MHz to 6.6 GHz Rates 50 Hz to 200 kHz Peak Dev 250 to 400 kHz  (6.6 to 13.2) GHz Rates 50 Hz to 200 kHz Peak Dev 250 to 400 kHz  (13.2 to 26.5) GHz Rates 50 Hz to 200 kHz Peak Dev 250 to 400 kHz	$\beta > 0.2 - 1.5$ % of reading + 2 Hz $\beta > 1.2 - 1$ % of reading + 2 Hz  $\beta > 0.20 - 1.5$ % of reading + 2 Hz $\beta > 0.45 - 1$ % of reading + 2 Hz  $\beta > 0.2 - 2.5$ % of reading + 4 Hz $\beta > 8.0 - 1$ % of reading + 4 Hz  $\beta > 0.2 - 3.8$ % of reading + 9 Hz $\beta > 16 - 1$ % of reading + 9 Hz	Agilent N5531S Measuring Receiver



ANSI National Accreditation Board

Electrical – RF/Microwave

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Phase Modulation – Measure <sup>1</sup>	100 kHz to 6.6 GHz Deviations > 0.3 rad Deviations > 0.7 rad (6.6 to 13.2) GHz Deviations > 0.6 rad Deviations > 2.0 rad (13.2 to 26.5) GHz Deviations: > 1.2 rad Deviations > 4.0 rad	3 % of reading 1 % of reading 3 % of reading 1 % of reading 3 % of reading 1 % of reading	Agilent N5531S Measuring Receiver
Phase Noise for Signal Sources <sup>1</sup> Offset Frequency, ≤ 100 kHz	5 MHz < f ≤ 18 GHz	2.3 dB	HP 3048A Phase Noise System with 866XA Source
RF Power – Generate <sup>1</sup>	(30 to 2 000) MHz (20 to 0) dB (0 to -58) dB (-58 to -78) dB (-78 to -110) dB (-110 to -120) dB (2 000 to 3 050) MHz (20 to 0) dB (0 to -58) dB (-58 to -78) dB (-78 to -110) dB (-110 to -120) dB (3 050 to 6 600) MHz (20 to 0) dB (0 to -58) dB (-58 to -78) dB (-78 to -110) dB (-110 to -120) dB (6 600 to 13 200) MHz (20 to 0) dB (0 to -58) dB (-58 to -78) dB (-78 to -110) dB (13 200 to 18 000) MHz (20 to 0) dB (0 to -58) dB (-58 to -78) dB (-78 to -110) dB	0.26 dB 0.28 dB 0.29 dB 0.3 dB 0.37 dB 0.37 dB 0.38 dB 0.38 dB 0.4 dB 0.52 dB 0.37 dB 0.38 dB 0.38 dB 0.4 dB 0.52 dB 0.37 dB 0.38 dB 0.38 dB 0.49 dB 0.37 dB 0.38 dB 0.38 dB 0.75 dB	Agilent N5531S Measuring Receiver w/N5532A Opt 526 Power Sensor, E8257D Signal Generator

Electrical – RF/Microwave

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Generate <sup>1</sup>	(18 000 to 19 200) MHz (20 to 0) dB (0 to -58) dB (-58 to -78) dB (-78 to -110) dB (19 200 to 26 500) MHz (20 to 0) dB (0 to -58) dB (-58 to -78) dB (-78 to -110) dB	0.49 dB 0.50 dB 0.5 dB 0.81 dB 0.49 dB 0.5 dB 0.5 dB 1.2 dB	Agilent N5531S Measuring Receiver w/N5532A Opt 526 Power Sensor, E8257D Signal Generator
AM Distortion Measure <sup>1</sup> Rate 20 Hz to 1 kHz	(0.1 to 10) MHz AM Depth > 1% (0 to -20) dB (-20 to -30) dB AM Depth > 3% (0 to -20) dB (-20 to -30) dB (-30 to -40) dB 10 MHz to 26.5 GHz AM Depth > 1% (0 to -20) dB (-20 to -30) dB AM Depth > 3% (0 to -20) dB (-20 to -30) dB (-30 to -40) dB	1.2 dB 2.2 dB 1 dB 1.3 dB 2.4 dB 1.3 dB 2.5 dB 1.1 dB 1.4 dB 3 dB	Agilent N5531S Measuring Receiver
FM Distortion Measure <sup>1</sup> Rate 20 Hz to 1 kHz	(1 to 6 600) MHz Dev 500 Hz to 2 kHz (0 to -20) dB (-20 to -30) dB (-30 to -40) dB Dev > 2 kHz (0 to -20) dB (-20 to -30) dB (-30 to -40) dB (-40 to -50) dB	0.26 dB 0.79 dB 2.3 dB 0.09 dB 0.27 dB 0.83 dB 2.4 dB	Agilent N5531S Measuring Receiver

**Electrical – RF/Microwave**

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
FM Distortion Measure <sup>1</sup> Rate 20 Hz to 1 kHz	(6.6 to 13.2) GHz Dev > 2.3 kHz (0 to -20) dB (-20 to -30) dB (-30 to -40) dB Dev > 4.5 kHz (0 to -20) dB (-20 to -30) dB (-30 to -40) dB (-40 to -50) dB (13.2 to 26.5) GHz Dev > 2.7 kHz (0 to -20) dB (-20 to -30) dB (-30 to -40) dB Dev > 6.0 kHz (0 to -20) dB (-20 to -30) dB (-30 to -40) dB (-40 to -50) dB	0.26 dB 0.79 dB 2.3 dB 0.09 dB 0.27 dB 0.83 dB 2.4 dB 0.26 dB 0.79 dB 2.3 dB 0.09 dB 0.27 dB 0.83 dB 2.4 dB	Agilent N5531S Measuring Receiver
PM Distortion Measure <sup>1</sup>	(1 to 6 600) MHz Rate (20 to 500) Hz Dev > 0.8 rad (0 to -20) dB (-20 to -30) dB (-30 to -40) dB Dev > 2.5 rad (0 to -20) dB (-20 to -30) dB (-30 to -40) dB (-40 to -50) dB Rate (500 to 1 000) Hz Dev > 0.4 rad (0 to -20) dB (-20 to -30) dB (-30 to -40) dB Dev > 1.0 rad (0 to -20) dB (-20 to -30) dB (-30 to -40) dB (-40 to -50) dB	0.26 dB 0.79 dB 2.3 dB 0.09 dB 0.27 dB 0.83 dB 2.3 dB 0.26 dB 0.79 dB 2.3 dB 0.09 dB 0.27 dB 0.83 dB 2.3 dB	Agilent N5531S Measuring Receiver

Electrical – RF/Microwave

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
PM Distortion Measure <sup>1</sup>	(6.6 to 13.2) GHz		
	Rate (20 to 500) Hz		
	Dev > 1.8 rad		
	(0 to -20) dB	0.26 dB	
	(-20 to -30) dB	0.79 dB	
	(-30 to -40) dB	2.3 dB	
	Dev > 5.5 rad		
	(0 to -20) dB	0.09 dB	
	(-20 to -30) dB	0.27 dB	
	(-30 to -40) dB	0.83 dB	
	(-40 to -50) dB	2.3 dB	
	Rate (500 to 1 000) Hz		
	Dev > 0.8 rad		
	(0 to -20) dB	0.26 dB	
	(-20 to -30) dB	0.79 dB	
	(-30 to -40) dB	2.3 dB	
	Rate (500 to 1 000) Hz		
	Dev > 2.5 rad		
	(0 to -20) dB	0.09 dB	
	(-20 to -30) dB	0.27 dB	
(-30 to -40) dB	0.83 dB		
(-40 to -50) dB	2.3 dB		
(13.2 to 26.5) GHz			
Rate 20 to 500 Hz			
Dev > 3.5 rad			
(0 to -20) dB	0.26 dB		
(-20 to -30) dB	0.79 dB		
(-30 to -40) dB	2.3 dB		
Dev > 10.0 rad			
(0 to -20) dB	0.09 dB		
(-20 to -30) dB	0.27 dB		
(-30 to -40) dB	0.83 dB		
(-40 to -50) dB	2.3 dB		
Agilent N5531S Measuring Receiver			



ANSI National Accreditation Board

Electrical – RF/Microwave

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
PM Distortion Measure <sup>1</sup>	(13.2 to 26.5) GHz		Agilent N5531S Measuring Receiver
	Rate 500 Hz to 1000 Hz		
	Dev > 3.0 rad	0.26 dB	
	(0 to -20) dB	0.79 dB	
	(-20 to -30) dB	2.3 dB	
	(-30 to -40) dB		
	Dev > 8.0 rad	0.09 dB	
	(0 to -20) dB	0.27 dB	
Total Harmonic Distortion (THD)	(-30 to -40) dB	0.83 dB	HP 8903B Audio Analyzer
	(-40 to -50) dB	2.3 dB	
	(0 to -60) dB		
	20 Hz to 20 kHz		
	(0 to -40) dB	1 dB	
	(-40 to -50) dB	1 dB	
	(-50 to -60) dB	1.3 dB	
	(-60 to -65) dB	1.7 dB	
	(20 to 50) kHz		
	(0 to -40) dB	2 dB	
	(-40 to -50) dB	2.1 dB	
	(-50 to -60) dB	3 dB	
Harmonics Measure <sup>1</sup>	(50 to 100) kHz		Agilent E4440A Measuring Receiver
	(0 to -40) dB	2 dB	
	(-40 to -50) dB	2.4 dB	
	(-80 to -10) dB		
	2 <sup>nd</sup> through 5 <sup>th</sup> Harmonic		
	1kHz to 600MHz	0.37 dB	
	(600 to 1 320) MHz	1.1 dB	
	(1 320 to 2 200) MHz	1.4 dB	
	(2 200 to 3 000) MHz	1.4 dB	
	(3 000 to 4 400) MHz	1.7 dB	
(4 400 to 5 300) MHz	1.9 dB		
2 <sup>nd</sup> through 4 <sup>th</sup> Harmonic			
(5 300 to 6 625) MHz	2.1 dB		
2 <sup>nd</sup> through 3 <sup>rd</sup> Harmonic			
(6 625 to 8 833) MHz	2.1 dB		
2 <sup>nd</sup> Harmonic			
(8 833 to 13 250) MHz	2.1 dB		

**Length – Dimensional Metrology**

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Gage Blocks <sup>2</sup>	Up to 13 in	$(3.8 + 2.4L) \mu\text{in}$	Master gage blocks, P&W Labmaster Universal
Micrometer Standards Length Rods	Up to 13 in	$(0.75 + 6.9L) \mu\text{in}$	Gage blocks (grade 2), P&W Labmaster Universal
Micrometers <sup>1,2</sup>	Up to 40 in	$(31 + 4.1L) \mu\text{in}$	Gage blocks (grade 2)
Calipers <sup>1,2</sup>	Up to 40 in	$(280 + 1.8L) \mu\text{in}$	
Dial Indicators <sup>1,2</sup> Resolution $\geq 50\mu\text{in}$ Resolution $< 50\mu\text{in}$	Up to 10 in Up to 0.1 in	$(27 + 3.3L) \mu\text{in}$ $7 \mu\text{in}$	Gage blocks (grade 2)
Height Gages <sup>1,2</sup>	Up to 40 in	$(98 + 2.6L) \mu\text{in}$	
Rulers <sup>1</sup>	Up to 46 in	0.009 1 in	
Feeler Gage <sup>1</sup>	Up to 1 in	45 $\mu\text{in}$	Pratt & Whitney Supermicrometer C
Radius Gages	(0.01 to 1) in	590 $\mu\text{in}$	Optical comparator
Bore Micrometers/Gages <sup>2</sup> 2 point 3 point	Up to 8 in	$(7 + 2L) \mu\text{in}$ $(42 + 1.4L) \mu\text{in}$	Master gage blocks, P&W universal measuring machine, Master Ring
Protractors <sup>1</sup>	(0 to 360) °	0.019 °	Angle blocks, Sine Bar, Gage Blocks Grade 0
Parallelism & Straightness	(0 to 4) in	110 $\mu\text{in}$	Gage Amplifier, Surface Plate
Cylindrical Gages <sup>2</sup> Plain Pins, Plugs Rings	(0 to 13) in (0.04 to 14) in	$(2 + 3D) \mu\text{in}$ $(6.5 + 3.2D) \mu\text{in}$	Master gage blocks, P&W universal measuring machine
Thread Plugs <sup>1</sup> Major Diameter Pitch Diameter	Up to 12 in Up to 12 in	52 $\mu\text{in}$ 110 $\mu\text{in}$	B & S 599-246-00, Van Keuren thread wire set, Gage blocks, P & W Model C
Thread Rings <sup>3</sup> Pitch Diameter (tactile fit)	Up to 12 in	$(350 + 47D) \mu\text{in}$	Thread setting plug gages
Surface Plates <sup>1</sup> – Overall Flatness Local Area Flatness	Up to 6 ft × 6 ft (-0.001 to 0.001) in	125 $\mu\text{in}$ 68 $\mu\text{in}$	Planekator Repeat-o-meter

**Length – Dimensional Metrology**

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Optical Comparators <sup>1</sup> Linearity	Up to 20 in (20 to 40) in	190 μin 350 μin	Gage blocks, SI Industries glass scales, angle blocks
Magnification Angle	(10 to 100) X (0 to 360)°	250 μin 0.02°	
Tape Measures <sup>2</sup>	Up to 100 ft	(0.000 26F + 0.026) in	Standard rule
Coating Thickness Gages <sup>1,2</sup> Eddy Current & Magnetic Induction	(0.9 to 20) mils	75 μin	Coating thickness Standards
Coating Thickness Shims <sup>2</sup>	(0 to 243) mils	61 μin	Pratt & Whitney Supermicrometer C
Ultrasonic Thickness Gauges <sup>1</sup>	Up to 10 in	590 μin	Gage blocks

**Mass and Mass Related**

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Scales & Balances <sup>1,2,4</sup>	(1 to 400) g	(0.001 5X + 0.019) mg	Class 0 Weights
Scales & Balances <sup>1,2,4</sup>	(1 to 1 000) g (1 000 to 40 000) g	(0.0031X + 0.04) mg (0.003 3X) mg	Class 1 Weights
Scales & Balances <sup>1,2,4</sup>	(1 to 9 100) g	(0.12X + 9.3) mg	Class F Weights
Scales & Balances <sup>1,2,4</sup>	(0.002 to 1 000) lb	(0.000 12W + 0.000 04) lb	Class F Weights
Force <sup>1</sup> Tension and Compression	(0.5 to 500) lbf	0.09 % of reading	Class F weights
Pressure - Generate	(10 to 500) psi (200 to 10 000) psi	0.008% of reading 0.01% of reading	Fluke P3224-PSI Pressure Calibrator



ANSI National Accreditation Board

Mass and Mass Related

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Pressure <sup>1</sup>	(-1 to 1) inH <sub>2</sub> O (-50 to 50) inH <sub>2</sub> O (-30 to 0) inHg (-15 to 30) psi (30 to 100) psi (100 to 500) psi (500 to 1 000) psi (1 000 to 5 000) psi (5 000 to 15 000) psi	0.001 2 inH <sub>2</sub> O 0.058 inH <sub>2</sub> O 0.015 inHg 0.007 2 psi 0.033 psi 0.14 psi 0.3 psi 1.5 psi 8.9 psi	Digital Pressure Gages Additel ADT681-05-DP1 Additel ADT681-05-DP50 Additel ADT681-02-CP30 Additel ADT681-02-CP30 Additel ADT681-02-GP100 Additel ADT681-02-GP500 Additel ADT681-02-GP1K Additel ADT681-02-GP5K Additel ADT681-05-GP15K
Torque Tools <sup>1</sup>	(1 to 10) ozf·in (10 to 100) ozf·in 4 lbf·in to 1 000 lbf·ft	0.63 % of reading 0.6 % of reading 0.32 % of reading	AWS QCMIO-10 CDI 1001-O-DDT CDI 5000 ST torque tester
Torque Analyzers	(1 to 20) ozf·in (20 to 100) ozf·in (4 to 150) lbf·in (12.5 to 250) lbf·ft	0.25 % of reading 0.12 % of reading 0.04 % of reading 0.16 % of reading	Torque wheels & weights
Rockwell Hardness Testers <sup>1</sup>	(0 to 59) HRBW (60 to 79) HRBW (80 to 100) HRBW (20 to 39) HRC (40 to 59) HRC (60 to 70) HRC (20 to 65) HRA (70 to 78) HRA (80 to 84) HRA	1.9 HRBW 1.4 HRBW 1.3 HRBW 1.3 HRC 1.3 HRC 0.68 HRC 1.3 HRA 1.3 HRA 0.68 HRA	Indirect verification per ASTM E18

Thermodynamic

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Relative Humidity Generate	(10 to 95) %RH	0.5 %RH	Thunder Scientific 2500 Humidity Chamber
Relative Humidity – Measure	(0 to 90) %RH	1.3 %RH	Vaisala MI70 / HMP75 Thermohygrometer
Temperature – Measuring Equipment <sup>1</sup>	(-25 to 375) °C	0.19 °C	Hart 1502A with 5615 PRT and dry block



ANSI National Accreditation Board

**Thermodynamic**

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Temperature – Measure <sup>1</sup>	(-196 to 232) °C (-321 to 450) °F (232 to 420) °C (450 to 788) °F	0.026 °C 0.047 °F 0.033 °C 0.059 °F	Hart 1502A Indicator with 5615 PRT
Infrared (IR) Thermometry <sup>1</sup>	(20 to 100) °C (100 to 300) °C (300 to 420) °C (420 to 500) °C	1.5 °C 4.3 °C 6 °C 7.7 °C	Fluke 9132 Blackbody $\epsilon = 0.95, \lambda = (8 \text{ to } 14) \mu\text{m}$

**Time and Frequency**

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency – Generate <sup>1</sup>	10 MHz	$1 \times 10^{-12}$ Hz/Hz	HP 58503A GPS Receiver
Frequency – Generate <sup>1</sup>	(1 to 10) Hz (10 to 100) Hz (100 to 1 000) Hz (1 to 10) kHz (10 to 100) kHz (0.1 to 1) MHz (1 to 10) MHz	$1 \times 10^{-12}$ Hz/Hz + 0.57 $\mu$ Hz $1 \times 10^{-12}$ Hz/Hz + 5.7 $\mu$ Hz $1 \times 10^{-12}$ Hz/Hz + 57 $\mu$ Hz $1 \times 10^{-12}$ Hz/Hz + 0.57 mHz $1 \times 10^{-12}$ Hz/Hz + 5.7 mHz $1 \times 10^{-12}$ Hz/Hz + 57 mHz $1 \times 10^{-12}$ Hz/Hz + 0.57 Hz	Agilent 33250A Function Generator / HP 58503A GPS Receiver
Frequency – Generate <sup>1</sup>	(10 to 40 000) MHz	$1 \times 10^{-12}$ Hz/Hz + 0.57mHz	Agilent E8257D Opt 540 Signal Generator / HP 58503A GPS Receiver
Time – Generate	1 pps	$1 \times 10^{-12}$ s/s + 750 ps	HP 58503A GPS Receiver
Frequency – Measure <sup>1</sup>	(1 to 10) Hz (10 to 100) Hz (100 to 1 000) Hz (1 to 10) kHz (10 to 100) kHz (100 to 200) kHz (0.2 to 3 000) MHz	$1.82 \times 10^{-9}$ Hz/Hz $0.59 \times 10^{-9}$ Hz/Hz $0.20 \times 10^{-9}$ Hz/Hz $74 \times 10^{-10}$ Hz/Hz $35 \times 10^{-10}$ Hz/Hz $23 \times 10^{-10}$ Hz/Hz $20 \times 10^{-10}$ Hz/Hz	Agilent 53132A Opt 030 Frequency Counter / HP 58503A GPS Receiver
Frequency – Measure <sup>1</sup>	(10 to 26 500) MHz	$1 \times 10^{-12}$ Hz/Hz + 0.1 Hz	Agilent E4440A Spectrum Analyzer / HP 58503A GPS Receiver
Type I (digital) Timers	(0 to 19.99) sec/day (0 to 599) sec/month	0.031 sec/day 1.1 sec/month	Helmut Klein Timometer 4500

**Time and Frequency**

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Type II (mechanical) Timers	(0 to 320) sec/day	0.6 sec/day	Helmut Klein Timometer 4500
Tachometers – RPM <sup>1</sup>	Up to 100 000 RPM	0.001 % of reading + 0.6R	HP 33250A Signal Generator & LED

**DIMENSIONAL MEASUREMENT**

**1 Dimensional**

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Length	X Axis (0.01 to 10) in Y Axis (0.01 to 6) in	162 μin 123 μin	Optical comparator

**2 Dimensional**

Melbourne, FL

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Angle	Up to 360°	0.006°	Optical comparator

[Return to Site listing](#) (top)    [Go to Notes](#) (bottom)

**Services performed at satellite laboratory**

**Technical Maintenance, Inc.**

**Toyo Tires**

3660 Highway 411 N.E.

White, GA 30184

Chris Hale (On-site Technical Services Operations Manager) Phone: (318) 805-8600

Scott Chamberlain (Quality Manager) Phone: (321) 242-0890

**CALIBRATION**

**Electrical – DC/Low Frequency**

White, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Voltage – Generate	(0 to 329.999) mV 330 mV to 3.299 99 V (3.3 to 32.999 9) V (33 to 329.999) V (330 to 1 020) V	0.028 % of reading + 10 $\mu$ V 0.013 % of reading + 15 $\mu$ V 0.016 % of reading + 150 $\mu$ V 0.015 % of reading + 1.5 mV 0.015 % of reading + 5.5 mV	Fluke 5080A Multiproduct Calibrator
DC Voltage – Measure	(0 to 100) mV (0.1 to 1) V (1 to 10) V (10 to 100) V (100 to 1 000) V	0.007 6 % of reading + 3.5 $\mu$ V 0.005 8 % of reading + 7 $\mu$ V 0.005 1 % of reading + 50 $\mu$ V 0.006 4 % of reading + 600 $\mu$ V 0.005 6 % of reading + 10 mV	Fluke 8845A Multimeter
DC Current – Generate	(0 to 329.99) $\mu$ A (.330 to 3.299 9) mA (3.3 to 32.999) mA (33 to 329.999) mA 330 mA to 1.099 9 A (1.1 to 2.999 9) A (3 to 10.999) A (11 to 20.5) A	0.1 % of reading + 0.1 $\mu$ A 0.095 % of reading + 0.25 $\mu$ A 0.083 % of reading + 1.25 $\mu$ A 0.1 % of reading + 16.5 $\mu$ A 0.18 % of reading + 220 $\mu$ A 0.19 % of reading + 220 $\mu$ A 0.29 % of reading + 2.5 mA 0.58 % of reading + 3.8 mA	Fluke 5080A Multiproduct Calibrator
	(20.5 to 150) A (150 to 1 050) A	0.44 % of reading + 0.14 A 0.55 % of reading + 0.14 A	Fluke 5080A Multiproduct Calibrator, 5500A/Coil x50
Resistance – Generate	0 $\Omega$ 100 $\Omega$ 1 000 $\Omega$	0.014 $\Omega$ 0.049 % of reading 0.029 % of reading	Fluke 5080A Multiproduct Calibrator
Resistance – Generate	10 k $\Omega$ 100 k $\Omega$ 1 M $\Omega$ 10 M $\Omega$	0.029 % of reading 0.044 % of reading 0.11 % of reading 0.12 % of reading	Fluke 5080A Multiproduct Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

White, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Resistance – Generate	100 MΩ 1 000 MΩ	1 % of reading 1 % of reading	Tsuruga 5804 Resistor
AC Voltage – Generate  45 Hz to 1 kHz	(1 to 32.99) mV (33 to 329.99) mV 330 mV to 3.299 9 V (3.3 to 32.999) V (33 to 101.99) V (102 to 329.99) V (330 to 1 020) V	0.42 % of reading + 60 μV 0.2 % of reading + 60 μV 0.15 % of reading + 180 μV 0.16 % of reading + 1.8 mV 0.19 % of reading + 18 mV 0.19 % of reading + 18 mV 0.2 % of reading + 180 mV	Fluke 5080A Multiproduct Calibrator
AC Voltage – Measure	(0 to 100) mV 10 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (0.1 to 1) V 10 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (1 to 10) V 10 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (10 to 100) V 10 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (100 to 750) V 10 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz	0.12 % of reading + 0.04 mV 0.19 % of reading + 0.05 mV 0.79 % of reading + 0.08 mV 4.8 % of reading + 0.5 mV 0.13 % of reading + 0.3 mV 0.18 % of reading + 0.5 mV 0.74 % of reading + 0.8 mV 4.7 % of reading + 5 mV 0.12 % of reading + 3 mV 0.18 % of reading + 5 mV 0.74 % of reading + 8 mV 4.7 % of reading + 50 mV 0.12 % of reading + 30 mV 0.18 % of reading + 50 mV 0.74 % of reading + 80 mV 4.7 % of reading + 500 mV 0.12 % of reading + 0.3 V 0.18 % of reading + 0.5 V 0.74 % of reading + 0.8 V 4.7 % of reading + 5 V	Fluke 8845A Multimeter



ANSI National Accreditation Board

**Electrical – DC/Low Frequency**

White, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Generate 45 Hz to 1 kHz	(29 to 329.9) $\mu$ A 330 $\mu$ A to 3.299 9 mA (3.3 to 32.999) mA (33 to 329.99) mA 330 mA to 1.099 9 A (1.1 to 2.999 9) A (3 to 10.999) A (11 to 20.5) A	0.31 % of reading + 0.75 $\mu$ A 0.27 % of reading + 0.9 $\mu$ A 0.23 % of reading + 12 $\mu$ A 0.32 % of reading + 120 $\mu$ A 0.28 % of reading + 1.2 mA 0.33 % of reading + 1.5 mA 0.47 % of reading + 6 mA 0.8 % of reading + 15 mA	Fluke 5080A Multiproduct Calibrator
AC Current – Generate 45 Hz to 1 kHz	(20.5 to 1 000) A (45 to 65) Hz (65 to 440) Hz	0.57 % of reading + 1 A 0.86 % of reading + 0.98 A	Fluke 5080A Multiproduct Calibrator, 5500A/Coil x50
Electrical Simulation of Thermocouple Indicators	Type J (-200 to -0) $^{\circ}$ C (0 to 1 200) $^{\circ}$ C	1.2 $^{\circ}$ C 0.83 $^{\circ}$ C	Fluke 725 Process Calibrator
Electrical Simulation of RTD Indicators	Pt100-385 (0 to 200) $^{\circ}$ C Pt100-392 (0 to 200) $^{\circ}$ C	0.39 $^{\circ}$ C 0.36 $^{\circ}$ C	

**Length – Dimensional Metrology**

White, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Calipers	Up to 300 mm	1.6 $\mu$ m	Gage Blocks
Micrometers	Up to 25 mm	0.63 $\mu$ m	Gage Blocks
Dial Indicators	Up to 50 mm	1 $\mu$ m	Gage Blocks
Linear Scales – Rulers	Up to 300 mm	0.29 mm	Gage Blocks
Tape Measures	Up to 300 cm	0.69 mm	Steel Rule

**Mass and Mass Related**

White, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Scales & Balances <sup>2,4</sup>	(0.1 to 250) kg	9 g + 0.6R	F Class Weights

**Mass and Mass Related**

White, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Pressure	Up to 100 psi Up to 689 kPa	0.09 psi 0.62 kPa	Fluke 700P06 Pressure Module
	Up to 300 psi Up to 2 068 kPa	0.22 psi 1.5 kPa	Fluke 700P27 Pressure Module
	Up to 1 000 psi Up to 6 895 kPa	0.72 psi 5 kPa	Fluke 700P08 Pressure Module
	Up to 10 000 psi Up to 68 948 kPa	10 psi 69 kPa	Fluke 700P31 Pressure Module

**Thermodynamic**

White, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RTD – Measure Pt100-385 Pt100-3926	(0 to 200) °C	0.39 °C	Fluke 725 Process Calibrator
	(0 to 200) °C	0.36 °C	
Type J Thermocouples Measure	(50 to 105) °C	1.2 °C	Fluke 725 Process Calibrator, Fluke 9140 Dry Block

**Time and Frequency**

White, GA

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency	(45 to 119.99) Hz (120 to 1 000) Hz	0.017 % of reading + 2 mHz 0.16 % of reading + 2 mHz	Fluke 5080A Multiproduct Calibrator

[Return to Site listing](#) (top)    [Go to Notes](#) (bottom)



ANSI National Accreditation Board

## Services performed at satellite laboratory

### Technical Maintenance, Inc.

1355 Garden of the Gods Road, Suite 100

Colorado Springs, CO 80907

Glenn Curtis (Branch Manager) Phone: (719) 424-7068

Scott Chamberlain (Quality Manager) Phone: (321) 242-0890

## CALIBRATION AND DIMENSIONAL MEASUREMENT

### CALIBRATION

#### Acoustics and Vibration

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Accelerometers – Acceleration	(0.01 to 10) g (5 to < 10) Hz (10 to < 30) Hz (30 to < 2 000) Hz (2 to 10) kHz	4 % of reading 3 % of reading 1.5 % of reading 4 % of reading	Accelerometer Calibrator

#### Chemical Quantities

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
pH meters <sup>1,5</sup>	4 pH 7 pH 10 pH	0.016 pH 0.025 pH 0.061 pH	pH buffer solutions
Conductivity Meters <sup>1,5</sup>	0.65 µS/cm 10 µS/cm 100 µS/cm 1 000 µS/cm 10 000 µS/cm 100 000 µS/cm	0.33 µS/cm 0.64 µS/cm 2.1 µS/cm 5 µS/cm 44 µS/cm 400 µS/cm	Conductivity solutions

**Electrical – DC/Low Frequency**

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Current – Generate <sup>1</sup>	(1 to 2.2) nA (2.2 to 22) nA (22 to 220) nA (0.22 to 2.2) $\mu$ A (2.2 to 10) $\mu$ A	93 $\mu$ A/A + 0.007 nA 92 $\mu$ A/A + 0.007 nA 92 $\mu$ A/A + 0.01 nA 36 $\mu$ A/A + 0.1 nA 15 $\mu$ A/A + 1 nA	Fluke 5730A Multifunction Calibrator, Fluke 5522A Multiproduct Calibrator
DC Current – Generate <sup>1</sup>	(10 to 220) $\mu$ A 220 $\mu$ A to 2.2 mA (2.2 to 22) mA (22 to 100) mA (100 to 220) mA (0.22 to 1) A (1 to 2.2) A	38 $\mu$ A/A + 5 nA 30 $\mu$ A/A + 6 nA 30 $\mu$ A/A + 44 nA 38 $\mu$ A/A + 0.7 $\mu$ A 45 $\mu$ A/A + 0.7 $\mu$ A 68 $\mu$ A/A + 12 $\mu$ A 105 $\mu$ A/A + 12 $\mu$ A	Fluke 5730A Multifunction Calibrator
DC Current – Generate <sup>1</sup>	(2.2 to 11) A	274 $\mu$ A/A + 371 $\mu$ A	Fluke 5730A Multifunction Calibrator 5725A Amplifier
DC Current – Generate <sup>1</sup>	(11 to 20.5) A	761 $\mu$ A/A + 631 $\mu$ A	Fluke 5522A Multiproduct Calibrator
DC Current – Generate 1 Clamp Meters	(0 to 200) A	0.21 % of output + 0.028 A	Fluke 5522A Multiproduct Calibrator / 9100-200 x10 Coil
DC Current – Generate 1 Clamp Meters	(0 to 1 000) A	0.21 % of output + 0.04 A	Fluke 5522A Multiproduct Calibrator / 9100-200 x50 Coil
DC Current – Measure <sup>1</sup>	(1 to 10) nA (10 to 100) nA (0.1 to 1) $\mu$ A (1 to 10) $\mu$ A	35 $\mu$ A/A + 0.12 pA 13 $\mu$ A/A + 1.2 pA 9.1 $\mu$ A/A + 0.01 nA 7.9 $\mu$ A/A + 0.12 nA	Fluke 5730A Multifunction Calibrator Agilent 3458A Option 002 Multimeter
DC Current – Measure <sup>1</sup>	(10 to 100) $\mu$ A (0.1 to 1) mA (1 to 10) mA (10 to 100) mA (0.1 to 1) A	20 $\mu$ A/A + 0.8 nA 20 $\mu$ A/A + 5 nA 20 $\mu$ A/A + 0.05 $\mu$ A 35 $\mu$ A/A + 0.5 $\mu$ A 110 $\mu$ A/A + 10 $\mu$ A	Agilent 3458A Multimeter
DC Current – Measure <sup>1</sup>	(1 to 3) A (3 to 10) A	761 $\mu$ A/A + 462 $\mu$ A 1.1 mA/A + 614 $\mu$ A	Fluke 8846A Multimeter
DC Current – Measure <sup>1</sup>	(1 to 1 000) A	0.25 % of reading	Agilent 3458A Option 002 Multimeter, Empro Current Shunt



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Voltage – Generate <sup>1</sup>	(0 to 220) mV 220 mV to 2.2 V (2.2 to 11) V (11 to 22) V (22 to 220) V (220 to 1 100) V	6.8 $\mu\text{V/V} + 0.8 \mu\text{V}$ 4.6 $\mu\text{V/V} + 0.9 \mu\text{V}$ 3 $\mu\text{V/V} + 2.5 \mu\text{V}$ 3 $\mu\text{V/V} + 3.9 \mu\text{V}$ 4.6 $\mu\text{V/V} + 38 \mu\text{V}$ 6.1 $\mu\text{V/V} + 385 \mu\text{V}$	Fluke 5730A Multifunction Calibrator
DC Voltage – Measure <sup>1</sup>	(0 to 100) mV (0.1 to 1) V (1 to 10) V (10 to 100) V (100 to 1 000) V	5.0 $\mu\text{V/V} + 1.2 \mu\text{V}$ 4.0 $\mu\text{V/V} + 1.2 \mu\text{V}$ 4.0 $\mu\text{V/V} + 2.1 \mu\text{V}$ 6.0 $\mu\text{V/V} + 30 \mu\text{V}$ 6 $\mu\text{V/V} + 100 \mu\text{V} + 12 \mu\text{V/V} \times (\text{Vin}/1\ 000) ^2$	Agilent 3458A Option 002 Multimeter
DC Voltage – Measure <sup>1</sup>	(1 to 30) kV	0.1 % of reading	Ross VD30 High Voltage Divider, Fluke 287 Multimeter
AC Voltage – Generate <sup>1</sup>	(0.22 to 2.2) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz (2.2 to 22) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	228 $\mu\text{V/V} + 3.9 \mu\text{V}$ 88 $\mu\text{V/V} + 3.9 \mu\text{V}$ 76 $\mu\text{V/V} + 3.9 \mu\text{V}$ 190 $\mu\text{V/V} + 3.9 \mu\text{V}$ 457 $\mu\text{V/V} + 4.6 \mu\text{V}$ 989 $\mu\text{V/V} + 9.2 \mu\text{V}$ 1.3 mV/V + 19 $\mu\text{V}$ 2.6 mV/V + 19 $\mu\text{V}$ 228 $\mu\text{V/V} + 3.9 \mu\text{V}$ 88 $\mu\text{V/V} + 3.9 \mu\text{V}$ 76 $\mu\text{V/V} + 3.9 \mu\text{V}$ 190 $\mu\text{V/V} + 3.9 \mu\text{V}$ 457 $\mu\text{V/V} + 4.6 \mu\text{V}$ 989 $\mu\text{V/V} + 9.2 \mu\text{V}$ 1.3 mV/V + 19 $\mu\text{V}$ 2.6 mV/V + 19 $\mu\text{V}$	Fluke 5730A Multifunction Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Generate <sup>1</sup>	(22 to 220) mV		Fluke 5730A Multifunction Calibrator
	(10 to 20) Hz	228 $\mu$ V/V + 11.4 $\mu$ V	
	(20 to 40) Hz	88 $\mu$ V/V + 6.1 $\mu$ V	
	40 Hz to 20 kHz	53 $\mu$ V/V + 6.1 $\mu$ V	
	(20 to 50) kHz	114 $\mu$ V/V + 6.1 $\mu$ V	
	(50 to 100) kHz	304 $\mu$ V/V + 15.2 $\mu$ V	
	(100 to 300) kHz	609 $\mu$ V/V + 19 $\mu$ V	
	(300 to 500) kHz	1.3 mV /V + 23 $\mu$ V	
	500 kHz to 1 MHz	2.5 mV /V + 46 $\mu$ V	
	220 mV to 2.2 V		
	(10 to 20) Hz	228 $\mu$ V/V + 38 $\mu$ V	
	(20 to 40) Hz	84 $\mu$ V/V + 15 $\mu$ V	
	40 Hz to 20 kHz	37 $\mu$ V/V + 8 $\mu$ V	
	(20 to 50) kHz	61 $\mu$ V/V + 9 $\mu$ V	
	(50 to 100) kHz	76 $\mu$ V/V + 30 $\mu$ V	
	(100 to 300) kHz	304 $\mu$ V/V + 76 $\mu$ V	
	(300 to 500) kHz	913 $\mu$ V /V + 190 $\mu$ V	
	500 kHz to 1 MHz	1.5 mV/V + 304 $\mu$ V	
	(2.2 to 22) V		
	(10 to 20) Hz	228 $\mu$ V/V + 380 $\mu$ V	
	(20 to 40) Hz	84 $\mu$ V/V + 152 $\mu$ V	
	40 Hz to 20 kHz	37 $\mu$ V/V + 54 $\mu$ V	
	(20 to 50) kHz	61 $\mu$ V/V + 91 $\mu$ V	
	(50 to 100) kHz	76 $\mu$ V/V + 190 $\mu$ V	
	(100 to 300) kHz	228 $\mu$ V/V + 609 $\mu$ V	
	(300 to 500) kHz	913 $\mu$ V/V + 1.9 mV	
	500 kHz to 1 MHz	1.4 mV/V + 3 mV	
(22 to 220) V			
(10 to 20) Hz	228 $\mu$ V/V + 3.8 mV		
(20 to 40) Hz	84 $\mu$ V/V + 1.5 mV		
40 Hz to 20 kHz	49 $\mu$ V/V + 0.6 mV		
(20 to 50) kHz	76 $\mu$ V/V + 0.9 mV		
(50 to 100) kHz	137 $\mu$ V/V + 2.3 mV		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Generate <sup>1</sup>	(220 to 750) V 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (750 to 1 000) V 40 Hz to 1 kHz (1 to 20) kHz (20 to 30) kHz	68 $\mu$ V/V + 3 mV 126 $\mu$ V/V + 5 mV 457 $\mu$ V/V + 8 mV 1.8 mV/V + 34 mV 68 $\mu$ V/V + 3 mV 126 $\mu$ V/V + 5 mV 457 $\mu$ V/V + 8 mV	Fluke 5730A Multifunction Calibrator, 5725A Amplifier
AC Voltage – Generate <sup>1</sup> Wideband Absolute	(0.3 to 1.1) mV (10 to 30 Hz) 30 Hz to 500 kHz (0.5 to 1.2) MHz (1.2 to 2) MHz (2 to 12) MHz (12 to 20) MHz (20 to 30) MHz (1.1 to 3.3) mV (10 to 30 Hz) 30 Hz to 500 kHz (0.5 to 1.2) MHz (1.2 to 2) MHz (2 to 12) MHz (12 to 20) MHz (20 to 30) MHz (3.3 to 11) mV (10 to 30 Hz) 30 Hz to 500 kHz (0.5 to 1.2) MHz (1.2 to 2) MHz (2 to 12) MHz (12 to 20) MHz (20 to 30) MHz (11 to 33) mV (10 to 30 Hz) 30 Hz to 500 kHz (0.5 to 1.2) MHz (1.2 to 2) MHz (2 to 12) MHz (12 to 20) MHz (20 to 30) MHz	0.65 % of output + 1.5 $\mu$ V 0.61 % of output + 1.5 $\mu$ V 0.63 % of output + 3.8 $\mu$ V 0.63 % of output + 3.8 $\mu$ V 0.68 % of output + 3.8 $\mu$ V 0.76 % of output + 3.8 $\mu$ V 1.3 % of output + 13 $\mu$ V 0.58 % of output + 2.3 $\mu$ V 0.53 % of output + 2.3 $\mu$ V 0.54 % of output + 4.6 $\mu$ V 0.54 % of output + 4.6 $\mu$ V 0.58 % of output + 4.6 $\mu$ V 0.65 % of output + 4.6 $\mu$ V 1.3 % of output + 4.6 $\mu$ V 0.58 % of output + 6.1 $\mu$ V 0.53 % of output + 6.1 $\mu$ V 0.54 % of output + 8.4 $\mu$ V 0.54 % of output + 8.4 $\mu$ V 0.55 % of output + 8.4 $\mu$ V 0.61 % of output + 8.4 $\mu$ V 0.93 % of output + 8.4 $\mu$ V 0.52 % of output + 12 $\mu$ V 0.46 % of output + 12 $\mu$ V 0.47 % of output + 14 $\mu$ V 0.47 % of output + 14 $\mu$ V 0.49 % of output + 14 $\mu$ V 0.55 % of output + 14 $\mu$ V 0.89 % of output + 14 $\mu$ V	Fluke 5730A Option 003 Multifunction Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Generate <sup>1</sup> Wideband Absolute	(33 to 110) mV		Fluke 5730A Option 003 Multifunction Calibrator
	(10 to 30 Hz)	0.52 % of output + 30 μV	
	30 Hz to 500 kHz	0.46 % of output + 30 μV	
	(0.5 to 1.2) MHz	0.47 % of output + 33 μV	
	(1.2 to 2) MHz	0.47 % of output + 33 μV	
	(2 to 12) MHz	0.49 % of output + 33 μV	
	(12 to 20) MHz	0.55 % of output + 33 μV	
	(20 to 30) MHz	0.89 % of output + 33 μV	
	(110 to 330) mV		
	(10 to 30 Hz)	0.45 % of output + 0.1 mV	
	30 Hz to 500 kHz	0.38 % of output + 0.1 mV	
	(0.5 to 1.2) MHz	0.4 % of output + 0.1 mV	
	(1.2 to 2) MHz	0.4 % of output + 0.1 mV	
	(2 to 12) MHz	0.42 % of output + 0.1 mV	
	(12 to 20) MHz	0.49 % of output + 0.1 mV	
	(20 to 30) MHz	0.85 % of output + 0.1 mV	
	(0.33 to 1.1) V		
	(10 to 30 Hz)	0.45 % of output + 0.3 mV	
	30 Hz to 500 kHz	0.38 % of output + 0.3 mV	
	(0.5 to 1.2) MHz	0.4 % of output + 0.3 mV	
(1.2 to 2) MHz	0.4 % of output + 0.3 mV		
(2 to 12) MHz	0.42 % of output + 0.3 mV		
(12 to 20) MHz	0.49 % of output + 0.3 mV		
(20 to 30) MHz	0.85 % of output + 0.3 mV		
(1.1 to 3.5) V			
(10 to 30 Hz)	0.39 % of output + 0.4 mV		
30 Hz to 500 kHz	0.3 % of output + 0.4 mV		
(0.5 to 1.2) MHz	0.32 % of output + 0.4 mV		
(1.2 to 2) MHz	0.32 % of output + 0.4 mV		
(2 to 12) MHz	0.35 % of output + 0.4 mV		
(12 to 20) MHz	0.44 % of output + 0.4 mV		
(20 to 30) MHz	0.82 % of output + 0.4 mV		
AC Voltage – Measure <sup>1</sup>	(1 to 10) mV		Agilent 3458A Multimeter
	(1 to 40) Hz	0.03 % of reading + 3 μV	
	40 Hz to 1 kHz	0.02 % of reading + 1.1 μV	
	(1 to 20) kHz	0.03 % of reading + 1.1 μV	
	(20 to 50) kHz	0.1 % of reading + 1.1 μV	
	(50 to 100) kHz	0.5 % of reading + 1.1 μV	
	100 kHz to 1 MHz	1.2 % of reading + 5 μV	
	(1 to 4) MHz	7 % of reading + 7 μV	
(4 to 8) MHz	20 % of reading + 8 μV		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure <sup>1</sup>	(10 to 100) mV		Agilent 3458A Multimeter
	(1 to 40) Hz	0.007 % of reading + 4 μV	
	40 Hz to 1 kHz	0.007 % of reading + 2 μV	
	(1 to 20) kHz	0.014 % of reading + 2 μV	
	(20 to 50) kHz	0.03 % of reading + 2 μV	
	(50 to 100) kHz	0.08 % of reading + 2 μV	
	(100 to 300) kHz	0.3 % of reading + 10 μV	
	300 kHz to 1 MHz	1 % of reading + 10 μV	
	(1 to 2) MHz	1.5 % of reading + 70 μV	
	(2 to 4) MHz	4 % of reading + 70 μV	
	(4 to 8) MHz	4 % of reading + 80 μV	
	(8 to 10) MHz	15 % of reading + 100 μV	
	(0.10 to 1) V		
	(1 to 40) Hz	0.007 % of reading + 40 μV	
	40 Hz to 1 kHz	0.07 % of reading + 20 μV	
	(1 to 20) kHz	0.014 % of reading + 20 μV	
	(20 to 50) kHz	0.03 % of reading + 20 μV	
	(50 to 100) kHz	0.08 % of reading + 20 μV	
	(100 to 300) kHz	0.3 % of reading + 100 μV	
	300 kHz to 1 MHz	1 % of reading + 100 μV	
	(1 to 2) MHz	1.5 % of reading + 0.7 mV	
	(2 to 4) MHz	4 % of reading + 0.7 mV	
	(4 to 8) MHz	4 % of reading + 0.8 mV	
	(8 to 10) MHz	15 % of reading + 1 mV	
	(1 to 10) V		
	(1 to 40) Hz	0.007 % of reading + 0.4 mV	
	40 Hz to 1 kHz	0.007 % of reading + 0.2 mV	
(1 to 20) kHz	0.014 % of reading + 0.2 mV		
(20 to 50) kHz	0.03 % of reading + 0.2 mV		
(50 to 100) kHz	0.08 % of reading + 0.2 mV		
(100 to 300) kHz	0.3 % of reading + 1 mV		
300 kHz to 1 MHz	1 % of reading + 1 mV		
(1 to 2) MHz	1.5 % of reading + 7 mV		
(2 to 4) MHz	4 % of reading + 7 mV		
(4 to 8) MHz	4 % of reading + 8 mV		
(8 to 10) MHz	15 % of reading + 10 mV		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure <sup>1</sup>	(10 to 100) V		Agilent 3458A/002 Precision Multimeter
	(1 to 40) Hz	0.02 % of reading + 4 mV	
	40 Hz to 1 kHz	0.02 % of reading + 2 mV	
	(1 to 20) kHz	0.02 % of reading + 2 mV	
	(20 to 50) kHz	0.035 % of reading + 2 mV	
	(50 to 100) kHz	0.12 % of reading + 2 mV	
	(100 to 300) kHz	0.4 % of reading + 10 mV	
	300 kHz to 1 MHz	1.5 % of reading + 10 mV	
	(100 to 700) V		
	(1 to 40) Hz	0.04 % of reading + 40 mV	
	40 Hz to 1 kHz	0.04 % of reading + 20 mV	
	(1 to 20) kHz	0.06 % of reading + 20 mV	
	(20 to 50) kHz	0.12 % of reading + 20 mV	
	(50 to 100) kHz	0.3 % of reading + 20 mV	
AC Voltage – Measure <sup>1</sup>	(1 to 5) kV	0.5 % of reading	Ross VD30 Voltage Divider, Fluke 287 Multimeter
	60 Hz		
	(5 to 21) kV	0.5 % of reading	
	60 Hz		
AC Current – Generate <sup>1</sup>	(9 to 220) $\mu$ A		Fluke 5730A Multifunction Calibrator
	(10 to 20) Hz	228 $\mu$ A/A + 15 nA	
	(20 to 40) Hz	152 $\mu$ A/A + 10 nA	
	40 Hz to 1 kHz	91 $\mu$ A/A + 8 nA	
	(1 to 5) kHz	266 $\mu$ A/A + 12 nA	
	(5 to 10) kHz	989 $\mu$ A/A + 61 nA	
	(0.22 to 2.2) mA		
	(10 to 20) Hz	228 $\mu$ A/A + 39 nA	
	(20 to 40) Hz	152 $\mu$ A/A + 31 nA	
	40 Hz to 1 kHz	91 $\mu$ A/A + 31 nA	
	(1 to 5) kHz	183 $\mu$ A/A + 99 nA	
	(5 to 10) kHz	989 $\mu$ A/A + 609 nA	
	(2.2 to 22) mA		
	(10 to 20) Hz	228 $\mu$ A/A + 385 nA	
	(20 to 40) Hz	152 $\mu$ A/A + 310 nA	
	40 Hz to 1 kHz	91 $\mu$ A/A + 310 nA	
	(1 to 5) kHz	183 $\mu$ A/A + 536 nA	
	(5 to 10) kHz	989 $\mu$ A/A + 4566 nA	



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Generate <sup>1</sup>	(22 to 220) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (0.22 to 2.2) A 20 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	228 $\mu$ A/A + 4 $\mu$ A 152 $\mu$ A/A + 3 $\mu$ A 91 $\mu$ A/A + 2 $\mu$ A 183 $\mu$ A/A + 3 $\mu$ A 989 $\mu$ A/A + 9 $\mu$ A 228 $\mu$ A/A + 31 $\mu$ A 380 $\mu$ A/A + 76 $\mu$ A 0.61 % of reading + 152 $\mu$ A	Fluke 5730A Multifunction Calibrator
AC Current – Generate <sup>1</sup>	(2.2 to 11) A 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	350 $\mu$ A /A + 141 $\mu$ A 723 $\mu$ A /A + 295 $\mu$ A 0.27 % of reading + 573 $\mu$ A	Fluke 5730A Multifunction Calibrator, 5725 Amplifier
AC Current – Generate <sup>1</sup>	(11 to 20.5) A (45 to 100) Hz 100 Hz to 1 kHz (1 to 5) kHz	0.09 % of output + 5 mA 0.11 % of output + 5 mA 2.3 % of output + 5 mA	Fluke 5522A Multiproduct Calibrator
AC Current – Generate <sup>1</sup> Clamp Meters	(3.3 to 30) A (10 to 100) Hz (100 to 440) Hz (30 to 200) A 10 to 100 Hz (100 to 440) Hz	0.22 % of output + 0.028 A 0.3 % of output + 0.07 A 0.22 % of output + 0.032 A 0.79 % of output + 0.08 A	Fluke 5522A Multiproduct Calibrator / 9100-200 x10 Coil
AC Current – Generate <sup>1</sup> Clamp Meters	(16.5 to 150) A (10 to 100) Hz (100 to 440) Hz (150 to 1 000) A (10 to 100) Hz (100 to 440) Hz	0.22 % of output + 0.029 A 0.3 % of output + 0.08 A 0.22 % of output + 0.081 A 0.79 % of output + 0.20 A	Fluke 5522A Multiproduct Calibrator / 9100-200 x50 Coil
AC Current – Measure <sup>1</sup>	(5 to 100) $\mu$ A (10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz 100 $\mu$ A to 1 mA (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz	0.4 % of reading + 0.03 $\mu$ A 0.15 % of reading + 0.03 $\mu$ A 0.06 % of reading + 0.03 $\mu$ A 0.4 % of reading + 0.2 $\mu$ A 0.15 % of reading + 0.2 $\mu$ A 0.06 % of reading + 0.2 $\mu$ A 0.03 % of reading + 0.2 $\mu$ A	Agilent 3458A Multimeter



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Measure <sup>1</sup>	(1 to 10) mA		Agilent 3458A Multimeter
	(10 to 20) Hz	0.4 % of reading + 2 μA	
	(20 to 45) Hz	0.15 % of reading + 2 μA	
	(45 to 100) Hz	0.06 % of reading + 2 μA	
	100 Hz to 5 kHz	0.03 % of reading + 2 μA	
	(10 to 100) mA		
	(10 to 20) Hz	0.4 % of reading + 20 μA	
	(20 to 45) Hz	0.15 % of reading + 20 μA	
	(45 to 100) Hz	0.06 % of reading + 20 μA	
	100 Hz to 5 kHz	0.03 % of reading + 20 μA	
	100 mA to 1 A		
	(10 to 20) Hz	0.4 % of reading + 0.2 mA	
(20 to 45) Hz	0.15 % of reading + 0.2 mA		
(45 to 100) Hz	0.06 % of reading + 0.2 mA		
100 Hz to 5 kHz	0.03 % of reading + 0.2 mA		
AC Current – Measure <sup>1</sup>	(1 to 3) A		Fluke 8846A Multimeter
	(3 to 5) Hz	0.84 % of reading + 1.4 mA	
	(5 to 10) Hz	0.27 % of reading + 1.4 mA	
	10 Hz to 5 kHz	0.11 % of reading + 1.4 mA	
	(5 to 10) kHz	0.27 % of reading + 16 mA	
	(3 to 10) A		
	(3 to 5) Hz	0.84 % of reading + 4.6 mA	
	(5 to 10) Hz	0.27 % of reading + 4.6 mA	
	10 Hz to 5 kHz	0.11 % of reading + 4.6 mA	
	(5 to 10) kHz	0.27 % of reading + 53 mA	
Resistance – Generate <sup>1</sup> Fixed Points	(1, 1.9) Ω	84 μΩ/Ω + 0.1 μΩ	Fluke 5730A Multifunction Calibrator
	(10, 19) Ω	21 μΩ/Ω + 1 μΩ	
	(100, 190) Ω	9.1 μΩ/Ω + 6 μΩ	
	(1, 1.9) kΩ	6.1 μΩ/Ω + 60 μΩ	
	(10, 19) kΩ	6.1 μΩ/Ω + 0.6 mΩ	
	100 kΩ	7.6 μΩ/Ω + 6 mΩ	
	190 kΩ	9.1 μΩ/Ω + 6 mΩ	
	1 MΩ	11.4 μΩ/Ω + 60 mΩ	
	1.9 MΩ	16 μΩ/Ω + 60 mΩ	
	10 MΩ	35 μΩ/Ω + 0.6 Ω	
	19 MΩ	42 mΩ/Ω + 0.6 Ω	
	100 MΩ	91 mΩ/Ω + 6 Ω	



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Resistance – Generate <sup>1</sup> Fixed Points	100 V 100 kΩ (100 to 1 000) V (1, 10, 100) MΩ (1, 10) GΩ 100 GΩ	1 % of reading 1 % of reading 1 % of reading 2 % of reading	TMI RB Resistance Standard
Resistance – Measure <sup>1</sup>	Up to 12 Ω (10 to 120) Ω (0.1 to 1.2) kΩ (1 to 12) kΩ (10 to 120) kΩ (0.1 to 1.2) MΩ (1 to 12) MΩ (10 to 120) MΩ (0.1 to 1.2) GΩ	15 μΩ/Ω + 56 μΩ 12 μΩ/Ω + 0.5 mΩ 10 μΩ/Ω + 0.6 mΩ 10 μΩ/Ω + 5.6 mΩ 10 μΩ/Ω + 56 mΩ 15 μΩ/Ω + 2.2 Ω 50 μΩ/Ω + 120 Ω 500 μΩ/Ω + 1.2 kΩ 0.5 % of reading + 70 kΩ	Agilent 3458A Multimeter
Resistance – Generate <sup>1</sup>	Up to 11 Ω (11 to 33) Ω (33 to 110) Ω (110 to 330) Ω (0.33 to 1.1) kΩ (1.1 to 3.3) kΩ (3.3 to 11) kΩ (11 to 33) kΩ (33 to 110) kΩ (110 to 330) kΩ (0.33 to 1.1) MΩ (1.1 to 3.3) MΩ (3.3 to 11) MΩ (11 to 33) MΩ (33 to 110) MΩ (110 to 330) MΩ (0.33 to 1.1) GΩ	30 μΩ/Ω + 0.001 Ω 23 μΩ/Ω + 0.001 Ω 21 μΩ/Ω + 0.001 Ω 21 μΩ/Ω + 0.002 Ω 21 μΩ/Ω + 0.002 Ω 21 μΩ/Ω + 0.02 Ω 21 μΩ/Ω + 0.02 Ω 21 μΩ/Ω + 0.2 Ω 21 μΩ/Ω + 0.2 Ω 24 μΩ/Ω + 2 Ω 24 μΩ/Ω + 2 Ω 46 μΩ/Ω + 23 Ω 99 μΩ/Ω + 38 Ω 190 μΩ/Ω + 1.9 kΩ 380 μΩ/Ω + 2.3 kΩ 0.23 % of setting + 76 kΩ 1.1 % of reading + 380 kΩ	Fluke 5522A Multiproduct Calibrator
Capacitance – Generate <sup>1</sup>	(220 to 400) pF (0.4 to 3.299 9) nF (3.3 to 10.999 9) nF (11 to 32.999 9) nF (33 to 109.999) nF (110 to 329.999) nF (0.33 to 1.099 99) μF (1.1 to 3.299 99) μF (3.3 to 10.999 9) μF	0.38% of output + 7.6 pF 0.38 % of output + 0.01 nF 0.19 % of output + 0.01 nF 0.19 % of output + 0.08 nF 0.19 % of output + 0.08 nF 0.19 % of output + 0.23 nF 0.19 % of output + 0.76 nF 0.19 % of output + 2.3 nF 0.19 % of output + 7.6 nF	Fluke 5522A Multiproduct Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Capacitance – Generate <sup>1</sup>	(11 to 32.999 9) $\mu$ F (33 to 109.999) $\mu$ F (110 to 329.999) $\mu$ F (0.33 to 1.099 99) mF (1.1 to 3.299 99) mF (3.3 to 10.999 9) mF (11 to 32.999 9) mF (33 to 110) mF	0.3 % of output + 23 nF 0.34 % of output + 76 nF 0.34 % of output + 228 nF 0.34 % of output + 0.76 $\mu$ F 0.34 % of output + 2.3 $\mu$ F 0.34 % of output + 7.6 $\mu$ F 0.57 % of output + 23 $\mu$ F 0.84 % of output + 76 $\mu$ F	Fluke 5522A Multiproduct Calibrator
Oscilloscopes Calibration <sup>1</sup> – Generate Voltage DC - 50 $\Omega$	(1 to 24.999) mV (25 to 109.99) mV (110mV to 2.1999) V (2.2 to 6.6) V	0.19 % of Output + 31 $\mu$ V 0.19 % of Output + 36 $\mu$ V 0.19 % of Output + 87 $\mu$ V 0.19 % of Output + 0.6 mV	Fluke 5820A Oscilloscope Calibrator w/ GHz Option
DC - 1M $\Omega$	(1 to 24.999) mV (25 to 109.99) mV (110mV to 2.1999) V (2.2 to 10.999) V (11 to 130) V	0.019 % of reading + 20 $\mu$ V 0.019 % of reading + 25 $\mu$ V 0.019 % of reading + 76 $\mu$ V 0.019 % of reading + 0.6 mV 0.019 % of reading + 6 mV	
Square Wave 10 Hz to 10 kHz - 50 $\Omega$	(1 to 24.999) mVpp (25 to 109.99) mVpp (110mV to 2.1999) Vpp (2.2 to 6.6) Vpp	0.19 % of Output + 31 $\mu$ V 0.19 % of Output + 36 $\mu$ V 0.19 % of Output + 87 $\mu$ V 0.19 % of Output + 0.6 mV	
Square Wave 10 Hz to 1 kHz - 1M $\Omega$	(1 to 24.999) mV (25 to 109.99) mV (110mV to 2.1999) V (2.2 to 10.999) V (11 to 130) V	0.038 % of reading + 4 $\mu$ V 0.038 % of reading + 9 $\mu$ V 0.038 % of reading + 60 $\mu$ V 0.038 % of reading + 0.6 mV 0.038 % of reading + 6 mV	
Square Wave (1 to 10) kHz - 1M $\Omega$	(1 to 24.999) mV (25 to 109.99) mV (110mV to 2.1999) V (2.2 to 10.999) V (11 to 130) V	0.19 % of reading + 31 $\mu$ V 0.19 % of reading + 36 $\mu$ V 0.19 % of reading + 87 $\mu$ V 0.19 % of reading + 0.6 mV 0.19 % of reading + 6 mV	



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Oscilloscopes Calibration – Generate <sup>1</sup>  Leveled Sine Flatness 50 kHz to 10 MHz Reference	3 dB Bandwidth (5 to 50) mVpp		Fluke 5820A Oscilloscope Calibrator w/ GHz Option
	50 kHz to 100 MHz	0.34 dB	
	(100 to 300) MHz	0.36 dB	
	(300 to 500) MHz	0.42 dB	
	(500 to 600) MHz	0.46 dB	
	(600 to 1 600) MHz	0.5 dB	
	(1 600 to 2 100) MHz	0.56 dB	
	50 mV to 3.5 Vpp		
	50 kHz to 100 MHz	0.24 dB	
	(100 to 300) MHz	0.24 dB	
	(300 to 500) MHz	0.32 dB	
	(500 to 600) MHz	0.34 dB	
(600 to 1 600) MHz	0.4 dB		
(1 600 to 2 100) MHz	0.44 dB		
(3.5 to 5) Vpp			
50 kHz to 100 MHz	0.24 dB		
(100 to 300) MHz	0.24 dB		
(300 to 500) MHz	0.32 dB		
(500 to 600) MHz	0.34 dB		
Oscilloscopes Calibration – Generate <sup>1</sup> Leveled Sine Flatness	3 dB Bandwidth 50 mV to 3.5Vpp (2 100 to 4 000) MHz (4 000 to 8 000) MHz (8 000 to 18 000) MHz	0.25 dB 0.35 dB 0.46 dB	EPM Power Meter w/ E Series Power Sensors
Oscilloscopes Calibration – Generate <sup>1</sup> Time Marker	500 ps to 20 ms 50 ms to 5 s	0.25 $\mu$ s/s 1.9 $\mu$ s/s + 3.8 $\mu$ Hz	Fluke 5820A Oscilloscope Calibrator w/ GHz Option
Oscilloscopes Calibration – Measure <sup>1</sup> Input Impedance Resistance Leakage	(40 to 60) $\Omega$ 500 k $\Omega$ to 1.5 M $\Omega$ (0 to 5.99) V	0.08 % of reading 0.08 % of reading 0.038 % of reading + 0.8 mV	Fluke 5820A Oscilloscope Calibrator w/ GHz Option
DC/Low Frequency Power – Generate <sup>1</sup>	11 $\mu$ W to 20 kW DC (45 to 65) Hz, PF = 1	0.21 % of reading 0.25 % of reading	Fluke 5522A Multiproduct Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Calibration of Thermocouple Indicators <sup>1</sup>	Type B		Fluke 7526A Process Calibrator
	(600 to 800) °C	0.27 °C	
	(800 to 1 550) °C	0.22 °C	
	(1 550 to 1 820) °C	0.17 °C	
	Type C		
	(0 to 1 000) °C	0.13 °C	
	(1 000 to 1 800) °C	0.18 °C	
	(1 800 to 2 000) °C	0.2 °C	
	(2 000 to 2 316) °C	0.27 °C	
	Type E		
	(-250 to -200) °C	0.19 °C	
	(-200 to -100) °C	0.1 °C	
	(-100 to 0) °C	0.07 °C	
	(0 to 600) °C	0.07 °C	
	(600 to 1 000) °C	0.08 °C	
	Type J		
	(-210 to -100) °C	0.11 °C	
	(-100 to 800) °C	0.07 °C	
	(800 to 1 200) °C	0.08 °C	
	Type K		
	(-250 to -200) °C	0.35 °C	
	(-200 to -100) °C	0.13 °C	
	(-100 to 800) °C	0.08 °C	
	(800 to 1 372) °C	0.1 °C	
Type L			
(-200 to -100) °C	0.08 °C		
(-100 to 900) °C	0.07 °C		
Type N			
(-250 to -200) °C	0.56 °C		
(-200 to -100) °C	0.18 °C		
(-100 to 0) °C	0.1 °C		
(0 to 100) °C	0.09 °C		
(100 to 800) °C	0.08 °C		
(800 to 1 300) °C	0.1 °C		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Calibration of Thermocouple Indicators <sup>1</sup>	Type R		Fluke 7526A Process Calibrator
	(-50 to -25) °C	0.42 °C	
	(-25 to 0) °C	0.34 °C	
	(0 to 100) °C	0.3 °C	
	(100 to 400) °C	0.22 °C	
	(400 to 600) °C	0.17 °C	
	(600 to 1 000) °C	0.16 °C	
	(1 000 to 1 600) °C	0.15 °C	
	(600 to 1 767) °C	0.18 °C	
	Type S		
	(50 to -25) °C	0.39 °C	
	(-25 to 0) °C	0.33 °C	
	(0 to 100) °C	0.29 °C	
	(100 to 400) °C	0.22 °C	
	(400 to 600) °C	0.18 °C	
	(600 to 1 600) °C	0.17 °C	
	(1 600 to 1 767) °C	0.2 °C	
	Type T		
(-250 to -200) °C	0.26 °C		
(-200 to -100) °C	0.13 °C		
(-100 to 0) °C	0.09 °C		
(0 to 400) °C	0.07 °C		
Type U			
(-200 to 0) °C	0.13 °C		
(0 to 600) °C	0.08 °C		
Electrical Calibration of RTD Indicators <sup>1</sup>	Pt 385, 100 Ω		Fluke 7526A Process Calibrator
	(-200 to 800) °C	0.05 °C	
	Pt 3926, 100 Ω		
	(-200 to 630) °C	0.05 °C	
	Pt 3916, 100 Ω		
	(-200 to 630) °C	0.05 °C	
	Pt 385, 200 Ω		
	(-200 to 400) °C	0.4 °C	
(400 to 630) °C	0.5 °C		
Pt 385, 500 Ω			
(-200 to 630) °C	0.17 °C		
Pt 385, 1 000 Ω			
(-200 to 630) °C	0.09 °C		
Power Meter Range Calibration <sup>1</sup>	3 μW to 100 mW	0.25 % of reading	HP 11683A Power Meter Calibrator



ANSI National Accreditation Board

Electrical – RF/Microwave

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Flatness – Measure <sup>1</sup>	9 kHz to 2 000 MHz (20 to -10) dBm (-10 to -30) dBm (-30 to -40) dBm (-40 to -42) dBm (2 to 14) GHz (20 to -10) dBm (-10 to -30) dBm (-30 to -40) dBm (-40 to -42) dBm (14 to 18) GHz (20 to -10) dBm (-10 to -30) dBm (-30 to -40) dBm (-40 to -42) dBm	0.1 dB 0.1 dB 0.11 dB 0.12 dB 0.1 dB 0.09 dB 0.1 dB 0.11 dB 0.11 dB 0.12 dB 0.12 dB 0.12 dB 0.13 dB	Agilent EPM Series Power Meter w/E9304A H18 Power Sensor
RF Power Measure <sup>1</sup>	9 kHz to 14 000 MHz (20 to 0) dB (0 to -40) dB (-40 to -50) dB (-50 to -55) dB (14 000 to 18 000) MHz (20 to 0) dB (0 to -40) dB (-40 to -50) dB (-50 to -55) dB	0.13 dB 0.15 dB 0.35 dB 0.93 dB 0.12 dB 0.16 dB 0.35 dB 0.93 dB	Agilent EPM Series Power Meter w/E9304A H18 Power Sensor
RF Power Measure <sup>1</sup>	(100 to 2 000) MHz (20 to 10) dB (10 to 0) dB (0 to -10) dB (-10 to -20) dB (-20 to -25) dB (2 000 to 12 400) MHz (20 to 10) dB (10 to 0) dB (0 to -10) dB (-10 to -20) dB (-20 to -25) dB	0.14 dB 0.06 dB 0.07 dB 0.08 dB 0.15 dB 0.15 dB 0.07 dB 0.07 dB 0.08 dB 0.15 dB	Agilent EPM Series Power Meter 8487A Power Sensor



ANSI National Accreditation Board

Electrical – RF/Microwave

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power Measure <sup>1</sup>	(12 400 to 18 000) MHz		Agilent EPM Series Power Meter 8487A Power Sensor
	(20 to 10) dB	0.15 dB	
	(10 to 0) dB	0.08 dB	
	(0 to -10) dB	0.08 dB	
	(-10 to -20) dB	0.09 dB	
	(-20 to -25) dB	0.15 dB	
	(18 000 to 26 500) MHz		
	(20 to 10) dB	0.16 dB	
	(10 to 0) dB	0.1 dB	
	(0 to -10) dB	0.1 dB	
	(-10 to -20) dB	0.11 dB	
	(-20 to -25) dB	0.16 dB	
	(26 5000 to 40 000) MHz		
	(20 to 10) dB	0.17 dB	
	(10 to 0) dB	0.12 dB	
	(0 to -10) dB	0.12 dB	
	(-10 to -20) dB	0.13 dB	
	(-20 to -25) dB	0.18 dB	
(40 000 to 50 000) MHz			
(20 to 10) dB	0.22 dB		
(10 to 0) dB	0.19 dB		
(0 to -10) dB	0.19 dB		
(-10 to -20) dB	0.19 dB		
(-20 to -25) dB	0.23 dB		
RF Power Measure <sup>1</sup>	100 kHz to 30 MHz		Agilent N5531S Measuring Receiver N1912A w/E9304A H18 Power Sensor
	(20 to 0) dB	0.12 dB	
	(0 to -58) dB	0.13 dB	
	(-58 to -78) dB	0.15 dB	
	(-78 to -110) dB	0.19 dB	
	(-110 to -115) dB	0.26 dB	
	(-115 to -120) dB	0.39 dB	
	(-120 to -125) dB	0.59 dB	
RF Power Measure <sup>1</sup>	(30 to 2 000) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor
	(30 to 20) dB	0.36 dB	
	(20 to 0) dB	0.2 dB	
	(0 to -58) dB	0.22 dB	
	(-58 to -78) dB	0.23 dB	
	(-78 to -110) dB	0.25 dB	
	(-110 to -115) dB	0.27 dB	
	(-115 to -120) dB	0.33 dB	
(-120 to -125) dB	0.46 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power Measure <sup>1</sup>	(2 000 to 3 050) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor
	(30 to 20) dB	0.37 dB	
	(20 to 0) dB	0.21 dB	
	(0 to -58) dB	0.23 dB	
	(-58 to -78) dB	0.24 dB	
	(-78 to -110) dB	0.26 dB	
	(-110 to -115) dB	0.28 dB	
	(-115 to -120) dB	0.34 dB	
	(-120 to -125) dB	0.7 dB	
	(3 050 to 6 600) MHz		
	(30 to 20) dB	0.37 dB	
	(20 to 0) dB	0.21 dB	
	(0 to -58) dB	0.23 dB	
	(-58 to -78) dB	0.24 dB	
	(-78 to -110) dB	0.29 dB	
	(-110 to -115) dB	0.38 dB	
	(-115 to -120) dB	0.53 dB	
	(6 600 to 13 200) MHz		
	(30 to 20) dB	0.37 dB	
	(20 to 0) dB	0.21 dB	
	(0 to -58) dB	0.23 dB	
	(-58 to -78) dB	0.24 dB	
	(-78 to -110) dB	0.34 dB	
	(-110 to -115) dB	0.46 dB	
	(-115 to -120) dB	0.65 dB	
	(13 200 to 18 000) MHz		
	(30 to 20) dB	0.37 dB	
	(20 to 0) dB	0.21 dB	
	(0 to -58) dB	0.23 dB	
	(-58 to -78) dB	0.24 dB	
	(-78 to -90) dB	0.26 dB	
	(-90 to -95) dB	0.26 dB	
	(-95 to -100) dB	0.26 dB	
	(-100 to -105) dB	0.29 dB	
	(-105 to -110) dB	0.38 dB	
	(-110 to -115) dB	0.53 dB	
(-115 to -120) dB	0.75 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power Measure <sup>1</sup>	(18 000 to 19 200) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor
	(30 to 20) dB	0.39 dB	
	(20 to 0) dB	0.25 dB	
	(0 to -58) dB	0.24 dB	
	(-58 to -78) dB	0.25 dB	
	(-78 to -90) dB	0.27 dB	
	(-90 to -95) dB	0.27 dB	
	(-95 to -100) dB	0.27 dB	
	(-100 to -105) dB	0.3 dB	
	(-105 to -110) dB	0.38 dB	
	(-110 to -115) dB	0.53 dB	
	(-115 to -120) dB	0.75 dB	
	(19 200 to 26 500) MHz		
	(30 to 20) dB	0.39 dB	
	(20 to 0) dB	0.25 dB	
	(0 to -58) dB	0.24 dB	
	(-58 to -78) dB	0.25 dB	
	(-78 to -90) dB	0.28 dB	
	(-90 to -95) dB	0.33 dB	
	(-95 to -100) dB	0.43 dB	
	(-100 to -105) dB	0.61 dB	
	(-105 to -110) dB	0.85 dB	
	(-110 to -115) dB	1.2 dB	
	(-115 to -120) dB	1.5 dB	
	(26 500 to 31 150) MHz		
	(30 to 20) dB	0.42 dB	
	(20 to 0) dB	0.3 dB	
(0 to -58) dB	0.34 dB		
(-58 to -78) dB	0.34 dB		
(-78 to -90) dB	0.36 dB		
(-90 to -95) dB	0.39 dB		
(-95 to -100) dB	0.46 dB		
(-100 to -105) dB	0.61 dB		
(-105 to -110) dB	0.82 dB		
(-110 to -115) dB	1.1 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power Measure <sup>1</sup>	(31 150 to 41 000) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor
	(30 to 20) dB	0.42 dB	
	(20 to 0) dB	0.3 dB	
	(0 to -58) dB	0.34 dB	
	(-58 to -78) dB	0.35 dB	
	(-78 to -90) dB	0.48 dB	
	(-90 to -95) dB	0.64 dB	
	(-95 to -100) dB	0.87 dB	
	(-100 to -105) dB	1.2 dB	
	(-105 to -110) dB	1.5 dB	
	(41 000 to 45 000) MHz		
	(30 to 20) dB	0.42 dB	
	(20 to 0) dB	0.3 dB	
	(0 to -58) dB	0.34 dB	
	(-58 to -78) dB	0.38 dB	
	(-78 to -90) dB	0.68 dB	
	(-90 to -95) dB	0.93 dB	
	(-95 to -100) dB	1.2 dB	
(-100 to -105) dB	1.6 dB		
RF Attenuation – Measure <sup>1</sup>	(45 000 to 50 000) MHz		
	(30 to 20) dB	0.42 dB	
	(20 to 0) dB	0.3 dB	
	(0 to -58) dB	0.34 dB	
	(-58 to -78) dB	0.67 dB	
	(-78 to -90) dB	1.4 dB	
	(30 to 3 050) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
(30 to 40) dB	0.035 dB		
(40 to 50) dB	0.04 dB		
(50 to 60) dB	0.076 dB		
(60 to 70) dB	0.081 dB		
(70 to 80) dB	0.12 dB		
(80 to 90) dB	0.12 dB		
(90 to 100) dB	0.13 dB		
(100 to 110) dB	0.13 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Attenuation – Measure <sup>1</sup>	(3 050 to 6 600) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.12 dB	
	(90 to 100) dB	0.13 dB	
	(100 to 110) dB	0.19 dB	
	(6 600 to 13 200) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.12 dB	
	(90 to 100) dB	0.13 dB	
	(100 to 110) dB	0.25 dB	
	(13 200 to 19 200) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
(50 to 60) dB	0.076 dB		
(60 to 70) dB	0.081 dB		
(70 to 80) dB	0.12 dB		
(80 to 90) dB	0.12 dB		
(90 to 100) dB	0.13 dB		
(100 to 110) dB	0.31 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Attenuation – Measure <sup>1</sup>	(19 200 to 26 500) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.14 dB	
	(90 to 100) dB	0.36 dB	
	(100 to 110) dB	0.82 dB	
	(26 500 to 31 150) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.13 dB	
	(90 to 100) dB	0.33 dB	
	(100 to 110) dB	0.77 dB	
	(31 150 to 41 000) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
(50 to 60) dB	0.076 dB		
(60 to 70) dB	0.081 dB		
(70 to 80) dB	0.14 dB		
(80 to 90) dB	0.36 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Attenuation – Measure <sup>1</sup>	(41 000 to 45 000) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.11 dB	
	(70 to 80) dB	0.24 dB	
	(80 to 90) dB	0.6 dB	
	(45 000 to 50 000) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.11 dB	
	(60 to 70) dB	0.29 dB	
(70 to 80) dB	0.7 dB		
(80 to 90) dB	1.4 dB		
Amplitude Modulation Measure <sup>1</sup>	100 kHz to 10 MHz Rate 50 Hz to 10 kHz (5 to 99) % Depth	0.75 % of reading + 0.3 digits	Agilent N5531S Measuring Receiver
	10 MHz to 3 GHz Rate 50 Hz to 100 kHz (5 to 20) % Depth	2.5 % of reading + 0.4 digits	
	(20 to 99) % Depth	1.5 % of reading + 0.4 digits	
	(3 to 26.5) GHz Rate 50 Hz to 100 kHz (5 to 20) % Depth	4.5 % of reading + 0.4 digits	
	(20 to 99) % Depth	1.5 % of reading + 0.4 digits	
	(26.5 to 31.15) GHz Rate 50 Hz to 100 kHz (5 to 20) % Depth	6.8 % of reading + 0.4 digits	
	(20 to 99) % Depth	1.9 % of reading + 0.4 digits	
	(31.15 to 50) GHz Rate 50 Hz to 100 kHz (5 to 20) % Depth	26 % of reading + 0.4 digits	
	(20 to 99) % Depth	6 % of reading + 0.4 digits	



ANSI National Accreditation Board

Electrical – RF/Microwave

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency Modulation Measure <sup>1</sup>  $\beta = \text{deviation} / \text{rate}$	250 kHz to 10 MHz Rates 20 Hz to 10 kHz Peak Dev 200 to 40 kHz 10 MHz to 6.6 GHz Rates 50 Hz to 200 kHz Peak Dev 250 to 400 kHz (6.6 to 13.2) GHz Rates 50 Hz to 200 kHz Peak Dev 250 to 400 kHz (13.2 to 31.15) GHz Rates 50 Hz to 200 kHz Peak Dev 250 to 400 kHz (31.15 to 50) GHz Rates 50 Hz to 200 kHz Peak Dev 250 to 400 kHz	$\beta > 0.2 - 1.5 \% \text{ of reading} + 2 \text{ Hz}$ $\beta > 1.2 - 1 \% \text{ of reading} + 2 \text{ Hz}$  $\beta > 0.2 - 1.5 \% \text{ of reading} + 2 \text{ Hz}$ $\beta > 0.45 - 1 \% \text{ of reading} + 2 \text{ Hz}$  $\beta > 0.2 - 2.5 \% \text{ of reading} + 4 \text{ Hz}$ $\beta > 8 - 1 \% \text{ of reading} + 4 \text{ Hz}$  $\beta > 0.2 - 3.8 \% \text{ of reading} + 9 \text{ Hz}$ $\beta > 16 - 1 \% \text{ of reading} + 9 \text{ Hz}$  $\beta > 0.2 - 8.5 \% \text{ of reading} + 17 \text{ Hz}$ $\beta > 16 - 1 \% \text{ of reading} + 17 \text{ Hz}$	Agilent N5531S Measuring Receiver
Phase Modulation – Measure <sup>1</sup>	100 kHz to 6.6 GHz Deviations > 0.3 rad Deviations > 0.7 rad (6.6 to 13.2) GHz Deviations > 0.6 rad Deviations > 2.0 rad (13.2 to 26.5) GHz Deviations: > 1.2 rad Deviations > 4.0 rad (26.5 to 31.15) GHz Deviations: > 1.3 rad Deviations > 4.0 rad (31.15 to 50) GHz Deviations: > 2.4 rad Deviations > 8.0 rad	3 % of reading + 0.002 rad 1 % of reading + 0.002 rad  3 % of reading + 0.005 rad 1 % of reading + 0.005 rad  3 % of reading + 0.009 rad 1 % of reading + 0.009 rad  3 % of reading + 0.009 rad 1 % of reading + 0.009 rad  3 % of reading + 0.018 rad 1 % of reading + 0.018 rad	Agilent N5531S Measuring Receiver
RF Power Generate <sup>1</sup>	0.25 to 30 MHz (10 to 0) dB (0 to -58) dB (-58 to -78) dB (-78 to -110) dB (-110 to -120) dB (30 to 2 000) MHz (20 to 0) dB (0 to -58) dB (-58 to -78) dB (-78 to -110) dB	0.23 dB 0.23 dB 0.25 dB 0.27 dB 0.43 dB  0.29 dB 0.3 dB 0.31 dB 0.32 dB	Agilent N5531S Measuring Receiver N1912A w/E9304A Power Sensor, E8257D Signal Generator

**Electrical – RF/Microwave**

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power Generate <sup>1</sup>	(2 000 to 3 050) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor, E8257D Signal Generator
	(20 to 0) dB	0.34 dB	
	(0 to -58) dB	0.35 dB	
	(-58 to -78) dB	0.36 dB	
	(-78 to -110) dB	0.39 dB	
	(3 050 to 6 600) MHz		
	(20 to 0) dB	0.34 dB	
	(0 to -58) dB	0.35 dB	
	(-58 to -78) dB	0.36 dB	
	(-78 to -110) dB	0.39 dB	
	(6 600 to 13 200) MHz		
	(20 to 0) dB	0.34 dB	
	(0 to -58) dB	0.35 dB	
	(-58 to -78) dB	0.36 dB	
	(-78 to -110) dB	0.42 dB	
	(13 200 to 18 000) MHz		
	(15 to 0) dB	0.34 dB	
	(0 to -58) dB	0.35 dB	
	(-58 to -78) dB	0.36 dB	
	(-78 to -110) dB	0.46 dB	
	(18 000 to 19 200) MHz		
	(15 to 0) dB	0.41 dB	
	(0 to -58) dB	0.41 dB	
	(-58 to -78) dB	0.42 dB	
	(-78 to -110) dB	0.5 dB	
	(19 200 to 26 500) MHz		
	(15 to 0) dB	0.41 dB	
	(0 to -58) dB	0.41 dB	
(-58 to -78) dB	0.42 dB		
(-78 to -110) dB	0.9 dB		
(26 500 to 31 150) MHz			
(15 to 0) dB	0.62 dB		
(0 to -58) dB	0.63 dB		
(-58 to -78) dB	0.64 dB		
(-78 to -110) dB	0.96 dB		
(31 150 to 41 000) MHz			
(10 to 0) dB	0.82 dB		
(0 to -58) dB	0.83 dB		
(-58 to -78) dB	0.84 dB		
(-78 to -100) dB	1.1 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power Generate <sup>1</sup>	(41 000 to 45 000) MHz (10 to 0) dB (0 to -58) dB (-58 to -78) dB (-78 to -100) dB (45 000 to 50 000) MHz (10 to 0) dB (0 to -58) dB (-58 to -78) dB (-78 to -90) dB	0.82 dB 0.83 dB 0.85 dB 1.4 dB 0.82 dB 0.83 dB 1 dB 1.5 dB	Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor, E8257D Signal Generator
AM Distortion Measure <sup>1</sup> Rate 20 Hz to 1 kHz	(0.1 to 10) MHz AM Depth > 1% (0 to -20) dB (-20 to -30) dB AM Depth > 3% (0 to -20) dB (-20 to -30) dB (-30 to -40) dB 10 MHz to 26.5 GHz AM Depth > 1% (0 to -20) dB (-20 to -30) dB AM Depth > 3% (0 to -20) dB (-20 to -30) dB (-30 to -40) dB (26.5 MHz to 50.0 GHz) AM Depth > 3% (0 to -20) dB AM Depth > 5% (0 to -20) dB (-20 to -30) dB	1.2 dB 2.2 dB 1 dB 1.3 dB 2.4 dB 1.3 dB 2.5 dB 1.1 dB 1.4 dB 3 dB 1.8 dB 1.5 dB 3.5 dB	Agilent N5531S Measuring Receiver



ANSI National Accreditation Board

Electrical – RF/Microwave

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
FM Distortion Measure <sup>1</sup> Rate 20 Hz to 1 kHz	(1 to 6 600) MHz Dev 500 Hz to 2 kHz		Agilent N5531S Measuring Receiver
	(0 to -20) dB	0.26 dB	
	(-20 to -30) dB	0.79 dB	
	(-30 to -40) dB	2.3 dB	
	Dev > 2 kHz		
	(0 to -20) dB	0.09 dB	
	(-20 to -30) dB	0.27 dB	
	(-30 to -40) dB	0.83 dB	
	(-40 to -50) dB	2.4 dB	
	(6.6 to 13.2) GHz		
	Dev > 2.3 kHz		
	(0 to -20) dB	0.26 dB	
	(-20 to -30) dB	0.79 dB	
	(-30 to -40) dB	2.3 dB	
	Dev > 4.5 kHz		
	(0 to -20) dB	0.09 dB	
	(-20 to -30) dB	0.27 dB	
	(-30 to -40) dB	0.83 dB	
	(-40 to -50) dB	2.4 dB	
	(13.2 to 31.15) GHz		
	Dev > 2.7 kHz		
	(0 to -20) dB	0.26 dB	
	(-20 to -30) dB	0.79 dB	
	(-30 to -40) dB	2.3 dB	
Dev > 6.0 kHz			
(0 to -20) dB	0.09 dB		
(-20 to -30) dB	0.27 dB		
(-30 to -40) dB	0.83 dB		
(-40 to -50) dB	2.4 dB		
(31.15 to 50.0) GHz			
Dev > 4 kHz			
(0 to -20) dB	0.26 dB		
(-20 to -30) dB	0.79 dB		
(-30 to -40) dB	2.3 dB		
Dev > 12.0 kHz			
(0 to -20) dB	0.09 dB		
(-20 to -30) dB	0.27 dB		
(-30 to -40) dB	0.83 dB		
(-40 to -50) dB	2.4 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
PM Distortion Measure <sup>1</sup>	(1 to 6 600) MHz		
	Rate (20 to 500) Hz		
	Dev > 0.8 rad		
	(0 to -20) dB	0.26 dB	
	(-20 to -30) dB	0.79 dB	
	(-30 to -40) dB	2.3 dB	
	Dev > 2.5 rad		
	(0 to -20) dB	0.09 dB	
	(-20 to -30) dB	0.27 dB	
	(-30 to -40) dB	0.83 dB	
	(-40 to -50) dB	2.3 dB	
	Rate (500 to 1 000) Hz		
	Dev > 0.4 rad	0.26 dB	
	(0 to -20) dB	0.79 dB	
	(-20 to -30) dB	2.3 dB	
	(-30 to -40) dB		
	Dev > 1.0 rad		
	(0 to -20) dB	0.09 dB	
	(-20 to -30) dB	0.27 dB	
	(-30 to -40) dB	0.83 dB	
	(-40 to -50) dB	2.3 dB	
	(6.6 to 13.2) GHz		
	Rate (20 to 500) Hz		
	Dev > 1.8 rad		
(0 to -20) dB	0.26 dB		
(-20 to -30) dB	0.79 dB		
(-30 to -40) dB	2.3 dB		
Dev > 5.5 rad			
(0 to -20) dB	0.09 dB		
(-20 to -30) dB	0.27 dB		
(-30 to -40) dB	0.83 dB		
(-40 to -50) dB	2.3 dB		
Rate (500 to 1 000) Hz			
Dev > 0.8 rad			
(0 to -20) dB	0.26 dB		
(-20 to -30) dB	0.79 dB		
(-30 to -40) dB	2.3 dB		
Dev > 2.5 rad			
(0 to -20) dB	0.09 dB		
(-20 to -30) dB	0.27 dB		
(-30 to -40) dB	0.83 dB		
(-40 to -50) dB	2.3 dB		

Agilent N5531S  
Measuring Receiver



ANSI National Accreditation Board

Electrical – RF/Microwave

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
PM Distortion Measure <sup>1</sup>	(13.2 to 31.15) GHz Rate (20 to 500) Hz Dev > 3.5 rad (0 to -20) dB (-20 to -30) dB (-30 to -40) dB Dev > 10.0 rad (0 to -20) dB (-20 to -30) dB (-30 to -40) dB (-40 to -50) dB Rate (500 to 1 000) Hz Dev > 3.0 rad (0 to -20) dB (-20 to -30) dB (-30 to -40) dB Dev > 8.0 rad (0 to -20) dB (-20 to -30) dB (-30 to -40) dB (-40 to -50) dB	0.26 dB 0.79 dB 2.3 dB 0.09 dB 0.27 dB 0.83 dB 2.3 dB 0.26 dB 0.79 dB 2.3 dB 0.09 dB 0.27 dB 0.83 dB 2.3 dB	Agilent N5531S Measuring Receiver
Total Harmonic Distortion (THD)	(0 to -60) dB 20 Hz to 20 kHz (0 to -40) dB (-40 to -50) dB (-50 to -60) dB (-60 to -65) dB (20 to 50) kHz (0 to -40) dB (-40 to -50) dB (-50 to -60) dB (50 to 100) kHz (0 to -40) dB (-40 to -50) dB	1 dB 1 dB 1.3 dB 1.7 dB 2 dB 2.1 dB 3 dB 2 dB 2.4 dB	HP 8903B Audio Analyzer



ANSI National Accreditation Board

Electrical – RF/Microwave

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Harmonics Measure <sup>1</sup>	(-80 to -10) dB 2 <sup>nd</sup> through 5 <sup>th</sup> Harmonic 1kHz to 600MHz (600 to 1 320) MHz (1 320 to 2 200) MHz (2 200 to 3 000) MHz (3 000 to 4 400) MHz (4 400 to 5 300) MHz (5 300 to 10 000) MHz 2 <sup>nd</sup> through 4 <sup>th</sup> Harmonic (10 000 to 12 500) MHz 2 <sup>nd</sup> through 3 <sup>rd</sup> Harmonic (12 500 to 16 667) MHz 2 <sup>nd</sup> Harmonic (16 667 to 25 000) MHz	0.37 dB 1.1 dB 1.4 dB 1.4 dB 1.7 dB 1.9 dB 2.1 dB 2.1 dB 2.1 dB 2.3 dB	Agilent E4440A Measuring Receiver
EFT/Burst Generator – Voltage (+/-) Risetime Impulse Duration Burst Duration Burst Period	10 V to 8 kV 5 ns +/- 30 % 50 ns +/- 30 % 15 ms +/- 20 % 300 ms +/- 20 %	2.6 % of reading 0.003 % of reading 0.003 % of reading 0.003 % of reading 0.003 % of reading	IEC/EN 61000-4-4 Tektronix TDS784C Oscilloscope, EFT Attenuator Set
Surge Generator – Front Time (+/-) Risetime Open/Short Circuit (+/-)  Time to Half Value (+/-)  Open Circuit Voltage (+/-) Short Circuit Voltage (+/-)  Ring Wave Voltage Ring Wave Risetime	(0.1 to 50) μs (0.1 to 50) μs  (20 to 1 500) μs  10 V to 18 kV (0.125 to 3) kA  1 kV +/- 10 % 1.5 μs +/- 0.5 μs	0.008 % of reading 0.008 % of reading  0.008 % of reading  2.9 % of reading 2.9 % of reading  2.9 % of reading 0.008 % of reading	IEC/EN 61000-4-5 IEC/EN 61000-4-8 IEC/EN 61000-4-9 IEC/EN 61000-4-10 IEC/EN 61000-4-11 IEC/EN 61000-4-12 Tektronix TDS784C Oscilloscope, Tektronix P5210 High Voltage Probe Pearson 411 Current Probe
ISN – Relative Amplitude (dB) Insertion Loss  Impedance Magnitude (Ω)  Impedance Phase (°)	(0 to 120) dB 150 kHz to 80 MHz (0 to 1) kΩ 150 kHz to 80 MHz (-180 to 180)° 150 kHz to 80 MHz	0.59 dB  2 % of reading  1.8°	CISPR 22, CISPR 32, CISPR 16-1-2 HP 8751A Network Analyzer & HP 87512A Transmission/Reflection Test Set, 85032B Calibration Kit



ANSI National Accreditation Board

Electrical – RF/Microwave

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
CDN's & Adapters – (50 to 150) Ω Relative Amplitude (dB) Adapter Insertion Loss Relative Amplitude (dB) Coupling Factor Impedance Magnitude (Ω) Relative Amplitude (dB) Voltage Division Factor	(0 to 120) dB 10 kHz to 230 MHz (0 to 120) dB 10 kHz to 230 MHz (0 to 1) kΩ 10 kHz to 230 MHz (0 to 120) dB 10 kHz to 230 MHz	1.3 dB 1.3 dB 2 % reading 0.59 dB	IEC/EN 61000-4-6 CISPR 16-1-2 HP 8751A Network Analyzer & HP 87512A Transmission/Reflection Test Set, 85032B Calibration Kit
LISN – Relative Amplitude (dB) Insertion Loss Impedance Magnitude (Ω) Impedance – Phase (°) Relative Amplitude (dB) Isolation	(0 to 120) dB 9 kHz to 100 MHz (100 to 300) MHz (300 to 500) MHz (0 to 1) kΩ Ω 9 kHz to 500 MHz (-180 to 180)° 9 kHz to 500 MHz (0 to 120) dB 9 kHz to 100 MHz (100 to 300) MHz (300 to 500) MHz	0.59 dB 1.3 dB 1.8 dB 2 % of reading 1.8° 0.59 dB 1.3 dB 1.8 dB	ANSI C63.4, CISPR 25, CISPR 16-1-2 HP 8751A Network Analyzer & HP 87512A Transmission/Reflection Test Set, 85032B Calibration Kit
Current Probes and Bulk Current Injection Probes Insertion Loss and Transfer Impedance Relative Amplitude (dB)	(0 to 120) dB 20 Hz to 100 MHz (100 to 300) MHz (300 to 500) MHz	0.59 dB 1.3 dB 1.8 dB	CISPR 16-1-2 IEC/EN 61000-4-6 HP 8751A Network Analyzer & HP 87512A Transmission/Reflection Test Set, 85032B Calibration Kit
Source Errors for CISPR Bands A, B, C and D for Impulse Spectral Amplitude Absolute Amplitude Pulse Response & Relative Ratio	Band A (0 to 120) dB (10 to 150) kHz Band B (0 to 120) dB (0.15 to 30) MHz	0.82 dB 0.82 dB	IGUU 2918 Pulse Generator CISPR 16-1-1 Signal Generator, Agilent 33250A Waveform Generator
Source Errors for CISPR Bands A, B, C and D for Impulse Spectral Amplitude Absolute Amplitude Pulse Response & Relative Ratio	Band C and D (0 to 120) dB (30 to 1 000) MHz (0 to 120) dB Band E (1 to 18) GHz	1.5 dB 0.77 dB	IGUU 2918 Pulse Generator CISPR 16-1-1 Signal Generator, Agilent 33250A Waveform Generator



ANSI National Accreditation Board

**Electrical – RF/Microwave**

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Source Errors for Sinewave output for CISPR Checks (at 60 dBμV) Absolute Amplitude	60 dBμV Band A through D Band E	0.58 dB 0.58 dB	CISPR 16-1-1 Agilent E8257D Signal Generator, Agilent 33250A Waveform Generator
QuasiPeak to Peak & Average Detector Response Relative Amplitude Ratio	(-60 to 60) dB Band A through D	1.3 dB	IGUU 2918 Pulse Generator
Return Loss (VSWR) 30 kHz to 6 GHz	(0 to 80) dB	2.4 dB	Agilent 8753ES Network Analyzers
6 GHz to 18 GHz	(0 to 50) dB	0.32 dB	Wiltron 87A50 VSWR Bridge
ESD Simulators			
Contact Voltage (Positive and Negative)	(0 to 20) kV (20 to 30) kV	1.2 % of reading + 2V 1.2 % of reading + 20V	ESVM
Risetime	(0 to 5) ns	31.5 ps	Agilent 54855A
Peak Current	(0 to 60) A	4.2 % of reading	Oscilloscope, IEC ESD
30 ns Current	(0 to 60) A	4.2 % of reading	Target
60 ns Current	(0 to 60) A	4.2 % of reading	
Air Discharge Voltage (Positive and Negative)	(1 to 20) kV (20 to 30) kV	1.2 % of reading + 2V 1.2 % of reading + 20V	IEC 61000-4-2 ISO 10605 MIL STD 331
Risetime	(0 to 5) ns	31.5 ps	
RC Time Constant (at +/- 15 kV)	600 ns +/- 130 ns (for 330 pF probe) 300 ns +/- 60 ns (for 150 pF probe)	7.2 ns 6 ns	

**Length – Dimensional Metrology**

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Gage Blocks <sup>2</sup>	Up to 13 in	(4.2 + 2.1L) μin	Master gage blocks P&W Universal Measuring Machine

**Length – Dimensional Metrology**

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Micrometers <sup>1,2</sup>	Up to 40 in	$(28 + 4.6L) \mu\text{in}$	Gage Blocks
Bore Micrometers <sup>2</sup> 2 point 3 point	Up to 12 in Up to 3 in	$(56 + 3.3L) \mu\text{in}$ $(57 + 2.8L) \mu\text{in}$	Master gage blocks, P&W Labmaster Universal, Master Ring
Calipers <sup>1,2</sup>	Up to 40 in	$(280 + 1.8L) \mu\text{in}$	Gage Blocks
Dial Indicators <sup>1,2</sup> Resolution $\geq 50\mu\text{in}$ Resolution $< 50\mu\text{in}$	Up to 10 in Up to 0.1 in	$(26 + 3.8L) \mu\text{in}$ 8.1 $\mu\text{in}$	Gage Blocks
Height Gages <sup>1,2</sup>	Up to 40 in	$(95 + 3L) \mu\text{in}$	Gage Blocks
Rulers <sup>1</sup>	Up to 46 in	0.009 1 in	Gage Blocks
Tape Measures <sup>2</sup>	Up to 100 ft	$(0.000 26F + 0.025) \text{ in}$	Standard rule
Protractors <sup>1</sup>	(0 to 360) <sup>o</sup>	0.013 <sup>o</sup>	Angle Blocks
Rotary Encoders - Angle <sup>1</sup>	(0 to 360) <sup>o</sup>	24 arc sec	Rotary Encoder
Feeler Gage	Up to 1 in	31 $\mu\text{in}$	Pratt & Whitney Supermicrometer C
Cylindrical Gages <sup>2</sup>  Plain Pins, Plugs  Plain Rings	  (0 to 13) in  (0.04 to 14) in	  $(10 + 3.1D) \mu\text{in}$  $(11 + 3.2D) \mu\text{in}$	  Master gage blocks P&W Universal Measuring Machine
Solid Thread Rings Pitch Diameter	Up to 12 in	98 $\mu\text{in}$	Pratt & Whitney Labmaster Measuring Machine
Thread Plugs Major Diameter Pitch Diameter	Up to 10 in Up to 10 in	44 $\mu\text{in}$ 79 $\mu\text{in}$	Thread Wires Gage Blocks Pratt & Whitney Supermicrometer C
Thread Rings <sup>2</sup> Pitch Diameter	Up to 1 in	$(350 + 47D) \mu\text{in}$	Thread Setting Plugs Tactile Fit
Thread Wires	Up to 0.5 in	11 $\mu\text{in}$	Master gage blocks, P&W universal measuring machine



ANSI National Accreditation Board

**Length – Dimensional Metrology**

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Surface Plates <sup>1</sup> Overall Flatness Local Area Flatness	(18 x 18) in to (6 x 6) ft (-0.001 to 0.001) in	86 μin 68 μin	Planekator Repeat-o-meter
Optical Comparators <sup>1</sup> – Angle Linearity	(0 to 360) ° Up to 20 in (20 to 40) in	0.016 ° 320 μin 630 μin	Gage blocks, Angle blocks,
Magnification	10x to 100x	430 μin	SI Industries glass scales

**Mass and Mass Related**

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Pressure <sup>1</sup>	(-150 to 150) inH <sub>2</sub> O Up to 30 psia Up to 100 psig (-15 to 300) psig Up to 1 000 psig Up to 5 000 psig Up to 10 000 psig Up to 15 000 psig	0.07 inH <sub>2</sub> O 0.035 psi 0.025 psi 0.077 psi 0.25 psi 1.5 psi 3.1 psi 3.6 psi	Pressure Gages ADT681-02-DP150 ADT681-10AP30 ADT681-02-GP100 ADT681-02-CP300 ADT681-02-GP1K ADT681-02-GP5K ADT681-02-GP10K ADT681-05-GP15K
Pressure <sup>1</sup>	(5 to 15 000) psi	0.04 % of reading	Ametek type T Deadweight Tester
Pipettes <sup>1,2</sup>	(10 to 100) μL (100 to 1000) μL (1 to 10) mL	(0.58 + 0.004V) μL (0.6 + 0.001V) μL (2.6 + 0.0012V) μL	Mass Balance
Torque Tools <sup>1</sup>	(10 to 100) ozf-in 4 lbf-in to 1 000 lbf-ft	0.59 % of reading 0.3 % of reading	Torque Testers CDI 1001-O-DDT CDI 5000 ST
Torque Analyzers <sup>1</sup>	Up to 100 ozf-in (4 to 150) lbf-in (12.5 to 1 000) lbf-ft	0.1 % of reading 0.064 % of reading 0.036 % of reading	Torque Arms and Class F Weights
Force <sup>1</sup> Tension and Compression	(0.5 to 500) lbf	0.02 % of reading	Class F Weights
Force Tension and Compression	(200 to 25 000) lbf	0.017 % of reading	Morehouse Load Cells
Scales and Balances <sup>1,2,4</sup>	1 mg to 420 g	(0.021 + 0.003 4X) mg	Class 1 Weights

**Mass and Mass Related**

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Scales and Balances <sup>1,2,4</sup>	500 g to 41 kg	(0.015 + 0.024X) mg	Class 4 Weights
Scales and Balances <sup>1,2,4</sup>	Up to 100 kg Up to 1 000 lb	(0.001 7 + 0.000 12X) g (0.002 1 + 0.000 12W) lb	Class F Weights
Mass - Fixed Points Metric	(1, 2, 3, 5, 10) mg (20, 30, 50, 100) mg (200, 300, 500) mg (1, 2, 3, 5) g (10, 20, 30) g 50 g 100 g 200 g	0.013 mg 0.013 mg 0.013 mg 0.049 mg 0.068 mg 0.072 mg 0.23 mg 0.38 mg	Comparison to ASTM E617 Class 1 weights
Mass - Fixed Points Metric	500 g (1, 2, 3) kg (5, 10, 20) kg	58 mg	Comparison to ASTM E617 Class 4 weights
Rockwell Hardness Testers <sup>1</sup>	(< 60) HRBW (≥ 60 to < 80) HRBW ≥ 80 HRBW  < 35 HRC (≥ 35 to < 60) HRC ≥ 60 HRC  < 70 HRA (≥ 70 and < 80) HRA ≥ 80 HRA	3 HRBW 3 HRBW 1.3 HRBW  1.3 HRC 1.2 HRC 0.7 HRC  1.3 HRA 1.2 HRA 0.7 HRA	Indirect verification per ASTM E18

**Thermodynamic**

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Relative Humidity Generate	(10 to 95) %RH	0.5 %RH	Thunder Scientific 2500 Humidity Chamber
Relative Humidity Measure <sup>1</sup>	(0 to 90) %RH (90 to 100) %RH	1.2 %RH 2 %RH	Vaisala MI70/HMP76 Humidity Indicator and Probe
Temperature Measuring Equipment <sup>1</sup>	(-40 to 660) °C	0.18 °C	Dry Well Calibrators

**Thermodynamic**

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Temperature Measure <sup>1</sup>	(-200 to 400) °C (400 to 660) °C	0.043 °C 0.13 °C	Secondary PRT
Infrared (IR) Thermometers <sup>1</sup>	(20 to 100) °C (100 to 300) °C (300 to 420) °C (420 to 500) °C	1.5 °C 4.3 °C 6 °C 7.7 °C	Fluke 9132 IR Calibrator $\epsilon = 0.95, \lambda = (8 \text{ to } 14) \mu\text{m}$

**Time and Frequency**

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency – Generate <sup>1</sup>	10 MHz	$1 \times 10^{-12}$ Hz/Hz	HP 58503A GPS Receiver
Frequency – Generate <sup>1</sup>	(1 to 10) Hz (10 to 100) Hz (100 to 1 000) Hz (1 to 10) kHz (10 to 100) kHz (0.1 to 1) MHz (1 to 10) MHz	$1 \times 10^{-12}$ Hz/Hz + 0.57 $\mu$ Hz $1 \times 10^{-12}$ Hz/Hz + 5.7 $\mu$ Hz $1 \times 10^{-12}$ Hz/Hz + 57 $\mu$ Hz $1 \times 10^{-12}$ Hz/Hz + 0.57 mHz $1 \times 10^{-12}$ Hz/Hz + 5.7 mHz $1 \times 10^{-12}$ Hz/Hz + 57 mHz $1 \times 10^{-12}$ Hz/Hz + 0.57 Hz	Agilent 33250A Function Generator / HP 58503A GPS Receiver
Frequency – Generate <sup>1</sup>	(10 to 40 000) MHz	$1 \times 10^{-12}$ Hz/Hz + 0.57 mHz	Agilent E8257D Opt 540 Signal Generator / HP 58503A GPS Receiver
Time – Generate	1 pps	$1 \times 10^{-12}$ s/s + 750 ps	HP 58503A GPS Receiver
Frequency – Measure <sup>1</sup>	(1 to 10) Hz (10 to 100) Hz (100 to 1 000) Hz (1 to 10) kHz (10 to 100) kHz (100 to 200) kHz (0.2 to 3 000) MHz	$4.2 \times 10^{-9}$ Hz/Hz $1.5 \times 10^{-9}$ Hz/Hz $0.6 \times 10^{-9}$ Hz/Hz $0.33 \times 10^{-9}$ Hz/Hz $0.24 \times 10^{-9}$ Hz/Hz $0.21 \times 10^{-9}$ Hz/Hz $0.21 \times 10^{-9}$ Hz/Hz	Agilent 53131A Opt 030 Frequency Counter / HP 58503A GPS Receiver
Frequency – Measure <sup>1</sup>	(10 to 50 000) MHz	$1 \times 10^{-12}$ Hz/Hz + 0.1 Hz	Agilent E4448A Spectrum Analyzer / HP 58503A GPS Receiver
Timer, Stopwatch <sup>1</sup>	10 s to 24 hr	34 ms	Totalize method with counter
Tachometers – RPM <sup>1</sup>	Up to 100 000 RPM	0.001 % of reading + 0.6R	HP 33250A Signal Generator & LED

## DIMENSIONAL MEASUREMENT

1 Dimensional

Colorado Springs, CO

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Length <sup>1,2</sup>	(0 to 480) in	$(0.8 + 2.6L) \mu\text{in}$	Laser measurement system

[Return to Site listing](#) (top)    [Go to Notes](#) (bottom)





ANSI National Accreditation Board

## Services performed at satellite laboratory

### Technical Maintenance, Inc.

2200 Gateway Centre Blvd. #208

Morrisville, NC 27560

Taylor Floyd (Branch Manager) Phone: (919) 234-1291

Scott Chamberlain (Quality Manager) Phone: (321) 242-0890

## CALIBRATION

### Chemical Quantities

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
pH meters <sup>1,5</sup>	4 pH 7 pH 10 pH	0.025 pH 0.027 pH 0.064 pH	pH buffer solutions
Conductivity Meters <sup>1,5</sup>	10 µS/cm 500 µS/cm 1 000 µS/cm	0.17 µS/cm 2.2 µS/cm 3.6 µS/cm	Conductivity solutions

### Electrical – DC/Low Frequency

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Current – Generate <sup>1</sup>	(1 to 1.2) nA (1.2 to 12) nA (12 to 120) nA (0.12 to 1.2) µA (1.2 to 10) µA	92 µA/A + 0.007 nA 92 µA/A + 0.007 nA 92 µA/A + 0.01 nA 36 µA/A + 0.1 nA 13 µA/A + 1 nA	Fluke 5730A Multifunction Calibrator, Fluke 5560A Multiproduct Calibrator
DC Current – Generate <sup>1</sup>	(10 to 220) µA 220 µA to 2.2 mA (2.2 to 22) mA (22 to 100) mA (100 to 220) mA (0.22 to 1) A (1 to 2.2) A	38 µA/A + 5 nA 30 µA/A + 7 nA 30 µA/A + 44 nA 38 µA/A + 0.7 µA 45 µA/A + 0.7 µA 68 µA/A + 12 µA 105 µA/A + 12 µA	Fluke 5730A Multifunction Calibrator
	(2.2 to 3.1) A (3.1 to 12) A (12 to 30) A	228 µA/A + 115 µA 228 µA/A + 191 µA 761 µA/A + 392 µA	Fluke 5560A Multiproduct Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Current – Generate <sup>1</sup> Clamp Meters	(3.1 to 300.2) A	0.65 % of output + 0.014 A	Fluke 5560A Multiproduct Calibrator 55xxA x1/2/10 Coil
DC Current – Measure <sup>1</sup>	(1 to 10) nA (10 to 100) nA (0.1 to 1) $\mu$ A (1 to 10) $\mu$ A	35 $\mu$ A/A + 0.1 pA 12 $\mu$ A/A + 1 pA 8.3 $\mu$ A/A + 0.01 nA 6.9 $\mu$ A/A + 0.1 nA	Fluke 5730A Multifunction Calibrator Agilent 3458A Option 002 Multimeter
	(0.1 to 20.2) $\mu$ A (20.2 to 202) $\mu$ A (0.202 to 2.02) mA (2.02 to 20.2) mA (20.2 to 202) mA (0.202 to 2.02) A	30.4 $\mu$ A/A + 0.4 nA 11.4 $\mu$ A/A + 0.5 nA 11.4 $\mu$ A/A + 5 nA 15.2 $\mu$ A/A + 0.05 $\mu$ A 58 $\mu$ A/A + 1.5 $\mu$ A 150 $\mu$ A/A + 0.2mA	Fluke 8588A Multimeter
	(1 to 3) A (3 to 10) A	0.08 % of Reading + 0.46 mA 0.11 % of Reading + 0.61 mA	Fluke 8845A Multimeter
	(10 to 250) A	0.25 % of Reading + 0.02A	Fluke 8845A Multimeter, Empro Current Shunt
	DC Voltage – Generate <sup>1</sup>	(0 to 220) mV 220 mV to 2.2 V (2.2 to 11) V (11 to 22) V (22 to 220) V (220 to 1 100) V	6.8 $\mu$ V/V + 0.8 $\mu$ V 4.6 $\mu$ V/V + 0.9 $\mu$ V 3 $\mu$ V/V + 2.5 $\mu$ V 3 $\mu$ V/V + 3.9 $\mu$ V 4.6 $\mu$ V/V + 38 $\mu$ V 6.1 $\mu$ V/V + 385 $\mu$ V
DC Voltage – Measure <sup>1</sup>	(0 to 220) mV (0.202 to 2.02) V (2.02 to 20.2) V (20.2 to 202) V (202 to 1 000) V	8.1 $\mu$ V/V + 0.7 $\mu$ V 5.2 $\mu$ V/V + 0.8 $\mu$ V 5.2 $\mu$ V/V + 0.9 $\mu$ V 8.3 $\mu$ V/V + 34 $\mu$ V 8.4 $\mu$ V/V + 1.3 mV	Fluke 8558A Multimeter
DC Voltage – Measure <sup>1</sup>	(0 to 100) mV (0.1 to 1) V (1 to 10) V (10 to 100) V (100 to 1 000) V	3.3 $\mu$ V/V + 1 $\mu$ V 2.6 $\mu$ V/V + 1 $\mu$ V 2.6 $\mu$ V/V + 1.5 $\mu$ V 3.9 $\mu$ V/V + 20 $\mu$ V 3.9 $\mu$ V/V + 66 $\mu$ V + 12 $\mu$ V/V x (Vin/1000) ^2	Agilent 3458A Multimeter
DC Voltage – Measure <sup>1</sup>	(1 to 30) kV	0.1 % of reading	Ross VD30 High Voltage Divider, HP 34401A Multimeter



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Generate <sup>1</sup>	(0.22 to 2.2) mV		Fluke 5730A Multifunction Calibrator
	(10 to 20) Hz	228 $\mu\text{V/V} + 3.9 \mu\text{V}$	
	(20 to 40) Hz	88 $\mu\text{V/V} + 3.9 \mu\text{V}$	
	40 Hz to 20 kHz	76 $\mu\text{V/V} + 3.9 \mu\text{V}$	
	(20 to 50) kHz	190 $\mu\text{V/V} + 3.9 \mu\text{V}$	
	(50 to 100) kHz	457 $\mu\text{V/V} + 4.6 \mu\text{V}$	
	(100 to 300) kHz	989 $\mu\text{V/V} + 9.2 \mu\text{V}$	
	(300 to 500) kHz	1.3 mV/V + 19 $\mu\text{V}$	
	500 kHz to 1 MHz	2.6 mV/V + 19 $\mu\text{V}$	
	(2.2 to 22) mV		
	(10 to 20) Hz	228 $\mu\text{V/V} + 3.9 \mu\text{V}$	
	(20 to 40) Hz	88 $\mu\text{V/V} + 3.9 \mu\text{V}$	
	40 Hz to 20 kHz	76 $\mu\text{V/V} + 3.9 \mu\text{V}$	
	(20 to 50) kHz	190 $\mu\text{V/V} + 3.9 \mu\text{V}$	
	(50 to 100) kHz	457 $\mu\text{V/V} + 4.6 \mu\text{V}$	
	(100 to 300) kHz	989 $\mu\text{V/V} + 9.2 \mu\text{V}$	
	(300 to 500) kHz	1.3 mV/V + 19 $\mu\text{V}$	
	500 kHz to 1 MHz	2.6 mV/V + 19 $\mu\text{V}$	
	(22 to 220) mV		
	(10 to 20) Hz	228 $\mu\text{V/V} + 11.4 \mu\text{V}$	
	(20 to 40) Hz	88 $\mu\text{V/V} + 6.1 \mu\text{V}$	
	40 Hz to 20 kHz	53 $\mu\text{V/V} + 6.1 \mu\text{V}$	
	(20 to 50) kHz	114 $\mu\text{V/V} + 6.1 \mu\text{V}$	
	(50 to 100) kHz	304 $\mu\text{V/V} + 15.2 \mu\text{V}$	
	(100 to 300) kHz	609 $\mu\text{V/V} + 19 \mu\text{V}$	
	(300 to 500) kHz	1.3 mV /V + 23 $\mu\text{V}$	
	500 kHz to 1 MHz	2.5 mV /V + 46 $\mu\text{V}$	
220 mV to 2.2 V			
(10 to 20) Hz	228 $\mu\text{V/V} + 38 \mu\text{V}$		
(20 to 40) Hz	84 $\mu\text{V/V} + 15 \mu\text{V}$		
40 Hz to 20 kHz	37 $\mu\text{V/V} + 8 \mu\text{V}$		
(20 to 50) kHz	61 $\mu\text{V/V} + 9 \mu\text{V}$		
(50 to 100) kHz	76 $\mu\text{V/V} + 30 \mu\text{V}$		
(100 to 300) kHz	304 $\mu\text{V/V} + 76 \mu\text{V}$		
(300 to 500) kHz	913 $\mu\text{V} /V + 190 \mu\text{V}$		
500 kHz to 1 MHz	1.5 mV/V + 304 $\mu\text{V}$		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Generate <sup>1</sup>	(2.2 to 22) V		Fluke 5730A Multifunction Calibrator
	(10 to 20) Hz	228 $\mu$ V/V + 380 $\mu$ V	
	(20 to 40) Hz	84 $\mu$ V/V + 152 $\mu$ V	
	40 Hz to 20 kHz	37 $\mu$ V/V + 54 $\mu$ V	
	(20 to 50) kHz	61 $\mu$ V/V + 91 $\mu$ V	
	(50 to 100) kHz	76 $\mu$ V/V + 190 $\mu$ V	
	(100 to 300) kHz	228 $\mu$ V/V + 609 $\mu$ V	
	(300 to 500) kHz	913 $\mu$ V/V + 1.9 mV	
	500 kHz to 1 MHz	1.4 mV/V + 3 mV	
	(22 to 220) V		
	(10 to 20) Hz	228 $\mu$ V/V + 3.8 mV	
	(20 to 40) Hz	84 $\mu$ V/V + 1.5 mV	
	40 Hz to 20 kHz	49 $\mu$ V/V + 0.6 mV	
	(20 to 50) kHz	76 $\mu$ V/V + 0.9 mV	
(50 to 100) kHz	137 $\mu$ V/V + 2.3 mV		
AC Voltage – Generate <sup>1</sup>	(220 to 750) V		Fluke 5730A Multifunction Calibrator, 5725A Amplifier
	40 Hz to 1 kHz	68 $\mu$ V/V + 3 mV	
	(1 to 20) kHz	126 $\mu$ V/V + 5 mV	
	(20 to 50) kHz	457 $\mu$ V/V + 8 mV	
	(50 to 100) kHz	1.8 mV/V + 34 mV	
	(750 to 1 000) V		
	40 Hz to 1 kHz	68 $\mu$ V/V + 3 mV	
	(1 to 20) kHz	126 $\mu$ V/V + 5 mV	
	(20 to 30) kHz	457 $\mu$ V/V + 8 mV	
	AC Voltage – Generate <sup>1</sup> Wideband Absolute	(0.3 to 1.1) mV	
(10 to 30 Hz)		0.65 % of output + 1.5 $\mu$ V	
30 Hz to 500 kHz		0.61 % of output + 1.5 $\mu$ V	
(0.5 to 1.2) MHz		0.63 % of output + 3.8 $\mu$ V	
(1.2 to 2) MHz		0.63 % of output + 3.8 $\mu$ V	
(2 to 12) MHz		0.68 % of output + 3.8 $\mu$ V	
(12 to 20) MHz		0.76 % of output + 3.8 $\mu$ V	
(20 to 30) MHz		1.3 % of output + 12.9 $\mu$ V	
(1.1 to 3.3) mV			
(10 to 30 Hz)		0.58 % of output + 2.3 $\mu$ V	
30 Hz to 500 kHz		0.53 % of output + 2.3 $\mu$ V	
(0.5 to 1.2) MHz		0.54 % of output + 4.6 $\mu$ V	
(1.2 to 2) MHz		0.54 % of output + 4.6 $\mu$ V	
(2 to 12) MHz		0.58 % of output + 4.6 $\mu$ V	
(12 to 20) MHz		0.65 % of output + 4.6 $\mu$ V	
(20 to 30) MHz		1.3 % of output + 4.6 $\mu$ V	



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Generate <sup>1</sup> Wideband Absolute	(3.3 to 11) mV		Fluke 5730A Option 003 Multifunction Calibrator
	(10 to 30 Hz)	0.58 % of output + 6.1 $\mu$ V	
	30 Hz to 500 kHz	0.53 % of output + 6.1 $\mu$ V	
	(0.5 to 1.2) MHz	0.54 % of output + 8.4 $\mu$ V	
	(1.2 to 2) MHz	0.54 % of output + 8.4 $\mu$ V	
	(2 to 12) MHz	0.55 % of output + 8.4 $\mu$ V	
	(12 to 20) MHz	0.61 % of output + 8.4 $\mu$ V	
	(20 to 30) MHz	0.93 % of output + 8.4 $\mu$ V	
	(11 to 33) mV		
	(10 to 30 Hz)	0.52 % of output + 12 $\mu$ V	
	30 Hz to 500 kHz	0.46 % of output + 12 $\mu$ V	
	(0.5 to 1.2) MHz	0.47 % of output + 14 $\mu$ V	
	(1.2 to 2) MHz	0.47 % of output + 14 $\mu$ V	
	(2 to 12) MHz	0.49 % of output + 14 $\mu$ V	
	(12 to 20) MHz	0.55 % of output + 14 $\mu$ V	
	(20 to 30) MHz	0.89 % of output + 14 $\mu$ V	
	(33 to 110) mV		
	(10 to 30 Hz)	0.52 % of output + 30 $\mu$ V	
	30 Hz to 500 kHz	0.46 % of output + 30 $\mu$ V	
	(0.5 to 1.2) MHz	0.47 % of output + 33 $\mu$ V	
	(1.2 to 2) MHz	0.47 % of output + 33 $\mu$ V	
	(2 to 12) MHz	0.49 % of output + 33 $\mu$ V	
	(12 to 20) MHz	0.55 % of output + 33 $\mu$ V	
	(20 to 30) MHz	0.89 % of output + 33 $\mu$ V	
	(110 to 330) mV		
	(10 to 30 Hz)	0.45 % of output + 0.1 mV	
	30 Hz to 500 kHz	0.38 % of output + 0.1 mV	
	(0.5 to 1.2) MHz	0.4 % of output + 0.1 mV	
(1.2 to 2) MHz	0.4 % of output + 0.1 mV		
(2 to 12) MHz	0.42 % of output + 0.1 mV		
(12 to 20) MHz	0.49 % of output + 0.1 mV		
(20 to 30) MHz	0.85 % of output + 0.1 mV		
(0.33 to 1.1) V			
(10 to 30 Hz)	0.45 % of output + 0.3 mV		
30 Hz to 500 kHz	0.38 % of output + 0.3 mV		
(0.5 to 1.2) MHz	0.4 % of output + 0.3 mV		
(1.2 to 2) MHz	0.4 % of output + 0.3 mV		
(2 to 12) MHz	0.42 % of output + 0.3 mV		
(12 to 20) MHz	0.49 % of output + 0.3 mV		
(20 to 30) MHz	0.85 % of output + 0.3 mV		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Generate <sup>1</sup> Wideband Absolute	(1.1 to 3.5) V (10 to 30 Hz) 30 Hz to 500 kHz (0.5 to 1.2) MHz (1.2 to 2) MHz (2 to 12) MHz (12 to 20) MHz (20 to 30) MHz	0.39 % of output + 0.4 mV 0.3 % of output + 0.4 mV 0.32 % of output + 0.4 mV 0.32 % of output + 0.4 mV 0.35 % of output + 0.4 mV 0.44 % of output + 0.4 mV 0.82 % of output + 0.4 mV	Fluke 5730A Option 003 Multifunction Calibrator
AC Voltage – Measure <sup>1</sup>	(1.2 to 12.12) mV 1 to 2000 Hz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (100 to 300) kHz (0.3 to 1 MHz) (12.12 to 121.2) mV 1 to 2000 Hz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (100 to 300) kHz (0.3 to 1 MHz) (1 to 2) MHz (2 to 4) MHz (4 to 8) MHz (8 to 10) MHz (12.12 to 121.2) mV 1 to 2000 Hz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (100 to 300) kHz (0.3 to 1 MHz) (1 to 2) MHz (2 to 4) MHz (4 to 8) MHz (8 to 10) MHz	0.06 % of reading + 2.0 μV 0.045 % of reading + 2.0 μV 0.046 % of reading + 2.0 μV 0.42 % of reading + 2.0 μV 1.58 % of reading + 5.9 μV 2.47 % of reading + 5.9 μV 0.011 % of reading + 1 μV 0.015 % of reading + 1 μV 0.025 % of reading + 2 μV 0.057 % of reading + 20 μV 0.34 % of reading + 50 μV 1.38 % of reading + 0.2mV 1.6 % of reading + 0.7mV 4.6 % of reading + 1.2 mV 9.2 % of reading + 1.2 mV 18 % of reading + 1.2 mV 0.011 % of reading + 1 μV 0.015 % of reading + 1 μV 0.025 % of reading + 2 μV 0.057 % of reading + 20 μV 0.34 % of reading + 49.5 μV 1.38 % of reading + 0.2mV 1.6 % of reading + 0.7mV 4.6 % of reading + 1.2 mV 9.2 % of reading + 1.2 mV 18 % of reading + 1.2 mV	Fluke 8588A Multimeter



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure <sup>1</sup>	(0.1212 to 1.212)V (1 to 2 000) Hz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (100 to 300) kHz (0.3 to 1 MHz) (1 to 2) MHz (2 to 4) MHz (4 to 8) MHz (8 to 10) MHz (1.212 to 12.12)V 1 to 2000 Hz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (100 to 300) kHz (0.3 to 1 MHz) (1 to 2) MHz (2 to 4) MHz (4 to 8) MHz (8 to 10) MHz (12.12 to 121.2)V (1 to 2 000) Hz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (100 to 200) kHz (121.2 to 1 050)V (1 to 1000 Hz (2 to 10) kHz (10 to 30) kHz	0.01 % of reading + 0.01mV 0.014 % of reading + 0.01 mV 0.025 % of reading + 0.02 mV 0.057 % of reading + 0.20 mV 0.26 % of reading + 0.5 mV 1.4 % of reading + 2mV 1.6 % of reading + 7mV 4.6 % of reading + 12 mV 9.1 % of reading + 12 mV 18 % of reading + 12 mV 0.01 % of reading + 0.1mV 0.014 % of reading + 0.1 mV 0.025 % of reading + 0.2 mV 0.057 % of reading + 2 mV 0.26 % of reading + 5 mV 1.4 % of reading + 20mV 1.6 % of reading + 70mV 4.6 % of reading + 120 mV 9.1 % of reading + 120 mV 18 % of reading + 120 mV 0.011 % of reading + 1mV 0.015 % of reading + 1 mV 0.025 % of reading + 2 mV 0.063 % of reading + 20 mV 0.43 % of reading + 100 mV 0.016 % of reading + 30mV 0.018 % of reading + 30 mV 0.032 % of reading + 30 mV	Fluke 8588A Multimeter
AC Voltage – Measure <sup>1</sup>	(1 to 10) mV (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz 100 kHz to 1 MHz (1 to 4) MHz (4 to 8) MHz	0.02 % of reading + 2 μV 0.013 % of reading + 0.7 μV 0.02 % of reading + 0.7 μV 0.065 % of reading + 0.7 μV 0.33 % of reading + 0.7 μV 0.78 % of reading + 3.3 μV 4.6 % of reading + 4.6 μV 13 % of reading + 5.3 μV	Agilent 3458A Multimeter



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure <sup>1</sup>	10 mV to 100 mV	0.005 % of reading + 2.6 μV	Agilent 3458A Multimeter
	(1 to 40) Hz	0.005 % of reading + 1.3 μV	
	40 Hz to 1 kHz	0.009 % of reading + 1.3 μV	
	(1 to 20) kHz	0.02 % of reading + 1.3 μV	
	(20 to 50) kHz	0.052 % of reading + 1.3 μV	
	(50 to 100) kHz	0.20 % of reading + 6.5 μV	
	(100 to 300) kHz	0.70 % of reading + 6.5 μV	
	300 kHz to 1 MHz	0.98 % of reading + 46 μV	
	(1 to 2) MHz	2.6 % of reading + 46 μV	
	(2 to 4) MHz	2.6 % of reading + 52 μV	
	(4 to 8) MHz	9.8 % of reading + 65 μV	
	(8 to 10) MHz		
	(0.10 to 1) V	0.005 % of reading + 26 μV	
	(1 to 40) Hz	0.005 % of reading + 13 μV	
	40 Hz to 1 kHz	0.009 % of reading + 13 μV	
	(1 to 20) kHz	0.02 % of reading + 13 μV	
	(20 to 50) kHz	0.052 % of reading + 13 μV	
	(50 to 100) kHz	0.2 % of reading + 65 μV	
	(100 to 300) kHz	0.65 % of reading + 65 μV	
	300 kHz to 1 MHz	0.98 % of reading + 0.5 mV	
	(1 to 2) MHz	2.6 % of reading + 0.5 mV	
	(2 to 4) MHz	2.6 % of reading + 0.5 mV	
	(4 to 8) MHz	9.8 % of reading + 0.7 mV	
	(8 to 10) MHz		
	(1 to 10) V	0.005 % of reading + 0.3 mV	
	(1 to 40) Hz	0.005 % of reading + 0.1 mV	
	40 Hz to 1 kHz	0.009 % of reading + 0.1 mV	
	(1 to 20) kHz	0.02 % of reading + 0.1 mV	
(20 to 50) kHz	0.052 % of reading + 0.1 mV		
(50 to 100) kHz	0.2 % of reading + 0.7 mV		
(100 to 300) kHz	0.65 % of reading + 0.7 mV		
300 kHz to 1 MHz	0.98 % of reading + 4.6 mV		
(1 to 2) MHz	2.6 % of reading + 4.6 mV		
(2 to 4) MHz	2.6 % of reading + 5.2 mV		
(4 to 8) MHz	9.8 % of reading + 6.5 mV		
(8 to 10) MHz			



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure <sup>1</sup>	(10 to 100) V (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (100 to 700) V (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz	0.013 % of reading + 2.6 mV 0.013 % of reading + 1.3 mV 0.013 % of reading + 1.3 mV 0.023 % of reading + 1.3 mV 0.08 % of reading + 1.3 mV 0.26 % of reading + 6.5 mV 0.98 % of reading + 6.5 mV 0.026 % of reading + 26 mV 0.026 % of reading + 13 mV 0.039 % of reading + 13 mV 0.078 % of reading + 13 mV 0.2 % of reading + 13 mV	Agilent 3458A Multimeter
AC Voltage – Measure <sup>1</sup>	(1 to 5) kV 60 Hz (5 to 21) kV 60 Hz	0.5 % of reading 0.5 % of reading	Ross VD30 Voltage Divider, HP 34401A Multimeter
AC Current – Generate <sup>1</sup>	(9 to 220) $\mu$ A (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (0.22 to 2.2) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (2.2 to 22) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	228 $\mu$ A/A + 15 nA 152 $\mu$ A/A + 10 nA 91 $\mu$ A/A + 8 nA 266 $\mu$ A/A + 12 nA 989 $\mu$ A/A + 61 nA 228 $\mu$ A/A + 39 nA 152 $\mu$ A/A + 31 nA 91 $\mu$ A/A + 31 nA 183 $\mu$ A/A + 99 nA 989 $\mu$ A/A + 609 nA 228 $\mu$ A/A + 385 nA 152 $\mu$ A/A + 310 nA 91 $\mu$ A/A + 310 nA 183 $\mu$ A/A + 536 nA 989 $\mu$ A/A + 4566 nA	Fluke 5730A Multifunction Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Generate <sup>1</sup>	(22 to 220) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	228 $\mu$ A/A + 4 $\mu$ A 152 $\mu$ A/A + 3 $\mu$ A 91 $\mu$ A/A + 2 $\mu$ A 183 $\mu$ A/A + 3 $\mu$ A 989 $\mu$ A/A + 9 $\mu$ A	Fluke 5730A Multifunction Calibrator
AC Current – Generate <sup>1</sup>	(0.22 to 1.2) A (3 to 45) Hz (45 to 1000) Hz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz (1.2 to 3.1) A (3 to 45) Hz (45 to 1000) Hz (1 to 5) kHz (5 to 10) kHz (3.1 to 12) A (3 to 45) Hz (45 to 1 000) Hz (1 to 5) kHz (5 to 10) kHz (12 to 30.2) A (3 to 45) Hz (45 to 1 000) Hz (1 to 5) kHz	0.019 % of output + 0.08 mA 0.019 % of output + 0.04 mA 0.019 % of output + 0.06 mA 0.19 % of output + 0.23 mA 0.38 % of output + 0.23 mA 0.029 % of output + 0.4 mA 0.023 % of output + 0.27 mA 0.029 % of output + 0.27 mA 0.19 % of output + 0.40 mA 0.029 % of output + 0.8 mA 0.023 % of output + 0.4 mA 0.029 % of output + 0.6 mA 0.19 % of output + 0.8 mA 0.076 % of output + 7.7 mA 0.053 % of output + 6.2 mA 0.38 % of output + 6.2 mA	Fluke 5560A Multiproduct Calibrator
AC Current – Generate <sup>1</sup> Clamp Meters	(3.1 to 300.2) A (3 to 440) Hz	0.65 % of output + 0.077 A	Fluke 5560A Multiproduct Calibrator 55xxA x1/2/10 Coil
AC Current – Measure <sup>1</sup>	(2.02 to 20.2) $\mu$ A (1 to 2 000) Hz (2 to 10) kHz (10 Hz to 30) kHz (2.02 to 20.2) $\mu$ A (1 to 2 000) Hz (2 to 10) kHz (10 Hz to 30) kHz (0.202 to 2.02) mA (1 to 2 000) Hz (2 to 10) kHz (10 Hz to 30) kHz	0.38 % of reading + 2.9 nA 0.38 % of reading + 2.9 nA 0.38 % of reading + 2.9 nA 0.044 % of reading + 9.8 nA 0.087 % of reading + 9.8 nA 0.11 % of reading + 9.8 nA 0.044 % of reading + 98 nA 0.087 % of reading + 98 nA 0.11 % of reading + 98 nA	Fluke 8588A Multimeter

**Electrical – DC/Low Frequency**

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Measure <sup>1</sup>	(2.02 to 20.2) mA (1 to 2 000) Hz (2 to 10) kHz (10 Hz to 30) kHz (20.2 to 202) mA (1 to 2 000) Hz (2 to 10) kHz (10 Hz to 30) kHz (0.202 to 2.02)A (1 to 2 000) Hz (2 to 10) kHz (10 Hz to 30) kHz	0.044 % of reading + 0.98 μA 0.087 % of reading + 0.98 μA 0.11 % of reading + 0.98 μA 0.044 % of reading + 9.8 μA 0.087 % of reading + 9.8 μA 0.11 % of reading + 9.8 μA 0.044 % of reading + 148 μA 0.076 % of reading + 148 μA 0.12 % of reading + 148 μA	Fluke 8588A Multimeter
AC Current – Measure <sup>1</sup>	(5 to 100) μA (10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz 100 μA to 1 mA (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz (1 to 10) mA (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz (10 to 100) mA (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz 100 mA to 1 A (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz	0.26 % of reading + 0.02 μA 0.1 % of reading + 0.02 μA 0.04 % of reading + 0.02 μA 0.26 % of reading + 0.13 μA 0.1 % of reading + 0.13 μA 0.04 % of reading + 0.13 μA 0.02 % of reading + 0.13 μA 0.26 % of reading + 1.3 μA 0.1 % of reading + 1.3 μA 0.04 % of reading + 1.3 μA 0.02 % of reading + 1.3 μA 0.26 % of reading + 13 μA 0.1 % of reading + 13 μA 0.04 % of reading + 13 μA 0.02 % of reading + 13 μA 0.26 % of reading + 0.13 mA 0.1 % of reading + 0.13 mA 0.05 % of reading + 0.13 mA 0.07 % of reading + 0.13 mA	Agilent 3458A Multimeter



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Measure <sup>1</sup>	(1 to 3) A (3 to 5) Hz (5 to 10) Hz 10 Hz to 5 kHz (5 to 10) kHz (3 to 10) A (3 to 5) Hz (5 to 10) Hz 10 Hz to 5 kHz (5 to 10) kHz	0.84 % of reading + 1.4 mA 0.27 % of reading + 1.4 mA 0.11 % of reading + 1.4 mA 0.27 % of reading + 16 mA 0.84 % of reading + 4.6 mA 0.27 % of reading + 4.6 mA 0.11 % of reading + 4.6 mA 0.27 % of reading + 53 mA	Fluke 8845A Multimeter
Resistance – Generate <sup>1</sup> Fixed Points	(1, 1.9) Ω (10, 19) Ω (100, 190) Ω (1, 1.9) kΩ (10, 19) kΩ 100 kΩ 190 kΩ 1 MΩ 1.9 MΩ 10 MΩ 19 MΩ 100 MΩ	84 μΩ/Ω + 0.1 μΩ 21 μΩ/Ω + 1 μΩ 9.1 μΩ/Ω + 6 μΩ 6.1 μΩ/Ω + 60 μΩ 6.1 μΩ/Ω + 0.6 mΩ 7.6 μΩ/Ω + 6 mΩ 9.1 μΩ/Ω + 6 mΩ 11.4 μΩ/Ω + 60 mΩ 16 μΩ/Ω + 60 mΩ 35 μΩ/Ω + 0.6 Ω 42 mΩ/Ω + 0.6 Ω 91 mΩ/Ω + 6 Ω	Fluke 5730A Multifunction Calibrator
Resistance – Generate <sup>1</sup>	Up to 12 Ω (12 to 120) Ω (0.12 to 1.20) kΩ (1.2 to 12.0) kΩ (12 to 120) kΩ (0.12 to 1.2) MΩ (1.2 to 12) MΩ (12 to 120) MΩ (120 to 1 200) MΩ	19 μΩ/Ω + 0.001 Ω 19 μΩ/Ω + 0.001 Ω 19 μΩ/Ω + 0.002 Ω 19 μΩ/Ω + 0.02 Ω 19 μΩ/Ω + 0.2 Ω 19 μΩ/Ω + 2.0Ω 27 μΩ/Ω + 24 Ω 327 μΩ/Ω + 2 kΩ 3 mΩ/Ω + 76 Ω	Fluke 5560A Multiproduct Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Resistance – Measure <sup>1</sup>	Up to 2.02 Ω 2.02 to 20.2 Ω (20.2 to 202) Ω (0.202 to 2.02) kΩ (2.02 to 20.2) kΩ (20.2 to 202) kΩ (0.202 to 2.02) MΩ (2.02 to 20.2) MΩ (20.2 to 202) MΩ (0.202 to 2.02) GΩ (2.02 to 20.2) GΩ	21 μΩ/Ω + 4.4 μΩ 15 μΩ/Ω + 20 μΩ 12 μΩ/Ω + 53 μΩ 12 μΩ/Ω + 0.53 mΩ 12 μΩ/Ω + 5.3 mΩ 12 μΩ/Ω + 53 mΩ 12.9 μΩ/Ω + 1.0 Ω 17 μΩ/Ω + 9.8 Ω 67 μΩ/Ω + 0.98 kΩ 228 μΩ/Ω + 98 kΩ 1.3 mΩ/Ω + 9.8 MΩ	Fluke 8588A Multimeter
Resistance – Measure <sup>1</sup>	Up to 12 Ω (10 to 120) Ω (0.1 to 1.2) kΩ (1 to 12) kΩ (10 to 120) kΩ (0.1 to 1.2) MΩ (1 to 12) MΩ (10 to 120) MΩ (0.1 to 1.2) GΩ	9.8 μΩ/Ω + 38 μΩ 7.8 μΩ/Ω + 0.3 mΩ 6.5 μΩ/Ω + 0.4 mΩ 6.5 μΩ/Ω + 3.8 mΩ 6.5 μΩ/Ω + 38 mΩ 9.8 μΩ/Ω + 1.5 Ω 33 μΩ/Ω + 100 Ω 327 μΩ/Ω + 1 kΩ 0.33 % of reading + 70 kΩ	Agilent 3458A Multimeter
Capacitance – Generate <sup>1</sup>	(0.2 to 1.2) nF (1.2 to 12.0) nF (12 to 120.0) nF (0.12 to 1.2) μF (1.2 to 12.0) μF (12 to 120.0) μF (0.12 to 1.2) mF (1.2 to 12.0) mF (12 to 120.0) mF	0.09% of output + 1.5 pF 0.09 % of output + 0.004 nF 0.1 % of output + 0.023 nF 0.1 % of output + 0.23 nF 0.1 % of output + 2.3 nF 0.11 % of output + 19 nF 0.19 % of output + 190 nF 0.19 % of output + 2.3 μF 0.38 % of output + 23 μF	Fluke 5560A Multiproduct Calibrator
Inductance – Generate <sup>1</sup>	(13 to 120.0) μH (0.12 to 1.2) mH (1.2 to 12.0) mH (12 to 120.0) mH (0.12 to 1.2) H (1.2 to 12.0) H (12 to 120.0) H	0.15% of output + 0.15 μH 0.09 % of output + 0.76 μH 0.09 % of output + 7.6 μH 0.09 % of output + 76 μH 0.11 % of output + 0.76 mH 0.15 % of output + 7.6 mH 0.19 % of output + 76 mH	Fluke 5560A Multiproduct Calibrator

**Electrical – DC/Low Frequency**

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Oscilloscopes Calibration <sup>1</sup> – Generate Voltage DC - 50Ω DC - 1MΩ  Square Wave 10Hz to 10kHz – 1MΩ	1 mV to 5 V 1 mV to 200 V  1 mV to 200 Vpp	0.025 % of output + 25 μV 0.025 % of output + 2.5 μV  0.1 % of output + 10 μV	Fluke 9500B with 9510 Active Head
Oscilloscopes Calibration <sup>1</sup> – Generate  Leveled Sine Flatness 50 kHz to 10 MHz Reference	(5mV to 5Vpp) 0.1 Hz to 300 MHz (300 to 550) MHz  5 mV to 3 Vpp (550 to 1 100) MHz	0.18 dB 0.22 dB  0.3 dB	Fluke 9500B with 9510 Active Head
Oscilloscopes Calibration <sup>1</sup> – Generate Time Marker	9 ns to 55 s	0.25 μs/s	Fluke 9500B with 9510 Active Head
Oscilloscopes Calibration <sup>1</sup> – Measure  Input Impedance Resistance	(10 to 40) Ω (40 to 90) Ω (90 to 150) Ω (50 to 800) KΩ (0.8 to 1.2) MΩ (1.2 to 12) MΩ	0.5 % of reading 0.1 % of reading 0.5 % of reading 0.5 % of reading 0.1 % of reading 0.5 % of reading	Fluke 9500B with 9510 Active Head
Electrical Calibration of Thermocouple Indicators/Sources <sup>1</sup>	Type B (600 to 800) °C (800 to 1000) °C (1 000 to 1 550) °C (1 550 to 1 820) °C Type C (0 to 150) °C (150 to 650) °C (650 to 1 000) °C (1 000 to 1 800) °C (1 800 to 2 315) °C Type D (0 to 150) °C (150 to 650) °C (650 to 1 000) °C (1 000 to 1 800) °C (1 800 to 2 315) °C	0.33 °C 0.26 °C 0.23 °C 0.25 °C  0.19 °C 0.16 °C 0.20 °C 0.34 °C 0.60 °C  0.19 °C 0.16 °C 0.2 °C 0.33 °C 0.59 °C	Fluke 5560A Multiproduct Calibrator

**Electrical – DC/Low Frequency**

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment		
<p>Electrical Calibration of Thermocouple Indicators/Sources <sup>1</sup></p>	<p>Type E            (-250 to -150) °C            (-150 to -25) °C            (-25 to 350) °C            (350 to 650) °C            (650 to 1 000) °C</p>	<p>0.3 °C            0.11 °C            0.08 °C            0.12 °C            0.16 °C</p>	<p>Fluke 5560A Multiproduct Calibrator</p>		
	<p>Type G            (0 to 150) °C            (150 to 650) °C            (650 to 1 000) °C            (1 000 to 1 800) °C            (1 800 to 2 315) °C</p>	<p>0.38 °C            0.25 °C            0.2 °C            0.33 °C            0.59 °C</p>			
	<p>Type J            (-210 to -100) °C            (-100 to -30) °C            (-30 to 150) °C            (150 to 760) °C            (760 to 1 200) °C</p>	<p>0.18 °C            0.1 °C            0.08 °C            0.11 °C            0.15 °C</p>			
	<p>Type K            (-200 to -100) °C            (-100 to -25) °C            (-25 to 120) °C            (120 to 1 000) °C            (1 000 to 1 372) °C</p>	<p>0.21 °C            0.1 °C            0.08 °C            0.16 °C            0.27 °C</p>			
	<p>Type L            (-200 to -100) °C            (-100 to 800) °C            (800 to 900) °C</p>	<p>0.24 °C            0.15 °C            0.08 °C</p>			
	<p>Electrical Calibration of Thermocouple Indicators <sup>1</sup></p>	<p>Type N            (-200 to -100) °C            (-100 to -25) °C            (-25 to 120) °C            (120 to 410) °C            (410 to 1 300) °C</p>		<p>0.25 °C            0.11 °C            0.09 °C            0.08 °C            0.15 °C</p>	<p>Fluke 7526A Process Calibrator</p>
		<p>Type R            (0 to 250) °C            (250 to 400) °C            (400 to 1 000) °C            (1 000 to 1 767) °C</p>		<p>0.39 °C            0.22 °C            0.21 °C            0.26 °C</p>	

**Electrical – DC/Low Frequency**

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Calibration of Thermocouple Indicators <sup>1</sup>	Type S (0 to 250) °C	0.32 °C	Fluke 7526A Process Calibrator
	(250 to 1 000) °C	0.24 °C	
	(1 000 to 1 400) °C	0.24 °C	
	(1 400 to 1 767) °C	0.31 °C	
	Type T (-250 to -150) °C	0.46 °C	
	(-150 to 0) °C	0.16 °C	
	(0 to 120) °C	0.1 °C	
	(120 to 400) °C	0.08 °C	
	Type U (-200 to 0) °C	0.3 °C	
	(0 to 600) °C	0.08 °C	
	Type BP (0 to 1 000) °C	0.3 °C	
	(1 000 to 2 000) °C	0.46 °C	
	(2 000 to 2 500) °C	0.61 °C	
Type XK (-200 to 300) °C	0.15 °C	Fluke 5560A Multiproduct Calibrator	
(300 to 800) °C	0.23 °C		
Cu 10 (427) (-80 to 260) °C	0.23 °C		
Electrical Calibration of RTD Indicators <sup>1</sup>	Cu 50 (428) (-180 to 200) °C	0.30 °C	Fluke 5560A Multiproduct Calibrator
	Cu 100 (428) (-180 to 40) °C	0.3 °C	
	(-40 to 200) °C	0.49 °C	
Electrical Calibration of RTD Indicators <sup>1</sup>	Ni 120 (672) (-80 to 0) °C	0.06 °C	Fluke 5560A Multiproduct Calibrator
	(0 to 100) °C	0.06 °C	
	(100 to 260) °C	0.11 °C	
	Pt 100 (385) (-200 to -80) °C	0.04 °C	
	(-80 to 0) °C	0.04 °C	
	(0 to 100) °C	0.05 °C	
	(100 to 300) °C	0.07 °C	
	(300 to 400) °C	0.08 °C	
	(400 to 630) °C	0.09 °C	
	(630 to 800) °C	0.18 °C	



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Calibration of RTD Indicators <sup>1</sup>	Pt 100 (3916) (-200 to -190)°C (-190 to -80)°C (-80 to 0)°C (0 to 100)°C (100 to 260)°C (260 to 300)°C (300 to 400)°C (400 to 600)°C (600 to 630)°C  Pt 100 (3926) (-200 to -80)°C (-80 to 0)°C (0 to 100)°C (100 to 300)°C (300 to 400)°C (400 to 630)°C  Pt 200 (385) (-200 to -80)°C (-80 to 0)°C (0 to 100)°C (100 to 260)°C (260 to 300)°C (300 to 400)°C (400 to 600)°C (600 to 630)°C	0.19 °C 0.03 °C 0.04 °C 0.05 °C 0.05 °C 0.06 °C 0.07 °C 0.08 °C 0.18 °C  0.04 °C 0.04 °C 0.05 °C 0.07 °C 0.08 °C 0.09 °C  0.03 °C 0.03 °C 0.03 °C 0.04 °C 0.09 °C 0.10 °C 0.11 °C 0.12 °C	Fluke 5560A Multiproduct Calibrator
Electrical Calibration of RTD Indicators <sup>1</sup>	Pt 500 (385) (-200 to -80)°C (-80 to 0)°C (0 to 100)°C (100 to 260)°C (260 to 300)°C (300 to 400)°C (400 to 600)°C (600 to 630)°C	0.03 °C 0.04 °C 0.04 °C 0.05 °C 0.06 °C 0.06 °C 0.07 °C 0.08 °C	Fluke 5560A Multiproduct Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Calibration of RTD Indicators <sup>1</sup>	Pt 1000 (385) (-200 to -80)°C (-80 to 0)°C (0 to 100)°C (100 to 260)°C (260 to 300)°C (300 to 400)°C (400 to 600)°C (600 to 630)°C	0.02 °C 0.02 °C 0.03 °C 0.04 °C 0.05 °C 0.05 °C 0.05 °C 0.18 °C	Fluke 5560A Multiproduct Calibrator
AC Power – Generate <sup>1</sup>  (12 to 120) mV <i>Power Factor = 1</i>	(1.2 to 12) mA (10 to 40) Hz (40 to 1000) Hz (1 to 5) kHz (5 to 10) kHz (12 to 120) mA (10 to 40) Hz (40 to 1000) Hz (1 to 5) kHz (5 to 10) kHz (0.12 to 1.2) A (10 to 40) Hz (40 to 1000) Hz (1 to 5) kHz (5 to 10) kHz (1.2 to 3.1) A (10 to 40) Hz (40 to 1000) Hz (1 to 5) kHz (5 to 10) kHz	0.022 % of Output + 0.9 μW 0.022 % of Output + 0.9 μW 0.022 % of Output + 0.9 μW 0.115 % of Output + 0.9 μW 0.022 % of Output + 9.2 μW 0.016 % of Output + 4.7 μW 0.022 % of Output + 7.4 μW 0.115 % of Output + 9.2 μW 0.022 % of Output + 92 μW 0.022 % of Output + 47 μW 0.022 % of Output + 74 μW 0.19 % of Output + 0.27 mW 0.03 % of Output + 0.46 mW 0.025 % of Output + 0.27 mW 0.03 % of Output + 0.27 mW 0.19 % of Output + 0.46 mW	Fluke 5560A Multiproduct Calibrator
AC Power – Generate <sup>1</sup>  (12 to 120) mV <i>Power Factor = 1</i>	(3.1 to 12) A (10 to 40) Hz (40 to 1000) Hz (1 to 5) kHz (5 to 10) kHz (12 to 30.2) A (10 to 40) Hz (40 to 1000) Hz (1 to 5) kHz	0.03 % of Output + 0.92 mW 0.025 % of Output + 0.47 mW 0.03 % of Output + 0.74 mW 0.19 % of Output + 0.92 mW 0.077 % of Output + 9.1 mW 0.054 % of Output + 7.3 mW 0.38 % of Output + 7.3 mW	Fluke 5560A Multiproduct Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment		
AC Power – Generate <sup>1</sup>  (0.12 to 1.2) V <i>Power Factor = 1</i>	(1.2 to 12) mA		Fluke 5560A Multiproduct Calibrator		
	(10 to 40) Hz	0.022 % of Output + 0.9 μW			
	(40 to 1000) Hz	0.022 % of Output + 0.9 μW			
	(1 to 5) kHz	0.022 % of Output + 0.9 μW			
	(5 to 10) kHz	0.12 % of Output + 0.9 μW			
	(12 to 120) mA				
	(10 to 40) Hz	0.022 % of Output + 9.2 μW			
	(40 to 1000) Hz	0.016 % of Output + 4.7 μW			
	(1 to 5) kHz	0.022 % of Output + 7.4 μW			
	(5 to 10) kHz	0.12 % of Output + 9.2 μW			
	(0.12 to 1.2) A				
	(10 to 40) Hz	0.022 % of Output + 92 μW			
	(40 to 1000) Hz	0.022 % of Output + 47 μW			
	(1 to 5) kHz	0.022 % of Output + 74 μW			
	(5 to 10) kHz	0.19 % of Output + 0.27 mW			
	AC Power – Generate <sup>1</sup>  (1.2 to 12) V <i>Power Factor = 1</i>	(1.2 to 3.1) A			Fluke 5560A Multiproduct Calibrator
(10 to 40) Hz		0.03 % of Output + 0.46 mW			
(40 to 1000) Hz		0.025 % of Output + 0.27 mW			
(1 to 5) kHz		0.03 % of Output + 0.27 mW			
(5 to 10) kHz		0.19 % of Output + 0.46 mW			
(3.1 to 12) A					
(10 to 40) Hz		0.03 % of Output + 0.92 mW			
(40 to 1000) Hz		0.025 % of Output + 0.47 mW			
(1 to 5) kHz		0.03 % of Output + 0.74 mW			
(5 to 10) kHz		0.19 % of Output + 0.92 mW			
(12 to 30.2) A					
(10 to 40) Hz		0.077 % of Output + 9.1 mW			
(40 to 1000) Hz		0.054 % of Output + 7.3 mW			
(1 to 5) kHz		0.38 % of Output + 7.3 mW			
AC Power – Generate <sup>1</sup>  (1.2 to 12) V <i>Power Factor = 1</i>		(1.2 to 12) mA		Fluke 5560A Multiproduct Calibrator	
		(10 to 40) Hz	0.022 % of Output + 9.2 μW		
	(40 to 1 000) Hz	0.022 % of Output + 9.2 μW			
	(1 to 5) kHz	0.022 % of Output + 9.2 μW			
	(5 to 10) kHz	0.12 % of Output + 9.2 μW			
	(12 to 120) mA				
	(10 to 40) Hz	0.022 % of Output + 92 μW			
	(40 to 1 000) Hz	0.016 % of Output + 46 μW			
(1 to 5) kHz	0.022 % of Output + 74 μW				
(5 to 10) kHz	0.12 % of Output + 92 μW				



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Power – Generate <sup>1</sup>  (1.2 to 12) V Power Factor = 1	(0.12 to 1.2) A (10 to 40) Hz (40 to 1 000) Hz (1 to 5) kHz (5 to 10) kHz (1.2 to 3.1) A (10 to 40) Hz (40 to 1 000) Hz (1 to 5) kHz (5 to 10) kHz (3.1 to 12) A (10 to 40) Hz (40 to 1 000) Hz (1 to 5) kHz (5 to 10) kHz (12 to 30.2) A (10 to 40) Hz (40 to 1 000) Hz (1 to 5) kHz	0.022 % of Output + 0.92 mW 0.022 % of Output + 0.46 mW 0.022 % of Output + 0.74 mW 0.19 % of Output + 2.7 mW 0.03 % of Output + 4.6 mW 0.025 % of Output + 2.7 mW 0.03 % of Output + 2.7 mW 0.19 % of Output + 4.6 mW 0.03 % of Output + 9.2 mW 0.025 % of Output + 4.6 mW 0.03 % of Output + 7.4 mW 0.19 % of Output + 9.2 mW 0.077 % of Output + 91 mW 0.054 % of Output + 73 mW 0.38 % of Output + 73 mW	Fluke 5560A Multiproduct Calibrator
AC Power – Generate <sup>1</sup>  (12 to 120) V Power Factor = 1	(1.2 to 12) mA (10 to 40) Hz (40 to 1 000) Hz (1 to 5) kHz (5 to 10) kHz (12 to 120) mA (10 to 40) Hz (40 to 1 000) Hz (1 to 5) kHz (5 to 10) kHz (0.12 to 1.2) A (10 to 40) Hz (40 to 1 000) Hz (1 to 5) kHz (5 to 10) kHz (1.2 to 3.1) A (10 to 40) Hz (40 to 1 000) Hz (1 to 5) kHz (5 to 10) kHz	0.022 % of Output + 92 μW 0.022 % of Output + 92 μW 0.022 % of Output + 92 μW 0.12 % of Output + 92 μW 0.022 % of Output + 0.92 mW 0.016 % of Output + 0.46 mW 0.022 % of Output + 0.74 mW 0.12 % of Output + 0.92 mW 0.022 % of Output + 9.2 mW 0.022 % of Output + 4.6 mW 0.022 % of Output + 7.4 mW 0.19 % of Output + 27 mW 0.03 % of Output + 46 mW 0.025 % of Output + 27 mW 0.03 % of Output + 27 mW 0.19% of Output + 46 mW	Fluke 5560A Multiproduct Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Power – Generate <sup>1</sup>  (12 to 120) V Power Factor = 1	(3.1 to 12) A (10 to 40) Hz (40 to 1 000) Hz (1 to 5) kHz (5 to 10) kHz (12 to 30.2) A (10 to 40) Hz (40 to 1 000) Hz (1 to 5) kHz	0.03 % of Output + 92 mW 0.025 % of Output + 46 mW 0.03 % of Output + 74 mW 0.19 % of Output + 92 mW 0.077 % of Output + 0.91 W 0.054 % of Output + 0.73 W 0.38 % of Output + 0.73 W	Fluke 5560A Multiproduct Calibrator
AC Power – Generate <sup>1</sup>  (120 to 330) V Power Factor = 1	(1.2 to 12) mA (10 to 40) Hz (40 to 1 000) Hz (1 to 5) kHz (5 to 10) kHz (12 to 120) mA (10 to 40) Hz (40 to 1 000) Hz (1 to 5) kHz (5 to 10) kHz (0.12 to 1.2) A (10 to 40) Hz (40 to 1 000) Hz (1 to 5) kHz (5 to 10) kHz (1.2 to 3.1) A (10 to 40) Hz (40 to 1 000) Hz (1 to 5) kHz (5 to 10) kHz (3.1 to 12) A (10 to 40) Hz (40 to 1 000) Hz (1 to 5) kHz (5 to 10) kHz (12 to 30.2) A (10 to 40) Hz (40 to 1 000) Hz (1 to 5) kHz	0.022 % of Output + 0.26 mW 0.022 % of Output + 0.26 mW 0.022 % of Output + 0.26 mW 0.12 % of Output + 0.26 mW 0.022 % of Output + 2.6 mW 0.016 % of Output + 1.5 mW 0.022 % of Output + 2.1 mW 0.12 % of Output + 2.6 mW 0.022 % of Output + 26 mW 0.022 % of Output + 15 mW 0.022 % of Output + 21 mW 0.19 % of Output + 76 mW 0.03 % of Output + 0.13 W 0.025 % of Output + 0.08 W 0.03 % of Output + 0.08 W 0.19 % of Output + 0.13 W 0.03 % of Output + 0.26 W 0.025 % of Output + 0.15 W 0.03 % of Output + 0.21 W 0.19 % of Output + 0.26 W 0.077 % of Output + 2.5 W 0.054 % of Output + 2.0 W 0.38 % of Output + 2.0 W	Fluke 5560A Multiproduct Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Power – Generate <sup>1</sup>  (330 to 1020) V <i>Power Factor = 1</i>	(1.2 to 12) mA		Fluke 5560A Multiproduct Calibrator
	(10 to 40) Hz	0.022 % of Output + 1.1 mW	
	(40 to 1 000) Hz	0.022 % of Output + 1.1 mW	
	(1 to 5) kHz	0.022 % of Output + 1.1 mW	
	(5 to 10) kHz	0.12 % of Output + 1.1 mW	
	(12 to 120) mA		
	(10 to 40) Hz	0.022 % of Output + 11 mW	
	(40 to 1 000) Hz	0.016 % of Output + 8.3 mW	
	(1 to 5) kHz	0.022 % of Output + 9.6 mW	
	(5 to 10) kHz	0.12 % of Output + 11 mW	
	(0.12 to 1.2) A		
	(10 to 40) Hz	0.022 % of Output + 107 mW	
	(40 to 1 000) Hz	0.022 % of Output + 83 mW	
	(1 to 5) kHz	0.022 % of Output + 96 mW	
	(5 to 10) kHz	0.19 % of Output + 244 mW	
(1.2 to 3.1) A			
(10 to 40) Hz	0.03 % of Output + 0.43 W		
(40 to 1 000) Hz	0.025 % of Output + 0.3 W		
(1 to 5) kHz	0.03 % of Output + 0.3 W		
(5 to 10) kHz	0.19 % of Output + 0.43 W		
(3.1 to 12) A			
(10 to 40) Hz	0.03 % of Output + 1.1 W		
(40 to 1 000) Hz	0.025 % of Output + 0.8 W		
(1 to 5) kHz	0.03 % of Output + 1 W		
(5 to 10) kHz	0.19 % of Output + 1.1 W		
(12 to 30.2) A			
(10 to 40) Hz	0.077 % of Output + 8 W		
(40 to 1 000) Hz	0.054 % of Output + 6.5 W		
(1 to 5) kHz	0.38 % of Output + 6.5 W		
Phase Angle – Generate <sup>1</sup>	(-180 to 180)°		Fluke 5560A Multiproduct Calibrator
	(10 to 65) Hz	0.08°	
	(65 to 500) Hz	0.19°	
	500 Hz to 1 kHz	0.38°	
	(1 to 5) kHz	1.9°	
	(5 to 10) kHz	3.8°	
(10 to 30) kHz	7.6°		
Power Meter Range Calibration <sup>1</sup>	3 μW to 100 mW	0.25 % of reading	HP 11683A Power Meter Calibrator



ANSI National Accreditation Board

Electrical – RF/Microwave

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Flatness – Measure <sup>1</sup>	9 kHz to 2 000 MHz		Agilent EPM Series Power Meter w/E9304A H18 Power Sensor
	(20 to -10) dBm	0.1 dB	
	(-10 to -30) dBm	0.1 dB	
	(-30 to -40) dBm	0.11 dB	
	(-40 to -42) dBm	0.12 dB	
	(2 to 14) GHz		
	(20 to -10) dBm	0.1 dB	
	(-10 to -30) dBm	0.09 dB	
	(-30 to -40) dBm	0.1 dB	
	(-40 to -42) dBm	0.11 dB	
	(14 to 18) GHz		
	(20 to -10) dBm	0.11 dB	
	(-10 to -30) dBm	0.12 dB	
	(-30 to -40) dBm	0.12 dB	
(-40 to -42) dBm	0.13 dB		
RF Power Measure <sup>1</sup>	9 kHz to 14 000 MHz		Agilent EPM Series Power Meter w/E9304A H18 Power Sensor
	(20 to 0) dB	0.13 dB	
	(0 to -40) dB	0.15 dB	
	(-40 to -50) dB	0.35 dB	
	(-50 to -55) dB	0.93 dB	
	(14 000 to 18 000) MHz		
	(20 to 0) dB	0.12 dB	
	(0 to -40) dB	0.16 dB	
	(-40 to -50) dB	0.35 dB	
	(-50 to -55) dB	0.93 dB	
RF Power Measure <sup>1</sup>	(50 to 100) MHz		Agilent EPM Series Power Meter w/N8487A Power Sensor
	(20 to 10) dB	0.08 dB	
	(10 to 0) dB	0.07 dB	
	(0 to -10) dB	0.07 dB	
	(-10 to -20) dB	0.07 dB	
	(-20 to -25) dB	0.11 dB	
	(100 to 6 000) MHz		
	(20 to 10) dB	0.08 dB	
	(10 to 0) dB	0.07 dB	
	(0 to -10) dB	0.07 dB	
	(-10 to -20) dB	0.08 dB	
	(-20 to -25) dB	0.11 dB	

Electrical – RF/Microwave

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power Measure <sup>1</sup>	(6 000 to 12 400) MHz		Agilent EPM Series Power Meter w/N8487A Power Sensor
	(20 to 10) dB	0.08 dB	
	(10 to 0) dB	0.08 dB	
	(0 to -10) dB	0.08 dB	
	(-10 to -20) dB	0.08 dB	
	(-20 to -25) dB	0.12 dB	
	(12 400 to 18 000) MHz		
	(20 to 10) dB	0.09 dB	
	(10 to 0) dB	0.09 dB	
	(0 to -10) dB	0.08 dB	
	(-10 to -20) dB	0.09 dB	
	(-20 to -25) dB	0.12 dB	
	(18 000 to 26 500) MHz		
	(20 to 10) dB	0.11 dB	
	(10 to 0) dB	0.11 dB	
	(0 to -10) dB	0.1 dB	
	(-10 to -20) dB	0.11 dB	
	(-20 to -25) dB	0.14 dB	
	(26 500 to 33 000) MHz		
	(20 to 10) dB	0.12 dB	
	(10 to 0) dB	0.12 dB	
	(0 to -10) dB	0.12 dB	
	(-10 to -20) dB	0.12 dB	
	(-20 to -25) dB	0.15 dB	
(33 000 to 40 000) MHz			
(20 to 10) dB	0.13 dB		
(10 to 0) dB	0.13 dB		
(0 to -10) dB	0.12 dB		
(-10 to -20) dB	0.13 dB		
(-20 to -25) dB	0.15 dB		
(40 000 to 50 000) MHz			
(20 to 10) dB	0.19 dB		
(10 to 0) dB	0.18 dB		
(0 to -10) dB	0.18 dB		
(-10 to -20) dB	0.19 dB		
(-20 to -25) dB	0.20 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power Measure <sup>1</sup>	100 kHz to 30 MHz (20 to 0) dB (0 to -58) dB (-58 to -78) dB (-78 to -110) dB (-110 to -115) dB (-115 to -120) dB (-120 to -125) dB	0.12 dB 0.13 dB 0.15 dB 0.19 dB 0.26 dB 0.39 dB 0.59 dB	Agilent N5531S Measuring Receiver N1912A w/E9304A H18 Power Sensor
RF Power Measure <sup>1</sup>	(30 to 2 000) MHz (30 to 20) dB (20 to 0) dB (0 to -58) dB (-58 to -78) dB (-78 to -110) dB (-110 to -115) dB (-115 to -120) dB (-120 to -125) dB (2 000 to 3 050) MHz (30 to 20) dB (20 to 0) dB (0 to -58) dB (-58 to -78) dB (-78 to -110) dB (-110 to -115) dB (-115 to -120) dB (-120 to -125) dB (3 050 to 6 600) MHz (30 to 20) dB (20 to 0) dB (0 to -58) dB (-58 to -78) dB (-78 to -110) dB (-110 to -115) dB (-115 to -120) dB	0.36 dB 0.2 dB 0.22 dB 0.23 dB 0.25 dB 0.27 dB 0.33 dB 0.7 dB 0.37 dB 0.21 dB 0.23 dB 0.24 dB 0.26 dB 0.28 dB 0.34 dB 0.7 dB 0.37 dB 0.21 dB 0.23 dB 0.24 dB 0.29 dB 0.38 dB 0.53 dB	Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor

**Electrical – RF/Microwave**

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power Measure <sup>1</sup>	(6 600 to 13 200) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor
	(30 to 20) dB	0.37 dB	
	(20 to 0) dB	0.21 dB	
	(0 to -58) dB	0.23 dB	
	(-58 to -78) dB	0.24 dB	
	(-78 to -110) dB	0.34 dB	
	(-110 to -115) dB	0.46 dB	
	(-115 to -120) dB	0.65 dB	
	(13 200 to 18 000) MHz		
	(30 to 20) dB	0.37 dB	
	(20 to 0) dB	0.21 dB	
	(0 to -58) dB	0.23 dB	
	(-58 to -78) dB	0.24 dB	
	(-78 to -90) dB	0.26 dB	
	(-90 to -95) dB	0.26 dB	
	(-95 to -100) dB	0.26 dB	
	(-100 to -105) dB	0.29 dB	
	(-105 to -110) dB	0.38 dB	
	(-110 to -115) dB	0.53 dB	
	(-115 to -120) dB	0.75 dB	
	(18 000 to 19 200) MHz		
	(30 to 20) dB	0.39 dB	
	(20 to 0) dB	0.25 dB	
	(0 to -58) dB	0.24 dB	
	(-58 to -78) dB	0.25 dB	
	(-78 to -90) dB	0.27 dB	
	(-90 to -95) dB	0.27 dB	
	(-95 to -100) dB	0.27 dB	
	(-100 to -105) dB	0.3 dB	
	(-105 to -110) dB	0.38 dB	
(-110 to -115) dB	0.53 dB		
(-115 to -120) dB	0.75 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power Measure <sup>1</sup>	(19 200 to 26 500) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor
	(30 to 20) dB	0.39 dB	
	(20 to 0) dB	0.25 dB	
	(0 to -58) dB	0.24 dB	
	(-58 to -78) dB	0.25 dB	
	(-78 to -90) dB	0.28 dB	
	(-90 to -95) dB	0.33 dB	
	(-95 to -100) dB	0.43 dB	
	(-100 to -105) dB	0.61 dB	
	(-105 to -110) dB	0.85 dB	
	(-110 to -115) dB	1.2 dB	
	(-115 to -120) dB	1.5 dB	
	(26 500 to 31 150) MHz		
	(30 to 20) dB	0.42 dB	
	(20 to 0) dB	0.3 dB	
	(0 to -58) dB	0.34 dB	
	(-58 to -78) dB	0.34 dB	
	(-78 to -90) dB	0.36 dB	
	(-90 to -95) dB	0.39 dB	
	(-95 to -100) dB	0.46 dB	
	(-100 to -105) dB	0.61 dB	
	(-105 to -110) dB	0.82 dB	
	(-110 to -115) dB	1.1 dB	
	(31 150 to 41 000) MHz		
	(30 to 20) dB	0.42 dB	
	(20 to 0) dB	0.3 dB	
	(0 to -58) dB	0.34 dB	
(-58 to -78) dB	0.35 dB		
(-78 to -90) dB	0.48 dB		
(-90 to -95) dB	0.64 dB		
(-95 to -100) dB	0.87 dB		
(-100 to -105) dB	1.2 dB		
(-105 to -110) dB	1.5 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power Measure <sup>1</sup>	(41 000 to 45 000) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor
	(30 to 20) dB	0.42 dB	
	(20 to 0) dB	0.3 dB	
	(0 to -58) dB	0.34 dB	
	(-58 to -78) dB	0.38 dB	
	(-78 to -90) dB	0.68 dB	
	(-90 to -95) dB	0.93 dB	
	(-95 to -100) dB	1.2 dB	
	(-100 to -105) dB	1.6 dB	
	(45 000 to 50 000) MHz		
	(30 to 20) dB	0.42 dB	
	(20 to 0) dB	0.3 dB	
	(0 to -58) dB	0.34 dB	
	(-58 to -78) dB	0.67 dB	
(-78 to -90) dB	1.4 dB		
RF Attenuation – Measure <sup>1</sup>	(30 to 3 050) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.12 dB	
	(90 to 100) dB	0.13 dB	
	(100 to 110) dB	0.13 dB	
	(3 050 to 6 600) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.12 dB	
(90 to 100) dB	0.13 dB		
(100 to 110) dB	0.19 dB		

Electrical – RF/Microwave

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Attenuation – Measure <sup>1</sup>	(6 600 to 13 200) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.12 dB	
	(90 to 100) dB	0.13 dB	
	(100 to 110) dB	0.25 dB	
	(13 200 to 19 200) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.12 dB	
	(90 to 100) dB	0.13 dB	
	(100 to 110) dB	0.31 dB	
	(19 200 to 26 500) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
(50 to 60) dB	0.076 dB		
(60 to 70) dB	0.081 dB		
(70 to 80) dB	0.12 dB		
(80 to 90) dB	0.14 dB		
(90 to 100) dB	0.36 dB		
(100 to 110) dB	0.82 dB		

Electrical – RF/Microwave

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Attenuation – Measure <sup>1</sup>	(26 500 to 31 150) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.12 dB	
	(80 to 90) dB	0.13 dB	
	(90 to 100) dB	0.33 dB	
	(100 to 110) dB	0.77 dB	
	(31 150 to 41 000) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
	(40 to 50) dB	0.04 dB	
	(50 to 60) dB	0.076 dB	
	(60 to 70) dB	0.081 dB	
	(70 to 80) dB	0.14 dB	
	(80 to 90) dB	0.36 dB	
	(41 000 to 45 000) MHz		
	(0 to 10) dB	0.02 dB	
	(10 to 20) dB	0.025 dB	
	(20 to 30) dB	0.03 dB	
	(30 to 40) dB	0.035 dB	
(40 to 50) dB	0.04 dB		
(50 to 60) dB	0.076 dB		
(60 to 70) dB	0.11 dB		
(70 to 80) dB	0.24 dB		
(80 to 90) dB	0.6 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Attenuation – Measure <sup>1</sup>	(45 000 to 50 000) MHz (0 to 10) dB (10 to 20) dB (20 to 30) dB (30 to 40) dB (40 to 50) dB (50 to 60) dB (60 to 70) dB (70 to 80) dB (80 to 90) dB	0.02 dB 0.025 dB 0.03 dB 0.035 dB 0.04 dB 0.11 dB 0.29 dB 0.7 dB 1.4 dB	Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor
Amplitude Modulation Measure <sup>1</sup>	100 kHz to 10 MHz Rate 50 Hz to 10 kHz (5 to 99) % Depth 10 MHz to 3 GHz Rate 50 Hz to 100 kHz (5 to 20) % Depth (20 to 99) % Depth (3 to 26.5) GHz Rate 50 Hz to 100 kHz (5 to 20) % Depth (20 to 99) % Depth (26.5 to 31.15) GHz Rate 50 Hz to 100 kHz (5 to 20) % Depth (20 to 99) % Depth (31.15 to 50) GHz Rate 50 Hz to 100 kHz (5 to 20) % Depth (20 to 99) % Depth	0.75 % of reading + 0.3 digits 2.5 % of reading + 0.4 digits 1.5 % of reading + 0.4 digits 4.5 % of reading + 0.4 digits 1.5 % of reading + 0.4 digits 6.8 % of reading + 0.4 digits 1.9 % of reading + 0.4 digits 2.6 % of reading + 0.4 digits 6 % of reading + 0.4 digits	Agilent N5531S Measuring Receiver



ANSI National Accreditation Board

Electrical – RF/Microwave

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency Modulation Measure <sup>1</sup>  $\beta = \text{deviation} / \text{rate}$	250 kHz to 10 MHz Rates 20 Hz to 10 kHz Peak Dev 200 to 40 kHz 10 MHz to 6.6 GHz Rates 50 Hz to 200 kHz Peak Dev 250 to 400 kHz (6.6 to 13.2) GHz Rates 50 Hz to 200 kHz Peak Dev 250 to 400 kHz (13.2 to 31.15) GHz Rates 50 Hz to 200 kHz Peak Dev 250 to 400 kHz (31.15 to 50) GHz Rates 50 Hz to 200 kHz Peak Dev 250 to 400 kHz	$\beta > 0.2 - 1.5 \% \text{ of reading} + 2 \text{ Hz}$ $\beta > 1.2 - 1 \% \text{ of reading} + 2 \text{ Hz}$  $\beta > 0.2 - 1.5 \% \text{ of reading} + 2 \text{ Hz}$ $\beta > 0.45 - 1 \% \text{ of reading} + 2 \text{ Hz}$  $\beta > 0.2 - 2.5 \% \text{ of reading} + 4 \text{ Hz}$ $\beta > 8 - 1 \% \text{ of reading} + 4 \text{ Hz}$  $\beta > 0.2 - 3.8 \% \text{ of reading} + 9 \text{ Hz}$ $\beta > 16 - 1 \% \text{ of reading} + 9 \text{ Hz}$  $\beta > 0.2 - 8.5 \% \text{ of reading} + 17 \text{ Hz}$ $\beta > 16 - 1 \% \text{ of reading} + 17 \text{ Hz}$	Agilent N5531S Measuring Receiver
Phase Modulation – Measure <sup>1</sup>	100 kHz to 6.6 GHz Deviations > 0.3 rad Deviations > 0.7 rad (6.6 to 13.2) GHz Deviations > 0.6 rad Deviations > 2.0 rad (13.2 to 26.5) GHz Deviations: > 1.2 rad Deviations > 4.0 rad (26.5 to 31.15) GHz Deviations: > 1.3 rad Deviations > 4.0 rad (31.15 to 50) GHz Deviations: > 2.4 rad Deviations > 8.0 rad	3 % of reading + 0.002 rad 1 % of reading + 0.002 rad  3 % of reading + 0.005 rad 1 % of reading + 0.005 rad  3 % of reading + 0.009 rad 1 % of reading + 0.009 rad  3 % of reading + 0.009 rad 1 % of reading + 0.009 rad  3 % of reading + 0.018 rad 1 % of reading + 0.018 rad	Agilent N5531S Measuring Receiver
RF Power Generate <sup>1</sup>	10 to 30 MHz (0 to -58) dB (-58 to -78) dB (-78 to -110) dB (-110 to -120) dB	0.23 dB 0.25 dB 0.27 dB 0.43 dB	Agilent N5531S Measuring Receiver N1912A w/E9304A Power Sensor, 83650B Signal Generator

Electrical – RF/Microwave

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power Generate <sup>1</sup>	(30 to 2 000) MHz		Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor, 83650B Signal Generator
	(0 to -58) dB	0.3 dB	
	(-58 to -78) dB	0.31 dB	
	(-78 to -110) dB	0.32 dB	
	(2 000 to 3 050) MHz		
	(0 to -58) dB	0.35 dB	
	(-58 to -78) dB	0.36 dB	
	(-78 to -110) dB	0.37 dB	
	(3 050 to 6 600) MHz		
	(0 to -58) dB	0.35 dB	
	(-58 to -78) dB	0.36 dB	
	(-78 to -110) dB	0.39 dB	
	(6 600 to 13 200) MHz		
	(0 to -58) dB	0.35 dB	
	(-58 to -78) dB	0.36 dB	
	(-78 to -110) dB	0.42 dB	
	(13 200 to 18 000) MHz		
	(0 to -58) dB	0.35 dB	
	(-58 to -78) dB	0.36 dB	
	(-78 to -110) dB	0.46 dB	
	(18 000 to 19 200) MHz		
	(0 to -58) dB	0.41 dB	
	(-58 to -78) dB	0.42 dB	
	(-78 to -110) dB	0.5 dB	
(19 200 to 26 500) MHz			
(0 to -58) dB	0.41 dB		
(-58 to -78) dB	0.42 dB		
(-78 to -110) dB	0.9 dB		
(26 500 to 31 150) MHz			
(-3 to -58) dB	0.63 dB		
(-58 to -78) dB	0.64 dB		
(-78 to -110) dB	0.96 dB		
(31 150 to 41 000) MHz			
(-6 to -58) dB	0.83 dB		
(-58 to -78) dB	0.84 dB		
(-78 to -100) dB	1.1 dB		
(41 000 to 45 000) MHz			
(-6 to -58) dB	0.83 dB		
(-58 to -78) dB	0.85 dB		
(-78 to -100) dB	1.4 dB		

**Electrical – RF/Microwave**

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power Generate <sup>1</sup>	(45 000 to 50 000) MHz (-6 to -58) dB (-58 to -78) dB (-78 to -90) dB	0.83 dB 1 dB 1.5 dB	Agilent N5531S Measuring Receiver w/N5532A Opt 550 Power Sensor, 83650B Signal Generator
AM Distortion Measure <sup>1</sup> Rate 20 Hz to 1 kHz	(0.1 to 10) MHz AM Depth > 1 % (0 to -20) dB (-20 to -30) dB AM Depth > 3 % (0 to -20) dB (-20 to -30) dB (-30 to -40) dB 10 MHz to 26.5 GHz AM Depth > 1 % (0 to -20) dB (-20 to -30) dB AM Depth > 3% (0 to -20) dB (-20 to -30) dB (-30 to -40) dB (26.5 MHz to 50.0 GHz) AM Depth > 3 % (0 to -20) dB AM Depth > 5% (0 to -20) dB (-20 to -30) dB	1.2 dB 2.2 dB 1 dB 1.3 dB 2.4 dB 1.3 dB 2.5 dB 1.1 dB 1.4 dB 3 dB 1.8 dB 1.5 dB 3.5 dB	Agilent N5531S Measuring Receiver



ANSI National Accreditation Board

Electrical – RF/Microwave

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment	
FM Distortion Measure <sup>1</sup> Rate 20 Hz to 1 kHz	(1 to 6 600) MHz		Agilent N5531S Measuring Receiver	
	Dev 500 Hz to 2 kHz			
	(0 to -20) dB	0.26 dB		
	(-20 to -30) dB	0.79 dB		
	(-30 to -40) dB	2.3 dB		
	Dev > 2 kHz			
	(0 to -20) dB	0.09 dB		
	(-20 to -30) dB	0.27 dB		
	(-30 to -40) dB	0.83 dB		
	(-40 to -50) dB	2.4 dB		
	(6.6 to 13.2) GHz			
	Dev > 2.3 kHz			
	(0 to -20) dB	0.26 dB		
	(-20 to -30) dB	0.79 dB		
	(-30 to -40) dB	2.3 dB		
	Dev > 4.5 kHz			
	(0 to -20) dB	0.09 dB		
	(-20 to -30) dB	0.27 dB		
	(-30 to -40) dB	0.83 dB		
	(-40 to -50) dB	2.4 dB		
	(13.2 to 31.15) GHz			
	Dev > 2.7 kHz			
	(0 to -20) dB	0.26 dB		
	(-20 to -30) dB	0.79 dB		
(-30 to -40) dB	2.3 dB			
Dev > 6.0 kHz				
(0 to -20) dB	0.09 dB			
(-20 to -30) dB	0.27 dB			
(-30 to -40) dB	0.83 dB			
(-40 to -50) dB	2.4 dB			
(31.15 to 50.0) GHz				
Dev > 4 kHz				
(0 to -20) dB	0.26 dB			
(-20 to -30) dB	0.79 dB			
(-30 to -40) dB	2.3 dB			
Dev > 12.0 kHz				
(0 to -20) dB	0.09 dB			
(-20 to -30) dB	0.27 dB			
(-30 to -40) dB	0.83 dB			
(-40 to -50) dB	2.4 dB			

Electrical – RF/Microwave

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
PM Distortion Measure <sup>1</sup>	(1 to 6 600) MHz		
	Rate (20 to 500) Hz		
	Dev > 0.8 rad		
	(0 to -20) dB	0.26 dB	
	(-20 to -30) dB	0.79 dB	
	(-30 to -40) dB	2.3 dB	
	Dev > 2.5 rad		
	(0 to -20) dB	0.09 dB	
	(-20 to -30) dB	0.27 dB	
	(-30 to -40) dB	0.83 dB	
	(-40 to -50) dB	2.3 dB	
	Rate (500 to 1 000) Hz		
	Dev > 0.4 rad	0.26 dB	
	(0 to -20) dB	0.79 dB	
	(-20 to -30) dB	2.3 dB	
	(-30 to -40) dB	0.09 dB	
	Dev > 1.0 rad		
	(0 to -20) dB	0.27 dB	
	(-20 to -30) dB	0.83 dB	
	(-30 to -40) dB	2.3 dB	
	(-40 to -50) dB	2.3 dB	
(6.6 to 13.2) GHz			
Rate (20 to 500) Hz			
Dev > 1.8 rad			
(0 to -20) dB	0.26 dB		
(-20 to -30) dB	0.79 dB		
(-30 to -40) dB	2.3 dB		
Dev > 5.5 rad			
(0 to -20) dB	0.09 dB		
(-20 to -30) dB	0.27 dB		
(-30 to -40) dB	0.83 dB		
(-40 to -50) dB	2.4 dB		

Agilent N5531S  
Measuring Receiver

Electrical – RF/Microwave

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
PM Distortion Measure <sup>1</sup>	(6.6 to 13.2) GHz		
	Rate (500 to 1 000) Hz		
	Dev > 0.8 rad		
	(0 to -20) dB	0.26 dB	
	(-20 to -30) dB	0.79 dB	
	(-30 to -40) dB	2.3 dB	
	Dev > 2.5 rad		
	(0 to -20) dB	0.09 dB	
	(-20 to -30) dB	0.27 dB	
	(-30 to -40) dB	0.83 dB	
	(-40 to -50) dB	2.3 dB	
	(13.2 to 31.15) GHz		
	Rate (20 to 500) Hz		
	Dev > 3.5 rad		
	(0 to -20) dB	0.26 dB	
	(-20 to -30) dB	0.79 dB	
	(-30 to -40) dB	2.3 dB	
	Dev > 10.0 rad		
	(0 to -20) dB	0.09 dB	
	(-20 to -30) dB	0.27 dB	
(-30 to -40) dB	0.83 dB		
(-40 to -50) dB	2.3 dB		
Rate (500 to 1 000) Hz			
Dev > 3.0 rad			
(0 to -20) dB	0.26 dB		
(-20 to -30) dB	0.79 dB		
(-30 to -40) dB	2.3 dB		
Dev > 8.0 rad			
(0 to -20) dB	0.09 dB		
(-20 to -30) dB	0.27 dB		
(-30 to -40) dB	0.83 dB		
(-40 to -50) dB	2.3 dB		
			Agilent N5531S Measuring Receiver

**Electrical – RF/Microwave**

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Total Harmonic Distortion (THD)	(0 to -60) dB		HP 8903B Audio Analyzer
	20 Hz to 20 kHz		
	(0 to -40) dB	1 dB	
	(-40 to -50) dB	1 dB	
	(-50 to -60) dB	1.3 dB	
	(-60 to -65) dB	1.7 dB	
	(20 to 50) kHz		
	(0 to -40) dB	2 dB	
	(-40 to -50) dB	2.1 dB	
	(-50 to -60) dB	3 dB	
Harmonics Measure <sup>1</sup>	(50 to 100) kHz		Agilent E4448A Measuring Receiver
	(0 to -40) dB	2 dB	
	(-40 to -50) dB	2.4 dB	
	(-80 to -10) dB		
	2 <sup>nd</sup> through 5 <sup>th</sup> Harmonic		
	1kHz to 600MHz	0.37 dB	
	(600 to 1 320) MHz	1.1 dB	
	(1 320 to 2 200) MHz	1.4 dB	
	(2 200 to 3 000) MHz	1.4 dB	
	(3 000 to 4 400) MHz	1.7 dB	
(4 400 to 5 300) MHz	1.9 dB		
(5 300 to 10 000) MHz	2.1 dB		
2 <sup>nd</sup> through 4 <sup>th</sup> Harmonic			
(10 000 to 12 500) MHz	2.1 dB		
2 <sup>nd</sup> through 3 <sup>rd</sup> Harmonic			
(12 500 to 16 667) MHz	2.1 dB		
2 <sup>nd</sup> Harmonic			
(16 667 to 25 000) MHz	2.3 dB		

**Length – Dimensional Metrology**

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Gage Blocks <sup>2</sup>	Up to 13 in	(4 + 1.5L) µin	Master gage blocks P&W Universal Measuring Machine
Micrometers <sup>1,2</sup>	Up to 40 in	(28 + 4.4L) µin	Gage Blocks

**Length – Dimensional Metrology**

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Bore Micrometers <sup>2</sup> 2 point 3 point	Up to 12 in Up to 3 in	$(57 + 2.8L) \mu\text{in}$ $(52 + 14L) \mu\text{in}$	Master gage blocks, P&W Labmaster Universal, Master Ring
Calipers <sup>1,2</sup>	Up to 40 in	$(280 + 1.7L) \mu\text{in}$	Gage Blocks
Dial Indicators <sup>1,2</sup> Resolution $\geq 50\mu\text{in}$ Resolution $< 50\mu\text{in}$	Up to 10 in Up to 0.1 in	$(26 + 3.8L) \mu\text{in}$ 8.2 $\mu\text{in}$	Gage Blocks
Height Gages <sup>1,2</sup>	Up to 40 in	$(96 + 2.8L) \mu\text{in}$	Gage Blocks
Rulers <sup>1</sup>	Up to 46 in	0.009 1 in	Gage Blocks
Tape Measures <sup>2</sup>	Up to 100 ft	$(0.000 26F + 0.025) \text{ in}$	Standard rule
Protractors <sup>1</sup>	(0 to 360) <sup>o</sup>	0.013 <sup>o</sup>	Angle Blocks
Feeler Gage	Up to 1 in	31 $\mu\text{in}$	Pratt & Whitney Supermicrometer C
Cylindrical Gages <sup>2</sup>  Plain Pins, Plugs  Plain Rings	  (0 to 13) in  (0.04 to 14) in	  $(6.8 + 3.3D) \mu\text{in}$  $(11 + 3.2D) \mu\text{in}$	  Master gage blocks P&W Universal Measuring Machine
Solid Thread Rings Pitch Diameter	Up to 12 in	120 $\mu\text{in}$	Pratt & Whitney Labmaster Measuring Machine
Thread Plugs Major Diameter Pitch Diameter	Up to 10 in Up to 10 in	53 $\mu\text{in}$ 98 $\mu\text{in}$	Thread Wires Gage Blocks Pratt & Whitney Supermicrometer C
Thread Rings <sup>2</sup>  Pitch Diameter	  Up to 1 in	  $(350 + 47D) \mu\text{in}$	  Thread Setting Plugs Tactile Fit
Thread Wires	Up to 0.5 in	11 $\mu\text{in}$	Master gage blocks, P&W universal measuring machine
Surface Plates <sup>1</sup> Overall Flatness Local Area Flatness	(18 x 18) in to (6 x 6) ft (-0.001 to 0.001) in	86 $\mu\text{in}$ 72 $\mu\text{in}$	Planekator Repeat-o-meter



ANSI National Accreditation Board

**Length – Dimensional Metrology**

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Optical Comparators <sup>1</sup> – Angle Linearity	(0 to 360) ° Up to 20 in (20 to 40) in	0.016 ° 320 μin 630 μin	Gage blocks, Angle blocks,
Magnification	10x to 100x	430 μin	SI Industries glass scales

**Mass and Mass Related**

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Pressure <sup>1</sup>	(-15 to <0) psi (0 to 100) psi	0.043 psi (0.003 6 psi or 0.12 %RD)* *Whichever is Greater	ADT783 Pressure Controller w/module ADT151-01RD-CP100M
Pressure <sup>1</sup>	(-15 to <0) psi (0 to 300) psi	0.043 psi (0.011 psi or 0.12 %RD)* *Whichever is Greater	ADT783 Pressure Controller w/module ADT151-01RD-CP300M
Pressure <sup>1</sup>	(-15 to <0) psi (0 to 1 000) psi	0.045 psi (0.036 psi or 0.12 %RD)* *Whichever is Greater	ADT783 Pressure Controller w/module ADT151-01RD-CP1KM
Pressure <sup>1</sup>	(-15 to <0) psi (0 to 3 600) psi	0.045 psi (0.14 psi or 0.13 %RD)* *Whichever is Greater	ADT783 Pressure Controller w/module ADT151-01RD-CP3.6KM
Pressure <sup>1</sup>	(0 to 6 000) psi (0 to 15 000) psi	0.72 psi 2 psi	ADT783 Pressure Controller w/module ADT151-01-GP15K
Pipettes <sup>1,2</sup>	(10 to 100) μL (100 to 1 000) μL (1 to 10) mL	(0.58 + 0.004 <i>V</i> ) μL (0.6 + 0.001 <i>V</i> ) μL (2.6 + 0.00 12 <i>V</i> ) μL	Mass Balance
Torque Tools <sup>1</sup>	(10 to 100) ozf-in 4 lbf-in to 1 000 lbf-ft	0.59 % of reading 0.32 % of reading	Torque Testers CDI 1001-O-DDT CDI 5000 ST
Torque Analyzers <sup>1</sup>	Up to 100 ozf-in (4 to 150) lbf-in (12.5 to 1 000) lbf-ft	0.1 % of reading 0.064 % of reading 0.036 % of reading	Torque Arms and Class F Weights
Force <sup>1</sup> Tension and Compression	(0.5 to 500) lbf	0.02% of reading	Class F Weights
Scales and Balances <sup>1,2,4</sup>	1 g to 11 kg (0.001 to 21) lb	(0.16 + 0.003 <i>X</i> ) mg (0.000 003 1 <i>W</i> ) lb	Class 1 Weights

**Mass and Mass Related**

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Scales and Balances <sup>1,2,4</sup>	1 g to 11 kg (0.001 to 531) lb	(0.75 + 0.12 <i>X</i> ) mg (0.000 12 <i>W</i> ) lb	Class 6 Weights
Mass - Fixed Points Metric	(1, 2, 5) g 10 g 20g 50 g 100 g 200 g 500 g (1, 2, 5) kg	0.14 mg 0.19 mg 0.18 mg 0.25 mg 0.45 mg 0.65 mg 29 mg 31 mg	Comparison to ASTM E617 Class 1 weights
	(0.001, 0.002) lb (0.005, 0.01, 0.02) lb 0.05 lb 0.1 lb 0.2 lb (0.5, 1, 2) lb (5, 10) lb 50 lb	0.13 mg 0.14 mg 0.2 mg 0.3 mg 0.34 mg 27 mg 28 mg 230 mg	
Rockwell Hardness Testers <sup>1</sup>	HRBW < 60 (≥ 60 to < 80) ≥ 80 HRC < 35 (≥ 35 to < 60) ≥ 60	0.8 HRBW 0.85 HRBW 0.69 HRBW 0.58 HRC 0.54 HRC 0.39 HRC	Indirect verification per ASTM E18

**Thermodynamic**

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Relative Humidity Generate	(10 to 95) %RH	0.5 %RH	Thunder Scientific 2500 Humidity Chamber
Relative Humidity Measure <sup>1</sup>	(0 to 90) %RH (90 to 100) %RH	1.2 %RH 2 %RH	Vaisala MI70/HMP76 Humidity Indicator and Probe
Temperature Measuring Equipment <sup>1</sup>	(-40 to 160) °C (33 to 700) °C	0.13 °C (0.21 + 0.00042 <i>T</i> ) °C	Dry Well Calibrators



ANSI National Accreditation Board

**Thermodynamic**

Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Temperature Measure <sup>1</sup>	(-200 to 200) °C (200 to 400) °C (400 to 660) °C	0.015 °C 0.023 °C 0.035 °C	Secondary PRT w/ Additel Reference Thermometer Readout
Infrared (IR) Thermometers <sup>1</sup>	(20 to 100) °C (100 to 300) °C (300 to 420) °C (420 to 500) °C	1.5 °C 4.3 °C 6 °C 7.7 °C	Fluke 9132 IR Calibrator $\varepsilon = 0.95, \lambda = (8 \text{ to } 14) \mu\text{m}$

**Time and Frequency**


Morrisville (Raleigh), NC

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency – Generate <sup>1</sup>	10 MHz	$1 \times 10^{-12}$ Hz/Hz	HP Z3805A GPS Receiver
Frequency – Generate <sup>1</sup>	(1 to 10) Hz (10 to 100) Hz (100 to 1 000) Hz (1 to 10) kHz (10 to 100) kHz (0.1 to 1) MHz (1 to 10) MHz	$1 \times 10^{-12}$ Hz/Hz + 0.57 $\mu$ Hz $1 \times 10^{-12}$ Hz/Hz + 5.7 $\mu$ Hz $1 \times 10^{-12}$ Hz/Hz + 57 $\mu$ Hz $1 \times 10^{-12}$ Hz/Hz + 0.57 mHz $1 \times 10^{-12}$ Hz/Hz + 5.7 mHz $1 \times 10^{-12}$ Hz/Hz + 57 mHz $1 \times 10^{-12}$ Hz/Hz + 0.57 Hz	Agilent 33250A Function Generator / HP Z3805A GPS Receiver
Frequency – Generate <sup>1</sup>	(10 to 50 000) MHz	$1 \times 10^{-12}$ Hz/Hz + 0.57 mHz	Agilent 83650B Signal Generator / HP Z3805A GPS Receiver
Time – Generate	1 pps	$1 \times 10^{-12}$ s/s + 750 ps	HP Z3805A GPS Receiver
Frequency – Measure <sup>1</sup>	(1 to 10) Hz (10 to 100) Hz (100 to 1 000) Hz (1 to 10) kHz (10 to 100) kHz (100 to 200) kHz (0.2 to 3 000) MHz	$4.2 \times 10^{-9}$ Hz/Hz $1.5 \times 10^{-9}$ Hz/Hz $0.6 \times 10^{-9}$ Hz/Hz $0.33 \times 10^{-9}$ Hz/Hz $0.24 \times 10^{-9}$ Hz/Hz $0.21 \times 10^{-9}$ Hz/Hz $0.21 \times 10^{-9}$ Hz/Hz	Agilent 53131A Opt 030 Frequency Counter / HP Z3805A GPS Receiver
Frequency – Measure <sup>1</sup>	(10 to 50 000) MHz	$1 \times 10^{-12}$ Hz/Hz + 0.1 Hz	Agilent E4448A Spectrum Analyzer / HP Z3805A GPS Receiver
Timer, Stopwatch <sup>1</sup>	10 s to 24 hr	34 ms	Totalize method with counter
Tachometers – RPM <sup>1</sup>	Up to 100 000 RPM	0.001 % of reading + 0.6R	HP 33250A Signal Generator & LED

[Return to Site listing](#) (top)

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ( $k=2$ ), corresponding to a confidence level of approximately 95%.

- Notes:
1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
  2.  $D$  = diameter in inches,  $DL$  = diagonal length in inches,  $L$  = length in inches (or millimeters as indicated),  $F$  = length in feet,  $R$  = resolution of device under test,  $t$  = time in seconds,  $V$  = Volume in microliters,  $X$  = weight in grams,  $W$  = weight in pounds, 1 mil = 0.001 inch.
  3. The tactile fit of an adjustable thread ring to a thread-setting plug is not a measurement of pitch diameter. The uncertainty for this pitch diameter setting is based on the contributors associated with the thread setting plug and environmental contributors only.
  4. The CMC for scales and balances is highly dependent upon the resolution of the unit under test. The uncertainties presented here do not include the resolution of the unit under test. The resolution will be included in the reported measurement uncertainty at the time of calibration.
  5. The nominal values listed are approximate.
  6. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-2080.



Jason Stine, Vice President

